

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502

Continuing Calib Date: 03/12/2025 **Initial Calibration Date(s):** 03/12/2025 03/12/2025

Continuing Calib Time: 11:48 **Initial Calibration Time(s):** 10:41 11:19

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	5.96	5.96	5.86	6.06	0.00
EDB	2.61	2.61	2.51	2.71	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502

Continuing Calib Date: 03/12/2025 **Initial Calibration Date(s):** 03/12/2025 03/12/2025

Continuing Calib Time: 11:48 **Initial Calibration Time(s):** 10:41 11:19

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	4.99	4.99	4.89	5.09	0.00
EDB	1.98	1.98	1.88	2.08	0.00