

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/12/2025 03/12/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: M8011.504.1 0.1 PF Data File : PQ070045.D Time Analyzed: 13:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.960	5.861	6.061	0.100	0.100	0.0
EDB	2.609	2.510	2.710	0.100	0.100	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/12/2025 03/12/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: M8011.504.1 0.1 PF Data File : PQ070045.D Time Analyzed: 13:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	4.995	4.894	5.094	0.100	0.100	0.0
EDB	1.979	1.879	2.079	0.100	0.100	0.0