284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q1502

Project ID : NJ Waste Water PT

Client : Alliance Technical Group, LLC - Newark

Lab Sample Number	Client Sample Number
Q1502-01	PT-VOA-WP
Q1502-02	PT-VOA-WP
Q1502-03	PT-BN-WP
Q1502-04	PT-BN-WP
Q1502-05	PT-BN-WP
Q1502-06	PT-ACIDS-WP
Q1502-07	PT-ACIDS-WP
Q1502-08	PT-ACIDS-WP
Q1502-09	PT-PEST-WP
Q1502-10	PT-PEST-WP
Q1502-11	PT-CHLR-WP
Q1502-12	PT-CHLR-WP
Q1502-13	PT-TXP-WP
Q1502-14	PT-TXP-WP
Q1502-15	PT-PCBW-WP
Q1502-16	PT-PCBW-WP
Q1502-17	PT-HERB-WP
Q1502-18	RR-GAS-WP
Q1502-19	RR-DIES-WP
Q1502-20	RR-8011-WP
Q1502-21	RR-PAH-WP
Q1502-22	RR-TRIAZINE-WP

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 4/1/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Alliance Technical Group, LLC - Newark
Project Name: NJ Waste Water PT
Project # N/A
Chemtech Project # Q1502
Test Name: SVOCMS Group2

A. Number of Samples and Date of Receipt:
21 Water samples were received on 03/05/2025.
1 Water sample was received on 03/11/2025.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: Diesel Range Organics, Gasoline Range Organics, Herbicide group1, PCB, PESTICIDE Group1, PESTICIDE Group2, PESTICIDE Group3, SVOCMS Group1, SVOCMS Group2, SVOCMS Group3, SVOCMS Group4, SVOCMS Group5, SVOCMS Group6, VOCGC Group 1 and VOCMS Group1. This data package contains results for SVOCMS Group2.

C. Analytical Techniques:
The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df The samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOCMS Group2 was based on method 625.1,8270E and extraction was done based on method 3510.

D. QA/ QC Samples:
The Holding Times were met for all analysis.
The Surrogate recoveries met the acceptable criteria.
The Internal Standards Areas met the acceptable requirements.
The Retention Times were acceptable for all samples.
The RPD met criteria .
The Blank Spike met requirements for all samples .
The Blank Spike Duplicate met requirements for all samples .
The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (8270-BF031025.M) for 2,4-Dinitrophenol and this compound is passing on Linear regression.

The Continuous Calibration met the requirements .
The Tuning criteria met requirements.

Sample PT-ACIDS-WP was diluted due to high concentration.

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:
The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.
Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:
Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value**.** This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1502 MATRIX: Water

METHOD: 625.1,8270E/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			✓
	The % RSD is greater than 20% in the Initial Calibration (8270-BF031025.M) for 2,4-Dinitrophenol and this compound is passing on Linear regression. The Continuous Calibration met the requirements .			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples .			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

		NA	NO	YES
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

Sample PT-ACIDS-WP was diluted due to high concentration.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

_____ _____

QA REVIEW Date

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: **Q1502**

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: **MOHAMMAD AHMED** **Date:** **04/01/2025**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1502	**OrderDate:**	3/6/2025 10:04:07 AM
Client:	Alliance Technical Group, LLC - Newark	**Project:**	NJ Waste Water PT
Contact:	Mohammad Ahmed	**Location:**	QA Office,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1502-03	**PT-BN-WP**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group1	8270E		03/12/25	03/13/25	
Q1502-03DL	**PT-BN-WPDL**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group1	8270E		03/12/25	03/13/25	
Q1502-05	**PT-BN-WP**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group3	8270-Modified		03/13/25	03/14/25	
Q1502-05DL	**PT-BN-WPDL**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group3	8270-Modified		03/13/25	03/14/25	
Q1502-06	**PT-ACIDS-WP**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group2	8270E		03/12/25	03/13/25	
Q1502-06DL	**PT-ACIDS-WPDL**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group2	8270E		03/12/25	03/13/25	
Q1502-08	**PT-ACIDS-WP**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group4	8270-Modified		03/13/25	03/14/25	
Q1502-08DL	**PT-ACIDS-WPDL**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group4	8270-Modified		03/13/25	03/14/25	
Q1502-21	**RR-PAH-WP**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group5	8270-Modified		03/13/25	03/14/25	
Q1502-21DL	**RR-PAH-WPDL**	**Water**			**03/03/25**			**03/05/25**
			SVOCMS Group5	8270-Modified		03/13/25	03/14/25	
Q1502-22	**RR-TRIAZINE-WP**	**Water**			**03/07/25**			**03/11/25**
			SVOCMS Group1			03/12/25	03/13/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Client ID :	**PT-ACIDS-WP**							
Q1502-06	PT-ACIDS-WP	WATER	Phenol	140.000	E	0.91	5	ug/L
Q1502-06	PT-ACIDS-WP	WATER	2-Chlorophenol	39.200		0.58	5	ug/L
Q1502-06	PT-ACIDS-WP	WATER	2-Methylphenol	180.000	E	1.1	5	ug/L
Q1502-06	PT-ACIDS-WP	WATER	3+4-Methylphenols	110.000	E	1.1	10	ug/L
Q1502-06	PT-ACIDS-WP	WATER	2-Nitrophenol	180.000	E	1.8	5	ug/L
Q1502-06	PT-ACIDS-WP	WATER	2,4-Dimethylphenol	200.000	E	1.9	5	ug/L
Q1502-06	PT-ACIDS-WP	WATER	2,4-Dichlorophenol	68.800		0.52	5	ug/L
Q1502-06	PT-ACIDS-WP	WATER	4-Chloro-3-methylphenol	63.200		0.59	5	ug/L
Q1502-06	PT-ACIDS-WP	WATER	2,4,6-Trichlorophenol	130.000	E	0.51	5	ug/L
Q1502-06	PT-ACIDS-WP	WATER	2,4-Dinitrophenol	200.000	E	6	10	ug/L
Q1502-06	PT-ACIDS-WP	WATER	4-Nitrophenol	140.000	E	2.4	10	ug/L
Q1502-06	PT-ACIDS-WP	WATER	4,6-Dinitro-2-methylphenol	96.900	E	2.9	10	ug/L
Q1502-06	PT-ACIDS-WP	WATER	Pentachlorophenol	82.500	E	1.6	10	ug/L
Q1502-06	PT-ACIDS-WP	WATER	2,3,4,6-Tetrachlorophenol	57.400		0.72	5	ug/L
			Total Svoc :	**1,688.00**				
			Total Concentration:	**1,688.00**				
Client ID :	**PT-ACIDS-WPDL**							
Q1502-06DL	PT-ACIDS-WPDL	WATER	Phenol	150.000	D	4.6	25	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	2-Chlorophenol	39.600	D	2.9	25	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	2-Methylphenol	190.000	D	5.6	25	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	3+4-Methylphenols	110.000	D	5.5	50	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	2-Nitrophenol	170.000	D	8.8	25	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	2,4-Dimethylphenol	210.000	D	9.3	25	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	2,4-Dichlorophenol	67.000	D	2.6	25	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	4-Chloro-3-methylphenol	61.500	D	3	25	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	2,4,6-Trichlorophenol	130.000	D	2.6	25	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	2,4-Dinitrophenol	160.000	D	29.9	50	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	4-Nitrophenol	130.000	D	11.9	50	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	4,6-Dinitro-2-methylphenol	75.600	D	14.4	50	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	Pentachlorophenol	73.800	D	7.9	50	ug/L
Q1502-06DL	PT-ACIDS-WPDL	WATER	2,3,4,6-Tetrachlorophenol	54.800	D	3.6	25	ug/L
			Total Svoc :	**1,622.30**				
			Total Concentration:	**1,622.30**				



QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%) Low	Limits (%) High
PB167097BL	PB167097BL	2-Fluorophenol	150	134	90		10	139
		Phenol-d6	150	130	86		10	134
		2,4,6-Tribromophenol	150	154	102		44	137
PB167097BS	PB167097BS	2-Fluorophenol	150	140	93		10	139
		Phenol-d6	150	137	91		10	134
		2,4,6-Tribromophenol	150	160	107		44	137
PB167097BSD	PB167097BSD	2-Fluorophenol	150	132	88		10	139
		Phenol-d6	150	129	86		10	134
		2,4,6-Tribromophenol	150	151	101		44	137
Q1502-06	PT-ACIDS-WP	2-Fluorophenol	150	145	97		10	139
		Phenol-d6	150	142	95		10	134
		2,4,6-Tribromophenol	150	151	101		44	137
Q1502-06DL	PT-ACIDS-WPDL	2-Fluorophenol	150	149	99		10	139
		Phenol-d6	150	148	99		10	134
		2,4,6-Tribromophenol	150	144	96		44	137

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E **DataFile:** BF141941.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB167097BS	Phenol	50	45.8	ug/L	92				66	118	
	2-Chlorophenol	50	46.9	ug/L	94				70	117	
	2-Methylphenol	50	46.7	ug/L	93				69	109	
	3+4-Methylphenols	50	46.3	ug/L	93				67	106	
	2-Nitrophenol	50	47.5	ug/L	95				70	115	
	2,4-Dimethylphenol	50	57.6	ug/L	115				42	142	
	2,4-Dichlorophenol	50	46.1	ug/L	92				66	115	
	Benzoic acid	50	52.7	ug/L	105				10	125	
	4-Chloro-3-methylphenol	50	45.9	ug/L	92				65	114	
	2,4,6-Trichlorophenol	50	49.2	ug/L	98				61	110	
	2,4,5-Trichlorophenol	50	45.4	ug/L	91				70	106	
	2,4-Dinitrophenol	100	99.2	ug/L	99				36	166	
	4-Nitrophenol	100	100	ug/L	100				45	147	
	4,6-Dinitro-2-methylphenol	50	51.0	ug/L	102				62	132	
	Pentachlorophenol	100	99.9	ug/L	100				47	114	
	2,3,4,6-Tetrachlorophenol	50	46.3	ug/L	93				63	116	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E **DataFile:** BF141942.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB167097BSD	Phenol	50	43.0	ug/L	86	6			66	118	20
	2-Chlorophenol	50	44.2	ug/L	88	6			70	117	20
	2-Methylphenol	50	44.5	ug/L	89	5			69	109	20
	3+4-Methylphenols	50	44.0	ug/L	88	5			67	106	20
	2-Nitrophenol	50	45.9	ug/L	92	3			70	115	20
	2,4-Dimethylphenol	50	54.8	ug/L	110	5			42	142	20
	2,4-Dichlorophenol	50	43.8	ug/L	88	5			66	115	20
	Benzoic acid	50	52.1	ug/L	104	1			10	125	20
	4-Chloro-3-methylphenol	50	43.1	ug/L	86	6			65	114	20
	2,4,6-Trichlorophenol	50	45.1	ug/L	90	9			61	110	20
	2,4,5-Trichlorophenol	50	45.7	ug/L	91	1			70	106	20
	2,4-Dinitrophenol	100	95.0	ug/L	95	4			36	166	20
	4-Nitrophenol	100	97.3	ug/L	97	3			45	147	20
	4,6-Dinitro-2-methylphenol	50	49.9	ug/L	100	2			62	132	20
	Pentachlorophenol	100	95.6	ug/L	96	4			47	114	20
	2,3,4,6-Tetrachlorophenol	50	43.9	ug/L	88	5			63	116	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167097BL

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Lab File ID: BF141940.D Lab Sample ID: PB167097BL

Instrument ID: BNA_F Date Extracted: 03/12/2025

Matrix: (soil/water) Water Date Analyzed: 03/13/2025

Level: (low/med) LOW Time Analyzed: 12:47

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167097BS	PB167097BS	BF141941.D	03/13/2025
PB167097BSD	PB167097BSD	BF141942.D	03/13/2025
PT-ACIDS-WP	Q1502-06	BP024175.D	03/13/2025

COMMENTS:

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1502 SDG NO.: Q1502

Lab File ID: BF141896.D DFTPP Injection Date: 03/10/2025

Instrument ID: BNA_F DFTPP Injection Time: 10:31

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	29.1
68	Less than 2.0% of mass 69	0.5 (1.9) 1
69	Mass 69 relative abundance	32.0
70	Less than 2.0% of mass 69	0.2 (0.7) 1
127	10.0 - 80.0% of mass 198	44.8
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	31.2
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	19.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (19.2) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF141897.D	03/10/2025	11:01
SSTDICC005	SSTDICC005	BF141898.D	03/10/2025	11:30
SSTDICC010	SSTDICC010	BF141899.D	03/10/2025	12:00
SSTDICC020	SSTDICC020	BF141900.D	03/10/2025	12:29
SSTDICCC040	SSTDICCC040	BF141901.D	03/10/2025	12:58
SSTDICC060	SSTDICC060	BF141903.D	03/10/2025	13:57
SSTDICC080	SSTDICC080	BF141904.D	03/10/2025	14:27
SSTDICC050	SSTDICC050	BF141905.D	03/10/2025	15:20

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1502 SDG NO.: Q1502

Lab File ID: BF141938.D DFTPP Injection Date: 03/13/2025

Instrument ID: BNA_F DFTPP Injection Time: 11:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.0
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	34.0
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	29.9
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	16.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141939.D	03/13/2025	12:17
PB167097BL	PB167097BL	BF141940.D	03/13/2025	12:47
PB167097BS	PB167097BS	BF141941.D	03/13/2025	13:17
PB167097BSD	PB167097BSD	BF141942.D	03/13/2025	13:47

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1502 SDG NO.: Q1502

Lab File ID: BP024159.D DFTPP Injection Date: 03/12/2025

Instrument ID: BNA_P DFTPP Injection Time: 15:36

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	28.5
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	33.9
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	46.1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	27.5
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	14.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.4 (19.5) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024160.D	03/12/2025	16:17
SSTDICC005	SSTDICC005	BP024161.D	03/12/2025	16:57
SSTDICC010	SSTDICC010	BP024162.D	03/12/2025	17:38
SSTDICC020	SSTDICC020	BP024163.D	03/12/2025	18:19
SSTDICCC040	SSTDICCC040	BP024164.D	03/12/2025	19:00
SSTDICC050	SSTDICC050	BP024165.D	03/12/2025	19:41
SSTDICC060	SSTDICC060	BP024166.D	03/12/2025	20:21
SSTDICC080	SSTDICC080	BP024167.D	03/12/2025	21:02

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1502 SDG NO.: Q1502

Lab File ID: BP024170.D DFTPP Injection Date: 03/13/2025

Instrument ID: BNA_P DFTPP Injection Time: 09:41

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	29
68	Less than 2.0% of mass 69	0.5 (1.5) 1
69	Mass 69 relative abundance	35.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	46.9
197	Less than 2.0% of mass 198	0.1
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	14.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.7 (19.2) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024171.D	03/13/2025	10:22
PT-ACIDS-WP	Q1502-06	BP024175.D	03/13/2025	13:51
PT-ACIDS-WPDL	Q1502-06DL	BP024176.D	03/13/2025	15:48

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/13/2025

Lab File ID: BF141939.D Time Analyzed: 12:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	198004	6.875	767232	8.16	426931	9.91
	UPPER LIMIT	396008	7.375	1534460	8.657	853862	10.41
	LOWER LIMIT	99002	6.375	383616	7.657	213466	9.41
	EPA SAMPLE NO.						
01	PB167097BL	176773	6.88	705846	8.16	407739	9.91
02	PB167097BS	167779	6.88	666660	8.16	381133	9.92
03	PB167097BSD	178366	6.88	702401	8.16	401749	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/13/2025

Lab File ID: BF141939.D Time Analyzed: 12:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	726402	11.398	451214	14.033	394917	15.504
	UPPER LIMIT	1452800	11.898	902428	14.533	789834	16.004
	LOWER LIMIT	363201	10.898	225607	13.533	197459	15.004
	EPA SAMPLE NO.						
01	PB167097BL	769835	11.39	601564	14.03	427331	15.50
02	PB167097BS	657343	11.40	424407	14.04	370791	15.50
03	PB167097BSD	683407	11.40	437316	14.03	392710	15.50

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/13/2025

Lab File ID: BP024171.D Time Analyzed: 10:22

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	330996	7.764	1338420	10.55	811311	14.41
	UPPER LIMIT	661992	8.264	2676840	11.046	1622620	14.91
	LOWER LIMIT	165498	7.264	669210	10.046	405656	13.91
	EPA SAMPLE NO.						
01	PT-ACIDS-WP	332785	7.74	1357430	10.53	822876	14.42
02	PT-ACIDS-WPDL	333252	7.75	1381030	10.54	845510	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/13/2025

Lab File ID: BP024171.D Time Analyzed: 10:22

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	1560240	17.216	1695900	21.675	1872320	25.063
	UPPER LIMIT	3120480	17.716	3391800	22.175	3744640	25.563
	LOWER LIMIT	780120	16.716	847950	21.175	936160	24.563
	EPA SAMPLE NO.						
01	PT-ACIDS-WP	1600350	17.22	1655510	21.70	1860600	25.12
02	PT-ACIDS-WPDL	1630080	17.21	1669060	21.67	1840330	25.09

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



SAMPLE DATA

Raw Data: **BP024175.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	03/03/25
Project:	NJ Waste Water PT	Date Received:	03/05/25
Client Sample ID:	PT-ACIDS-WP	SDG No.:	Q1502
Lab Sample ID:	Q1502-06	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024175.D	1	03/12/25 08:45	03/13/25 13:51	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
108-95-2	Phenol	140	E	0.91	5.00	ug/L
95-57-8	2-Chlorophenol	39.2		0.58	5.00	ug/L
95-48-7	2-Methylphenol	180	E	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	110	E	1.10	10.0	ug/L
88-75-5	2-Nitrophenol	180	E	1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	200	E	1.90	5.00	ug/L
120-83-2	2,4-Dichlorophenol	68.8		0.52	5.00	ug/L
65-85-0	Benzoic acid	4.20	U	4.20	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	63.2		0.59	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	130	E	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	0.62	U	0.62	5.00	ug/L
51-28-5	2,4-Dinitrophenol	200	E	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	140	E	2.40	10.0	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	96.9	E	2.90	10.0	ug/L
87-86-5	Pentachlorophenol	82.5	E	1.60	10.0	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	57.4		0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	145		10 - 139	97%	SPK: 150
13127-88-3	Phenol-d6	142		10 - 134	95%	SPK: 150
118-79-6	2,4,6-Tribromophenol	151		44 - 137	101%	SPK: 150
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	333000	7.74			
1146-65-2	Naphthalene-d8	1360000	10.534			
15067-26-2	Acenaphthene-d10	823000	14.416			
1517-22-2	Phenanthrene-d10	1600000	17.222			
1719-03-5	Chrysene-d12	1660000	21.698			
1520-96-3	Perylene-d12	1860000	25.121			

Raw Data: **BP024175.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	03/03/25
Project:	NJ Waste Water PT	Date Received:	03/05/25
Client Sample ID:	PT-ACIDS-WP	SDG No.:	Q1502
Lab Sample ID:	Q1502-06	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024175.D	1	03/12/25 08:45	03/13/25 13:51	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024175.D
Acq On : 13 Mar 2025 13:51
Operator : RC/JU
Sample : Q1502-06
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Quant Time: Mar 13 14:15:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.740	152	332785	20.000	ng	-0.03
21) Naphthalene-d8	10.534	136	1357432	20.000	ng	-0.02
39) Acenaphthene-d10	14.416	164	822876	20.000	ng	0.01
64) Phenanthrene-d10	17.222	188	1600351	20.000	ng	0.02
76) Chrysene-d12	21.698	240	1655513	20.000	ng	0.04
86) Perylene-d12	25.121	264	1860601	20.000	ng	0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.352	112	3029498	145.317	ng	-0.03
7) Phenol-d6	6.922	99	3965401	142.238	ng	-0.03
23) Nitrobenzene-d5	8.899	82	2292555	95.289	ng	-0.02
42) 2,4,6-Tribromophenol	15.934	330	1564249	150.914	ng	0.02
45) 2-Fluorobiphenyl	13.016	172	5235368	93.813	ng	0.00
79) Terphenyl-d14	19.957	244	8356039	104.190	ng	0.04
Target Compounds						Qvalue
8) 2-Chlorophenol	7.316	128	883671	39.242	ng	99
10) Phenol	6.952	94	3907245	138.832	ng	98
17) 2-Methylphenol	8.187	107	3207081	181.723	ng	99
20) 3+4-Methylphenols	8.505	107	2618812	108.010	ng	91
26) 2-Nitrophenol	9.646	139	1995159	175.987	ng	99
27) 2,4-Dimethylphenol	9.710	122	2866370	196.624	ng	100
29) 2,4-Dichlorophenol	10.181	162	1356868	68.805	ng	100
36) 4-Chloro-3-methylphenol	11.822	107	1345188	63.231	ng	98
43) 2,4,6-Trichlorophenol	12.822	196	1916696	126.791	ng	100
54) 2,4-Dinitrophenol	14.522	184	1304329	203.475	ng	96
56) 4-Nitrophenol	14.646	139	1490824	136.357	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.051	232	840235	57.365	ng	99
65) 4,6-Dinitro-2-methylph...	15.563	198	920727	96.912	ng	99
70) Pentachlorophenol	16.869	266	1109698	82.513	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024175.D
Acq On : 13 Mar 2025 13:51
Operator : RC/JU
Sample : Q1502-06
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Quant Time: Mar 13 14:15:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.740 min Scan# 7
Delta R.T. -0.029 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:152 Resp: 332785

Ion	Ratio	Lower	Upper
152	100		
150	156.2	125.5	188.3
115	57.9	46.3	69.5







#5
2-Fluorophenol
Concen: 145.317 ng
RT: 5.352 min Scan# 385
Delta R.T. -0.029 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:112 Resp: 3029498

Ion	Ratio	Lower	Upper
112	100		
64	51.7	41.3	61.9
63	25.8	20.6	31.0







#7
Phenol-d6
Concen: 142.238 ng
RT: 6.922 min Scan# 6
Delta R.T. -0.029 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion: 99 Resp: 3965401

Ion	Ratio	Lower	Upper
99	100		
42	13.1	10.6	16.0
71	29.0	23.3	34.9







#8
2-Chlorophenol
Concen: 39.242 ng
RT: 7.316 min Scan# 719
Delta R.T. -0.029 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:128 Resp: 883671

Ion	Ratio	Lower	Upper
128	100		
130	32.6	12.0	52.0
64	41.6	21.6	61.6







#10
Phenol
Concen: 138.832 ng
RT: 6.952 min Scan# 6
Delta R.T. -0.023 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion: 94 Resp: 3907245

Ion	Ratio	Lower	Upper
94	100		
65	24.3	4.9	44.9
66	30.7	11.9	51.9







#17
2-Methylphenol
Concen: 181.723 ng
RT: 8.187 min Scan# 867
Delta R.T. -0.029 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:107 Resp: 3207081

Ion	Ratio	Lower	Upper
107	100		
108	113.2	90.0	135.0
77	42.1	34.0	51.0
79	40.7	32.7	49.1







#20
3+4-Methylphenols
Concen: 108.010 ng
RT: 8.505 min Scan# 9
Delta R.T. -0.035 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:107 Resp: 2618812

Ion	Ratio	Lower	Upper
107	100		
108	80.2	69.7	109.7
77	28.0	10.3	50.3
79	19.4	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.534 min Scan# 1266
Delta R.T. -0.018 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:136 Resp: 1357432

Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	6.4	5.3	7.9
68	6.3	5.0	7.4







#23
Nitrobenzene-d5
Concen: 95.289 ng
RT: 8.899 min Scan# 9
Delta R.T. -0.024 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion: 82 Resp: 2292555

Ion	Ratio	Lower	Upper
82	100		
128	46.2	37.4	56.0
54	40.2	32.5	48.7







#26
2-Nitrophenol
Concen: 175.987 ng
RT: 9.646 min Scan# 1115
Delta R.T. -0.024 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:139 Resp: 1995159

Ion	Ratio	Lower	Upper
139	100		
109	26.6	21.7	32.5
65	44.2	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 196.624 ng
RT: 9.710 min Scan# 11
Delta R.T. -0.024 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:122 Resp: 2866370

Ion	Ratio	Lower	Upper
122	100		
107	112.0	89.3	133.9
121	59.0	47.5	71.3







#29
2,4-Dichlorophenol
Concen: 68.805 ng
RT: 10.181 min Scan# 1206
Delta R.T. -0.024 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:162 Resp: 1356868

Ion	Ratio	Lower	Upper
162	100		
164	64.2	44.2	84.2
98	35.8	15.1	55.1







#36
4-Chloro-3-methylphenol
Concen: 63.231 ng
RT: 11.822 min Scan# 14
Delta R.T. -0.018 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:107 Resp: 1345188

Ion	Ratio	Lower	Upper
107	100		
144	27.0	21.6	32.4
142	84.1	65.4	98.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.416 min Scan# 1926
Delta R.T. 0.012 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:164 Resp: 822876

Ion	Ratio	Lower	Upper
164	100		
162	100.3	80.6	121.0
160	44.6	35.4	53.2







#42
2,4,6-Tribromophenol
Concen: 150.914 ng
RT: 15.934 min Scan# 21
Delta R.T. 0.018 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:330 Resp: 1564249

Ion	Ratio	Lower	Upper
330	100		
332	96.6	77.6	116.4
141	30.7	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 126.791 ng
RT: 12.822 min Scan# 1655
Delta R.T. -0.006 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:196 Resp: 1916696

Ion	Ratio	Lower	Upper
196	100		
198	95.9	76.6	115.0
200	31.4	25.4	38.0







#45
2-Fluorobiphenyl
Concen: 93.813 ng
RT: 13.016 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:172 Resp: 5235368

Ion	Ratio	Lower	Upper
172	100		
171	35.3	28.3	42.5
170	23.9	19.0	28.4







#54
2,4-Dinitrophenol
Concen: 203.475 ng
RT: 14.522 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:184 Resp: 1304329

Ion	Ratio	Lower	Upper
184	100		
63	57.2	42.1	63.1
154	57.7	44.6	67.0







#56
4-Nitrophenol
Concen: 136.357 ng
RT: 14.646 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:139 Resp: 1490824

Ion	Ratio	Lower	Upper
139	100		
109	63.3	43.3	83.3
65	80.3	61.2	101.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 57.365 ng
RT: 15.051 min Scan# 2034
Delta R.T. 0.012 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:232 Resp: 840235

Ion	Ratio	Lower	Upper
232	100		
131	45.4	35.9	53.9
130	2.3	1.8	2.6
166	30.9	24.2	36.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.222 min Scan# 24
Delta R.T. 0.018 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:188 Resp: 1600351

Ion	Ratio	Lower	Upper
188	100		
94	10.3	8.1	12.1
80	11.0	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 96.912 ng
RT: 15.563 min Scan# 2121
Delta R.T. 0.012 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:198 Resp: 920727

Ion	Ratio	Lower	Upper
198	100		
51	34.2	14.7	54.7
105	43.0	23.8	63.8







#70
Pentachlorophenol
Concen: 82.513 ng
RT: 16.869 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:266 Resp: 1109698

Ion	Ratio	Lower	Upper
266	100		
268	64.3	51.3	76.9
264	63.4	50.6	76.0







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.698 min Scan# 3164
Delta R.T. 0.041 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:240 Resp: 1655513

Ion	Ratio	Lower	Upper
240	100		
120	12.0	9.2	13.8
236	25.1	20.4	30.6







#79
Terphenyl-d14
Concen: 104.190 ng
RT: 19.957 min Scan# 2
Delta R.T. 0.035 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WP

Tgt Ion:244 Resp: 8356039

Ion	Ratio	Lower	Upper
244	100		
212	7.4	5.7	8.5
122	12.1	9.6	14.4






#86
Perylene-d12
Concen: 20.000 ng
RT: 25.121 min Scan# 3746
Delta R.T. 0.065 min
Lab File: BP024175.D
Acq: 13 Mar 2025 13:51

Tgt Ion:264 Resp: 1860601

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.5	27.7
265	21.9	17.7	26.5



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	03/03/25
Project:	NJ Waste Water PT	Date Received:	03/05/25
Client Sample ID:	PT-ACIDS-WPDL	SDG No.:	Q1502
Lab Sample ID:	Q1502-06DL	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024176.D	5	03/12/25 08:45	03/13/25 15:48	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
108-95-2	Phenol	150	D	4.60	25.0	ug/L
95-57-8	2-Chlorophenol	39.6	D	2.90	25.0	ug/L
95-48-7	2-Methylphenol	190	D	5.60	25.0	ug/L
65794-96-9	3+4-Methylphenols	110	D	5.50	50.0	ug/L
88-75-5	2-Nitrophenol	170	D	8.80	25.0	ug/L
105-67-9	2,4-Dimethylphenol	210	D	9.30	25.0	ug/L
120-83-2	2,4-Dichlorophenol	67.0	D	2.60	25.0	ug/L
65-85-0	Benzoic acid	21.0	UD	21.0	50.0	ug/L
59-50-7	4-Chloro-3-methylphenol	61.5	D	3.00	25.0	ug/L
88-06-2	2,4,6-Trichlorophenol	130	D	2.60	25.0	ug/L
95-95-4	2,4,5-Trichlorophenol	3.10	UD	3.10	25.0	ug/L
51-28-5	2,4-Dinitrophenol	160	D	29.9	50.0	ug/L
100-02-7	4-Nitrophenol	130	D	11.9	50.0	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	75.6	D	14.4	50.0	ug/L
87-86-5	Pentachlorophenol	73.8	D	7.90	50.0	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	54.8	D	3.60	25.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	149		10 - 139	99%	SPK: 150
13127-88-3	Phenol-d6	148		10 - 134	99%	SPK: 150
118-79-6	2,4,6-Tribromophenol	144		44 - 137	96%	SPK: 150
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	333000	7.752			
1146-65-2	Naphthalene-d8	1380000	10.54			
15067-26-2	Acenaphthene-d10	846000	14.398			
1517-22-2	Phenanthrene-d10	1630000	17.21			
1719-03-5	Chrysene-d12	1670000	21.669			
1520-96-3	Perylene-d12	1840000	25.086			

Raw Data: **BP024176.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	03/03/25
Project:	NJ Waste Water PT	Date Received:	03/05/25
Client Sample ID:	PT-ACIDS-WPDL	SDG No.:	Q1502
Lab Sample ID:	Q1502-06DL	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024176.D	5	03/12/25 08:45	03/13/25 15:48	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024176.D
Acq On : 13 Mar 2025 15:48
Operator : RC/JU
Sample : Q1502-06DL 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Quant Time: Mar 13 16:13:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.752	152	333252	20.000	ng	-0.02
21) Naphthalene-d8	10.540	136	1381029	20.000	ng	-0.01
39) Acenaphthene-d10	14.398	164	845510	20.000	ng	0.00
64) Phenanthrene-d10	17.210	188	1630079	20.000	ng	0.00
76) Chrysene-d12	21.669	240	1669058	20.000	ng	0.01
86) Perylene-d12	25.086	264	1840328	20.000	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.358	112	622174	29.802	ng	-0.02
7) Phenol-d6	6.928	99	827403	29.637	ng	-0.02
23) Nitrobenzene-d5	8.899	82	484698	19.802	ng	-0.02
42) 2,4,6-Tribromophenol	15.916	330	306659	28.794	ng	0.00
45) 2-Fluorobiphenyl	13.004	172	1185484	20.674	ng	-0.01
79) Terphenyl-d14	19.933	244	1827756	22.605	ng	0.01
Target Compounds						Qvalue
8) 2-Chlorophenol	7.322	128	178656	7.923	ng	99
10) Phenol	6.952	94	837578	29.719	ng	100
17) 2-Methylphenol	8.193	107	685521	38.789	ng	99
20) 3+4-Methylphenols	8.510	107	551986	22.734	ng	91
26) 2-Nitrophenol	9.646	139	393632	34.128	ng	98
27) 2,4-Dimethylphenol	9.710	122	611752	41.247	ng	98
29) 2,4-Dichlorophenol	10.187	162	268731	13.394	ng	99
36) 4-Chloro-3-methylphenol	11.828	107	266244	12.301	ng	97
43) 2,4,6-Trichlorophenol	12.816	196	392630	25.277	ng	99
54) 2,4-Dinitrophenol	14.516	184	204675	31.075	ng	97
56) 4-Nitrophenol	14.628	139	290976	25.901	ng	97
59) 2,3,4,6-Tetrachlorophenol	15.040	232	164836	10.953	ng	99
65) 4,6-Dinitro-2-methylph...	15.545	198	146285	15.117	ng	95
70) Pentachlorophenol	16.857	266	202265	14.765	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024176.D
Acq On : 13 Mar 2025 15:48
Operator : RC/JU
Sample : Q1502-06DL 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Quant Time: Mar 13 16:13:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.752 min Scan# 7
Delta R.T. -0.017 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:152 Resp: 333252

Ion	Ratio	Lower	Upper
152	100		
150	153.8	125.5	188.3
115	57.2	46.3	69.5







#5
2-Fluorophenol
Concen: 29.802 ng
RT: 5.358 min Scan# 386
Delta R.T. -0.024 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:112 Resp: 622174

Ion	Ratio	Lower	Upper
112	100		
64	52.9	41.3	61.9
63	27.0	20.6	31.0







#7
Phenol-d6
Concen: 29.637 ng
RT: 6.928 min Scan# 6
Delta R.T. -0.024 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion: 99 Resp: 827403

Ion	Ratio	Lower	Upper
99	100		
42	13.9	10.6	16.0
71	29.0	23.3	34.9







#8
2-Chlorophenol
Concen: 7.923 ng
RT: 7.322 min Scan# 720
Delta R.T. -0.024 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:128 Resp: 178656

Ion	Ratio	Lower	Upper
128	100		
130	31.7	12.0	52.0
64	41.9	21.6	61.6







#10
Phenol
Concen: 29.719 ng
RT: 6.952 min Scan# 6
Delta R.T. -0.024 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion: 94 Resp: 837578

Ion	Ratio	Lower	Upper
94	100		
65	25.0	4.9	44.9
66	32.2	11.9	51.9







#17
2-Methylphenol
Concen: 38.789 ng
RT: 8.193 min Scan# 868
Delta R.T. -0.024 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:107 Resp: 685521

Ion	Ratio	Lower	Upper
107	100		
108	110.3	90.0	135.0
77	42.2	34.0	51.0
79	40.5	32.7	49.1







#20
3+4-Methylphenols
Concen: 22.734 ng
RT: 8.510 min Scan# 9
Delta R.T. -0.029 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:107 Resp: 551986

Ion	Ratio	Lower	Upper
107	100		
108	80.6	69.7	109.7
77	28.4	10.3	50.3
79	19.9	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.540 min Scan# 1267
Delta R.T. -0.012 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:136 Resp: 1381029

Ion	Ratio	Lower	Upper
136	100		
137	10.6	9.0	13.4
54	6.6	5.3	7.9
68	6.3	5.0	7.4







#23
Nitrobenzene-d5
Concen: 19.802 ng
RT: 8.899 min Scan# 9
Delta R.T. -0.024 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion: 82 Resp: 484698

Ion	Ratio	Lower	Upper
82	100		
128	44.6	37.4	56.0
54	40.2	32.5	48.7







#26
2-Nitrophenol
Concen: 34.128 ng
RT: 9.646 min Scan# 1115
Delta R.T. -0.024 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:139 Resp: 393632

Ion	Ratio	Lower	Upper
139	100		
109	26.8	21.7	32.5
65	46.3	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 41.247 ng
RT: 9.710 min Scan# 11
Delta R.T. -0.024 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:122 Resp: 611752

Ion	Ratio	Lower	Upper
122	100		
107	109.8	89.3	133.9
121	57.6	47.5	71.3







#29
2,4-Dichlorophenol
Concen: 13.394 ng
RT: 10.187 min Scan# 1207
Delta R.T. -0.018 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:162 Resp: 268731

Ion	Ratio	Lower	Upper
162	100		
164	65.4	44.2	84.2
98	35.3	15.1	55.1







#36
4-Chloro-3-methylphenol
Concen: 12.301 ng
RT: 11.828 min Scan# 14
Delta R.T. -0.012 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:107 Resp: 266244

Ion	Ratio	Lower	Upper
107	100		
144	26.3	21.6	32.4
142	84.4	65.4	98.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1923
Delta R.T. -0.006 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:164 Resp: 845510

Ion	Ratio	Lower	Upper
164	100		
162	100.3	80.6	121.0
160	44.7	35.4	53.2







#42
2,4,6-Tribromophenol
Concen: 28.794 ng
RT: 15.916 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:330 Resp: 306659

Ion	Ratio	Lower	Upper
330	100		
332	96.9	77.6	116.4
141	33.0	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 25.277 ng
RT: 12.816 min Scan# 1654
Delta R.T. -0.012 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:196 Resp: 392630

Ion	Ratio	Lower	Upper
196	100		
198	94.7	76.6	115.0
200	30.9	25.4	38.0







#45
2-Fluorobiphenyl
Concen: 20.674 ng
RT: 13.004 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:172 Resp: 1185484

Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.3	42.5
170	23.6	19.0	28.4







#54
2,4-Dinitrophenol
Concen: 31.075 ng
RT: 14.516 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:184 Resp: 204675

Ion	Ratio	Lower	Upper
184	100		
63	56.7	42.1	63.1
154	55.4	44.6	67.0







#56
4-Nitrophenol
Concen: 25.901 ng
RT: 14.628 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:139 Resp: 290976

Ion	Ratio	Lower	Upper
139	100		
109	65.8	43.3	83.3
65	83.5	61.2	101.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 10.953 ng
RT: 15.040 min Scan# 2032
Delta R.T. -0.000 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:232 Resp: 164836

Ion	Ratio	Lower	Upper
232	100		
131	45.3	35.9	53.9
130	2.3	1.8	2.6
166	31.4	24.2	36.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.210 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:188 Resp: 1630079

Ion	Ratio	Lower	Upper
188	100		
94	10.7	8.1	12.1
80	11.7	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 15.117 ng
RT: 15.545 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:198 Resp: 146285

Ion	Ratio	Lower	Upper
198	100		
51	39.2	14.7	54.7
105	46.0	23.8	63.8







#70
Pentachlorophenol
Concen: 14.765 ng
RT: 16.857 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:266 Resp: 202265

Ion	Ratio	Lower	Upper
266	100		
268	65.5	51.3	76.9
264	63.8	50.6	76.0







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.669 min Scan# 3159
Delta R.T. 0.012 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:240 Resp: 1669058

Ion	Ratio	Lower	Upper
240	100		
120	12.1	9.2	13.8
236	25.8	20.4	30.6







#79
Terphenyl-d14
Concen: 22.605 ng
RT: 19.933 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Instrument :
BNA_P
ClientSampleId :
PT-ACIDS-WPDL

Tgt Ion:244 Resp: 1827756

Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.7	8.5
122	13.3	9.6	14.4






#86
Perylene-d12
Concen: 20.000 ng
RT: 25.086 min Scan# 3740
Delta R.T. 0.029 min
Lab File: BP024176.D
Acq: 13 Mar 2025 15:48

Tgt Ion:264 Resp: 1840328

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.5	27.7
265	21.9	17.7	26.5



CALIBRATION SUMMARY

Response Factor Report BNA_F

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF031025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Mar 10 15:46:22 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BF141897.D 5 =BF141898.D 10 =BF141899.D 20 =BF141900.D 40 =BF141901.D 50 =BF141905.D 60 =BF141903.D 80 =BF1419
04.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------	ISTD	----------							
2)		1,4-Dioxane		0.490	0.482	0.536	0.504	0.495	0.501	0.508	0.502	3.43
3)		Pyridine		1.217	1.204	1.334	1.221	1.203	1.209	1.233	1.232	3.77
4)		n-Nitrosodimet...		0.554	0.556	0.621	0.583	0.595	0.594	0.607	0.587	4.26
5)	S	2-Fluorophenol		1.267	1.243	1.298	1.148	1.162	1.146	1.124	1.198	5.79
6)		Aniline		1.599	1.602	1.640	1.465	1.443	1.435	1.351	1.505	7.20
7)	S	Phenol-d6		1.623	1.593	1.644	1.442	1.483	1.471	1.424	1.526	5.99
8)		2-Chlorophenol		1.421	1.354	1.434	1.259	1.295	1.284	1.256	1.329	5.63
9)		Benzaldehyde			1.021	0.975	0.778	0.778	0.716		0.853	15.85
10)	C	Phenol		1.708	1.675	1.747	1.536	1.542	1.532	1.496	1.605	6.30
11)		bis(2-Chloroet...		1.287	1.246	1.281	1.144	1.175	1.176	1.127	1.205	5.43
12)		1,3-Dichlorobe...		1.475	1.489	1.538	1.373	1.401	1.401	1.367	1.435	4.59
13)	C	1,4-Dichlorobe...		1.514	1.503	1.548	1.393	1.408	1.420	1.376	1.452	4.71
14)		1,2-Dichlorobe...		1.461	1.417	1.460	1.303	1.315	1.343	1.273	1.367	5.69
15)		Benzyl Alcohol		1.236	1.252	1.322	1.185	1.214	1.246	1.180	1.233	3.91
16)		2,2'-oxybis(1-...		1.591	1.525	1.534	1.363	1.387	1.472	1.313	1.455	7.07
17)		2-Methylphenol		1.080	1.050	1.122	1.008	1.038	1.079	1.021	1.057	3.73
18)		Hexachloroethane		0.545	0.551	0.576	0.517	0.540	0.560	0.522	0.544	3.76
19)	P	n-Nitroso-di-n...	1.020	1.024	1.000	1.043	0.913	0.940	0.991	0.912	0.980	5.29
20)		3+4-Methylphenols		1.449	1.394	1.470	1.294	1.311	1.334	1.223	1.354	6.55
21)	I	Naphthalene-d8	----------	ISTD	----------							
22)		Acetophenone		0.514	0.493	0.513	0.460	0.450	0.469	0.453	0.479	5.76
23)	S	Nitrobenzene-d5		0.345	0.356	0.379	0.348	0.344	0.363	0.353	0.355	3.47
24)		Nitrobenzene		0.347	0.351	0.375	0.346	0.343	0.359	0.352	0.353	3.07
25)		Isophorone		0.647	0.620	0.655	0.600	0.605	0.648	0.622	0.628	3.49
26)	C	2-Nitrophenol		0.150	0.159	0.182	0.175	0.179	0.188	0.181	0.173	7.88
27)		2,4-Dimethylph...		0.247	0.234	0.251	0.231	0.233	0.244	0.232	0.239	3.45
28)		bis(2-Chloroet...		0.411	0.406	0.416	0.372	0.377	0.397	0.378	0.394	4.63
29)	C	2,4-Dichloroph...		0.292	0.295	0.309	0.286	0.295	0.303	0.294	0.296	2.54
30)		1,2,4-Trichlor...		0.337	0.325	0.342	0.310	0.318	0.333	0.322	0.327	3.50
31)		Naphthalene		1.104	1.075	1.101	0.984	1.001	0.973	0.954	1.027	6.22
32)		Benzoic acid			0.149	0.191	0.208	0.228	0.242	0.241	0.210	17.07
33)		4-Chloroaniline		0.383	0.378	0.393	0.351	0.354	0.352	0.329	0.363	6.14
34)	C	Hexachlorobuta...		0.211	0.211	0.221	0.203	0.214	0.214	0.216	0.213	2.52
35)		Caprolactam		0.092	0.089	0.094	0.084	0.088	0.090	0.094	0.090	4.00
36)	C	4-Chloro-3-met...		0.334	0.331	0.349	0.318	0.329	0.334	0.320	0.331	3.13
37)		2-Methylnaphth...		0.746	0.709	0.733	0.653	0.655	0.667	0.645	0.687	6.09
38)		1-Methylnaphth...		0.717	0.691	0.721	0.623	0.619	0.647	0.620	0.663	6.95


Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF031025.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.615	0.604	0.626	0.592	0.597	0.606	0.623	0.609	2.11
41)	P	Hexachlorocycl...	0.204	0.221	0.249	0.250	0.237	0.252	0.256	0.238	8.03
42)	S	2,4,6-Tribromo...	0.234	0.236	0.259	0.250	0.259	0.265	0.272	0.254	5.67
43)	C	2,4,6-Trichlor...	0.388	0.389	0.402	0.398	0.400	0.401	0.409	0.398	1.91
44)		2,4,5-Trichlor...	0.396	0.391	0.422	0.389	0.400	0.415	0.407	0.403	3.05
45)	S	2-Fluorobiphenyl	1.430	1.379	1.379	1.254	1.272	1.246	1.246	1.315	5.95
46)		1,1'-Biphenyl	1.636	1.573	1.591	1.454	1.502	1.431	1.450	1.520	5.29
47)		2-Chloronaphth...	1.197	1.141	1.181	1.091	1.129	1.090	1.100	1.133	3.82
48)		2-Nitroaniline	0.296	0.314	0.346	0.322	0.356	0.332	0.330	0.328	6.08
49)		Acenaphthylene	1.765	1.740	1.778	1.635	1.664	1.595	1.589	1.681	4.74
50)		Dimethylphthalate	1.481	1.414	1.469	1.329	1.425	1.345	1.341	1.401	4.47
51)		2,6-Dinitrotol...	0.277	0.279	0.305	0.285	0.307	0.296	0.292	0.292	4.11
52)	C	Acenaphthene	1.236	1.195	1.231	1.146	1.144	1.165	1.152	1.181	3.37
53)		3-Nitroaniline	0.291	0.295	0.316	0.293	0.283	0.301	0.291	0.296	3.61
54)	P	2,4-Dinitrophenol		0.085	0.124	0.142	0.153	0.161	0.169	0.139	22.02
55)		Dibenzofuran	1.849	1.787	1.791	1.614	1.607	1.596	1.570	1.688	6.87
56)	P	4-Nitrophenol	0.192	0.216	0.234	0.230	0.229	0.230	0.230	0.223	6.60
57)		2,4-Dinitrotol...	0.365	0.382	0.402	0.375	0.383	0.383	0.376	0.381	2.95
58)		Fluorene	1.443	1.361	1.381	1.229	1.243	1.234	1.220	1.302	7.01
59)		2,3,4,6-Tetrac...	0.355	0.365	0.390	0.359	0.364	0.366	0.369	0.367	3.05
60)		Diethylphthalate	1.475	1.471	1.497	1.336	1.357	1.338	1.302	1.397	5.80
61)		4-Chlorophenyl...	0.730	0.688	0.702	0.643	0.664	0.656	0.667	0.679	4.41
62)		4-Nitroaniline	0.284	0.288	0.304	0.288	0.287	0.286	0.278	0.288	2.73
63)		Azobenzene	1.392	1.363	1.404	1.245	1.261	1.226	1.197	1.298	6.57
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.083	0.111	0.121	0.130	0.133	0.137	0.119	16.61
66)	c	n-Nitrosodiphe...	0.675	0.651	0.671	0.620	0.638	0.638	0.637	0.647	3.03
67)		4-Bromophenyl-...	0.240	0.249	0.255	0.244	0.263	0.245	0.262	0.251	3.55
68)		Hexachlorobenzene	0.262	0.267	0.284	0.269	0.283	0.271	0.292	0.275	3.92
69)		Atrazine	0.216	0.213	0.193	0.164	0.206			0.198	10.72
70)	C	Pentachlorophenol	0.130	0.151	0.172	0.177	0.180	0.186	0.191	0.170	12.66
71)		Phenanthrene	1.155	1.151	1.138	1.038	1.026	1.032	1.023	1.080	5.90
72)		Anthracene	1.173	1.139	1.139	1.046	1.026	1.036	1.026	1.084	5.89
73)		Carbazole	0.996	0.994	0.985	0.908	0.903	0.892	0.864	0.935	5.91
74)		Di-n-butylphth...	1.320	1.321	1.319	1.177	1.238	1.175	1.132	1.240	6.50
75)	C	Fluoranthene	1.234	1.226	1.228	1.108	1.148	1.059	1.017	1.146	7.66
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine	0.347	0.309	0.190	0.360	0.220	0.275	0.252	0.279	22.73
78)		Pyrene	1.695	1.651	1.727	1.611	1.877	1.823	1.726	1.730	5.38
79)	S	Terphenyl-d14	1.343	1.309	1.360	1.276	1.380	1.417	1.384	1.353	3.56
80)		Butylbenzylpht...	0.629	0.656	0.702	0.655	0.703	0.717	0.677	0.677	4.72
81)		Benzo(a)anthra...	1.370	1.304	1.334	1.273	1.294	1.317	1.294	1.312	2.41
82)		3,3'-Dichlorob...	0.370	0.389	0.380	0.372	0.365	0.394	0.384	0.379	2.78
83)		Chrysene	1.143	1.194	1.260	1.136	1.192	1.213	1.189	1.190	3.53
84)		Bis(2-ethylhex...	0.897	0.912	0.966	0.904	0.978	0.952	0.927	0.934	3.43
85)	c	Di-n-octyl pht...	1.145	1.219	1.338	1.265	1.376	1.390	1.361	1.299	7.10

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF031025.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.097	1.142	1.308	1.287	1.322	1.472	1.429	1.294	10.60
88)		Benzo(b)fluora...	1.420	1.310	1.485	1.210	1.346	1.390	1.434	1.371	6.65
89)		Benzo(k)fluora...	1.162	1.268	1.197	1.205	1.129	1.210	1.041	1.173	6.16
90)	C	Benzo(a)pyrene	1.040	1.042	1.123	1.036	1.083	1.083	1.101	1.073	3.16
91)		Dibenzo(a,h)an...	0.910	0.953	1.073	1.068	1.095	1.195	1.178	1.067	9.91
92)		Benzo(g,h,i)pe...	0.912	0.949	1.061	1.065	1.064	1.167	1.167	1.055	9.24

(#) = Out of Range



Response = 1.805e-001 * Amt - 5.933e-002
Coef of Det (r^2) = 0.997436 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF031025.M
Calibration Table Last Updated: Mon Mar 10 15:46:22 2025

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141897.D
Acq On : 10 Mar 2025 11:01
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Mar 10 15:31:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	218163	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	918034	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	548310	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	1010378	20.000	ng	0.00
76) Chrysene-d12	14.033	240	740344	20.000	ng	0.00
86) Perylene-d12	15.509	264	520745	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.275	70	27816	2.599	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141897.D
Acq On : 10 Mar 2025 11:01
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Mar 10 15:31:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 218163

Ion	Ratio	Lower	Upper
152	100		
150	158.0	127.4	191.2
115	65.4	51.9	77.9







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.49 min

Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	58.1
63	31.8






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.50 min

Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99


Sig	Exp Ratio
99	100
42	18.2
71	38.5




#19
n-Nitroso-di-n-propylamine
Concen: 2.599 ng
RT: 7.275 min Scan# 881
Delta R.T. -0.024 min
Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Tgt Ion: 70 Resp: 27816

Ion	Ratio	Lower	Upper
70	100		
42	56.8	41.4	62.0
101	10.3	8.2	12.2
130	19.8	18.2	27.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 918034

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	7.5	5.8	8.8
68	5.2	4.1	6.1







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 7.45 min

Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	45.0
54	49.5







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 548310

Ion	Ratio	Lower	Upper
164	100		
162	101.3	81.8	122.6
160	45.5	36.7	55.1







#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 10.70 min

Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	97.0
141	30.9






#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 9.24 min

Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172


Sig	Exp Ratio
172	100
171	36.6
170	24.2




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.398 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Tgt Ion:188 Resp: 1010378

Ion	Ratio	Lower	Upper
188	100		
94	8.9	6.8	10.2
80	10.0	7.6	11.4







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.033 min Scan# 2(
Delta R.T. -0.006 min
Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 740344

Ion	Ratio	Lower	Upper
240	100		
120	10.5	8.4	12.6
236	26.2	20.5	30.7







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 12.99 min

Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	7.5
122	9.6






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.509 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF141897.D
Acq: 10 Mar 2025 11:01

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 520745

Ion	Ratio	Lower	Upper
264	100		
260	23.7	19.1	28.7
265	21.9	17.5	26.3



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141898.D
Acq On : 10 Mar 2025 11:30
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Mar 10 15:32:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	230406	20.000	ng		0.00
21) Naphthalene-d8	8.157	136	926118	20.000	ng		0.00
39) Acenaphthene-d10	9.916	164	546702	20.000	ng		0.00
64) Phenanthrene-d10	11.398	188	1002901	20.000	ng		0.00
76) Chrysene-d12	14.033	240	745624	20.000	ng		0.00
86) Perylene-d12	15.504	264	527129	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.487	112	145959	10.570	ng		0.00
7) Phenol-d6	6.486	99	186951	10.630	ng		-0.02
23) Nitrobenzene-d5	7.433	82	159777	9.670	ng		-0.01
42) 2,4,6-Tribromophenol	10.698	330	63996	9.238	ng		0.00
45) 2-Fluorobiphenyl	9.233	172	390913	10.888	ng		0.00
79) Terphenyl-d14	12.980	244	500607	9.924	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.675	88	28210	4.874	ng		99
3) Pyridine	3.451	79	70090	4.928	ng		99
4) n-Nitrosodimethylamine	3.369	42	31893	4.717	ng	#	99
6) Aniline	6.534	93	92116	5.305	ng		100
8) 2-Chlorophenol	6.657	128	81845	5.347	ng		98
10) Phenol	6.504	94	98369	5.301	ng		96
11) bis(2-Chloroethyl)ether	6.604	93	74158	5.334	ng		98
12) 1,3-Dichlorobenzene	6.816	146	84952	5.136	ng		100
13) 1,4-Dichlorobenzene	6.898	146	87237	5.213	ng		98
14) 1,2-Dichlorobenzene	7.045	146	84183	5.342	ng		99
15) Benzyl Alcohol	7.010	79	71182	5.000	ng		98
16) 2,2'-oxybis(1-Chloropr...	7.151	45	91659	5.470	ng		97
17) 2-Methylphenol	7.116	107	62211	5.114	ng		99
18) Hexachloroethane	7.392	117	31373	5.000	ng		96
19) n-Nitroso-di-n-propyla...	7.281	70	58981	5.218	ng		98
20) 3+4-Methylphenols	7.269	107	83461	5.349	ng		97
22) Acetophenone	7.281	105	119040	5.358	ng		99
24) Nitrobenzene	7.451	77	80338	4.904	ng		99
25) Isophorone	7.692	82	149807	5.135	ng		99
26) 2-Nitrophenol	7.775	139	34757	4.329	ng		98
27) 2,4-Dimethylphenol	7.804	122	57106	5.164	ng		96
28) bis(2-Chloroethoxy)met...	7.904	93	95214	5.210	ng		97
29) 2,4-Dichlorophenol	8.010	162	67710	4.934	ng		99
30) 1,2,4-Trichlorobenzene	8.098	180	78127	5.160	ng		98
31) Naphthalene	8.180	128	255632	5.388	ng		98
33) 4-Chloroaniline	8.228	127	88635	5.294	ng		99
34) Hexachlorobutadiene	8.298	225	48885	4.966	ng		99
35) Caprolactam	8.557	113	21358	5.118	ng		94
36) 4-Chloro-3-methylphenol	8.692	107	77370	5.070	ng		99
37) 2-Methylnaphthalene	8.869	142	172734	5.427	ng		99
38) 1-Methylnaphthalene	8.969	142	166050	5.387	ng		99
40) 1,2,4,5-Tetrachloroben...	9.033	216	84034	5.041	ng		99
41) Hexachlorocyclopentadiene	9.022	237	27931	4.246	ng		96
43) 2,4,6-Trichlorophenol	9.145	196	52998	4.877	ng		99
44) 2,4,5-Trichlorophenol	9.180	196	54111	4.901	ng		97
46) 1,1'-Biphenyl	9.333	154	223637	5.395	ng		98

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141898.D
Acq On : 10 Mar 2025 11:30
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Mar 10 15:32:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
47) 2-Chloronaphthalene	9.357	162	163642	5.298	ng	97
48) 2-Nitroaniline	9.451	65	40475	4.561	ng	98
49) Acenaphthylene	9.774	152	241253	5.268	ng	99
50) Dimethylphthalate	9.627	163	202373	5.328	ng	100
51) 2,6-Dinitrotoluene	9.686	165	37796	4.785	ng	99
52) Acenaphthene	9.945	154	168975	5.225	ng	98
53) 3-Nitroaniline	9.857	138	39761	4.902	ng	98
55) Dibenzofuran	10.116	168	252715	5.481	ng	97
56) 4-Nitrophenol	10.004	139	26252	4.309	ng	99
57) 2,4-Dinitrotoluene	10.092	165	49913	4.800	ng	98
58) Fluorene	10.457	166	197266	5.542	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.233	232	48472	4.839	ng	98
60) Diethylphthalate	10.327	149	201554	5.286	ng	99
61) 4-Chlorophenyl-phenyle...	10.451	204	99747	5.380	ng	98
62) 4-Nitroaniline	10.463	138	38863	4.945	ng	94
63) Azobenzene	10.610	77	190233	5.370	ng	98
66) n-Nitrosodiphenylamine	10.569	169	169225	5.223	ng	98
67) 4-Bromophenyl-phenylether	10.945	248	60238	4.823	ng	98
68) Hexachlorobenzene	11.004	284	65620	4.753	ng	98
69) Atrazine	11.092	200	54251	6.867	ng	98
70) Pentachlorophenol	11.198	266	32716	3.843	ng	100
71) Phenanthrene	11.421	178	289580	5.340	ng	98
72) Anthracene	11.474	178	294198	5.407	ng	98
73) Carbazole	11.627	167	249634	5.329	ng	99
74) Di-n-butylphthalate	11.957	149	330999	5.352	ng	99
75) Fluoranthene	12.610	202	309273	5.432	ng	99
77) Benzidine	12.733	184	64673	6.138	ng	99
78) Pyrene	12.839	202	315969	4.941	ng	99
80) Butylbenzylphthalate	13.457	149	117256	4.653	ng	99
81) Benzo(a)anthracene	14.021	228	255308	5.205	ng	100
82) 3,3'-Dichlorobenzidine	13.986	252	68969	4.865	ng	100
83) Chrysene	14.062	228	213144	4.837	ng	99
84) Bis(2-ethylhexyl)phtha...	14.015	149	167165	4.836	ng	99
85) Di-n-octyl phthalate	14.627	149	213395	4.435	ng	99
87) Indeno(1,2,3-cd)pyrene	16.980	276	144536	4.228	ng	99
88) Benzo(b)fluoranthene	15.068	252	187154	5.234	ng	99
89) Benzo(k)fluoranthene	15.098	252	153142	4.903	ng	99
90) Benzo(a)pyrene	15.439	252	137043	4.860	ng	99
91) Dibenzo(a,h)anthracene	16.998	278	119937	4.252	ng	99
92) Benzo(g,h,i)perylene	17.427	276	120142	4.294	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141898.D
Acq On : 10 Mar 2025 11:30
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Mar 10 15:32:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:152 Resp: 230406

Ion	Ratio	Lower	Upper
152	100		
150	155.9	127.4	191.2
115	64.6	51.9	77.9







#2
1,4-Dioxane
Concen: 4.874 ng
RT: 2.675 min Scan# 99
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 88 Resp: 28210

Ion	Ratio	Lower	Upper
88	100		
58	71.4	57.4	86.2
43	27.5	21.5	32.3







#3
Pyridine
Concen: 4.928 ng
RT: 3.451 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 70090

Ion	Ratio	Lower	Upper
79	100		
52	67.3	53.3	79.9
51	35.2	27.6	41.4







#4
n-Nitrosodimethylamine
Concen: 4.717 ng
RT: 3.369 min Scan# 217
Delta R.T. -0.024 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 42 Resp: 31893

Ion	Ratio	Lower	Upper
42	100		
74	129.9	103.4	155.0
44	9.5	5.0	7.4#







#5
2-Fluorophenol
Concen: 10.570 ng
RT: 5.487 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:112 Resp: 145959

Ion	Ratio	Lower	Upper
112	100		
64	57.8	46.5	69.7
63	29.9	25.4	38.2







#6
Aniline
Concen: 5.305 ng
RT: 6.534 min Scan# 755
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 93 Resp: 92116

Ion	Ratio	Lower	Upper
93	100		
66	39.6	31.9	47.9
65	19.6	15.8	23.6







#7
Phenol-d6
Concen: 10.630 ng
RT: 6.486 min Scan# 7
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 99 Resp: 186951

Ion	Ratio	Lower	Upper
99	100		
42	17.5	14.6	21.8
71	36.4	30.8	46.2







#8
2-Chlorophenol
Concen: 5.347 ng
RT: 6.657 min Scan# 776
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:128 Resp: 81845

Ion	Ratio	Lower	Upper
128	100		
130	31.5	12.8	52.8
64	45.5	24.4	64.4







#10
Phenol
Concen: 5.301 ng
RT: 6.504 min Scan# 7
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 94 Resp: 98369

Ion	Ratio	Lower	Upper
94	100		
65	28.4	10.5	50.5
66	43.3	25.8	65.8







#11
bis(2-Chloroethyl)ether
Concen: 5.334 ng
RT: 6.604 min Scan# 767
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 93 Resp: 74158

Ion	Ratio	Lower	Upper
93	100		
63	75.3	53.0	93.0
95	32.9	12.9	52.9







#12
1,3-Dichlorobenzene
Concen: 5.136 ng
RT: 6.816 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:146 Resp: 84952

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.3	76.9
75	33.3	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 5.213 ng
RT: 6.898 min Scan# 817
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:146 Resp: 87237

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.0	76.6
111	42.2	35.7	53.5







#14
1,2-Dichlorobenzene
Concen: 5.342 ng
RT: 7.045 min Scan# 84
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:146 Resp: 84183

Ion	Ratio	Lower	Upper
146	100		
148	64.7	50.8	76.2
111	47.4	38.0	57.0







#15
Benzyl Alcohol
Concen: 5.000 ng
RT: 7.010 min Scan# 836
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 79 Resp: 71182

Ion	Ratio	Lower	Upper
79	100		
108	76.1	59.7	89.5
77	64.8	51.0	76.6







#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.470 ng
RT: 7.151 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 45 Resp: 91659

Ion	Ratio	Lower	Upper
45	100		
77	16.5	0.0	37.8
79	13.0	0.0	34.2







#17
2-Methylphenol
Concen: 5.114 ng
RT: 7.116 min Scan# 854
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:107 Resp: 62211

Ion	Ratio	Lower	Upper
107	100		
108	112.2	90.2	135.2
77	47.6	39.6	59.4
79	48.5	39.1	58.7







#18
Hexachloroethane
Concen: 5.000 ng
RT: 7.392 min Scan# 9(
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:117 Resp: 31373

Ion	Ratio	Lower	Upper
117	100		
119	99.3	76.9	115.3
201	104.7	79.9	119.9







#19
n-Nitroso-di-n-propylamine
Concen: 5.218 ng
RT: 7.281 min Scan# 882
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 70 Resp: 58981

Ion	Ratio	Lower	Upper
70	100		
42	53.7	41.4	62.0
101	9.9	8.2	12.2
130	22.5	18.2	27.4







#20
3+4-Methylphenols
Concen: 5.349 ng
RT: 7.269 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 83461

Ion	Ratio	Lower	Upper
107	100		
108	88.4	68.4	108.4
77	83.1	69.7	109.7
79	30.9	11.4	51.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:136 Resp: 926118

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.8	13.2
54	7.7	5.8	8.8
68	5.3	4.1	6.1







#22
Acetophenone
Concen: 5.358 ng
RT: 7.281 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:105 Resp: 119040

Ion	Ratio	Lower	Upper
105	100		
71	8.0	5.8	8.8
51	29.0	23.4	35.0
120	22.0	17.3	25.9







#23
Nitrobenzene-d5
Concen: 9.670 ng
RT: 7.433 min Scan# 908
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 82 Resp: 159777

Ion	Ratio	Lower	Upper
82	100		
128	42.0	36.0	54.0
54	49.2	39.6	59.4







#24
Nitrobenzene
Concen: 4.904 ng
RT: 7.451 min Scan# 91
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 77 Resp: 80338

Ion	Ratio	Lower	Upper
77	100		
123	45.2	37.0	55.4
65	13.6	11.3	16.9







#25
Isophorone
Concen: 5.135 ng
RT: 7.692 min Scan# 952
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 82 Resp: 149807

Ion	Ratio	Lower	Upper
82	100		
95	7.8	6.4	9.6
138	18.3	14.9	22.3







#26
2-Nitrophenol
Concen: 4.329 ng
RT: 7.775 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:139 Resp: 34757

Ion	Ratio	Lower	Upper
139	100		
109	30.8	26.5	39.7
65	49.4	40.3	60.5







#27
2,4-Dimethylphenol
Concen: 5.164 ng
RT: 7.804 min Scan# 971
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:122 Resp: 57106

Ion	Ratio	Lower	Upper
122	100		
107	113.8	94.8	142.2
121	56.3	47.8	71.6







#28
bis(2-Chloroethoxy)methane
Concen: 5.210 ng
RT: 7.904 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 95214

Ion	Ratio	Lower	Upper
93	100		
95	32.1	26.6	40.0
123	11.3	10.2	15.4







#29
2,4-Dichlorophenol
Concen: 4.934 ng
RT: 8.010 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:162 Resp: 67710

Ion	Ratio	Lower	Upper
162	100		
164	64.4	43.8	83.8
98	38.2	18.1	58.1







#30
1,2,4-Trichlorobenzene
Concen: 5.160 ng
RT: 8.098 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:180 Resp: 78127

Ion	Ratio	Lower	Upper
180	100		
182	94.3	77.5	116.3
145	31.6	25.8	38.6







#31
Naphthalene
Concen: 5.388 ng
RT: 8.180 min Scan# 1035
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:128 Resp: 255632

Ion	Ratio	Lower	Upper
128	100		
129	10.5	9.0	13.6
127	12.9	10.8	16.2







#33
4-Chloroaniline
Concen: 5.294 ng
RT: 8.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:127 Resp: 88635

Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.2	39.4
65	30.5	25.4	38.2
92	19.4	16.2	24.4







#34
Hexachlorobutadiene
Concen: 4.966 ng
RT: 8.298 min Scan# 1055
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:225 Resp: 48885

Ion	Ratio	Lower	Upper
225	100		
223	63.6	50.6	75.8
227	64.4	52.2	78.2







#35
Caprolactam
Concen: 5.118 ng
RT: 8.557 min Scan# 1
Delta R.T. -0.047 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:113 Resp: 21358

Ion	Ratio	Lower	Upper
113	100		
55	145.9	129.6	169.6
56	113.1	104.0	144.0







#36
4-Chloro-3-methylphenol
Concen: 5.070 ng
RT: 8.692 min Scan# 1122
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:107 Resp: 77370

Ion	Ratio	Lower	Upper
107	100		
144	22.8	19.0	28.6
142	74.2	59.2	88.8







#37
2-Methylnaphthalene
Concen: 5.427 ng
RT: 8.869 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:142 Resp: 172734

Ion	Ratio	Lower	Upper
142	100		
141	88.1	71.2	106.8







#38
1-Methylnaphthalene
Concen: 5.387 ng
RT: 8.969 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:142 Resp: 166050

Ion	Ratio	Lower	Upper
142	100		
141	93.9	74.2	111.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:164 Resp: 546702

Ion	Ratio	Lower	Upper
164	100		
162	101.8	81.8	122.6
160	45.2	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.041 ng
RT: 9.033 min Scan# 1180
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:216 Resp: 84034

Ion	Ratio	Lower	Upper
216	100		
214	78.6	61.9	92.9
179	22.8	18.1	27.1
108	18.7	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 4.246 ng
RT: 9.022 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:237 Resp: 27931

Ion	Ratio	Lower	Upper
237	100		
235	66.6	43.1	83.1
272	13.2	0.0	32.0







#42
2,4,6-Tribromophenol
Concen: 9.238 ng
RT: 10.698 min Scan# 1463
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:330 Resp: 63996

Ion	Ratio	Lower	Upper
330	100		
332	96.3	77.6	116.4
141	33.4	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 4.877 ng
RT: 9.145 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:196 Resp: 52998

Ion	Ratio	Lower	Upper
196	100		
198	99.7	80.6	121.0
200	31.6	25.8	38.6







#44
2,4,5-Trichlorophenol
Concen: 4.901 ng
RT: 9.180 min Scan# 1205
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:196 Resp: 54111

Ion	Ratio	Lower	Upper
196	100		
198	98.4	80.7	121.1
97	49.4	42.1	63.1
132	27.2	23.4	35.2







#45
2-Fluorobiphenyl
Concen: 10.888 ng
RT: 9.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:172 Resp: 390913

Ion	Ratio	Lower	Upper
172	100		
171	34.9	29.3	43.9
170	23.0	19.4	29.0







#46
1,1'-Biphenyl
Concen: 5.395 ng
RT: 9.333 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:154 Resp: 223637

Ion	Ratio	Lower	Upper
154	100		
153	39.8	21.1	61.1
76	11.9	0.0	31.3







#47
2-Chloronaphthalene
Concen: 5.298 ng
RT: 9.357 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:162 Resp: 163642

Ion	Ratio	Lower	Upper
162	100		
127	41.7	35.5	53.3
164	32.8	26.6	40.0







#48
2-Nitroaniline
Concen: 4.561 ng
RT: 9.451 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 65 Resp: 40475

Ion	Ratio	Lower	Upper
65	100		
92	71.9	56.6	84.8
138	112.5	88.6	133.0







#49
Acenaphthylene
Concen: 5.268 ng
RT: 9.774 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:152 Resp: 241253

Ion	Ratio	Lower	Upper
152	100		
151	20.1	16.3	24.5
153	13.6	11.1	16.7







#50
Dimethylphthalate
Concen: 5.328 ng
RT: 9.627 min Scan# 1281
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:163 Resp: 202373

Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.1	4.7
164	10.1	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 4.785 ng
RT: 9.686 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:165 Resp: 37796

Ion	Ratio	Lower	Upper
165	100		
63	59.1	47.0	70.4
89	61.0	47.4	71.2







#52
Acenaphthene
Concen: 5.225 ng
RT: 9.945 min Scan# 1335
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:154 Resp: 168975

Ion	Ratio	Lower	Upper
154	100		
153	111.0	90.7	136.1
152	50.2	41.1	61.7







#53
3-Nitroaniline
Concen: 4.902 ng
RT: 9.857 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:138 Resp: 39761

Ion	Ratio	Lower	Upper
138	100		
108	12.6	9.3	13.9
92	118.6	93.5	140.3







#55
Dibenzofuran
Concen: 5.481 ng
RT: 10.116 min Scan# 1364
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:168 Resp: 252715

Ion	Ratio	Lower	Upper
168	100		
139	39.8	33.9	50.9
169	13.6	11.0	16.4







#56
4-Nitrophenol
Concen: 4.309 ng
RT: 10.004 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:139 Resp: 26252

Ion	Ratio	Lower	Upper
139	100		
109	81.1	61.6	101.6
65	100.7	79.4	119.4







#57
2,4-Dinitrotoluene
Concen: 4.800 ng
RT: 10.092 min Scan# 1360
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:165 Resp: 49913

Ion	Ratio	Lower	Upper
165	100		
63	41.2	35.4	53.2
89	70.9	56.6	84.8
182	2.8	2.2	3.4







#58
Fluorene
Concen: 5.542 ng
RT: 10.457 min Scan# 14
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:166 Resp: 197266

Ion	Ratio	Lower	Upper
166	100		
165	98.0	79.0	118.4
167	13.2	11.0	16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 4.839 ng
RT: 10.233 min Scan# 1384
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:232 Resp: 48472

Ion	Ratio	Lower	Upper
232	100		
131	44.9	37.3	55.9
130	2.1	1.9	2.9
166	30.3	24.8	37.2







#60
Diethylphthalate
Concen: 5.286 ng
RT: 10.327 min Scan# 14
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 201554

Ion	Ratio	Lower	Upper
149	100		
177	22.1	18.0	27.0
150	12.1	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 5.380 ng
RT: 10.451 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:204 Resp: 99747

Ion	Ratio	Lower	Upper
204	100		
206	31.1	26.8	40.2
141	64.5	52.4	78.6







#62
4-Nitroaniline
Concen: 4.945 ng
RT: 10.463 min Scan# 14
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:138 Resp: 38863

Ion	Ratio	Lower	Upper
138	100		
92	50.5	32.6	72.6
108	77.1	65.0	105.0







#63
Azobenzene
Concen: 5.370 ng
RT: 10.610 min Scan# 1448
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 77 Resp: 190233

Ion	Ratio	Lower	Upper
77	100		
182	25.8	6.6	46.6
105	21.0	1.3	41.3
51	29.1	10.4	50.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.398 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:188 Resp: 1002901

Ion	Ratio	Lower	Upper
188	100		
94	9.0	6.8	10.2
80	10.2	7.6	11.4







#66
n-Nitrosodiphenylamine
Concen: 5.223 ng
RT: 10.569 min Scan# 1441
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:169 Resp: 169225

Ion	Ratio	Lower	Upper
169	100		
168	65.7	53.7	80.5
167	35.5	29.2	43.8







#67
4-Bromophenyl-phenylether
Concen: 4.823 ng
RT: 10.945 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:248 Resp: 60238

Ion	Ratio	Lower	Upper
248	100		
250	97.1	77.8	116.8
141	74.7	62.6	93.8







#68
Hexachlorobenzene
Concen: 4.753 ng
RT: 11.004 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:284 Resp: 65620

Ion	Ratio	Lower	Upper
284	100		
142	35.6	27.0	40.6
249	31.9	25.2	37.8







#69
Atrazine
Concen: 6.867 ng
RT: 11.092 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:200 Resp: 54251

Ion	Ratio	Lower	Upper
200	100		
173	25.1	4.7	44.7
215	47.1	25.4	65.4







#70
Pentachlorophenol
Concen: 3.843 ng
RT: 11.198 min Scan# 1548
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:266 Resp: 32716

Ion	Ratio	Lower	Upper
266	100		
268	64.2	51.0	76.6
264	62.8	50.2	75.2







#71
Phenanthrene
Concen: 5.340 ng
RT: 11.421 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:178 Resp: 289580

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.8	23.8
179	15.3	12.6	19.0







#72
Anthracene
Concen: 5.407 ng
RT: 11.474 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:178 Resp: 294198

Ion	Ratio	Lower	Upper
178	100		
176	18.1	15.1	22.7
179	15.1	12.6	19.0







#73
Carbazole
Concen: 5.329 ng
RT: 11.627 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:167 Resp: 249634

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.4	26.2
139	13.3	11.0	16.6







#74
Di-n-butylphthalate
Concen: 5.352 ng
RT: 11.957 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:149 Resp: 330999

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.6	11.4
104	5.0	4.4	6.6







#75
Fluoranthene
Concen: 5.432 ng
RT: 12.610 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:202 Resp: 309273

Ion	Ratio	Lower	Upper
202	100		
101	9.8	0.0	29.7
203	17.2	0.0	37.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.033 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:240 Resp: 745624

Ion	Ratio	Lower	Upper
240	100		
120	10.4	8.4	12.6
236	25.3	20.5	30.7







#77
Benzidine
Concen: 6.138 ng
RT: 12.733 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:184 Resp: 64673

Ion	Ratio	Lower	Upper
184	100		
185	14.0	11.5	17.3
183	12.3	9.7	14.5







#78
Pyrene
Concen: 4.941 ng
RT: 12.839 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:202 Resp: 315969

Ion	Ratio	Lower	Upper
202	100		
200	20.9	17.3	25.9
203	17.5	14.4	21.6







#79
Terphenyl-d14
Concen: 9.924 ng
RT: 12.980 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:244 Resp: 500607

Ion	Ratio	Lower	Upper
244	100		
212	6.6	6.0	9.0
122	9.7	7.7	11.5







#80
Butylbenzylphthalate
Concen: 4.653 ng
RT: 13.457 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:149 Resp: 117256

Ion	Ratio	Lower	Upper
149	100		
91	71.4	56.5	84.7
206	21.4	17.3	25.9







#81
Benzo(a)anthracene
Concen: 5.205 ng
RT: 14.021 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:228 Resp: 255308

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.8	32.8
229	19.6	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 4.865 ng
RT: 13.986 min Scan# 2022
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:252 Resp: 68969

Ion	Ratio	Lower	Upper
252	100		
254	64.0	51.0	76.6
126	10.7	8.4	12.6







#83
Chrysene
Concen: 4.837 ng
RT: 14.062 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:228 Resp: 213144

Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.1	36.1
229	19.9	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 4.836 ng
RT: 14.015 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:149 Resp: 167165

Ion	Ratio	Lower	Upper
149	100		
167	26.0	21.4	32.0
279	4.4	3.8	5.6







#85
Di-n-octyl phthalate
Concen: 4.435 ng
RT: 14.627 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 213395

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	8.4	6.9	10.3







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.504 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:264 Resp: 527129

Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.1	28.7
265	21.7	17.5	26.3







#87
Indeno(1,2,3-cd)pyrene
Concen: 4.228 ng
RT: 16.980 min Scan# 2
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:276 Resp: 144536

Ion	Ratio	Lower	Upper
276	100		
138	23.7	18.4	27.6
277	25.1	20.3	30.5







#88
Benzo(b)fluoranthene
Concen: 5.234 ng
RT: 15.068 min Scan# 2206
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:252 Resp: 187154

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.5	26.3
125	9.1	7.0	10.4







#89
Benzo(k)fluoranthene
Concen: 4.903 ng
RT: 15.098 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 153142

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	9.4	7.2	10.8







#90
Benzo(a)pyrene
Concen: 4.860 ng
RT: 15.439 min Scan# 2269
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:252 Resp: 137043

Ion	Ratio	Lower	Upper
252	100		
253	21.1	17.4	26.0
125	10.4	7.8	11.8







#91
Dibenzo(a,h)anthracene
Concen: 4.252 ng
RT: 16.998 min Scan# 2
Delta R.T. -0.024 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:278 Resp: 119937

Ion	Ratio	Lower	Upper
278	100		
139	14.8	11.4	17.2
279	23.9	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 4.294 ng
RT: 17.427 min Scan# 2607
Delta R.T. -0.024 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:276 Resp: 120142

Ion	Ratio	Lower	Upper
276	100		
277	23.1	18.7	28.1
138	19.6	15.8	23.8



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141899.D
Acq On : 10 Mar 2025 12:00
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Mar 10 15:33:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	224499	20.000	ng		0.00
21) Naphthalene-d8	8.163	136	921725	20.000	ng		0.00
39) Acenaphthene-d10	9.916	164	541295	20.000	ng		0.00
64) Phenanthrene-d10	11.398	188	966771	20.000	ng		0.00
76) Chrysene-d12	14.039	240	703100	20.000	ng		0.00
86) Perylene-d12	15.510	264	510790	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.487	112	279139	20.746	ng		0.00
7) Phenol-d6	6.493	99	357672	20.872	ng		-0.01
23) Nitrobenzene-d5	7.434	82	328098	19.951	ng		-0.01
42) 2,4,6-Tribromophenol	10.698	330	127785	18.630	ng		0.00
45) 2-Fluorobiphenyl	9.234	172	746589	21.003	ng		0.00
79) Terphenyl-d14	12.980	244	920159	19.345	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.675	88	54068	9.588	ng		98
3) Pyridine	3.440	79	135148	9.752	ng		99
4) n-Nitrosodimethylamine	3.369	42	62411	9.474	ng	#	98
6) Aniline	6.540	93	179854	10.631	ng		99
8) 2-Chlorophenol	6.657	128	151998	10.192	ng		97
9) Benzaldehyde	6.428	77	114586	11.963	ng		99
10) Phenol	6.504	94	187967	10.396	ng		98
11) bis(2-Chloroethyl)ether	6.610	93	139813	10.321	ng		99
12) 1,3-Dichlorobenzene	6.822	146	167192	10.374	ng		98
13) 1,4-Dichlorobenzene	6.898	146	168717	10.347	ng		98
14) 1,2-Dichlorobenzene	7.051	146	159032	10.357	ng		98
15) Benzyl Alcohol	7.010	79	140530	10.132	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.151	45	171225	10.487	ng		98
17) 2-Methylphenol	7.116	107	117869	9.945	ng		99
18) Hexachloroethane	7.392	117	61809	10.111	ng		100
19) n-Nitroso-di-n-propyla...	7.281	70	112281	10.194	ng		97
20) 3+4-Methylphenols	7.269	107	156526	10.296	ng		92
22) Acetophenone	7.281	105	227350	10.282	ng		99
24) Nitrobenzene	7.457	77	161777	9.922	ng		99
25) Isophorone	7.692	82	285760	9.842	ng		99
26) 2-Nitrophenol	7.775	139	73301	9.174	ng		98
27) 2,4-Dimethylphenol	7.804	122	107669	9.784	ng		99
28) bis(2-Chloroethoxy)met...	7.904	93	187307	10.298	ng		98
29) 2,4-Dichlorophenol	8.010	162	135772	9.941	ng		98
30) 1,2,4-Trichlorobenzene	8.098	180	149790	9.941	ng		99
31) Naphthalene	8.181	128	495470	10.493	ng		99
32) Benzoic acid	7.869	122	68640	7.136	ng		98
33) 4-Chloroaniline	8.228	127	173977	10.441	ng		98
34) Hexachlorobutadiene	8.298	225	97298	9.931	ng		98
35) Caprolactam	8.563	113	41208	9.922	ng		97
36) 4-Chloro-3-methylphenol	8.698	107	152340	10.029	ng		96
37) 2-Methylnaphthalene	8.869	142	326733	10.313	ng		99
38) 1-Methylnaphthalene	8.969	142	318324	10.376	ng		98
40) 1,2,4,5-Tetrachloroben...	9.034	216	163481	9.905	ng		99
41) Hexachlorocyclopentadiene	9.022	237	59769	9.177	ng		99
43) 2,4,6-Trichlorophenol	9.145	196	105191	9.777	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141899.D
Acq On : 10 Mar 2025 12:00
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Mar 10 15:33:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.181	196	105943	9.692	ng		99
46) 1,1'-Biphenyl	9.334	154	425640	10.372	ng		98
47) 2-Chloronaphthalene	9.357	162	308729	10.094	ng		98
48) 2-Nitroaniline	9.451	65	84854	9.658	ng		98
49) Acenaphthylene	9.775	152	470801	10.383	ng		99
50) Dimethylphthalate	9.628	163	382722	10.176	ng		100
51) 2,6-Dinitrotoluene	9.692	165	75577	9.663	ng		97
52) Acenaphthene	9.945	154	323406	10.101	ng		100
53) 3-Nitroaniline	9.857	138	79944	9.955	ng		96
54) 2,4-Dinitrophenol	9.963	184	23121	11.386	ng	#	71
55) Dibenzofuran	10.116	168	483596	10.593	ng		98
56) 4-Nitrophenol	10.004	139	58549	9.706	ng		98
57) 2,4-Dinitrotoluene	10.092	165	103313	10.035	ng		98
58) Fluorene	10.463	166	368370	10.452	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.233	232	98725	9.954	ng		99
60) Diethylphthalate	10.328	149	398003	10.542	ng		100
61) 4-Chlorophenyl-phenyle...	10.451	204	186237	10.145	ng		99
62) 4-Nitroaniline	10.463	138	78048	10.030	ng		97
63) Azobenzene	10.610	77	368982	10.521	ng		99
65) 4,6-Dinitro-2-methylph...	10.498	198	40348	7.040	ng		99
66) n-Nitrosodiphenylamine	10.569	169	314758	10.078	ng		99
67) 4-Bromophenyl-phenylether	10.945	248	120534	10.011	ng		97
68) Hexachlorobenzene	11.010	284	129242	9.712	ng		98
69) Atrazine	11.092	200	102822	13.502	ng		99
70) Pentachlorophenol	11.198	266	73112	8.909	ng		99
71) Phenanthrene	11.422	178	556546	10.646	ng		98
72) Anthracene	11.475	178	550767	10.501	ng		100
73) Carbazole	11.628	167	480520	10.641	ng		99
74) Di-n-butylphthalate	11.957	149	638429	10.708	ng		99
75) Fluoranthene	12.610	202	592806	10.801	ng		99
77) Benzidine	12.733	184	108761	10.946	ng		99
78) Pyrene	12.839	202	580442	9.626	ng		99
80) Butylbenzylphthalate	13.457	149	230714	9.710	ng		98
81) Benzo(a)anthracene	14.027	228	458422	9.911	ng		99
82) 3,3'-Dichlorobenzidine	13.986	252	136843	10.236	ng		99
83) Chrysene	14.063	228	419687	10.100	ng		99
84) Bis(2-ethylhexyl)phtha...	14.016	149	320536	9.833	ng		99
85) Di-n-octyl phthalate	14.627	149	428377	9.442	ng		100
87) Indeno(1,2,3-cd)pyrene	16.986	276	291723	8.807	ng		99
88) Benzo(b)fluoranthene	15.074	252	334664	9.659	ng		99
89) Benzo(k)fluoranthene	15.104	252	323772	10.698	ng		99
90) Benzo(a)pyrene	15.439	252	266149	9.740	ng		99
91) Dibenzo(a,h)anthracene	17.004	278	243443	8.907	ng		98
92) Benzo(g,h,i)perylene	17.433	276	242380	8.941	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141899.D
Acq On : 10 Mar 2025 12:00
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Mar 10 15:33:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 224499

Ion	Ratio	Lower	Upper
152	100		
150	157.0	127.4	191.2
115	63.6	51.9	77.9







#2
1,4-Dioxane
Concen: 9.588 ng
RT: 2.675 min Scan# 99
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion: 88 Resp: 54068

Ion	Ratio	Lower	Upper
88	100		
58	72.9	57.4	86.2
43	28.8	21.5	32.3







#3
Pyridine
Concen: 9.752 ng
RT: 3.440 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 135148

Ion	Ratio	Lower	Upper
79	100		
52	66.0	53.3	79.9
51	34.7	27.6	41.4







#4
n-Nitrosodimethylamine
Concen: 9.474 ng
RT: 3.369 min Scan# 217
Delta R.T. -0.023 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion: 42 Resp: 62411

Ion	Ratio	Lower	Upper
42	100		
74	131.7	103.4	155.0
44	7.7	5.0	7.4#







#5
2-Fluorophenol
Concen: 20.746 ng
RT: 5.487 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:112 Resp: 279139

Ion	Ratio	Lower	Upper
112	100		
64	58.3	46.5	69.7
63	31.6	25.4	38.2







#6
Aniline
Concen: 10.631 ng
RT: 6.540 min Scan# 756
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion: 93 Resp: 179854

Ion	Ratio	Lower	Upper
93	100		
66	39.1	31.9	47.9
65	19.4	15.8	23.6







#7
Phenol-d6
Concen: 20.872 ng
RT: 6.493 min Scan# 7
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 99 Resp: 357672

Ion	Ratio	Lower	Upper
99	100		
42	17.6	14.6	21.8
71	36.0	30.8	46.2







#8
2-Chlorophenol
Concen: 10.192 ng
RT: 6.657 min Scan# 776
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:128 Resp: 151998

Ion	Ratio	Lower	Upper
128	100		
130	33.1	12.8	52.8
64	47.3	24.4	64.4







#9
Benzaldehyde
Concen: 11.963 ng
RT: 6.428 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 114586

Ion	Ratio	Lower	Upper
77	100		
105	98.6	80.0	120.0
106	94.3	74.5	114.5







#10
Phenol
Concen: 10.396 ng
RT: 6.504 min Scan# 750
Delta R.T. -0.017 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion: 94 Resp: 187967

Ion	Ratio	Lower	Upper
94	100		
65	29.9	10.5	50.5
66	43.8	25.8	65.8







#11
bis(2-Chloroethyl)ether
Concen: 10.321 ng
RT: 6.610 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 139813

Ion	Ratio	Lower	Upper
93	100		
63	73.7	53.0	93.0
95	33.0	12.9	52.9







#12
1,3-Dichlorobenzene
Concen: 10.374 ng
RT: 6.822 min Scan# 804
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:146 Resp: 167192

Ion	Ratio	Lower	Upper
146	100		
148	63.5	51.3	76.9
75	30.8	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 10.347 ng
RT: 6.898 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:146 Resp: 168717

Ion	Ratio	Lower	Upper
146	100		
148	63.3	51.0	76.6
111	42.1	35.7	53.5







#14
1,2-Dichlorobenzene
Concen: 10.357 ng
RT: 7.051 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:146 Resp: 159032

Ion	Ratio	Lower	Upper
146	100		
148	64.4	50.8	76.2
111	45.8	38.0	57.0







#15
Benzyl Alcohol
Concen: 10.132 ng
RT: 7.010 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 140530

Ion	Ratio	Lower	Upper
79	100		
108	75.1	59.7	89.5
77	63.3	51.0	76.6







#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.487 ng
RT: 7.151 min Scan# 860
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion: 45 Resp: 171225

Ion	Ratio	Lower	Upper
45	100		
77	16.9	0.0	37.8
79	13.7	0.0	34.2







#17
2-Methylphenol
Concen: 9.945 ng
RT: 7.116 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:107 Resp: 117869

Ion	Ratio	Lower	Upper
107	100		
108	111.2	90.2	135.2
77	48.5	39.6	59.4
79	49.6	39.1	58.7







#18
Hexachloroethane
Concen: 10.111 ng
RT: 7.392 min Scan# 901
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:117 Resp: 61809

Ion	Ratio	Lower	Upper
117	100		
119	96.1	76.9	115.3
201	99.9	79.9	119.9







#19
n-Nitroso-di-n-propylamine
Concen: 10.194 ng
RT: 7.281 min Scan# 8
Delta R.T. -0.018 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 70 Resp: 112281

Ion	Ratio	Lower	Upper
70	100		
42	54.9	41.4	62.0
101	10.2	8.2	12.2
130	22.0	18.2	27.4







#20
3+4-Methylphenols
Concen: 10.296 ng
RT: 7.269 min Scan# 880
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:107 Resp: 156526

Ion	Ratio	Lower	Upper
107	100		
108	89.9	68.4	108.4
77	75.0	69.7	109.7
79	30.8	11.4	51.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:136 Resp: 921725

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.8	13.2
54	7.0	5.8	8.8
68	5.0	4.1	6.1







#22
Acetophenone
Concen: 10.282 ng
RT: 7.281 min Scan# 882
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:105 Resp: 227350

Ion	Ratio	Lower	Upper
105	100		
71	8.0	5.8	8.8
51	30.3	23.4	35.0
120	21.8	17.3	25.9







#23
Nitrobenzene-d5
Concen: 19.951 ng
RT: 7.434 min Scan# 9(
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 328098

Ion	Ratio	Lower	Upper
82	100		
128	42.5	36.0	54.0
54	49.6	39.6	59.4







#24
Nitrobenzene
Concen: 9.922 ng
RT: 7.457 min Scan# 912
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion: 77 Resp: 161777

Ion	Ratio	Lower	Upper
77	100		
123	46.3	37.0	55.4
65	13.0	11.3	16.9







#25
Isophorone
Concen: 9.842 ng
RT: 7.692 min Scan# 9
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 285760

Ion	Ratio	Lower	Upper
82	100		
95	7.8	6.4	9.6
138	18.2	14.9	22.3







#26
2-Nitrophenol
Concen: 9.174 ng
RT: 7.775 min Scan# 966
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:139 Resp: 73301

Ion	Ratio	Lower	Upper
139	100		
109	32.2	26.5	39.7
65	49.3	40.3	60.5







#27
2,4-Dimethylphenol
Concen: 9.784 ng
RT: 7.804 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:122 Resp: 107669

Ion	Ratio	Lower	Upper
122	100		
107	120.0	94.8	142.2
121	59.5	47.8	71.6







#28
bis(2-Chloroethoxy)methane
Concen: 10.298 ng
RT: 7.904 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion: 93 Resp: 187307

Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.6	40.0
123	11.6	10.2	15.4







#29
2,4-Dichlorophenol
Concen: 9.941 ng
RT: 8.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 135772

Ion	Ratio	Lower	Upper
162	100		
164	63.7	43.8	83.8
98	35.6	18.1	58.1







#30
1,2,4-Trichlorobenzene
Concen: 9.941 ng
RT: 8.098 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:180 Resp: 149790

Ion	Ratio	Lower	Upper
180	100		
182	96.1	77.5	116.3
145	31.4	25.8	38.6







#31
Naphthalene
Concen: 10.493 ng
RT: 8.181 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:128 Resp: 495470

Ion	Ratio	Lower	Upper
128	100		
129	10.9	9.0	13.6
127	13.4	10.8	16.2







#32
Benzoic acid
Concen: 7.136 ng
RT: 7.869 min Scan# 982
Delta R.T. -0.059 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:122 Resp: 68640

Ion	Ratio	Lower	Upper
122	100		
105	124.8	107.9	147.9
77	97.8	79.0	119.0







#33
4-Chloroaniline
Concen: 10.441 ng
RT: 8.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:127 Resp: 173977

Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.2	39.4
65	30.3	25.4	38.2
92	19.5	16.2	24.4







#34
Hexachlorobutadiene
Concen: 9.931 ng
RT: 8.298 min Scan# 1055
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:225 Resp: 97298

Ion	Ratio	Lower	Upper
225	100		
223	62.0	50.6	75.8
227	62.7	52.2	78.2







#35
Caprolactam
Concen: 9.922 ng
RT: 8.563 min Scan# 11
Delta R.T. -0.041 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:113 Resp: 41208

Ion	Ratio	Lower	Upper
113	100		
55	151.6	129.6	169.6
56	119.3	104.0	144.0







#36
4-Chloro-3-methylphenol
Concen: 10.029 ng
RT: 8.698 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:107 Resp: 152340

Ion	Ratio	Lower	Upper
107	100		
144	25.1	19.0	28.6
142	77.3	59.2	88.8







#37
2-Methylnaphthalene
Concen: 10.313 ng
RT: 8.869 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:142 Resp: 326733

Ion	Ratio	Lower	Upper
142	100		
141	89.6	71.2	106.8







#38
1-Methylnaphthalene
Concen: 10.376 ng
RT: 8.969 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:142 Resp: 318324

Ion	Ratio	Lower	Upper
142	100		
141	91.1	74.2	111.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:164 Resp: 541295

Ion	Ratio	Lower	Upper
164	100		
162	102.7	81.8	122.6
160	45.9	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 9.905 ng
RT: 9.034 min Scan# 1180
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:216 Resp: 163481

Ion	Ratio	Lower	Upper
216	100		
214	77.1	61.9	92.9
179	21.5	18.1	27.1
108	18.3	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 9.177 ng
RT: 9.022 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:237 Resp: 59769

Ion	Ratio	Lower	Upper
237	100		
235	62.1	43.1	83.1
272	12.0	0.0	32.0







#42
2,4,6-Tribromophenol
Concen: 18.630 ng
RT: 10.698 min Scan# 1463
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:330 Resp: 127785

Ion	Ratio	Lower	Upper
330	100		
332	97.8	77.6	116.4
141	33.6	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 9.777 ng
RT: 9.145 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:196 Resp: 105191

Ion	Ratio	Lower	Upper
196	100		
198	99.4	80.6	121.0
200	31.9	25.8	38.6







#44
2,4,5-Trichlorophenol
Concen: 9.692 ng
RT: 9.181 min Scan# 1205
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:196 Resp: 105943

Ion	Ratio	Lower	Upper
196	100		
198	102.0	80.7	121.1
97	53.7	42.1	63.1
132	29.7	23.4	35.2







#45
2-Fluorobiphenyl
Concen: 21.003 ng
RT: 9.234 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:172 Resp: 746589

Ion	Ratio	Lower	Upper
172	100		
171	34.9	29.3	43.9
170	23.1	19.4	29.0







#46
1,1'-Biphenyl
Concen: 10.372 ng
RT: 9.334 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:154 Resp: 425640

Ion	Ratio	Lower	Upper
154	100		
153	39.6	21.1	61.1
76	11.1	0.0	31.3







#47
2-Chloronaphthalene
Concen: 10.094 ng
RT: 9.357 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 308729

Ion	Ratio	Lower	Upper
162	100		
127	42.1	35.5	53.3
164	32.9	26.6	40.0







#48
2-Nitroaniline
Concen: 9.658 ng
RT: 9.451 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion: 65 Resp: 84854

Ion	Ratio	Lower	Upper
65	100		
92	72.5	56.6	84.8
138	109.4	88.6	133.0







#49
Acenaphthylene
Concen: 10.383 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 470801

Ion	Ratio	Lower	Upper
152	100		
151	20.0	16.3	24.5
153	13.7	11.1	16.7







#50
Dimethylphthalate
Concen: 10.176 ng
RT: 9.628 min Scan# 1281
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:163 Resp: 382722

Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.1	4.7
164	10.3	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 9.663 ng
RT: 9.692 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 75577

Ion	Ratio	Lower	Upper
165	100		
63	56.0	47.0	70.4
89	57.4	47.4	71.2







#52
Acenaphthene
Concen: 10.101 ng
RT: 9.945 min Scan# 1335
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:154 Resp: 323406

Ion	Ratio	Lower	Upper
154	100		
153	113.6	90.7	136.1
152	50.9	41.1	61.7







#53
3-Nitroaniline
Concen: 9.955 ng
RT: 9.857 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:138 Resp: 79944

Ion	Ratio	Lower	Upper
138	100		
108	12.2	9.3	13.9
92	121.1	93.5	140.3







#54
2,4-Dinitrophenol
Concen: 11.386 ng
RT: 9.963 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:184 Resp: 23121

Ion	Ratio	Lower	Upper
184	100		
63	75.2	60.2	90.4
154	222.1	132.9	199.3#







#55
Dibenzofuran
Concen: 10.593 ng
RT: 10.116 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:168 Resp: 483596

Ion	Ratio	Lower	Upper
168	100		
139	40.9	33.9	50.9
169	13.2	11.0	16.4







#56
4-Nitrophenol
Concen: 9.706 ng
RT: 10.004 min Scan# 1345
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:139 Resp: 58549

Ion	Ratio	Lower	Upper
139	100		
109	78.8	61.6	101.6
65	97.5	79.4	119.4







#57
2,4-Dinitrotoluene
Concen: 10.035 ng
RT: 10.092 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 103313

Ion	Ratio	Lower	Upper
165	100		
63	41.3	35.4	53.2
89	70.9	56.6	84.8
182	3.0	2.2	3.4







#58
Fluorene
Concen: 10.452 ng
RT: 10.463 min Scan# 1423
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:166 Resp: 368370

Ion	Ratio	Lower	Upper
166	100		
165	97.8	79.0	118.4
167	13.3	11.0	16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 9.954 ng
RT: 10.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:232 Resp: 98725

Ion	Ratio	Lower	Upper
232	100		
131	46.2	37.3	55.9
130	2.3	1.9	2.9
166	30.1	24.8	37.2







#60
Diethylphthalate
Concen: 10.542 ng
RT: 10.328 min Scan# 1400
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:149 Resp: 398003

Ion	Ratio	Lower	Upper
149	100		
177	22.2	18.0	27.0
150	12.7	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 10.145 ng
RT: 10.451 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:204 Resp: 186237

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.8	40.2
141	65.3	52.4	78.6







#62
4-Nitroaniline
Concen: 10.030 ng
RT: 10.463 min Scan# 1423
Delta R.T. -0.017 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:138 Resp: 78048

Ion	Ratio	Lower	Upper
138	100		
92	51.4	32.6	72.6
108	81.1	65.0	105.0







#63
Azobenzene
Concen: 10.521 ng
RT: 10.610 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 368982

Ion	Ratio	Lower	Upper
77	100		
182	26.6	6.6	46.6
105	21.3	1.3	41.3
51	29.6	10.4	50.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.398 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:188 Resp: 966771

Ion	Ratio	Lower	Upper
188	100		
94	9.1	6.8	10.2
80	10.0	7.6	11.4







#65
4,6-Dinitro-2-methylphenol
Concen: 7.040 ng
RT: 10.498 min Scan# 14
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:198 Resp: 40348

Ion	Ratio	Lower	Upper
198	100		
51	42.7	22.3	62.3
105	44.3	23.0	63.0







#66
n-Nitrosodiphenylamine
Concen: 10.078 ng
RT: 10.569 min Scan# 1441
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:169 Resp: 314758

Ion	Ratio	Lower	Upper
169	100		
168	65.8	53.7	80.5
167	35.8	29.2	43.8







#67
4-Bromophenyl-phenylether
Concen: 10.011 ng
RT: 10.945 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:248 Resp: 120534

Ion	Ratio	Lower	Upper
248	100		
250	96.8	77.8	116.8
141	72.9	62.6	93.8







#68
Hexachlorobenzene
Concen: 9.712 ng
RT: 11.010 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:284 Resp: 129242

Ion	Ratio	Lower	Upper
284	100		
142	31.4	27.0	40.6
249	31.1	25.2	37.8







#69
Atrazine
Concen: 13.502 ng
RT: 11.092 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:200 Resp: 102822

Ion	Ratio	Lower	Upper
200	100		
173	24.3	4.7	44.7
215	45.0	25.4	65.4







#70
Pentachlorophenol
Concen: 8.909 ng
RT: 11.198 min Scan# 1548
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:266 Resp: 73112

Ion	Ratio	Lower	Upper
266	100		
268	64.0	51.0	76.6
264	64.3	50.2	75.2







#71
Phenanthrene
Concen: 10.646 ng
RT: 11.422 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:178 Resp: 556546

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.8	23.8
179	15.2	12.6	19.0







#72
Anthracene
Concen: 10.501 ng
RT: 11.475 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:178 Resp: 550767

Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.1	22.7
179	15.7	12.6	19.0







#73
Carbazole
Concen: 10.641 ng
RT: 11.628 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:167 Resp: 480520

Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.4	26.2
139	13.3	11.0	16.6







#74
Di-n-butylphthalate
Concen: 10.708 ng
RT: 11.957 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:149 Resp: 638429

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.6	11.4
104	5.2	4.4	6.6







#75
Fluoranthene
Concen: 10.801 ng
RT: 12.610 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:202 Resp: 592806

Ion	Ratio	Lower	Upper
202	100		
101	10.1	0.0	29.7
203	17.4	0.0	37.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.039 min Scan# 2031
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:240 Resp: 703100

Ion	Ratio	Lower	Upper
240	100		
120	9.7	8.4	12.6
236	25.5	20.5	30.7







#77
Benzidine
Concen: 10.946 ng
RT: 12.733 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:184 Resp: 108761

Ion	Ratio	Lower	Upper
184	100		
185	13.9	11.5	17.3
183	11.6	9.7	14.5







#78
Pyrene
Concen: 9.626 ng
RT: 12.839 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:202 Resp: 580442

Ion	Ratio	Lower	Upper
202	100		
200	20.9	17.3	25.9
203	17.6	14.4	21.6







#79
Terphenyl-d14
Concen: 19.345 ng
RT: 12.980 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:244 Resp: 920159

Ion	Ratio	Lower	Upper
244	100		
212	6.9	6.0	9.0
122	10.3	7.7	11.5







#80
Butylbenzylphthalate
Concen: 9.710 ng
RT: 13.457 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:149 Resp: 230714

Ion	Ratio	Lower	Upper
149	100		
91	71.8	56.5	84.7
206	20.7	17.3	25.9







#81
Benzo(a)anthracene
Concen: 9.911 ng
RT: 14.027 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 458422

Ion	Ratio	Lower	Upper
228	100		
226	26.6	21.8	32.8
229	19.6	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 10.236 ng
RT: 13.986 min Scan# 2022
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:252 Resp: 136843

Ion	Ratio	Lower	Upper
252	100		
254	64.2	51.0	76.6
126	10.7	8.4	12.6







#83
Chrysene
Concen: 10.100 ng
RT: 14.063 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 419687

Ion	Ratio	Lower	Upper
228	100		
226	29.4	24.1	36.1
229	19.4	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 9.833 ng
RT: 14.016 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:149 Resp: 320536

Ion	Ratio	Lower	Upper
149	100		
167	26.4	21.4	32.0
279	4.1	3.8	5.6







#85
Di-n-octyl phthalate
Concen: 9.442 ng
RT: 14.627 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:149 Resp: 428377

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.6	6.9	10.3







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.510 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:264 Resp: 510790

Ion	Ratio	Lower	Upper
264	100		
260	23.1	19.1	28.7
265	21.2	17.5	26.3







#87
Indeno(1,2,3-cd)pyrene
Concen: 8.807 ng
RT: 16.986 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:276 Resp: 291723

Ion	Ratio	Lower	Upper
276	100		
138	23.8	18.4	27.6
277	25.3	20.3	30.5







#88
Benzo(b)fluoranthene
Concen: 9.659 ng
RT: 15.074 min Scan# 2207
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:252 Resp: 334664

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.5	26.3
125	9.1	7.0	10.4







#89
Benzo(k)fluoranthene
Concen: 10.698 ng
RT: 15.104 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:252 Resp: 323772

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	9.3	7.2	10.8







#90
Benzo(a)pyrene
Concen: 9.740 ng
RT: 15.439 min Scan# 2269
Delta R.T. -0.012 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:252 Resp: 266149

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.0
125	10.7	7.8	11.8







#91
Dibenzo(a,h)anthracene
Concen: 8.907 ng
RT: 17.004 min Scan# 2
Delta R.T. -0.018 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 243443

Ion	Ratio	Lower	Upper
278	100		
139	15.5	11.4	17.2
279	23.0	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 8.941 ng
RT: 17.433 min Scan# 2608
Delta R.T. -0.018 min
Lab File: BF141899.D
Acq: 10 Mar 2025 12:00

Tgt Ion:276 Resp: 242380

Ion	Ratio	Lower	Upper
276	100		
277	23.3	18.7	28.1
138	19.2	15.8	23.8



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141900.D
Acq On : 10 Mar 2025 12:29
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Mar 10 15:34:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	198053	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	797548	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	470833	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	826085	20.000	ng	0.00
76) Chrysene-d12	14.039	240	581591	20.000	ng	0.00
86) Perylene-d12	15.509	264	428960	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	514189	43.318	ng	0.00
7) Phenol-d6	6.498	99	651272	43.079	ng	0.00
23) Nitrobenzene-d5	7.439	82	604270	42.465	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	243832	40.869	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	1298709	42.002	ng	0.00
79) Terphenyl-d14	12.986	244	1582373	40.218	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.669	88	106079	21.323	ng	99
3) Pyridine	3.434	79	264229	21.612	ng	100
4) n-Nitrosodimethylamine	3.375	42	122948	21.155	ng	99
6) Aniline	6.539	93	324802	21.763	ng	99
8) 2-Chlorophenol	6.663	128	284082	21.592	ng	99
9) Benzaldehyde	6.428	77	193024	22.843	ng	98
10) Phenol	6.510	94	346008	21.692	ng	97
11) bis(2-Chloroethyl)ether	6.610	93	253793	21.236	ng	98
12) 1,3-Dichlorobenzene	6.822	146	304632	21.426	ng	98
13) 1,4-Dichlorobenzene	6.898	146	306665	21.319	ng	100
14) 1,2-Dichlorobenzene	7.051	146	289199	21.349	ng	99
15) Benzyl Alcohol	7.016	79	261893	21.402	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.151	45	303847	21.095	ng	99
17) 2-Methylphenol	7.122	107	222173	21.248	ng	97
18) Hexachloroethane	7.392	117	114019	21.141	ng	99
19) n-Nitroso-di-n-propyla...	7.286	70	206654	21.268	ng	97
20) 3+4-Methylphenols	7.275	107	291065	21.702	ng	99
22) Acetophenone	7.286	105	409149	21.385	ng	99
24) Nitrobenzene	7.457	77	299046	21.196	ng	99
25) Isophorone	7.698	82	522239	20.788	ng	100
26) 2-Nitrophenol	7.775	139	144979	20.970	ng	99
27) 2,4-Dimethylphenol	7.804	122	200058	21.009	ng	100
28) bis(2-Chloroethoxy)met...	7.904	93	331966	21.093	ng	99
29) 2,4-Dichlorophenol	8.010	162	246810	20.884	ng	100
30) 1,2,4-Trichlorobenzene	8.098	180	272773	20.922	ng	97
31) Naphthalene	8.181	128	877897	21.487	ng	99
32) Benzoic acid	7.892	122	152420	18.313	ng	99
33) 4-Chloroaniline	8.228	127	313249	21.726	ng	98
34) Hexachlorobutadiene	8.298	225	176253	20.791	ng	99
35) Caprolactam	8.580	113	75048	20.884	ng	99
36) 4-Chloro-3-methylphenol	8.698	107	278207	21.168	ng	100
37) 2-Methylnaphthalene	8.875	142	584552	21.324	ng	99
38) 1-Methylnaphthalene	8.975	142	575428	21.677	ng	99
40) 1,2,4,5-Tetrachloroben...	9.039	216	294845	20.538	ng	99
41) Hexachlorocyclopentadiene	9.028	237	117174	20.685	ng	100
43) 2,4,6-Trichlorophenol	9.145	196	189069	20.203	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141900.D
Acq On : 10 Mar 2025 12:29
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Mar 10 15:34:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.180	196	198841	20.912	ng		99
46) 1,1'-Biphenyl	9.333	154	749087	20.985	ng		100
47) 2-Chloronaphthalene	9.363	162	555946	20.898	ng		98
48) 2-Nitroaniline	9.451	65	162790	21.302	ng		97
49) Acenaphthylene	9.775	152	837016	21.222	ng		99
50) Dimethylphthalate	9.633	163	691454	21.137	ng		100
51) 2,6-Dinitrotoluene	9.692	165	143513	21.096	ng		100
52) Acenaphthene	9.951	154	579623	20.813	ng		98
53) 3-Nitroaniline	9.863	138	148977	21.327	ng		98
54) 2,4-Dinitrophenol	9.963	184	58223	20.368	ng	#	33
55) Dibenzofuran	10.122	168	843359	21.238	ng		98
56) 4-Nitrophenol	10.010	139	109958	20.955	ng		100
57) 2,4-Dinitrotoluene	10.098	165	189234	21.132	ng		98
58) Fluorene	10.463	166	650245	21.211	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.233	232	183513	21.272	ng		99
60) Diethylphthalate	10.333	149	704993	21.468	ng		100
61) 4-Chlorophenyl-phenyle...	10.457	204	330489	20.698	ng		99
62) 4-Nitroaniline	10.474	138	142995	21.125	ng		99
63) Azobenzene	10.616	77	660935	21.665	ng		99
65) 4,6-Dinitro-2-methylph...	10.504	198	91852	18.755	ng		98
66) n-Nitrosodiphenylamine	10.569	169	553905	20.756	ng		99
67) 4-Bromophenyl-phenylether	10.945	248	210241	20.436	ng		100
68) Hexachlorobenzene	11.010	284	234399	20.613	ng		99
69) Atrazine	11.092	200	159306	24.481	ng		99
70) Pentachlorophenol	11.198	266	141842	20.227	ng		100
71) Phenanthrene	11.427	178	939782	21.038	ng		99
72) Anthracene	11.474	178	940875	20.993	ng		99
73) Carbazole	11.627	167	813998	21.096	ng		100
74) Di-n-butylphthalate	11.963	149	1089304	21.382	ng		99
75) Fluoranthene	12.610	202	1014464	21.632	ng		100
77) Benzidine	12.733	184	110648	13.463	ng		99
78) Pyrene	12.839	202	1004157	20.133	ng		99
80) Butylbenzylphthalate	13.457	149	408044	20.761	ng		98
81) Benzo(a)anthracene	14.027	228	776120	20.285	ng		99
82) 3,3'-Dichlorobenzidine	13.992	252	221168	19.999	ng		99
83) Chrysene	14.062	228	732966	21.325	ng		100
84) Bis(2-ethylhexyl)phtha...	14.015	149	562062	20.845	ng		99
85) Di-n-octyl phthalate	14.633	149	778373	20.740	ng		100
87) Indeno(1,2,3-cd)pyrene	16.992	276	560913	20.163	ng		99
88) Benzo(b)fluoranthene	15.074	252	636808	21.886	ng		100
89) Benzo(k)fluoranthene	15.104	252	513311	20.196	ng		99
90) Benzo(a)pyrene	15.445	252	481683	20.991	ng		100
91) Dibenzo(a,h)anthracene	17.009	278	460109	20.046	ng		99
92) Benzo(g,h,i)perylene	17.439	276	455034	19.987	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141900.D
Acq On : 10 Mar 2025 12:29
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Mar 10 15:34:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 198053

Ion	Ratio	Lower	Upper
152	100		
150	158.7	127.4	191.2
115	64.5	51.9	77.9







#2
1,4-Dioxane
Concen: 21.323 ng
RT: 2.669 min Scan# 98
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion: 88 Resp: 106079

Ion	Ratio	Lower	Upper
88	100		
58	72.3	57.4	86.2
43	27.4	21.5	32.3







#3
Pyridine
Concen: 21.612 ng
RT: 3.434 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 264229

Ion	Ratio	Lower	Upper
79	100		
52	66.8	53.3	79.9
51	34.7	27.6	41.4







#4
n-Nitrosodimethylamine
Concen: 21.155 ng
RT: 3.375 min Scan# 218
Delta R.T. -0.018 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion: 42 Resp: 122948

Ion	Ratio	Lower	Upper
42	100		
74	129.9	103.4	155.0
44	6.5	5.0	7.4







#5
2-Fluorophenol
Concen: 43.318 ng
RT: 5.487 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:112 Resp: 514189

Ion	Ratio	Lower	Upper
112	100		
64	59.5	46.5	69.7
63	31.4	25.4	38.2







#6
Aniline
Concen: 21.763 ng
RT: 6.539 min Scan# 756
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion: 93 Resp: 324802

Ion	Ratio	Lower	Upper
93	100		
66	40.8	31.9	47.9
65	19.4	15.8	23.6







#7
Phenol-d6
Concen: 43.079 ng
RT: 6.498 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 99 Resp: 651272

Ion	Ratio	Lower	Upper
99	100		
42	17.6	14.6	21.8
71	36.9	30.8	46.2







#8
2-Chlorophenol
Concen: 21.592 ng
RT: 6.663 min Scan# 777
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:128 Resp: 284082

Ion	Ratio	Lower	Upper
128	100		
130	32.6	12.8	52.8
64	45.4	24.4	64.4







#9
Benzaldehyde
Concen: 22.843 ng
RT: 6.428 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 193024

Ion	Ratio	Lower	Upper
77	100		
105	97.1	80.0	120.0
106	92.6	74.5	114.5







#10
Phenol
Concen: 21.692 ng
RT: 6.510 min Scan# 751
Delta R.T. -0.012 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion: 94 Resp: 346008

Ion	Ratio	Lower	Upper
94	100		
65	29.5	10.5	50.5
66	43.6	25.8	65.8







#11
bis(2-Chloroethyl)ether
Concen: 21.236 ng
RT: 6.610 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 93 Resp: 253793

Ion	Ratio	Lower	Upper
93	100		
63	74.8	53.0	93.0
95	32.4	12.9	52.9







#12
1,3-Dichlorobenzene
Concen: 21.426 ng
RT: 6.822 min Scan# 804
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:146 Resp: 304632

Ion	Ratio	Lower	Upper
146	100		
148	65.2	51.3	76.9
75	31.9	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 21.319 ng
RT: 6.898 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:146 Resp: 306665

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.0	76.6
111	44.3	35.7	53.5







#14
1,2-Dichlorobenzene
Concen: 21.349 ng
RT: 7.051 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:146 Resp: 289199

Ion	Ratio	Lower	Upper
146	100		
148	64.4	50.8	76.2
111	46.6	38.0	57.0







#15
Benzyl Alcohol
Concen: 21.402 ng
RT: 7.016 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 261893

Ion	Ratio	Lower	Upper
79	100		
108	73.9	59.7	89.5
77	63.3	51.0	76.6







#16
2,2'-oxybis(1-Chloropropane)
Concen: 21.095 ng
RT: 7.151 min Scan# 860
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion: 45 Resp: 303847

Ion	Ratio	Lower	Upper
45	100		
77	17.3	0.0	37.8
79	13.5	0.0	34.2







#17
2-Methylphenol
Concen: 21.248 ng
RT: 7.122 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:107 Resp: 222173

Ion	Ratio	Lower	Upper
107	100		
108	109.5	90.2	135.2
77	48.1	39.6	59.4
79	47.7	39.1	58.7







#18
Hexachloroethane
Concen: 21.141 ng
RT: 7.392 min Scan# 901
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:117 Resp: 114019

Ion	Ratio	Lower	Upper
117	100		
119	97.8	76.9	115.3
201	100.3	79.9	119.9







#19
n-Nitroso-di-n-propylamine
Concen: 21.268 ng
RT: 7.286 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 70 Resp: 206654

Ion	Ratio	Lower	Upper
70	100		
42	53.7	41.4	62.0
101	10.0	8.2	12.2
130	21.6	18.2	27.4







#20
3+4-Methylphenols
Concen: 21.702 ng
RT: 7.275 min Scan# 881
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:107 Resp: 291065

Ion	Ratio	Lower	Upper
107	100		
108	88.1	68.4	108.4
77	87.9	69.7	109.7
79	30.8	11.4	51.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:136 Resp: 797548

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	7.1	5.8	8.8
68	4.9	4.1	6.1







#22
Acetophenone
Concen: 21.385 ng
RT: 7.286 min Scan# 883
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:105 Resp: 409149

Ion	Ratio	Lower	Upper
105	100		
71	8.2	5.8	8.8
51	28.8	23.4	35.0
120	22.1	17.3	25.9







#23
Nitrobenzene-d5
Concen: 42.465 ng
RT: 7.439 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 604270

Ion	Ratio	Lower	Upper
82	100		
128	43.8	36.0	54.0
54	49.3	39.6	59.4







#24
Nitrobenzene
Concen: 21.196 ng
RT: 7.457 min Scan# 912
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion: 77 Resp: 299046

Ion	Ratio	Lower	Upper
77	100		
123	45.6	37.0	55.4
65	13.7	11.3	16.9







#25
Isophorone
Concen: 20.788 ng
RT: 7.698 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 522239

Ion	Ratio	Lower	Upper
82	100		
95	8.2	6.4	9.6
138	18.5	14.9	22.3







#26
2-Nitrophenol
Concen: 20.970 ng
RT: 7.775 min Scan# 966
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:139 Resp: 144979

Ion	Ratio	Lower	Upper
139	100		
109	32.6	26.5	39.7
65	51.2	40.3	60.5







#27
2,4-Dimethylphenol
Concen: 21.009 ng
RT: 7.804 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:122 Resp: 200058

Ion	Ratio	Lower	Upper
122	100		
107	118.9	94.8	142.2
121	60.1	47.8	71.6







#28
bis(2-Chloroethoxy)methane
Concen: 21.093 ng
RT: 7.904 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion: 93 Resp: 331966

Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.6	40.0
123	12.6	10.2	15.4







#29
2,4-Dichlorophenol
Concen: 20.884 ng
RT: 8.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 246810

Ion	Ratio	Lower	Upper
162	100		
164	64.0	43.8	83.8
98	38.4	18.1	58.1







#30
1,2,4-Trichlorobenzene
Concen: 20.922 ng
RT: 8.098 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:180 Resp: 272773

Ion	Ratio	Lower	Upper
180	100		
182	93.6	77.5	116.3
145	31.9	25.8	38.6







#31
Naphthalene
Concen: 21.487 ng
RT: 8.181 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:128 Resp: 877897

Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.0	13.6
127	13.4	10.8	16.2







#32
Benzoic acid
Concen: 18.313 ng
RT: 7.892 min Scan# 986
Delta R.T. -0.036 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:122 Resp: 152420

Ion	Ratio	Lower	Upper
122	100		
105	127.6	107.9	147.9
77	97.6	79.0	119.0







#33
4-Chloroaniline
Concen: 21.726 ng
RT: 8.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:127 Resp: 313249

Ion	Ratio	Lower	Upper
127	100		
129	31.9	26.2	39.4
65	30.0	25.4	38.2
92	19.5	16.2	24.4







#34
Hexachlorobutadiene
Concen: 20.791 ng
RT: 8.298 min Scan# 1055
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:225 Resp: 176253

Ion	Ratio	Lower	Upper
225	100		
223	63.7	50.6	75.8
227	63.5	52.2	78.2







#35
Caprolactam
Concen: 20.884 ng
RT: 8.580 min Scan# 11
Delta R.T. -0.024 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:113 Resp: 75048

Ion	Ratio	Lower	Upper
113	100		
55	148.8	129.6	169.6
56	121.8	104.0	144.0







#36
4-Chloro-3-methylphenol
Concen: 21.168 ng
RT: 8.698 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:107 Resp: 278207

Ion	Ratio	Lower	Upper
107	100		
144	23.9	19.0	28.6
142	74.3	59.2	88.8







#37
2-Methylnaphthalene
Concen: 21.324 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:142 Resp: 584552

Ion	Ratio	Lower	Upper
142	100		
141	90.0	71.2	106.8







#38
1-Methylnaphthalene
Concen: 21.677 ng
RT: 8.975 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:142 Resp: 575428

Ion	Ratio	Lower	Upper
142	100		
141	91.8	74.2	111.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:164 Resp: 470833

Ion	Ratio	Lower	Upper
164	100		
162	101.6	81.8	122.6
160	45.1	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.538 ng
RT: 9.039 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:216 Resp: 294845

Ion	Ratio	Lower	Upper
216	100		
214	78.7	61.9	92.9
179	22.3	18.1	27.1
108	18.6	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 20.685 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:237 Resp: 117174

Ion	Ratio	Lower	Upper
237	100		
235	63.1	43.1	83.1
272	12.3	0.0	32.0







#42
2,4,6-Tribromophenol
Concen: 40.869 ng
RT: 10.704 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:330 Resp: 243832

Ion	Ratio	Lower	Upper
330	100		
332	95.8	77.6	116.4
141	32.4	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 20.203 ng
RT: 9.145 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:196 Resp: 189069

Ion	Ratio	Lower	Upper
196	100		
198	100.9	80.6	121.0
200	32.3	25.8	38.6







#44
2,4,5-Trichlorophenol
Concen: 20.912 ng
RT: 9.180 min Scan# 1205
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:196 Resp: 198841

Ion	Ratio	Lower	Upper
196	100		
198	99.6	80.7	121.1
97	54.3	42.1	63.1
132	29.4	23.4	35.2







#45
2-Fluorobiphenyl
Concen: 42.002 ng
RT: 9.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:172 Resp: 1298709

Ion	Ratio	Lower	Upper
172	100		
171	35.7	29.3	43.9
170	23.7	19.4	29.0







#46
1,1'-Biphenyl
Concen: 20.985 ng
RT: 9.333 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:154 Resp: 749087

Ion	Ratio	Lower	Upper
154	100		
153	41.0	21.1	61.1
76	11.9	0.0	31.3







#47
2-Chloronaphthalene
Concen: 20.898 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 555946

Ion	Ratio	Lower	Upper
162	100		
127	42.8	35.5	53.3
164	33.1	26.6	40.0







#48
2-Nitroaniline
Concen: 21.302 ng
RT: 9.451 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion: 65 Resp: 162790

Ion	Ratio	Lower	Upper
65	100		
92	69.7	56.6	84.8
138	106.7	88.6	133.0







#49
Acenaphthylene
Concen: 21.222 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 837016

Ion	Ratio	Lower	Upper
152	100		
151	20.1	16.3	24.5
153	13.7	11.1	16.7







#50
Dimethylphthalate
Concen: 21.137 ng
RT: 9.633 min Scan# 1282
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:163 Resp: 691454

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.1	4.7
164	10.3	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 21.096 ng
RT: 9.692 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 143513

Ion	Ratio	Lower	Upper
165	100		
63	58.7	47.0	70.4
89	59.5	47.4	71.2







#52
Acenaphthene
Concen: 20.813 ng
RT: 9.951 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:154 Resp: 579623

Ion	Ratio	Lower	Upper
154	100		
153	110.9	90.7	136.1
152	50.3	41.1	61.7







#53
3-Nitroaniline
Concen: 21.327 ng
RT: 9.863 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:138 Resp: 148977

Ion	Ratio	Lower	Upper
138	100		
108	10.6	9.3	13.9
92	114.5	93.5	140.3







#54
2,4-Dinitrophenol
Concen: 20.368 ng
RT: 9.963 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:184 Resp: 58223

Ion	Ratio	Lower	Upper
184	100		
63	82.2	60.2	90.4
154	293.5	132.9	199.3#







#55
Dibenzofuran
Concen: 21.238 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:168 Resp: 843359

Ion	Ratio	Lower	Upper
168	100		
139	40.9	33.9	50.9
169	13.3	11.0	16.4







#56
4-Nitrophenol
Concen: 20.955 ng
RT: 10.010 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:139 Resp: 109958

Ion	Ratio	Lower	Upper
139	100		
109	81.5	61.6	101.6
65	99.4	79.4	119.4







#57
2,4-Dinitrotoluene
Concen: 21.132 ng
RT: 10.098 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 189234

Ion	Ratio	Lower	Upper
165	100		
63	42.1	35.4	53.2
89	69.2	56.6	84.8
182	2.8	2.2	3.4







#58
Fluorene
Concen: 21.211 ng
RT: 10.463 min Scan# 1423
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:166 Resp: 650245

Ion	Ratio	Lower	Upper
166	100		
165	99.2	79.0	118.4
167	13.4	11.0	16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 21.272 ng
RT: 10.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:232 Resp: 183513

Ion	Ratio	Lower	Upper
232	100		
131	46.0	37.3	55.9
130	2.4	1.9	2.9
166	31.2	24.8	37.2







#60
Diethylphthalate
Concen: 21.468 ng
RT: 10.333 min Scan# 1401
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:149 Resp: 704993

Ion	Ratio	Lower	Upper
149	100		
177	22.3	18.0	27.0
150	12.6	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 20.698 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:204 Resp: 330489

Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.8	40.2
141	65.1	52.4	78.6







#62
4-Nitroaniline
Concen: 21.125 ng
RT: 10.474 min Scan# 1425
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:138 Resp: 142995

Ion	Ratio	Lower	Upper
138	100		
92	51.4	32.6	72.6
108	84.4	65.0	105.0







#63
Azobenzene
Concen: 21.665 ng
RT: 10.616 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 660935

Ion	Ratio	Lower	Upper
77	100		
182	27.3	6.6	46.6
105	21.2	1.3	41.3
51	29.3	10.4	50.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:188 Resp: 826085

Ion	Ratio	Lower	Upper
188	100		
94	8.2	6.8	10.2
80	9.3	7.6	11.4







#65
4,6-Dinitro-2-methylphenol
Concen: 18.755 ng
RT: 10.504 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:198 Resp: 91852

Ion	Ratio	Lower	Upper
198	100		
51	40.7	22.3	62.3
105	41.6	23.0	63.0







#66
n-Nitrosodiphenylamine
Concen: 20.756 ng
RT: 10.569 min Scan# 1441
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:169 Resp: 553905

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.7	80.5
167	35.8	29.2	43.8







#67
4-Bromophenyl-phenylether
Concen: 20.436 ng
RT: 10.945 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:248 Resp: 210241

Ion	Ratio	Lower	Upper
248	100		
250	97.7	77.8	116.8
141	77.7	62.6	93.8







#68
Hexachlorobenzene
Concen: 20.613 ng
RT: 11.010 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:284 Resp: 234399

Ion	Ratio	Lower	Upper
284	100		
142	32.7	27.0	40.6
249	31.1	25.2	37.8







#69
Atrazine
Concen: 24.481 ng
RT: 11.092 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:200 Resp: 159306

Ion	Ratio	Lower	Upper
200	100		
173	24.6	4.7	44.7
215	44.5	25.4	65.4







#70
Pentachlorophenol
Concen: 20.227 ng
RT: 11.198 min Scan# 1548
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:266 Resp: 141842

Ion	Ratio	Lower	Upper
266	100		
268	63.4	51.0	76.6
264	62.9	50.2	75.2







#71
Phenanthrene
Concen: 21.038 ng
RT: 11.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:178 Resp: 939782

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.8	23.8
179	15.5	12.6	19.0







#72
Anthracene
Concen: 20.993 ng
RT: 11.474 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:178 Resp: 940875

Ion	Ratio	Lower	Upper
178	100		
176	18.4	15.1	22.7
179	15.5	12.6	19.0







#73
Carbazole
Concen: 21.096 ng
RT: 11.627 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:167 Resp: 813998

Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.4	26.2
139	13.8	11.0	16.6







#74
Di-n-butylphthalate
Concen: 21.382 ng
RT: 11.963 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:149 Resp: 1089304

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.6	11.4
104	5.2	4.4	6.6







#75
Fluoranthene
Concen: 21.632 ng
RT: 12.610 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:202 Resp: 1014464

Ion	Ratio	Lower	Upper
202	100		
101	9.9	0.0	29.7
203	17.7	0.0	37.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.039 min Scan# 2031
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:240 Resp: 581591

Ion	Ratio	Lower	Upper
240	100		
120	9.9	8.4	12.6
236	25.6	20.5	30.7







#77
Benzidine
Concen: 13.463 ng
RT: 12.733 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:184 Resp: 110648

Ion	Ratio	Lower	Upper
184	100		
185	14.4	11.5	17.3
183	11.5	9.7	14.5







#78
Pyrene
Concen: 20.133 ng
RT: 12.839 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:202 Resp: 1004157

Ion	Ratio	Lower	Upper
202	100		
200	21.1	17.3	25.9
203	17.7	14.4	21.6







#79
Terphenyl-d14
Concen: 40.218 ng
RT: 12.986 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:244 Resp: 1582373

Ion	Ratio	Lower	Upper
244	100		
212	7.0	6.0	9.0
122	9.4	7.7	11.5







#80
Butylbenzylphthalate
Concen: 20.761 ng
RT: 13.457 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:149 Resp: 408044

Ion	Ratio	Lower	Upper
149	100		
91	72.5	56.5	84.7
206	20.5	17.3	25.9







#81
Benzo(a)anthracene
Concen: 20.285 ng
RT: 14.027 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 776120

Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.8	32.8
229	19.6	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 19.999 ng
RT: 13.992 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:252 Resp: 221168

Ion	Ratio	Lower	Upper
252	100		
254	65.0	51.0	76.6
126	10.2	8.4	12.6







#83
Chrysene
Concen: 21.325 ng
RT: 14.062 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 732966

Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.1	36.1
229	19.8	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 20.845 ng
RT: 14.015 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:149 Resp: 562062

Ion	Ratio	Lower	Upper
149	100		
167	26.5	21.4	32.0
279	4.2	3.8	5.6







#85
Di-n-octyl phthalate
Concen: 20.740 ng
RT: 14.633 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:149 Resp: 778373

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	8.5	6.9	10.3







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.509 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:264 Resp: 428960

Ion	Ratio	Lower	Upper
264	100		
260	23.8	19.1	28.7
265	21.7	17.5	26.3







#87
Indeno(1,2,3-cd)pyrene
Concen: 20.163 ng
RT: 16.992 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:276 Resp: 560913

Ion	Ratio	Lower	Upper
276	100		
138	22.6	18.4	27.6
277	24.9	20.3	30.5







#88
Benzo(b)fluoranthene
Concen: 21.886 ng
RT: 15.074 min Scan# 2207
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:252 Resp: 636808

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.5	26.3
125	9.2	7.0	10.4







#89
Benzo(k)fluoranthene
Concen: 20.196 ng
RT: 15.104 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:252 Resp: 513311

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	9.1	7.2	10.8







#90
Benzo(a)pyrene
Concen: 20.991 ng
RT: 15.445 min Scan# 2270
Delta R.T. -0.006 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:252 Resp: 481683

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.0
125	9.8	7.8	11.8







#91
Dibenzo(a,h)anthracene
Concen: 20.046 ng
RT: 17.009 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:278 Resp: 460109

Ion	Ratio	Lower	Upper
278	100		
139	14.8	11.4	17.2
279	23.6	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 19.987 ng
RT: 17.439 min Scan# 2609
Delta R.T. -0.012 min
Lab File: BF141900.D
Acq: 10 Mar 2025 12:29

Tgt Ion:276 Resp: 455034

Ion	Ratio	Lower	Upper
276	100		
277	23.3	18.7	28.1
138	19.2	15.8	23.8



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141901.D
Acq On : 10 Mar 2025 12:58
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Quant Time: Mar 10 15:35:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	222904	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	874166	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	493744	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	839154	20.000	ng	0.00
76) Chrysene-d12	14.039	240	569865	20.000	ng	0.00
86) Perylene-d12	15.509	264	469286	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	1023722	76.628	ng	0.00
7) Phenol-d6	6.504	99	1285980	75.580	ng	0.00
23) Nitrobenzene-d5	7.445	82	1216280	77.983	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	493945	78.949	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	2475837	76.357	ng	0.00
79) Terphenyl-d14	12.986	244	2908357	75.440	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.675	88	224795	40.148	ng	100
3) Pyridine	3.440	79	544434	39.565	ng	100
4) n-Nitrosodimethylamine	3.393	42	259739	39.710	ng	100
6) Aniline	6.545	93	653060	38.879	ng	100
8) 2-Chlorophenol	6.669	128	561230	37.902	ng	100
9) Benzaldehyde	6.434	77	346818	36.467	ng	100
10) Phenol	6.522	94	684724	38.141	ng	100
11) bis(2-Chloroethyl)ether	6.616	93	510214	37.932	ng	100
12) 1,3-Dichlorobenzene	6.822	146	611938	38.242	ng	100
13) 1,4-Dichlorobenzene	6.898	146	620951	38.355	ng	100
14) 1,2-Dichlorobenzene	7.051	146	581043	38.111	ng	100
15) Benzyl Alcohol	7.022	79	528444	38.371	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.157	45	607522	37.476	ng	100
17) 2-Methylphenol	7.128	107	449433	38.190	ng	100
18) Hexachloroethane	7.392	117	230613	37.993	ng	100
19) n-Nitroso-di-n-propyla...	7.298	70	407069	37.223	ng	100
20) 3+4-Methylphenols	7.281	107	576902	38.219	ng	100
22) Acetophenone	7.292	105	804747	38.376	ng	100
24) Nitrobenzene	7.463	77	604738	39.106	ng	100
25) Isophorone	7.704	82	1049049	38.097	ng	100
26) 2-Nitrophenol	7.781	139	305349	40.295	ng	100
27) 2,4-Dimethylphenol	7.810	122	403621	38.671	ng	100
28) bis(2-Chloroethoxy)met...	7.910	93	650344	37.702	ng	100
29) 2,4-Dichlorophenol	8.016	162	500239	38.619	ng	100
30) 1,2,4-Trichlorobenzene	8.104	180	541182	37.870	ng	100
31) Naphthalene	8.186	128	1719643	38.400	ng	100
32) Benzoic acid	7.928	122	363328	39.827	ng	100
33) 4-Chloroaniline	8.228	127	613535	38.824	ng	100
34) Hexachlorobutadiene	8.304	225	355740	38.285	ng	100
35) Caprolactam	8.604	113	147056	37.335	ng	100
36) 4-Chloro-3-methylphenol	8.704	107	555328	38.549	ng	100
37) 2-Methylnaphthalene	8.875	142	1141228	37.983	ng	100
38) 1-Methylnaphthalene	8.975	142	1089576	37.447	ng	100
40) 1,2,4,5-Tetrachloroben...	9.039	216	584322	38.814	ng	100
41) Hexachlorocyclopentadiene	9.028	237	246666	41.523	ng	100
43) 2,4,6-Trichlorophenol	9.145	196	392696	40.014	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141901.D
Acq On : 10 Mar 2025 12:58
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Quant Time: Mar 10 15:35:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.186	196	384539	38.566	ng	100
46) 1,1'-Biphenyl	9.339	154	1436040	38.362	ng	100
47) 2-Chloronaphthalene	9.363	162	1077099	38.609	ng	100
48) 2-Nitroaniline	9.457	65	318068	39.690	ng	100
49) Acenaphthylene	9.780	152	1614250	39.029	ng	100
50) Dimethylphthalate	9.639	163	1312856	38.270	ng	100
51) 2,6-Dinitrotoluene	9.698	165	281695	39.486	ng	100
52) Acenaphthene	9.951	154	1132087	38.764	ng	100
53) 3-Nitroaniline	9.869	138	289208	39.481	ng	100
54) 2,4-Dinitrophenol	9.969	184	140024	38.116	ng	100
55) Dibenzofuran	10.122	168	1593454	38.266	ng	100
56) 4-Nitrophenol	10.016	139	227070	41.266	ng	100
57) 2,4-Dinitrotoluene	10.104	165	370337	39.437	ng	100
58) Fluorene	10.469	166	1213316	37.742	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.239	232	354863	39.225	ng	100
60) Diethylphthalate	10.339	149	1319116	38.304	ng	100
61) 4-Chlorophenyl-phenyle...	10.457	204	635445	37.950	ng	100
62) 4-Nitroaniline	10.480	138	284615	40.097	ng	100
63) Azobenzene	10.616	77	1229473	38.432	ng	100
65) 4,6-Dinitro-2-methylph...	10.510	198	203106	40.827	ng	100
66) n-Nitrosodiphenylamine	10.574	169	1041349	38.415	ng	100
67) 4-Bromophenyl-phenylether	10.945	248	409602	39.195	ng	100
68) Hexachlorobenzene	11.010	284	452189	39.146	ng	100
69) Atrazine	11.098	200	275084	41.614	ng	100
70) Pentachlorophenol	11.204	266	297893	41.818	ng	100
71) Phenanthrene	11.427	178	1741309	38.373	ng	100
72) Anthracene	11.480	178	1756035	38.571	ng	100
73) Carbazole	11.633	167	1524663	38.898	ng	100
74) Di-n-butylphthalate	11.963	149	1975625	38.175	ng	100
75) Fluoranthene	12.616	202	1860202	39.049	ng	100
77) Benzidine	12.733	184	410026	50.915	ng	100
78) Pyrene	12.845	202	1835917	37.567	ng	100
80) Butylbenzylphthalate	13.463	149	746613	38.769	ng	100
81) Benzo(a)anthracene	14.027	228	1451332	38.713	ng	100
82) 3,3'-Dichlorobenzidine	13.992	252	423835	39.114	ng	100
83) Chrysene	14.068	228	1294538	38.439	ng	100
84) Bis(2-ethylhexyl)phtha...	14.021	149	1029987	38.985	ng	100
85) Di-n-octyl phthalate	14.633	149	1441248	39.193	ng	100
87) Indeno(1,2,3-cd)pyrene	16.998	276	1207616	39.681	ng	100
88) Benzo(b)fluoranthene	15.080	252	1135547	35.673	ng	100
89) Benzo(k)fluoranthene	15.109	252	1130615	40.661	ng	100
90) Benzo(a)pyrene	15.451	252	972220	38.727	ng	100
91) Dibenzo(a,h)anthracene	17.021	278	1002747	39.933	ng	100
92) Benzo(g,h,i)perylene	17.451	276	999947	40.149	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141901.D
Acq On : 10 Mar 2025 12:58
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Quant Time: Mar 10 15:35:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 222904

Ion	Ratio	Lower	Upper
152	100		
150	159.3	127.4	191.2
115	64.9	51.9	77.9







#2
1,4-Dioxane
Concen: 40.148 ng
RT: 2.675 min Scan# 99
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion: 88 Resp: 224795

Ion	Ratio	Lower	Upper
88	100		
58	71.8	57.4	86.2
43	26.9	21.5	32.3







#3
Pyridine
Concen: 39.565 ng
RT: 3.440 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 544434

Ion	Ratio	Lower	Upper
79	100		
52	66.6	53.3	79.9
51	34.5	27.6	41.4







#4
n-Nitrosodimethylamine
Concen: 39.710 ng
RT: 3.393 min Scan# 221
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion: 42 Resp: 259739

Ion	Ratio	Lower	Upper
42	100		
74	129.2	103.4	155.0
44	6.2	5.0	7.4







#5
2-Fluorophenol
Concen: 76.628 ng
RT: 5.492 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:112 Resp: 1023722

Ion	Ratio	Lower	Upper
112	100		
64	58.1	46.5	69.7
63	31.8	25.4	38.2







#6
Aniline
Concen: 38.879 ng
RT: 6.545 min Scan# 757
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion: 93 Resp: 653060

Ion	Ratio	Lower	Upper
93	100		
66	39.9	31.9	47.9
65	19.7	15.8	23.6







#7
Phenol-d6
Concen: 75.580 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 99 Resp: 1285980

Ion	Ratio	Lower	Upper
99	100		
42	18.2	14.6	21.8
71	38.5	30.8	46.2







#8
2-Chlorophenol
Concen: 37.902 ng
RT: 6.669 min Scan# 778
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:128 Resp: 561230

Ion	Ratio	Lower	Upper
128	100		
130	32.8	12.8	52.8
64	44.4	24.4	64.4







#9
Benzaldehyde
Concen: 36.467 ng
RT: 6.434 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 346818

Ion	Ratio	Lower	Upper
77	100		
105	100.0	80.0	120.0
106	94.5	74.5	114.5







#10
Phenol
Concen: 38.141 ng
RT: 6.522 min Scan# 753
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion: 94 Resp: 684724

Ion	Ratio	Lower	Upper
94	100		
65	30.5	10.5	50.5
66	45.8	25.8	65.8







#11
bis(2-Chloroethyl)ether
Concen: 37.932 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 93 Resp: 510214

Ion	Ratio	Lower	Upper
93	100		
63	73.0	53.0	93.0
95	32.9	12.9	52.9







#12
1,3-Dichlorobenzene
Concen: 38.242 ng
RT: 6.822 min Scan# 804
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:146 Resp: 611938

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.3	76.9
75	33.3	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 38.355 ng
RT: 6.898 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:146 Resp: 620951

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.0	76.6
111	44.6	35.7	53.5







#14
1,2-Dichlorobenzene
Concen: 38.111 ng
RT: 7.051 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:146 Resp: 581043

Ion	Ratio	Lower	Upper
146	100		
148	63.5	50.8	76.2
111	47.5	38.0	57.0







#15
Benzyl Alcohol
Concen: 38.371 ng
RT: 7.022 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 528444

Ion	Ratio	Lower	Upper
79	100		
108	74.6	59.7	89.5
77	63.8	51.0	76.6







#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.476 ng
RT: 7.157 min Scan# 861
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion: 45 Resp: 607522

Ion	Ratio	Lower	Upper
45	100		
77	17.8	0.0	37.8
79	14.2	0.0	34.2







#17
2-Methylphenol
Concen: 38.190 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:107 Resp: 449433

Ion	Ratio	Lower	Upper
107	100		
108	112.7	90.2	135.2
77	49.5	39.6	59.4
79	48.9	39.1	58.7







#18
Hexachloroethane
Concen: 37.993 ng
RT: 7.392 min Scan# 901
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:117 Resp: 230613

Ion	Ratio	Lower	Upper
117	100		
119	96.1	76.9	115.3
201	99.9	79.9	119.9







#19
n-Nitroso-di-n-propylamine
Concen: 37.223 ng
RT: 7.298 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 70 Resp: 407069

Ion	Ratio	Lower	Upper
70	100		
42	51.7	41.4	62.0
101	10.2	8.2	12.2
130	22.8	18.2	27.4







#20
3+4-Methylphenols
Concen: 38.219 ng
RT: 7.281 min Scan# 882
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:107 Resp: 576902

Ion	Ratio	Lower	Upper
107	100		
108	88.4	68.4	108.4
77	89.7	69.7	109.7
79	31.4	11.4	51.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:136 Resp: 874166

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	7.3	5.8	8.8
68	5.1	4.1	6.1







#22
Acetophenone
Concen: 38.376 ng
RT: 7.292 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:105 Resp: 804747

Ion	Ratio	Lower	Upper
105	100		
71	7.3	5.8	8.8
51	29.2	23.4	35.0
120	21.6	17.3	25.9







#23
Nitrobenzene-d5
Concen: 77.983 ng
RT: 7.445 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1216280

Ion	Ratio	Lower	Upper
82	100		
128	45.0	36.0	54.0
54	49.5	39.6	59.4







#24
Nitrobenzene
Concen: 39.106 ng
RT: 7.463 min Scan# 913
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion: 77 Resp: 604738

Ion	Ratio	Lower	Upper
77	100		
123	46.2	37.0	55.4
65	14.1	11.3	16.9







#25
Isophorone
Concen: 38.097 ng
RT: 7.704 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1049049

Ion	Ratio	Lower	Upper
82	100		
95	8.0	6.4	9.6
138	18.6	14.9	22.3







#26
2-Nitrophenol
Concen: 40.295 ng
RT: 7.781 min Scan# 967
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:139 Resp: 305349

Ion	Ratio	Lower	Upper
139	100		
109	33.1	26.5	39.7
65	50.4	40.3	60.5







#27
2,4-Dimethylphenol
Concen: 38.671 ng
RT: 7.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:122 Resp: 403621

Ion	Ratio	Lower	Upper
122	100		
107	118.5	94.8	142.2
121	59.7	47.8	71.6







#28
bis(2-Chloroethoxy)methane
Concen: 37.702 ng
RT: 7.910 min Scan# 989
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion: 93 Resp: 650344

Ion	Ratio	Lower	Upper
93	100		
95	33.3	26.6	40.0
123	12.8	10.2	15.4







#29
2,4-Dichlorophenol
Concen: 38.619 ng
RT: 8.016 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 500239

Ion	Ratio	Lower	Upper
162	100		
164	63.8	43.8	83.8
98	38.1	18.1	58.1







#30
1,2,4-Trichlorobenzene
Concen: 37.870 ng
RT: 8.104 min Scan# 1022
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:180 Resp: 541182

Ion	Ratio	Lower	Upper
180	100		
182	96.9	77.5	116.3
145	32.2	25.8	38.6







#31
Naphthalene
Concen: 38.400 ng
RT: 8.186 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:128 Resp: 1719643

Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.6
127	13.5	10.8	16.2







#32
Benzoic acid
Concen: 39.827 ng
RT: 7.928 min Scan# 992
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:122 Resp: 363328

Ion	Ratio	Lower	Upper
122	100		
105	127.9	107.9	147.9
77	99.0	79.0	119.0







#33
4-Chloroaniline
Concen: 38.824 ng
RT: 8.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:127 Resp: 613535

Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.2	39.4
65	31.8	25.4	38.2
92	20.3	16.2	24.4







#34
Hexachlorobutadiene
Concen: 38.285 ng
RT: 8.304 min Scan# 1056
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:225 Resp: 355740

Ion	Ratio	Lower	Upper
225	100		
223	63.2	50.6	75.8
227	65.2	52.2	78.2







#35
Caprolactam
Concen: 37.335 ng
RT: 8.604 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:113 Resp: 147056

Ion	Ratio	Lower	Upper
113	100		
55	149.6	129.6	169.6
56	124.0	104.0	144.0







#36
4-Chloro-3-methylphenol
Concen: 38.549 ng
RT: 8.704 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:107 Resp: 555328

Ion	Ratio	Lower	Upper
107	100		
144	23.8	19.0	28.6
142	74.0	59.2	88.8







#37
2-Methylnaphthalene
Concen: 37.983 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:142 Resp: 1141228

Ion	Ratio	Lower	Upper
142	100		
141	89.0	71.2	106.8







#38
1-Methylnaphthalene
Concen: 37.447 ng
RT: 8.975 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:142 Resp: 1089576

Ion	Ratio	Lower	Upper
142	100		
141	92.8	74.2	111.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:164 Resp: 493744

Ion	Ratio	Lower	Upper
164	100		
162	102.2	81.8	122.6
160	45.9	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.814 ng
RT: 9.039 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:216 Resp: 584322

Ion	Ratio	Lower	Upper
216	100		
214	77.4	61.9	92.9
179	22.6	18.1	27.1
108	18.4	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 41.523 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:237 Resp: 246666

Ion	Ratio	Lower	Upper
237	100		
235	63.1	43.1	83.1
272	12.0	0.0	32.0







#42
2,4,6-Tribromophenol
Concen: 78.949 ng
RT: 10.704 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:330 Resp: 493945

Ion	Ratio	Lower	Upper
330	100		
332	97.0	77.6	116.4
141	30.9	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 40.014 ng
RT: 9.145 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:196 Resp: 392696

Ion	Ratio	Lower	Upper
196	100		
198	100.8	80.6	121.0
200	32.2	25.8	38.6







#44
2,4,5-Trichlorophenol
Concen: 38.566 ng
RT: 9.186 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:196 Resp: 384539

Ion	Ratio	Lower	Upper
196	100		
198	100.9	80.7	121.1
97	52.6	42.1	63.1
132	29.3	23.4	35.2







#45
2-Fluorobiphenyl
Concen: 76.357 ng
RT: 9.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:172 Resp: 2475837

Ion	Ratio	Lower	Upper
172	100		
171	36.6	29.3	43.9
170	24.2	19.4	29.0







#46
1,1'-Biphenyl
Concen: 38.362 ng
RT: 9.339 min Scan# 1232
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:154 Resp: 1436040

Ion	Ratio	Lower	Upper
154	100		
153	41.1	21.1	61.1
76	11.3	0.0	31.3







#47
2-Chloronaphthalene
Concen: 38.609 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 1077099

Ion	Ratio	Lower	Upper
162	100		
127	44.4	35.5	53.3
164	33.3	26.6	40.0







#48
2-Nitroaniline
Concen: 39.690 ng
RT: 9.457 min Scan# 1252
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion: 65 Resp: 318068

Ion	Ratio	Lower	Upper
65	100		
92	70.7	56.6	84.8
138	110.8	88.6	133.0







#49
Acenaphthylene
Concen: 39.029 ng
RT: 9.780 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 1614250

Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.3	24.5
153	13.9	11.1	16.7







#50
Dimethylphthalate
Concen: 38.270 ng
RT: 9.639 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:163 Resp: 1312856

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.1	4.7
164	10.3	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 39.486 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 281695

Ion	Ratio	Lower	Upper
165	100		
63	58.7	47.0	70.4
89	59.3	47.4	71.2







#52
Acenaphthene
Concen: 38.764 ng
RT: 9.951 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:154 Resp: 1132087

Ion	Ratio	Lower	Upper
154	100		
153	113.4	90.7	136.1
152	51.4	41.1	61.7







#53
3-Nitroaniline
Concen: 39.481 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:138 Resp: 289208

Ion	Ratio	Lower	Upper
138	100		
108	11.6	9.3	13.9
92	116.9	93.5	140.3







#54
2,4-Dinitrophenol
Concen: 38.116 ng
RT: 9.969 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:184 Resp: 140024

Ion	Ratio	Lower	Upper
184	100		
63	75.3	60.2	90.4
154	166.1	132.9	199.3







#55
Dibenzofuran
Concen: 38.266 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:168 Resp: 1593454

Ion	Ratio	Lower	Upper
168	100		
139	42.4	33.9	50.9
169	13.7	11.0	16.4







#56
4-Nitrophenol
Concen: 41.266 ng
RT: 10.016 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:139 Resp: 227070

Ion	Ratio	Lower	Upper
139	100		
109	81.6	61.6	101.6
65	99.4	79.4	119.4







#57
2,4-Dinitrotoluene
Concen: 39.437 ng
RT: 10.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 370337

Ion	Ratio	Lower	Upper
165	100		
63	44.3	35.4	53.2
89	70.7	56.6	84.8
182	2.8	2.2	3.4







#58
Fluorene
Concen: 37.742 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:166 Resp: 1213316

Ion	Ratio	Lower	Upper
166	100		
165	98.7	79.0	118.4
167	13.7	11.0	16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 39.225 ng
RT: 10.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:232 Resp: 354863

Ion	Ratio	Lower	Upper
232	100		
131	46.6	37.3	55.9
130	2.4	1.9	2.9
166	31.0	24.8	37.2







#60
Diethylphthalate
Concen: 38.304 ng
RT: 10.339 min Scan# 1402
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:149 Resp: 1319116

Ion	Ratio	Lower	Upper
149	100		
177	22.5	18.0	27.0
150	12.8	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 37.950 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:204 Resp: 635445

Ion	Ratio	Lower	Upper
204	100		
206	33.5	26.8	40.2
141	65.5	52.4	78.6







#62
4-Nitroaniline
Concen: 40.097 ng
RT: 10.480 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:138 Resp: 284615

Ion	Ratio	Lower	Upper
138	100		
92	52.6	32.6	72.6
108	85.0	65.0	105.0







#63
Azobenzene
Concen: 38.432 ng
RT: 10.616 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 1229473

Ion	Ratio	Lower	Upper
77	100		
182	26.6	6.6	46.6
105	21.3	1.3	41.3
51	30.4	10.4	50.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:188 Resp: 839154

Ion	Ratio	Lower	Upper
188	100		
94	8.5	6.8	10.2
80	9.5	7.6	11.4







#65
4,6-Dinitro-2-methylphenol
Concen: 40.827 ng
RT: 10.510 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:198 Resp: 203106

Ion	Ratio	Lower	Upper
198	100		
51	42.3	22.3	62.3
105	43.0	23.0	63.0







#66
n-Nitrosodiphenylamine
Concen: 38.415 ng
RT: 10.574 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:169 Resp: 1041349

Ion	Ratio	Lower	Upper
169	100		
168	67.1	53.7	80.5
167	36.5	29.2	43.8







#67
4-Bromophenyl-phenylether
Concen: 39.195 ng
RT: 10.945 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:248 Resp: 409602

Ion	Ratio	Lower	Upper
248	100		
250	97.3	77.8	116.8
141	78.2	62.6	93.8







#68
Hexachlorobenzene
Concen: 39.146 ng
RT: 11.010 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:284 Resp: 452189

Ion	Ratio	Lower	Upper
284	100		
142	33.8	27.0	40.6
249	31.5	25.2	37.8







#69
Atrazine
Concen: 41.614 ng
RT: 11.098 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:200 Resp: 275084

Ion	Ratio	Lower	Upper
200	100		
173	24.7	4.7	44.7
215	45.4	25.4	65.4







#70
Pentachlorophenol
Concen: 41.818 ng
RT: 11.204 min Scan# 1549
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:266 Resp: 297893

Ion	Ratio	Lower	Upper
266	100		
268	63.8	51.0	76.6
264	62.7	50.2	75.2







#71
Phenanthrene
Concen: 38.373 ng
RT: 11.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:178 Resp: 1741309

Ion	Ratio	Lower	Upper
178	100		
176	19.8	15.8	23.8
179	15.8	12.6	19.0







#72
Anthracene
Concen: 38.571 ng
RT: 11.480 min Scan# 1596
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:178 Resp: 1756035

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.1	22.7
179	15.8	12.6	19.0







#73
Carbazole
Concen: 38.898 ng
RT: 11.633 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:167 Resp: 1524663

Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.4	26.2
139	13.8	11.0	16.6







#74
Di-n-butylphthalate
Concen: 38.175 ng
RT: 11.963 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:149 Resp: 1975625

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.6	11.4
104	5.5	4.4	6.6







#75
Fluoranthene
Concen: 39.049 ng
RT: 12.616 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:202 Resp: 1860202

Ion	Ratio	Lower	Upper
202	100		
101	9.7	0.0	29.7
203	17.8	0.0	37.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.039 min Scan# 2031
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:240 Resp: 569865

Ion	Ratio	Lower	Upper
240	100		
120	10.5	8.4	12.6
236	25.6	20.5	30.7







#77
Benzidine
Concen: 50.915 ng
RT: 12.733 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:184 Resp: 410026

Ion	Ratio	Lower	Upper
184	100		
185	14.4	11.5	17.3
183	12.1	9.7	14.5







#78
Pyrene
Concen: 37.567 ng
RT: 12.845 min Scan# 1828
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:202 Resp: 1835917

Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.3	25.9
203	18.0	14.4	21.6







#79
Terphenyl-d14
Concen: 75.440 ng
RT: 12.986 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:244 Resp: 2908357

Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.0	9.0
122	9.6	7.7	11.5







#80
Butylbenzylphthalate
Concen: 38.769 ng
RT: 13.463 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:149 Resp: 746613

Ion	Ratio	Lower	Upper
149	100		
91	70.6	56.5	84.7
206	21.6	17.3	25.9







#81
Benzo(a)anthracene
Concen: 38.713 ng
RT: 14.027 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 1451332

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.8	32.8
229	20.0	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 39.114 ng
RT: 13.992 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:252 Resp: 423835

Ion	Ratio	Lower	Upper
252	100		
254	63.8	51.0	76.6
126	10.5	8.4	12.6







#83
Chrysene
Concen: 38.439 ng
RT: 14.068 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 1294538

Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.1	36.1
229	19.6	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 38.985 ng
RT: 14.021 min Scan# 2028
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:149 Resp: 1029987

Ion	Ratio	Lower	Upper
149	100		
167	26.7	21.4	32.0
279	4.7	3.8	5.6







#85
Di-n-octyl phthalate
Concen: 39.193 ng
RT: 14.633 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:149 Resp: 1441248

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.6	6.9	10.3







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.509 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:264 Resp: 469286

Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.1	28.7
265	21.9	17.5	26.3







#87
Indeno(1,2,3-cd)pyrene
Concen: 39.681 ng
RT: 16.998 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:276 Resp: 1207616

Ion	Ratio	Lower	Upper
276	100		
138	23.0	18.4	27.6
277	25.4	20.3	30.5







#88
Benzo(b)fluoranthene
Concen: 35.673 ng
RT: 15.080 min Scan# 2208
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:252 Resp: 1135547

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.5	26.3
125	8.7	7.0	10.4







#89
Benzo(k)fluoranthene
Concen: 40.661 ng
RT: 15.109 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:252 Resp: 1130615

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	9.0	7.2	10.8







#90
Benzo(a)pyrene
Concen: 38.727 ng
RT: 15.451 min Scan# 2271
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:252 Resp: 972220

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.0
125	9.8	7.8	11.8







#91
Dibenzo(a,h)anthracene
Concen: 39.933 ng
RT: 17.021 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:278 Resp: 1002747

Ion	Ratio	Lower	Upper
278	100		
139	14.3	11.4	17.2
279	23.5	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 40.149 ng
RT: 17.451 min Scan# 2611
Delta R.T. 0.000 min
Lab File: BF141901.D
Acq: 10 Mar 2025 12:58

Tgt Ion:276 Resp: 999947

Ion	Ratio	Lower	Upper
276	100		
277	23.4	18.7	28.1
138	19.8	15.8	23.8



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141903.D
Acq On : 10 Mar 2025 13:57
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Quant Time: Mar 10 15:37:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	216861	20.000	ng		0.00
21) Naphthalene-d8	8.169	136	872222	20.000	ng		0.00
39) Acenaphthene-d10	9.922	164	497194	20.000	ng		0.00
64) Phenanthrene-d10	11.404	188	827202	20.000	ng		0.00
76) Chrysene-d12	14.045	240	468454	20.000	ng		0.00
86) Perylene-d12	15.509	264	407346	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.498	112	1490872	114.705	ng		0.00
7) Phenol-d6	6.510	99	1913610	115.601	ng		0.00
23) Nitrobenzene-d5	7.451	82	1899726	122.074	ng		0.00
42) 2,4,6-Tribromophenol	10.710	330	791365	125.609	ng		0.00
45) 2-Fluorobiphenyl	9.239	172	3716590	113.827	ng		0.00
79) Terphenyl-d14	12.992	244	3982780	125.675	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.669	88	325927	59.832	ng		99
3) Pyridine	3.440	79	786537	58.753	ng		99
4) n-Nitrosodimethylamine	3.404	42	386646	60.759	ng		100
6) Aniline	6.551	93	933875	57.146	ng		99
8) 2-Chlorophenol	6.669	128	835549	58.000	ng		97
9) Benzaldehyde	6.434	77	465511	50.311	ng		100
10) Phenol	6.528	94	996848	57.074	ng		95
11) bis(2-Chloroethyl)ether	6.622	93	765071	58.464	ng		99
12) 1,3-Dichlorobenzene	6.828	146	911340	58.540	ng		99
13) 1,4-Dichlorobenzene	6.904	146	923545	58.635	ng		99
14) 1,2-Dichlorobenzene	7.057	146	873604	58.898	ng		99
15) Benzyl Alcohol	7.028	79	810347	60.480	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.157	45	957808	60.730	ng		99
17) 2-Methylphenol	7.134	107	701676	61.285	ng		99
18) Hexachloroethane	7.398	117	364561	61.734	ng		96
19) n-Nitroso-di-n-propyla...	7.304	70	644501	60.576	ng		99
20) 3+4-Methylphenols	7.292	107	868157	59.116	ng	#	81
22) Acetophenone	7.298	105	1226859	58.635	ng		97
24) Nitrobenzene	7.469	77	939508	60.889	ng		100
25) Isophorone	7.710	82	1695749	61.720	ng		100
26) 2-Nitrophenol	7.781	139	491301	64.978	ng		98
27) 2,4-Dimethylphenol	7.816	122	639505	61.408	ng		99
28) bis(2-Chloroethoxy)met...	7.916	93	1039951	60.422	ng		99
29) 2,4-Dichlorophenol	8.022	162	791904	61.272	ng		99
30) 1,2,4-Trichlorobenzene	8.104	180	870403	61.044	ng		99
31) Naphthalene	8.186	128	2545181	56.961	ng		99
32) Benzoic acid	7.951	122	634105	69.664	ng		99
33) 4-Chloroaniline	8.233	127	920015	58.348	ng		99
34) Hexachlorobutadiene	8.304	225	559955	60.397	ng		99
35) Caprolactam	8.628	113	235861	60.015	ng		95
36) 4-Chloro-3-methylphenol	8.716	107	874022	60.807	ng		98
37) 2-Methylnaphthalene	8.875	142	1744374	58.187	ng		99
38) 1-Methylnaphthalene	8.975	142	1692113	58.285	ng		99
40) 1,2,4,5-Tetrachloroben...	9.045	216	903928	59.627	ng		99
41) Hexachlorocyclopentadiene	9.028	237	375919	62.842	ng		100
43) 2,4,6-Trichlorophenol	9.151	196	597686	60.479	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141903.D
Acq On : 10 Mar 2025 13:57
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Quant Time: Mar 10 15:37:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	618930	61.642	ng		99
46) 1,1'-Biphenyl	9.345	154	2134829	56.634	ng		99
47) 2-Chloronaphthalene	9.369	162	1625973	57.879	ng		100
48) 2-Nitroaniline	9.463	65	495229	61.368	ng		99
49) Acenaphthylene	9.786	152	2379216	57.125	ng		99
50) Dimethylphthalate	9.651	163	2006480	58.083	ng		100
51) 2,6-Dinitrotoluene	9.704	165	441383	61.441	ng		99
52) Acenaphthene	9.957	154	1737049	59.065	ng		99
53) 3-Nitroaniline	9.875	138	449553	60.945	ng		100
54) 2,4-Dinitrophenol	9.975	184	240530	60.329	ng	#	74
55) Dibenzofuran	10.127	168	2380973	56.781	ng		100
56) 4-Nitrophenol	10.027	139	343657	62.021	ng		98
57) 2,4-Dinitrotoluene	10.110	165	571731	60.461	ng		98
58) Fluorene	10.469	166	1840654	56.859	ng		100
59) 2,3,4,6-Tetrachlorophenol	10.239	232	546621	60.002	ng		100
60) Diethylphthalate	10.345	149	1995827	57.552	ng		99
61) 4-Chlorophenyl-phenyle...	10.463	204	977819	57.993	ng		97
62) 4-Nitroaniline	10.492	138	426741	59.702	ng		100
63) Azobenzene	10.622	77	1829282	56.784	ng		98
65) 4,6-Dinitro-2-methylph...	10.516	198	330315	67.356	ng		98
66) n-Nitrosodiphenylamine	10.580	169	1583832	59.270	ng		100
67) 4-Bromophenyl-phenylether	10.951	248	608246	59.044	ng		98
68) Hexachlorobenzene	11.016	284	673465	59.144	ng		97
70) Pentachlorophenol	11.204	266	462709	65.893	ng		100
71) Phenanthrene	11.433	178	2560092	57.232	ng		100
72) Anthracene	11.486	178	2570717	57.281	ng		100
73) Carbazole	11.633	167	2213331	57.284	ng		99
74) Di-n-butylphthalate	11.963	149	2916048	57.161	ng		99
75) Fluoranthene	12.616	202	2628993	55.984	ng		99
77) Benzidine	12.733	184	386007	58.309	ng		99
78) Pyrene	12.845	202	2562337	63.781	ng		99
80) Butylbenzylphthalate	13.463	149	1008043	63.676	ng		100
81) Benzo(a)anthracene	14.033	228	1850347	60.040	ng		100
82) 3,3'-Dichlorobenzidine	13.992	252	553078	62.091	ng		100
83) Chrysene	14.068	228	1705257	61.596	ng		99
84) Bis(2-ethylhexyl)phtha...	14.021	149	1337474	61.582	ng		99
85) Di-n-octyl phthalate	14.633	149	1953184	64.613	ng		99
87) Indeno(1,2,3-cd)pyrene	17.009	276	1799372	68.115	ng		100
88) Benzo(b)fluoranthene	15.080	252	1698744	61.481	ng		99
89) Benzo(k)fluoranthene	15.115	252	1478389	61.253	ng		100
90) Benzo(a)pyrene	15.451	252	1323726	60.747	ng		99
91) Dibenzo(a,h)anthracene	17.027	278	1459934	66.981	ng		99
92) Benzo(g,h,i)perylene	17.462	276	1425582	65.942	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141903.D
Acq On : 10 Mar 2025 13:57
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Quant Time: Mar 10 15:37:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 216861

Ion	Ratio	Lower	Upper
152	100		
150	159.8	127.4	191.2
115	65.7	51.9	77.9







#2
1,4-Dioxane
Concen: 59.832 ng
RT: 2.669 min Scan# 98
Delta R.T. -0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion: 88 Resp: 325927

Ion	Ratio	Lower	Upper
88	100		
58	72.5	57.4	86.2
43	27.4	21.5	32.3







#3
Pyridine
Concen: 58.753 ng
RT: 3.440 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 786537

Ion	Ratio	Lower	Upper
79	100		
52	65.9	53.3	79.9
51	34.4	27.6	41.4







#4
n-Nitrosodimethylamine
Concen: 60.759 ng
RT: 3.404 min Scan# 223
Delta R.T. 0.012 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion: 42 Resp: 386646

Ion	Ratio	Lower	Upper
42	100		
74	129.0	103.4	155.0
44	5.8	5.0	7.4







#5
2-Fluorophenol
Concen: 114.705 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:112 Resp: 1490872

Ion	Ratio	Lower	Upper
112	100		
64	57.2	46.5	69.7
63	31.6	25.4	38.2







#6
Aniline
Concen: 57.146 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion: 93 Resp: 933875

Ion	Ratio	Lower	Upper
93	100		
66	40.5	31.9	47.9
65	19.9	15.8	23.6







#7
Phenol-d6
Concen: 115.601 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 99 Resp: 1913610

Ion	Ratio	Lower	Upper
99	100		
42	18.2	14.6	21.8
71	39.1	30.8	46.2







#8
2-Chlorophenol
Concen: 58.000 ng
RT: 6.669 min Scan# 778
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:128 Resp: 835549

Ion	Ratio	Lower	Upper
128	100		
130	32.9	12.8	52.8
64	47.6	24.4	64.4







#9
Benzaldehyde
Concen: 50.311 ng
RT: 6.434 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 465511

Ion	Ratio	Lower	Upper
77	100		
105	99.3	80.0	120.0
106	94.7	74.5	114.5







#10
Phenol
Concen: 57.074 ng
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion: 94 Resp: 996848

Ion	Ratio	Lower	Upper
94	100		
65	32.6	10.5	50.5
66	49.5	25.8	65.8







#11
bis(2-Chloroethyl)ether
Concen: 58.464 ng
RT: 6.622 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 93 Resp: 765071

Ion	Ratio	Lower	Upper
93	100		
63	73.9	53.0	93.0
95	32.4	12.9	52.9







#12
1,3-Dichlorobenzene
Concen: 58.540 ng
RT: 6.828 min Scan# 805
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:146 Resp: 911340

Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.3	76.9
75	32.6	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 58.635 ng
RT: 6.904 min Scan# 81
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:146 Resp: 923545

Ion	Ratio	Lower	Upper
146	100		
148	64.7	51.0	76.6
111	44.4	35.7	53.5







#14
1,2-Dichlorobenzene
Concen: 58.898 ng
RT: 7.057 min Scan# 844
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:146 Resp: 873604

Ion	Ratio	Lower	Upper
146	100		
148	64.4	50.8	76.2
111	46.7	38.0	57.0







#15
Benzyl Alcohol
Concen: 60.480 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 810347

Ion	Ratio	Lower	Upper
79	100		
108	73.6	59.7	89.5
77	63.4	51.0	76.6







#16
2,2'-oxybis(1-Chloropropane)
Concen: 60.730 ng
RT: 7.157 min Scan# 861
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion: 45 Resp: 957808

Ion	Ratio	Lower	Upper
45	100		
77	18.2	0.0	37.8
79	13.9	0.0	34.2







#17
2-Methylphenol
Concen: 61.285 ng
RT: 7.134 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:107 Resp: 701676

Ion	Ratio	Lower	Upper
107	100		
108	112.1	90.2	135.2
77	49.8	39.6	59.4
79	50.1	39.1	58.7







#18
Hexachloroethane
Concen: 61.734 ng
RT: 7.398 min Scan# 902
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:117 Resp: 364561

Ion	Ratio	Lower	Upper
117	100		
119	96.8	76.9	115.3
201	107.1	79.9	119.9







#19
n-Nitroso-di-n-propylamine
Concen: 60.576 ng
RT: 7.304 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 70 Resp: 644501

Ion	Ratio	Lower	Upper
70	100		
42	52.8	41.4	62.0
101	10.4	8.2	12.2
130	22.4	18.2	27.4







#20
3+4-Methylphenols
Concen: 59.116 ng
RT: 7.292 min Scan# 884
Delta R.T. 0.012 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:107 Resp: 868157

Ion	Ratio	Lower	Upper
107	100		
108	90.1	68.4	108.4
77	128.9	69.7	109.7#
79	33.3	11.4	51.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:136 Resp: 872222

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	7.0	5.8	8.8
68	4.9	4.1	6.1







#22
Acetophenone
Concen: 58.635 ng
RT: 7.298 min Scan# 885
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:105 Resp: 1226859

Ion	Ratio	Lower	Upper
105	100		
71	6.2	5.8	8.8
51	31.1	23.4	35.0
120	22.2	17.3	25.9







#23
Nitrobenzene-d5
Concen: 122.074 ng
RT: 7.451 min Scan# 91
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 1899726

Ion	Ratio	Lower	Upper
82	100		
128	44.9	36.0	54.0
54	50.3	39.6	59.4







#24
Nitrobenzene
Concen: 60.889 ng
RT: 7.469 min Scan# 914
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion: 77 Resp: 939508

Ion	Ratio	Lower	Upper
77	100		
123	45.9	37.0	55.4
65	14.3	11.3	16.9







#25
Isophorone
Concen: 61.720 ng
RT: 7.710 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 1695749

Ion	Ratio	Lower	Upper
82	100		
95	7.9	6.4	9.6
138	18.4	14.9	22.3







#26
2-Nitrophenol
Concen: 64.978 ng
RT: 7.781 min Scan# 967
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:139 Resp: 491301

Ion	Ratio	Lower	Upper
139	100		
109	33.1	26.5	39.7
65	52.3	40.3	60.5







#27
2,4-Dimethylphenol
Concen: 61.408 ng
RT: 7.816 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:122 Resp: 639505

Ion	Ratio	Lower	Upper
122	100		
107	119.4	94.8	142.2
121	59.7	47.8	71.6







#28
bis(2-Chloroethoxy)methane
Concen: 60.422 ng
RT: 7.916 min Scan# 990
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion: 93 Resp: 1039951

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.6	40.0
123	12.5	10.2	15.4







#29
2,4-Dichlorophenol
Concen: 61.272 ng
RT: 8.022 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 791904

Ion	Ratio	Lower	Upper
162	100		
164	64.9	43.8	83.8
98	38.3	18.1	58.1







#30
1,2,4-Trichlorobenzene
Concen: 61.044 ng
RT: 8.104 min Scan# 1022
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:180 Resp: 870403

Ion	Ratio	Lower	Upper
180	100		
182	95.1	77.5	116.3
145	32.2	25.8	38.6







#31
Naphthalene
Concen: 56.961 ng
RT: 8.186 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:128 Resp: 2545181

Ion	Ratio	Lower	Upper
128	100		
129	11.5	9.0	13.6
127	13.8	10.8	16.2







#32
Benzoic acid
Concen: 69.664 ng
RT: 7.951 min Scan# 996
Delta R.T. 0.023 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:122 Resp: 634105

Ion	Ratio	Lower	Upper
122	100		
105	127.5	107.9	147.9
77	96.2	79.0	119.0







#33
4-Chloroaniline
Concen: 58.348 ng
RT: 8.233 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 920015

Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.2	39.4
65	31.5	25.4	38.2
92	20.7	16.2	24.4







#34
Hexachlorobutadiene
Concen: 60.397 ng
RT: 8.304 min Scan# 1056
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:225 Resp: 559955

Ion	Ratio	Lower	Upper
225	100		
223	63.8	50.6	75.8
227	63.8	52.2	78.2







#35
Caprolactam
Concen: 60.015 ng
RT: 8.628 min Scan# 11
Delta R.T. 0.024 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:113 Resp: 235861

Ion	Ratio	Lower	Upper
113	100		
55	146.7	129.6	169.6
56	115.4	104.0	144.0







#36
4-Chloro-3-methylphenol
Concen: 60.807 ng
RT: 8.716 min Scan# 1126
Delta R.T. 0.012 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:107 Resp: 874022

Ion	Ratio	Lower	Upper
107	100		
144	24.6	19.0	28.6
142	75.4	59.2	88.8







#37
2-Methylnaphthalene
Concen: 58.187 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:142 Resp: 1744374

Ion	Ratio	Lower	Upper
142	100		
141	89.8	71.2	106.8







#38
1-Methylnaphthalene
Concen: 58.285 ng
RT: 8.975 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:142 Resp: 1692113

Ion	Ratio	Lower	Upper
142	100		
141	92.0	74.2	111.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:164 Resp: 497194

Ion	Ratio	Lower	Upper
164	100		
162	102.3	81.8	122.6
160	45.4	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 59.627 ng
RT: 9.045 min Scan# 1182
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:216 Resp: 903928

Ion	Ratio	Lower	Upper
216	100		
214	78.9	61.9	92.9
179	22.3	18.1	27.1
108	18.2	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 62.842 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:237 Resp: 375919

Ion	Ratio	Lower	Upper
237	100		
235	63.2	43.1	83.1
272	12.6	0.0	32.0







#42
2,4,6-Tribromophenol
Concen: 125.609 ng
RT: 10.710 min Scan# 1465
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:330 Resp: 791365

Ion	Ratio	Lower	Upper
330	100		
332	96.6	77.6	116.4
141	28.7	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 60.479 ng
RT: 9.151 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:196 Resp: 597686

Ion	Ratio	Lower	Upper
196	100		
198	99.8	80.6	121.0
200	32.1	25.8	38.6







#44
2,4,5-Trichlorophenol
Concen: 61.642 ng
RT: 9.192 min Scan# 1207
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:196 Resp: 618930

Ion	Ratio	Lower	Upper
196	100		
198	101.7	80.7	121.1
97	52.1	42.1	63.1
132	29.2	23.4	35.2







#45
2-Fluorobiphenyl
Concen: 113.827 ng
RT: 9.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:172 Resp: 3716590

Ion	Ratio	Lower	Upper
172	100		
171	37.3	29.3	43.9
170	24.7	19.4	29.0







#46
1,1'-Biphenyl
Concen: 56.634 ng
RT: 9.345 min Scan# 1233
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:154 Resp: 2134829

Ion	Ratio	Lower	Upper
154	100		
153	41.8	21.1	61.1
76	11.4	0.0	31.3







#47
2-Chloronaphthalene
Concen: 57.879 ng
RT: 9.369 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 1625973

Ion	Ratio	Lower	Upper
162	100		
127	44.5	35.5	53.3
164	33.5	26.6	40.0







#48
2-Nitroaniline
Concen: 61.368 ng
RT: 9.463 min Scan# 1253
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion: 65 Resp: 495229

Ion	Ratio	Lower	Upper
65	100		
92	71.7	56.6	84.8
138	112.3	88.6	133.0







#49
Acenaphthylene
Concen: 57.125 ng
RT: 9.786 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 2379216

Ion	Ratio	Lower	Upper
152	100		
151	20.8	16.3	24.5
153	14.2	11.1	16.7







#50
Dimethylphthalate
Concen: 58.083 ng
RT: 9.651 min Scan# 1285
Delta R.T. 0.012 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:163 Resp: 2006480

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.1	4.7
164	10.4	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 61.441 ng
RT: 9.704 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 441383

Ion	Ratio	Lower	Upper
165	100		
63	58.6	47.0	70.4
89	58.6	47.4	71.2







#52
Acenaphthene
Concen: 59.065 ng
RT: 9.957 min Scan# 1337
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:154 Resp: 1737049

Ion	Ratio	Lower	Upper
154	100		
153	112.2	90.7	136.1
152	51.6	41.1	61.7







#53
3-Nitroaniline
Concen: 60.945 ng
RT: 9.875 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:138 Resp: 449553

Ion	Ratio	Lower	Upper
138	100		
108	11.5	9.3	13.9
92	116.5	93.5	140.3







#54
2,4-Dinitrophenol
Concen: 60.329 ng
RT: 9.975 min Scan# 1340
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:184 Resp: 240530

Ion	Ratio	Lower	Upper
184	100		
63	70.8	60.2	90.4
154	118.6	132.9	199.3#







#55
Dibenzofuran
Concen: 56.781 ng
RT: 10.127 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:168 Resp: 2380973

Ion	Ratio	Lower	Upper
168	100		
139	42.4	33.9	50.9
169	14.0	11.0	16.4







#56
4-Nitrophenol
Concen: 62.021 ng
RT: 10.027 min Scan# 1349
Delta R.T. 0.012 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:139 Resp: 343657

Ion	Ratio	Lower	Upper
139	100		
109	80.6	61.6	101.6
65	97.3	79.4	119.4







#57
2,4-Dinitrotoluene
Concen: 60.461 ng
RT: 10.110 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 571731

Ion	Ratio	Lower	Upper
165	100		
63	45.8	35.4	53.2
89	72.1	56.6	84.8
182	2.9	2.2	3.4







#58
Fluorene
Concen: 56.859 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:166 Resp: 1840654

Ion	Ratio	Lower	Upper
166	100		
165	98.3	79.0	118.4
167	13.9	11.0	16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 60.002 ng
RT: 10.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:232 Resp: 546621

Ion	Ratio	Lower	Upper
232	100		
131	46.3	37.3	55.9
130	2.4	1.9	2.9
166	31.1	24.8	37.2







#60
Diethylphthalate
Concen: 57.552 ng
RT: 10.345 min Scan# 1403
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:149 Resp: 1995827

Ion	Ratio	Lower	Upper
149	100		
177	23.0	18.0	27.0
150	13.0	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 57.993 ng
RT: 10.463 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:204 Resp: 977819

Ion	Ratio	Lower	Upper
204	100		
206	33.6	26.8	40.2
141	62.2	52.4	78.6







#62
4-Nitroaniline
Concen: 59.702 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.012 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:138 Resp: 426741

Ion	Ratio	Lower	Upper
138	100		
92	52.2	32.6	72.6
108	85.2	65.0	105.0







#63
Azobenzene
Concen: 56.784 ng
RT: 10.622 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 1829282

Ion	Ratio	Lower	Upper
77	100		
182	28.1	6.6	46.6
105	22.0	1.3	41.3
51	30.6	10.4	50.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:188 Resp: 827202

Ion	Ratio	Lower	Upper
188	100		
94	8.8	6.8	10.2
80	10.0	7.6	11.4







#65
4,6-Dinitro-2-methylphenol
Concen: 67.356 ng
RT: 10.516 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:198 Resp: 330315

Ion	Ratio	Lower	Upper
198	100		
51	44.2	22.3	62.3
105	43.4	23.0	63.0







#66
n-Nitrosodiphenylamine
Concen: 59.270 ng
RT: 10.580 min Scan# 1443
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:169 Resp: 1583832

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.7	80.5
167	36.6	29.2	43.8







#67
4-Bromophenyl-phenylether
Concen: 59.044 ng
RT: 10.951 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:248 Resp: 608246

Ion	Ratio	Lower	Upper
248	100		
250	97.0	77.8	116.8
141	75.1	62.6	93.8







#68
Hexachlorobenzene
Concen: 59.144 ng
RT: 11.016 min Scan# 1517
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:284 Resp: 673465

Ion	Ratio	Lower	Upper
284	100		
142	31.1	27.0	40.6
249	30.8	25.2	37.8







#70
Pentachlorophenol
Concen: 65.893 ng
RT: 11.204 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:266 Resp: 462709

Ion	Ratio	Lower	Upper
266	100		
268	63.5	51.0	76.6
264	63.1	50.2	75.2







#71
Phenanthrene
Concen: 57.232 ng
RT: 11.433 min Scan# 1588
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:178 Resp: 2560092

Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.8	23.8
179	15.8	12.6	19.0







#72
Anthracene
Concen: 57.281 ng
RT: 11.486 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:178 Resp: 2570717

Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.1	22.7
179	15.9	12.6	19.0







#73
Carbazole
Concen: 57.284 ng
RT: 11.633 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:167 Resp: 2213331

Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.4	26.2
139	14.2	11.0	16.6







#74
Di-n-butylphthalate
Concen: 57.161 ng
RT: 11.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 2916048

Ion	Ratio	Lower	Upper
149	100		
150	9.8	7.6	11.4
104	5.6	4.4	6.6







#75
Fluoranthene
Concen: 55.984 ng
RT: 12.616 min Scan# 1789
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:202 Resp: 2628993

Ion	Ratio	Lower	Upper
202	100		
101	10.0	0.0	29.7
203	18.0	0.0	37.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:240 Resp: 468454

Ion	Ratio	Lower	Upper
240	100		
120	9.7	8.4	12.6
236	25.3	20.5	30.7







#77
Benzidine
Concen: 58.309 ng
RT: 12.733 min Scan# 1809
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:184 Resp: 386007

Ion	Ratio	Lower	Upper
184	100		
185	15.1	11.5	17.3
183	11.9	9.7	14.5







#78
Pyrene
Concen: 63.781 ng
RT: 12.845 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:202 Resp: 2562337

Ion	Ratio	Lower	Upper
202	100		
200	22.0	17.3	25.9
203	18.4	14.4	21.6







#79
Terphenyl-d14
Concen: 125.675 ng
RT: 12.992 min Scan# 1853
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:244 Resp: 3982780

Ion	Ratio	Lower	Upper
244	100		
212	7.3	6.0	9.0
122	9.4	7.7	11.5







#80
Butylbenzylphthalate
Concen: 63.676 ng
RT: 13.463 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 1008043

Ion	Ratio	Lower	Upper
149	100		
91	70.7	56.5	84.7
206	21.6	17.3	25.9







#81
Benzo(a)anthracene
Concen: 60.040 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:228 Resp: 1850347

Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.8	32.8
229	20.1	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 62.091 ng
RT: 13.992 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 553078

Ion	Ratio	Lower	Upper
252	100		
254	63.6	51.0	76.6
126	10.7	8.4	12.6







#83
Chrysene
Concen: 61.596 ng
RT: 14.068 min Scan# 2036
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:228 Resp: 1705257

Ion	Ratio	Lower	Upper
228	100		
226	30.6	24.1	36.1
229	20.2	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 61.582 ng
RT: 14.021 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 1337474

Ion	Ratio	Lower	Upper
149	100		
167	27.1	21.4	32.0
279	4.9	3.8	5.6







#85
Di-n-octyl phthalate
Concen: 64.613 ng
RT: 14.633 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:149 Resp: 1953184

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.9	6.9	10.3







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.509 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:264 Resp: 407346

Ion	Ratio	Lower	Upper
264	100		
260	23.1	19.1	28.7
265	21.4	17.5	26.3







#87
Indeno(1,2,3-cd)pyrene
Concen: 68.115 ng
RT: 17.009 min Scan# 2536
Delta R.T. 0.012 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:276 Resp: 1799372

Ion	Ratio	Lower	Upper
276	100		
138	22.7	18.4	27.6
277	25.2	20.3	30.5







#88
Benzo(b)fluoranthene
Concen: 61.481 ng
RT: 15.080 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 1698744

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.5	26.3
125	9.3	7.0	10.4







#89
Benzo(k)fluoranthene
Concen: 61.253 ng
RT: 15.115 min Scan# 2214
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:252 Resp: 1478389

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	8.7	7.2	10.8







#90
Benzo(a)pyrene
Concen: 60.747 ng
RT: 15.451 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 1323726

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.4	26.0
125	10.0	7.8	11.8







#91
Dibenzo(a,h)anthracene
Concen: 66.981 ng
RT: 17.027 min Scan# 2539
Delta R.T. 0.006 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Tgt Ion:278 Resp: 1459934

Ion	Ratio	Lower	Upper
278	100		
139	14.7	11.4	17.2
279	24.3	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 65.942 ng
RT: 17.462 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF141903.D
Acq: 10 Mar 2025 13:57

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:276 Resp: 1425582

Ion	Ratio	Lower	Upper
276	100		
277	23.6	18.7	28.1
138	19.0	15.8	23.8



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141904.D
Acq On : 10 Mar 2025 14:27
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED
Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025

Quant Time: Mar 10 15:37:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.887	152	214737	20.000	ng		0.00
21) Naphthalene-d8	8.169	136	826030	20.000	ng		0.00
39) Acenaphthene-d10	9.922	164	452864	20.000	ng		0.00
64) Phenanthrene-d10	11.404	188	738472	20.000	ng		0.00
76) Chrysene-d12	14.045	240	424370	20.000	ng		0.00
86) Perylene-d12	15.510	264	391768	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.498	112	1930406	149.991	ng		0.00
7) Phenol-d6	6.522	99	2446183	149.235	ng		0.02
23) Nitrobenzene-d5	7.457	82	2335417	158.463	ng		0.01
42) 2,4,6-Tribromophenol	10.716	330	986206	171.858	ng		0.01
45) 2-Fluorobiphenyl	9.245	172	4513418	151.763	ng		0.00
79) Terphenyl-d14	12.992	244	4700026	163.713	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.669	88	436007	80.831	ng		98
3) Pyridine	3.440	79	1059473	79.923	ng		99
4) n-Nitrosodimethylamine	3.416	42	521442	82.752	ng		99
6) Aniline	6.563	93	1160437	71.712	ng		99
8) 2-Chlorophenol	6.675	128	1078659	75.616	ng		98
10) Phenol	6.534	94	1285145	74.308	ng		89
11) bis(2-Chloroethyl)ether	6.628	93	968280	74.725	ng		99
12) 1,3-Dichlorobenzene	6.828	146	1173924	76.153	ng		100
13) 1,4-Dichlorobenzene	6.904	146	1181975	75.784	ng		99
14) 1,2-Dichlorobenzene	7.063	146	1093316	74.440	ng		99
15) Benzyl Alcohol	7.034	79	1013314	76.376	ng		98
16) 2,2'-oxybis(1-Chloropr...	7.163	45	1127534	72.198	ng		98
17) 2-Methylphenol	7.134	107	876665	77.326	ng		98
18) Hexachloroethane	7.398	117	448765	76.745	ng		96
19) n-Nitroso-di-n-propyla...	7.316	70	783318	74.351	ng		100
20) 3+4-Methylphenols	7.298	107	1050191	72.219	ng	#	84
22) Acetophenone	7.304	105	1496035	75.498	ng	#	95
24) Nitrobenzene	7.475	77	1162753	79.572	ng		100
25) Isophorone	7.722	82	2056405	79.032	ng		100
26) 2-Nitrophenol	7.787	139	599261	83.688	ng		99
27) 2,4-Dimethylphenol	7.822	122	767514	77.821	ng		99
28) bis(2-Chloroethoxy)met...	7.916	93	1248549	76.598	ng		100
29) 2,4-Dichlorophenol	8.028	162	971886	79.403	ng		99
30) 1,2,4-Trichlorobenzene	8.110	180	1062793	78.705	ng		99
31) Naphthalene	8.192	128	3152444	74.497	ng		99
32) Benzoic acid	7.975	122	795850	92.323	ng		98
33) 4-Chloroaniline	8.239	127	1088546	72.896	ng		100
34) Hexachlorobutadiene	8.304	225	713827	81.299	ng		99
35) Caprolactam	8.639	113	311758m	83.763	ng		
36) 4-Chloro-3-methylphenol	8.716	107	1056917	77.644	ng		99
37) 2-Methylnaphthalene	8.881	142	2129990	75.023	ng		100
38) 1-Methylnaphthalene	8.981	142	2050191	74.568	ng		99
40) 1,2,4,5-Tetrachloroben...	9.045	216	1127671	81.667	ng		98
41) Hexachlorocyclopentadiene	9.028	237	463300	85.031	ng		99
43) 2,4,6-Trichlorophenol	9.157	196	741710	82.399	ng		100
44) 2,4,5-Trichlorophenol	9.198	196	736926	80.579	ng		100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141904.D
Acq On : 10 Mar 2025 14:27
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED
Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025

Quant Time: Mar 10 15:37:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

	Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
46)	1,1'-Biphenyl	9.345	154	2626697	76.503	ng		99
47)	2-Chloronaphthalene	9.369	162	1992242	77.859	ng		99
48)	2-Nitroaniline	9.463	65	597778	81.326	ng		98
49)	Acenaphthylene	9.786	152	2878496	75.877	ng		99
50)	Dimethylphthalate	9.657	163	2428843	77.192	ng		100
51)	2,6-Dinitrotoluene	9.710	165	529139	80.867	ng		99
52)	Acenaphthene	9.957	154	2085941	77.872	ng		99
53)	3-Nitroaniline	9.881	138	526980	78.435	ng		100
54)	2,4-Dinitrophenol	9.981	184	305399	81.479	ng	#	55
55)	Dibenzofuran	10.128	168	2844432	74.474	ng		99
56)	4-Nitrophenol	10.033	139	416153	82.456	ng		97
57)	2,4-Dinitrotoluene	10.116	165	681278	79.099	ng		95
58)	Fluorene	10.475	166	2210653	74.974	ng		99
59)	2,3,4,6-Tetrachlorophenol	10.245	232	668027	80.507	ng		99
60)	Diethylphthalate	10.345	149	2358921	74.681	ng		99
61)	4-Chlorophenyl-phenyle...	10.463	204	1209021	78.724	ng		97
62)	4-Nitroaniline	10.504	138	502894	77.243	ng		98
63)	Azobenzene	10.622	77	2168742	73.911	ng		99
65)	4,6-Dinitro-2-methylph...	10.522	198	404505	92.396	ng		98
66)	n-Nitrosodiphenylamine	10.586	169	1882098	78.895	ng		99
67)	4-Bromophenyl-phenylether	10.957	248	774645	84.232	ng		95
68)	Hexachlorobenzene	11.016	284	862357	84.832	ng		97
70)	Pentachlorophenol	11.210	266	563030	89.813	ng		100
71)	Phenanthrene	11.433	178	3020950	75.649	ng		99
72)	Anthracene	11.486	178	3031014	75.653	ng		99
73)	Carbazole	11.639	167	2552563	74.002	ng		98
74)	Di-n-butylphthalate	11.969	149	3342776	73.399	ng		99
75)	Fluoranthene	12.622	202	3005566	71.694	ng		99
77)	Benzidine	12.733	184	427770	71.330	ng		99
78)	Pyrene	12.851	202	2930115	80.513	ng		99
80)	Butylbenzylphthalate	13.463	149	1149989	80.189	ng		99
81)	Benzo(a)anthracene	14.033	228	2196935	78.692	ng		100
82)	3,3'-Dichlorobenzidine	13.998	252	652678	80.884	ng		99
83)	Chrysene	14.074	228	2018832	80.498	ng		99
84)	Bis(2-ethylhexyl)phtha...	14.021	149	1573231	79.962	ng		98
85)	Di-n-octyl phthalate	14.633	149	2310836	84.386	ng		100
87)	Indeno(1,2,3-cd)pyrene	17.009	276	2238588	88.111	ng		99
88)	Benzo(b)fluoranthene	15.086	252	2246428	84.536	ng		99
89)	Benzo(k)fluoranthene	15.116	252	1631675	70.292	ng		99
90)	Benzo(a)pyrene	15.457	252	1725632	82.339	ng		100
91)	Dibenzo(a,h)anthracene	17.033	278	1846778	88.098	ng		99
92)	Benzo(g,h,i)perylene	17.468	276	1828662	87.950	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141904.D
Acq On : 10 Mar 2025 14:27
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1


Quant Time: Mar 10 15:37:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:152 Resp: 21473


Ion	Ratio	Lower	Upper
152	100		
150	164.7	127.4	191.2
115	64.2	51.9	77.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#2
1,4-Dioxane
Concen: 80.831 ng
RT: 2.669 min Scan# 98
Delta R.T. -0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 88 Resp: 436007

Ion	Ratio	Lower	Upper
88	100		
58	70.3	57.4	86.2
43	27.6	21.5	32.3






#3
Pyridine
Concen: 79.923 ng
RT: 3.440 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 79 Resp: 105947


Ion	Ratio	Lower	Upper
79	100		
52	66.9	53.3	79.9
51	35.0	27.6	41.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#4
n-Nitrosodimethylamine
Concen: 82.752 ng
RT: 3.416 min Scan# 225
Delta R.T. 0.024 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 42 Resp: 521442

Ion	Ratio	Lower	Upper
42	100		
74	130.6	103.4	155.0
44	6.0	5.0	7.4






#5
2-Fluorophenol
Concen: 149.991 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:112 Resp: 193040


Ion	Ratio	Lower	Upper
112	100		
64	58.7	46.5	69.7
63	32.5	25.4	38.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#6
Aniline
Concen: 71.712 ng
RT: 6.563 min Scan# 760
Delta R.T. 0.018 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 93 Resp: 1160437

Ion	Ratio	Lower	Upper
93	100		
66	40.9	31.9	47.9
65	19.7	15.8	23.6






#7
Phenol-d6
Concen: 149.235 ng
RT: 6.522 min Scan# 7
Delta R.T. 0.018 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 99 Resp: 244618


Ion	Ratio	Lower	Upper
99	100		
42	18.5	14.6	21.8
71	39.2	30.8	46.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#8
2-Chlorophenol
Concen: 75.616 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:128 Resp: 1078659

Ion	Ratio	Lower	Upper
128	100		
130	32.7	12.8	52.8
64	46.4	24.4	64.4






#10
Phenol
Concen: 74.308 ng
RT: 6.534 min Scan# 7
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 94 Resp: 128514


Ion	Ratio	Lower	Upper
94	100		
65	34.4	10.5	50.5
66	54.3	25.8	65.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#11
bis(2-Chloroethyl)ether
Concen: 74.725 ng
RT: 6.628 min Scan# 771
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 93 Resp: 968280

Ion	Ratio	Lower	Upper
93	100		
63	73.9	53.0	93.0
95	33.4	12.9	52.9






#12
1,3-Dichlorobenzene
Concen: 76.153 ng
RT: 6.828 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:146 Resp: 117392


Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.3	76.9
75	33.6	26.6	40.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#13
1,4-Dichlorobenzene
Concen: 75.784 ng
RT: 6.904 min Scan# 818
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:146 Resp: 1181975

Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.0	76.6
111	44.6	35.7	53.5






#14
1,2-Dichlorobenzene
Concen: 74.440 ng
RT: 7.063 min Scan# 84
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:146 Resp: 109331


Ion	Ratio	Lower	Upper
146	100		
148	64.1	50.8	76.2
111	46.1	38.0	57.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#15
Benzyl Alcohol
Concen: 76.376 ng
RT: 7.034 min Scan# 840
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 79 Resp: 1013314

Ion	Ratio	Lower	Upper
79	100		
108	72.0	59.7	89.5
77	63.1	51.0	76.6





#16
2,2'-oxybis(1-Chloropropane)
Concen: 72.198 ng
RT: 7.163 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Tgt Ion: 45 Resp: 112753


Ion	Ratio	Lower	Upper
45	100		
77	19.3	0.0	37.8
79	14.8	0.0	34.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#17
2-Methylphenol
Concen: 77.326 ng
RT: 7.134 min Scan# 857
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:107 Resp: 876665

Ion	Ratio	Lower	Upper
107	100		
108	111.3	90.2	135.2
77	50.5	39.6	59.4
79	50.8	39.1	58.7






#18
Hexachloroethane
Concen: 76.745 ng
RT: 7.398 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:117 Resp: 44876


Ion	Ratio	Lower	Upper
117	100		
119	96.4	76.9	115.3
201	107.9	79.9	119.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#19
n-Nitroso-di-n-propylamine
Concen: 74.351 ng
RT: 7.316 min Scan# 888
Delta R.T. 0.018 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 70 Resp: 783318

Ion	Ratio	Lower	Upper
70	100		
42	51.4	41.4	62.0
101	10.2	8.2	12.2
130	23.0	18.2	27.4






#20
3+4-Methylphenols
Concen: 72.219 ng
RT: 7.298 min Scan# 8
Delta R.T. 0.018 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:107 Resp: 105019


Ion	Ratio	Lower	Upper
107	100		
108	89.5	68.4	108.4
77	123.4	69.7	109.7
79	32.9	11.4	51.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:136 Resp: 826030

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	7.0	5.8	8.8
68	4.7	4.1	6.1






#22
Acetophenone
Concen: 75.498 ng
RT: 7.304 min Scan# 8
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:105 Resp: 149603


Ion	Ratio	Lower	Upper
105	100		
71	5.4	5.8	8.8#
51	33.0	23.4	35.0
120	22.2	17.3	25.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#23
Nitrobenzene-d5
Concen: 158.463 ng
RT: 7.457 min Scan# 912
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 82 Resp: 2335417

Ion	Ratio	Lower	Upper
82	100		
128	45.3	36.0	54.0
54	50.1	39.6	59.4






#24
Nitrobenzene
Concen: 79.572 ng
RT: 7.475 min Scan# 9
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 77 Resp: 116275


Ion	Ratio	Lower	Upper
77	100		
123	46.2	37.0	55.4
65	14.4	11.3	16.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#25
Isophorone
Concen: 79.032 ng
RT: 7.722 min Scan# 957
Delta R.T. 0.018 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 82 Resp: 2056405

Ion	Ratio	Lower	Upper
82	100		
95	8.2	6.4	9.6
138	18.8	14.9	22.3






#26
2-Nitrophenol
Concen: 83.688 ng
RT: 7.787 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:139 Resp: 59926


Ion	Ratio	Lower	Upper
139	100		
109	34.4	26.5	39.7
65	50.8	40.3	60.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#27
2,4-Dimethylphenol
Concen: 77.821 ng
RT: 7.822 min Scan# 974
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:122 Resp: 767514

Ion	Ratio	Lower	Upper
122	100		
107	120.5	94.8	142.2
121	60.1	47.8	71.6






#28
bis(2-Chloroethoxy)methane
Concen: 76.598 ng
RT: 7.916 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 93 Resp: 124854


Ion	Ratio	Lower	Upper
93	100		
95	33.2	26.6	40.0
123	12.4	10.2	15.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#29
2,4-Dichlorophenol
Concen: 79.403 ng
RT: 8.028 min Scan# 1009
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:162 Resp: 971886

Ion	Ratio	Lower	Upper
162	100		
164	64.4	43.8	83.8
98	37.1	18.1	58.1






#30
1,2,4-Trichlorobenzene
Concen: 78.705 ng
RT: 8.110 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:180 Resp: 106279


Ion	Ratio	Lower	Upper
180	100		
182	96.1	77.5	116.3
145	32.2	25.8	38.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#31
Naphthalene
Concen: 74.497 ng
RT: 8.192 min Scan# 1037
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:128 Resp: 3152444

Ion	Ratio	Lower	Upper
128	100		
129	11.6	9.0	13.6
127	14.0	10.8	16.2






#32
Benzoic acid
Concen: 92.323 ng
RT: 7.975 min Scan# 1
Delta R.T. 0.047 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:122 Resp: 79585


Ion	Ratio	Lower	Upper
122	100		
105	131.3	107.9	147.9
77	98.9	79.0	119.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#33
4-Chloroaniline
Concen: 72.896 ng
RT: 8.239 min Scan# 1045
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:127 Resp: 1088546

Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.2	39.4
65	31.8	25.4	38.2
92	20.8	16.2	24.4





#34
Hexachlorobutadiene
Concen: 81.299 ng
RT: 8.304 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Tgt Ion:225 Resp: 71382


Ion	Ratio	Lower	Upper
225	100		
223	63.1	50.6	75.8
227	64.3	52.2	78.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#35
Caprolactam
Concen: 83.763 ng m
RT: 8.639 min Scan# 1113
Delta R.T. 0.035 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:113 Resp: 311758

Ion	Ratio	Lower	Upper
113	100		
55	146.1	129.6	169.6
56	118.7	104.0	144.0





#36
4-Chloro-3-methylphenol
Concen: 77.644 ng
RT: 8.716 min Scan# 1
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Tgt Ion:107 Resp: 105691


Ion	Ratio	Lower	Upper
107	100		
144	24.0	19.0	28.6
142	72.9	59.2	88.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#37
2-Methylnaphthalene
Concen: 75.023 ng
RT: 8.881 min Scan# 1154
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:142 Resp: 2129990

Ion	Ratio	Lower	Upper
142	100		
141	89.1	71.2	106.8






#38
1-Methylnaphthalene
Concen: 74.568 ng
RT: 8.981 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:142 Resp: 205019


Ion	Ratio	Lower	Upper
142	100		
141	91.8	74.2	111.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:164 Resp: 452864

Ion	Ratio	Lower	Upper
164	100		
162	103.9	81.8	122.6
160	45.9	36.7	55.1






#40
1,2,4,5-Tetrachlorobenzene
Concen: 81.667 ng
RT: 9.045 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:216 Resp: 112767


Ion	Ratio	Lower	Upper
216	100		
214	79.3	61.9	92.9
179	22.7	18.1	27.1
108	18.2	14.7	22.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#41
Hexachlorocyclopentadiene
Concen: 85.031 ng
RT: 9.028 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:237 Resp: 463300

Ion	Ratio	Lower	Upper
237	100		
235	62.6	43.1	83.1
272	12.4	0.0	32.0






#42
2,4,6-Tribromophenol
Concen: 171.858 ng
RT: 10.716 min Scan# 14
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:330 Resp: 98620


Ion	Ratio	Lower	Upper
330	100		
332	96.4	77.6	116.4
141	27.3	24.7	37.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#43
2,4,6-Trichlorophenol
Concen: 82.399 ng
RT: 9.157 min Scan# 1201
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:196 Resp: 741710

Ion	Ratio	Lower	Upper
196	100		
198	100.7	80.6	121.0
200	32.7	25.8	38.6






#44
2,4,5-Trichlorophenol
Concen: 80.579 ng
RT: 9.198 min Scan# 1
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:196 Resp: 73692


Ion	Ratio	Lower	Upper
196	100		
198	100.6	80.7	121.1
97	52.2	42.1	63.1
132	29.5	23.4	35.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#45
2-Fluorobiphenyl
Concen: 151.763 ng
RT: 9.245 min Scan# 1216
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:172 Resp: 4513418

Ion	Ratio	Lower	Upper
172	100		
171	38.4	29.3	43.9
170	25.5	19.4	29.0






#46
1,1'-Biphenyl
Concen: 76.503 ng
RT: 9.345 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:154 Resp: 262669


Ion	Ratio	Lower	Upper
154	100		
153	41.9	21.1	61.1
76	11.7	0.0	31.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#47
2-Chloronaphthalene
Concen: 77.859 ng
RT: 9.369 min Scan# 1237
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:162 Resp: 1992242

Ion	Ratio	Lower	Upper
162	100		
127	45.0	35.5	53.3
164	33.4	26.6	40.0





#48
2-Nitroaniline
Concen: 81.326 ng
RT: 9.463 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Tgt Ion: 65 Resp: 59777


Ion	Ratio	Lower	Upper
65	100		
92	70.8	56.6	84.8
138	106.9	88.6	133.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#49
Acenaphthylene
Concen: 75.877 ng
RT: 9.786 min Scan# 1308
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:152 Resp: 2878496

Ion	Ratio	Lower	Upper
152	100		
151	20.9	16.3	24.5
153	14.2	11.1	16.7





#50
Dimethylphthalate
Concen: 77.192 ng
RT: 9.657 min Scan# 1
Delta R.T. 0.018 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Tgt Ion:163 Resp: 242884


Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.1	4.7
164	10.4	8.2	12.4

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#51
2,6-Dinitrotoluene
Concen: 80.867 ng
RT: 9.710 min Scan# 1295
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:165 Resp: 529139

Ion	Ratio	Lower	Upper
165	100		
63	57.7	47.0	70.4
89	58.8	47.4	71.2






#52
Acenaphthene
Concen: 77.872 ng
RT: 9.957 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:154 Resp: 208594


Ion	Ratio	Lower	Upper
154	100		
153	112.4	90.7	136.1
152	52.0	41.1	61.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#53
3-Nitroaniline
Concen: 78.435 ng
RT: 9.881 min Scan# 1324
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:138 Resp: 526980

Ion	Ratio	Lower	Upper
138	100		
108	11.6	9.3	13.9
92	116.4	93.5	140.3





#54
2,4-Dinitrophenol
Concen: 81.479 ng
RT: 9.981 min Scan# 1
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Tgt Ion:184 Resp: 30539


Ion	Ratio	Lower	Upper
184	100		
63	68.1	60.2	90.4
154	82.9	132.9	199.3#

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#55
Dibenzofuran
Concen: 74.474 ng
RT: 10.128 min Scan# 1366
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:168 Resp: 2844432

Ion	Ratio	Lower	Upper
168	100		
139	43.4	33.9	50.9
169	14.1	11.0	16.4





#56
4-Nitrophenol
Concen: 82.456 ng
RT: 10.033 min Scan# 1
Delta R.T. 0.018 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Tgt Ion:139 Resp: 41615


Ion	Ratio	Lower	Upper
139	100		
109	84.3	61.6	101.6
65	96.7	79.4	119.4

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#57
2,4-Dinitrotoluene
Concen: 79.099 ng
RT: 10.116 min Scan# 1364
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:165 Resp: 681278

Ion	Ratio	Lower	Upper
165	100		
63	49.4	35.4	53.2
89	73.9	56.6	84.8
182	2.7	2.2	3.4






#58
Fluorene
Concen: 74.974 ng
RT: 10.475 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:166 Resp: 221065


Ion	Ratio	Lower	Upper
166	100		
165	97.5	79.0	118.4
167	13.8	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#59
2,3,4,6-Tetrachlorophenol
Concen: 80.507 ng
RT: 10.245 min Scan# 1386
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:232 Resp: 668027

Ion	Ratio	Lower	Upper
232	100		
131	45.4	37.3	55.9
130	2.3	1.9	2.9
166	30.4	24.8	37.2






#60
Diethylphthalate
Concen: 74.681 ng
RT: 10.345 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:149 Resp: 235892


Ion	Ratio	Lower	Upper
149	100		
177	23.1	18.0	27.0
150	13.3	10.2	15.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#61
4-Chlorophenyl-phenylether
Concen: 78.724 ng
RT: 10.463 min Scan# 1423
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:204 Resp: 1209021

Ion	Ratio	Lower	Upper
204	100		
206	33.0	26.8	40.2
141	62.1	52.4	78.6






#62
4-Nitroaniline
Concen: 77.243 ng
RT: 10.504 min Scan# 14
Delta R.T. 0.024 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:138 Resp: 50289


Ion	Ratio	Lower	Upper
138	100		
92	52.8	32.6	72.6
108	88.1	65.0	105.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#63
Azobenzene
Concen: 73.911 ng
RT: 10.622 min Scan# 1450
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion: 77 Resp: 2168742

Ion	Ratio	Lower	Upper
77	100		
182	27.2	6.6	46.6
105	21.7	1.3	41.3
51	31.5	10.4	50.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Tgt Ion:188 Resp: 73847


Ion	Ratio	Lower	Upper
188	100		
94	8.5	6.8	10.2
80	9.7	7.6	11.4

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#65
4,6-Dinitro-2-methylphenol
Concen: 92.396 ng
RT: 10.522 min Scan# 1433
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:198 Resp: 404505

Ion	Ratio	Lower	Upper
198	100		
51	44.2	22.3	62.3
105	43.9	23.0	63.0






#66
n-Nitrosodiphenylamine
Concen: 78.895 ng
RT: 10.586 min Scan# 14
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:169 Resp: 188209


Ion	Ratio	Lower	Upper
169	100		
168	67.8	53.7	80.5
167	36.9	29.2	43.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#67
4-Bromophenyl-phenylether
Concen: 84.232 ng
RT: 10.957 min Scan# 1507
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:248 Resp: 774645

Ion	Ratio	Lower	Upper
248	100		
250	98.1	77.8	116.8
141	69.9	62.6	93.8





#68
Hexachlorobenzene
Concen: 84.832 ng
RT: 11.016 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Instrument :
BNA_F
ClientSampleId :
SSTDICC080


Tgt Ion:284 Resp: 86235

Ion	Ratio	Lower	Upper
284	100		
142	31.2	27.0	40.6
249	31.0	25.2	37.8

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#70
Pentachlorophenol
Concen: 89.813 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:266 Resp: 563030

Ion	Ratio	Lower	Upper
266	100		
268	63.9	51.0	76.6
264	62.9	50.2	75.2






#71
Phenanthrene
Concen: 75.649 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:178 Resp: 302095


Ion	Ratio	Lower	Upper
178	100		
176	20.2	15.8	23.8
179	16.2	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#72
Anthracene
Concen: 75.653 ng
RT: 11.486 min Scan# 1597
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:178 Resp: 3031014

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.1	22.7
179	16.1	12.6	19.0






#73
Carbazole
Concen: 74.002 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:167 Resp: 255256


Ion	Ratio	Lower	Upper
167	100		
166	22.4	17.4	26.2
139	14.6	11.0	16.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#74
Di-n-butylphthalate
Concen: 73.399 ng
RT: 11.969 min Scan# 1679
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:149 Resp: 3342776

Ion	Ratio	Lower	Upper
149	100		
150	10.0	7.6	11.4
104	5.7	4.4	6.6






#75
Fluoranthene
Concen: 71.694 ng
RT: 12.622 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:202 Resp: 300556


Ion	Ratio	Lower	Upper
202	100		
101	9.4	0.0	29.7
203	18.4	0.0	37.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.045 min Scan# 2032
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:240 Resp: 424370

Ion	Ratio	Lower	Upper
240	100		
120	9.3	8.4	12.6
236	25.8	20.5	30.7






#77
Benzidine
Concen: 71.330 ng
RT: 12.733 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:184 Resp: 42777


Ion	Ratio	Lower	Upper
184	100		
185	14.8	11.5	17.3
183	11.9	9.7	14.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#78
Pyrene
Concen: 80.513 ng
RT: 12.851 min Scan# 1829
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:202 Resp: 2930115

Ion	Ratio	Lower	Upper
202	100		
200	22.1	17.3	25.9
203	18.2	14.4	21.6






#79
Terphenyl-d14
Concen: 163.713 ng
RT: 12.992 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:244 Resp: 470002


Ion	Ratio	Lower	Upper
244	100		
212	7.4	6.0	9.0
122	9.5	7.7	11.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#80
Butylbenzylphthalate
Concen: 80.189 ng
RT: 13.463 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:149 Resp: 1149989

Ion	Ratio	Lower	Upper
149	100		
91	71.4	56.5	84.7
206	21.7	17.3	25.9






#81
Benzo(a)anthracene
Concen: 78.692 ng
RT: 14.033 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:228 Resp: 219693


Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.8	32.8
229	20.3	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#82
3,3'-Dichlorobenzidine
Concen: 80.884 ng
RT: 13.998 min Scan# 2024
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:252 Resp: 652678

Ion	Ratio	Lower	Upper
252	100		
254	64.8	51.0	76.6
126	9.9	8.4	12.6





#83
Chrysene
Concen: 80.498 ng
RT: 14.074 min Scan# 2(
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27


Instrument :
BNA_F
ClientSampleId :
SSTDICC080


Tgt Ion:228 Resp: 201883

Ion	Ratio	Lower	Upper
228	100		
226	30.6	24.1	36.1
229	20.2	15.7	23.5

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025







#84
Bis(2-ethylhexyl)phthalate
Concen: 79.962 ng
RT: 14.021 min Scan# 2028
Delta R.T. 0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:149 Resp: 1573231

Ion	Ratio	Lower	Upper
149	100		
167	27.7	21.4	32.0
279	5.2	3.8	5.6






#85
Di-n-octyl phthalate
Concen: 84.386 ng
RT: 14.633 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:149 Resp: 231083


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.7	6.9	10.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.510 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:264 Resp: 391768

Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.1	28.7
265	21.7	17.5	26.3






#87
Indeno(1,2,3-cd)pyrene
Concen: 88.111 ng
RT: 17.009 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:276 Resp: 223858


Ion	Ratio	Lower	Upper
276	100		
138	21.6	18.4	27.6
277	25.5	20.3	30.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#88
Benzo(b)fluoranthene
Concen: 84.536 ng
RT: 15.086 min Scan# 2209
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:252 Resp: 2246428

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.5	26.3
125	8.7	7.0	10.4






#89
Benzo(k)fluoranthene
Concen: 70.292 ng
RT: 15.116 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:252 Resp: 163167


Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.7	26.5
125	8.8	7.2	10.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025




#90
Benzo(a)pyrene
Concen: 82.339 ng
RT: 15.457 min Scan# 2272
Delta R.T. 0.006 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:252 Resp: 1725632

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.0
125	9.7	7.8	11.8






#91
Dibenzo(a,h)anthracene
Concen: 88.098 ng
RT: 17.033 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:278 Resp: 184677


Ion	Ratio	Lower	Upper
278	100		
139	14.1	11.4	17.2
279	24.1	18.8	28.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/11/2025
Supervised By :Jagrut Upadhyay 03/11/2025



#92
Benzo(g,h,i)perylene
Concen: 87.950 ng
RT: 17.468 min Scan# 2614
Delta R.T. 0.018 min
Lab File: BF141904.D
Acq: 10 Mar 2025 14:27

Tgt Ion:276 Resp: 1828662

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.7	28.1
138	19.2	15.8	23.8



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141905.D
Acq On : 10 Mar 2025 15:20
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Quant Time: Mar 10 15:39:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	215017	20.000	ng		0.00
21) Naphthalene-d8	8.163	136	877329	20.000	ng		0.00
39) Acenaphthene-d10	9.922	164	481971	20.000	ng		0.00
64) Phenanthrene-d10	11.404	188	811142	20.000	ng		0.00
76) Chrysene-d12	14.045	240	491943	20.000	ng		0.00
86) Perylene-d12	15.509	264	413564	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.492	112	1249601	96.967	ng		0.00
7) Phenol-d6	6.504	99	1594608	97.156	ng		0.00
23) Nitrobenzene-d5	7.445	82	1507471	96.304	ng		0.00
42) 2,4,6-Tribromophenol	10.710	330	625035	102.342	ng		0.00
45) 2-Fluorobiphenyl	9.239	172	3064989	96.836	ng		0.00
79) Terphenyl-d14	12.992	244	3394803	102.007	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.675	88	266045	49.258	ng		99
3) Pyridine	3.434	79	646627	48.716	ng		99
4) n-Nitrosodimethylamine	3.398	42	320008	50.718	ng		99
6) Aniline	6.545	93	775854	47.884	ng		99
8) 2-Chlorophenol	6.669	128	696154	48.738	ng		100
9) Benzaldehyde	6.434	77	418049	45.569	ng		99
10) Phenol	6.522	94	829086	47.876	ng		96
11) bis(2-Chloroethyl)ether	6.616	93	631516	48.672	ng		100
12) 1,3-Dichlorobenzene	6.822	146	753352	48.807	ng		99
13) 1,4-Dichlorobenzene	6.898	146	756875	48.465	ng		100
14) 1,2-Dichlorobenzene	7.051	146	706635	48.049	ng		100
15) Benzyl Alcohol	7.022	79	652454	49.113	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.157	45	745710	47.687	ng		100
17) 2-Methylphenol	7.128	107	557981	49.153	ng		99
18) Hexachloroethane	7.392	117	290127	49.551	ng		99
19) n-Nitroso-di-n-propyla...	7.298	70	505166	47.887	ng		99
20) 3+4-Methylphenols	7.286	107	704563	48.388	ng	#	86
22) Acetophenone	7.292	105	987577	46.924	ng		97
24) Nitrobenzene	7.463	77	752936	48.514	ng		99
25) Isophorone	7.704	82	1327254	48.027	ng		99
26) 2-Nitrophenol	7.781	139	392772	51.644	ng		100
27) 2,4-Dimethylphenol	7.810	122	510395	48.725	ng		99
28) bis(2-Chloroethoxy)met...	7.910	93	826776	47.757	ng		100
29) 2,4-Dichlorophenol	8.016	162	647319	49.793	ng		100
30) 1,2,4-Trichlorobenzene	8.104	180	697274	48.617	ng		99
31) Naphthalene	8.186	128	2196111	48.863	ng		100
32) Benzoic acid	7.939	122	500593	54.676	ng		100
33) 4-Chloroaniline	8.233	127	775699	48.909	ng		99
34) Hexachlorobutadiene	8.304	225	469787	50.376	ng		99
35) Caprolactam	8.616	113	193267	48.891	ng		95
36) 4-Chloro-3-methylphenol	8.710	107	721800	49.925	ng		100
37) 2-Methylnaphthalene	8.875	142	1436485	47.638	ng		100
38) 1-Methylnaphthalene	8.975	142	1357356	46.482	ng		100
40) 1,2,4,5-Tetrachloroben...	9.039	216	719431	48.955	ng		99
41) Hexachlorocyclopentadiene	9.027	237	284968	49.142	ng		99
43) 2,4,6-Trichlorophenol	9.151	196	481782	50.291	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141905.D
Acq On : 10 Mar 2025 15:20
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Quant Time: Mar 10 15:39:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.186	196	482570	49.580	ng		99
46) 1,1'-Biphenyl	9.339	154	1809746	49.526	ng		99
47) 2-Chloronaphthalene	9.369	162	1360528	49.960	ng		99
48) 2-Nitroaniline	9.457	65	429111	54.854	ng		97
49) Acenaphthylene	9.780	152	2005538	49.674	ng		100
50) Dimethylphthalate	9.645	163	1716995	51.273	ng		100
51) 2,6-Dinitrotoluene	9.704	165	370368	53.184	ng		96
52) Acenaphthene	9.957	154	1377901	48.333	ng		98
53) 3-Nitroaniline	9.874	138	340633	47.638	ng		100
54) 2,4-Dinitrophenol	9.975	184	184410	49.104	ng	#	58
55) Dibenzofuran	10.127	168	1936780	47.647	ng		99
56) 4-Nitrophenol	10.022	139	276487	51.474	ng		99
57) 2,4-Dinitrotoluene	10.104	165	461276	50.321	ng		97
58) Fluorene	10.469	166	1497123	47.708	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.239	232	438025	49.600	ng		100
60) Diethylphthalate	10.339	149	1635073	48.639	ng		100
61) 4-Chlorophenyl-phenyle...	10.457	204	799744	48.929	ng		99
62) 4-Nitroaniline	10.486	138	346378	49.990	ng		99
63) Azobenzene	10.622	77	1519356	48.653	ng		98
65) 4,6-Dinitro-2-methylph...	10.516	198	262910	54.673	ng		99
66) n-Nitrosodiphenylamine	10.580	169	1293745	49.373	ng		100
67) 4-Bromophenyl-phenylether	10.951	248	532762	52.740	ng		97
68) Hexachlorobenzene	11.016	284	573467	51.359	ng		97
69) Atrazine	11.104	200	418147	65.441	ng		99
70) Pentachlorophenol	11.204	266	365292	53.050	ng		99
71) Phenanthrene	11.433	178	2080633	47.434	ng		100
72) Anthracene	11.480	178	2081490	47.299	ng		99
73) Carbazole	11.633	167	1831424	48.338	ng		99
74) Di-n-butylphthalate	11.963	149	2510447	50.185	ng		99
75) Fluoranthene	12.616	202	2328357	50.564	ng		99
77) Benzidine	12.733	184	270893	38.966	ng		99
78) Pyrene	12.845	202	2308671	54.723	ng		100
80) Butylbenzylphthalate	13.463	149	864849	52.022	ng		99
81) Benzo(a)anthracene	14.033	228	1592035	49.192	ng		100
82) 3,3'-Dichlorobenzidine	13.992	252	449408	48.044	ng		99
83) Chrysene	14.068	228	1465389	50.404	ng		100
84) Bis(2-ethylhexyl)phtha...	14.021	149	1203021	52.747	ng		99
85) Di-n-octyl phthalate	14.633	149	1692206	53.307	ng		100
87) Indeno(1,2,3-cd)pyrene	17.003	276	1366974	50.969	ng		99
88) Benzo(b)fluoranthene	15.080	252	1392083	49.625	ng		99
89) Benzo(k)fluoranthene	15.109	252	1167351	47.638	ng		100
90) Benzo(a)pyrene	15.451	252	1119256	50.591	ng		99
91) Dibenzo(a,h)anthracene	17.021	278	1131875	51.149	ng		99
92) Benzo(g,h,i)perylene	17.456	276	1099760	50.106	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141905.D
Acq On : 10 Mar 2025 15:20
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Quant Time: Mar 10 15:39:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 215017

Ion	Ratio	Lower	Upper
152	100		
150	158.8	127.4	191.2
115	64.3	51.9	77.9







#2
1,4-Dioxane
Concen: 49.258 ng
RT: 2.675 min Scan# 99
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion: 88 Resp: 266045

Ion	Ratio	Lower	Upper
88	100		
58	72.1	57.4	86.2
43	27.9	21.5	32.3







#3
Pyridine
Concen: 48.716 ng
RT: 3.434 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 646627

Ion	Ratio	Lower	Upper
79	100		
52	67.0	53.3	79.9
51	34.8	27.6	41.4







#4
n-Nitrosodimethylamine
Concen: 50.718 ng
RT: 3.398 min Scan# 222
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion: 42 Resp: 320008

Ion	Ratio	Lower	Upper
42	100		
74	128.4	103.4	155.0
44	6.2	5.0	7.4







#5
2-Fluorophenol
Concen: 96.967 ng
RT: 5.492 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:112 Resp: 1249601

Ion	Ratio	Lower	Upper
112	100		
64	57.5	46.5	69.7
63	31.2	25.4	38.2







#6
Aniline
Concen: 47.884 ng
RT: 6.545 min Scan# 757
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion: 93 Resp: 775854

Ion	Ratio	Lower	Upper
93	100		
66	40.4	31.9	47.9
65	19.9	15.8	23.6







#7
Phenol-d6
Concen: 97.156 ng
RT: 6.504 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 99 Resp: 1594608

Ion	Ratio	Lower	Upper
99	100		
42	18.3	14.6	21.8
71	38.7	30.8	46.2







#8
2-Chlorophenol
Concen: 48.738 ng
RT: 6.669 min Scan# 778
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:128 Resp: 696154

Ion	Ratio	Lower	Upper
128	100		
130	32.9	12.8	52.8
64	44.8	24.4	64.4







#9
Benzaldehyde
Concen: 45.569 ng
RT: 6.434 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 418049

Ion	Ratio	Lower	Upper
77	100		
105	98.6	80.0	120.0
106	95.5	74.5	114.5







#10
Phenol
Concen: 47.876 ng
RT: 6.522 min Scan# 753
Delta R.T. 0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion: 94 Resp: 829086

Ion	Ratio	Lower	Upper
94	100		
65	32.3	10.5	50.5
66	48.7	25.8	65.8







#11
bis(2-Chloroethyl)ether
Concen: 48.672 ng
RT: 6.616 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 93 Resp: 631516

Ion	Ratio	Lower	Upper
93	100		
63	72.7	53.0	93.0
95	33.0	12.9	52.9







#12
1,3-Dichlorobenzene
Concen: 48.807 ng
RT: 6.822 min Scan# 804
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:146 Resp: 753352

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.3	76.9
75	32.9	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 48.465 ng
RT: 6.898 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:146 Resp: 756875

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.0	76.6
111	44.6	35.7	53.5







#14
1,2-Dichlorobenzene
Concen: 48.049 ng
RT: 7.051 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:146 Resp: 706635

Ion	Ratio	Lower	Upper
146	100		
148	63.7	50.8	76.2
111	47.6	38.0	57.0







#15
Benzyl Alcohol
Concen: 49.113 ng
RT: 7.022 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 652454

Ion	Ratio	Lower	Upper
79	100		
108	73.0	59.7	89.5
77	63.1	51.0	76.6







#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.687 ng
RT: 7.157 min Scan# 861
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion: 45 Resp: 745710

Ion	Ratio	Lower	Upper
45	100		
77	17.9	0.0	37.8
79	14.0	0.0	34.2







#17
2-Methylphenol
Concen: 49.153 ng
RT: 7.128 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:107 Resp: 557981

Ion	Ratio	Lower	Upper
107	100		
108	113.1	90.2	135.2
77	50.3	39.6	59.4
79	49.7	39.1	58.7







#18
Hexachloroethane
Concen: 49.551 ng
RT: 7.392 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:117 Resp: 290127

Ion	Ratio	Lower	Upper
117	100		
119	94.5	76.9	115.3
201	100.9	79.9	119.9







#19
n-Nitroso-di-n-propylamine
Concen: 47.887 ng
RT: 7.298 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 70 Resp: 505166

Ion	Ratio	Lower	Upper
70	100		
42	52.8	41.4	62.0
101	10.2	8.2	12.2
130	22.8	18.2	27.4







#20
3+4-Methylphenols
Concen: 48.388 ng
RT: 7.286 min Scan# 883
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:107 Resp: 704563

Ion	Ratio	Lower	Upper
107	100		
108	88.9	68.4	108.4
77	118.3	69.7	109.7#
79	33.1	11.4	51.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:136 Resp: 877329

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	7.2	5.8	8.8
68	5.1	4.1	6.1







#22
Acetophenone
Concen: 46.924 ng
RT: 7.292 min Scan# 884
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:105 Resp: 987577

Ion	Ratio	Lower	Upper
105	100		
71	6.3	5.8	8.8
51	31.7	23.4	35.0
120	22.0	17.3	25.9







#23
Nitrobenzene-d5
Concen: 96.304 ng
RT: 7.445 min Scan# 91
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 1507471

Ion	Ratio	Lower	Upper
82	100		
128	44.8	36.0	54.0
54	49.8	39.6	59.4







#24
Nitrobenzene
Concen: 48.514 ng
RT: 7.463 min Scan# 913
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion: 77 Resp: 752936

Ion	Ratio	Lower	Upper
77	100		
123	45.6	37.0	55.4
65	14.2	11.3	16.9







#25
Isophorone
Concen: 48.027 ng
RT: 7.704 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 1327254

Ion	Ratio	Lower	Upper
82	100		
95	8.0	6.4	9.6
138	18.1	14.9	22.3







#26
2-Nitrophenol
Concen: 51.644 ng
RT: 7.781 min Scan# 967
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:139 Resp: 392772

Ion	Ratio	Lower	Upper
139	100		
109	33.1	26.5	39.7
65	50.4	40.3	60.5







#27
2,4-Dimethylphenol
Concen: 48.725 ng
RT: 7.810 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:122 Resp: 510395

Ion	Ratio	Lower	Upper
122	100		
107	119.4	94.8	142.2
121	59.1	47.8	71.6







#28
bis(2-Chloroethoxy)methane
Concen: 47.757 ng
RT: 7.910 min Scan# 989
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion: 93 Resp: 826776

Ion	Ratio	Lower	Upper
93	100		
95	33.2	26.6	40.0
123	12.4	10.2	15.4







#29
2,4-Dichlorophenol
Concen: 49.793 ng
RT: 8.016 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 647319

Ion	Ratio	Lower	Upper
162	100		
164	63.8	43.8	83.8
98	38.1	18.1	58.1







#30
1,2,4-Trichlorobenzene
Concen: 48.617 ng
RT: 8.104 min Scan# 1022
Delta R.T. 0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:180 Resp: 697274

Ion	Ratio	Lower	Upper
180	100		
182	97.8	77.5	116.3
145	32.3	25.8	38.6







#31
Naphthalene
Concen: 48.863 ng
RT: 8.186 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:128 Resp: 2196111

Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.6
127	13.5	10.8	16.2







#32
Benzoic acid
Concen: 54.676 ng
RT: 7.939 min Scan# 994
Delta R.T. 0.011 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:122 Resp: 500593

Ion	Ratio	Lower	Upper
122	100		
105	128.0	107.9	147.9
77	99.6	79.0	119.0







#33
4-Chloroaniline
Concen: 48.909 ng
RT: 8.233 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:127 Resp: 775699

Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.2	39.4
65	30.9	25.4	38.2
92	20.2	16.2	24.4







#34
Hexachlorobutadiene
Concen: 50.376 ng
RT: 8.304 min Scan# 1056
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:225 Resp: 469787

Ion	Ratio	Lower	Upper
225	100		
223	62.2	50.6	75.8
227	64.9	52.2	78.2







#35
Caprolactam
Concen: 48.891 ng
RT: 8.616 min Scan# 11
Delta R.T. 0.012 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:113 Resp: 193267

Ion	Ratio	Lower	Upper
113	100		
55	144.7	129.6	169.6
56	117.1	104.0	144.0







#36
4-Chloro-3-methylphenol
Concen: 49.925 ng
RT: 8.710 min Scan# 1125
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:107 Resp: 721800

Ion	Ratio	Lower	Upper
107	100		
144	23.8	19.0	28.6
142	74.4	59.2	88.8







#37
2-Methylnaphthalene
Concen: 47.638 ng
RT: 8.875 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:142 Resp: 1436485

Ion	Ratio	Lower	Upper
142	100		
141	88.7	71.2	106.8







#38
1-Methylnaphthalene
Concen: 46.482 ng
RT: 8.975 min Scan# 1170
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:142 Resp: 1357356

Ion	Ratio	Lower	Upper
142	100		
141	92.4	74.2	111.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:164 Resp: 481971

Ion	Ratio	Lower	Upper
164	100		
162	101.3	81.8	122.6
160	45.2	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 48.955 ng
RT: 9.039 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:216 Resp: 719431

Ion	Ratio	Lower	Upper
216	100		
214	78.9	61.9	92.9
179	22.7	18.1	27.1
108	18.5	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 49.142 ng
RT: 9.027 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:237 Resp: 284968

Ion	Ratio	Lower	Upper
237	100		
235	62.1	43.1	83.1
272	12.4	0.0	32.0







#42
2,4,6-Tribromophenol
Concen: 102.342 ng
RT: 10.710 min Scan# 1465
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:330 Resp: 625035

Ion	Ratio	Lower	Upper
330	100		
332	96.6	77.6	116.4
141	30.1	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 50.291 ng
RT: 9.151 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:196 Resp: 481782

Ion	Ratio	Lower	Upper
196	100		
198	99.9	80.6	121.0
200	32.3	25.8	38.6







#44
2,4,5-Trichlorophenol
Concen: 49.580 ng
RT: 9.186 min Scan# 1206
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:196 Resp: 482570

Ion	Ratio	Lower	Upper
196	100		
198	100.5	80.7	121.1
97	54.9	42.1	63.1
132	30.3	23.4	35.2







#45
2-Fluorobiphenyl
Concen: 96.836 ng
RT: 9.239 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:172 Resp: 3064989

Ion	Ratio	Lower	Upper
172	100		
171	36.5	29.3	43.9
170	24.3	19.4	29.0







#46
1,1'-Biphenyl
Concen: 49.526 ng
RT: 9.339 min Scan# 1232
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:154 Resp: 1809746

Ion	Ratio	Lower	Upper
154	100		
153	41.3	21.1	61.1
76	11.8	0.0	31.3







#47
2-Chloronaphthalene
Concen: 49.960 ng
RT: 9.369 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 1360528

Ion	Ratio	Lower	Upper
162	100		
127	43.9	35.5	53.3
164	33.2	26.6	40.0







#48
2-Nitroaniline
Concen: 54.854 ng
RT: 9.457 min Scan# 1252
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion: 65 Resp: 429111

Ion	Ratio	Lower	Upper
65	100		
92	70.4	56.6	84.8
138	106.4	88.6	133.0







#49
Acenaphthylene
Concen: 49.674 ng
RT: 9.780 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 2005538

Ion	Ratio	Lower	Upper
152	100		
151	20.5	16.3	24.5
153	14.1	11.1	16.7







#50
Dimethylphthalate
Concen: 51.273 ng
RT: 9.645 min Scan# 1284
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:163 Resp: 1716995

Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.1	4.7
164	10.2	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 53.184 ng
RT: 9.704 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 370368

Ion	Ratio	Lower	Upper
165	100		
63	55.0	47.0	70.4
89	56.6	47.4	71.2







#52
Acenaphthene
Concen: 48.333 ng
RT: 9.957 min Scan# 1337
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:154 Resp: 1377901

Ion	Ratio	Lower	Upper
154	100		
153	110.7	90.7	136.1
152	51.0	41.1	61.7







#53
3-Nitroaniline
Concen: 47.638 ng
RT: 9.874 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:138 Resp: 340633

Ion	Ratio	Lower	Upper
138	100		
108	11.6	9.3	13.9
92	116.5	93.5	140.3







#54
2,4-Dinitrophenol
Concen: 49.104 ng
RT: 9.975 min Scan# 1340
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:184 Resp: 184410

Ion	Ratio	Lower	Upper
184	100		
63	66.3	60.2	90.4
154	90.1	132.9	199.3#







#55
Dibenzofuran
Concen: 47.647 ng
RT: 10.127 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:168 Resp: 1936780

Ion	Ratio	Lower	Upper
168	100		
139	41.7	33.9	50.9
169	13.5	11.0	16.4







#56
4-Nitrophenol
Concen: 51.474 ng
RT: 10.022 min Scan# 1348
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:139 Resp: 276487

Ion	Ratio	Lower	Upper
139	100		
109	82.6	61.6	101.6
65	99.1	79.4	119.4







#57
2,4-Dinitrotoluene
Concen: 50.321 ng
RT: 10.104 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 461276

Ion	Ratio	Lower	Upper
165	100		
63	45.8	35.4	53.2
89	74.3	56.6	84.8
182	2.8	2.2	3.4







#58
Fluorene
Concen: 47.708 ng
RT: 10.469 min Scan# 1424
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:166 Resp: 1497123

Ion	Ratio	Lower	Upper
166	100		
165	97.6	79.0	118.4
167	13.7	11.0	16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 49.600 ng
RT: 10.239 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:232 Resp: 438025

Ion	Ratio	Lower	Upper
232	100		
131	46.2	37.3	55.9
130	2.4	1.9	2.9
166	31.0	24.8	37.2







#60
Diethylphthalate
Concen: 48.639 ng
RT: 10.339 min Scan# 1402
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:149 Resp: 1635073

Ion	Ratio	Lower	Upper
149	100		
177	22.5	18.0	27.0
150	13.2	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 48.929 ng
RT: 10.457 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:204 Resp: 799744

Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.8	40.2
141	64.8	52.4	78.6







#62
4-Nitroaniline
Concen: 49.990 ng
RT: 10.486 min Scan# 1427
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:138 Resp: 346378

Ion	Ratio	Lower	Upper
138	100		
92	52.9	32.6	72.6
108	86.4	65.0	105.0







#63
Azobenzene
Concen: 48.653 ng
RT: 10.622 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 1519356

Ion	Ratio	Lower	Upper
77	100		
182	28.3	6.6	46.6
105	22.0	1.3	41.3
51	30.2	10.4	50.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:188 Resp: 811142

Ion	Ratio	Lower	Upper
188	100		
94	8.6	6.8	10.2
80	9.6	7.6	11.4







#65
4,6-Dinitro-2-methylphenol
Concen: 54.673 ng
RT: 10.516 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:198 Resp: 262910

Ion	Ratio	Lower	Upper
198	100		
51	41.7	22.3	62.3
105	42.0	23.0	63.0







#66
n-Nitrosodiphenylamine
Concen: 49.373 ng
RT: 10.580 min Scan# 1443
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:169 Resp: 1293745

Ion	Ratio	Lower	Upper
169	100		
168	67.2	53.7	80.5
167	36.6	29.2	43.8







#67
4-Bromophenyl-phenylether
Concen: 52.740 ng
RT: 10.951 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:248 Resp: 532762

Ion	Ratio	Lower	Upper
248	100		
250	97.5	77.8	116.8
141	72.5	62.6	93.8







#68
Hexachlorobenzene
Concen: 51.359 ng
RT: 11.016 min Scan# 1517
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:284 Resp: 573467

Ion	Ratio	Lower	Upper
284	100		
142	30.9	27.0	40.6
249	30.8	25.2	37.8







#69
Atrazine
Concen: 65.441 ng
RT: 11.104 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:200 Resp: 418147

Ion	Ratio	Lower	Upper
200	100		
173	25.0	4.7	44.7
215	46.3	25.4	65.4







#70
Pentachlorophenol
Concen: 53.050 ng
RT: 11.204 min Scan# 1549
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:266 Resp: 365292

Ion	Ratio	Lower	Upper
266	100		
268	62.9	51.0	76.6
264	63.3	50.2	75.2







#71
Phenanthrene
Concen: 47.434 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:178 Resp: 2080633

Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.8	23.8
179	15.9	12.6	19.0







#72
Anthracene
Concen: 47.299 ng
RT: 11.480 min Scan# 1596
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:178 Resp: 2081490

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.1	22.7
179	16.1	12.6	19.0







#73
Carbazole
Concen: 48.338 ng
RT: 11.633 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:167 Resp: 1831424

Ion	Ratio	Lower	Upper
167	100		
166	22.0	17.4	26.2
139	14.1	11.0	16.6







#74
Di-n-butylphthalate
Concen: 50.185 ng
RT: 11.963 min Scan# 1678
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:149 Resp: 2510447

Ion	Ratio	Lower	Upper
149	100		
150	9.8	7.6	11.4
104	5.6	4.4	6.6







#75
Fluoranthene
Concen: 50.564 ng
RT: 12.616 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:202 Resp: 2328357

Ion	Ratio	Lower	Upper
202	100		
101	10.1	0.0	29.7
203	18.1	0.0	37.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.045 min Scan# 2032
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:240 Resp: 491943

Ion	Ratio	Lower	Upper
240	100		
120	9.5	8.4	12.6
236	25.8	20.5	30.7







#77
Benzidine
Concen: 38.966 ng
RT: 12.733 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:184 Resp: 270893

Ion	Ratio	Lower	Upper
184	100		
185	15.1	11.5	17.3
183	12.0	9.7	14.5







#78
Pyrene
Concen: 54.723 ng
RT: 12.845 min Scan# 1828
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:202 Resp: 2308671

Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.3	25.9
203	18.2	14.4	21.6







#79
Terphenyl-d14
Concen: 102.007 ng
RT: 12.992 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:244 Resp: 3394803

Ion	Ratio	Lower	Upper
244	100		
212	7.3	6.0	9.0
122	9.2	7.7	11.5







#80
Butylbenzylphthalate
Concen: 52.022 ng
RT: 13.463 min Scan# 1933
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:149 Resp: 864849

Ion	Ratio	Lower	Upper
149	100		
91	71.1	56.5	84.7
206	21.7	17.3	25.9







#81
Benzo(a)anthracene
Concen: 49.192 ng
RT: 14.033 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 1592035

Ion	Ratio	Lower	Upper
228	100		
226	27.6	21.8	32.8
229	20.1	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 48.044 ng
RT: 13.992 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:252 Resp: 449408

Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.0	76.6
126	10.8	8.4	12.6







#83
Chrysene
Concen: 50.404 ng
RT: 14.068 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 1465389

Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.1	36.1
229	19.9	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 52.747 ng
RT: 14.021 min Scan# 2028
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:149 Resp: 1203021

Ion	Ratio	Lower	Upper
149	100		
167	27.1	21.4	32.0
279	4.9	3.8	5.6







#85
Di-n-octyl phthalate
Concen: 53.307 ng
RT: 14.633 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:149 Resp: 1692206

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.7	6.9	10.3







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.509 min Scan# 2281
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:264 Resp: 413564

Ion	Ratio	Lower	Upper
264	100		
260	23.1	19.1	28.7
265	21.4	17.5	26.3







#87
Indeno(1,2,3-cd)pyrene
Concen: 50.969 ng
RT: 17.003 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:276 Resp: 1366974

Ion	Ratio	Lower	Upper
276	100		
138	21.9	18.4	27.6
277	25.3	20.3	30.5







#88
Benzo(b)fluoranthene
Concen: 49.625 ng
RT: 15.080 min Scan# 2208
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:252 Resp: 1392083

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.5	26.3
125	8.9	7.0	10.4







#89
Benzo(k)fluoranthene
Concen: 47.638 ng
RT: 15.109 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:252 Resp: 1167351

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	9.0	7.2	10.8







#90
Benzo(a)pyrene
Concen: 50.591 ng
RT: 15.451 min Scan# 2271
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:252 Resp: 1119256

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	9.6	7.8	11.8







#91
Dibenzo(a,h)anthracene
Concen: 51.149 ng
RT: 17.021 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 1131875

Ion	Ratio	Lower	Upper
278	100		
139	14.4	11.4	17.2
279	23.9	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 50.106 ng
RT: 17.456 min Scan# 2612
Delta R.T. 0.006 min
Lab File: BF141905.D
Acq: 10 Mar 2025 15:20

Tgt Ion:276 Resp: 1099760

Ion	Ratio	Lower	Upper
276	100		
277	24.1	18.7	28.1
138	19.2	15.8	23.8



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	209747	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	840616	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	483165	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	813672	20.000	ng	0.00
76) Chrysene-d12	14.039	240	523890	20.000	ng	0.00
86) Perylene-d12	15.510	264	426422	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	965473	76.823	ng	0.00
7) Phenol-d6	6.504	99	1231487	76.962	ng	0.00
23) Nitrobenzene-d5	7.445	82	1179480	78.962	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	486334	79.342	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	2399526	75.528	ng	0.00
79) Terphenyl-d14	12.986	244	2825870	79.748	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.669	88	203628	38.670	ng	99
3) Pyridine	3.428	79	504398	39.050	ng	99
4) n-Nitrosodimethylamine	3.387	42	239879	38.959	ng	97
6) Aniline	6.545	93	614962	38.958	ng	98
8) 2-Chlorophenol	6.669	128	539701	38.721	ng	99
9) Benzaldehyde	6.434	77	339615	37.950	ng	99
10) Phenol	6.516	94	658449	39.115	ng	97
11) bis(2-Chloroethyl)ether	6.616	93	483438	38.246	ng	99
12) 1,3-Dichlorobenzene	6.822	146	577700	38.391	ng	100
13) 1,4-Dichlorobenzene	6.898	146	584132	38.366	ng	99
14) 1,2-Dichlorobenzene	7.051	146	555621	38.744	ng	100
15) Benzyl Alcohol	7.022	79	509792	39.408	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.157	45	569783	37.338	ng	100
17) 2-Methylphenol	7.128	107	429715	38.775	ng	99
18) Hexachloroethane	7.392	117	221507	38.797	ng	100
19) n-Nitroso-di-n-propyla...	7.298	70	391096	38.038	ng	100
20) 3+4-Methylphenols	7.281	107	548208	38.619	ng	99
22) Acetophenone	7.292	105	757917	37.650	ng	98
24) Nitrobenzene	7.463	77	578836	38.980	ng	99
25) Isophorone	7.704	82	1007659	38.163	ng	100
26) 2-Nitrophenol	7.775	139	294657	40.429	ng	96
27) 2,4-Dimethylphenol	7.810	122	385504	38.414	ng	99
28) bis(2-Chloroethoxy)met...	7.910	93	628379	37.943	ng	100
29) 2,4-Dichlorophenol	8.016	162	485205	38.952	ng	99
30) 1,2,4-Trichlorobenzene	8.104	180	526250	38.336	ng	100
31) Naphthalene	8.187	128	1633401	37.827	ng	100
32) Benzoic acid	7.928	122	373179	42.303	ng	99
33) 4-Chloroaniline	8.228	127	592172	38.847	ng	99
34) Hexachlorobutadiene	8.304	225	347789	38.849	ng	99
35) Caprolactam	8.604	113	141562	37.276	ng	98
36) 4-Chloro-3-methylphenol	8.704	107	535661	38.550	ng	99
37) 2-Methylnaphthalene	8.875	142	1094895	37.935	ng	100
38) 1-Methylnaphthalene	8.975	142	1071850	38.484	ng	100
40) 1,2,4,5-Tetrachloroben...	9.039	216	561804	38.191	ng	99
41) Hexachlorocyclopentadiene	9.028	237	220386	38.281	ng	98
43) 2,4,6-Trichlorophenol	9.151	196	377743	39.292	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.186	196	372564	38.263	ng	99
46) 1,1'-Biphenyl	9.339	154	1401736	38.182	ng	100
47) 2-Chloronaphthalene	9.363	162	1042241	38.090	ng	99
48) 2-Nitroaniline	9.457	65	308259	38.908	ng	99
49) Acenaphthylene	9.781	152	1531852	37.725	ng	100
50) Dimethylphthalate	9.639	163	1272572	37.611	ng	100
51) 2,6-Dinitrotoluene	9.698	165	274531	38.970	ng	99
52) Acenaphthene	9.951	154	1095922	38.405	ng	100
53) 3-Nitroaniline	9.869	138	275555	38.561	ng	99
54) 2,4-Dinitrophenol	9.969	184	142831	39.326	ng	95
55) Dibenzofuran	10.122	168	1540381	37.778	ng	100
56) 4-Nitrophenol	10.016	139	212462	39.425	ng	99
57) 2,4-Dinitrotoluene	10.104	165	361518	39.291	ng	99
58) Fluorene	10.469	166	1175392	37.381	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.239	232	346168	39.070	ng	99
60) Diethylphthalate	10.339	149	1293401	38.337	ng	100
61) 4-Chlorophenyl-phenyle...	10.457	204	627404	38.272	ng	99
62) 4-Nitroaniline	10.480	138	270993	38.953	ng	99
63) Azobenzene	10.622	77	1189121	37.911	ng	98
65) 4,6-Dinitro-2-methylph...	10.510	198	200888	41.415	ng	99
66) n-Nitrosodiphenylamine	10.575	169	1008029	38.283	ng	99
67) 4-Bromophenyl-phenylether	10.951	248	398683	39.015	ng	98
68) Hexachlorobenzene	11.016	284	445938	39.789	ng	96
69) Atrazine	11.098	200	257856	31.945	ng	99
70) Pentachlorophenol	11.204	266	289713	41.955	ng	100
71) Phenanthrene	11.427	178	1685770	38.357	ng	100
72) Anthracene	11.480	178	1675103	37.991	ng	100
73) Carbazole	11.633	167	1451333	38.167	ng	99
74) Di-n-butylphthalate	11.963	149	1953516	38.717	ng	100
75) Fluoranthene	12.616	202	1750317	37.545	ng	100
77) Benzidine	12.733	184	420377	57.513	ng	100
78) Pyrene	12.845	202	1746456	38.539	ng	100
80) Butylbenzylphthalate	13.463	149	723288	40.778	ng	99
81) Benzo(a)anthracene	14.027	228	1347977	39.211	ng	100
82) 3,3'-Dichlorobenzidine	13.992	252	380866	38.337	ng	100
83) Chrysene	14.068	228	1159074	37.195	ng	100
84) Bis(2-ethylhexyl)phtha...	14.021	149	1006170	41.142	ng	99
85) Di-n-octyl phthalate	14.633	149	1394136	40.971	ng	100
87) Indeno(1,2,3-cd)pyrene	17.004	276	1061096	38.467	ng	99
88) Benzo(b)fluoranthene	15.080	252	1166263	39.906	ng	99
89) Benzo(k)fluoranthene	15.110	252	869471	34.765	ng	100
90) Benzo(a)pyrene	15.451	252	882967	38.612	ng	99
91) Dibenzo(a,h)anthracene	17.021	278	886038	38.931	ng	99
92) Benzo(g,h,i)perylene	17.451	276	855812	38.051	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:152 Resp: 209747

Ion	Ratio	Lower	Upper
152	100		
150	159.4	127.4	191.2
115	65.1	51.9	77.9







#2
1,4-Dioxane
Concen: 38.670 ng
RT: 2.669 min Scan# 98
Delta R.T. -0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion: 88 Resp: 203628

Ion	Ratio	Lower	Upper
88	100		
58	71.2	57.4	86.2
43	27.6	21.5	32.3







#3
Pyridine
Concen: 39.050 ng
RT: 3.428 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 79 Resp: 504398

Ion	Ratio	Lower	Upper
79	100		
52	67.8	53.3	79.9
51	34.5	27.6	41.4







#4
n-Nitrosodimethylamine
Concen: 38.959 ng
RT: 3.387 min Scan# 220
Delta R.T. -0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion: 42 Resp: 239879

Ion	Ratio	Lower	Upper
42	100		
74	133.0	103.4	155.0
44	6.0	5.0	7.4







#5
2-Fluorophenol
Concen: 76.823 ng
RT: 5.493 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:112 Resp: 965473

Ion	Ratio	Lower	Upper
112	100		
64	58.1	46.5	69.7
63	31.7	25.4	38.2







#6
Aniline
Concen: 38.958 ng
RT: 6.545 min Scan# 757
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion: 93 Resp: 614962

Ion	Ratio	Lower	Upper
93	100		
66	41.1	31.9	47.9
65	20.0	15.8	23.6







#7
Phenol-d6
Concen: 76.962 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 99 Resp: 1231487

Ion	Ratio	Lower	Upper
99	100		
42	18.0	14.6	21.8
71	38.3	30.8	46.2







#8
2-Chlorophenol
Concen: 38.721 ng
RT: 6.669 min Scan# 778
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:128 Resp: 539701

Ion	Ratio	Lower	Upper
128	100		
130	33.0	12.8	52.8
64	43.6	24.4	64.4







#9
Benzaldehyde
Concen: 37.950 ng
RT: 6.434 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 77 Resp: 339615

Ion	Ratio	Lower	Upper
77	100		
105	99.2	80.0	120.0
106	95.7	74.5	114.5







#10
Phenol
Concen: 39.115 ng
RT: 6.516 min Scan# 752
Delta R.T. -0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion: 94 Resp: 658449

Ion	Ratio	Lower	Upper
94	100		
65	31.6	10.5	50.5
66	47.8	25.8	65.8







#11
bis(2-Chloroethyl)ether
Concen: 38.246 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 93 Resp: 483438

Ion	Ratio	Lower	Upper
93	100		
63	73.5	53.0	93.0
95	32.6	12.9	52.9







#12
1,3-Dichlorobenzene
Concen: 38.391 ng
RT: 6.822 min Scan# 804
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:146 Resp: 577700

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.3	76.9
75	33.3	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 38.366 ng
RT: 6.898 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:146 Resp: 584132

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.0	76.6
111	43.8	35.7	53.5







#14
1,2-Dichlorobenzene
Concen: 38.744 ng
RT: 7.051 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:146 Resp: 555621

Ion	Ratio	Lower	Upper
146	100		
148	63.3	50.8	76.2
111	47.2	38.0	57.0







#15
Benzyl Alcohol
Concen: 39.408 ng
RT: 7.022 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 79 Resp: 509792

Ion	Ratio	Lower	Upper
79	100		
108	72.4	59.7	89.5
77	63.0	51.0	76.6







#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.338 ng
RT: 7.157 min Scan# 861
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion: 45 Resp: 569783

Ion	Ratio	Lower	Upper
45	100		
77	17.8	0.0	37.8
79	13.8	0.0	34.2







#17
2-Methylphenol
Concen: 38.775 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:107 Resp: 429715

Ion	Ratio	Lower	Upper
107	100		
108	111.0	90.2	135.2
77	48.9	39.6	59.4
79	49.0	39.1	58.7







#18
Hexachloroethane
Concen: 38.797 ng
RT: 7.392 min Scan# 901
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:117 Resp: 221507

Ion	Ratio	Lower	Upper
117	100		
119	95.5	76.9	115.3
201	99.7	79.9	119.9







#19
n-Nitroso-di-n-propylamine
Concen: 38.038 ng
RT: 7.298 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 70 Resp: 391096

Ion	Ratio	Lower	Upper
70	100		
42	51.8	41.4	62.0
101	10.5	8.2	12.2
130	22.7	18.2	27.4







#20
3+4-Methylphenols
Concen: 38.619 ng
RT: 7.281 min Scan# 882
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:107 Resp: 548208

Ion	Ratio	Lower	Upper
107	100		
108	88.3	68.4	108.4
77	91.7	69.7	109.7
79	31.7	11.4	51.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:136 Resp: 840616

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.8	13.2
54	7.1	5.8	8.8
68	4.9	4.1	6.1







#22
Acetophenone
Concen: 37.650 ng
RT: 7.292 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:105 Resp: 757917

Ion	Ratio	Lower	Upper
105	100		
71	7.4	5.8	8.8
51	30.2	23.4	35.0
120	22.4	17.3	25.9







#23
Nitrobenzene-d5
Concen: 78.962 ng
RT: 7.445 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 82 Resp: 1179480

Ion	Ratio	Lower	Upper
82	100		
128	44.1	36.0	54.0
54	49.3	39.6	59.4







#24
Nitrobenzene
Concen: 38.980 ng
RT: 7.463 min Scan# 913
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion: 77 Resp: 578836

Ion	Ratio	Lower	Upper
77	100		
123	45.6	37.0	55.4
65	14.2	11.3	16.9







#25
Isophorone
Concen: 38.163 ng
RT: 7.704 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 82 Resp: 1007659

Ion	Ratio	Lower	Upper
82	100		
95	8.1	6.4	9.6
138	18.6	14.9	22.3







#26
2-Nitrophenol
Concen: 40.429 ng
RT: 7.775 min Scan# 966
Delta R.T. -0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:139 Resp: 294657

Ion	Ratio	Lower	Upper
139	100		
109	34.3	26.5	39.7
65	54.1	40.3	60.5







#27
2,4-Dimethylphenol
Concen: 38.414 ng
RT: 7.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:122 Resp: 385504

Ion	Ratio	Lower	Upper
122	100		
107	120.5	94.8	142.2
121	60.1	47.8	71.6







#28
bis(2-Chloroethoxy)methane
Concen: 37.943 ng
RT: 7.910 min Scan# 989
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion: 93 Resp: 628379

Ion	Ratio	Lower	Upper
93	100		
95	33.3	26.6	40.0
123	12.9	10.2	15.4







#29
2,4-Dichlorophenol
Concen: 38.952 ng
RT: 8.016 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:162 Resp: 485205

Ion	Ratio	Lower	Upper
162	100		
164	63.3	43.8	83.8
98	37.6	18.1	58.1







#30
1,2,4-Trichlorobenzene
Concen: 38.336 ng
RT: 8.104 min Scan# 1022
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:180 Resp: 526250

Ion	Ratio	Lower	Upper
180	100		
182	96.7	77.5	116.3
145	32.5	25.8	38.6







#31
Naphthalene
Concen: 37.827 ng
RT: 8.187 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:128 Resp: 1633401

Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.6
127	13.7	10.8	16.2







#32
Benzoic acid
Concen: 42.303 ng
RT: 7.928 min Scan# 992
Delta R.T. -0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:122 Resp: 373179

Ion	Ratio	Lower	Upper
122	100		
105	127.1	107.9	147.9
77	97.0	79.0	119.0







#33
4-Chloroaniline
Concen: 38.847 ng
RT: 8.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:127 Resp: 592172

Ion	Ratio	Lower	Upper
127	100		
129	32.2	26.2	39.4
65	31.9	25.4	38.2
92	20.7	16.2	24.4







#34
Hexachlorobutadiene
Concen: 38.849 ng
RT: 8.304 min Scan# 1056
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:225 Resp: 347789

Ion	Ratio	Lower	Upper
225	100		
223	63.2	50.6	75.8
227	63.8	52.2	78.2







#35
Caprolactam
Concen: 37.276 ng
RT: 8.604 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:113 Resp: 141562

Ion	Ratio	Lower	Upper
113	100		
55	148.0	129.6	169.6
56	121.8	104.0	144.0







#36
4-Chloro-3-methylphenol
Concen: 38.550 ng
RT: 8.704 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:107 Resp: 535661

Ion	Ratio	Lower	Upper
107	100		
144	23.9	19.0	28.6
142	73.3	59.2	88.8







#37
2-Methylnaphthalene
Concen: 37.935 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:142 Resp: 1094895

Ion	Ratio	Lower	Upper
142	100		
141	89.3	71.2	106.8







#38
1-Methylnaphthalene
Concen: 38.484 ng
RT: 8.975 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:142 Resp: 1071850

Ion	Ratio	Lower	Upper
142	100		
141	92.3	74.2	111.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:164 Resp: 483165

Ion	Ratio	Lower	Upper
164	100		
162	102.2	81.8	122.6
160	44.8	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.191 ng
RT: 9.039 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:216 Resp: 561804

Ion	Ratio	Lower	Upper
216	100		
214	79.2	61.9	92.9
179	22.7	18.1	27.1
108	18.5	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 38.281 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:237 Resp: 220386

Ion	Ratio	Lower	Upper
237	100		
235	61.7	43.1	83.1
272	12.3	0.0	32.0







#42
2,4,6-Tribromophenol
Concen: 79.342 ng
RT: 10.710 min Scan# 1465
Delta R.T. 0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:330 Resp: 486334

Ion	Ratio	Lower	Upper
330	100		
332	96.2	77.6	116.4
141	30.9	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 39.292 ng
RT: 9.151 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:196 Resp: 377743

Ion	Ratio	Lower	Upper
196	100		
198	100.6	80.6	121.0
200	32.8	25.8	38.6







#44
2,4,5-Trichlorophenol
Concen: 38.263 ng
RT: 9.186 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:196 Resp: 372564

Ion	Ratio	Lower	Upper
196	100		
198	100.3	80.7	121.1
97	53.5	42.1	63.1
132	29.9	23.4	35.2







#45
2-Fluorobiphenyl
Concen: 75.528 ng
RT: 9.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:172 Resp: 2399526

Ion	Ratio	Lower	Upper
172	100		
171	36.4	29.3	43.9
170	24.1	19.4	29.0







#46
1,1'-Biphenyl
Concen: 38.182 ng
RT: 9.339 min Scan# 1232
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:154 Resp: 1401736

Ion	Ratio	Lower	Upper
154	100		
153	41.3	21.1	61.1
76	11.1	0.0	31.3







#47
2-Chloronaphthalene
Concen: 38.090 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:162 Resp: 1042241

Ion	Ratio	Lower	Upper
162	100		
127	44.1	35.5	53.3
164	32.8	26.6	40.0







#48
2-Nitroaniline
Concen: 38.908 ng
RT: 9.457 min Scan# 1252
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion: 65 Resp: 308259

Ion	Ratio	Lower	Upper
65	100		
92	71.2	56.6	84.8
138	110.1	88.6	133.0







#49
Acenaphthylene
Concen: 37.725 ng
RT: 9.781 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:152 Resp: 1531852

Ion	Ratio	Lower	Upper
152	100		
151	20.7	16.3	24.5
153	13.9	11.1	16.7







#50
Dimethylphthalate
Concen: 37.611 ng
RT: 9.639 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:163 Resp: 1272572

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.1	4.7
164	10.4	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 38.970 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:165 Resp: 274531

Ion	Ratio	Lower	Upper
165	100		
63	59.5	47.0	70.4
89	59.1	47.4	71.2







#52
Acenaphthene
Concen: 38.405 ng
RT: 9.951 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:154 Resp: 1095922

Ion	Ratio	Lower	Upper
154	100		
153	113.2	90.7	136.1
152	51.8	41.1	61.7







#53
3-Nitroaniline
Concen: 38.561 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:138 Resp: 275555

Ion	Ratio	Lower	Upper
138	100		
108	11.3	9.3	13.9
92	117.7	93.5	140.3







#54
2,4-Dinitrophenol
Concen: 39.326 ng
RT: 9.969 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:184 Resp: 142831

Ion	Ratio	Lower	Upper
184	100		
63	73.1	60.2	90.4
154	173.5	132.9	199.3







#55
Dibenzofuran
Concen: 37.778 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:168 Resp: 1540381

Ion	Ratio	Lower	Upper
168	100		
139	42.2	33.9	50.9
169	13.8	11.0	16.4







#56
4-Nitrophenol
Concen: 39.425 ng
RT: 10.016 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:139 Resp: 212462

Ion	Ratio	Lower	Upper
139	100		
109	83.2	61.6	101.6
65	100.4	79.4	119.4







#57
2,4-Dinitrotoluene
Concen: 39.291 ng
RT: 10.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:165 Resp: 361518

Ion	Ratio	Lower	Upper
165	100		
63	44.7	35.4	53.2
89	72.2	56.6	84.8
182	2.7	2.2	3.4







#58
Fluorene
Concen: 37.381 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:166 Resp: 1175392

Ion	Ratio	Lower	Upper
166	100		
165	98.8	79.0	118.4
167	13.8	11.0	16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 39.070 ng
RT: 10.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:232 Resp: 346168

Ion	Ratio	Lower	Upper
232	100		
131	45.8	37.3	55.9
130	2.4	1.9	2.9
166	30.7	24.8	37.2







#60
Diethylphthalate
Concen: 38.337 ng
RT: 10.339 min Scan# 1402
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:149 Resp: 1293401

Ion	Ratio	Lower	Upper
149	100		
177	22.5	18.0	27.0
150	12.8	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 38.272 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:204 Resp: 627404

Ion	Ratio	Lower	Upper
204	100		
206	33.0	26.8	40.2
141	64.8	52.4	78.6







#62
4-Nitroaniline
Concen: 38.953 ng
RT: 10.480 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:138 Resp: 270993

Ion	Ratio	Lower	Upper
138	100		
92	52.3	32.6	72.6
108	86.3	65.0	105.0







#63
Azobenzene
Concen: 37.911 ng
RT: 10.622 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion: 77 Resp: 1189121

Ion	Ratio	Lower	Upper
77	100		
182	28.5	6.6	46.6
105	22.0	1.3	41.3
51	30.2	10.4	50.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:188 Resp: 813672

Ion	Ratio	Lower	Upper
188	100		
94	8.5	6.8	10.2
80	9.1	7.6	11.4







#65
4,6-Dinitro-2-methylphenol
Concen: 41.415 ng
RT: 10.510 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:198 Resp: 200888

Ion	Ratio	Lower	Upper
198	100		
51	43.7	22.3	62.3
105	43.4	23.0	63.0







#66
n-Nitrosodiphenylamine
Concen: 38.283 ng
RT: 10.575 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:169 Resp: 1008029

Ion	Ratio	Lower	Upper
169	100		
168	66.7	53.7	80.5
167	36.9	29.2	43.8







#67
4-Bromophenyl-phenylether
Concen: 39.015 ng
RT: 10.951 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:248 Resp: 398683

Ion	Ratio	Lower	Upper
248	100		
250	98.1	77.8	116.8
141	74.4	62.6	93.8







#68
Hexachlorobenzene
Concen: 39.789 ng
RT: 11.016 min Scan# 1517
Delta R.T. 0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:284 Resp: 445938

Ion	Ratio	Lower	Upper
284	100		
142	30.0	27.0	40.6
249	30.5	25.2	37.8







#69
Atrazine
Concen: 31.945 ng
RT: 11.098 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:200 Resp: 257856

Ion	Ratio	Lower	Upper
200	100		
173	25.1	4.7	44.7
215	45.7	25.4	65.4







#70
Pentachlorophenol
Concen: 41.955 ng
RT: 11.204 min Scan# 1549
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:266 Resp: 289713

Ion	Ratio	Lower	Upper
266	100		
268	63.8	51.0	76.6
264	62.9	50.2	75.2







#71
Phenanthrene
Concen: 38.357 ng
RT: 11.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:178 Resp: 1685770

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.8	23.8
179	15.9	12.6	19.0







#72
Anthracene
Concen: 37.991 ng
RT: 11.480 min Scan# 1596
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:178 Resp: 1675103

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.1	22.7
179	15.7	12.6	19.0







#73
Carbazole
Concen: 38.167 ng
RT: 11.633 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:167 Resp: 1451333

Ion	Ratio	Lower	Upper
167	100		
166	22.0	17.4	26.2
139	14.2	11.0	16.6







#74
Di-n-butylphthalate
Concen: 38.717 ng
RT: 11.963 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:149 Resp: 1953516

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.6	11.4
104	5.5	4.4	6.6







#75
Fluoranthene
Concen: 37.545 ng
RT: 12.616 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:202 Resp: 1750317

Ion	Ratio	Lower	Upper
202	100		
101	9.6	0.0	29.7
203	17.9	0.0	37.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.039 min Scan# 2031
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:240 Resp: 523890

Ion	Ratio	Lower	Upper
240	100		
120	9.8	8.4	12.6
236	25.3	20.5	30.7







#77
Benzidine
Concen: 57.513 ng
RT: 12.733 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:184 Resp: 420377

Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.5	17.3
183	12.0	9.7	14.5







#78
Pyrene
Concen: 38.539 ng
RT: 12.845 min Scan# 1828
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:202 Resp: 1746456

Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.3	25.9
203	17.7	14.4	21.6







#79
Terphenyl-d14
Concen: 79.748 ng
RT: 12.986 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:244 Resp: 2825870

Ion	Ratio	Lower	Upper
244	100		
212	7.4	6.0	9.0
122	9.7	7.7	11.5







#80
Butylbenzylphthalate
Concen: 40.778 ng
RT: 13.463 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:149 Resp: 723288

Ion	Ratio	Lower	Upper
149	100		
91	69.3	56.5	84.7
206	21.8	17.3	25.9







#81
Benzo(a)anthracene
Concen: 39.211 ng
RT: 14.027 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:228 Resp: 1347977

Ion	Ratio	Lower	Upper
228	100		
226	26.9	21.8	32.8
229	20.0	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 38.337 ng
RT: 13.992 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:252 Resp: 380866

Ion	Ratio	Lower	Upper
252	100		
254	63.8	51.0	76.6
126	9.9	8.4	12.6







#83
Chrysene
Concen: 37.195 ng
RT: 14.068 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:228 Resp: 1159074

Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.1	36.1
229	20.0	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 41.142 ng
RT: 14.021 min Scan# 2028
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:149 Resp: 1006170

Ion	Ratio	Lower	Upper
149	100		
167	27.2	21.4	32.0
279	4.8	3.8	5.6







#85
Di-n-octyl phthalate
Concen: 40.971 ng
RT: 14.633 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:149 Resp: 1394136

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.5	6.9	10.3







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.510 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:264 Resp: 426422

Ion	Ratio	Lower	Upper
264	100		
260	23.8	19.1	28.7
265	21.8	17.5	26.3







#87
Indeno(1,2,3-cd)pyrene
Concen: 38.467 ng
RT: 17.004 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:276 Resp: 1061096

Ion	Ratio	Lower	Upper
276	100		
138	22.2	18.4	27.6
277	25.3	20.3	30.5







#88
Benzo(b)fluoranthene
Concen: 39.906 ng
RT: 15.080 min Scan# 2208
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:252 Resp: 1166263

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.5	26.3
125	8.7	7.0	10.4







#89
Benzo(k)fluoranthene
Concen: 34.765 ng
RT: 15.110 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:252 Resp: 869471

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	8.9	7.2	10.8







#90
Benzo(a)pyrene
Concen: 38.612 ng
RT: 15.451 min Scan# 2271
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:252 Resp: 882967

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.4	26.0
125	9.6	7.8	11.8







#91
Dibenzo(a,h)anthracene
Concen: 38.931 ng
RT: 17.021 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Tgt Ion:278 Resp: 886038

Ion	Ratio	Lower	Upper
278	100		
139	13.7	11.4	17.2
279	23.9	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 38.051 ng
RT: 17.451 min Scan# 2611
Delta R.T. 0.000 min
Lab File: BF141906.D
Acq: 10 Mar 2025 15:53

Tgt Ion:276 Resp: 855812

Ion	Ratio	Lower	Upper
276	100		
277	23.3	18.7	28.1
138	18.7	15.8	23.8



Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2		1,4-Dioxane	0.502	0.485	3.4	91	0.00
3		Pyridine	1.232	1.202	2.4	93	-0.01
4		n-Nitrosodimethylamine	0.587	0.572	2.6	92	0.00
5	S	2-Fluorophenol	1.198	1.151	3.9	94	0.00
6		Aniline	1.505	1.466	2.6	94	0.00
7	S	Phenol-d6	1.526	1.468	3.8	96	0.00
8		2-Chlorophenol	1.329	1.287	3.2	96	0.00
9		Benzaldehyde	0.853	0.810	5.0	98	0.00
10	C	Phenol	1.605	1.570	2.2	96	0.00
11		bis(2-Chloroethyl)ether	1.205	1.152	4.4	95	0.00
12		1,3-Dichlorobenzene	1.435	1.377	4.0	94	0.00
13	C	1,4-Dichlorobenzene	1.452	1.392	4.1	94	0.00
14		1,2-Dichlorobenzene	1.367	1.325	3.1	96	0.00
15		Benzyl Alcohol	1.233	1.215	1.5	96	0.00
16		2,2'-oxybis(1-Chloropropane	1.455	1.358	6.7	94	0.00
17		2-Methylphenol	1.057	1.024	3.1	96	0.00
18		Hexachloroethane	0.544	0.528	2.9	96	0.00
19	P	n-Nitroso-di-n-propylamine	0.980	0.932	4.9	96	0.00
20		3+4-Methylphenols	1.354	1.307	3.5	95	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
22		Acetophenone	0.479	0.451	5.8	94	0.00
23	S	Nitrobenzene-d5	0.355	0.351	1.1	97	0.00
24		Nitrobenzene	0.353	0.344	2.5	96	0.00
25		Isophorone	0.628	0.599	4.6	96	0.00
26	C	2-Nitrophenol	0.173	0.175	-1.2	96	0.00
27		2,4-Dimethylphenol	0.239	0.229	4.2	96	0.00
28		bis(2-Chloroethoxy)methane	0.394	0.374	5.1	97	0.00
29	C	2,4-Dichlorophenol	0.296	0.289	2.4	97	0.00
30		1,2,4-Trichlorobenzene	0.327	0.313	4.3	97	0.00
31		Naphthalene	1.027	0.972	5.4	95	0.00
32		Benzoic acid	0.210	0.222	-5.7	103	0.00
33		4-Chloroaniline	0.363	0.352	3.0	97	0.00
34	C	Hexachlorobutadiene	0.213	0.207	2.8	98	0.00
35		Caprolactam	0.090	0.084	6.7	96	0.00
36	C	4-Chloro-3-methylphenol	0.331	0.319	3.6	96	0.00
37		2-Methylnaphthalene	0.687	0.651	5.2	96	0.00
38		1-Methylnaphthalene	0.663	0.638	3.8	98	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
40		1,2,4,5-Tetrachlorobenzene	0.609	0.581	4.6	96	0.00
41	P	Hexachlorocyclopentadiene	0.238	0.228	4.2	89	0.00
42	S	2,4,6-Tribromophenol	0.254	0.252	0.8	98	0.00
43	C	2,4,6-Trichlorophenol	0.398	0.391	1.8	96	0.00
44		2,4,5-Trichlorophenol	0.403	0.386	4.2	97	0.00
45	S	2-Fluorobiphenyl	1.315	1.242	5.6	97	0.00
46		1,1'-Biphenyl	1.520	1.451	4.5	98	0.00
47		2-Chloronaphthalene	1.133	1.079	4.8	97	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.328	0.319	2.7	97	0.00
49		Acenaphthylene	1.681	1.585	5.7	95	0.00
50		Dimethylphthalate	1.401	1.317	6.0	97	0.00
51		2,6-Dinitrotoluene	0.292	0.284	2.7	97	0.00
52	C	Acenaphthene	1.181	1.134	4.0	97	0.00
53		3-Nitroaniline	0.296	0.285	3.7	95	0.00
54	P	2,4-Dinitrophenol	0.139	0.148	-6.5	102	0.00
55		Dibenzofuran	1.688	1.594	5.6	97	0.00
56	P	4-Nitrophenol	0.223	0.220	1.3	94	0.00
57		2,4-Dinitrotoluene	0.381	0.374	1.8	98	0.00
58		Fluorene	1.302	1.216	6.6	97	0.00
59		2,3,4,6-Tetrachlorophenol	0.367	0.358	2.5	98	0.00
60		Diethylphthalate	1.397	1.338	4.2	98	0.00
61		4-Chlorophenyl-phenylether	0.679	0.649	4.4	99	0.00
62		4-Nitroaniline	0.288	0.280	2.8	95	0.00
63		Azobenzene	1.298	1.231	5.2	97	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
65		4,6-Dinitro-2-methylphenol	0.119	0.123	-3.4	99	0.00
66	c	n-Nitrosodiphenylamine	0.647	0.619	4.3	97	0.00
67		4-Bromophenyl-phenylether	0.251	0.245	2.4	97	0.00
68		Hexachlorobenzene	0.275	0.274	0.4	99	0.00
69		Atrazine	0.198	0.158	20.2	94	0.00
70	C	Pentachlorophenol	0.170	0.178	-4.7	97	0.00
71		Phenanthrene	1.080	1.036	4.1	97	0.00
72		Anthracene	1.084	1.029	5.1	95	0.00
73		Carbazole	0.935	0.892	4.6	95	0.00
74		Di-n-butylphthalate	1.240	1.200	3.2	99	0.00
75	C	Fluoranthene	1.146	1.076	6.1	94	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	92	0.00
77		Benzidine	0.279	0.401	-43.7#	103	0.00
78		Pyrene	1.730	1.667	3.6	95	0.00
79	S	Terphenyl-d14	1.353	1.349	0.3	97	0.00
80		Butylbenzylphthalate	0.677	0.690	-1.9	97	0.00
81		Benzo(a)anthracene	1.312	1.287	1.9	93	0.00
82		3,3'-Dichlorobenzidine	0.379	0.363	4.2	90	0.00
83		Chrysene	1.190	1.106	7.1	90	0.00
84		Bis(2-ethylhexyl)phthalate	0.934	0.960	-2.8	98	0.00
85	c	Di-n-octyl phthalate	1.299	1.331	-2.5	97	0.00
86	I	Perylene-d12	1.000	1.000	0.0	91	0.00
87		Indeno(1,2,3-cd)pyrene	1.294	1.244	3.9	88	0.00
88		Benzo(b)fluoranthene	1.371	1.367	0.3	103	0.00
89		Benzo(k)fluoranthene	1.173	1.019	13.1	77	0.00
90	C	Benzo(a)pyrene	1.073	1.035	3.5	91	0.00
91		Dibenzo(a,h)anthracene	1.067	1.039	2.6	88	0.00
92		Benzo(g,h,i)perylene	1.055	1.003	4.9	86	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2		1,4-Dioxane	40.000	38.670	3.3	91	0.00
3		Pyridine	40.000	39.050	2.4	93	-0.01
4		n-Nitrosodimethylamine	40.000	38.959	2.6	92	0.00
5	S	2-Fluorophenol	80.000	76.823	4.0	94	0.00
6		Aniline	40.000	38.958	2.6	94	0.00
7	S	Phenol-d6	80.000	76.962	3.8	96	0.00
8		2-Chlorophenol	40.000	38.721	3.2	96	0.00
9		Benzaldehyde	40.000	37.950	5.1	98	0.00
10	C	Phenol	40.000	39.115	2.2	96	0.00
11		bis(2-Chloroethyl)ether	40.000	38.246	4.4	95	0.00
12		1,3-Dichlorobenzene	40.000	38.391	4.0	94	0.00
13	C	1,4-Dichlorobenzene	40.000	38.366	4.1	94	0.00
14		1,2-Dichlorobenzene	40.000	38.744	3.1	96	0.00
15		Benzyl Alcohol	40.000	39.408	1.5	96	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	37.338	6.7	94	0.00
17		2-Methylphenol	40.000	38.775	3.1	96	0.00
18		Hexachloroethane	40.000	38.797	3.0	96	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	38.038	4.9	96	0.00
20		3+4-Methylphenols	40.000	38.619	3.5	95	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22		Acetophenone	40.000	37.650	5.9	94	0.00
23	S	Nitrobenzene-d5	80.000	78.962	1.3	97	0.00
24		Nitrobenzene	40.000	38.980	2.6	96	0.00
25		Isophorone	40.000	38.163	4.6	96	0.00
26	C	2-Nitrophenol	40.000	40.429	-1.1	96	0.00
27		2,4-Dimethylphenol	40.000	38.414	4.0	96	0.00
28		bis(2-Chloroethoxy)methane	40.000	37.943	5.1	97	0.00
29	C	2,4-Dichlorophenol	40.000	38.952	2.6	97	0.00
30		1,2,4-Trichlorobenzene	40.000	38.336	4.2	97	0.00
31		Naphthalene	40.000	37.827	5.4	95	0.00
32		Benzoic acid	40.000	42.303	-5.8	103	0.00
33		4-Chloroaniline	40.000	38.847	2.9	97	0.00
34	C	Hexachlorobutadiene	40.000	38.849	2.9	98	0.00
35		Caprolactam	40.000	37.276	6.8	96	0.00
36	C	4-Chloro-3-methylphenol	40.000	38.550	3.6	96	0.00
37		2-Methylnaphthalene	40.000	37.935	5.2	96	0.00
38		1-Methylnaphthalene	40.000	38.484	3.8	98	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	38.191	4.5	96	0.00
41	P	Hexachlorocyclopentadiene	40.000	38.281	4.3	89	0.00
42	S	2,4,6-Tribromophenol	80.000	79.342	0.8	98	0.00
43	C	2,4,6-Trichlorophenol	40.000	39.292	1.8	96	0.00
44		2,4,5-Trichlorophenol	40.000	38.263	4.3	97	0.00
45	S	2-Fluorobiphenyl	80.000	75.528	5.6	97	0.00
46		1,1'-Biphenyl	40.000	38.182	4.5	98	0.00
47		2-Chloronaphthalene	40.000	38.090	4.8	97	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	38.908	2.7	97	0.00
49		Acenaphthylene	40.000	37.725	5.7	95	0.00
50		Dimethylphthalate	40.000	37.611	6.0	97	0.00
51		2,6-Dinitrotoluene	40.000	38.970	2.6	97	0.00
52	C	Acenaphthene	40.000	38.405	4.0	97	0.00
53		3-Nitroaniline	40.000	38.561	3.6	95	0.00
54	P	2,4-Dinitrophenol	40.000	39.326	1.7	102	0.00
55		Dibenzofuran	40.000	37.778	5.6	97	0.00
56	P	4-Nitrophenol	40.000	39.425	1.4	94	0.00
57		2,4-Dinitrotoluene	40.000	39.291	1.8	98	0.00
58		Fluorene	40.000	37.381	6.5	97	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	39.070	2.3	98	0.00
60		Diethylphthalate	40.000	38.337	4.2	98	0.00
61		4-Chlorophenyl-phenylether	40.000	38.272	4.3	99	0.00
62		4-Nitroaniline	40.000	38.953	2.6	95	0.00
63		Azobenzene	40.000	37.911	5.2	97	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65		4,6-Dinitro-2-methylphenol	40.000	41.415	-3.5	99	0.00
66	c	n-Nitrosodiphenylamine	40.000	38.283	4.3	97	0.00
67		4-Bromophenyl-phenylether	40.000	39.015	2.5	97	0.00
68		Hexachlorobenzene	40.000	39.789	0.5	99	0.00
69		Atrazine	40.000	31.945	20.1	94	0.00
70	C	Pentachlorophenol	40.000	41.955	-4.9	97	0.00
71		Phenanthrene	40.000	38.357	4.1	97	0.00
72		Anthracene	40.000	37.991	5.0	95	0.00
73		Carbazole	40.000	38.167	4.6	95	0.00
74		Di-n-butylphthalate	40.000	38.717	3.2	99	0.00
75	C	Fluoranthene	40.000	37.545	6.1	94	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77		Benzidine	40.000	57.513	-43.8#	103	0.00
78		Pyrene	40.000	38.539	3.7	95	0.00
79	S	Terphenyl-d14	80.000	79.748	0.3	97	0.00
80		Butylbenzylphthalate	40.000	40.778	-1.9	97	0.00
81		Benzo(a)anthracene	40.000	39.211	2.0	93	0.00
82		3,3'-Dichlorobenzidine	40.000	38.337	4.2	90	0.00
83		Chrysene	40.000	37.195	7.0	90	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	41.142	-2.9	98	0.00
85	c	Di-n-octyl phthalate	40.000	40.971	-2.4	97	0.00
86	I	Perylene-d12	20.000	20.000	0.0	91	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	38.467	3.8	88	0.00
88		Benzo(b)fluoranthene	40.000	39.906	0.2	103	0.00
89		Benzo(k)fluoranthene	40.000	34.765	13.1	77	0.00
90	C	Benzo(a)pyrene	40.000	38.612	3.5	91	0.00
91		Dibenzo(a,h)anthracene	40.000	38.931	2.7	88	0.00
92		Benzo(g,h,i)perylene	40.000	38.051	4.9	86	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141906.D
Acq On : 10 Mar 2025 15:53
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031025

Quant Time: Mar 10 16:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Response Factor Report MSBNA_P

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP031225.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Mar 13 05:55:58 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BP024160.D 5 =BP024161.D 10 =BP024162.D 20 =BP024163.D 40 =BP024164.D 50 =BP024165.D 60 =BP024166.D 80 =BP0241
67.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------ISTD----------									
2)		1,4-Dioxane		0.503	0.499	0.518	0.469	0.496	0.498	0.493	0.497	2.93
3)		Pyridine		1.296	1.269	1.406	1.311	1.396	1.398	1.429	1.358	4.69
4)		n-Nitrosodimet...		0.444	0.436	0.476	0.431	0.461	0.466	0.461	0.454	3.66
5)	S	2-Fluorophenol		1.158	1.194	1.307	1.211	1.291	1.307	1.303	1.253	5.06
6)		Aniline		1.508	1.531	1.656	1.484	1.548	1.512	1.447	1.527	4.31
7)	S	Phenol-d6		1.518	1.581	1.747	1.631	1.749	1.754	1.748	1.675	5.86
8)		2-Chlorophenol		1.257	1.310	1.413	1.306	1.391	1.400	1.397	1.353	4.52
9)		Benzaldehyde		0.971	0.949	0.930	0.780	0.751	0.702	0.577	0.809	18.20
10)	C	Phenol		1.579	1.632	1.754	1.628	1.744	1.752	1.751	1.691	4.46
11)		bis(2-Chloroet...		1.286	1.317	1.407	1.273	1.341	1.358	1.356	1.334	3.45
12)		1,3-Dichlorobe...		1.466	1.448	1.522	1.387	1.477	1.471	1.445	1.459	2.79
13)	C	1,4-Dichlorobe...		1.507	1.474	1.562	1.413	1.493	1.503	1.479	1.490	3.01
14)		1,2-Dichlorobe...		1.446	1.426	1.494	1.358	1.435	1.440	1.411	1.430	2.85
15)		Benzyl Alcohol		0.934	0.966	1.096	1.042	1.122	1.126	1.139	1.061	7.77
16)		2,2'-oxybis(1-...		1.374	1.367	1.434	1.293	1.365	1.362	1.336	1.361	3.11
17)		2-Methylphenol		0.932	0.997	1.120	1.038	1.116	1.113	1.109	1.061	6.96
18)		Hexachloroethane		0.533	0.512	0.547	0.508	0.533	0.538	0.529	0.529	2.67
19)	P	n-Nitroso-di-n...	0.869	0.910	0.956	1.048	0.962	1.026	1.007	0.997	0.972	6.18
20)		3+4-Methylphenols		1.265	1.360	1.530	1.433	1.547	1.530	1.534	1.457	7.50
21)	I	Naphthalene-d8	----------ISTD----------									
22)		Acetophenone		0.483	0.490	0.531	0.496	0.508	0.519	0.512	0.506	3.37
23)	S	Nitrobenzene-d5		0.330	0.338	0.366	0.351	0.362	0.368	0.367	0.354	4.34
24)		Nitrobenzene		0.330	0.336	0.362	0.341	0.355	0.363	0.361	0.350	3.88
25)		Isophorone		0.532	0.562	0.631	0.602	0.635	0.642	0.649	0.608	7.33
26)	C	2-Nitrophenol		0.126	0.141	0.167	0.171	0.182	0.189	0.192	0.167	14.92
27)		2,4-Dimethylph...		0.186	0.198	0.220	0.215	0.223	0.229	0.232	0.215	7.94
28)		bis(2-Chloroet...		0.414	0.418	0.438	0.418	0.429	0.437	0.439	0.427	2.55
29)	C	2,4-Dichloroph...		0.237	0.263	0.298	0.296	0.306	0.314	0.319	0.291	10.21
30)		1,2,4-Trichlor...		0.310	0.312	0.329	0.312	0.322	0.331	0.331	0.321	2.94
31)		Naphthalene		1.041	1.037	1.099	1.029	1.069	1.078	1.069	1.060	2.38
32)		Benzoic acid			0.148	0.190	0.218	0.239	0.252		0.209	19.82
33)		4-Chloroaniline		0.344	0.356	0.397	0.377	0.394	0.389	0.398	0.379	5.63
34)	C	Hexachlorobuta...		0.177	0.178	0.187	0.180	0.187	0.192	0.191	0.185	3.36
35)		Caprolactam		0.072	0.079	0.097	0.097	0.107	0.104	0.111	0.095	15.56
36)	C	4-Chloro-3-met...		0.261	0.283	0.324	0.314	0.333	0.332	0.348	0.313	9.83
37)		2-Methylnaphth...		0.689	0.684	0.737	0.704	0.731	0.734	0.735	0.716	3.27
38)		1-Methylnaphth...		0.674	0.673	0.728	0.673	0.704	0.705	0.720	0.697	3.32

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP031225.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.548	0.560	0.596	0.558	0.580	0.608	0.596	0.578	3.98
41)	P	Hexachlorocycl...	0.169	0.189	0.214	0.212	0.220	0.234	0.222	0.208	10.65
42)	S	2,4,6-Tribromo...	0.203	0.222	0.256	0.252	0.270	0.275	0.285	0.252	11.69
43)	C	2,4,6-Trichlor...	0.291	0.324	0.379	0.372	0.391	0.409	0.406	0.367	11.98
44)		2,4,5-Trichlor...	0.331	0.362	0.414	0.409	0.422	0.439	0.443	0.403	10.23
45)	S	2-Fluorobiphenyl	1.364	1.358	1.420	1.319	1.335	1.367	1.332	1.356	2.44
46)		1,1'-Biphenyl	1.467	1.475	1.566	1.462	1.493	1.536	1.521	1.503	2.60
47)		2-Chloronaphth...	1.108	1.100	1.180	1.095	1.115	1.164	1.152	1.130	3.01
48)		2-Nitroaniline	0.228	0.252	0.300	0.292	0.311	0.317	0.324	0.289	12.34
49)		Acenaphthylene	1.586	1.628	1.776	1.661	1.749	1.789	1.784	1.710	4.91
50)		Dimethylphthalate	1.351	1.349	1.459	1.358	1.417	1.424	1.439	1.399	3.27
51)		2,6-Dinitrotol...	0.248	0.266	0.303	0.294	0.312	0.316	0.322	0.294	9.40
52)	C	Acenaphthene	1.057	1.056	1.136	1.061	1.096	1.123	1.121	1.093	3.18
53)		3-Nitroaniline	0.253	0.276	0.328	0.322	0.344	0.343	0.359	0.318	12.24
54)	P	2,4-Dinitrophenol		0.108	0.146	0.160	0.179	0.186		0.156	19.97
55)		Dibenzofuran	1.759	1.751	1.835	1.711	1.768	1.799	1.769	1.770	2.18
56)	P	4-Nitrophenol	0.187	0.218	0.271	0.277	0.299	0.296	0.312	0.266	17.41
57)		2,4-Dinitrotol...	0.295	0.334	0.400	0.394	0.423	0.431	0.446	0.389	14.19
58)		Fluorene	1.309	1.343	1.442	1.345	1.407	1.407	1.402	1.379	3.43
59)		2,3,4,6-Tetrac...	0.316	0.323	0.365	0.356	0.371	0.376	0.386	0.356	7.54
60)		Diethylphthalate	1.297	1.336	1.437	1.358	1.422	1.435	1.469	1.394	4.53
61)		4-Chlorophenyl...	0.658	0.660	0.696	0.658	0.685	0.689	0.690	0.676	2.53
62)		4-Nitroaniline	0.246	0.292	0.348	0.343	0.362	0.365	0.379	0.334	14.22
63)		Azobenzene	1.215	1.251	1.369	1.291	1.341	1.363	1.365	1.314	4.73
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.085	0.111	0.120	0.127	0.132	0.137	0.119	15.85
66)	c	n-Nitrosodiphe...	0.564	0.583	0.629	0.600	0.617	0.632	0.626	0.607	4.27
67)		4-Bromophenyl-...	0.206	0.207	0.222	0.217	0.224	0.229	0.231	0.220	4.60
68)		Hexachlorobenzene	0.248	0.249	0.263	0.254	0.263	0.270	0.272	0.260	3.76
69)		Atrazine	0.165	0.171	0.157	0.135	0.105			0.147	18.49
70)	C	Pentachlorophenol	0.121	0.140	0.166	0.177	0.184	0.190	0.197	0.168	16.61
71)		Phenanthrene	1.068	1.059	1.126	1.083	1.100	1.111	1.102	1.093	2.19
72)		Anthracene	0.993	1.012	1.101	1.063	1.092	1.090	1.098	1.064	4.16
73)		Carbazole	0.920	0.975	1.060	1.041	1.060	1.061	1.066	1.026	5.50
74)		Di-n-butylphth...	1.026	1.121	1.247	1.260	1.272	1.297	1.304	1.218	8.57
75)	C	Fluoranthene	1.173	1.220	1.287	1.264	1.305	1.298	1.312	1.265	4.07
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine	0.222	0.199	0.178	0.258	0.173	0.278	0.222	0.219	17.88
78)		Pyrene	1.177	1.190	1.299	1.204	1.266	1.293	1.274	1.243	4.12
79)	S	Terphenyl-d14	0.967	0.973	1.033	0.948	0.966	0.971	0.924	0.969	3.45
80)		Butylbenzylpht...	0.399	0.441	0.513	0.524	0.554	0.581	0.576	0.513	13.49
81)		Benzo(a)anthra...	1.194	1.194	1.291	1.208	1.261	1.285	1.271	1.243	3.47
82)		3,3'-Dichlorob...	0.340	0.369	0.429	0.428	0.434	0.467	0.450	0.417	10.94
83)		Chrysene	1.185	1.159	1.224	1.158	1.212	1.220	1.209	1.195	2.36
84)		Bis(2-ethylhex...	0.613	0.678	0.784	0.792	0.819	0.878	0.851	0.774	12.30
85)	c	Di-n-octyl pht...	0.922	1.033	1.224	1.272	1.357	1.435	1.433	1.240	15.95

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP031225.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.274	1.289	1.421	1.366	1.442	1.479	1.499	1.396	6.38
88)		Benzo(b)fluora...	1.103	1.116	1.258	1.190	1.244	1.266	1.264	1.206	5.85
89)		Benzo(k)fluora...	1.092	1.136	1.213	1.130	1.203	1.251	1.232	1.180	5.07
90)	C	Benzo(a)pyrene	0.953	0.959	1.065	1.029	1.094	1.119	1.122	1.049	6.77
91)		Dibenzo(a,h)an...	1.063	1.083	1.184	1.137	1.191	1.226	1.244	1.161	5.97
92)		Benzo(g,h,i)pe...	1.097	1.107	1.199	1.154	1.202	1.246	1.259	1.180	5.39

(#) = Out of Range

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024160.D
Acq On : 12 Mar 2025 16:17
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Mar 13 04:04:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	323931	20.000	ng	0.00
21) Naphthalene-d8	10.546	136	1324397	20.000	ng	0.00
39) Acenaphthene-d10	14.410	164	805172	20.000	ng	0.00
64) Phenanthrene-d10	17.216	188	1551813	20.000	ng	0.01
76) Chrysene-d12	21.669	240	1615554	20.000	ng	0.01
86) Perylene-d12	25.074	264	1740877	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	8.569	70	35203	2.454	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024160.D
Acq On : 12 Mar 2025 16:17
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Mar 13 04:04:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 323931

Ion	Ratio	Lower	Upper
152	100		
150	155.9	125.5	188.3
115	57.6	46.3	69.5







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.38 min

Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	51.6
63	25.8






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.95 min

Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99


Sig	Exp Ratio
99	100
42	13.3
71	29.1




#19
n-Nitroso-di-n-propylamine
Concen: 2.454 ng
RT: 8.569 min Scan# 932
Delta R.T. -0.006 min
Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Tgt Ion: 70 Resp: 35203

Ion	Ratio	Lower	Upper
70	100		
42	43.9	36.0	54.0
101	10.2	8.3	12.5
130	24.1	18.6	28.0







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.546 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 1324397

Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	6.4	5.3	7.9
68	6.2	5.0	7.4







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 8.92 min

Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	46.7
54	40.6







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 805172

Ion	Ratio	Lower	Upper
164	100		
162	100.1	80.6	121.0
160	44.9	35.4	53.2







#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.92 min

Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	97.0
141	31.3






#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 13.02 min

Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172


Sig	Exp Ratio
172	100
171	35.4
170	23.7




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.216 min Scan# 2402
Delta R.T. 0.012 min
Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Tgt Ion:188 Resp: 1551813

Ion	Ratio	Lower	Upper
188	100		
94	9.8	8.1	12.1
80	10.9	8.9	13.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.669 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 1615554

Ion	Ratio	Lower	Upper
240	100		
120	11.3	9.2	13.8
236	25.5	20.4	30.6







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 19.92 min

Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	7.1
122	12.0






#86
Perylene-d12
Concen: 20.000 ng
RT: 25.074 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024160.D
Acq: 12 Mar 2025 16:17

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 1740877

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.5	27.7
265	22.6	17.7	26.5



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024161.D
Acq On : 12 Mar 2025 16:57
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Mar 13 04:05:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.769	152	318045	20.000	ng		0.00
21) Naphthalene-d8	10.552	136	1309940	20.000	ng		0.00
39) Acenaphthene-d10	14.410	164	799469	20.000	ng		0.00
64) Phenanthrene-d10	17.222	188	1516956	20.000	ng		0.02
76) Chrysene-d12	21.669	240	1578855	20.000	ng		0.01
86) Perylene-d12	25.080	264	1694453	20.000	ng		0.02
System Monitoring Compounds							
5) 2-Fluorophenol	5.381	112	184089	9.032	ng		0.00
7) Phenol-d6	6.952	99	241350	9.213	ng		0.00
23) Nitrobenzene-d5	8.916	82	216074	9.202	ng		0.00
42) 2,4,6-Tribromophenol	15.934	330	81159	8.288	ng		0.02
45) 2-Fluorobiphenyl	13.016	172	545120	9.532	ng		0.00
79) Terphenyl-d14	19.951	244	763023	9.563	ng		0.03
Target Compounds							Qvalue
2) 1,4-Dioxane	3.323	88	40006	5.003	ng		97
3) Pyridine	3.723	79	103023	4.956	ng		97
4) n-Nitrosodimethylamine	3.628	42	35306	4.788	ng	#	92
6) Aniline	7.105	93	119908	4.998	ng		100
8) 2-Chlorophenol	7.340	128	99927	4.632	ng		98
9) Benzaldehyde	6.922	77	77223m	6.456	ng		
10) Phenol	6.975	94	125574	4.767	ng		98
11) bis(2-Chloroethyl)ether	7.199	93	102214	4.926	ng		98
12) 1,3-Dichlorobenzene	7.664	146	116531	4.917	ng		100
13) 1,4-Dichlorobenzene	7.805	146	119798	4.998	ng		99
14) 1,2-Dichlorobenzene	8.122	146	114960	4.998	ng		99
15) Benzyl Alcohol	8.011	79	74252	4.630	ng		96
16) 2,2'-oxybis(1-Chloropr...	8.287	45	109209	5.422	ng		98
17) 2-Methylphenol	8.216	107	74107	4.616	ng		96
18) Hexachloroethane	8.846	117	42376	4.909	ng		96
19) n-Nitroso-di-n-propyla...	8.569	70	72383	5.139	ng		98
20) 3+4-Methylphenols	8.540	107	100588	4.606	ng		98
22) Acetophenone	8.587	105	158185	4.699	ng	#	99
24) Nitrobenzene	8.958	77	108086	4.677	ng		96
25) Isophorone	9.487	82	174338	4.540	ng		100
26) 2-Nitrophenol	9.669	139	41305	3.880	ng		98
27) 2,4-Dimethylphenol	9.734	122	60843	4.354	ng		96
28) bis(2-Chloroethoxy)met...	9.957	93	135529	4.872	ng		99
29) 2,4-Dichlorophenol	10.205	162	77653	4.072	ng		99
30) 1,2,4-Trichlorobenzene	10.416	180	101518	4.646	ng		99
31) Naphthalene	10.599	128	340918	4.840	ng		99
33) 4-Chloroaniline	10.710	127	112776	4.545	ng		98
34) Hexachlorobutadiene	10.887	225	58080	4.589	ng		97
35) Caprolactam	11.475	113	23432	4.057	ng		95
36) 4-Chloro-3-methylphenol	11.840	107	85333	4.352	ng		99
37) 2-Methylnaphthalene	12.216	142	225485	4.817	ng		100
38) 1-Methylnaphthalene	12.434	142	220784	4.848	ng		99
40) 1,2,4,5-Tetrachloroben...	12.587	216	109528	4.459	ng		99
41) Hexachlorocyclopentadiene	12.569	237	33755	3.693	ng		100
43) 2,4,6-Trichlorophenol	12.822	196	58172	3.883	ng		98
44) 2,4,5-Trichlorophenol	12.904	196	66250	4.044	ng		98

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024161.D
Acq On : 12 Mar 2025 16:57
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Mar 13 04:05:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

	Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46)	1,1'-Biphenyl	13.234	154	293289	4.721	ng	98
47)	2-Chloronaphthalene	13.275	162	221413	4.715	ng	98
48)	2-Nitroaniline	13.475	65	45667	4.117	ng	88
49)	Acenaphthylene	14.128	152	316902	4.580	ng	99
50)	Dimethylphthalate	13.857	163	270053	4.789	ng	100
51)	2,6-Dinitrotoluene	13.975	165	49593	4.317	ng	97
52)	Acenaphthene	14.475	154	211290	4.742	ng	99
53)	3-Nitroaniline	14.310	138	50611	4.129	ng	99
55)	Dibenzofuran	14.822	168	351651	4.858	ng	100
56)	4-Nitrophenol	14.645	139	37281	3.521	ng	93
57)	2,4-Dinitrotoluene	14.775	165	58882	3.946	ng	90
58)	Fluorene	15.481	166	261676	4.629	ng	99
59)	2,3,4,6-Tetrachlorophenol	15.045	232	63120	4.441	ng	98
60)	Diethylphthalate	15.251	149	259242	4.631	ng	98
61)	4-Chlorophenyl-phenyle...	15.481	204	131441	4.724	ng	99
62)	4-Nitroaniline	15.498	138	49239	3.871	ng	95
63)	Azobenzene	15.775	77	242768	4.628	ng	97
66)	n-Nitrosodiphenylamine	15.692	169	213927	4.522	ng	99
67)	4-Bromophenyl-phenylether	16.392	248	78309	4.597	ng	93
68)	Hexachlorobenzene	16.510	284	94130	4.696	ng	98
69)	Atrazine	16.669	200	62551	5.152	ng	98
70)	Pentachlorophenol	16.869	266	45885	3.525	ng	99
71)	Phenanthrene	17.263	178	404881	4.734	ng	99
72)	Anthracene	17.357	178	376601	4.595	ng	99
73)	Carbazole	17.634	167	349000	4.469	ng	99
74)	Di-n-butylphthalate	18.234	149	389057	4.270	ng	99
75)	Fluoranthene	19.351	202	444670	4.597	ng	97
77)	Benzidine	19.545	184	87777	4.562	ng	98
78)	Pyrene	19.722	202	464704	4.634	ng	99
80)	Butylbenzylphthalate	20.675	149	157601	4.101	ng	99
81)	Benzo(a)anthracene	21.651	228	471347	4.761	ng	99
82)	3,3'-Dichlorobenzidine	21.580	252	134146	4.336	ng	99
83)	Chrysene	21.722	228	467600	4.898	ng	99
84)	Bis(2-ethylhexyl)phtha...	21.598	149	242048	4.153	ng	99
85)	Di-n-octyl phthalate	22.898	149	363786	3.905	ng	98
87)	Indeno(1,2,3-cd)pyrene	28.945	276	539523	4.476	ng	97
88)	Benzo(b)fluoranthene	24.004	252	467406	4.344	ng	99
89)	Benzo(k)fluoranthene	24.069	252	462653	4.417	ng	99
90)	Benzo(a)pyrene	24.910	252	403692	4.421	ng	99
91)	Dibenzo(a,h)anthracene	29.015	278	450461	4.444	ng	98
92)	Benzo(g,h,i)perylene	30.133	276	464721	4.518	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024161.D
Acq On : 12 Mar 2025 16:57
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Mar 13 04:05:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:152 Resp: 318045

Ion	Ratio	Lower	Upper
152	100		
150	156.0	125.5	188.3
115	57.9	46.3	69.5







#2
1,4-Dioxane
Concen: 5.003 ng
RT: 3.323 min Scan# 40
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion: 88 Resp: 40006

Ion	Ratio	Lower	Upper
88	100		
58	64.7	49.8	74.6
43	22.3	17.3	25.9







#3
Pyridine
Concen: 4.956 ng
RT: 3.723 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 103023

Ion	Ratio	Lower	Upper
79	100		
52	55.7	43.8	65.8
51	30.1	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 4.788 ng
RT: 3.628 min Scan# 92
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion: 42 Resp: 35306

Ion	Ratio	Lower	Upper
42	100		
74	165.1	139.8	209.6
44	15.9	5.3	7.9#







#5
2-Fluorophenol
Concen: 9.032 ng
RT: 5.381 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:112 Resp: 184089

Ion	Ratio	Lower	Upper
112	100		
64	52.7	41.3	61.9
63	26.7	20.6	31.0







#6
Aniline
Concen: 4.998 ng
RT: 7.105 min Scan# 683
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion: 93 Resp: 119908

Ion	Ratio	Lower	Upper
93	100		
66	35.3	28.2	42.2
65	17.6	14.0	21.0







#7
Phenol-d6
Concen: 9.213 ng
RT: 6.952 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 99 Resp: 241350

Ion	Ratio	Lower	Upper
99	100		
42	13.3	10.6	16.0
71	27.9	23.3	34.9







#8
2-Chlorophenol
Concen: 4.632 ng
RT: 7.340 min Scan# 723
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:128 Resp: 99927

Ion	Ratio	Lower	Upper
128	100		
130	32.4	12.0	52.0
64	43.0	21.6	61.6







#9
Benzaldehyde
Concen: 6.456 ng m
RT: 6.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 77 Resp: 77223

Ion	Ratio	Lower	Upper
77	100		
105	102.6	81.4	121.4
106	96.7	81.1	121.1







#10
Phenol
Concen: 4.767 ng
RT: 6.975 min Scan# 661
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion: 94 Resp: 125574

Ion	Ratio	Lower	Upper
94	100		
65	24.9	4.9	44.9
66	30.1	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 4.926 ng
RT: 7.199 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 102214

Ion	Ratio	Lower	Upper
93	100		
63	68.8	47.4	87.4
95	31.5	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 4.917 ng
RT: 7.664 min Scan# 778
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:146 Resp: 116531

Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.6	77.4
75	28.6	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 4.998 ng
RT: 7.805 min Scan# 8(
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:146 Resp: 119798

Ion	Ratio	Lower	Upper
146	100		
148	65.2	51.9	77.9
111	38.3	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 4.998 ng
RT: 8.122 min Scan# 856
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:146 Resp: 114960

Ion	Ratio	Lower	Upper
146	100		
148	65.7	51.4	77.2
111	41.4	33.2	49.8







#15
Benzyl Alcohol
Concen: 4.630 ng
RT: 8.011 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 74252

Ion	Ratio	Lower	Upper
79	100		
108	77.9	66.5	99.7
77	64.9	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.422 ng
RT: 8.287 min Scan# 884
Delta R.T. -0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion: 45 Resp: 109209

Ion	Ratio	Lower	Upper
45	100		
77	19.2	0.0	39.0
79	14.0	0.0	35.6







#17
2-Methylphenol
Concen: 4.616 ng
RT: 8.216 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 74107

Ion	Ratio	Lower	Upper
107	100		
108	105.7	90.0	135.0
77	42.2	34.0	51.0
79	40.3	32.7	49.1







#18
Hexachloroethane
Concen: 4.909 ng
RT: 8.846 min Scan# 979
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:117 Resp: 42376

Ion	Ratio	Lower	Upper
117	100		
119	93.0	77.4	116.0
201	91.0	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 5.139 ng
RT: 8.569 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 70 Resp: 72383

Ion	Ratio	Lower	Upper
70	100		
42	46.0	36.0	54.0
101	10.0	8.3	12.5
130	21.5	18.6	28.0







#20
3+4-Methylphenols
Concen: 4.606 ng
RT: 8.540 min Scan# 927
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:107 Resp: 100588

Ion	Ratio	Lower	Upper
107	100		
108	91.8	69.7	109.7
77	31.0	10.3	50.3
79	27.0	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.552 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:136 Resp: 1309940

Ion	Ratio	Lower	Upper
136	100		
137	10.6	9.0	13.4
54	6.1	5.3	7.9
68	6.0	5.0	7.4







#22
Acetophenone
Concen: 4.699 ng
RT: 8.587 min Scan# 935
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:105 Resp: 158185

Ion	Ratio	Lower	Upper
105	100		
71	1.3	2.4	3.6#
51	20.5	16.2	24.2
120	23.5	19.0	28.6







#23
Nitrobenzene-d5
Concen: 9.202 ng
RT: 8.916 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 82 Resp: 216074

Ion	Ratio	Lower	Upper
82	100		
128	43.3	37.4	56.0
54	40.8	32.5	48.7







#24
Nitrobenzene
Concen: 4.677 ng
RT: 8.958 min Scan# 998
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion: 77 Resp: 108086

Ion	Ratio	Lower	Upper
77	100		
123	46.0	39.4	59.2
65	12.4	9.9	14.9







#25
Isophorone
Concen: 4.540 ng
RT: 9.487 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 82 Resp: 174338

Ion	Ratio	Lower	Upper
82	100		
95	6.4	5.3	7.9
138	18.6	14.8	22.2







#26
2-Nitrophenol
Concen: 3.880 ng
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:139 Resp: 41305

Ion	Ratio	Lower	Upper
139	100		
109	27.7	21.7	32.5
65	46.6	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 4.354 ng
RT: 9.734 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:122 Resp: 60843

Ion	Ratio	Lower	Upper
122	100		
107	107.7	89.3	133.9
121	55.4	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 4.872 ng
RT: 9.957 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion: 93 Resp: 135529

Ion	Ratio	Lower	Upper
93	100		
95	31.5	25.9	38.9
123	11.9	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 4.072 ng
RT: 10.205 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:162 Resp: 77653

Ion	Ratio	Lower	Upper
162	100		
164	63.6	44.2	84.2
98	34.2	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 4.646 ng
RT: 10.416 min Scan# 1246
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:180 Resp: 101518

Ion	Ratio	Lower	Upper
180	100		
182	96.2	76.1	114.1
145	31.2	24.2	36.2







#31
Naphthalene
Concen: 4.840 ng
RT: 10.599 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:128 Resp: 340918

Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.7	13.1
127	13.3	10.4	15.6







#33
4-Chloroaniline
Concen: 4.545 ng
RT: 10.710 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:127 Resp: 112776

Ion	Ratio	Lower	Upper
127	100		
129	32.6	25.8	38.6
65	28.7	21.2	31.8
92	18.7	14.4	21.6







#34
Hexachlorobutadiene
Concen: 4.589 ng
RT: 10.887 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:225 Resp: 58080

Ion	Ratio	Lower	Upper
225	100		
223	65.6	49.9	74.9
227	66.1	52.2	78.2







#35
Caprolactam
Concen: 4.057 ng
RT: 11.475 min Scan# 1426
Delta R.T. -0.024 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:113 Resp: 23432

Ion	Ratio	Lower	Upper
113	100		
55	136.8	113.6	153.6
56	108.1	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 4.352 ng
RT: 11.840 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 85333

Ion	Ratio	Lower	Upper
107	100		
144	26.0	21.6	32.4
142	81.1	65.4	98.0







#37
2-Methylnaphthalene
Concen: 4.817 ng
RT: 12.216 min Scan# 1552
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:142 Resp: 225485

Ion	Ratio	Lower	Upper
142	100		
141	88.0	70.3	105.5
115	32.4	26.0	39.0







#38
1-Methylnaphthalene
Concen: 4.848 ng
RT: 12.434 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:142 Resp: 220784

Ion	Ratio	Lower	Upper
142	100		
141	91.6	73.9	110.9
116	3.7	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1925
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:164 Resp: 799469

Ion	Ratio	Lower	Upper
164	100		
162	100.8	80.6	121.0
160	44.9	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.459 ng
RT: 12.587 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:216 Resp: 109528

Ion	Ratio	Lower	Upper
216	100		
214	79.5	62.8	94.2
179	21.1	16.8	25.2
108	18.8	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 3.693 ng
RT: 12.569 min Scan# 1612
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:237 Resp: 33755

Ion	Ratio	Lower	Upper
237	100		
235	61.6	41.8	81.8
272	14.2	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 8.288 ng
RT: 15.934 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:330 Resp: 81159

Ion	Ratio	Lower	Upper
330	100		
332	96.2	77.6	116.4
141	33.0	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 3.883 ng
RT: 12.822 min Scan# 1655
Delta R.T. -0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:196 Resp: 58172

Ion	Ratio	Lower	Upper
196	100		
198	93.4	76.6	115.0
200	30.9	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 4.044 ng
RT: 12.904 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:196 Resp: 66250

Ion	Ratio	Lower	Upper
196	100		
198	91.9	76.5	114.7
97	49.3	38.7	58.1
132	27.7	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 9.532 ng
RT: 13.016 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:172 Resp: 545120

Ion	Ratio	Lower	Upper
172	100		
171	34.6	28.3	42.5
170	23.7	19.0	28.4







#46
1,1'-Biphenyl
Concen: 4.721 ng
RT: 13.234 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 293289

Ion	Ratio	Lower	Upper
154	100		
153	39.3	20.3	60.3
76	13.7	0.0	32.9







#47
2-Chloronaphthalene
Concen: 4.715 ng
RT: 13.275 min Scan# 1732
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:162 Resp: 221413

Ion	Ratio	Lower	Upper
162	100		
127	38.3	31.5	47.3
164	31.8	26.6	39.8







#48
2-Nitroaniline
Concen: 4.117 ng
RT: 13.475 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 65 Resp: 45667

Ion	Ratio	Lower	Upper
65	100		
92	70.4	62.1	93.1
138	108.8	99.5	149.3







#49
Acenaphthylene
Concen: 4.580 ng
RT: 14.128 min Scan# 1877
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:152 Resp: 316902

Ion	Ratio	Lower	Upper
152	100		
151	19.6	16.0	24.0
153	13.0	10.6	15.8







#50
Dimethylphthalate
Concen: 4.789 ng
RT: 13.857 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:163 Resp: 270053

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.5	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 4.317 ng
RT: 13.975 min Scan# 1851
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:165 Resp: 49593

Ion	Ratio	Lower	Upper
165	100		
63	48.6	37.7	56.5
89	55.1	41.9	62.9







#52
Acenaphthene
Concen: 4.742 ng
RT: 14.475 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 211290

Ion	Ratio	Lower	Upper
154	100		
153	113.6	90.5	135.7
152	54.7	43.1	64.7







#53
3-Nitroaniline
Concen: 4.129 ng
RT: 14.310 min Scan# 1908
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:138 Resp: 50611

Ion	Ratio	Lower	Upper
138	100		
108	11.4	8.2	12.2
92	107.5	86.7	130.1







#55
Dibenzofuran
Concen: 4.858 ng
RT: 14.822 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:168 Resp: 351651

Ion	Ratio	Lower	Upper
168	100		
139	36.4	29.1	43.7
169	13.0	10.8	16.2







#56
4-Nitrophenol
Concen: 3.521 ng
RT: 14.645 min Scan# 1965
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:139 Resp: 37281

Ion	Ratio	Lower	Upper
139	100		
109	65.1	43.3	83.3
65	91.1	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 3.946 ng
RT: 14.775 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:165 Resp: 58882

Ion	Ratio	Lower	Upper
165	100		
63	39.8	29.0	43.6
89	78.5	54.2	81.4
182	3.0	2.3	3.5







#58
Fluorene
Concen: 4.629 ng
RT: 15.481 min Scan# 2107
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:166 Resp: 261676

Ion	Ratio	Lower	Upper
166	100		
165	98.2	77.6	116.4
167	13.1	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 4.441 ng
RT: 15.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:232 Resp: 63120

Ion	Ratio	Lower	Upper
232	100		
131	43.5	35.9	53.9
130	2.6	1.8	2.6
166	29.2	24.2	36.2







#60
Diethylphthalate
Concen: 4.631 ng
RT: 15.251 min Scan# 2068
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:149 Resp: 259242

Ion	Ratio	Lower	Upper
149	100		
177	23.3	17.7	26.5
150	12.3	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 4.724 ng
RT: 15.481 min Scan# 21
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:204 Resp: 131441

Ion	Ratio	Lower	Upper
204	100		
206	33.2	26.4	39.6
141	56.6	45.8	68.8







#62
4-Nitroaniline
Concen: 3.871 ng
RT: 15.498 min Scan# 2110
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:138 Resp: 49239

Ion	Ratio	Lower	Upper
138	100		
92	53.6	28.7	68.7
108	72.5	49.5	89.5







#63
Azobenzene
Concen: 4.628 ng
RT: 15.775 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 77 Resp: 242768

Ion	Ratio	Lower	Upper
77	100		
182	25.5	7.7	47.7
105	21.1	2.3	42.3
51	23.7	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.222 min Scan# 2403
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:188 Resp: 1516956

Ion	Ratio	Lower	Upper
188	100		
94	10.4	8.1	12.1
80	11.2	8.9	13.3







#66
n-Nitrosodiphenylamine
Concen: 4.522 ng
RT: 15.692 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:169 Resp: 213927

Ion	Ratio	Lower	Upper
169	100		
168	66.9	52.9	79.3
167	36.4	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 4.597 ng
RT: 16.392 min Scan# 2262
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:248 Resp: 78309

Ion	Ratio	Lower	Upper
248	100		
250	92.7	78.4	117.6
141	66.4	59.4	89.0







#68
Hexachlorobenzene
Concen: 4.696 ng
RT: 16.510 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:284 Resp: 94130

Ion	Ratio	Lower	Upper
284	100		
142	34.1	25.4	38.2
249	29.3	23.5	35.3







#69
Atrazine
Concen: 5.152 ng
RT: 16.669 min Scan# 2309
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:200 Resp: 62551

Ion	Ratio	Lower	Upper
200	100		
173	26.6	7.6	47.6
215	47.9	29.0	69.0







#70
Pentachlorophenol
Concen: 3.525 ng
RT: 16.869 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:266 Resp: 45885

Ion	Ratio	Lower	Upper
266	100		
268	65.2	51.3	76.9
264	63.3	50.6	76.0







#71
Phenanthrene
Concen: 4.734 ng
RT: 17.263 min Scan# 2410
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:178 Resp: 404881

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.6	23.4
179	15.5	12.5	18.7







#72
Anthracene
Concen: 4.595 ng
RT: 17.357 min Scan# 24
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:178 Resp: 376601

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.2	22.8
179	14.8	12.6	18.8







#73
Carbazole
Concen: 4.469 ng
RT: 17.634 min Scan# 2473
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:167 Resp: 349000

Ion	Ratio	Lower	Upper
167	100		
166	20.9	17.1	25.7
139	13.3	10.2	15.4







#74
Di-n-butylphthalate
Concen: 4.270 ng
RT: 18.234 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 389057

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.6	11.4
104	5.2	3.9	5.9







#75
Fluoranthene
Concen: 4.597 ng
RT: 19.351 min Scan# 2765
Delta R.T. 0.024 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:202 Resp: 444670

Ion	Ratio	Lower	Upper
202	100		
101	11.5	0.0	32.5
203	16.8	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.669 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:240 Resp: 1578855

Ion	Ratio	Lower	Upper
240	100		
120	12.2	9.2	13.8
236	25.5	20.4	30.6







#77
Benzidine
Concen: 4.562 ng
RT: 19.545 min Scan# 2798
Delta R.T. 0.024 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:184 Resp: 87777

Ion	Ratio	Lower	Upper
184	100		
185	15.7	11.8	17.6
183	10.9	9.1	13.7







#78
Pyrene
Concen: 4.634 ng
RT: 19.722 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:202 Resp: 464704

Ion	Ratio	Lower	Upper
202	100		
200	20.8	17.1	25.7
203	17.6	14.2	21.4







#79
Terphenyl-d14
Concen: 9.563 ng
RT: 19.951 min Scan# 2867
Delta R.T. 0.029 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:244 Resp: 763023

Ion	Ratio	Lower	Upper
244	100		
212	6.2	5.7	8.5
122	11.1	9.6	14.4







#80
Butylbenzylphthalate
Concen: 4.101 ng
RT: 20.675 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 157601

Ion	Ratio	Lower	Upper
149	100		
91	69.9	55.3	82.9
206	20.6	16.1	24.1







#81
Benzo(a)anthracene
Concen: 4.761 ng
RT: 21.651 min Scan# 3156
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:228 Resp: 471347

Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.8	32.8
229	19.6	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 4.336 ng
RT: 21.580 min Scan# 3
Delta R.T. 0.029 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 134146

Ion	Ratio	Lower	Upper
252	100		
254	65.3	51.4	77.2
126	12.4	9.3	13.9







#83
Chrysene
Concen: 4.898 ng
RT: 21.722 min Scan# 3168
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:228 Resp: 467600

Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.5	36.7
229	19.9	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 4.153 ng
RT: 21.598 min Scan# 3
Delta R.T. 0.029 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 242048

Ion	Ratio	Lower	Upper
149	100		
167	27.5	22.5	33.7
279	4.8	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 3.905 ng
RT: 22.898 min Scan# 3368
Delta R.T. 0.029 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:149 Resp: 363786

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.2	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.080 min Scan# 3
Delta R.T. 0.024 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:264 Resp: 1694453

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.5	27.7
265	21.5	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 4.476 ng
RT: 28.945 min Scan# 4396
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:276 Resp: 539523

Ion	Ratio	Lower	Upper
276	100		
138	20.0	13.7	20.5
277	25.2	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 4.344 ng
RT: 24.004 min Scan# 3
Delta R.T. 0.024 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 467406

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.2	25.8
125	10.9	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 4.417 ng
RT: 24.069 min Scan# 3567
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:252 Resp: 462653

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.7	26.5
125	11.3	9.1	13.7







#90
Benzo(a)pyrene
Concen: 4.421 ng
RT: 24.910 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 403692

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.4	26.0
125	13.4	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 4.444 ng
RT: 29.015 min Scan# 4408
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Tgt Ion:278 Resp: 450461

Ion	Ratio	Lower	Upper
278	100		
139	15.4	13.0	19.4
279	24.8	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 4.518 ng
RT: 30.133 min Scan# 4
Delta R.T. 0.006 min
Lab File: BP024161.D
Acq: 12 Mar 2025 16:57

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:276 Resp: 464721

Ion	Ratio	Lower	Upper
276	100		
277	24.7	19.2	28.8
138	21.4	17.4	26.2



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024162.D
Acq On : 12 Mar 2025 17:38
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Mar 13 04:06:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.769	152	350891	20.000	ng		0.00
21) Naphthalene-d8	10.546	136	1459823	20.000	ng		0.00
39) Acenaphthene-d10	14.404	164	878356	20.000	ng		0.00
64) Phenanthrene-d10	17.210	188	1659965	20.000	ng		0.00
76) Chrysene-d12	21.680	240	1774754	20.000	ng		0.02
86) Perylene-d12	25.074	264	1895165	20.000	ng		0.02
System Monitoring Compounds							
5) 2-Fluorophenol	5.381	112	419093	18.636	ng		0.00
7) Phenol-d6	6.946	99	554588	19.188	ng		0.00
23) Nitrobenzene-d5	8.916	82	493399	18.856	ng		0.00
42) 2,4,6-Tribromophenol	15.916	330	195393	18.161	ng		0.00
45) 2-Fluorobiphenyl	13.010	172	1192744	18.984	ng		0.00
79) Terphenyl-d14	19.939	244	1726609	19.251	ng		0.02
Target Compounds							Qvalue
2) 1,4-Dioxane	3.323	88	87630	9.934	ng		99
3) Pyridine	3.723	79	222685	9.709	ng		99
4) n-Nitrosodimethylamine	3.623	42	76580	9.413	ng	#	93
6) Aniline	7.105	93	268532	10.145	ng		99
8) 2-Chlorophenol	7.340	128	229815	9.655	ng		99
9) Benzaldehyde	6.922	77	166528	12.620	ng		99
10) Phenol	6.975	94	286246	9.849	ng		99
11) bis(2-Chloroethyl)ether	7.199	93	231087	10.095	ng		98
12) 1,3-Dichlorobenzene	7.658	146	254039	9.715	ng		99
13) 1,4-Dichlorobenzene	7.805	146	258553	9.777	ng		100
14) 1,2-Dichlorobenzene	8.122	146	250124	9.856	ng		100
15) Benzyl Alcohol	8.005	79	169563	9.584	ng		100
16) 2,2'-oxybis(1-Chloropr...	8.293	45	239855	10.794	ng		99
17) 2-Methylphenol	8.211	107	174913	9.876	ng		99
18) Hexachloroethane	8.840	117	89894	9.439	ng		98
19) n-Nitroso-di-n-propyla...	8.563	70	167789	10.797	ng		98
20) 3+4-Methylphenols	8.534	107	238685	9.906	ng		97
22) Acetophenone	8.581	105	357738	9.536	ng	#	99
24) Nitrobenzene	8.958	77	245585	9.536	ng		96
25) Isophorone	9.481	82	410357	9.590	ng		100
26) 2-Nitrophenol	9.669	139	103123	8.692	ng		98
27) 2,4-Dimethylphenol	9.728	122	144292	9.265	ng		99
28) bis(2-Chloroethoxy)met...	9.957	93	304956	9.838	ng		99
29) 2,4-Dichlorophenol	10.199	162	192263	9.046	ng		99
30) 1,2,4-Trichlorobenzene	10.410	180	227800	9.355	ng		98
31) Naphthalene	10.599	128	756826	9.642	ng		99
32) Benzoic acid	9.816	122	108330	7.587	ng		94
33) 4-Chloroaniline	10.705	127	260033	9.404	ng		98
34) Hexachlorobutadiene	10.887	225	130104	9.224	ng		98
35) Caprolactam	11.481	113	57479	8.931	ng		96
36) 4-Chloro-3-methylphenol	11.840	107	206744	9.462	ng		98
37) 2-Methylnaphthalene	12.210	142	498967	9.564	ng		99
38) 1-Methylnaphthalene	12.434	142	491304	9.680	ng		100
40) 1,2,4,5-Tetrachloroben...	12.581	216	246009	9.116	ng		100
41) Hexachlorocyclopentadiene	12.557	237	82803	8.245	ng		100
43) 2,4,6-Trichlorophenol	12.822	196	142353	8.649	ng		100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024162.D
Acq On : 12 Mar 2025 17:38
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Mar 13 04:06:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.904	196	159135	8.841	ng		98
46) 1,1'-Biphenyl	13.228	154	647980	9.494	ng		99
47) 2-Chloronaphthalene	13.269	162	483199	9.365	ng		98
48) 2-Nitroaniline	13.475	65	110660	9.079	ng		94
49) Acenaphthylene	14.122	152	714937	9.405	ng		99
50) Dimethylphthalate	13.857	163	592520	9.563	ng		100
51) 2,6-Dinitrotoluene	13.975	165	116661	9.243	ng		95
52) Acenaphthene	14.469	154	463749	9.473	ng		99
53) 3-Nitroaniline	14.304	138	121229	9.001	ng		90
54) 2,4-Dinitrophenol	14.516	184	47409	7.053	ng		94
55) Dibenzofuran	14.810	168	769059	9.671	ng		99
56) 4-Nitrophenol	14.634	139	95936	8.246	ng		92
57) 2,4-Dinitrotoluene	14.775	165	146673	8.946	ng		93
58) Fluorene	15.475	166	589720	9.495	ng		100
59) 2,3,4,6-Tetrachlorophenol	15.045	232	141666	9.071	ng		99
60) Diethylphthalate	15.240	149	586933	9.543	ng		100
61) 4-Chlorophenyl-phenyle...	15.463	204	289728	9.478	ng		99
62) 4-Nitroaniline	15.487	138	128152	9.169	ng		99
63) Azobenzene	15.763	77	549569	9.535	ng		97
65) 4,6-Dinitro-2-methylph...	15.545	198	70717	7.361	ng		92
66) n-Nitrosodiphenylamine	15.687	169	484111	9.352	ng		98
67) 4-Bromophenyl-phenylether	16.381	248	171620	9.207	ng		97
68) Hexachlorobenzene	16.492	284	206475	9.414	ng		98
69) Atrazine	16.663	200	142226	10.706	ng		99
70) Pentachlorophenol	16.857	266	116455	8.175	ng		100
71) Phenanthrene	17.257	178	878834	9.391	ng		99
72) Anthracene	17.351	178	839953	9.366	ng		99
73) Carbazole	17.634	167	809266	9.471	ng		99
74) Di-n-butylphthalate	18.228	149	930693	9.335	ng		99
75) Fluoranthene	19.351	202	1012600	9.566	ng		98
77) Benzidine	19.545	184	176282	8.151	ng		99
78) Pyrene	19.722	202	1056004	9.367	ng		99
80) Butylbenzylphthalate	20.675	149	390890	9.049	ng		99
81) Benzo(a)anthracene	21.657	228	1059397	9.520	ng		99
82) 3,3'-Dichlorobenzidine	21.569	252	327195	9.409	ng		99
83) Chrysene	21.727	228	1028579	9.585	ng		98
84) Bis(2-ethylhexyl)phtha...	21.592	149	601388	9.181	ng		99
85) Di-n-octyl phthalate	22.904	149	916941	8.757	ng		99
87) Indeno(1,2,3-cd)pyrene	28.951	276	1221650	9.061	ng	#	93
88) Benzo(b)fluoranthene	24.004	252	1057737	8.789	ng		99
89) Benzo(k)fluoranthene	24.074	252	1076560	9.190	ng		99
90) Benzo(a)pyrene	24.927	252	908542	8.896	ng		99
91) Dibenzo(a,h)anthracene	29.021	278	1026140	9.052	ng		99
92) Benzo(g,h,i)perylene	30.145	276	1049241	9.121	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024162.D
Acq On : 12 Mar 2025 17:38
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Mar 13 04:06:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 350891

Ion	Ratio	Lower	Upper
152	100		
150	156.9	125.5	188.3
115	59.8	46.3	69.5







#2
1,4-Dioxane
Concen: 9.934 ng
RT: 3.323 min Scan# 40
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion: 88 Resp: 87630

Ion	Ratio	Lower	Upper
88	100		
58	62.8	49.8	74.6
43	21.8	17.3	25.9







#3
Pyridine
Concen: 9.709 ng
RT: 3.723 min Scan# 1(
Delta R.T. 0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 222685

Ion	Ratio	Lower	Upper
79	100		
52	55.7	43.8	65.8
51	28.0	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 9.413 ng
RT: 3.623 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion: 42 Resp: 76580

Ion	Ratio	Lower	Upper
42	100		
74	165.6	139.8	209.6
44	9.7	5.3	7.9#







#5
2-Fluorophenol
Concen: 18.636 ng
RT: 5.381 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:112 Resp: 419093

Ion	Ratio	Lower	Upper
112	100		
64	51.8	41.3	61.9
63	26.0	20.6	31.0







#6
Aniline
Concen: 10.145 ng
RT: 7.105 min Scan# 683
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion: 93 Resp: 268532

Ion	Ratio	Lower	Upper
93	100		
66	35.2	28.2	42.2
65	18.4	14.0	21.0







#7
Phenol-d6
Concen: 19.188 ng
RT: 6.946 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 99 Resp: 554588

Ion	Ratio	Lower	Upper
99	100		
42	13.9	10.6	16.0
71	29.3	23.3	34.9







#8
2-Chlorophenol
Concen: 9.655 ng
RT: 7.340 min Scan# 723
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:128 Resp: 229815

Ion	Ratio	Lower	Upper
128	100		
130	31.9	12.0	52.0
64	42.2	21.6	61.6







#9
Benzaldehyde
Concen: 12.620 ng
RT: 6.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 166528

Ion	Ratio	Lower	Upper
77	100		
105	102.0	81.4	121.4
106	101.9	81.1	121.1







#10
Phenol
Concen: 9.849 ng
RT: 6.975 min Scan# 661
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion: 94 Resp: 286246

Ion	Ratio	Lower	Upper
94	100		
65	25.0	4.9	44.9
66	31.3	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 10.095 ng
RT: 7.199 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 231087

Ion	Ratio	Lower	Upper
93	100		
63	66.0	47.4	87.4
95	33.0	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 9.715 ng
RT: 7.658 min Scan# 777
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:146 Resp: 254039

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.6	77.4
75	29.0	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 9.777 ng
RT: 7.805 min Scan# 8(
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:146 Resp: 258553

Ion	Ratio	Lower	Upper
146	100		
148	64.8	51.9	77.9
111	39.9	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 9.856 ng
RT: 8.122 min Scan# 856
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:146 Resp: 250124

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.2
111	41.1	33.2	49.8







#15
Benzyl Alcohol
Concen: 9.584 ng
RT: 8.005 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 169563

Ion	Ratio	Lower	Upper
79	100		
108	82.9	66.5	99.7
77	64.1	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.794 ng
RT: 8.293 min Scan# 885
Delta R.T. -0.012 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion: 45 Resp: 239855

Ion	Ratio	Lower	Upper
45	100		
77	19.1	0.0	39.0
79	14.8	0.0	35.6







#17
2-Methylphenol
Concen: 9.876 ng
RT: 8.211 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:107 Resp: 174913

Ion	Ratio	Lower	Upper
107	100		
108	112.0	90.0	135.0
77	44.0	34.0	51.0
79	41.0	32.7	49.1







#18
Hexachloroethane
Concen: 9.439 ng
RT: 8.840 min Scan# 978
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:117 Resp: 89894

Ion	Ratio	Lower	Upper
117	100		
119	96.5	77.4	116.0
201	92.8	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 10.797 ng
RT: 8.563 min Scan# 9
Delta R.T. -0.012 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 70 Resp: 167789

Ion	Ratio	Lower	Upper
70	100		
42	46.2	36.0	54.0
101	10.0	8.3	12.5
130	22.3	18.6	28.0







#20
3+4-Methylphenols
Concen: 9.906 ng
RT: 8.534 min Scan# 926
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:107 Resp: 238685

Ion	Ratio	Lower	Upper
107	100		
108	85.5	69.7	109.7
77	30.8	10.3	50.3
79	25.8	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.546 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:136 Resp: 1459823

Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	6.5	5.3	7.9
68	6.3	5.0	7.4







#22
Acetophenone
Concen: 9.536 ng
RT: 8.581 min Scan# 934
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:105 Resp: 357738

Ion	Ratio	Lower	Upper
105	100		
71	1.8	2.4	3.6#
51	20.4	16.2	24.2
120	23.9	19.0	28.6







#23
Nitrobenzene-d5
Concen: 18.856 ng
RT: 8.916 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 493399

Ion	Ratio	Lower	Upper
82	100		
128	45.7	37.4	56.0
54	41.7	32.5	48.7







#24
Nitrobenzene
Concen: 9.536 ng
RT: 8.958 min Scan# 998
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion: 77 Resp: 245585

Ion	Ratio	Lower	Upper
77	100		
123	45.7	39.4	59.2
65	12.0	9.9	14.9







#25
Isophorone
Concen: 9.590 ng
RT: 9.481 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 410357

Ion	Ratio	Lower	Upper
82	100		
95	6.8	5.3	7.9
138	18.6	14.8	22.2







#26
2-Nitrophenol
Concen: 8.692 ng
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:139 Resp: 103123

Ion	Ratio	Lower	Upper
139	100		
109	28.1	21.7	32.5
65	46.4	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 9.265 ng
RT: 9.728 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:122 Resp: 144292

Ion	Ratio	Lower	Upper
122	100		
107	111.4	89.3	133.9
121	57.9	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 9.838 ng
RT: 9.957 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion: 93 Resp: 304956

Ion	Ratio	Lower	Upper
93	100		
95	32.2	25.9	38.9
123	11.3	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 9.046 ng
RT: 10.199 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 192263

Ion	Ratio	Lower	Upper
162	100		
164	63.4	44.2	84.2
98	35.8	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 9.355 ng
RT: 10.410 min Scan# 1245
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:180 Resp: 227800

Ion	Ratio	Lower	Upper
180	100		
182	98.1	76.1	114.1
145	30.0	24.2	36.2







#31
Naphthalene
Concen: 9.642 ng
RT: 10.599 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:128 Resp: 756826

Ion	Ratio	Lower	Upper
128	100		
129	10.7	8.7	13.1
127	12.8	10.4	15.6







#32
Benzoic acid
Concen: 7.587 ng
RT: 9.816 min Scan# 1144
Delta R.T. -0.059 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:122 Resp: 108330

Ion	Ratio	Lower	Upper
122	100		
105	127.1	100.3	140.3
77	94.3	69.1	109.1







#33
4-Chloroaniline
Concen: 9.404 ng
RT: 10.705 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:127 Resp: 260033

Ion	Ratio	Lower	Upper
127	100		
129	32.7	25.8	38.6
65	28.2	21.2	31.8
92	18.3	14.4	21.6







#34
Hexachlorobutadiene
Concen: 9.224 ng
RT: 10.887 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:225 Resp: 130104

Ion	Ratio	Lower	Upper
225	100		
223	60.6	49.9	74.9
227	64.6	52.2	78.2







#35
Caprolactam
Concen: 8.931 ng
RT: 11.481 min Scan# 14
Delta R.T. -0.018 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:113 Resp: 57479

Ion	Ratio	Lower	Upper
113	100		
55	137.0	113.6	153.6
56	105.5	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 9.462 ng
RT: 11.840 min Scan# 1488
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:107 Resp: 206744

Ion	Ratio	Lower	Upper
107	100		
144	25.9	21.6	32.4
142	80.4	65.4	98.0







#37
2-Methylnaphthalene
Concen: 9.564 ng
RT: 12.210 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:142 Resp: 498967

Ion	Ratio	Lower	Upper
142	100		
141	89.2	70.3	105.5
115	32.8	26.0	39.0







#38
1-Methylnaphthalene
Concen: 9.680 ng
RT: 12.434 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:142 Resp: 491304

Ion	Ratio	Lower	Upper
142	100		
141	92.5	73.9	110.9
116	3.7	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.404 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:164 Resp: 878356

Ion	Ratio	Lower	Upper
164	100		
162	100.1	80.6	121.0
160	44.4	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 9.116 ng
RT: 12.581 min Scan# 1614
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:216 Resp: 246009

Ion	Ratio	Lower	Upper
216	100		
214	78.2	62.8	94.2
179	21.0	16.8	25.2
108	18.9	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 8.245 ng
RT: 12.557 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:237 Resp: 82803

Ion	Ratio	Lower	Upper
237	100		
235	61.6	41.8	81.8
272	13.3	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 18.161 ng
RT: 15.916 min Scan# 2181
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:330 Resp: 195393

Ion	Ratio	Lower	Upper
330	100		
332	96.1	77.6	116.4
141	32.6	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 8.649 ng
RT: 12.822 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:196 Resp: 142353

Ion	Ratio	Lower	Upper
196	100		
198	96.3	76.6	115.0
200	31.6	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 8.841 ng
RT: 12.904 min Scan# 1669
Delta R.T. 0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:196 Resp: 159135

Ion	Ratio	Lower	Upper
196	100		
198	96.5	76.5	114.7
97	46.4	38.7	58.1
132	26.9	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 18.984 ng
RT: 13.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:172 Resp: 1192744

Ion	Ratio	Lower	Upper
172	100		
171	34.8	28.3	42.5
170	22.9	19.0	28.4







#46
1,1'-Biphenyl
Concen: 9.494 ng
RT: 13.228 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:154 Resp: 647980

Ion	Ratio	Lower	Upper
154	100		
153	39.9	20.3	60.3
76	13.2	0.0	32.9







#47
2-Chloronaphthalene
Concen: 9.365 ng
RT: 13.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 483199

Ion	Ratio	Lower	Upper
162	100		
127	38.5	31.5	47.3
164	32.2	26.6	39.8







#48
2-Nitroaniline
Concen: 9.079 ng
RT: 13.475 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion: 65 Resp: 110660

Ion	Ratio	Lower	Upper
65	100		
92	75.5	62.1	93.1
138	115.9	99.5	149.3







#49
Acenaphthylene
Concen: 9.405 ng
RT: 14.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 714937

Ion	Ratio	Lower	Upper
152	100		
151	19.4	16.0	24.0
153	13.3	10.6	15.8







#50
Dimethylphthalate
Concen: 9.563 ng
RT: 13.857 min Scan# 1831
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:163 Resp: 592520

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.1	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 9.243 ng
RT: 13.975 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 116661

Ion	Ratio	Lower	Upper
165	100		
63	49.5	37.7	56.5
89	56.4	41.9	62.9







#52
Acenaphthene
Concen: 9.473 ng
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:154 Resp: 463749

Ion	Ratio	Lower	Upper
154	100		
153	113.7	90.5	135.7
152	54.4	43.1	64.7







#53
3-Nitroaniline
Concen: 9.001 ng
RT: 14.304 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:138 Resp: 121229

Ion	Ratio	Lower	Upper
138	100		
108	11.1	8.2	12.2
92	120.0	86.7	130.1







#54
2,4-Dinitrophenol
Concen: 7.053 ng
RT: 14.516 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:184 Resp: 47409

Ion	Ratio	Lower	Upper
184	100		
63	58.9	42.1	63.1
154	57.9	44.6	67.0







#55
Dibenzofuran
Concen: 9.671 ng
RT: 14.810 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:168 Resp: 769059

Ion	Ratio	Lower	Upper
168	100		
139	37.1	29.1	43.7
169	13.4	10.8	16.2







#56
4-Nitrophenol
Concen: 8.246 ng
RT: 14.634 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:139 Resp: 95936

Ion	Ratio	Lower	Upper
139	100		
109	67.0	43.3	83.3
65	90.3	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 8.946 ng
RT: 14.775 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 146673

Ion	Ratio	Lower	Upper
165	100		
63	39.3	29.0	43.6
89	74.6	54.2	81.4
182	2.8	2.3	3.5







#58
Fluorene
Concen: 9.495 ng
RT: 15.475 min Scan# 2106
Delta R.T. 0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:166 Resp: 589720

Ion	Ratio	Lower	Upper
166	100		
165	97.4	77.6	116.4
167	13.3	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 9.071 ng
RT: 15.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:232 Resp: 141666

Ion	Ratio	Lower	Upper
232	100		
131	45.1	35.9	53.9
130	2.4	1.8	2.6
166	30.9	24.2	36.2







#60
Diethylphthalate
Concen: 9.543 ng
RT: 15.240 min Scan# 2066
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:149 Resp: 586933

Ion	Ratio	Lower	Upper
149	100		
177	22.2	17.7	26.5
150	12.7	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 9.478 ng
RT: 15.463 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:204 Resp: 289728

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.4	39.6
141	57.9	45.8	68.8







#62
4-Nitroaniline
Concen: 9.169 ng
RT: 15.487 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:138 Resp: 128152

Ion	Ratio	Lower	Upper
138	100		
92	50.2	28.7	68.7
108	70.1	49.5	89.5







#63
Azobenzene
Concen: 9.535 ng
RT: 15.763 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 549569

Ion	Ratio	Lower	Upper
77	100		
182	25.2	7.7	47.7
105	21.6	2.3	42.3
51	24.2	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.210 min Scan# 2401
Delta R.T. 0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:188 Resp: 1659965

Ion	Ratio	Lower	Upper
188	100		
94	10.1	8.1	12.1
80	11.2	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 7.361 ng
RT: 15.545 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:198 Resp: 70717

Ion	Ratio	Lower	Upper
198	100		
51	39.3	14.7	54.7
105	48.9	23.8	63.8







#66
n-Nitrosodiphenylamine
Concen: 9.352 ng
RT: 15.687 min Scan# 2142
Delta R.T. 0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:169 Resp: 484111

Ion	Ratio	Lower	Upper
169	100		
168	68.3	52.9	79.3
167	36.0	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 9.207 ng
RT: 16.381 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:248 Resp: 171620

Ion	Ratio	Lower	Upper
248	100		
250	97.8	78.4	117.6
141	69.4	59.4	89.0







#68
Hexachlorobenzene
Concen: 9.414 ng
RT: 16.492 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:284 Resp: 206475

Ion	Ratio	Lower	Upper
284	100		
142	33.7	25.4	38.2
249	29.9	23.5	35.3







#69
Atrazine
Concen: 10.706 ng
RT: 16.663 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:200 Resp: 142226

Ion	Ratio	Lower	Upper
200	100		
173	26.8	7.6	47.6
215	48.3	29.0	69.0







#70
Pentachlorophenol
Concen: 8.175 ng
RT: 16.857 min Scan# 2341
Delta R.T. 0.012 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:266 Resp: 116455

Ion	Ratio	Lower	Upper
266	100		
268	64.1	51.3	76.9
264	63.7	50.6	76.0







#71
Phenanthrene
Concen: 9.391 ng
RT: 17.257 min Scan# 24
Delta R.T. 0.012 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:178 Resp: 878834

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.6	23.4
179	15.2	12.5	18.7







#72
Anthracene
Concen: 9.366 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.012 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:178 Resp: 839953

Ion	Ratio	Lower	Upper
178	100		
176	18.2	15.2	22.8
179	15.4	12.6	18.8







#73
Carbazole
Concen: 9.471 ng
RT: 17.634 min Scan# 24
Delta R.T. 0.018 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:167 Resp: 809266

Ion	Ratio	Lower	Upper
167	100		
166	20.8	17.1	25.7
139	12.9	10.2	15.4







#74
Di-n-butylphthalate
Concen: 9.335 ng
RT: 18.228 min Scan# 2574
Delta R.T. 0.018 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:149 Resp: 930693

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.6	11.4
104	4.9	3.9	5.9







#75
Fluoranthene
Concen: 9.566 ng
RT: 19.351 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:202 Resp: 1012600

Ion	Ratio	Lower	Upper
202	100		
101	11.7	0.0	32.5
203	17.2	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.680 min Scan# 3161
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:240 Resp: 1774754

Ion	Ratio	Lower	Upper
240	100		
120	11.7	9.2	13.8
236	25.6	20.4	30.6







#77
Benzidine
Concen: 8.151 ng
RT: 19.545 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:184 Resp: 176282

Ion	Ratio	Lower	Upper
184	100		
185	15.1	11.8	17.6
183	11.4	9.1	13.7







#78
Pyrene
Concen: 9.367 ng
RT: 19.722 min Scan# 2828
Delta R.T. 0.018 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:202 Resp: 1056004

Ion	Ratio	Lower	Upper
202	100		
200	20.6	17.1	25.7
203	17.7	14.2	21.4







#79
Terphenyl-d14
Concen: 19.251 ng
RT: 19.939 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:244 Resp: 1726609

Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.7	8.5
122	11.5	9.6	14.4







#80
Butylbenzylphthalate
Concen: 9.049 ng
RT: 20.675 min Scan# 2990
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:149 Resp: 390890

Ion	Ratio	Lower	Upper
149	100		
91	70.4	55.3	82.9
206	20.0	16.1	24.1







#81
Benzo(a)anthracene
Concen: 9.520 ng
RT: 21.657 min Scan# 3
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 1059397

Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.8	32.8
229	20.0	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 9.409 ng
RT: 21.569 min Scan# 3142
Delta R.T. 0.018 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:252 Resp: 327195

Ion	Ratio	Lower	Upper
252	100		
254	63.8	51.4	77.2
126	12.5	9.3	13.9







#83
Chrysene
Concen: 9.585 ng
RT: 21.727 min Scan# 3
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 1028579

Ion	Ratio	Lower	Upper
228	100		
226	29.3	24.5	36.7
229	19.1	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 9.181 ng
RT: 21.592 min Scan# 3146
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:149 Resp: 601388

Ion	Ratio	Lower	Upper
149	100		
167	27.4	22.5	33.7
279	4.4	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 8.757 ng
RT: 22.904 min Scan# 3
Delta R.T. 0.035 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:149 Resp: 916941

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	7.8	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.074 min Scan# 3738
Delta R.T. 0.018 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:264 Resp: 1895165

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.5	27.7
265	22.1	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 9.061 ng
RT: 28.951 min Scan# 4
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:276 Resp: 1221650

Ion	Ratio	Lower	Upper
276	100		
138	17.3	13.7	20.5
277	19.7	20.2	30.2#







#88
Benzo(b)fluoranthene
Concen: 8.789 ng
RT: 24.004 min Scan# 3556
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:252 Resp: 1057737

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.2	25.8
125	11.1	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 9.190 ng
RT: 24.074 min Scan# 3
Delta R.T. 0.024 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:252 Resp: 1076560

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	10.3	9.1	13.7







#90
Benzo(a)pyrene
Concen: 8.896 ng
RT: 24.927 min Scan# 3713
Delta R.T. 0.035 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:252 Resp: 908542

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.4	26.0
125	11.6	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 9.052 ng
RT: 29.021 min Scan# 4
Delta R.T. 0.012 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 1026140

Ion	Ratio	Lower	Upper
278	100		
139	15.6	13.0	19.4
279	23.7	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 9.121 ng
RT: 30.145 min Scan# 4600
Delta R.T. 0.018 min
Lab File: BP024162.D
Acq: 12 Mar 2025 17:38

Tgt Ion:276 Resp: 1049241

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.2	28.8
138	22.3	17.4	26.2



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024163.D
Acq On : 12 Mar 2025 18:19
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Mar 13 04:07:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.775	152	353514	20.000	ng		0.00
21) Naphthalene-d8	10.545	136	1479701	20.000	ng		0.00
39) Acenaphthene-d10	14.398	164	895871	20.000	ng		0.00
64) Phenanthrene-d10	17.204	188	1698224	20.000	ng		0.00
76) Chrysene-d12	21.657	240	1778727	20.000	ng		0.00
86) Perylene-d12	25.062	264	1914395	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.381	112	924270	40.796	ng		0.00
7) Phenol-d6	6.952	99	1235081	42.415	ng		0.00
23) Nitrobenzene-d5	8.916	82	1081750	40.785	ng		0.00
42) 2,4,6-Tribromophenol	15.922	330	459354	41.861	ng		0.00
45) 2-Fluorobiphenyl	13.016	172	2543735	39.695	ng		0.00
79) Terphenyl-d14	19.927	244	3676388	40.898	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.316	88	183072	20.599	ng		99
3) Pyridine	3.716	79	496891	21.504	ng		99
4) n-Nitrosodimethylamine	3.622	42	168162	20.517	ng		96
6) Aniline	7.104	93	585428	21.953	ng		99
8) 2-Chlorophenol	7.340	128	499424	20.827	ng		99
9) Benzaldehyde	6.922	77	328740	24.728	ng		99
10) Phenol	6.975	94	619974	21.173	ng		100
11) bis(2-Chloroethyl)ether	7.199	93	497357	21.565	ng		100
12) 1,3-Dichlorobenzene	7.663	146	537900	20.418	ng		99
13) 1,4-Dichlorobenzene	7.804	146	552358	20.733	ng		99
14) 1,2-Dichlorobenzene	8.122	146	528080	20.654	ng		99
15) Benzyl Alcohol	8.004	79	387410	21.735	ng		98
16) 2,2'-oxybis(1-Chloropr...	8.287	45	506774	22.636	ng		97
17) 2-Methylphenol	8.210	107	395947	22.190	ng		99
18) Hexachloroethane	8.846	117	193543	20.172	ng		98
19) n-Nitroso-di-n-propyla...	8.569	70	370486	23.663	ng		100
20) 3+4-Methylphenols	8.534	107	541047	22.289	ng		98
22) Acetophenone	8.587	105	786239	20.677	ng	#	100
24) Nitrobenzene	8.963	77	535855	20.529	ng		99
25) Isophorone	9.487	82	933509	21.523	ng		99
26) 2-Nitrophenol	9.669	139	246943	20.534	ng		99
27) 2,4-Dimethylphenol	9.728	122	325916	20.646	ng		100
28) bis(2-Chloroethoxy)met...	9.957	93	648792	20.649	ng		100
29) 2,4-Dichlorophenol	10.204	162	440705	20.457	ng		99
30) 1,2,4-Trichlorobenzene	10.416	180	486293	19.702	ng		99
31) Naphthalene	10.598	128	1625925	20.436	ng		99
32) Benzoic acid	9.845	122	280702	19.396	ng		99
33) 4-Chloroaniline	10.704	127	588052	20.980	ng		99
34) Hexachlorobutadiene	10.887	225	276905	19.369	ng		98
35) Caprolactam	11.487	113	143023	21.924	ng		98
36) 4-Chloro-3-methylphenol	11.839	107	479996	21.674	ng		99
37) 2-Methylnaphthalene	12.210	142	1090957	20.630	ng		98
38) 1-Methylnaphthalene	12.434	142	1077045	20.936	ng		99
40) 1,2,4,5-Tetrachloroben...	12.581	216	533843	19.394	ng		100
41) Hexachlorocyclopentadiene	12.563	237	191514	18.697	ng		99
43) 2,4,6-Trichlorophenol	12.822	196	339357	20.216	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024163.D
Acq On : 12 Mar 2025 18:19
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Mar 13 04:07:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.904	196	370561	20.184	ng		98
46) 1,1'-Biphenyl	13.228	154	1402874	20.152	ng		99
47) 2-Chloronaphthalene	13.269	162	1056861	20.083	ng		99
48) 2-Nitroaniline	13.469	65	268455	21.595	ng		98
49) Acenaphthylene	14.122	152	1590772	20.517	ng		100
50) Dimethylphthalate	13.857	163	1306638	20.677	ng		99
51) 2,6-Dinitrotoluene	13.969	165	271891	21.122	ng		97
52) Acenaphthene	14.469	154	1017503	20.379	ng		99
53) 3-Nitroaniline	14.304	138	293921	21.397	ng		97
54) 2,4-Dinitrophenol	14.516	184	130759	19.074	ng		99
55) Dibenzofuran	14.810	168	1643604	20.264	ng		99
56) 4-Nitrophenol	14.628	139	242843	20.466	ng		99
57) 2,4-Dinitrotoluene	14.769	165	358405	21.433	ng		97
58) Fluorene	15.469	166	1291744	20.392	ng		100
59) 2,3,4,6-Tetrachlorophenol	15.039	232	326758	20.514	ng		99
60) Diethylphthalate	15.233	149	1287503	20.524	ng		99
61) 4-Chlorophenyl-phenyle...	15.469	204	623454	19.996	ng		99
62) 4-Nitroaniline	15.486	138	311347	21.841	ng		99
63) Azobenzene	15.763	77	1226441	20.864	ng		99
65) 4,6-Dinitro-2-methylph...	15.551	198	188346	19.162	ng		97
66) n-Nitrosodiphenylamine	15.680	169	1068735	20.181	ng		100
67) 4-Bromophenyl-phenylether	16.375	248	377534	19.797	ng		99
68) Hexachlorobenzene	16.498	284	447016	19.921	ng		96
69) Atrazine	16.657	200	265940	19.568	ng		99
70) Pentachlorophenol	16.851	266	282344	19.373	ng		99
71) Phenanthrene	17.251	178	1911548	19.965	ng		99
72) Anthracene	17.345	178	1870471	20.386	ng		100
73) Carbazole	17.622	167	1799842	20.589	ng		100
74) Di-n-butylphthalate	18.210	149	2118089	20.765	ng		100
75) Fluoranthene	19.327	202	2186012	20.186	ng		99
77) Benzidine	19.527	184	317361	14.642	ng		99
78) Pyrene	19.704	202	2310992	20.454	ng		100
80) Butylbenzylphthalate	20.657	149	912738	21.082	ng		99
81) Benzo(a)anthracene	21.639	228	2295776	20.583	ng		99
82) 3,3'-Dichlorobenzidine	21.557	252	762665	21.882	ng		100
83) Chrysene	21.704	228	2177924	20.250	ng		99
84) Bis(2-ethylhexyl)phtha...	21.574	149	1395096	21.249	ng		99
85) Di-n-octyl phthalate	22.874	149	2177750	20.752	ng		100
87) Indeno(1,2,3-cd)pyrene	28.939	276	2721144	19.980	ng	#	92
88) Benzo(b)fluoranthene	23.980	252	2407773	19.806	ng		100
89) Benzo(k)fluoranthene	24.051	252	2322149	19.624	ng		99
90) Benzo(a)pyrene	24.892	252	2038106	19.755	ng		99
91) Dibenzo(a,h)anthracene	28.997	278	2267016	19.797	ng		99
92) Benzo(g,h,i)perylene	30.127	276	2294574	19.746	ng		100


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024163.D
Acq On : 12 Mar 2025 18:19
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Mar 13 04:07:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.775 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 353514

Ion	Ratio	Lower	Upper
152	100		
150	153.2	125.5	188.3
115	56.6	46.3	69.5







#2
1,4-Dioxane
Concen: 20.599 ng
RT: 3.316 min Scan# 39
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion: 88 Resp: 183072

Ion	Ratio	Lower	Upper
88	100		
58	62.8	49.8	74.6
43	21.9	17.3	25.9







#3
Pyridine
Concen: 21.504 ng
RT: 3.716 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 496891

Ion	Ratio	Lower	Upper
79	100		
52	54.5	43.8	65.8
51	28.0	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 20.517 ng
RT: 3.622 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion: 42 Resp: 168162

Ion	Ratio	Lower	Upper
42	100		
74	169.5	139.8	209.6
44	7.6	5.3	7.9







#5
2-Fluorophenol
Concen: 40.796 ng
RT: 5.381 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:112 Resp: 924270

Ion	Ratio	Lower	Upper
112	100		
64	52.6	41.3	61.9
63	26.2	20.6	31.0







#6
Aniline
Concen: 21.953 ng
RT: 7.104 min Scan# 683
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion: 93 Resp: 585428

Ion	Ratio	Lower	Upper
93	100		
66	35.0	28.2	42.2
65	18.1	14.0	21.0







#7
Phenol-d6
Concen: 42.415 ng
RT: 6.952 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 99 Resp: 1235081

Ion	Ratio	Lower	Upper
99	100		
42	13.1	10.6	16.0
71	29.2	23.3	34.9







#8
2-Chlorophenol
Concen: 20.827 ng
RT: 7.340 min Scan# 723
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:128 Resp: 499424

Ion	Ratio	Lower	Upper
128	100		
130	33.0	12.0	52.0
64	42.2	21.6	61.6







#9
Benzaldehyde
Concen: 24.728 ng
RT: 6.922 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 328740

Ion	Ratio	Lower	Upper
77	100		
105	102.3	81.4	121.4
106	100.0	81.1	121.1







#10
Phenol
Concen: 21.173 ng
RT: 6.975 min Scan# 661
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion: 94 Resp: 619974

Ion	Ratio	Lower	Upper
94	100		
65	25.0	4.9	44.9
66	31.8	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 21.565 ng
RT: 7.199 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 93 Resp: 497357

Ion	Ratio	Lower	Upper
93	100		
63	67.4	47.4	87.4
95	32.6	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 20.418 ng
RT: 7.663 min Scan# 778
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:146 Resp: 537900

Ion	Ratio	Lower	Upper
146	100		
148	65.0	51.6	77.4
75	28.1	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 20.733 ng
RT: 7.804 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:146 Resp: 552358

Ion	Ratio	Lower	Upper
146	100		
148	63.6	51.9	77.9
111	39.5	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 20.654 ng
RT: 8.122 min Scan# 856
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:146 Resp: 528080

Ion	Ratio	Lower	Upper
146	100		
148	63.4	51.4	77.2
111	41.0	33.2	49.8







#15
Benzyl Alcohol
Concen: 21.735 ng
RT: 8.004 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 387410

Ion	Ratio	Lower	Upper
79	100		
108	81.2	66.5	99.7
77	64.5	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 22.636 ng
RT: 8.287 min Scan# 884
Delta R.T. -0.018 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion: 45 Resp: 506774

Ion	Ratio	Lower	Upper
45	100		
77	18.2	0.0	39.0
79	14.1	0.0	35.6







#17
2-Methylphenol
Concen: 22.190 ng
RT: 8.210 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:107 Resp: 395947

Ion	Ratio	Lower	Upper
107	100		
108	111.2	90.0	135.0
77	42.6	34.0	51.0
79	42.3	32.7	49.1







#18
Hexachloroethane
Concen: 20.172 ng
RT: 8.846 min Scan# 979
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:117 Resp: 193543

Ion	Ratio	Lower	Upper
117	100		
119	95.9	77.4	116.0
201	98.5	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 23.663 ng
RT: 8.569 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 70 Resp: 370486

Ion	Ratio	Lower	Upper
70	100		
42	45.3	36.0	54.0
101	10.5	8.3	12.5
130	23.0	18.6	28.0







#20
3+4-Methylphenols
Concen: 22.289 ng
RT: 8.534 min Scan# 926
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:107 Resp: 541047

Ion	Ratio	Lower	Upper
107	100		
108	87.6	69.7	109.7
77	29.6	10.3	50.3
79	25.3	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.545 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:136 Resp: 1479701

Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.7	5.3	7.9
68	6.3	5.0	7.4







#22
Acetophenone
Concen: 20.677 ng
RT: 8.587 min Scan# 935
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:105 Resp: 786239

Ion	Ratio	Lower	Upper
105	100		
71	1.3	2.4	3.6#
51	20.2	16.2	24.2
120	23.8	19.0	28.6







#23
Nitrobenzene-d5
Concen: 40.785 ng
RT: 8.916 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 1081750

Ion	Ratio	Lower	Upper
82	100		
128	45.3	37.4	56.0
54	40.3	32.5	48.7







#24
Nitrobenzene
Concen: 20.529 ng
RT: 8.963 min Scan# 999
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion: 77 Resp: 535855

Ion	Ratio	Lower	Upper
77	100		
123	48.4	39.4	59.2
65	12.3	9.9	14.9







#25
Isophorone
Concen: 21.523 ng
RT: 9.487 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 933509

Ion	Ratio	Lower	Upper
82	100		
95	6.7	5.3	7.9
138	18.8	14.8	22.2







#26
2-Nitrophenol
Concen: 20.534 ng
RT: 9.669 min Scan# 1119
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:139 Resp: 246943

Ion	Ratio	Lower	Upper
139	100		
109	26.2	21.7	32.5
65	44.2	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 20.646 ng
RT: 9.728 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:122 Resp: 325916

Ion	Ratio	Lower	Upper
122	100		
107	111.6	89.3	133.9
121	58.5	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 20.649 ng
RT: 9.957 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion: 93 Resp: 648792

Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.9	38.9
123	11.8	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 20.457 ng
RT: 10.204 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 440705

Ion	Ratio	Lower	Upper
162	100		
164	64.9	44.2	84.2
98	36.2	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 19.702 ng
RT: 10.416 min Scan# 1246
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:180 Resp: 486293

Ion	Ratio	Lower	Upper
180	100		
182	94.3	76.1	114.1
145	28.7	24.2	36.2







#31
Naphthalene
Concen: 20.436 ng
RT: 10.598 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:128 Resp: 1625925

Ion	Ratio	Lower	Upper
128	100		
129	10.6	8.7	13.1
127	12.7	10.4	15.6







#32
Benzoic acid
Concen: 19.396 ng
RT: 9.845 min Scan# 1149
Delta R.T. -0.030 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:122 Resp: 280702

Ion	Ratio	Lower	Upper
122	100		
105	119.4	100.3	140.3
77	87.5	69.1	109.1







#33
4-Chloroaniline
Concen: 20.980 ng
RT: 10.704 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:127 Resp: 588052

Ion	Ratio	Lower	Upper
127	100		
129	31.9	25.8	38.6
65	27.0	21.2	31.8
92	17.9	14.4	21.6







#34
Hexachlorobutadiene
Concen: 19.369 ng
RT: 10.887 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:225 Resp: 276905

Ion	Ratio	Lower	Upper
225	100		
223	63.9	49.9	74.9
227	67.1	52.2	78.2







#35
Caprolactam
Concen: 21.924 ng
RT: 11.487 min Scan# 14
Delta R.T. -0.012 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:113 Resp: 143023

Ion	Ratio	Lower	Upper
113	100		
55	134.9	113.6	153.6
56	102.9	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 21.674 ng
RT: 11.839 min Scan# 1488
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:107 Resp: 479996

Ion	Ratio	Lower	Upper
107	100		
144	26.8	21.6	32.4
142	82.7	65.4	98.0







#37
2-Methylnaphthalene
Concen: 20.630 ng
RT: 12.210 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:142 Resp: 1090957

Ion	Ratio	Lower	Upper
142	100		
141	90.2	70.3	105.5
115	33.1	26.0	39.0







#38
1-Methylnaphthalene
Concen: 20.936 ng
RT: 12.434 min Scan# 1589
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:142 Resp: 1077045

Ion	Ratio	Lower	Upper
142	100		
141	91.7	73.9	110.9
116	3.6	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:164 Resp: 895871

Ion	Ratio	Lower	Upper
164	100		
162	100.1	80.6	121.0
160	45.2	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 19.394 ng
RT: 12.581 min Scan# 1614
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:216 Resp: 533843

Ion	Ratio	Lower	Upper
216	100		
214	78.2	62.8	94.2
179	21.0	16.8	25.2
108	18.7	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 18.697 ng
RT: 12.563 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:237 Resp: 191514

Ion	Ratio	Lower	Upper
237	100		
235	60.5	41.8	81.8
272	13.3	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 41.861 ng
RT: 15.922 min Scan# 2182
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:330 Resp: 459354

Ion	Ratio	Lower	Upper
330	100		
332	97.6	77.6	116.4
141	32.3	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 20.216 ng
RT: 12.822 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:196 Resp: 339357

Ion	Ratio	Lower	Upper
196	100		
198	96.1	76.6	115.0
200	30.7	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 20.184 ng
RT: 12.904 min Scan# 1669
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:196 Resp: 370561

Ion	Ratio	Lower	Upper
196	100		
198	97.0	76.5	114.7
97	45.4	38.7	58.1
132	27.5	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 39.695 ng
RT: 13.016 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:172 Resp: 2543735

Ion	Ratio	Lower	Upper
172	100		
171	34.7	28.3	42.5
170	23.4	19.0	28.4







#46
1,1'-Biphenyl
Concen: 20.152 ng
RT: 13.228 min Scan# 1724
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:154 Resp: 1402874

Ion	Ratio	Lower	Upper
154	100		
153	39.7	20.3	60.3
76	13.0	0.0	32.9







#47
2-Chloronaphthalene
Concen: 20.083 ng
RT: 13.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 1056861

Ion	Ratio	Lower	Upper
162	100		
127	38.2	31.5	47.3
164	33.0	26.6	39.8







#48
2-Nitroaniline
Concen: 21.595 ng
RT: 13.469 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion: 65 Resp: 268455

Ion	Ratio	Lower	Upper
65	100		
92	76.8	62.1	93.1
138	122.2	99.5	149.3







#49
Acenaphthylene
Concen: 20.517 ng
RT: 14.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 1590772

Ion	Ratio	Lower	Upper
152	100		
151	19.8	16.0	24.0
153	13.1	10.6	15.8







#50
Dimethylphthalate
Concen: 20.677 ng
RT: 13.857 min Scan# 1831
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:163 Resp: 1306638

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.0
164	10.0	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 21.122 ng
RT: 13.969 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 271891

Ion	Ratio	Lower	Upper
165	100		
63	49.2	37.7	56.5
89	54.7	41.9	62.9







#52
Acenaphthene
Concen: 20.379 ng
RT: 14.469 min Scan# 1935
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:154 Resp: 1017503

Ion	Ratio	Lower	Upper
154	100		
153	111.9	90.5	135.7
152	53.5	43.1	64.7







#53
3-Nitroaniline
Concen: 21.397 ng
RT: 14.304 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:138 Resp: 293921

Ion	Ratio	Lower	Upper
138	100		
108	10.2	8.2	12.2
92	111.4	86.7	130.1







#54
2,4-Dinitrophenol
Concen: 19.074 ng
RT: 14.516 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:184 Resp: 130759

Ion	Ratio	Lower	Upper
184	100		
63	54.4	42.1	63.1
154	56.0	44.6	67.0







#55
Dibenzofuran
Concen: 20.264 ng
RT: 14.810 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:168 Resp: 1643604

Ion	Ratio	Lower	Upper
168	100		
139	37.0	29.1	43.7
169	13.2	10.8	16.2







#56
4-Nitrophenol
Concen: 20.466 ng
RT: 14.628 min Scan# 1962
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:139 Resp: 242843

Ion	Ratio	Lower	Upper
139	100		
109	64.1	43.3	83.3
65	82.7	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 21.433 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 358405

Ion	Ratio	Lower	Upper
165	100		
63	38.7	29.0	43.6
89	69.6	54.2	81.4
182	2.9	2.3	3.5







#58
Fluorene
Concen: 20.392 ng
RT: 15.469 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:166 Resp: 1291744

Ion	Ratio	Lower	Upper
166	100		
165	97.1	77.6	116.4
167	13.5	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 20.514 ng
RT: 15.039 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:232 Resp: 326758

Ion	Ratio	Lower	Upper
232	100		
131	45.3	35.9	53.9
130	2.3	1.8	2.6
166	31.1	24.2	36.2







#60
Diethylphthalate
Concen: 20.524 ng
RT: 15.233 min Scan# 2065
Delta R.T. -0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:149 Resp: 1287503

Ion	Ratio	Lower	Upper
149	100		
177	21.5	17.7	26.5
150	12.6	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 19.996 ng
RT: 15.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:204 Resp: 623454

Ion	Ratio	Lower	Upper
204	100		
206	32.8	26.4	39.6
141	55.9	45.8	68.8







#62
4-Nitroaniline
Concen: 21.841 ng
RT: 15.486 min Scan# 2108
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:138 Resp: 311347

Ion	Ratio	Lower	Upper
138	100		
92	49.3	28.7	68.7
108	68.2	49.5	89.5







#63
Azobenzene
Concen: 20.864 ng
RT: 15.763 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 1226441

Ion	Ratio	Lower	Upper
77	100		
182	26.9	7.7	47.7
105	21.9	2.3	42.3
51	23.5	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.204 min Scan# 2400
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:188 Resp: 1698224

Ion	Ratio	Lower	Upper
188	100		
94	10.1	8.1	12.1
80	11.1	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 19.162 ng
RT: 15.551 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:198 Resp: 188346

Ion	Ratio	Lower	Upper
198	100		
51	32.4	14.7	54.7
105	42.4	23.8	63.8







#66
n-Nitrosodiphenylamine
Concen: 20.181 ng
RT: 15.680 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:169 Resp: 1068735

Ion	Ratio	Lower	Upper
169	100		
168	65.9	52.9	79.3
167	36.2	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 19.797 ng
RT: 16.375 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:248 Resp: 377534

Ion	Ratio	Lower	Upper
248	100		
250	97.5	78.4	117.6
141	74.9	59.4	89.0







#68
Hexachlorobenzene
Concen: 19.921 ng
RT: 16.498 min Scan# 2280
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:284 Resp: 447016

Ion	Ratio	Lower	Upper
284	100		
142	34.8	25.4	38.2
249	30.5	23.5	35.3







#69
Atrazine
Concen: 19.568 ng
RT: 16.657 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:200 Resp: 265940

Ion	Ratio	Lower	Upper
200	100		
173	27.6	7.6	47.6
215	49.8	29.0	69.0







#70
Pentachlorophenol
Concen: 19.373 ng
RT: 16.851 min Scan# 2340
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:266 Resp: 282344

Ion	Ratio	Lower	Upper
266	100		
268	63.7	51.3	76.9
264	62.7	50.6	76.0







#71
Phenanthrene
Concen: 19.965 ng
RT: 17.251 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:178 Resp: 1911548

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.6	23.4
179	15.2	12.5	18.7







#72
Anthracene
Concen: 20.386 ng
RT: 17.345 min Scan# 2424
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:178 Resp: 1870471

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.2	22.8
179	15.5	12.6	18.8







#73
Carbazole
Concen: 20.589 ng
RT: 17.622 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:167 Resp: 1799842

Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.1	25.7
139	12.8	10.2	15.4







#74
Di-n-butylphthalate
Concen: 20.765 ng
RT: 18.210 min Scan# 2571
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:149 Resp: 2118089

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.6	11.4
104	5.0	3.9	5.9







#75
Fluoranthene
Concen: 20.186 ng
RT: 19.327 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:202 Resp: 2186012

Ion	Ratio	Lower	Upper
202	100		
101	13.1	0.0	32.5
203	17.5	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.657 min Scan# 3157
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:240 Resp: 1778727

Ion	Ratio	Lower	Upper
240	100		
120	12.1	9.2	13.8
236	25.9	20.4	30.6







#77
Benzidine
Concen: 14.642 ng
RT: 19.527 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:184 Resp: 317361

Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.8	17.6
183	11.4	9.1	13.7







#78
Pyrene
Concen: 20.454 ng
RT: 19.704 min Scan# 2825
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:202 Resp: 2310992

Ion	Ratio	Lower	Upper
202	100		
200	21.1	17.1	25.7
203	17.6	14.2	21.4







#79
Terphenyl-d14
Concen: 40.898 ng
RT: 19.927 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:244 Resp: 3676388

Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.7	8.5
122	10.7	9.6	14.4







#80
Butylbenzylphthalate
Concen: 21.082 ng
RT: 20.657 min Scan# 2987
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:149 Resp: 912738

Ion	Ratio	Lower	Upper
149	100		
91	69.4	55.3	82.9
206	20.9	16.1	24.1







#81
Benzo(a)anthracene
Concen: 20.583 ng
RT: 21.639 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 2295776

Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.8	32.8
229	19.6	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 21.882 ng
RT: 21.557 min Scan# 3140
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:252 Resp: 762665

Ion	Ratio	Lower	Upper
252	100		
254	64.1	51.4	77.2
126	11.4	9.3	13.9







#83
Chrysene
Concen: 20.250 ng
RT: 21.704 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 2177924

Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.5	36.7
229	19.8	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 21.249 ng
RT: 21.574 min Scan# 3143
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:149 Resp: 1395096

Ion	Ratio	Lower	Upper
149	100		
167	27.7	22.5	33.7
279	4.5	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 20.752 ng
RT: 22.874 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:149 Resp: 2177750

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	7.5	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.062 min Scan# 3736
Delta R.T. 0.006 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:264 Resp: 1914395

Ion	Ratio	Lower	Upper
264	100		
260	23.5	18.5	27.7
265	22.0	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 19.980 ng
RT: 28.939 min Scan# 4
Delta R.T. 0.012 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:276 Resp: 2721144

Ion	Ratio	Lower	Upper
276	100		
138	16.2	13.7	20.5
277	19.5	20.2	30.2#







#88
Benzo(b)fluoranthene
Concen: 19.806 ng
RT: 23.980 min Scan# 3552
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:252 Resp: 2407773

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.2	25.8
125	10.6	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 19.624 ng
RT: 24.051 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:252 Resp: 2322149

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	10.7	9.1	13.7







#90
Benzo(a)pyrene
Concen: 19.755 ng
RT: 24.892 min Scan# 3707
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:252 Resp: 2038106

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	12.5	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 19.797 ng
RT: 28.997 min Scan# 4
Delta R.T. -0.012 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:278 Resp: 2267016

Ion	Ratio	Lower	Upper
278	100		
139	16.8	13.0	19.4
279	23.8	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 19.746 ng
RT: 30.127 min Scan# 4597
Delta R.T. -0.000 min
Lab File: BP024163.D
Acq: 12 Mar 2025 18:19

Tgt Ion:276 Resp: 2294574

Ion	Ratio	Lower	Upper
276	100		
277	24.3	19.2	28.8
138	21.9	17.4	26.2



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024164.D
Acq On : 12 Mar 2025 19:00
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Mar 13 04:08:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	383350	20.000	ng	0.00
21) Naphthalene-d8	10.552	136	1538579	20.000	ng	0.00
39) Acenaphthene-d10	14.404	164	930584	20.000	ng	0.00
64) Phenanthrene-d10	17.204	188	1720919	20.000	ng	0.00
76) Chrysene-d12	21.657	240	1881172	20.000	ng	0.00
86) Perylene-d12	25.057	264	2043442	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	1856351	75.560	ng	0.00
7) Phenol-d6	6.952	99	2501405	79.217	ng	0.00
23) Nitrobenzene-d5	8.922	82	2157744	78.240	ng	0.00
42) 2,4,6-Tribromophenol	15.916	330	938209	82.309	ng	0.00
45) 2-Fluorobiphenyl	13.016	172	4911554	73.785	ng	0.00
79) Terphenyl-d14	19.922	244	7133778	75.038	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.323	88	359722	37.325	ng	100
3) Pyridine	3.717	79	1004903	40.105	ng	100
4) n-Nitrosodimethylamine	3.623	42	330244	37.156	ng	100
6) Aniline	7.111	93	1138144	39.358	ng	100
8) 2-Chlorophenol	7.346	128	1001335	38.507	ng	100
9) Benzaldehyde	6.922	77	597740	41.462	ng	100
10) Phenol	6.975	94	1248029	39.305	ng	100
11) bis(2-Chloroethyl)ether	7.199	93	976195	39.033	ng	100
12) 1,3-Dichlorobenzene	7.664	146	1063428	37.225	ng	100
13) 1,4-Dichlorobenzene	7.811	146	1083221	37.495	ng	100
14) 1,2-Dichlorobenzene	8.122	146	1041511	37.564	ng	100
15) Benzyl Alcohol	8.011	79	798656	41.320	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.305	45	991099	40.824	ng	100
17) 2-Methylphenol	8.216	107	795971	41.136	ng	100
18) Hexachloroethane	8.846	117	389110	37.399	ng	100
19) n-Nitroso-di-n-propyla...	8.575	70	737233	43.423	ng	100
20) 3+4-Methylphenols	8.540	107	1098375	41.727	ng	100
22) Acetophenone	8.587	105	1525122	38.574	ng	100
24) Nitrobenzene	8.963	77	1049347	38.662	ng	100
25) Isophorone	9.487	82	1853580	41.100	ng	100
26) 2-Nitrophenol	9.669	139	525614	42.034	ng	100
27) 2,4-Dimethylphenol	9.734	122	661755	40.316	ng	100
28) bis(2-Chloroethoxy)met...	9.963	93	1285698	39.353	ng	100
29) 2,4-Dichlorophenol	10.205	162	912137	40.720	ng	100
30) 1,2,4-Trichlorobenzene	10.416	180	960872	37.439	ng	100
31) Naphthalene	10.605	128	3167478	38.287	ng	100
32) Benzoic acid	9.875	122	670600	44.564	ng	100
33) 4-Chloroaniline	10.710	127	1159633	39.789	ng	100
34) Hexachlorobutadiene	10.893	225	553609	37.241	ng	100
35) Caprolactam	11.499	113	299017	44.083	ng	100
36) 4-Chloro-3-methylphenol	11.840	107	966642	41.977	ng	100
37) 2-Methylnaphthalene	12.210	142	2167502	39.419	ng	100
38) 1-Methylnaphthalene	12.434	142	2071946	38.734	ng	100
40) 1,2,4,5-Tetrachloroben...	12.587	216	1039275	36.348	ng	100
41) Hexachlorocyclopentadiene	12.563	237	395145	37.139	ng	100
43) 2,4,6-Trichlorophenol	12.828	196	692810	39.731	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024164.D
Acq On : 12 Mar 2025 19:00
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Mar 13 04:08:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.899	196	760360	39.872	ng	100
46) 1,1'-Biphenyl	13.228	154	2720465	37.621	ng	100
47) 2-Chloronaphthalene	13.269	162	2037284	37.269	ng	100
48) 2-Nitroaniline	13.475	65	543429	42.085	ng	100
49) Acenaphthylene	14.128	152	3091813	38.389	ng	100
50) Dimethylphthalate	13.857	163	2526732	38.493	ng	100
51) 2,6-Dinitrotoluene	13.975	165	546912	40.901	ng	100
52) Acenaphthene	14.469	154	1973969	38.061	ng	100
53) 3-Nitroaniline	14.310	138	599822	42.038	ng	100
54) 2,4-Dinitrophenol	14.522	184	297379	41.760	ng	100
55) Dibenzofuran	14.810	168	3184940	37.802	ng	100
56) 4-Nitrophenol	14.634	139	515335	41.811	ng	100
57) 2,4-Dinitrotoluene	14.775	165	732669	42.180	ng	100
58) Fluorene	15.469	166	2502924	38.039	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.040	232	662947	40.068	ng	100
60) Diethylphthalate	15.240	149	2527904	38.794	ng	100
61) 4-Chlorophenyl-phenyle...	15.463	204	1225168	37.829	ng	100
62) 4-Nitroaniline	15.487	138	638694	43.134	ng	100
63) Azobenzene	15.763	77	2401881	39.335	ng	100
65) 4,6-Dinitro-2-methylph...	15.551	198	412313	41.395	ng	100
66) n-Nitrosodiphenylamine	15.681	169	2063452	38.450	ng	100
67) 4-Bromophenyl-phenylether	16.375	248	745376	38.570	ng	100
68) Hexachlorobenzene	16.492	284	873989	38.435	ng	100
69) Atrazine	16.657	200	466258	33.855	ng	100
70) Pentachlorophenol	16.845	266	610729	41.352	ng	100
71) Phenanthrene	17.245	178	3727050	38.414	ng	100
72) Anthracene	17.339	178	3658135	39.344	ng	100
73) Carbazole	17.616	167	3584494	40.463	ng	100
74) Di-n-butylphthalate	18.210	149	4336969	41.958	ng	100
75) Fluoranthene	19.328	202	4349734	39.636	ng	100
77) Benzidine	19.522	184	968916	42.268	ng	100
78) Pyrene	19.704	202	4529960	37.911	ng	100
80) Butylbenzylphthalate	20.651	149	1972887	43.087	ng	100
81) Benzo(a)anthracene	21.633	228	4546432	38.542	ng	100
82) 3,3'-Dichlorobenzidine	21.551	252	1610565	43.694	ng	100
83) Chrysene	21.704	228	4356257	38.298	ng	100
84) Bis(2-ethylhexyl)phtha...	21.569	149	2980735	42.929	ng	100
85) Di-n-octyl phthalate	22.869	149	4786749	43.129	ng	100
87) Indeno(1,2,3-cd)pyrene	28.927	276	5583803	38.410	ng	100
88) Benzo(b)fluoranthene	23.980	252	4864855	37.490	ng	100
89) Benzo(k)fluoranthene	24.051	252	4618537	36.566	ng	100
90) Benzo(a)pyrene	24.892	252	4204610	38.180	ng	100
91) Dibenzo(a,h)anthracene	29.009	278	4646524	38.014	ng	100
92) Benzo(g,h,i)perylene	30.127	276	4714749	38.011	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024164.D
Acq On : 12 Mar 2025 19:00
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Mar 13 04:08:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 383350

Ion	Ratio	Lower	Upper
152	100		
150	156.9	125.5	188.3
115	57.9	46.3	69.5







#2
1,4-Dioxane
Concen: 37.325 ng
RT: 3.323 min Scan# 40
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion: 88 Resp: 359722

Ion	Ratio	Lower	Upper
88	100		
58	62.2	49.8	74.6
43	21.6	17.3	25.9







#3
Pyridine
Concen: 40.105 ng
RT: 3.717 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 1004903

Ion	Ratio	Lower	Upper
79	100		
52	54.8	43.8	65.8
51	27.0	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 37.156 ng
RT: 3.623 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion: 42 Resp: 330244

Ion	Ratio	Lower	Upper
42	100		
74	174.7	139.8	209.6
44	6.6	5.3	7.9







#5
2-Fluorophenol
Concen: 75.560 ng
RT: 5.381 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:112 Resp: 1856351

Ion	Ratio	Lower	Upper
112	100		
64	51.6	41.3	61.9
63	25.8	20.6	31.0







#6
Aniline
Concen: 39.358 ng
RT: 7.111 min Scan# 684
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion: 93 Resp: 1138144

Ion	Ratio	Lower	Upper
93	100		
66	35.2	28.2	42.2
65	17.5	14.0	21.0







#7
Phenol-d6
Concen: 79.217 ng
RT: 6.952 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 99 Resp: 2501405

Ion	Ratio	Lower	Upper
99	100		
42	13.3	10.6	16.0
71	29.1	23.3	34.9







#8
2-Chlorophenol
Concen: 38.507 ng
RT: 7.346 min Scan# 724
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:128 Resp: 1001335

Ion	Ratio	Lower	Upper
128	100		
130	32.0	12.0	52.0
64	41.6	21.6	61.6







#9
Benzaldehyde
Concen: 41.462 ng
RT: 6.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 597740

Ion	Ratio	Lower	Upper
77	100		
105	101.4	81.4	121.4
106	101.1	81.1	121.1







#10
Phenol
Concen: 39.305 ng
RT: 6.975 min Scan# 661
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion: 94 Resp: 1248029

Ion	Ratio	Lower	Upper
94	100		
65	24.9	4.9	44.9
66	31.9	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 39.033 ng
RT: 7.199 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 93 Resp: 976195

Ion	Ratio	Lower	Upper
93	100		
63	67.4	47.4	87.4
95	32.1	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 37.225 ng
RT: 7.664 min Scan# 778
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:146 Resp: 1063428

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.6	77.4
75	28.5	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 37.495 ng
RT: 7.811 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:146 Resp: 1083221

Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.9	77.9
111	39.3	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 37.564 ng
RT: 8.122 min Scan# 856
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:146 Resp: 1041511

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.2
111	41.5	33.2	49.8







#15
Benzyl Alcohol
Concen: 41.320 ng
RT: 8.011 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 798656

Ion	Ratio	Lower	Upper
79	100		
108	83.1	66.5	99.7
77	63.6	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.824 ng
RT: 8.305 min Scan# 887
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion: 45 Resp: 991099

Ion	Ratio	Lower	Upper
45	100		
77	19.0	0.0	39.0
79	15.6	0.0	35.6







#17
2-Methylphenol
Concen: 41.136 ng
RT: 8.216 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:107 Resp: 795971

Ion	Ratio	Lower	Upper
107	100		
108	112.5	90.0	135.0
77	42.5	34.0	51.0
79	40.9	32.7	49.1







#18
Hexachloroethane
Concen: 37.399 ng
RT: 8.846 min Scan# 979
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:117 Resp: 389110

Ion	Ratio	Lower	Upper
117	100		
119	96.7	77.4	116.0
201	95.7	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 43.423 ng
RT: 8.575 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 70 Resp: 737233

Ion	Ratio	Lower	Upper
70	100		
42	45.0	36.0	54.0
101	10.4	8.3	12.5
130	23.3	18.6	28.0







#20
3+4-Methylphenols
Concen: 41.727 ng
RT: 8.540 min Scan# 927
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:107 Resp: 1098375

Ion	Ratio	Lower	Upper
107	100		
108	89.7	69.7	109.7
77	30.3	10.3	50.3
79	26.1	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.552 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:136 Resp: 1538579

Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.6	5.3	7.9
68	6.2	5.0	7.4







#22
Acetophenone
Concen: 38.574 ng
RT: 8.587 min Scan# 935
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:105 Resp: 1525122

Ion	Ratio	Lower	Upper
105	100		
71	3.0	2.4	3.6
51	20.2	16.2	24.2
120	23.8	19.0	28.6







#23
Nitrobenzene-d5
Concen: 78.240 ng
RT: 8.922 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 2157744

Ion	Ratio	Lower	Upper
82	100		
128	46.7	37.4	56.0
54	40.6	32.5	48.7







#24
Nitrobenzene
Concen: 38.662 ng
RT: 8.963 min Scan# 999
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion: 77 Resp: 1049347

Ion	Ratio	Lower	Upper
77	100		
123	49.3	39.4	59.2
65	12.4	9.9	14.9







#25
Isophorone
Concen: 41.100 ng
RT: 9.487 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1853580

Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.3	7.9
138	18.5	14.8	22.2







#26
2-Nitrophenol
Concen: 42.034 ng
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:139 Resp: 525614

Ion	Ratio	Lower	Upper
139	100		
109	27.1	21.7	32.5
65	44.8	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 40.316 ng
RT: 9.734 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:122 Resp: 661755

Ion	Ratio	Lower	Upper
122	100		
107	111.6	89.3	133.9
121	59.4	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 39.353 ng
RT: 9.963 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion: 93 Resp: 1285698

Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.9	38.9
123	11.9	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 40.720 ng
RT: 10.205 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 912137

Ion	Ratio	Lower	Upper
162	100		
164	64.2	44.2	84.2
98	35.1	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 37.439 ng
RT: 10.416 min Scan# 1246
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:180 Resp: 960872

Ion	Ratio	Lower	Upper
180	100		
182	95.1	76.1	114.1
145	30.2	24.2	36.2







#31
Naphthalene
Concen: 38.287 ng
RT: 10.605 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:128 Resp: 3167478

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.7	13.1
127	13.0	10.4	15.6







#32
Benzoic acid
Concen: 44.564 ng
RT: 9.875 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:122 Resp: 670600

Ion	Ratio	Lower	Upper
122	100		
105	120.3	100.3	140.3
77	89.1	69.1	109.1







#33
4-Chloroaniline
Concen: 39.789 ng
RT: 10.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:127 Resp: 1159633

Ion	Ratio	Lower	Upper
127	100		
129	32.2	25.8	38.6
65	26.5	21.2	31.8
92	18.0	14.4	21.6







#34
Hexachlorobutadiene
Concen: 37.241 ng
RT: 10.893 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:225 Resp: 553609

Ion	Ratio	Lower	Upper
225	100		
223	62.4	49.9	74.9
227	65.2	52.2	78.2







#35
Caprolactam
Concen: 44.083 ng
RT: 11.499 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:113 Resp: 299017

Ion	Ratio	Lower	Upper
113	100		
55	133.6	113.6	153.6
56	100.0	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 41.977 ng
RT: 11.840 min Scan# 1488
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:107 Resp: 966642

Ion	Ratio	Lower	Upper
107	100		
144	27.0	21.6	32.4
142	81.7	65.4	98.0







#37
2-Methylnaphthalene
Concen: 39.419 ng
RT: 12.210 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:142 Resp: 2167502

Ion	Ratio	Lower	Upper
142	100		
141	87.9	70.3	105.5
115	32.5	26.0	39.0







#38
1-Methylnaphthalene
Concen: 38.734 ng
RT: 12.434 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:142 Resp: 2071946

Ion	Ratio	Lower	Upper
142	100		
141	92.4	73.9	110.9
116	3.5	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.404 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:164 Resp: 930584

Ion	Ratio	Lower	Upper
164	100		
162	100.8	80.6	121.0
160	44.3	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 36.348 ng
RT: 12.587 min Scan# 1615
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:216 Resp: 1039275

Ion	Ratio	Lower	Upper
216	100		
214	78.5	62.8	94.2
179	21.0	16.8	25.2
108	18.4	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 37.139 ng
RT: 12.563 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:237 Resp: 395145

Ion	Ratio	Lower	Upper
237	100		
235	61.8	41.8	81.8
272	13.5	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 82.309 ng
RT: 15.916 min Scan# 2181
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:330 Resp: 938209

Ion	Ratio	Lower	Upper
330	100		
332	97.0	77.6	116.4
141	31.3	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 39.731 ng
RT: 12.828 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:196 Resp: 692810

Ion	Ratio	Lower	Upper
196	100		
198	95.8	76.6	115.0
200	31.7	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 39.872 ng
RT: 12.899 min Scan# 1668
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:196 Resp: 760360

Ion	Ratio	Lower	Upper
196	100		
198	95.6	76.5	114.7
97	48.4	38.7	58.1
132	27.7	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 73.785 ng
RT: 13.016 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:172 Resp: 4911554

Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.3	42.5
170	23.7	19.0	28.4







#46
1,1'-Biphenyl
Concen: 37.621 ng
RT: 13.228 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:154 Resp: 2720465

Ion	Ratio	Lower	Upper
154	100		
153	40.3	20.3	60.3
76	12.9	0.0	32.9







#47
2-Chloronaphthalene
Concen: 37.269 ng
RT: 13.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 2037284

Ion	Ratio	Lower	Upper
162	100		
127	39.4	31.5	47.3
164	33.2	26.6	39.8







#48
2-Nitroaniline
Concen: 42.085 ng
RT: 13.475 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion: 65 Resp: 543429

Ion	Ratio	Lower	Upper
65	100		
92	77.6	62.1	93.1
138	124.4	99.5	149.3







#49
Acenaphthylene
Concen: 38.389 ng
RT: 14.128 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 3091813

Ion	Ratio	Lower	Upper
152	100		
151	20.0	16.0	24.0
153	13.2	10.6	15.8







#50
Dimethylphthalate
Concen: 38.493 ng
RT: 13.857 min Scan# 1831
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:163 Resp: 2526732

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.3	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 40.901 ng
RT: 13.975 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 546912

Ion	Ratio	Lower	Upper
165	100		
63	47.1	37.7	56.5
89	52.4	41.9	62.9







#52
Acenaphthene
Concen: 38.061 ng
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:154 Resp: 1973969

Ion	Ratio	Lower	Upper
154	100		
153	113.1	90.5	135.7
152	53.9	43.1	64.7







#53
3-Nitroaniline
Concen: 42.038 ng
RT: 14.310 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:138 Resp: 599822

Ion	Ratio	Lower	Upper
138	100		
108	10.2	8.2	12.2
92	108.4	86.7	130.1







#54
2,4-Dinitrophenol
Concen: 41.760 ng
RT: 14.522 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:184 Resp: 297379

Ion	Ratio	Lower	Upper
184	100		
63	52.6	42.1	63.1
154	55.8	44.6	67.0







#55
Dibenzofuran
Concen: 37.802 ng
RT: 14.810 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:168 Resp: 3184940

Ion	Ratio	Lower	Upper
168	100		
139	36.4	29.1	43.7
169	13.5	10.8	16.2







#56
4-Nitrophenol
Concen: 41.811 ng
RT: 14.634 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:139 Resp: 515335

Ion	Ratio	Lower	Upper
139	100		
109	63.3	43.3	83.3
65	81.2	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 42.180 ng
RT: 14.775 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 732669

Ion	Ratio	Lower	Upper
165	100		
63	36.3	29.0	43.6
89	67.8	54.2	81.4
182	2.9	2.3	3.5







#58
Fluorene
Concen: 38.039 ng
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:166 Resp: 2502924

Ion	Ratio	Lower	Upper
166	100		
165	97.0	77.6	116.4
167	13.5	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 40.068 ng
RT: 15.040 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:232 Resp: 662947

Ion	Ratio	Lower	Upper
232	100		
131	44.9	35.9	53.9
130	2.2	1.8	2.6
166	30.2	24.2	36.2







#60
Diethylphthalate
Concen: 38.794 ng
RT: 15.240 min Scan# 2066
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:149 Resp: 2527904

Ion	Ratio	Lower	Upper
149	100		
177	22.1	17.7	26.5
150	12.8	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 37.829 ng
RT: 15.463 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:204 Resp: 1225168

Ion	Ratio	Lower	Upper
204	100		
206	33.0	26.4	39.6
141	57.3	45.8	68.8







#62
4-Nitroaniline
Concen: 43.134 ng
RT: 15.487 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:138 Resp: 638694

Ion	Ratio	Lower	Upper
138	100		
92	48.7	28.7	68.7
108	69.5	49.5	89.5







#63
Azobenzene
Concen: 39.335 ng
RT: 15.763 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 2401881

Ion	Ratio	Lower	Upper
77	100		
182	27.7	7.7	47.7
105	22.3	2.3	42.3
51	23.2	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.204 min Scan# 2400
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:188 Resp: 1720919

Ion	Ratio	Lower	Upper
188	100		
94	10.1	8.1	12.1
80	11.1	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 41.395 ng
RT: 15.551 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:198 Resp: 412313

Ion	Ratio	Lower	Upper
198	100		
51	34.7	14.7	54.7
105	43.8	23.8	63.8







#66
n-Nitrosodiphenylamine
Concen: 38.450 ng
RT: 15.681 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:169 Resp: 2063452

Ion	Ratio	Lower	Upper
169	100		
168	66.1	52.9	79.3
167	36.0	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 38.570 ng
RT: 16.375 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:248 Resp: 745376

Ion	Ratio	Lower	Upper
248	100		
250	98.0	78.4	117.6
141	74.2	59.4	89.0







#68
Hexachlorobenzene
Concen: 38.435 ng
RT: 16.492 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:284 Resp: 873989

Ion	Ratio	Lower	Upper
284	100		
142	31.8	25.4	38.2
249	29.4	23.5	35.3







#69
Atrazine
Concen: 33.855 ng
RT: 16.657 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:200 Resp: 466258

Ion	Ratio	Lower	Upper
200	100		
173	27.6	7.6	47.6
215	49.0	29.0	69.0







#70
Pentachlorophenol
Concen: 41.352 ng
RT: 16.845 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:266 Resp: 610729

Ion	Ratio	Lower	Upper
266	100		
268	64.1	51.3	76.9
264	63.3	50.6	76.0







#71
Phenanthrene
Concen: 38.414 ng
RT: 17.245 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:178 Resp: 3727050

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.6	23.4
179	15.6	12.5	18.7







#72
Anthracene
Concen: 39.344 ng
RT: 17.339 min Scan# 2423
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:178 Resp: 3658135

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.2	22.8
179	15.7	12.6	18.8







#73
Carbazole
Concen: 40.463 ng
RT: 17.616 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:167 Resp: 3584494

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.1	25.7
139	12.8	10.2	15.4







#74
Di-n-butylphthalate
Concen: 41.958 ng
RT: 18.210 min Scan# 2571
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:149 Resp: 4336969

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.6	11.4
104	4.9	3.9	5.9







#75
Fluoranthene
Concen: 39.636 ng
RT: 19.328 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:202 Resp: 4349734

Ion	Ratio	Lower	Upper
202	100		
101	12.5	0.0	32.5
203	17.9	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.657 min Scan# 3157
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:240 Resp: 1881172

Ion	Ratio	Lower	Upper
240	100		
120	11.5	9.2	13.8
236	25.5	20.4	30.6







#77
Benzidine
Concen: 42.268 ng
RT: 19.522 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:184 Resp: 968916

Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.8	17.6
183	11.4	9.1	13.7







#78
Pyrene
Concen: 37.911 ng
RT: 19.704 min Scan# 2825
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:202 Resp: 4529960

Ion	Ratio	Lower	Upper
202	100		
200	21.4	17.1	25.7
203	17.8	14.2	21.4







#79
Terphenyl-d14
Concen: 75.038 ng
RT: 19.922 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:244 Resp: 7133778

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.7	8.5
122	12.0	9.6	14.4







#80
Butylbenzylphthalate
Concen: 43.087 ng
RT: 20.651 min Scan# 2986
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:149 Resp: 1972887

Ion	Ratio	Lower	Upper
149	100		
91	69.1	55.3	82.9
206	20.1	16.1	24.1







#81
Benzo(a)anthracene
Concen: 38.542 ng
RT: 21.633 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 4546432

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.8	32.8
229	20.0	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 43.694 ng
RT: 21.551 min Scan# 3139
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:252 Resp: 1610565

Ion	Ratio	Lower	Upper
252	100		
254	64.3	51.4	77.2
126	11.6	9.3	13.9







#83
Chrysene
Concen: 38.298 ng
RT: 21.704 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 4356257

Ion	Ratio	Lower	Upper
228	100		
226	30.6	24.5	36.7
229	20.2	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 42.929 ng
RT: 21.569 min Scan# 3142
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:149 Resp: 2980735

Ion	Ratio	Lower	Upper
149	100		
167	28.1	22.5	33.7
279	4.5	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 43.129 ng
RT: 22.869 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:149 Resp: 4786749

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.3	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.057 min Scan# 3735
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:264 Resp: 2043442

Ion	Ratio	Lower	Upper
264	100		
260	23.1	18.5	27.7
265	22.1	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 38.410 ng
RT: 28.927 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:276 Resp: 5583803

Ion	Ratio	Lower	Upper
276	100		
138	17.1	13.7	20.5
277	25.2	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 37.490 ng
RT: 23.980 min Scan# 3552
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:252 Resp: 4864855

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.2	25.8
125	10.3	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 36.566 ng
RT: 24.051 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:252 Resp: 4618537

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.4	9.1	13.7







#90
Benzo(a)pyrene
Concen: 38.180 ng
RT: 24.892 min Scan# 3707
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:252 Resp: 4204610

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.0
125	12.0	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 38.014 ng
RT: 29.009 min Scan# 44
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:278 Resp: 4646524

Ion	Ratio	Lower	Upper
278	100		
139	16.2	13.0	19.4
279	23.5	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 38.011 ng
RT: 30.127 min Scan# 4597
Delta R.T. 0.000 min
Lab File: BP024164.D
Acq: 12 Mar 2025 19:00

Tgt Ion:276 Resp: 4714749

Ion	Ratio	Lower	Upper
276	100		
277	24.0	19.2	28.8
138	21.8	17.4	26.2



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024165.D
Acq On : 12 Mar 2025 19:41
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Mar 13 04:09:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.775	152	362382	20.000	ng		0.00
21) Naphthalene-d8	10.551	136	1504882	20.000	ng		0.00
39) Acenaphthene-d10	14.410	164	918204	20.000	ng		0.00
64) Phenanthrene-d10	17.216	188	1730200	20.000	ng		0.01
76) Chrysene-d12	21.674	240	1856010	20.000	ng		0.02
86) Perylene-d12	25.086	264	2039908	20.000	ng		0.03
System Monitoring Compounds							
5) 2-Fluorophenol	5.381	112	2339783	100.747	ng		0.00
7) Phenol-d6	6.952	99	3169724	106.191	ng		0.00
23) Nitrobenzene-d5	8.928	82	2723412	100.962	ng		0.00
42) 2,4,6-Tribromophenol	15.928	330	1240173	110.267	ng		0.01
45) 2-Fluorobiphenyl	13.022	172	6129086	93.318	ng		0.00
79) Terphenyl-d14	19.945	244	8967575	95.606	ng		0.02
Target Compounds							Qvalue
2) 1,4-Dioxane	3.316	88	449736	49.365	ng		100
3) Pyridine	3.716	79	1264797m	53.398	ng		
4) n-Nitrosodimethylamine	3.622	42	417922	49.741	ng		99
6) Aniline	7.110	93	1402509	51.307	ng		100
8) 2-Chlorophenol	7.346	128	1260095	51.262	ng		98
9) Benzaldehyde	6.922	77	680024	49.899	ng		99
10) Phenol	6.981	94	1580281	52.648	ng		98
11) bis(2-Chloroethyl)ether	7.199	93	1214752	51.383	ng		99
12) 1,3-Dichlorobenzene	7.663	146	1338137	49.552	ng		99
13) 1,4-Dichlorobenzene	7.810	146	1352373	49.520	ng		100
14) 1,2-Dichlorobenzene	8.122	146	1299613	49.586	ng		99
15) Benzyl Alcohol	8.010	79	1016261	55.621	ng		100
16) 2,2'-oxybis(1-Chloropr...	8.299	45	1236313	53.871	ng		99
17) 2-Methylphenol	8.216	107	1010720	55.257	ng		100
18) Hexachloroethane	8.846	117	482437	49.052	ng		97
19) n-Nitroso-di-n-propyla...	8.575	70	929876	57.939	ng		99
20) 3+4-Methylphenols	8.540	107	1401754	56.333	ng		99
22) Acetophenone	8.587	105	1912289	49.450	ng	#	99
24) Nitrobenzene	8.969	77	1335288	50.299	ng		99
25) Isophorone	9.487	82	2389410	54.168	ng		100
26) 2-Nitrophenol	9.669	139	686089	56.096	ng		99
27) 2,4-Dimethylphenol	9.734	122	840657	52.362	ng		99
28) bis(2-Chloroethoxy)met...	9.963	93	1612977	50.476	ng		100
29) 2,4-Dichlorophenol	10.210	162	1152247	52.591	ng		98
30) 1,2,4-Trichlorobenzene	10.416	180	1212617	48.306	ng		99
31) Naphthalene	10.604	128	4020722	49.689	ng		100
32) Benzoic acid	9.893	122	900054	61.152	ng		99
33) 4-Chloroaniline	10.710	127	1481634	51.976	ng		99
34) Hexachlorobutadiene	10.893	225	703606	48.392	ng		99
35) Caprolactam	11.510	113	402356	60.646	ng		98
36) 4-Chloro-3-methylphenol	11.845	107	1251209	55.551	ng		100
37) 2-Methylnaphthalene	12.216	142	2749915	51.131	ng		100
38) 1-Methylnaphthalene	12.434	142	2647375	50.600	ng		99
40) 1,2,4,5-Tetrachloroben...	12.587	216	1331287	47.188	ng		100
41) Hexachlorocyclopentadiene	12.569	237	503901	47.999	ng		99
43) 2,4,6-Trichlorophenol	12.828	196	896606	52.112	ng		100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024165.D
Acq On : 12 Mar 2025 19:41
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Mar 13 04:09:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.904	196	969372	51.517	ng	99
46) 1,1'-Biphenyl	13.234	154	3427883	48.043	ng	100
47) 2-Chloronaphthalene	13.275	162	2559098	47.446	ng	99
48) 2-Nitroaniline	13.475	65	714343	56.066	ng	97
49) Acenaphthylene	14.128	152	4015684	50.533	ng	100
50) Dimethylphthalate	13.869	163	3252950	50.224	ng	100
51) 2,6-Dinitrotoluene	13.981	165	715171	54.206	ng	99
52) Acenaphthene	14.481	154	2516832	49.182	ng	99
53) 3-Nitroaniline	14.310	138	790217	56.128	ng	98
54) 2,4-Dinitrophenol	14.528	184	410616	58.439	ng	99
55) Dibenzofuran	14.816	168	4058172	48.816	ng	99
56) 4-Nitrophenol	14.639	139	685687	56.382	ng	98
57) 2,4-Dinitrotoluene	14.781	165	971087	56.659	ng	99
58) Fluorene	15.481	166	3230571	49.759	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.045	232	851401	52.152	ng	100
60) Diethylphthalate	15.251	149	3265289	50.786	ng	99
61) 4-Chlorophenyl-phenyle...	15.475	204	1572773	49.216	ng	99
62) 4-Nitroaniline	15.504	138	831169	56.889	ng	99
63) Azobenzene	15.769	77	3077821	51.085	ng	98
65) 4,6-Dinitro-2-methylph...	15.563	198	551280	55.051	ng	97
66) n-Nitrosodiphenylamine	15.686	169	2670636	49.497	ng	99
67) 4-Bromophenyl-phenylether	16.386	248	968698	49.858	ng	98
68) Hexachlorobenzene	16.510	284	1136445	49.709	ng	99
69) Atrazine	16.669	200	452391	32.672	ng	99
70) Pentachlorophenol	16.863	266	796545	53.644	ng	99
71) Phenanthrene	17.257	178	4759945	48.797	ng	100
72) Anthracene	17.351	178	4723119	50.526	ng	100
73) Carbazole	17.633	167	4585466	51.485	ng	100
74) Di-n-butylphthalate	18.222	149	5502824	52.951	ng	100
75) Fluoranthene	19.345	202	5644914	51.163	ng	99
77) Benzidine	19.539	184	804810	35.585	ng	99
78) Pyrene	19.727	202	5876091	49.843	ng	100
80) Butylbenzylphthalate	20.674	149	2571381	56.920	ng	99
81) Benzo(a)anthracene	21.657	228	5851357	50.277	ng	100
82) 3,3'-Dichlorobenzidine	21.580	252	2015249	55.414	ng	99
83) Chrysene	21.727	228	5625314	50.126	ng	100
84) Bis(2-ethylhexyl)phtha...	21.592	149	3802016	55.499	ng	100
85) Di-n-octyl phthalate	22.904	149	6298049	57.516	ng	100
87) Indeno(1,2,3-cd)pyrene	28.956	276	7354717	50.680	ng	99
88) Benzo(b)fluoranthene	24.004	252	6342083	48.959	ng	98
89) Benzo(k)fluoranthene	24.080	252	6136938	48.672	ng	99
90) Benzo(a)pyrene	24.927	252	5581393	50.770	ng	99
91) Dibenzo(a,h)anthracene	29.045	278	6072440	49.766	ng	99
92) Benzo(g,h,i)perylene	30.168	276	6130219	49.509	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024165.D
Acq On : 12 Mar 2025 19:41
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Mar 13 04:09:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.775 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 362382

Ion	Ratio	Lower	Upper
152	100		
150	157.9	125.5	188.3
115	60.0	46.3	69.5







#2
1,4-Dioxane
Concen: 49.365 ng
RT: 3.316 min Scan# 39
Delta R.T. -0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion: 88 Resp: 449736

Ion	Ratio	Lower	Upper
88	100		
58	62.6	49.8	74.6
43	21.6	17.3	25.9







#3
Pyridine
Concen: 53.398 ng m
RT: 3.716 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 1264797

Ion	Ratio	Lower	Upper
79	100		
52	54.1	43.8	65.8
51	27.0	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 49.741 ng
RT: 3.622 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion: 42 Resp: 417922

Ion	Ratio	Lower	Upper
42	100		
74	176.4	139.8	209.6
44	6.9	5.3	7.9







#5
2-Fluorophenol
Concen: 100.747 ng
RT: 5.381 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:112 Resp: 2339783

Ion	Ratio	Lower	Upper
112	100		
64	51.2	41.3	61.9
63	25.7	20.6	31.0







#6
Aniline
Concen: 51.307 ng
RT: 7.110 min Scan# 684
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion: 93 Resp: 1402509

Ion	Ratio	Lower	Upper
93	100		
66	35.5	28.2	42.2
65	17.7	14.0	21.0







#7
Phenol-d6
Concen: 106.191 ng
RT: 6.952 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 99 Resp: 3169724

Ion	Ratio	Lower	Upper
99	100		
42	12.9	10.6	16.0
71	29.1	23.3	34.9







#8
2-Chlorophenol
Concen: 51.262 ng
RT: 7.346 min Scan# 724
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:128 Resp: 1260095

Ion	Ratio	Lower	Upper
128	100		
130	32.7	12.0	52.0
64	40.4	21.6	61.6







#9
Benzaldehyde
Concen: 49.899 ng
RT: 6.922 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 680024

Ion	Ratio	Lower	Upper
77	100		
105	102.9	81.4	121.4
106	101.1	81.1	121.1







#10
Phenol
Concen: 52.648 ng
RT: 6.981 min Scan# 662
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion: 94 Resp: 1580281

Ion	Ratio	Lower	Upper
94	100		
65	24.4	4.9	44.9
66	30.8	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 51.383 ng
RT: 7.199 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 93 Resp: 1214752

Ion	Ratio	Lower	Upper
93	100		
63	67.9	47.4	87.4
95	32.4	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 49.552 ng
RT: 7.663 min Scan# 778
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:146 Resp: 1338137

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.6	77.4
75	27.6	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 49.520 ng
RT: 7.810 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:146 Resp: 1352373

Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.9	77.9
111	39.5	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 49.586 ng
RT: 8.122 min Scan# 856
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:146 Resp: 1299613

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.4	77.2
111	41.2	33.2	49.8







#15
Benzyl Alcohol
Concen: 55.621 ng
RT: 8.010 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 1016261

Ion	Ratio	Lower	Upper
79	100		
108	83.4	66.5	99.7
77	63.7	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 53.871 ng
RT: 8.299 min Scan# 886
Delta R.T. -0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion: 45 Resp: 1236313

Ion	Ratio	Lower	Upper
45	100		
77	19.4	0.0	39.0
79	14.8	0.0	35.6







#17
2-Methylphenol
Concen: 55.257 ng
RT: 8.216 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:107 Resp: 1010720

Ion	Ratio	Lower	Upper
107	100		
108	112.3	90.0	135.0
77	41.9	34.0	51.0
79	41.0	32.7	49.1







#18
Hexachloroethane
Concen: 49.052 ng
RT: 8.846 min Scan# 979
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:117 Resp: 482437

Ion	Ratio	Lower	Upper
117	100		
119	96.3	77.4	116.0
201	100.3	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 57.939 ng
RT: 8.575 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 70 Resp: 929876

Ion	Ratio	Lower	Upper
70	100		
42	45.2	36.0	54.0
101	10.5	8.3	12.5
130	24.8	18.6	28.0







#20
3+4-Methylphenols
Concen: 56.333 ng
RT: 8.540 min Scan# 927
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:107 Resp: 1401754

Ion	Ratio	Lower	Upper
107	100		
108	88.4	69.7	109.7
77	30.3	10.3	50.3
79	26.3	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.551 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:136 Resp: 1504882

Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	6.4	5.3	7.9
68	6.2	5.0	7.4







#22
Acetophenone
Concen: 49.450 ng
RT: 8.587 min Scan# 935
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:105 Resp: 1912289

Ion	Ratio	Lower	Upper
105	100		
71	3.9	2.4	3.6#
51	20.8	16.2	24.2
120	23.8	19.0	28.6







#23
Nitrobenzene-d5
Concen: 100.962 ng
RT: 8.928 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 2723412

Ion	Ratio	Lower	Upper
82	100		
128	47.4	37.4	56.0
54	40.1	32.5	48.7







#24
Nitrobenzene
Concen: 50.299 ng
RT: 8.969 min Scan# 1000
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion: 77 Resp: 1335288

Ion	Ratio	Lower	Upper
77	100		
123	50.4	39.4	59.2
65	12.3	9.9	14.9







#25
Isophorone
Concen: 54.168 ng
RT: 9.487 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 2389410

Ion	Ratio	Lower	Upper
82	100		
95	6.7	5.3	7.9
138	18.3	14.8	22.2







#26
2-Nitrophenol
Concen: 56.096 ng
RT: 9.669 min Scan# 1119
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:139 Resp: 686089

Ion	Ratio	Lower	Upper
139	100		
109	26.8	21.7	32.5
65	44.2	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 52.362 ng
RT: 9.734 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:122 Resp: 840657

Ion	Ratio	Lower	Upper
122	100		
107	111.8	89.3	133.9
121	58.2	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 50.476 ng
RT: 9.963 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion: 93 Resp: 1612977

Ion	Ratio	Lower	Upper
93	100		
95	32.6	25.9	38.9
123	11.8	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 52.591 ng
RT: 10.210 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 1152247

Ion	Ratio	Lower	Upper
162	100		
164	65.6	44.2	84.2
98	34.5	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 48.306 ng
RT: 10.416 min Scan# 1246
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:180 Resp: 1212617

Ion	Ratio	Lower	Upper
180	100		
182	95.7	76.1	114.1
145	30.7	24.2	36.2







#31
Naphthalene
Concen: 49.689 ng
RT: 10.604 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:128 Resp: 4020722

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.7	13.1
127	13.2	10.4	15.6







#32
Benzoic acid
Concen: 61.152 ng
RT: 9.893 min Scan# 1157
Delta R.T. 0.017 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:122 Resp: 900054

Ion	Ratio	Lower	Upper
122	100		
105	121.3	100.3	140.3
77	87.4	69.1	109.1







#33
4-Chloroaniline
Concen: 51.976 ng
RT: 10.710 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:127 Resp: 1481634

Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.8	38.6
65	27.3	21.2	31.8
92	18.3	14.4	21.6







#34
Hexachlorobutadiene
Concen: 48.392 ng
RT: 10.893 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:225 Resp: 703606

Ion	Ratio	Lower	Upper
225	100		
223	63.2	49.9	74.9
227	63.9	52.2	78.2







#35
Caprolactam
Concen: 60.646 ng
RT: 11.510 min Scan# 14
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:113 Resp: 402356

Ion	Ratio	Lower	Upper
113	100		
55	131.8	113.6	153.6
56	98.2	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 55.551 ng
RT: 11.845 min Scan# 1489
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:107 Resp: 1251209

Ion	Ratio	Lower	Upper
107	100		
144	27.1	21.6	32.4
142	82.1	65.4	98.0







#37
2-Methylnaphthalene
Concen: 51.131 ng
RT: 12.216 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:142 Resp: 2749915

Ion	Ratio	Lower	Upper
142	100		
141	88.3	70.3	105.5
115	32.4	26.0	39.0







#38
1-Methylnaphthalene
Concen: 50.600 ng
RT: 12.434 min Scan# 1589
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:142 Resp: 2647375

Ion	Ratio	Lower	Upper
142	100		
141	91.4	73.9	110.9
116	3.7	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:164 Resp: 918204

Ion	Ratio	Lower	Upper
164	100		
162	99.9	80.6	121.0
160	44.2	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 47.188 ng
RT: 12.587 min Scan# 1615
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:216 Resp: 1331287

Ion	Ratio	Lower	Upper
216	100		
214	78.5	62.8	94.2
179	20.9	16.8	25.2
108	18.5	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 47.999 ng
RT: 12.569 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:237 Resp: 503901

Ion	Ratio	Lower	Upper
237	100		
235	63.2	41.8	81.8
272	13.5	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 110.267 ng
RT: 15.928 min Scan# 2183
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:330 Resp: 1240173

Ion	Ratio	Lower	Upper
330	100		
332	96.7	77.6	116.4
141	30.3	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 52.112 ng
RT: 12.828 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:196 Resp: 896606

Ion	Ratio	Lower	Upper
196	100		
198	95.7	76.6	115.0
200	30.9	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 51.517 ng
RT: 12.904 min Scan# 1669
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:196 Resp: 969372

Ion	Ratio	Lower	Upper
196	100		
198	94.7	76.5	114.7
97	48.5	38.7	58.1
132	28.0	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 93.318 ng
RT: 13.022 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:172 Resp: 6129086

Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.3	42.5
170	23.7	19.0	28.4







#46
1,1'-Biphenyl
Concen: 48.043 ng
RT: 13.234 min Scan# 1725
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:154 Resp: 3427883

Ion	Ratio	Lower	Upper
154	100		
153	40.0	20.3	60.3
76	12.9	0.0	32.9







#47
2-Chloronaphthalene
Concen: 47.446 ng
RT: 13.275 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 2559098

Ion	Ratio	Lower	Upper
162	100		
127	38.4	31.5	47.3
164	33.3	26.6	39.8







#48
2-Nitroaniline
Concen: 56.066 ng
RT: 13.475 min Scan# 1766
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion: 65 Resp: 714343

Ion	Ratio	Lower	Upper
65	100		
92	75.3	62.1	93.1
138	120.9	99.5	149.3







#49
Acenaphthylene
Concen: 50.533 ng
RT: 14.128 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 4015684

Ion	Ratio	Lower	Upper
152	100		
151	19.9	16.0	24.0
153	13.3	10.6	15.8







#50
Dimethylphthalate
Concen: 50.224 ng
RT: 13.869 min Scan# 1833
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:163 Resp: 3252950

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.0
164	10.2	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 54.206 ng
RT: 13.981 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 715171

Ion	Ratio	Lower	Upper
165	100		
63	47.9	37.7	56.5
89	53.0	41.9	62.9







#52
Acenaphthene
Concen: 49.182 ng
RT: 14.481 min Scan# 1937
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:154 Resp: 2516832

Ion	Ratio	Lower	Upper
154	100		
153	112.2	90.5	135.7
152	53.2	43.1	64.7







#53
3-Nitroaniline
Concen: 56.128 ng
RT: 14.310 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:138 Resp: 790217

Ion	Ratio	Lower	Upper
138	100		
108	9.9	8.2	12.2
92	110.1	86.7	130.1







#54
2,4-Dinitrophenol
Concen: 58.439 ng
RT: 14.528 min Scan# 1945
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:184 Resp: 410616

Ion	Ratio	Lower	Upper
184	100		
63	53.9	42.1	63.1
154	56.4	44.6	67.0







#55
Dibenzofuran
Concen: 48.816 ng
RT: 14.816 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:168 Resp: 4058172

Ion	Ratio	Lower	Upper
168	100		
139	37.5	29.1	43.7
169	13.5	10.8	16.2







#56
4-Nitrophenol
Concen: 56.382 ng
RT: 14.639 min Scan# 1964
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:139 Resp: 685687

Ion	Ratio	Lower	Upper
139	100		
109	62.5	43.3	83.3
65	78.1	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 56.659 ng
RT: 14.781 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 971087

Ion	Ratio	Lower	Upper
165	100		
63	36.0	29.0	43.6
89	68.4	54.2	81.4
182	3.0	2.3	3.5







#58
Fluorene
Concen: 49.759 ng
RT: 15.481 min Scan# 2107
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:166 Resp: 3230571

Ion	Ratio	Lower	Upper
166	100		
165	95.5	77.6	116.4
167	13.4	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 52.152 ng
RT: 15.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:232 Resp: 851401

Ion	Ratio	Lower	Upper
232	100		
131	44.9	35.9	53.9
130	2.2	1.8	2.6
166	30.4	24.2	36.2







#60
Diethylphthalate
Concen: 50.786 ng
RT: 15.251 min Scan# 2068
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:149 Resp: 3265289

Ion	Ratio	Lower	Upper
149	100		
177	22.5	17.7	26.5
150	12.5	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 49.216 ng
RT: 15.475 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:204 Resp: 1572773

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.4	39.6
141	58.0	45.8	68.8







#62
4-Nitroaniline
Concen: 56.889 ng
RT: 15.504 min Scan# 2111
Delta R.T. 0.017 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:138 Resp: 831169

Ion	Ratio	Lower	Upper
138	100		
92	48.1	28.7	68.7
108	68.9	49.5	89.5







#63
Azobenzene
Concen: 51.085 ng
RT: 15.769 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 3077821

Ion	Ratio	Lower	Upper
77	100		
182	25.6	7.7	47.7
105	22.2	2.3	42.3
51	23.3	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.216 min Scan# 2402
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:188 Resp: 1730200

Ion	Ratio	Lower	Upper
188	100		
94	10.2	8.1	12.1
80	10.8	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 55.051 ng
RT: 15.563 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:198 Resp: 551280

Ion	Ratio	Lower	Upper
198	100		
51	33.1	14.7	54.7
105	41.5	23.8	63.8







#66
n-Nitrosodiphenylamine
Concen: 49.497 ng
RT: 15.686 min Scan# 2142
Delta R.T. 0.006 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:169 Resp: 2670636

Ion	Ratio	Lower	Upper
169	100		
168	66.9	52.9	79.3
167	36.4	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 49.858 ng
RT: 16.386 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:248 Resp: 968698

Ion	Ratio	Lower	Upper
248	100		
250	97.0	78.4	117.6
141	71.8	59.4	89.0







#68
Hexachlorobenzene
Concen: 49.709 ng
RT: 16.510 min Scan# 2282
Delta R.T. 0.017 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:284 Resp: 1136445

Ion	Ratio	Lower	Upper
284	100		
142	31.4	25.4	38.2
249	29.6	23.5	35.3







#69
Atrazine
Concen: 32.672 ng
RT: 16.669 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:200 Resp: 452391

Ion	Ratio	Lower	Upper
200	100		
173	26.3	7.6	47.6
215	48.6	29.0	69.0







#70
Pentachlorophenol
Concen: 53.644 ng
RT: 16.863 min Scan# 2342
Delta R.T. 0.017 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:266 Resp: 796545

Ion	Ratio	Lower	Upper
266	100		
268	63.1	51.3	76.9
264	63.4	50.6	76.0







#71
Phenanthrene
Concen: 48.797 ng
RT: 17.257 min Scan# 24
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:178 Resp: 4759945

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.6	23.4
179	15.8	12.5	18.7







#72
Anthracene
Concen: 50.526 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:178 Resp: 4723119

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.2	22.8
179	15.6	12.6	18.8







#73
Carbazole
Concen: 51.485 ng
RT: 17.633 min Scan# 24
Delta R.T. 0.017 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:167 Resp: 4585466

Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.1	25.7
139	12.9	10.2	15.4







#74
Di-n-butylphthalate
Concen: 52.951 ng
RT: 18.222 min Scan# 2573
Delta R.T. 0.012 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:149 Resp: 5502824

Ion	Ratio	Lower	Upper
149	100		
150	9.7	7.6	11.4
104	5.0	3.9	5.9







#75
Fluoranthene
Concen: 51.163 ng
RT: 19.345 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:202 Resp: 5644914

Ion	Ratio	Lower	Upper
202	100		
101	13.0	0.0	32.5
203	18.2	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.674 min Scan# 3160
Delta R.T. 0.017 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:240 Resp: 1856010

Ion	Ratio	Lower	Upper
240	100		
120	12.2	9.2	13.8
236	25.7	20.4	30.6







#77
Benzidine
Concen: 35.585 ng
RT: 19.539 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:184 Resp: 804810

Ion	Ratio	Lower	Upper
184	100		
185	14.5	11.8	17.6
183	11.7	9.1	13.7







#78
Pyrene
Concen: 49.843 ng
RT: 19.727 min Scan# 2829
Delta R.T. 0.023 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:202 Resp: 5876091

Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.1	25.7
203	17.7	14.2	21.4







#79
Terphenyl-d14
Concen: 95.606 ng
RT: 19.945 min Scan# 2
Delta R.T. 0.023 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:244 Resp: 8967575

Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.7	8.5
122	11.3	9.6	14.4







#80
Butylbenzylphthalate
Concen: 56.920 ng
RT: 20.674 min Scan# 2990
Delta R.T. 0.023 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:149 Resp: 2571381

Ion	Ratio	Lower	Upper
149	100		
91	69.7	55.3	82.9
206	20.9	16.1	24.1







#81
Benzo(a)anthracene
Concen: 50.277 ng
RT: 21.657 min Scan# 3
Delta R.T. 0.023 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 5851357

Ion	Ratio	Lower	Upper
228	100		
226	27.6	21.8	32.8
229	19.9	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 55.414 ng
RT: 21.580 min Scan# 3144
Delta R.T. 0.029 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:252 Resp: 2015249

Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.4	77.2
126	10.8	9.3	13.9







#83
Chrysene
Concen: 50.126 ng
RT: 21.727 min Scan# 31
Delta R.T. 0.023 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 5625314

Ion	Ratio	Lower	Upper
228	100		
226	30.7	24.5	36.7
229	19.9	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 55.499 ng
RT: 21.592 min Scan# 3146
Delta R.T. 0.023 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:149 Resp: 3802016

Ion	Ratio	Lower	Upper
149	100		
167	28.2	22.5	33.7
279	4.9	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 57.516 ng
RT: 22.904 min Scan# 3
Delta R.T. 0.035 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:149 Resp: 6298049

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.3	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.086 min Scan# 3740
Delta R.T. 0.029 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:264 Resp: 2039908

Ion	Ratio	Lower	Upper
264	100		
260	23.3	18.5	27.7
265	21.9	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 50.680 ng
RT: 28.956 min Scan# 4
Delta R.T. 0.029 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:276 Resp: 7354717

Ion	Ratio	Lower	Upper
276	100		
138	16.5	13.7	20.5
277	25.2	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 48.959 ng
RT: 24.004 min Scan# 3556
Delta R.T. 0.023 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:252 Resp: 6342083

Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.2	25.8
125	11.3	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 48.672 ng
RT: 24.080 min Scan# 3
Delta R.T. 0.029 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:252 Resp: 6136938

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	10.4	9.1	13.7







#90
Benzo(a)pyrene
Concen: 50.770 ng
RT: 24.927 min Scan# 3713
Delta R.T. 0.035 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:252 Resp: 5581393

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	11.4	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 49.766 ng
RT: 29.045 min Scan# 44
Delta R.T. 0.035 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 6072440

Ion	Ratio	Lower	Upper
278	100		
139	16.0	13.0	19.4
279	24.4	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 49.509 ng
RT: 30.168 min Scan# 4604
Delta R.T. 0.041 min
Lab File: BP024165.D
Acq: 12 Mar 2025 19:41

Tgt Ion:276 Resp: 6130219

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.2	28.8
138	22.7	17.4	26.2



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024166.D
Acq On : 12 Mar 2025 20:21
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Mar 13 04:09:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	394376	20.000	ng	0.00
21) Naphthalene-d8	10.552	136	1591612	20.000	ng	0.00
39) Acenaphthene-d10	14.404	164	936110	20.000	ng	0.00
64) Phenanthrene-d10	17.210	188	1759587	20.000	ng	0.00
76) Chrysene-d12	21.669	240	1817869	20.000	ng	0.01
86) Perylene-d12	25.074	264	2003229	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	3091896	122.332	ng	0.00
7) Phenol-d6	6.958	99	4151063	127.785	ng	0.00
23) Nitrobenzene-d5	8.922	82	3516139	123.247	ng	0.00
42) 2,4,6-Tribromophenol	15.922	330	1542341	134.511	ng	0.00
45) 2-Fluorobiphenyl	13.016	172	7675412	114.626	ng	0.00
79) Terphenyl-d14	19.933	244	10595349	115.330	ng	0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.322	88	588784	59.385	ng	100
3) Pyridine	3.717	79	1653707m	64.153	ng	
4) n-Nitrosodimethylamine	3.622	42	551093	60.270	ng	99
6) Aniline	7.111	93	1788370	60.115	ng	99
8) 2-Chlorophenol	7.346	128	1656453	61.919	ng	99
9) Benzaldehyde	6.922	77	830475	55.995	ng	99
10) Phenol	6.981	94	2072585	63.448	ng	99
11) bis(2-Chloroethyl)ether	7.199	93	1606110	62.425	ng	99
12) 1,3-Dichlorobenzene	7.663	146	1740829	59.234	ng	100
13) 1,4-Dichlorobenzene	7.805	146	1778400	59.836	ng	100
14) 1,2-Dichlorobenzene	8.122	146	1703894	59.737	ng	100
15) Benzyl Alcohol	8.010	79	1332675	67.021	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.293	45	1610981	64.502	ng	99
17) 2-Methylphenol	8.210	107	1316438	66.132	ng	99
18) Hexachloroethane	8.846	117	636840	59.498	ng	96
19) n-Nitroso-di-n-propyla...	8.575	70	1191669	68.227	ng	99
20) 3+4-Methylphenols	8.540	107	1809984	66.838	ng	100
22) Acetophenone	8.587	105	2478568	60.600	ng	99
24) Nitrobenzene	8.963	77	1731589	61.673	ng	99
25) Isophorone	9.487	82	3063158	65.658	ng	100
26) 2-Nitrophenol	9.669	139	904714	69.941	ng	97
27) 2,4-Dimethylphenol	9.734	122	1095778	64.534	ng	99
28) bis(2-Chloroethoxy)met...	9.963	93	2086687	61.742	ng	99
29) 2,4-Dichlorophenol	10.204	162	1497152	64.609	ng	99
30) 1,2,4-Trichlorobenzene	10.416	180	1578453	59.453	ng	100
31) Naphthalene	10.604	128	5145859	60.129	ng	100
32) Benzoic acid	9.904	122	1203679	77.325	ng	98
33) 4-Chloroaniline	10.710	127	1859148	61.665	ng	99
34) Hexachlorobutadiene	10.887	225	918034	59.699	ng	98
35) Caprolactam	11.510	113	496683	70.784	ng	99
36) 4-Chloro-3-methylphenol	11.846	107	1582952	66.451	ng	98
37) 2-Methylnaphthalene	12.216	142	3504891	61.618	ng	100
38) 1-Methylnaphthalene	12.434	142	3364504	60.802	ng	100
40) 1,2,4,5-Tetrachloroben...	12.587	216	1707961	59.382	ng	100
41) Hexachlorocyclopentadiene	12.563	237	656781	61.365	ng	98
43) 2,4,6-Trichlorophenol	12.828	196	1148545	65.478	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024166.D
Acq On : 12 Mar 2025 20:21
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Mar 13 04:09:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.904	196	1233147	64.282	ng	99
46) 1,1'-Biphenyl	13.234	154	4312903	59.291	ng	99
47) 2-Chloronaphthalene	13.275	162	3269176	59.452	ng	100
48) 2-Nitroaniline	13.475	65	890207	68.533	ng	99
49) Acenaphthylene	14.128	152	5025011	62.024	ng	100
50) Dimethylphthalate	13.863	163	3997694	60.542	ng	100
51) 2,6-Dinitrotoluene	13.981	165	887706	65.996	ng	96
52) Acenaphthene	14.475	154	3153799	60.450	ng	99
53) 3-Nitroaniline	14.316	138	963434	67.123	ng	99
54) 2,4-Dinitrophenol	14.522	184	523590	73.093	ng	97
55) Dibenzofuran	14.816	168	5051493	59.602	ng	100
56) 4-Nitrophenol	14.640	139	831765	67.086	ng	99
57) 2,4-Dinitrotoluene	14.775	165	1211692	69.346	ng	98
58) Fluorene	15.469	166	3951671	59.702	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.045	232	1055860	63.439	ng	99
60) Diethylphthalate	15.245	149	4030341	61.486	ng	99
61) 4-Chlorophenyl-phenyle...	15.469	204	1934265	59.370	ng	100
62) 4-Nitroaniline	15.498	138	1025149	68.824	ng	98
63) Azobenzene	15.769	77	3829148	62.340	ng	98
65) 4,6-Dinitro-2-methylph...	15.557	198	698260	68.563	ng	93
66) n-Nitrosodiphenylamine	15.687	169	3334914	60.776	ng	100
67) 4-Bromophenyl-phenylether	16.381	248	1211468	61.311	ng	97
68) Hexachlorobenzene	16.498	284	1425714	61.320	ng	97
70) Pentachlorophenol	16.857	266	1002726	66.402	ng	100
71) Phenanthrene	17.251	178	5863198	59.103	ng	99
72) Anthracene	17.345	178	5754537	60.532	ng	100
73) Carbazole	17.622	167	5598533	61.810	ng	99
74) Di-n-butylphthalate	18.216	149	6844468	64.761	ng	100
75) Fluoranthene	19.333	202	6849338	61.042	ng	99
77) Benzidine	19.528	184	1518742	68.561	ng	100
78) Pyrene	19.710	202	7049649	61.052	ng	99
80) Butylbenzylphthalate	20.663	149	3170269	71.649	ng	99
81) Benzo(a)anthracene	21.645	228	7010330	61.500	ng	99
82) 3,3'-Dichlorobenzidine	21.569	252	2547922	71.531	ng	99
83) Chrysene	21.716	228	6653811	60.535	ng	99
84) Bis(2-ethylhexyl)phtha...	21.586	149	4788368	71.363	ng	98
85) Di-n-octyl phthalate	22.886	149	7825427	72.964	ng	100
87) Indeno(1,2,3-cd)pyrene	28.951	276	8887645m	62.364	ng	
88) Benzo(b)fluoranthene	23.998	252	7606878	59.798	ng	99
89) Benzo(k)fluoranthene	24.068	252	7515459	60.696	ng	99
90) Benzo(a)pyrene	24.915	252	6723161	62.275	ng	99
91) Dibenzo(a,h)anthracene	29.056	278	7370514	61.510	ng	99
92) Benzo(g,h,i)perylene	30.156	276	7489359	61.593	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024166.D
Acq On : 12 Mar 2025 20:21
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Mar 13 04:09:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 394376

Ion	Ratio	Lower	Upper
152	100		
150	155.1	125.5	188.3
115	57.8	46.3	69.5







#2
1,4-Dioxane
Concen: 59.385 ng
RT: 3.322 min Scan# 40
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion: 88 Resp: 588784

Ion	Ratio	Lower	Upper
88	100		
58	62.5	49.8	74.6
43	21.3	17.3	25.9







#3
Pyridine
Concen: 64.153 ng m
RT: 3.717 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 1653707

Ion	Ratio	Lower	Upper
79	100		
52	54.2	43.8	65.8
51	26.6	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 60.270 ng
RT: 3.622 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion: 42 Resp: 551093

Ion	Ratio	Lower	Upper
42	100		
74	175.9	139.8	209.6
44	6.8	5.3	7.9







#5
2-Fluorophenol
Concen: 122.332 ng
RT: 5.381 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:112 Resp: 3091896

Ion	Ratio	Lower	Upper
112	100		
64	51.7	41.3	61.9
63	25.7	20.6	31.0







#6
Aniline
Concen: 60.115 ng
RT: 7.111 min Scan# 684
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion: 93 Resp: 1788370

Ion	Ratio	Lower	Upper
93	100		
66	35.6	28.2	42.2
65	17.9	14.0	21.0







#7
Phenol-d6
Concen: 127.785 ng
RT: 6.958 min Scan# 6
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 99 Resp: 4151063

Ion	Ratio	Lower	Upper
99	100		
42	12.7	10.6	16.0
71	28.9	23.3	34.9







#8
2-Chlorophenol
Concen: 61.919 ng
RT: 7.346 min Scan# 724
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:128 Resp: 1656453

Ion	Ratio	Lower	Upper
128	100		
130	32.7	12.0	52.0
64	40.9	21.6	61.6







#9
Benzaldehyde
Concen: 55.995 ng
RT: 6.922 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 830475

Ion	Ratio	Lower	Upper
77	100		
105	102.0	81.4	121.4
106	98.7	81.1	121.1







#10
Phenol
Concen: 63.448 ng
RT: 6.981 min Scan# 662
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion: 94 Resp: 2072585

Ion	Ratio	Lower	Upper
94	100		
65	24.5	4.9	44.9
66	31.3	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 62.425 ng
RT: 7.199 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 93 Resp: 1606110

Ion	Ratio	Lower	Upper
93	100		
63	68.3	47.4	87.4
95	32.6	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 59.234 ng
RT: 7.663 min Scan# 778
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:146 Resp: 1740829

Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.6	77.4
75	28.1	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 59.836 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:146 Resp: 1778400

Ion	Ratio	Lower	Upper
146	100		
148	64.8	51.9	77.9
111	39.4	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 59.737 ng
RT: 8.122 min Scan# 856
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:146 Resp: 1703894

Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.4	77.2
111	41.8	33.2	49.8







#15
Benzyl Alcohol
Concen: 67.021 ng
RT: 8.010 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 1332675

Ion	Ratio	Lower	Upper
79	100		
108	82.9	66.5	99.7
77	63.1	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 64.502 ng
RT: 8.293 min Scan# 885
Delta R.T. -0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion: 45 Resp: 1610981

Ion	Ratio	Lower	Upper
45	100		
77	18.8	0.0	39.0
79	14.7	0.0	35.6







#17
2-Methylphenol
Concen: 66.132 ng
RT: 8.210 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:107 Resp: 1316438

Ion	Ratio	Lower	Upper
107	100		
108	111.0	90.0	135.0
77	42.5	34.0	51.0
79	40.9	32.7	49.1







#18
Hexachloroethane
Concen: 59.498 ng
RT: 8.846 min Scan# 979
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:117 Resp: 636840

Ion	Ratio	Lower	Upper
117	100		
119	96.0	77.4	116.0
201	103.0	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 68.227 ng
RT: 8.575 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 70 Resp: 1191669

Ion	Ratio	Lower	Upper
70	100		
42	44.3	36.0	54.0
101	10.6	8.3	12.5
130	24.5	18.6	28.0







#20
3+4-Methylphenols
Concen: 66.838 ng
RT: 8.540 min Scan# 927
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:107 Resp: 1809984

Ion	Ratio	Lower	Upper
107	100		
108	90.1	69.7	109.7
77	30.4	10.3	50.3
79	26.5	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.552 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:136 Resp: 1591612

Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	6.5	5.3	7.9
68	6.3	5.0	7.4







#22
Acetophenone
Concen: 60.600 ng
RT: 8.587 min Scan# 935
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:105 Resp: 2478568

Ion	Ratio	Lower	Upper
105	100		
71	3.0	2.4	3.6
51	19.7	16.2	24.2
120	24.0	19.0	28.6







#23
Nitrobenzene-d5
Concen: 123.247 ng
RT: 8.922 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 3516139

Ion	Ratio	Lower	Upper
82	100		
128	48.4	37.4	56.0
54	40.0	32.5	48.7







#24
Nitrobenzene
Concen: 61.673 ng
RT: 8.963 min Scan# 999
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion: 77 Resp: 1731589

Ion	Ratio	Lower	Upper
77	100		
123	50.3	39.4	59.2
65	12.3	9.9	14.9







#25
Isophorone
Concen: 65.658 ng
RT: 9.487 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 3063158

Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.3	7.9
138	18.7	14.8	22.2







#26
2-Nitrophenol
Concen: 69.941 ng
RT: 9.669 min Scan# 1119
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:139 Resp: 904714

Ion	Ratio	Lower	Upper
139	100		
109	26.2	21.7	32.5
65	42.7	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 64.534 ng
RT: 9.734 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:122 Resp: 1095778

Ion	Ratio	Lower	Upper
122	100		
107	112.0	89.3	133.9
121	57.7	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 61.742 ng
RT: 9.963 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion: 93 Resp: 2086687

Ion	Ratio	Lower	Upper
93	100		
95	32.8	25.9	38.9
123	12.0	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 64.609 ng
RT: 10.204 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 1497152

Ion	Ratio	Lower	Upper
162	100		
164	64.8	44.2	84.2
98	36.6	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 59.453 ng
RT: 10.416 min Scan# 1246
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:180 Resp: 1578453

Ion	Ratio	Lower	Upper
180	100		
182	95.2	76.1	114.1
145	29.7	24.2	36.2







#31
Naphthalene
Concen: 60.129 ng
RT: 10.604 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:128 Resp: 5145859

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.7	13.1
127	13.0	10.4	15.6







#32
Benzoic acid
Concen: 77.325 ng
RT: 9.904 min Scan# 1159
Delta R.T. 0.029 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:122 Resp: 1203679

Ion	Ratio	Lower	Upper
122	100		
105	119.3	100.3	140.3
77	85.8	69.1	109.1







#33
4-Chloroaniline
Concen: 61.665 ng
RT: 10.710 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 1859148

Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.8	38.6
65	27.3	21.2	31.8
92	18.5	14.4	21.6







#34
Hexachlorobutadiene
Concen: 59.699 ng
RT: 10.887 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:225 Resp: 918034

Ion	Ratio	Lower	Upper
225	100		
223	63.6	49.9	74.9
227	63.8	52.2	78.2







#35
Caprolactam
Concen: 70.784 ng
RT: 11.510 min Scan# 14
Delta R.T. 0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:113 Resp: 496683

Ion	Ratio	Lower	Upper
113	100		
55	133.6	113.6	153.6
56	98.8	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 66.451 ng
RT: 11.846 min Scan# 1489
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:107 Resp: 1582952

Ion	Ratio	Lower	Upper
107	100		
144	27.1	21.6	32.4
142	84.6	65.4	98.0







#37
2-Methylnaphthalene
Concen: 61.618 ng
RT: 12.216 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:142 Resp: 3504891

Ion	Ratio	Lower	Upper
142	100		
141	88.0	70.3	105.5
115	32.6	26.0	39.0







#38
1-Methylnaphthalene
Concen: 60.802 ng
RT: 12.434 min Scan# 1589
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:142 Resp: 3364504

Ion	Ratio	Lower	Upper
142	100		
141	92.2	73.9	110.9
116	3.6	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.404 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:164 Resp: 936110

Ion	Ratio	Lower	Upper
164	100		
162	100.0	80.6	121.0
160	44.5	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 59.382 ng
RT: 12.587 min Scan# 1615
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:216 Resp: 1707961

Ion	Ratio	Lower	Upper
216	100		
214	78.5	62.8	94.2
179	20.9	16.8	25.2
108	18.1	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 61.365 ng
RT: 12.563 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:237 Resp: 656781

Ion	Ratio	Lower	Upper
237	100		
235	63.5	41.8	81.8
272	13.1	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 134.511 ng
RT: 15.922 min Scan# 2182
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:330 Resp: 1542341

Ion	Ratio	Lower	Upper
330	100		
332	96.0	77.6	116.4
141	30.6	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 65.478 ng
RT: 12.828 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:196 Resp: 1148545

Ion	Ratio	Lower	Upper
196	100		
198	95.6	76.6	115.0
200	30.9	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 64.282 ng
RT: 12.904 min Scan# 1669
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:196 Resp: 1233147

Ion	Ratio	Lower	Upper
196	100		
198	95.4	76.5	114.7
97	49.0	38.7	58.1
132	28.7	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 114.626 ng
RT: 13.016 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:172 Resp: 7675412

Ion	Ratio	Lower	Upper
172	100		
171	36.2	28.3	42.5
170	24.5	19.0	28.4







#46
1,1'-Biphenyl
Concen: 59.291 ng
RT: 13.234 min Scan# 1725
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:154 Resp: 4312903

Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.3	60.3
76	13.4	0.0	32.9







#47
2-Chloronaphthalene
Concen: 59.452 ng
RT: 13.275 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 3269176

Ion	Ratio	Lower	Upper
162	100		
127	39.2	31.5	47.3
164	33.2	26.6	39.8







#48
2-Nitroaniline
Concen: 68.533 ng
RT: 13.475 min Scan# 1766
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion: 65 Resp: 890207

Ion	Ratio	Lower	Upper
65	100		
92	76.3	62.1	93.1
138	123.9	99.5	149.3







#49
Acenaphthylene
Concen: 62.024 ng
RT: 14.128 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 5025011

Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.0	24.0
153	13.1	10.6	15.8







#50
Dimethylphthalate
Concen: 60.542 ng
RT: 13.863 min Scan# 1832
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:163 Resp: 3997694

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.0
164	10.4	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 65.996 ng
RT: 13.981 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 887706

Ion	Ratio	Lower	Upper
165	100		
63	44.1	37.7	56.5
89	49.9	41.9	62.9







#52
Acenaphthene
Concen: 60.450 ng
RT: 14.475 min Scan# 1936
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:154 Resp: 3153799

Ion	Ratio	Lower	Upper
154	100		
153	111.1	90.5	135.7
152	53.5	43.1	64.7







#53
3-Nitroaniline
Concen: 67.123 ng
RT: 14.316 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:138 Resp: 963434

Ion	Ratio	Lower	Upper
138	100		
108	9.7	8.2	12.2
92	107.7	86.7	130.1







#54
2,4-Dinitrophenol
Concen: 73.093 ng
RT: 14.522 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:184 Resp: 523590

Ion	Ratio	Lower	Upper
184	100		
63	55.6	42.1	63.1
154	56.9	44.6	67.0







#55
Dibenzofuran
Concen: 59.602 ng
RT: 14.816 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:168 Resp: 5051493

Ion	Ratio	Lower	Upper
168	100		
139	36.6	29.1	43.7
169	13.4	10.8	16.2







#56
4-Nitrophenol
Concen: 67.086 ng
RT: 14.640 min Scan# 1964
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:139 Resp: 831765

Ion	Ratio	Lower	Upper
139	100		
109	63.5	43.3	83.3
65	82.1	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 69.346 ng
RT: 14.775 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 1211692

Ion	Ratio	Lower	Upper
165	100		
63	35.7	29.0	43.6
89	66.4	54.2	81.4
182	2.8	2.3	3.5







#58
Fluorene
Concen: 59.702 ng
RT: 15.469 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:166 Resp: 3951671

Ion	Ratio	Lower	Upper
166	100		
165	96.9	77.6	116.4
167	13.7	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 63.439 ng
RT: 15.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:232 Resp: 1055860

Ion	Ratio	Lower	Upper
232	100		
131	44.7	35.9	53.9
130	2.3	1.8	2.6
166	30.8	24.2	36.2







#60
Diethylphthalate
Concen: 61.486 ng
RT: 15.245 min Scan# 2067
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:149 Resp: 4030341

Ion	Ratio	Lower	Upper
149	100		
177	23.0	17.7	26.5
150	12.7	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 59.370 ng
RT: 15.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:204 Resp: 1934265

Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.4	39.6
141	57.5	45.8	68.8







#62
4-Nitroaniline
Concen: 68.824 ng
RT: 15.498 min Scan# 2110
Delta R.T. 0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:138 Resp: 1025149

Ion	Ratio	Lower	Upper
138	100		
92	47.5	28.7	68.7
108	67.1	49.5	89.5







#63
Azobenzene
Concen: 62.340 ng
RT: 15.769 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 3829148

Ion	Ratio	Lower	Upper
77	100		
182	28.8	7.7	47.7
105	23.3	2.3	42.3
51	22.9	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.210 min Scan# 2401
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:188 Resp: 1759587

Ion	Ratio	Lower	Upper
188	100		
94	9.5	8.1	12.1
80	10.5	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 68.563 ng
RT: 15.557 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:198 Resp: 698260

Ion	Ratio	Lower	Upper
198	100		
51	30.2	14.7	54.7
105	39.9	23.8	63.8







#66
n-Nitrosodiphenylamine
Concen: 60.776 ng
RT: 15.687 min Scan# 2142
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:169 Resp: 3334914

Ion	Ratio	Lower	Upper
169	100		
168	66.2	52.9	79.3
167	36.2	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 61.311 ng
RT: 16.381 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:248 Resp: 1211468

Ion	Ratio	Lower	Upper
248	100		
250	97.4	78.4	117.6
141	69.6	59.4	89.0







#68
Hexachlorobenzene
Concen: 61.320 ng
RT: 16.498 min Scan# 2280
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:284 Resp: 1425714

Ion	Ratio	Lower	Upper
284	100		
142	34.1	25.4	38.2
249	30.9	23.5	35.3







#70
Pentachlorophenol
Concen: 66.402 ng
RT: 16.857 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:266 Resp: 1002726

Ion	Ratio	Lower	Upper
266	100		
268	64.2	51.3	76.9
264	63.1	50.6	76.0







#71
Phenanthrene
Concen: 59.103 ng
RT: 17.251 min Scan# 2408
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:178 Resp: 5863198

Ion	Ratio	Lower	Upper
178	100		
176	19.8	15.6	23.4
179	15.9	12.5	18.7







#72
Anthracene
Concen: 60.532 ng
RT: 17.345 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:178 Resp: 5754537

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.2	22.8
179	15.6	12.6	18.8







#73
Carbazole
Concen: 61.810 ng
RT: 17.622 min Scan# 2471
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:167 Resp: 5598533

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.1	25.7
139	13.1	10.2	15.4







#74
Di-n-butylphthalate
Concen: 64.761 ng
RT: 18.216 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 6844468

Ion	Ratio	Lower	Upper
149	100		
150	9.6	7.6	11.4
104	4.9	3.9	5.9







#75
Fluoranthene
Concen: 61.042 ng
RT: 19.333 min Scan# 2762
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:202 Resp: 6849338

Ion	Ratio	Lower	Upper
202	100		
101	12.9	0.0	32.5
203	18.2	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.669 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:240 Resp: 1817869

Ion	Ratio	Lower	Upper
240	100		
120	11.0	9.2	13.8
236	25.7	20.4	30.6







#77
Benzidine
Concen: 68.561 ng
RT: 19.528 min Scan# 2795
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:184 Resp: 1518742

Ion	Ratio	Lower	Upper
184	100		
185	14.8	11.8	17.6
183	11.7	9.1	13.7







#78
Pyrene
Concen: 61.052 ng
RT: 19.710 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:202 Resp: 7049649

Ion	Ratio	Lower	Upper
202	100		
200	21.9	17.1	25.7
203	18.2	14.2	21.4







#79
Terphenyl-d14
Concen: 115.330 ng
RT: 19.933 min Scan# 2864
Delta R.T. 0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:244 Resp:10595349

Ion	Ratio	Lower	Upper
244	100		
212	7.5	5.7	8.5
122	12.3	9.6	14.4







#80
Butylbenzylphthalate
Concen: 71.649 ng
RT: 20.663 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 3170269

Ion	Ratio	Lower	Upper
149	100		
91	69.6	55.3	82.9
206	21.0	16.1	24.1







#81
Benzo(a)anthracene
Concen: 61.500 ng
RT: 21.645 min Scan# 3155
Delta R.T. 0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:228 Resp: 7010330

Ion	Ratio	Lower	Upper
228	100		
226	27.8	21.8	32.8
229	20.4	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 71.531 ng
RT: 21.569 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 2547922

Ion	Ratio	Lower	Upper
252	100		
254	65.2	51.4	77.2
126	11.1	9.3	13.9







#83
Chrysene
Concen: 60.535 ng
RT: 21.716 min Scan# 3167
Delta R.T. 0.012 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:228 Resp: 6653811

Ion	Ratio	Lower	Upper
228	100		
226	31.0	24.5	36.7
229	20.3	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 71.363 ng
RT: 21.586 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 4788368

Ion	Ratio	Lower	Upper
149	100		
167	28.9	22.5	33.7
279	5.3	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 72.964 ng
RT: 22.886 min Scan# 3366
Delta R.T. 0.018 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:149 Resp: 7825427

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.3	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.074 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:264 Resp: 2003229

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.5	27.7
265	21.8	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 62.364 ng m
RT: 28.951 min Scan# 4397
Delta R.T. 0.023 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:276 Resp: 8887645

Ion	Ratio	Lower	Upper
276	100		
138	17.7	13.7	20.5
277	20.7	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 59.798 ng
RT: 23.998 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 7606878

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.2	25.8
125	10.5	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 60.696 ng
RT: 24.068 min Scan# 3567
Delta R.T. 0.018 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:252 Resp: 7515459

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	10.5	9.1	13.7







#90
Benzo(a)pyrene
Concen: 62.275 ng
RT: 24.915 min Scan# 3
Delta R.T. 0.023 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 6723161

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	11.8	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 61.510 ng
RT: 29.056 min Scan# 4415
Delta R.T. 0.047 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Tgt Ion:278 Resp: 7370514

Ion	Ratio	Lower	Upper
278	100		
139	16.8	13.0	19.4
279	23.8	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 61.593 ng
RT: 30.156 min Scan# 4
Delta R.T. 0.029 min
Lab File: BP024166.D
Acq: 12 Mar 2025 20:21

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:276 Resp: 7489359

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.2	28.8
138	22.5	17.4	26.2



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024167.D
Acq On : 12 Mar 2025 21:02
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Mar 13 04:10:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.769	152	364999	20.000	ng		0.00
21) Naphthalene-d8	10.551	136	1468727	20.000	ng		0.00
39) Acenaphthene-d10	14.404	164	888748	20.000	ng		0.00
64) Phenanthrene-d10	17.210	188	1690241	20.000	ng		0.00
76) Chrysene-d12	21.674	240	1785678	20.000	ng		0.02
86) Perylene-d12	25.080	264	1974592	20.000	ng		0.02
System Monitoring Compounds							
5) 2-Fluorophenol	5.381	112	3803317	162.591	ng		0.00
7) Phenol-d6	6.957	99	5104893	169.796	ng		0.00
23) Nitrobenzene-d5	8.922	82	4314346	163.878	ng		0.00
42) 2,4,6-Tribromophenol	15.922	330	2025211	186.035	ng		0.00
45) 2-Fluorobiphenyl	13.016	172	9471932	148.993	ng		0.00
79) Terphenyl-d14	19.939	244	13193577	146.201	ng		0.02
Target Compounds							Qvalue
2) 1,4-Dioxane	3.317	88	719424	78.401	ng		99
3) Pyridine	3.716	79	2086180m	87.444	ng		
4) n-Nitrosodimethylamine	3.622	42	672598	79.479	ng		98
6) Aniline	7.110	93	2112268	76.717	ng		98
8) 2-Chlorophenol	7.346	128	2039550	82.376	ng		99
9) Benzaldehyde	6.922	77	842969	61.412	ng		98
10) Phenol	6.987	94	2556863	84.573	ng		99
11) bis(2-Chloroethyl)ether	7.199	93	1979282	83.121	ng		100
12) 1,3-Dichlorobenzene	7.663	146	2110081	77.577	ng		99
13) 1,4-Dichlorobenzene	7.810	146	2159502	78.507	ng		99
14) 1,2-Dichlorobenzene	8.122	146	2060216	78.042	ng		99
15) Benzyl Alcohol	8.010	79	1663098	90.370	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.299	45	1950889	84.398	ng		99
17) 2-Methylphenol	8.216	107	1619006	87.877	ng		99
18) Hexachloroethane	8.846	117	773060	78.037	ng		97
19) n-Nitroso-di-n-propyla...	8.575	70	1455297	90.027	ng		99
20) 3+4-Methylphenols	8.546	107	2240243	89.384	ng		98
22) Acetophenone	8.587	105	3007941	79.697	ng	#	99
24) Nitrobenzene	8.969	77	2118548	81.768	ng		98
25) Isophorone	9.493	82	3813082	88.570	ng		98
26) 2-Nitrophenol	9.669	139	1129770	94.646	ng		97
27) 2,4-Dimethylphenol	9.734	122	1361694	86.904	ng		99
28) bis(2-Chloroethoxy)met...	9.963	93	2576867	82.625	ng		100
29) 2,4-Dichlorophenol	10.204	162	1875825	87.724	ng		99
30) 1,2,4-Trichlorobenzene	10.416	180	1946046	79.431	ng		99
31) Naphthalene	10.604	128	6278869	79.507	ng		100
33) 4-Chloroaniline	10.716	127	2336710	83.990	ng		99
34) Hexachlorobutadiene	10.887	225	1124081	79.213	ng		99
35) Caprolactam	11.528	113	654999	101.156	ng		97
36) 4-Chloro-3-methylphenol	11.845	107	2042687	92.924	ng		99
37) 2-Methylnaphthalene	12.210	142	4316966	82.244	ng		99
38) 1-Methylnaphthalene	12.434	142	4227550	82.791	ng		99
40) 1,2,4,5-Tetrachloroben...	12.581	216	2120251	77.644	ng		100
41) Hexachlorocyclopentadiene	12.563	237	788854	77.633	ng		99
43) 2,4,6-Trichlorophenol	12.822	196	1443797	86.697	ng		99
44) 2,4,5-Trichlorophenol	12.904	196	1575527	86.506	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024167.D
Acq On : 12 Mar 2025 21:02
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Mar 13 04:10:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
46) 1,1'-Biphenyl	13.228	154	5406424	78.284	ng		99
47) 2-Chloronaphthalene	13.269	162	4094698	78.432	ng		100
48) 2-Nitroaniline	13.475	65	1150637	93.303	ng		98
49) Acenaphthylene	14.122	152	6343598	82.473	ng		99
50) Dimethylphthalate	13.857	163	5115780	81.603	ng		100
51) 2,6-Dinitrotoluene	13.975	165	1146015	89.740	ng		97
52) Acenaphthene	14.469	154	3985981	80.473	ng		99
53) 3-Nitroaniline	14.310	138	1277036	93.713	ng		98
55) Dibenzofuran	14.810	168	6287848	78.144	ng		99
56) 4-Nitrophenol	14.639	139	1110260	94.320	ng		99
57) 2,4-Dinitrotoluene	14.775	165	1584658	95.524	ng		97
58) Fluorene	15.469	166	4983161	79.298	ng		100
59) 2,3,4,6-Tetrachlorophenol	15.045	232	1371535	86.797	ng		100
60) Diethylphthalate	15.245	149	5221159	83.898	ng		98
61) 4-Chlorophenyl-phenyle...	15.463	204	2453010	79.305	ng		99
62) 4-Nitroaniline	15.498	138	1346814	95.238	ng		95
63) Azobenzene	15.763	77	4851665	83.196	ng		99
65) 4,6-Dinitro-2-methylph...	15.557	198	924643	94.517	ng		97
66) n-Nitrosodiphenylamine	15.686	169	4231796	80.285	ng		99
67) 4-Bromophenyl-phenylether	16.381	248	1565037	82.455	ng		94
68) Hexachlorobenzene	16.498	284	1841149	82.437	ng		99
70) Pentachlorophenol	16.851	266	1334567	92.003	ng		100
71) Phenanthrene	17.251	178	7450840	78.188	ng		99
72) Anthracene	17.345	178	7420971	81.264	ng		99
73) Carbazole	17.628	167	7209600	82.862	ng		98
74) Di-n-butylphthalate	18.216	149	8819030	86.868	ng		99
75) Fluoranthene	19.339	202	8868502	82.280	ng		99
77) Benzidine	19.533	184	1583965	72.794	ng		100
78) Pyrene	19.722	202	9100075	80.230	ng		99
80) Butylbenzylphthalate	20.669	149	4117551	94.735	ng		97
81) Benzo(a)anthracene	21.657	228	9075269	81.050	ng		98
82) 3,3'-Dichlorobenzidine	21.574	252	3216378	91.925	ng		99
83) Chrysene	21.721	228	8635685	79.981	ng		99
84) Bis(2-ethylhexyl)phtha...	21.592	149	6077881	92.215	ng	#	98
85) Di-n-octyl phthalate	22.898	149	10239008	97.189	ng		100
87) Indeno(1,2,3-cd)pyrene	28.962	276	11842118	84.301	ng		100
88) Benzo(b)fluoranthene	24.004	252	9984031	79.623	ng		98
89) Benzo(k)fluoranthene	24.080	252	9728938	79.712	ng		98
90) Benzo(a)pyrene	24.921	252	8860063	83.259	ng		99
91) Dibenzo(a,h)anthracene	29.039	278	9828241	83.210	ng		98
92) Benzo(g,h,i)perylene	30.180	276	9940803	82.940	ng		98


(#) = qualifier out of range (m) = manual integration (+) = signals summed


ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024167.D
Acq On : 12 Mar 2025 21:02
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Mar 13 04:10:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:01:12 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 364999

Ion	Ratio	Lower	Upper
152	100		
150	157.3	125.5	188.3
115	59.3	46.3	69.5







#2
1,4-Dioxane
Concen: 78.401 ng
RT: 3.317 min Scan# 39
Delta R.T. -0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion: 88 Resp: 719424

Ion	Ratio	Lower	Upper
88	100		
58	62.6	49.8	74.6
43	21.2	17.3	25.9







#3
Pyridine
Concen: 87.444 ng m
RT: 3.716 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 2086180

Ion	Ratio	Lower	Upper
79	100		
52	54.6	43.8	65.8
51	26.9	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 79.479 ng
RT: 3.622 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion: 42 Resp: 672598

Ion	Ratio	Lower	Upper
42	100		
74	177.9	139.8	209.6
44	6.7	5.3	7.9







#5
2-Fluorophenol
Concen: 162.591 ng
RT: 5.381 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:112 Resp: 3803317

Ion	Ratio	Lower	Upper
112	100		
64	51.2	41.3	61.9
63	25.7	20.6	31.0







#6
Aniline
Concen: 76.717 ng
RT: 7.110 min Scan# 684
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion: 93 Resp: 2112268

Ion	Ratio	Lower	Upper
93	100		
66	36.1	28.2	42.2
65	18.1	14.0	21.0







#7
Phenol-d6
Concen: 169.796 ng
RT: 6.957 min Scan# 6
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 99 Resp: 5104893

Ion	Ratio	Lower	Upper
99	100		
42	12.6	10.6	16.0
71	29.1	23.3	34.9







#8
2-Chlorophenol
Concen: 82.376 ng
RT: 7.346 min Scan# 724
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:128 Resp: 2039550

Ion	Ratio	Lower	Upper
128	100		
130	32.4	12.0	52.0
64	41.0	21.6	61.6







#9
Benzaldehyde
Concen: 61.412 ng
RT: 6.922 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 842969

Ion	Ratio	Lower	Upper
77	100		
105	102.7	81.4	121.4
106	98.8	81.1	121.1







#10
Phenol
Concen: 84.573 ng
RT: 6.987 min Scan# 663
Delta R.T. 0.012 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion: 94 Resp: 2556863

Ion	Ratio	Lower	Upper
94	100		
65	24.4	4.9	44.9
66	31.0	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 83.121 ng
RT: 7.199 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 93 Resp: 1979282

Ion	Ratio	Lower	Upper
93	100		
63	67.3	47.4	87.4
95	32.3	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 77.577 ng
RT: 7.663 min Scan# 778
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:146 Resp: 2110081

Ion	Ratio	Lower	Upper
146	100		
148	64.8	51.6	77.4
75	27.8	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 78.507 ng
RT: 7.810 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:146 Resp: 2159502

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.9	77.9
111	38.9	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 78.042 ng
RT: 8.122 min Scan# 856
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:146 Resp: 2060216

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.2
111	42.5	33.2	49.8







#15
Benzyl Alcohol
Concen: 90.370 ng
RT: 8.010 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 1663098

Ion	Ratio	Lower	Upper
79	100		
108	82.8	66.5	99.7
77	62.9	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 84.398 ng
RT: 8.299 min Scan# 886
Delta R.T. -0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion: 45 Resp: 1950889

Ion	Ratio	Lower	Upper
45	100		
77	18.8	0.0	39.0
79	15.0	0.0	35.6







#17
2-Methylphenol
Concen: 87.877 ng
RT: 8.216 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:107 Resp: 1619006

Ion	Ratio	Lower	Upper
107	100		
108	112.5	90.0	135.0
77	41.7	34.0	51.0
79	40.2	32.7	49.1







#18
Hexachloroethane
Concen: 78.037 ng
RT: 8.846 min Scan# 979
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:117 Resp: 773060

Ion	Ratio	Lower	Upper
117	100		
119	96.8	77.4	116.0
201	101.0	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 90.027 ng
RT: 8.575 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 70 Resp: 1455297

Ion	Ratio	Lower	Upper
70	100		
42	44.5	36.0	54.0
101	10.6	8.3	12.5
130	24.4	18.6	28.0







#20
3+4-Methylphenols
Concen: 89.384 ng
RT: 8.546 min Scan# 928
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:107 Resp: 2240243

Ion	Ratio	Lower	Upper
107	100		
108	91.2	69.7	109.7
77	30.6	10.3	50.3
79	27.5	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.551 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:136 Resp: 1468727

Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	6.5	5.3	7.9
68	6.1	5.0	7.4







#22
Acetophenone
Concen: 79.697 ng
RT: 8.587 min Scan# 935
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:105 Resp: 3007941

Ion	Ratio	Lower	Upper
105	100		
71	4.9	2.4	3.6#
51	20.0	16.2	24.2
120	23.8	19.0	28.6







#23
Nitrobenzene-d5
Concen: 163.878 ng
RT: 8.922 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 4314346

Ion	Ratio	Lower	Upper
82	100		
128	47.4	37.4	56.0
54	40.4	32.5	48.7







#24
Nitrobenzene
Concen: 81.768 ng
RT: 8.969 min Scan# 1000
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion: 77 Resp: 2118548

Ion	Ratio	Lower	Upper
77	100		
123	51.4	39.4	59.2
65	12.5	9.9	14.9







#25
Isophorone
Concen: 88.570 ng
RT: 9.493 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 3813082

Ion	Ratio	Lower	Upper
82	100		
95	6.8	5.3	7.9
138	19.5	14.8	22.2







#26
2-Nitrophenol
Concen: 94.646 ng
RT: 9.669 min Scan# 1119
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:139 Resp: 1129770

Ion	Ratio	Lower	Upper
139	100		
109	25.7	21.7	32.5
65	42.6	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 86.904 ng
RT: 9.734 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:122 Resp: 1361694

Ion	Ratio	Lower	Upper
122	100		
107	112.4	89.3	133.9
121	59.0	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 82.625 ng
RT: 9.963 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion: 93 Resp: 2576867

Ion	Ratio	Lower	Upper
93	100		
95	32.6	25.9	38.9
123	11.8	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 87.724 ng
RT: 10.204 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 1875825

Ion	Ratio	Lower	Upper
162	100		
164	64.5	44.2	84.2
98	36.0	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 79.431 ng
RT: 10.416 min Scan# 1246
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:180 Resp: 1946046

Ion	Ratio	Lower	Upper
180	100		
182	96.3	76.1	114.1
145	30.0	24.2	36.2







#31
Naphthalene
Concen: 79.507 ng
RT: 10.604 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:128 Resp: 6278869

Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.7	13.1
127	13.2	10.4	15.6







#33
4-Chloroaniline
Concen: 83.990 ng
RT: 10.716 min Scan# 1297
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:127 Resp: 2336710

Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.8	38.6
65	26.8	21.2	31.8
92	18.3	14.4	21.6







#34
Hexachlorobutadiene
Concen: 79.213 ng
RT: 10.887 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:225 Resp: 1124081

Ion	Ratio	Lower	Upper
225	100		
223	63.9	49.9	74.9
227	64.6	52.2	78.2







#35
Caprolactam
Concen: 101.156 ng
RT: 11.528 min Scan# 1435
Delta R.T. 0.029 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:113 Resp: 654999

Ion	Ratio	Lower	Upper
113	100		
55	130.2	113.6	153.6
56	97.1	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 92.924 ng
RT: 11.845 min Scan# 14
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:107 Resp: 2042687

Ion	Ratio	Lower	Upper
107	100		
144	26.7	21.6	32.4
142	82.4	65.4	98.0







#37
2-Methylnaphthalene
Concen: 82.244 ng
RT: 12.210 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:142 Resp: 4316966

Ion	Ratio	Lower	Upper
142	100		
141	88.5	70.3	105.5
115	33.0	26.0	39.0







#38
1-Methylnaphthalene
Concen: 82.791 ng
RT: 12.434 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:142 Resp: 4227550

Ion	Ratio	Lower	Upper
142	100		
141	91.6	73.9	110.9
116	3.7	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.404 min Scan# 1924
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:164 Resp: 888748

Ion	Ratio	Lower	Upper
164	100		
162	100.5	80.6	121.0
160	44.1	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 77.644 ng
RT: 12.581 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:216 Resp: 2120251

Ion	Ratio	Lower	Upper
216	100		
214	78.8	62.8	94.2
179	21.0	16.8	25.2
108	18.3	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 77.633 ng
RT: 12.563 min Scan# 1611
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:237 Resp: 788854

Ion	Ratio	Lower	Upper
237	100		
235	62.5	41.8	81.8
272	14.0	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 186.035 ng
RT: 15.922 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:330 Resp: 2025211

Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.6	116.4
141	29.9	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 86.697 ng
RT: 12.822 min Scan# 1655
Delta R.T. -0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:196 Resp: 1443797

Ion	Ratio	Lower	Upper
196	100		
198	95.3	76.6	115.0
200	30.8	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 86.506 ng
RT: 12.904 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:196 Resp: 1575527

Ion	Ratio	Lower	Upper
196	100		
198	95.4	76.5	114.7
97	46.4	38.7	58.1
132	27.8	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 148.993 ng
RT: 13.016 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:172 Resp: 9471932

Ion	Ratio	Lower	Upper
172	100		
171	36.4	28.3	42.5
170	24.6	19.0	28.4







#46
1,1'-Biphenyl
Concen: 78.284 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:154 Resp: 5406424

Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.3	60.3
76	13.4	0.0	32.9







#47
2-Chloronaphthalene
Concen: 78.432 ng
RT: 13.269 min Scan# 1731
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:162 Resp: 4094698

Ion	Ratio	Lower	Upper
162	100		
127	39.6	31.5	47.3
164	33.0	26.6	39.8







#48
2-Nitroaniline
Concen: 93.303 ng
RT: 13.475 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 65 Resp: 1150637

Ion	Ratio	Lower	Upper
65	100		
92	78.6	62.1	93.1
138	127.1	99.5	149.3







#49
Acenaphthylene
Concen: 82.473 ng
RT: 14.122 min Scan# 1876
Delta R.T. -0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:152 Resp: 6343598

Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.0	24.0
153	13.5	10.6	15.8







#50
Dimethylphthalate
Concen: 81.603 ng
RT: 13.857 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:163 Resp: 5115780

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.0
164	10.5	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 89.740 ng
RT: 13.975 min Scan# 1851
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:165 Resp: 1146015

Ion	Ratio	Lower	Upper
165	100		
63	44.9	37.7	56.5
89	50.2	41.9	62.9







#52
Acenaphthene
Concen: 80.473 ng
RT: 14.469 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:154 Resp: 3985981

Ion	Ratio	Lower	Upper
154	100		
153	111.1	90.5	135.7
152	53.5	43.1	64.7







#53
3-Nitroaniline
Concen: 93.713 ng
RT: 14.310 min Scan# 1908
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:138 Resp: 1277036

Ion	Ratio	Lower	Upper
138	100		
108	10.2	8.2	12.2
92	110.4	86.7	130.1







#55
Dibenzofuran
Concen: 78.144 ng
RT: 14.810 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:168 Resp: 6287848

Ion	Ratio	Lower	Upper
168	100		
139	37.4	29.1	43.7
169	13.7	10.8	16.2







#56
4-Nitrophenol
Concen: 94.320 ng
RT: 14.639 min Scan# 1964
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:139 Resp: 1110260

Ion	Ratio	Lower	Upper
139	100		
109	63.5	43.3	83.3
65	80.2	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 95.524 ng
RT: 14.775 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 1584658

Ion	Ratio	Lower	Upper
165	100		
63	34.8	29.0	43.6
89	65.7	54.2	81.4
182	2.9	2.3	3.5







#58
Fluorene
Concen: 79.298 ng
RT: 15.469 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:166 Resp: 4983161

Ion	Ratio	Lower	Upper
166	100		
165	96.7	77.6	116.4
167	13.7	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 86.797 ng
RT: 15.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:232 Resp: 1371535

Ion	Ratio	Lower	Upper
232	100		
131	44.8	35.9	53.9
130	2.2	1.8	2.6
166	30.5	24.2	36.2







#60
Diethylphthalate
Concen: 83.898 ng
RT: 15.245 min Scan# 2067
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:149 Resp: 5221159

Ion	Ratio	Lower	Upper
149	100		
177	23.3	17.7	26.5
150	12.7	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 79.305 ng
RT: 15.463 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:204 Resp: 2453010

Ion	Ratio	Lower	Upper
204	100		
206	32.8	26.4	39.6
141	56.8	45.8	68.8







#62
4-Nitroaniline
Concen: 95.238 ng
RT: 15.498 min Scan# 2110
Delta R.T. 0.012 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:138 Resp: 1346814

Ion	Ratio	Lower	Upper
138	100		
92	46.0	28.7	68.7
108	65.1	49.5	89.5







#63
Azobenzene
Concen: 83.196 ng
RT: 15.763 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 4851665

Ion	Ratio	Lower	Upper
77	100		
182	26.5	7.7	47.7
105	22.7	2.3	42.3
51	23.3	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.210 min Scan# 2401
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:188 Resp: 1690241

Ion	Ratio	Lower	Upper
188	100		
94	9.3	8.1	12.1
80	10.1	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 94.517 ng
RT: 15.557 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:198 Resp: 924643

Ion	Ratio	Lower	Upper
198	100		
51	32.8	14.7	54.7
105	42.5	23.8	63.8







#66
n-Nitrosodiphenylamine
Concen: 80.285 ng
RT: 15.686 min Scan# 2142
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:169 Resp: 4231796

Ion	Ratio	Lower	Upper
169	100		
168	66.6	52.9	79.3
167	36.7	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 82.455 ng
RT: 16.381 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:248 Resp: 1565037

Ion	Ratio	Lower	Upper
248	100		
250	97.1	78.4	117.6
141	64.3	59.4	89.0







#68
Hexachlorobenzene
Concen: 82.437 ng
RT: 16.498 min Scan# 2280
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:284 Resp: 1841149

Ion	Ratio	Lower	Upper
284	100		
142	31.7	25.4	38.2
249	30.0	23.5	35.3







#70
Pentachlorophenol
Concen: 92.003 ng
RT: 16.851 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:266 Resp: 1334567

Ion	Ratio	Lower	Upper
266	100		
268	64.1	51.3	76.9
264	63.4	50.6	76.0







#71
Phenanthrene
Concen: 78.188 ng
RT: 17.251 min Scan# 2408
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:178 Resp: 7450840

Ion	Ratio	Lower	Upper
178	100		
176	20.0	15.6	23.4
179	16.1	12.5	18.7







#72
Anthracene
Concen: 81.264 ng
RT: 17.345 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:178 Resp: 7420971

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.2	22.8
179	16.0	12.6	18.8







#73
Carbazole
Concen: 82.862 ng
RT: 17.628 min Scan# 2472
Delta R.T. 0.012 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:167 Resp: 7209600

Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.1	25.7
139	13.6	10.2	15.4







#74
Di-n-butylphthalate
Concen: 86.868 ng
RT: 18.216 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 8819030

Ion	Ratio	Lower	Upper
149	100		
150	10.0	7.6	11.4
104	5.1	3.9	5.9







#75
Fluoranthene
Concen: 82.280 ng
RT: 19.339 min Scan# 2763
Delta R.T. 0.012 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:202 Resp: 8868502

Ion	Ratio	Lower	Upper
202	100		
101	13.0	0.0	32.5
203	18.5	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.674 min Scan# 3
Delta R.T. 0.017 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:240 Resp: 1785678

Ion	Ratio	Lower	Upper
240	100		
120	11.7	9.2	13.8
236	25.6	20.4	30.6







#77
Benzidine
Concen: 72.794 ng
RT: 19.533 min Scan# 2796
Delta R.T. 0.012 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:184 Resp: 1583965

Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.8	17.6
183	11.7	9.1	13.7







#78
Pyrene
Concen: 80.230 ng
RT: 19.722 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:202 Resp: 9100075

Ion	Ratio	Lower	Upper
202	100		
200	22.3	17.1	25.7
203	18.3	14.2	21.4







#79
Terphenyl-d14
Concen: 146.201 ng
RT: 19.939 min Scan# 2865
Delta R.T. 0.017 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:244 Resp:13193577

Ion	Ratio	Lower	Upper
244	100		
212	8.0	5.7	8.5
122	12.2	9.6	14.4







#80
Butylbenzylphthalate
Concen: 94.735 ng
RT: 20.669 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 4117551

Ion	Ratio	Lower	Upper
149	100		
91	66.9	55.3	82.9
206	22.2	16.1	24.1







#81
Benzo(a)anthracene
Concen: 81.050 ng
RT: 21.657 min Scan# 3157
Delta R.T. 0.023 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:228 Resp: 9075269

Ion	Ratio	Lower	Upper
228	100		
226	28.4	21.8	32.8
229	20.5	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 91.925 ng
RT: 21.574 min Scan# 3
Delta R.T. 0.023 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 3216378

Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.4	77.2
126	10.8	9.3	13.9







#83
Chrysene
Concen: 79.981 ng
RT: 21.721 min Scan# 3168
Delta R.T. 0.018 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:228 Resp: 8635685

Ion	Ratio	Lower	Upper
228	100		
226	31.2	24.5	36.7
229	20.6	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 92.215 ng
RT: 21.592 min Scan# 3
Delta R.T. 0.023 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 6077881

Ion	Ratio	Lower	Upper
149	100		
167	29.2	22.5	33.7
279	5.5	3.6	5.4#







#85
Di-n-octyl phthalate
Concen: 97.189 ng
RT: 22.898 min Scan# 3368
Delta R.T. 0.029 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:149 Resp:10239008

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.1	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.080 min Scan# 3
Delta R.T. 0.023 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:264 Resp: 1974592

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.5	27.7
265	22.3	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 84.301 ng
RT: 28.962 min Scan# 4399
Delta R.T. 0.035 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:276 Resp:11842118

Ion	Ratio	Lower	Upper
276	100		
138	16.9	13.7	20.5
277	25.4	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 79.623 ng
RT: 24.004 min Scan# 3
Delta R.T. 0.023 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 9984031

Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.2	25.8
125	11.1	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 79.712 ng
RT: 24.080 min Scan# 3569
Delta R.T. 0.029 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:252 Resp: 9728938

Ion	Ratio	Lower	Upper
252	100		
253	22.6	17.7	26.5
125	10.5	9.1	13.7







#90
Benzo(a)pyrene
Concen: 83.259 ng
RT: 24.921 min Scan# 3
Delta R.T. 0.029 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 8860063

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.4	26.0
125	11.8	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 83.210 ng
RT: 29.039 min Scan# 4412
Delta R.T. 0.029 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Tgt Ion:278 Resp: 9828241

Ion	Ratio	Lower	Upper
278	100		
139	17.1	13.0	19.4
279	24.3	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 82.940 ng
RT: 30.180 min Scan# 4
Delta R.T. 0.053 min
Lab File: BP024167.D
Acq: 12 Mar 2025 21:02

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:276 Resp: 9940803

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.2	28.8
138	23.0	17.4	26.2



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	399821	20.000	ng	0.00
21) Naphthalene-d8	10.552	136	1631941	20.000	ng	0.00
39) Acenaphthene-d10	14.404	164	966525	20.000	ng	0.00
64) Phenanthrene-d10	17.204	188	1810172	20.000	ng	0.00
76) Chrysene-d12	21.669	240	1903114	20.000	ng	0.01
86) Perylene-d12	25.069	264	2127367	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	1959434	78.230	ng	0.00
7) Phenol-d6	6.952	99	2636154	78.704	ng	0.00
23) Nitrobenzene-d5	8.922	82	2274955	78.652	ng	0.00
42) 2,4,6-Tribromophenol	15.922	330	995466	81.766	ng	0.00
45) 2-Fluorobiphenyl	13.022	172	5105223	77.885	ng	0.00
79) Terphenyl-d14	19.933	244	7241807	78.549	ng	0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.317	88	376087	37.880	ng	99
3) Pyridine	3.717	79	1050879	38.718	ng	99
4) n-Nitrosodimethylamine	3.623	42	350385	38.646	ng	98
6) Aniline	7.111	93	1202738	39.413	ng	100
8) 2-Chlorophenol	7.346	128	1059150	39.149	ng	100
9) Benzaldehyde	6.922	77	652848	40.389	ng	99
10) Phenol	6.981	94	1314283	38.869	ng	99
11) bis(2-Chloroethyl)ether	7.199	93	1018670	38.203	ng	99
12) 1,3-Dichlorobenzene	7.664	146	1118950	38.353	ng	99
13) 1,4-Dichlorobenzene	7.805	146	1130770	37.960	ng	99
14) 1,2-Dichlorobenzene	8.122	146	1089564	38.116	ng	100
15) Benzyl Alcohol	8.011	79	848187	39.999	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.299	45	1043834	38.356	ng	99
17) 2-Methylphenol	8.216	107	840413	39.636	ng	100
18) Hexachloroethane	8.846	117	407804	38.587	ng	98
19) n-Nitroso-di-n-propyla...	8.575	70	771380	39.697	ng	99
20) 3+4-Methylphenols	8.540	107	1154401	39.629	ng	100
22) Acetophenone	8.587	105	1593044	38.611	ng	100
24) Nitrobenzene	8.963	77	1114096	39.045	ng	99
25) Isophorone	9.487	82	1967247	39.677	ng	99
26) 2-Nitrophenol	9.669	139	570971	41.892	ng	98
27) 2,4-Dimethylphenol	9.734	122	697163	39.779	ng	98
28) bis(2-Chloroethoxy)met...	9.963	93	1346656	38.608	ng	100
29) 2,4-Dichlorophenol	10.205	162	961583	40.558	ng	99
30) 1,2,4-Trichlorobenzene	10.416	180	1002302	38.264	ng	99
31) Naphthalene	10.605	128	3306535	38.223	ng	100
32) Benzoic acid	9.881	122	745543	43.618	ng	98
33) 4-Chloroaniline	10.710	127	1222820	39.499	ng	100
34) Hexachlorobutadiene	10.887	225	585728	38.854	ng	98
35) Caprolactam	11.499	113	317484	40.859	ng	97
36) 4-Chloro-3-methylphenol	11.840	107	1015423	39.702	ng	99
37) 2-Methylnaphthalene	12.210	142	2265320	38.762	ng	99
38) 1-Methylnaphthalene	12.434	142	2209719	38.874	ng	99
40) 1,2,4,5-Tetrachloroben...	12.581	216	1089303	38.988	ng	100
41) Hexachlorocyclopentadiene	12.569	237	413181	41.026	ng	100
43) 2,4,6-Trichlorophenol	12.828	196	727105	40.950	ng	99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.904	196	798076	40.985	ng	98
46) 1,1'-Biphenyl	13.228	154	2821370	38.846	ng	99
47) 2-Chloronaphthalene	13.269	162	2130879	39.006	ng	100
48) 2-Nitroaniline	13.475	65	574806	41.137	ng	99
49) Acenaphthylene	14.122	152	3245488	39.262	ng	100
50) Dimethylphthalate	13.863	163	2606961	38.548	ng	99
51) 2,6-Dinitrotoluene	13.975	165	563651	39.611	ng	98
52) Acenaphthene	14.475	154	2053476	38.881	ng	98
53) 3-Nitroaniline	14.310	138	625803	40.719	ng	99
54) 2,4-Dinitrophenol	14.516	184	322185	42.791	ng	98
55) Dibenzofuran	14.810	168	3273448	38.264	ng	100
56) 4-Nitrophenol	14.634	139	543119	42.293	ng	98
57) 2,4-Dinitrotoluene	14.769	165	772629	41.106	ng	99
58) Fluorene	15.469	166	2603595	39.061	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.040	232	685479	39.844	ng	100
60) Diethylphthalate	15.240	149	2595505	38.539	ng	100
61) 4-Chlorophenyl-phenyle...	15.469	204	1261005	38.572	ng	99
62) 4-Nitroaniline	15.493	138	664956	41.253	ng	99
63) Azobenzene	15.769	77	2497600	39.347	ng	99
65) 4,6-Dinitro-2-methylph...	15.551	198	434596	40.442	ng	95
66) n-Nitrosodiphenylamine	15.681	169	2156617	39.234	ng	99
67) 4-Bromophenyl-phenylether	16.375	248	772937	38.892	ng	98
68) Hexachlorobenzene	16.498	284	919105	39.073	ng	97
69) Atrazine	16.657	200	481443	36.287	ng	99
70) Pentachlorophenol	16.851	266	626146	41.161	ng	100
71) Phenanthrene	17.251	178	3795212	38.379	ng	99
72) Anthracene	17.351	178	3790779	39.358	ng	99
73) Carbazole	17.628	167	3684202	39.665	ng	100
74) Di-n-butylphthalate	18.222	149	4505984	40.866	ng	100
75) Fluoranthene	19.345	202	4467370	39.006	ng	99
77) Benzidine	19.534	184	827654	39.777	ng	99
78) Pyrene	19.716	202	4681182	39.565	ng	100
80) Butylbenzylphthalate	20.669	149	1991838	40.824	ng	98
81) Benzo(a)anthracene	21.651	228	4652183	39.318	ng	99
82) 3,3'-Dichlorobenzidine	21.569	252	1655655	41.750	ng	99
83) Chrysene	21.716	228	4393252	38.624	ng	99
84) Bis(2-ethylhexyl)phtha...	21.586	149	3060349	41.569	ng	99
85) Di-n-octyl phthalate	22.886	149	5007462	42.452	ng	100
87) Indeno(1,2,3-cd)pyrene	28.957	276	5699470	38.387	ng	100
88) Benzo(b)fluoranthene	23.992	252	4988392	38.891	ng	98
89) Benzo(k)fluoranthene	24.069	252	4803010	38.280	ng	99
90) Benzo(a)pyrene	24.916	252	4306060	38.607	ng	100
91) Dibenzo(a,h)anthracene	29.027	278	4752357	38.474	ng	99
92) Benzo(g,h,i)perylene	30.127	276	4782473m	38.087	ng	


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:152 Resp: 399821

Ion	Ratio	Lower	Upper
152	100		
150	155.8	125.5	188.3
115	58.8	46.3	69.5







#2
1,4-Dioxane
Concen: 37.880 ng
RT: 3.317 min Scan# 39
Delta R.T. -0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion: 88 Resp: 376087

Ion	Ratio	Lower	Upper
88	100		
58	63.2	49.8	74.6
43	21.7	17.3	25.9







#3
Pyridine
Concen: 38.718 ng
RT: 3.717 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 79 Resp: 1050879

Ion	Ratio	Lower	Upper
79	100		
52	54.2	43.8	65.8
51	26.8	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 38.646 ng
RT: 3.623 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion: 42 Resp: 350385

Ion	Ratio	Lower	Upper
42	100		
74	177.7	139.8	209.6
44	6.7	5.3	7.9







#5
2-Fluorophenol
Concen: 78.230 ng
RT: 5.381 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:112 Resp: 1959434

Ion	Ratio	Lower	Upper
112	100		
64	51.4	41.3	61.9
63	25.6	20.6	31.0







#6
Aniline
Concen: 39.413 ng
RT: 7.111 min Scan# 684
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion: 93 Resp: 1202738

Ion	Ratio	Lower	Upper
93	100		
66	35.3	28.2	42.2
65	17.8	14.0	21.0







#7
Phenol-d6
Concen: 78.704 ng
RT: 6.952 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 99 Resp: 2636154

Ion	Ratio	Lower	Upper
99	100		
42	13.1	10.6	16.0
71	29.3	23.3	34.9







#8
2-Chlorophenol
Concen: 39.149 ng
RT: 7.346 min Scan# 724
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:128 Resp: 1059150

Ion	Ratio	Lower	Upper
128	100		
130	32.1	12.0	52.0
64	41.3	21.6	61.6







#9
Benzaldehyde
Concen: 40.389 ng
RT: 6.922 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 77 Resp: 652848

Ion	Ratio	Lower	Upper
77	100		
105	103.7	81.4	121.4
106	100.8	81.1	121.1







#10
Phenol
Concen: 38.869 ng
RT: 6.981 min Scan# 662
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion: 94 Resp: 1314283

Ion	Ratio	Lower	Upper
94	100		
65	24.6	4.9	44.9
66	31.5	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 38.203 ng
RT: 7.199 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 93 Resp: 1018670

Ion	Ratio	Lower	Upper
93	100		
63	68.3	47.4	87.4
95	32.2	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 38.353 ng
RT: 7.664 min Scan# 778
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:146 Resp: 1118950

Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.6	77.4
75	28.2	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 37.960 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:146 Resp: 1130770

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.9	77.9
111	40.1	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 38.116 ng
RT: 8.122 min Scan# 856
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:146 Resp: 1089564

Ion	Ratio	Lower	Upper
146	100		
148	64.7	51.4	77.2
111	41.7	33.2	49.8







#15
Benzyl Alcohol
Concen: 39.999 ng
RT: 8.011 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 79 Resp: 848187

Ion	Ratio	Lower	Upper
79	100		
108	82.9	66.5	99.7
77	63.0	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.356 ng
RT: 8.299 min Scan# 886
Delta R.T. -0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion: 45 Resp: 1043834

Ion	Ratio	Lower	Upper
45	100		
77	18.9	0.0	39.0
79	14.7	0.0	35.6







#17
2-Methylphenol
Concen: 39.636 ng
RT: 8.216 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:107 Resp: 840413

Ion	Ratio	Lower	Upper
107	100		
108	112.5	90.0	135.0
77	41.8	34.0	51.0
79	40.5	32.7	49.1







#18
Hexachloroethane
Concen: 38.587 ng
RT: 8.846 min Scan# 979
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:117 Resp: 407804

Ion	Ratio	Lower	Upper
117	100		
119	95.5	77.4	116.0
201	97.7	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 39.697 ng
RT: 8.575 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 70 Resp: 771380

Ion	Ratio	Lower	Upper
70	100		
42	45.4	36.0	54.0
101	10.7	8.3	12.5
130	24.4	18.6	28.0







#20
3+4-Methylphenols
Concen: 39.629 ng
RT: 8.540 min Scan# 927
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:107 Resp: 1154401

Ion	Ratio	Lower	Upper
107	100		
108	89.8	69.7	109.7
77	30.5	10.3	50.3
79	26.4	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.552 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:136 Resp: 1631941

Ion	Ratio	Lower	Upper
136	100		
137	10.8	9.0	13.4
54	6.3	5.3	7.9
68	6.0	5.0	7.4







#22
Acetophenone
Concen: 38.611 ng
RT: 8.587 min Scan# 935
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:105 Resp: 1593044

Ion	Ratio	Lower	Upper
105	100		
71	2.9	2.4	3.6
51	20.0	16.2	24.2
120	23.7	19.0	28.6







#23
Nitrobenzene-d5
Concen: 78.652 ng
RT: 8.922 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 82 Resp: 2274955

Ion	Ratio	Lower	Upper
82	100		
128	46.8	37.4	56.0
54	40.5	32.5	48.7







#24
Nitrobenzene
Concen: 39.045 ng
RT: 8.963 min Scan# 999
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion: 77 Resp: 1114096

Ion	Ratio	Lower	Upper
77	100		
123	48.6	39.4	59.2
65	12.4	9.9	14.9







#25
Isophorone
Concen: 39.677 ng
RT: 9.487 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 82 Resp: 1967247

Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.3	7.9
138	18.8	14.8	22.2







#26
2-Nitrophenol
Concen: 41.892 ng
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:139 Resp: 570971

Ion	Ratio	Lower	Upper
139	100		
109	26.2	21.7	32.5
65	43.8	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 39.779 ng
RT: 9.734 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:122 Resp: 697163

Ion	Ratio	Lower	Upper
122	100		
107	114.1	89.3	133.9
121	59.2	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 38.608 ng
RT: 9.963 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion: 93 Resp: 1346656

Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.9	38.9
123	11.8	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 40.558 ng
RT: 10.205 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:162 Resp: 961583

Ion	Ratio	Lower	Upper
162	100		
164	64.9	44.2	84.2
98	35.8	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 38.264 ng
RT: 10.416 min Scan# 1246
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:180 Resp: 1002302

Ion	Ratio	Lower	Upper
180	100		
182	96.2	76.1	114.1
145	30.5	24.2	36.2







#31
Naphthalene
Concen: 38.223 ng
RT: 10.605 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:128 Resp: 3306535

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.7	13.1
127	13.2	10.4	15.6







#32
Benzoic acid
Concen: 43.618 ng
RT: 9.881 min Scan# 1155
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:122 Resp: 745543

Ion	Ratio	Lower	Upper
122	100		
105	122.4	100.3	140.3
77	86.8	69.1	109.1







#33
4-Chloroaniline
Concen: 39.499 ng
RT: 10.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:127 Resp: 1222820

Ion	Ratio	Lower	Upper
127	100		
129	32.3	25.8	38.6
65	26.9	21.2	31.8
92	18.4	14.4	21.6







#34
Hexachlorobutadiene
Concen: 38.854 ng
RT: 10.887 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:225 Resp: 585728

Ion	Ratio	Lower	Upper
225	100		
223	63.8	49.9	74.9
227	63.8	52.2	78.2







#35
Caprolactam
Concen: 40.859 ng
RT: 11.499 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:113 Resp: 317484

Ion	Ratio	Lower	Upper
113	100		
55	130.8	113.6	153.6
56	97.1	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 39.702 ng
RT: 11.840 min Scan# 1488
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:107 Resp: 1015423

Ion	Ratio	Lower	Upper
107	100		
144	26.4	21.6	32.4
142	82.2	65.4	98.0







#37
2-Methylnaphthalene
Concen: 38.762 ng
RT: 12.210 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:142 Resp: 2265320

Ion	Ratio	Lower	Upper
142	100		
141	89.0	70.3	105.5
115	32.5	26.0	39.0







#38
1-Methylnaphthalene
Concen: 38.874 ng
RT: 12.434 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:142 Resp: 2209719

Ion	Ratio	Lower	Upper
142	100		
141	91.7	73.9	110.9
116	3.6	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.404 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:164 Resp: 966525

Ion	Ratio	Lower	Upper
164	100		
162	100.8	80.6	121.0
160	44.9	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.988 ng
RT: 12.581 min Scan# 1614
Delta R.T. -0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:216 Resp: 1089303

Ion	Ratio	Lower	Upper
216	100		
214	78.3	62.8	94.2
179	21.0	16.8	25.2
108	18.7	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 41.026 ng
RT: 12.569 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:237 Resp: 413181

Ion	Ratio	Lower	Upper
237	100		
235	61.8	41.8	81.8
272	13.4	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 81.766 ng
RT: 15.922 min Scan# 2182
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:330 Resp: 995466

Ion	Ratio	Lower	Upper
330	100		
332	96.9	77.6	116.4
141	30.6	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 40.950 ng
RT: 12.828 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:196 Resp: 727105

Ion	Ratio	Lower	Upper
196	100		
198	96.5	76.6	115.0
200	31.7	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 40.985 ng
RT: 12.904 min Scan# 1669
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:196 Resp: 798076

Ion	Ratio	Lower	Upper
196	100		
198	96.1	76.5	114.7
97	45.3	38.7	58.1
132	27.3	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 77.885 ng
RT: 13.022 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:172 Resp: 5105223

Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.3	42.5
170	23.6	19.0	28.4







#46
1,1'-Biphenyl
Concen: 38.846 ng
RT: 13.228 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:154 Resp: 2821370

Ion	Ratio	Lower	Upper
154	100		
153	40.2	20.3	60.3
76	13.7	0.0	32.9







#47
2-Chloronaphthalene
Concen: 39.006 ng
RT: 13.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:162 Resp: 2130879

Ion	Ratio	Lower	Upper
162	100		
127	39.4	31.5	47.3
164	32.8	26.6	39.8







#48
2-Nitroaniline
Concen: 41.137 ng
RT: 13.475 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion: 65 Resp: 574806

Ion	Ratio	Lower	Upper
65	100		
92	75.5	62.1	93.1
138	123.5	99.5	149.3







#49
Acenaphthylene
Concen: 39.262 ng
RT: 14.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:152 Resp: 3245488

Ion	Ratio	Lower	Upper
152	100		
151	20.1	16.0	24.0
153	13.3	10.6	15.8







#50
Dimethylphthalate
Concen: 38.548 ng
RT: 13.863 min Scan# 1832
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:163 Resp: 2606961

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.0	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 39.611 ng
RT: 13.975 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:165 Resp: 563651

Ion	Ratio	Lower	Upper
165	100		
63	45.8	37.7	56.5
89	51.5	41.9	62.9







#52
Acenaphthene
Concen: 38.881 ng
RT: 14.475 min Scan# 1936
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:154 Resp: 2053476

Ion	Ratio	Lower	Upper
154	100		
153	110.8	90.5	135.7
152	53.7	43.1	64.7







#53
3-Nitroaniline
Concen: 40.719 ng
RT: 14.310 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:138 Resp: 625803

Ion	Ratio	Lower	Upper
138	100		
108	9.9	8.2	12.2
92	109.7	86.7	130.1







#54
2,4-Dinitrophenol
Concen: 42.791 ng
RT: 14.516 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:184 Resp: 322185

Ion	Ratio	Lower	Upper
184	100		
63	54.2	42.1	63.1
154	56.8	44.6	67.0







#55
Dibenzofuran
Concen: 38.264 ng
RT: 14.810 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:168 Resp: 3273448

Ion	Ratio	Lower	Upper
168	100		
139	36.1	29.1	43.7
169	13.4	10.8	16.2







#56
4-Nitrophenol
Concen: 42.293 ng
RT: 14.634 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:139 Resp: 543119

Ion	Ratio	Lower	Upper
139	100		
109	63.4	43.3	83.3
65	78.3	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 41.106 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:165 Resp: 772629

Ion	Ratio	Lower	Upper
165	100		
63	36.9	29.0	43.6
89	66.9	54.2	81.4
182	2.9	2.3	3.5







#58
Fluorene
Concen: 39.061 ng
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:166 Resp: 2603595

Ion	Ratio	Lower	Upper
166	100		
165	97.5	77.6	116.4
167	13.6	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 39.844 ng
RT: 15.040 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:232 Resp: 685479

Ion	Ratio	Lower	Upper
232	100		
131	45.1	35.9	53.9
130	2.2	1.8	2.6
166	30.5	24.2	36.2







#60
Diethylphthalate
Concen: 38.539 ng
RT: 15.240 min Scan# 2066
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:149 Resp: 2595505

Ion	Ratio	Lower	Upper
149	100		
177	22.1	17.7	26.5
150	12.6	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 38.572 ng
RT: 15.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:204 Resp: 1261005

Ion	Ratio	Lower	Upper
204	100		
206	33.7	26.4	39.6
141	57.5	45.8	68.8







#62
4-Nitroaniline
Concen: 41.253 ng
RT: 15.493 min Scan# 2109
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:138 Resp: 664956

Ion	Ratio	Lower	Upper
138	100		
92	49.7	28.7	68.7
108	69.1	49.5	89.5







#63
Azobenzene
Concen: 39.347 ng
RT: 15.769 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion: 77 Resp: 2497600

Ion	Ratio	Lower	Upper
77	100		
182	27.1	7.7	47.7
105	21.8	2.3	42.3
51	23.2	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.204 min Scan# 2400
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:188 Resp: 1810172

Ion	Ratio	Lower	Upper
188	100		
94	10.1	8.1	12.1
80	11.1	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 40.442 ng
RT: 15.551 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:198 Resp: 434596

Ion	Ratio	Lower	Upper
198	100		
51	31.7	14.7	54.7
105	41.0	23.8	63.8







#66
n-Nitrosodiphenylamine
Concen: 39.234 ng
RT: 15.681 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:169 Resp: 2156617

Ion	Ratio	Lower	Upper
169	100		
168	66.7	52.9	79.3
167	35.9	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 38.892 ng
RT: 16.375 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:248 Resp: 772937

Ion	Ratio	Lower	Upper
248	100		
250	97.3	78.4	117.6
141	71.7	59.4	89.0







#68
Hexachlorobenzene
Concen: 39.073 ng
RT: 16.498 min Scan# 2280
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:284 Resp: 919105

Ion	Ratio	Lower	Upper
284	100		
142	29.3	25.4	38.2
249	29.1	23.5	35.3







#69
Atrazine
Concen: 36.287 ng
RT: 16.657 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:200 Resp: 481443

Ion	Ratio	Lower	Upper
200	100		
173	28.1	7.6	47.6
215	48.5	29.0	69.0







#70
Pentachlorophenol
Concen: 41.161 ng
RT: 16.851 min Scan# 2340
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:266 Resp: 626146

Ion	Ratio	Lower	Upper
266	100		
268	64.6	51.3	76.9
264	63.1	50.6	76.0







#71
Phenanthrene
Concen: 38.379 ng
RT: 17.251 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:178 Resp: 3795212

Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.6	23.4
179	15.2	12.5	18.7







#72
Anthracene
Concen: 39.358 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:178 Resp: 3790779

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.2	22.8
179	15.5	12.6	18.8







#73
Carbazole
Concen: 39.665 ng
RT: 17.628 min Scan# 24
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:167 Resp: 3684202

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.1	25.7
139	13.0	10.2	15.4







#74
Di-n-butylphthalate
Concen: 40.866 ng
RT: 18.222 min Scan# 2573
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:149 Resp: 4505984

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.6	11.4
104	4.8	3.9	5.9







#75
Fluoranthene
Concen: 39.006 ng
RT: 19.345 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:202 Resp: 4467370

Ion	Ratio	Lower	Upper
202	100		
101	12.1	0.0	32.5
203	17.6	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.669 min Scan# 3159
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:240 Resp: 1903114

Ion	Ratio	Lower	Upper
240	100		
120	11.3	9.2	13.8
236	25.4	20.4	30.6







#77
Benzidine
Concen: 39.777 ng
RT: 19.534 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:184 Resp: 827654

Ion	Ratio	Lower	Upper
184	100		
185	15.0	11.8	17.6
183	11.3	9.1	13.7







#78
Pyrene
Concen: 39.565 ng
RT: 19.716 min Scan# 2827
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:202 Resp: 4681182

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.1	25.7
203	17.9	14.2	21.4







#79
Terphenyl-d14
Concen: 78.549 ng
RT: 19.933 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:244 Resp: 7241807

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.7	8.5
122	12.5	9.6	14.4







#80
Butylbenzylphthalate
Concen: 40.824 ng
RT: 20.669 min Scan# 2989
Delta R.T. 0.018 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:149 Resp: 1991838

Ion	Ratio	Lower	Upper
149	100		
91	68.1	55.3	82.9
206	22.0	16.1	24.1







#81
Benzo(a)anthracene
Concen: 39.318 ng
RT: 21.651 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:228 Resp: 4652183

Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.8	32.8
229	19.5	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 41.750 ng
RT: 21.569 min Scan# 3142
Delta R.T. 0.018 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:252 Resp: 1655655

Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.4	77.2
126	11.9	9.3	13.9







#83
Chrysene
Concen: 38.624 ng
RT: 21.716 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:228 Resp: 4393252

Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.5	36.7
229	20.0	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 41.569 ng
RT: 21.586 min Scan# 3145
Delta R.T. 0.018 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:149 Resp: 3060349

Ion	Ratio	Lower	Upper
149	100		
167	28.6	22.5	33.7
279	5.1	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 42.452 ng
RT: 22.886 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:149 Resp: 5007462

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.2	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.069 min Scan# 3737
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:264 Resp: 2127367

Ion	Ratio	Lower	Upper
264	100		
260	23.9	18.5	27.7
265	21.8	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 38.387 ng
RT: 28.957 min Scan# 4
Delta R.T. 0.029 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:276 Resp: 5699470

Ion	Ratio	Lower	Upper
276	100		
138	16.8	13.7	20.5
277	25.3	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 38.891 ng
RT: 23.992 min Scan# 3554
Delta R.T. 0.012 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:252 Resp: 4988392

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.2	25.8
125	11.0	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 38.280 ng
RT: 24.069 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:252 Resp: 4803010

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	10.2	9.1	13.7







#90
Benzo(a)pyrene
Concen: 38.607 ng
RT: 24.916 min Scan# 3711
Delta R.T. 0.024 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:252 Resp: 4306060

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.0
125	11.8	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 38.474 ng
RT: 29.027 min Scan# 4
Delta R.T. 0.018 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Tgt Ion:278 Resp: 4752357

Ion	Ratio	Lower	Upper
278	100		
139	16.5	13.0	19.4
279	24.0	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 38.087 ng m
RT: 30.127 min Scan# 4597
Delta R.T. 0.000 min
Lab File: BP024168.D
Acq: 12 Mar 2025 21:43

Tgt Ion:276 Resp: 4782473

Ion	Ratio	Lower	Upper
276	100		
277	24.1	19.2	28.8
138	22.4	17.4	26.2



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2		1,4-Dioxane	0.497	0.470	5.4	105	0.00
3		Pyridine	1.358	1.314	3.2	105	0.00
4		n-Nitrosodimethylamine	0.454	0.438	3.5	106	0.00
5	S	2-Fluorophenol	1.253	1.225	2.2	106	0.00
6		Aniline	1.527	1.504	1.5	106	0.00
7	S	Phenol-d6	1.675	1.648	1.6	105	0.00
8		2-Chlorophenol	1.353	1.325	2.1	106	0.00
9		Benzaldehyde	0.809	0.816	-0.9	109	0.00
10	C	Phenol	1.691	1.644	2.8	105	0.00
11		bis(2-Chloroethyl)ether	1.334	1.274	4.5	104	0.00
12		1,3-Dichlorobenzene	1.459	1.399	4.1	105	0.00
13	C	1,4-Dichlorobenzene	1.490	1.414	5.1	104	0.00
14		1,2-Dichlorobenzene	1.430	1.363	4.7	105	0.00
15		Benzyl Alcohol	1.061	1.061	0.0	106	0.00
16		2,2'-oxybis(1-Chloropropane	1.361	1.305	4.1	105	0.00
17		2-Methylphenol	1.061	1.051	0.9	106	0.00
18		Hexachloroethane	0.529	0.510	3.6	105	0.00
19	P	n-Nitroso-di-n-propylamine	0.972	0.965	0.7	105	0.00
20		3+4-Methylphenols	1.457	1.444	0.9	105	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
22		Acetophenone	0.506	0.488	3.6	104	0.00
23	S	Nitrobenzene-d5	0.354	0.349	1.4	105	0.00
24		Nitrobenzene	0.350	0.341	2.6	106	0.00
25		Isophorone	0.608	0.603	0.8	106	0.00
26	C	2-Nitrophenol	0.167	0.175	-4.8	109	0.00
27		2,4-Dimethylphenol	0.215	0.214	0.5	105	0.00
28		bis(2-Chloroethoxy)methane	0.427	0.413	3.3	105	0.00
29	C	2,4-Dichlorophenol	0.291	0.295	-1.4	105	0.00
30		1,2,4-Trichlorobenzene	0.321	0.307	4.4	104	0.00
31		Naphthalene	1.060	1.013	4.4	104	0.00
32		Benzoic acid	0.209	0.228	-9.1	111	0.00
33		4-Chloroaniline	0.379	0.375	1.1	105	0.00
34	C	Hexachlorobutadiene	0.185	0.179	3.2	106	0.00
35		Caprolactam	0.095	0.097	-2.1	106	0.00
36	C	4-Chloro-3-methylphenol	0.313	0.311	0.6	105	0.00
37		2-Methylnaphthalene	0.716	0.694	3.1	105	0.00
38		1-Methylnaphthalene	0.697	0.677	2.9	107	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40		1,2,4,5-Tetrachlorobenzene	0.578	0.564	2.4	105	0.00
41	P	Hexachlorocyclopentadiene	0.208	0.214	-2.9	105	0.00
42	S	2,4,6-Tribromophenol	0.252	0.257	-2.0	106	0.00
43	C	2,4,6-Trichlorophenol	0.367	0.376	-2.5	105	0.00
44		2,4,5-Trichlorophenol	0.403	0.413	-2.5	105	0.00
45	S	2-Fluorobiphenyl	1.356	1.321	2.6	104	0.00
46		1,1'-Biphenyl	1.503	1.460	2.9	104	0.00
47		2-Chloronaphthalene	1.130	1.102	2.5	105	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.289	0.297	-2.8	106	0.00
49		Acenaphthylene	1.710	1.679	1.8	105	0.00
50		Dimethylphthalate	1.399	1.349	3.6	103	0.00
51		2,6-Dinitrotoluene	0.294	0.292	0.7	103	0.00
52	C	Acenaphthene	1.093	1.062	2.8	104	0.00
53		3-Nitroaniline	0.318	0.324	-1.9	104	0.00
54	P	2,4-Dinitrophenol	0.156	0.167	-7.1	108	0.00
55		Dibenzofuran	1.770	1.693	4.4	103	0.00
56	P	4-Nitrophenol	0.266	0.281	-5.6	105	0.00
57		2,4-Dinitrotoluene	0.389	0.400	-2.8	105	0.00
58		Fluorene	1.379	1.347	2.3	104	0.00
59		2,3,4,6-Tetrachlorophenol	0.356	0.355	0.3	103	0.00
60		Diethylphthalate	1.394	1.343	3.7	103	0.00
61		4-Chlorophenyl-phenylether	0.676	0.652	3.6	103	0.00
62		4-Nitroaniline	0.334	0.344	-3.0	104	0.00
63		Azobenzene	1.314	1.292	1.7	104	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
65		4,6-Dinitro-2-methylphenol	0.119	0.120	-0.8	105	0.00
66	c	n-Nitrosodiphenylamine	0.607	0.596	1.8	105	0.00
67		4-Bromophenyl-phenylether	0.220	0.213	3.2	104	0.00
68		Hexachlorobenzene	0.260	0.254	2.3	105	0.00
69		Atrazine	0.147	0.133	9.5	103	0.00
70	C	Pentachlorophenol	0.168	0.173	-3.0	103	0.00
71		Phenanthrene	1.093	1.048	4.1	102	0.00
72		Anthracene	1.064	1.047	1.6	104	0.01
73		Carbazole	1.026	1.018	0.8	103	0.01
74		Di-n-butylphthalate	1.218	1.245	-2.2	104	0.01
75	C	Fluoranthene	1.265	1.234	2.5	103	0.02
76	I	Chrysene-d12	1.000	1.000	0.0	101	0.01
77		Benzidine	0.219	0.217	0.9	85	0.01
78		Pyrene	1.243	1.230	1.0	103	0.01
79	S	Terphenyl-d14	0.969	0.951	1.9	102	0.01
80		Butylbenzylphthalate	0.513	0.523	-1.9	101	0.02
81		Benzo(a)anthracene	1.243	1.222	1.7	102	0.02
82		3,3'-Dichlorobenzidine	0.417	0.435	-4.3	103	0.02
83		Chrysene	1.195	1.154	3.4	101	0.01
84		Bis(2-ethylhexyl)phthalate	0.774	0.804	-3.9	103	0.02
85	c	Di-n-octyl phthalate	1.240	1.316	-6.1	105	0.02
86	I	Perylene-d12	1.000	1.000	0.0	104	0.01
87		Indeno(1,2,3-cd)pyrene	1.396	1.340	4.0	102	0.03
88		Benzo(b)fluoranthene	1.206	1.172	2.8	103	0.01
89		Benzo(k)fluoranthene	1.180	1.129	4.3	104	0.02
90	C	Benzo(a)pyrene	1.049	1.012	3.5	102	0.02
91		Dibenzo(a,h)anthracene	1.161	1.117	3.8	102	0.02
92		Benzo(g,h,i)perylene	1.180	1.124	4.7	101	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2		1,4-Dioxane	40.000	37.880	5.3	105	0.00
3		Pyridine	40.000	38.718	3.2	105	0.00
4		n-Nitrosodimethylamine	40.000	38.646	3.4	106	0.00
5	S	2-Fluorophenol	80.000	78.230	2.2	106	0.00
6		Aniline	40.000	39.413	1.5	106	0.00
7	S	Phenol-d6	80.000	78.704	1.6	105	0.00
8		2-Chlorophenol	40.000	39.149	2.1	106	0.00
9		Benzaldehyde	40.000	40.389	-1.0	109	0.00
10	C	Phenol	40.000	38.869	2.8	105	0.00
11		bis(2-Chloroethyl)ether	40.000	38.203	4.5	104	0.00
12		1,3-Dichlorobenzene	40.000	38.353	4.1	105	0.00
13	C	1,4-Dichlorobenzene	40.000	37.960	5.1	104	0.00
14		1,2-Dichlorobenzene	40.000	38.116	4.7	105	0.00
15		Benzyl Alcohol	40.000	39.999	0.0	106	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	38.356	4.1	105	0.00
17		2-Methylphenol	40.000	39.636	0.9	106	0.00
18		Hexachloroethane	40.000	38.587	3.5	105	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	39.697	0.8	105	0.00
20		3+4-Methylphenols	40.000	39.629	0.9	105	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22		Acetophenone	40.000	38.611	3.5	104	0.00
23	S	Nitrobenzene-d5	80.000	78.652	1.7	105	0.00
24		Nitrobenzene	40.000	39.045	2.4	106	0.00
25		Isophorone	40.000	39.677	0.8	106	0.00
26	C	2-Nitrophenol	40.000	41.892	-4.7	109	0.00
27		2,4-Dimethylphenol	40.000	39.779	0.6	105	0.00
28		bis(2-Chloroethoxy)methane	40.000	38.608	3.5	105	0.00
29	C	2,4-Dichlorophenol	40.000	40.558	-1.4	105	0.00
30		1,2,4-Trichlorobenzene	40.000	38.264	4.3	104	0.00
31		Naphthalene	40.000	38.223	4.4	104	0.00
32		Benzoic acid	40.000	43.618	-9.0	111	0.00
33		4-Chloroaniline	40.000	39.499	1.3	105	0.00
34	C	Hexachlorobutadiene	40.000	38.854	2.9	106	0.00
35		Caprolactam	40.000	40.859	-2.1	106	0.00
36	C	4-Chloro-3-methylphenol	40.000	39.702	0.7	105	0.00
37		2-Methylnaphthalene	40.000	38.762	3.1	105	0.00
38		1-Methylnaphthalene	40.000	38.874	2.8	107	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	38.988	2.5	105	0.00
41	P	Hexachlorocyclopentadiene	40.000	41.026	-2.6	105	0.00
42	S	2,4,6-Tribromophenol	80.000	81.766	-2.2	106	0.00
43	C	2,4,6-Trichlorophenol	40.000	40.950	-2.4	105	0.00
44		2,4,5-Trichlorophenol	40.000	40.985	-2.5	105	0.00
45	S	2-Fluorobiphenyl	80.000	77.885	2.6	104	0.00
46		1,1'-Biphenyl	40.000	38.846	2.9	104	0.00
47		2-Chloronaphthalene	40.000	39.006	2.5	105	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	41.137	-2.8	106	0.00
49		Acenaphthylene	40.000	39.262	1.8	105	0.00
50		Dimethylphthalate	40.000	38.548	3.6	103	0.00
51		2,6-Dinitrotoluene	40.000	39.611	1.0	103	0.00
52	C	Acenaphthene	40.000	38.881	2.8	104	0.00
53		3-Nitroaniline	40.000	40.719	-1.8	104	0.00
54	P	2,4-Dinitrophenol	40.000	42.791	-7.0	108	0.00
55		Dibenzofuran	40.000	38.264	4.3	103	0.00
56	P	4-Nitrophenol	40.000	42.293	-5.7	105	0.00
57		2,4-Dinitrotoluene	40.000	41.106	-2.8	105	0.00
58		Fluorene	40.000	39.061	2.3	104	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	39.844	0.4	103	0.00
60		Diethylphthalate	40.000	38.539	3.7	103	0.00
61		4-Chlorophenyl-phenylether	40.000	38.572	3.6	103	0.00
62		4-Nitroaniline	40.000	41.253	-3.1	104	0.00
63		Azobenzene	40.000	39.347	1.6	104	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65		4,6-Dinitro-2-methylphenol	40.000	40.442	-1.1	105	0.00
66	c	n-Nitrosodiphenylamine	40.000	39.234	1.9	105	0.00
67		4-Bromophenyl-phenylether	40.000	38.892	2.8	104	0.00
68		Hexachlorobenzene	40.000	39.073	2.3	105	0.00
69		Atrazine	40.000	36.287	9.3	103	0.00
70	C	Pentachlorophenol	40.000	41.161	-2.9	103	0.00
71		Phenanthrene	40.000	38.379	4.1	102	0.00
72		Anthracene	40.000	39.358	1.6	104	0.01
73		Carbazole	40.000	39.665	0.8	103	0.01
74		Di-n-butylphthalate	40.000	40.866	-2.2	104	0.01
75	C	Fluoranthene	40.000	39.006	2.5	103	0.02
76	I	Chrysene-d12	20.000	20.000	0.0	101	0.01
77		Benzidine	40.000	39.777	0.6	85	0.01
78		Pyrene	40.000	39.565	1.1	103	0.01
79	S	Terphenyl-d14	80.000	78.549	1.8	102	0.01
80		Butylbenzylphthalate	40.000	40.824	-2.1	101	0.02
81		Benzo(a)anthracene	40.000	39.318	1.7	102	0.02
82		3,3'-Dichlorobenzidine	40.000	41.750	-4.4	103	0.02
83		Chrysene	40.000	38.624	3.4	101	0.01
84		Bis(2-ethylhexyl)phthalate	40.000	41.569	-3.9	103	0.02
85	c	Di-n-octyl phthalate	40.000	42.452	-6.1	105	0.02
86	I	Perylene-d12	20.000	20.000	0.0	104	0.01
87		Indeno(1,2,3-cd)pyrene	40.000	38.387	4.0	102	0.03
88		Benzo(b)fluoranthene	40.000	38.891	2.8	103	0.01
89		Benzo(k)fluoranthene	40.000	38.280	4.3	104	0.02
90	C	Benzo(a)pyrene	40.000	38.607	3.5	102	0.02
91		Dibenzo(a,h)anthracene	40.000	38.474	3.8	102	0.02
92		Benzo(g,h,i)perylene	40.000	38.087	4.8	101	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024168.D
Acq On : 12 Mar 2025 21:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP031225

Quant Time: Mar 13 04:44:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 04:38:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG No.: Q1502

Instrument ID: BNA_F Calibration Date/Time: 03/13/2025 12:17

Lab File ID: BF141939.D Init. Calib. Date(s): 03/10/2025 03/10/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:01 15:20

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.198	1.154		-3.7	
Phenol-d6	1.526	1.446		-5.2	
Phenol	1.605	1.528		-4.8	20.0
2-Chlorophenol	1.329	1.276		-4.0	
2-Methylphenol	1.057	0.994		-6.0	
3+4-Methylphenols	1.354	1.268		-6.4	
Nitrobenzene-d5	0.355	0.344		-3.1	
2-Nitrophenol	0.173	0.171		-1.2	20.0
2,4-Dimethylphenol	0.239	0.230		-3.8	
2,4-Dichlorophenol	0.296	0.284		-4.1	20.0
Benzoic acid	0.210	0.221		5.2	
4-Chloro-3-methylphenol	0.331	0.312		-5.7	20.0
2,4,6-Trichlorophenol	0.398	0.393		-1.3	20.0
2-Fluorobiphenyl	1.315	1.265		-3.8	
2,4,5-Trichlorophenol	0.403	0.393		-2.5	
2,4-Dinitrophenol	0.139	0.155	0.050	11.5	
4-Nitrophenol	0.223	0.230	0.050	3.1	
4,6-Dinitro-2-methylphenol	0.119	0.125		5.0	
2,4,6-Tribromophenol	0.254	0.253		-0.4	
Pentachlorophenol	0.170	0.180		5.9	20.0
Terphenyl-d14	1.353	1.361		0.6	
2,3,4,6-Tetrachlorophenol	0.367	0.364		-0.8	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	198004	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	767232	20.000	ng	0.00
39) Acenaphthene-d10	9.910	164	426931	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	726402	20.000	ng	0.00
76) Chrysene-d12	14.033	240	451214	20.000	ng	0.00
86) Perylene-d12	15.504	264	394917	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	914193	77.057	ng	0.00
7) Phenol-d6	6.504	99	1145375	75.826	ng	0.00
23) Nitrobenzene-d5	7.439	82	1055101	77.391	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	432513	79.856	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	2159613	76.930	ng	0.00
79) Terphenyl-d14	12.980	244	2456339	80.485	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.681	88	195530	39.334	ng	99
3) Pyridine	3.446	79	472100	38.717	ng	99
4) n-Nitrosodimethylamine	3.399	42	226886	39.034	ng	98
6) Aniline	6.540	93	558500	37.480	ng	100
8) 2-Chlorophenol	6.663	128	505200	38.395	ng	100
9) Benzaldehyde	6.428	77	297311	35.193	ng	99
10) Phenol	6.516	94	605095	38.077	ng	100
11) bis(2-Chloroethyl)ether	6.616	93	446408	37.411	ng	99
12) 1,3-Dichlorobenzene	6.816	146	542765	38.208	ng	99
13) 1,4-Dichlorobenzene	6.892	146	551564	38.375	ng	99
14) 1,2-Dichlorobenzene	7.051	146	522830	38.620	ng	99
15) Benzyl Alcohol	7.016	79	463386	37.946	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.151	45	508939	35.329	ng	99
17) 2-Methylphenol	7.122	107	393438	37.607	ng	99
18) Hexachloroethane	7.392	117	205179	38.069	ng	95
19) n-Nitroso-di-n-propyla...	7.292	70	347590	35.812	ng	100
20) 3+4-Methylphenols	7.275	107	502037	37.464	ng	98
22) Acetophenone	7.287	105	704297	38.332	ng	99
24) Nitrobenzene	7.457	77	523612	38.634	ng	100
25) Isophorone	7.698	82	889235	36.899	ng	100
26) 2-Nitrophenol	7.775	139	263074	39.549	ng	100
27) 2,4-Dimethylphenol	7.804	122	352630	38.499	ng	100
28) bis(2-Chloroethoxy)met...	7.904	93	559932	37.044	ng	99
29) 2,4-Dichlorophenol	8.010	162	435680	38.322	ng	99
30) 1,2,4-Trichlorobenzene	8.098	180	482699	38.527	ng	99
31) Naphthalene	8.181	128	1512818	38.385	ng	100
32) Benzoic acid	7.922	122	338815	42.081	ng	99
33) 4-Chloroaniline	8.228	127	523526	37.629	ng	99
34) Hexachlorobutadiene	8.298	225	319154	39.060	ng	99
35) Caprolactam	8.598	113	131311	37.884	ng	94
36) 4-Chloro-3-methylphenol	8.698	107	479000	37.769	ng	100
37) 2-Methylnaphthalene	8.869	142	1002209	38.045	ng	100
38) 1-Methylnaphthalene	8.969	142	966389	38.016	ng	99
40) 1,2,4,5-Tetrachloroben...	9.034	216	506575	38.972	ng	98
41) Hexachlorocyclopentadiene	9.022	237	213556	41.981	ng	99
43) 2,4,6-Trichlorophenol	9.145	196	335717	39.520	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.181	196	335620	39.009	ng	100
46) 1,1'-Biphenyl	9.333	154	1256074	38.721	ng	100
47) 2-Chloronaphthalene	9.357	162	946087	39.130	ng	99
48) 2-Nitroaniline	9.451	65	273545	39.074	ng	100
49) Acenaphthylene	9.775	152	1399495	39.005	ng	100
50) Dimethylphthalate	9.633	163	1139250	38.106	ng	100
51) 2,6-Dinitrotoluene	9.692	165	243366	39.096	ng	97
52) Acenaphthene	9.945	154	991587	39.326	ng	99
53) 3-Nitroaniline	9.863	138	243288	38.530	ng	99
54) 2,4-Dinitrophenol	9.963	184	132695	41.009	ng	88
55) Dibenzofuran	10.116	168	1389049	38.554	ng	99
56) 4-Nitrophenol	10.010	139	196664	41.301	ng	100
57) 2,4-Dinitrotoluene	10.098	165	318648	39.193	ng	98
58) Fluorene	10.463	166	1061221	38.195	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.233	232	310873	39.708	ng	98
60) Diethylphthalate	10.333	149	1134119	38.044	ng	100
61) 4-Chlorophenyl-phenyle...	10.451	204	559798	38.646	ng	98
62) 4-Nitroaniline	10.475	138	242993	39.528	ng	99
63) Azobenzene	10.610	77	1050098	37.888	ng	99
65) 4,6-Dinitro-2-methylph...	10.504	198	181277	41.862	ng	99
66) n-Nitrosodiphenylamine	10.569	169	906795	38.576	ng	100
67) 4-Bromophenyl-phenylether	10.939	248	355291	38.946	ng	100
68) Hexachlorobenzene	11.004	284	395519	39.530	ng	99
69) Atrazine	11.092	200	238990	33.164	ng	99
70) Pentachlorophenol	11.198	266	261878	42.480	ng	100
71) Phenanthrene	11.422	178	1501673	38.274	ng	99
72) Anthracene	11.475	178	1507154	38.288	ng	100
73) Carbazole	11.627	167	1302310	38.363	ng	100
74) Di-n-butylphthalate	11.957	149	1710894	37.983	ng	99
75) Fluoranthene	12.610	202	1559838	37.478	ng	99
77) Benzidine	12.727	184	397420	63.130	ng	99
78) Pyrene	12.839	202	1536542	39.368	ng	99
80) Butylbenzylphthalate	13.451	149	585688	38.339	ng	99
81) Benzo(a)anthracene	14.021	228	1129185	38.137	ng	100
82) 3,3'-Dichlorobenzidine	13.986	252	353908	41.361	ng	99
83) Chrysene	14.063	228	1052493	39.214	ng	100
84) Bis(2-ethylhexyl)phtha...	14.010	149	770893	36.599	ng	100
85) Di-n-octyl phthalate	14.621	149	1091714	37.251	ng	99
87) Indeno(1,2,3-cd)pyrene	16.986	276	1036676	40.580	ng	98
88) Benzo(b)fluoranthene	15.074	252	995633	36.786	ng	99
89) Benzo(k)fluoranthene	15.098	252	862161	37.223	ng	100
90) Benzo(a)pyrene	15.439	252	809840	38.240	ng	100
91) Dibenzo(a,h)anthracene	17.004	278	863026	40.945	ng	99
92) Benzo(g,h,i)perylene	17.439	276	863842	41.473	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 198004

Ion	Ratio	Lower	Upper
152	100		
150	158.4	127.4	191.2
115	66.9	51.9	77.9







#2
1,4-Dioxane
Concen: 39.334 ng
RT: 2.681 min Scan# 100
Delta R.T. 0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion: 88 Resp: 195530

Ion	Ratio	Lower	Upper
88	100		
58	70.9	57.4	86.2
43	27.3	21.5	32.3







#3
Pyridine
Concen: 38.717 ng
RT: 3.446 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 472100

Ion	Ratio	Lower	Upper
79	100		
52	65.2	53.3	79.9
51	34.4	27.6	41.4







#4
n-Nitrosodimethylamine
Concen: 39.034 ng
RT: 3.399 min Scan# 222
Delta R.T. 0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion: 42 Resp: 226886

Ion	Ratio	Lower	Upper
42	100		
74	126.9	103.4	155.0
44	6.2	5.0	7.4







#5
2-Fluorophenol
Concen: 77.057 ng
RT: 5.493 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 914193

Ion	Ratio	Lower	Upper
112	100		
64	56.5	46.5	69.7
63	31.2	25.4	38.2







#6
Aniline
Concen: 37.480 ng
RT: 6.540 min Scan# 756
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion: 93 Resp: 558500

Ion	Ratio	Lower	Upper
93	100		
66	40.2	31.9	47.9
65	19.7	15.8	23.6







#7
Phenol-d6
Concen: 75.826 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 1145375

Ion	Ratio	Lower	Upper
99	100		
42	17.3	14.6	21.8
71	37.7	30.8	46.2







#8
2-Chlorophenol
Concen: 38.395 ng
RT: 6.663 min Scan# 777
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:128 Resp: 505200

Ion	Ratio	Lower	Upper
128	100		
130	32.5	12.8	52.8
64	44.4	24.4	64.4







#9
Benzaldehyde
Concen: 35.193 ng
RT: 6.428 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 297311

Ion	Ratio	Lower	Upper
77	100		
105	98.3	80.0	120.0
106	94.8	74.5	114.5







#10
Phenol
Concen: 38.077 ng
RT: 6.516 min Scan# 752
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion: 94 Resp: 605095

Ion	Ratio	Lower	Upper
94	100		
65	30.7	10.5	50.5
66	45.6	25.8	65.8







#11
bis(2-Chloroethyl)ether
Concen: 37.411 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 446408

Ion	Ratio	Lower	Upper
93	100		
63	71.5	53.0	93.0
95	32.6	12.9	52.9







#12
1,3-Dichlorobenzene
Concen: 38.208 ng
RT: 6.816 min Scan# 803
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:146 Resp: 542765

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.3	76.9
75	34.2	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 38.375 ng
RT: 6.892 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 551564

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.0	76.6
111	45.1	35.7	53.5







#14
1,2-Dichlorobenzene
Concen: 38.620 ng
RT: 7.051 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:146 Resp: 522830

Ion	Ratio	Lower	Upper
146	100		
148	63.9	50.8	76.2
111	46.0	38.0	57.0







#15
Benzyl Alcohol
Concen: 37.946 ng
RT: 7.016 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 463386

Ion	Ratio	Lower	Upper
79	100		
108	73.8	59.7	89.5
77	63.0	51.0	76.6







#16
2,2'-oxybis(1-Chloropropane)
Concen: 35.329 ng
RT: 7.151 min Scan# 860
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion: 45 Resp: 508939

Ion	Ratio	Lower	Upper
45	100		
77	18.6	0.0	37.8
79	14.2	0.0	34.2







#17
2-Methylphenol
Concen: 37.607 ng
RT: 7.122 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 393438

Ion	Ratio	Lower	Upper
107	100		
108	111.6	90.2	135.2
77	49.0	39.6	59.4
79	48.8	39.1	58.7







#18
Hexachloroethane
Concen: 38.069 ng
RT: 7.392 min Scan# 901
Delta R.T. 0.000 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:117 Resp: 205179

Ion	Ratio	Lower	Upper
117	100		
119	96.1	76.9	115.3
201	109.1	79.9	119.9







#19
n-Nitroso-di-n-propylamine
Concen: 35.812 ng
RT: 7.292 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 347590

Ion	Ratio	Lower	Upper
70	100		
42	52.0	41.4	62.0
101	10.3	8.2	12.2
130	23.1	18.2	27.4







#20
3+4-Methylphenols
Concen: 37.464 ng
RT: 7.275 min Scan# 881
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:107 Resp: 502037

Ion	Ratio	Lower	Upper
107	100		
108	88.8	68.4	108.4
77	86.2	69.7	109.7
79	31.5	11.4	51.4







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 767232

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	6.8	5.8	8.8
68	4.8	4.1	6.1







#22
Acetophenone
Concen: 38.332 ng
RT: 7.287 min Scan# 883
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:105 Resp: 704297

Ion	Ratio	Lower	Upper
105	100		
71	7.1	5.8	8.8
51	29.9	23.4	35.0
120	21.6	17.3	25.9







#23
Nitrobenzene-d5
Concen: 77.391 ng
RT: 7.439 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 1055101

Ion	Ratio	Lower	Upper
82	100		
128	44.2	36.0	54.0
54	49.6	39.6	59.4







#24
Nitrobenzene
Concen: 38.634 ng
RT: 7.457 min Scan# 912
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion: 77 Resp: 523612

Ion	Ratio	Lower	Upper
77	100		
123	46.1	37.0	55.4
65	14.1	11.3	16.9







#25
Isophorone
Concen: 36.899 ng
RT: 7.698 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 889235

Ion	Ratio	Lower	Upper
82	100		
95	8.3	6.4	9.6
138	18.7	14.9	22.3







#26
2-Nitrophenol
Concen: 39.549 ng
RT: 7.775 min Scan# 966
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:139 Resp: 263074

Ion	Ratio	Lower	Upper
139	100		
109	33.2	26.5	39.7
65	50.1	40.3	60.5







#27
2,4-Dimethylphenol
Concen: 38.499 ng
RT: 7.804 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 352630

Ion	Ratio	Lower	Upper
122	100		
107	118.4	94.8	142.2
121	58.9	47.8	71.6







#28
bis(2-Chloroethoxy)methane
Concen: 37.044 ng
RT: 7.904 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion: 93 Resp: 559932

Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.6	40.0
123	13.1	10.2	15.4







#29
2,4-Dichlorophenol
Concen: 38.322 ng
RT: 8.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 435680

Ion	Ratio	Lower	Upper
162	100		
164	64.6	43.8	83.8
98	38.6	18.1	58.1







#30
1,2,4-Trichlorobenzene
Concen: 38.527 ng
RT: 8.098 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:180 Resp: 482699

Ion	Ratio	Lower	Upper
180	100		
182	96.0	77.5	116.3
145	32.5	25.8	38.6







#31
Naphthalene
Concen: 38.385 ng
RT: 8.181 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 1512818

Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.0	13.6
127	13.5	10.8	16.2







#32
Benzoic acid
Concen: 42.081 ng
RT: 7.922 min Scan# 991
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:122 Resp: 338815

Ion	Ratio	Lower	Upper
122	100		
105	127.7	107.9	147.9
77	96.2	79.0	119.0







#33
4-Chloroaniline
Concen: 37.629 ng
RT: 8.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 523526

Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.2	39.4
65	30.1	25.4	38.2
92	20.4	16.2	24.4







#34
Hexachlorobutadiene
Concen: 39.060 ng
RT: 8.298 min Scan# 1055
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:225 Resp: 319154

Ion	Ratio	Lower	Upper
225	100		
223	62.6	50.6	75.8
227	64.8	52.2	78.2







#35
Caprolactam
Concen: 37.884 ng
RT: 8.598 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 131311

Ion	Ratio	Lower	Upper
113	100		
55	141.3	129.6	169.6
56	118.1	104.0	144.0







#36
4-Chloro-3-methylphenol
Concen: 37.769 ng
RT: 8.698 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:107 Resp: 479000

Ion	Ratio	Lower	Upper
107	100		
144	23.6	19.0	28.6
142	73.6	59.2	88.8







#37
2-Methylnaphthalene
Concen: 38.045 ng
RT: 8.869 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 1002209

Ion	Ratio	Lower	Upper
142	100		
141	89.4	71.2	106.8







#38
1-Methylnaphthalene
Concen: 38.016 ng
RT: 8.969 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:142 Resp: 966389

Ion	Ratio	Lower	Upper
142	100		
141	92.0	74.2	111.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.910 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 426931

Ion	Ratio	Lower	Upper
164	100		
162	103.3	81.8	122.6
160	46.3	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.972 ng
RT: 9.034 min Scan# 1180
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:216 Resp: 506575

Ion	Ratio	Lower	Upper
216	100		
214	79.5	61.9	92.9
179	22.4	18.1	27.1
108	18.3	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 41.981 ng
RT: 9.022 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 213556

Ion	Ratio	Lower	Upper
237	100		
235	62.7	43.1	83.1
272	12.8	0.0	32.0







#42
2,4,6-Tribromophenol
Concen: 79.856 ng
RT: 10.698 min Scan# 1463
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:330 Resp: 432513

Ion	Ratio	Lower	Upper
330	100		
332	97.4	77.6	116.4
141	29.8	24.7	37.1







#43
2,4,6-Trichlorophenol
Concen: 39.520 ng
RT: 9.145 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 335717

Ion	Ratio	Lower	Upper
196	100		
198	100.8	80.6	121.0
200	32.1	25.8	38.6







#44
2,4,5-Trichlorophenol
Concen: 39.009 ng
RT: 9.181 min Scan# 1205
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:196 Resp: 335620

Ion	Ratio	Lower	Upper
196	100		
198	100.8	80.7	121.1
97	53.0	42.1	63.1
132	29.4	23.4	35.2







#45
2-Fluorobiphenyl
Concen: 76.930 ng
RT: 9.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 2159613

Ion	Ratio	Lower	Upper
172	100		
171	36.6	29.3	43.9
170	24.3	19.4	29.0







#46
1,1'-Biphenyl
Concen: 38.721 ng
RT: 9.333 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:154 Resp: 1256074

Ion	Ratio	Lower	Upper
154	100		
153	40.9	21.1	61.1
76	11.1	0.0	31.3







#47
2-Chloronaphthalene
Concen: 39.130 ng
RT: 9.357 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 946087

Ion	Ratio	Lower	Upper
162	100		
127	43.4	35.5	53.3
164	33.0	26.6	40.0







#48
2-Nitroaniline
Concen: 39.074 ng
RT: 9.451 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion: 65 Resp: 273545

Ion	Ratio	Lower	Upper
65	100		
92	70.3	56.6	84.8
138	110.3	88.6	133.0







#49
Acenaphthylene
Concen: 39.005 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 1399495

Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.3	24.5
153	13.6	11.1	16.7







#50
Dimethylphthalate
Concen: 38.106 ng
RT: 9.633 min Scan# 1282
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:163 Resp: 1139250

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.1	4.7
164	10.4	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 39.096 ng
RT: 9.692 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 243366

Ion	Ratio	Lower	Upper
165	100		
63	56.0	47.0	70.4
89	57.6	47.4	71.2







#52
Acenaphthene
Concen: 39.326 ng
RT: 9.945 min Scan# 1335
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:154 Resp: 991587

Ion	Ratio	Lower	Upper
154	100		
153	112.3	90.7	136.1
152	51.1	41.1	61.7







#53
3-Nitroaniline
Concen: 38.530 ng
RT: 9.863 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 243288

Ion	Ratio	Lower	Upper
138	100		
108	11.7	9.3	13.9
92	116.2	93.5	140.3







#54
2,4-Dinitrophenol
Concen: 41.009 ng
RT: 9.963 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:184 Resp: 132695

Ion	Ratio	Lower	Upper
184	100		
63	72.2	60.2	90.4
154	144.1	132.9	199.3







#55
Dibenzofuran
Concen: 38.554 ng
RT: 10.116 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 1389049

Ion	Ratio	Lower	Upper
168	100		
139	41.9	33.9	50.9
169	13.7	11.0	16.4







#56
4-Nitrophenol
Concen: 41.301 ng
RT: 10.010 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:139 Resp: 196664

Ion	Ratio	Lower	Upper
139	100		
109	81.7	61.6	101.6
65	99.4	79.4	119.4







#57
2,4-Dinitrotoluene
Concen: 39.193 ng
RT: 10.098 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 318648

Ion	Ratio	Lower	Upper
165	100		
63	42.8	35.4	53.2
89	69.7	56.6	84.8
182	2.7	2.2	3.4







#58
Fluorene
Concen: 38.195 ng
RT: 10.463 min Scan# 1423
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:166 Resp: 1061221

Ion	Ratio	Lower	Upper
166	100		
165	98.0	79.0	118.4
167	13.5	11.0	16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 39.708 ng
RT: 10.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 310873

Ion	Ratio	Lower	Upper
232	100		
131	44.6	37.3	55.9
130	2.3	1.9	2.9
166	30.4	24.8	37.2







#60
Diethylphthalate
Concen: 38.044 ng
RT: 10.333 min Scan# 1401
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:149 Resp: 1134119

Ion	Ratio	Lower	Upper
149	100		
177	22.7	18.0	27.0
150	12.9	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 38.646 ng
RT: 10.451 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 559798

Ion	Ratio	Lower	Upper
204	100		
206	33.0	26.8	40.2
141	63.0	52.4	78.6







#62
4-Nitroaniline
Concen: 39.528 ng
RT: 10.475 min Scan# 1425
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:138 Resp: 242993

Ion	Ratio	Lower	Upper
138	100		
92	52.1	32.6	72.6
108	85.8	65.0	105.0







#63
Azobenzene
Concen: 37.888 ng
RT: 10.610 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 1050098

Ion	Ratio	Lower	Upper
77	100		
182	27.1	6.6	46.6
105	21.4	1.3	41.3
51	30.0	10.4	50.4







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.398 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:188 Resp: 726402

Ion	Ratio	Lower	Upper
188	100		
94	8.1	6.8	10.2
80	9.0	7.6	11.4







#65
4,6-Dinitro-2-methylphenol
Concen: 41.862 ng
RT: 10.504 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 181277

Ion	Ratio	Lower	Upper
198	100		
51	41.2	22.3	62.3
105	43.1	23.0	63.0







#66
n-Nitrosodiphenylamine
Concen: 38.576 ng
RT: 10.569 min Scan# 1441
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:169 Resp: 906795

Ion	Ratio	Lower	Upper
169	100		
168	66.9	53.7	80.5
167	36.4	29.2	43.8







#67
4-Bromophenyl-phenylether
Concen: 38.946 ng
RT: 10.939 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 355291

Ion	Ratio	Lower	Upper
248	100		
250	97.6	77.8	116.8
141	78.1	62.6	93.8







#68
Hexachlorobenzene
Concen: 39.530 ng
RT: 11.004 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:284 Resp: 395519

Ion	Ratio	Lower	Upper
284	100		
142	32.9	27.0	40.6
249	31.1	25.2	37.8







#69
Atrazine
Concen: 33.164 ng
RT: 11.092 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 238990

Ion	Ratio	Lower	Upper
200	100		
173	24.9	4.7	44.7
215	46.2	25.4	65.4







#70
Pentachlorophenol
Concen: 42.480 ng
RT: 11.198 min Scan# 1548
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:266 Resp: 261878

Ion	Ratio	Lower	Upper
266	100		
268	64.0	51.0	76.6
264	63.0	50.2	75.2







#71
Phenanthrene
Concen: 38.274 ng
RT: 11.422 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 1501673

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.8	23.8
179	15.6	12.6	19.0







#72
Anthracene
Concen: 38.288 ng
RT: 11.475 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:178 Resp: 1507154

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.1	22.7
179	15.6	12.6	19.0







#73
Carbazole
Concen: 38.363 ng
RT: 11.627 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 1302310

Ion	Ratio	Lower	Upper
167	100		
166	22.0	17.4	26.2
139	14.0	11.0	16.6







#74
Di-n-butylphthalate
Concen: 37.983 ng
RT: 11.957 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:149 Resp: 1710894

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.6	11.4
104	5.3	4.4	6.6







#75
Fluoranthene
Concen: 37.478 ng
RT: 12.610 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 1559838

Ion	Ratio	Lower	Upper
202	100		
101	9.1	0.0	29.7
203	17.6	0.0	37.8







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.033 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:240 Resp: 451214

Ion	Ratio	Lower	Upper
240	100		
120	9.6	8.4	12.6
236	25.4	20.5	30.7







#77
Benzidine
Concen: 63.130 ng
RT: 12.727 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 397420

Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.5	17.3
183	11.8	9.7	14.5







#78
Pyrene
Concen: 39.368 ng
RT: 12.839 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:202 Resp: 1536542

Ion	Ratio	Lower	Upper
202	100		
200	21.3	17.3	25.9
203	17.6	14.4	21.6







#79
Terphenyl-d14
Concen: 80.485 ng
RT: 12.980 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 2456339

Ion	Ratio	Lower	Upper
244	100		
212	7.0	6.0	9.0
122	9.2	7.7	11.5







#80
Butylbenzylphthalate
Concen: 38.339 ng
RT: 13.451 min Scan# 1931
Delta R.T. -0.012 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:149 Resp: 585688

Ion	Ratio	Lower	Upper
149	100		
91	71.3	56.5	84.7
206	21.0	17.3	25.9







#81
Benzo(a)anthracene
Concen: 38.137 ng
RT: 14.021 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 1129185

Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.8	32.8
229	19.7	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 41.361 ng
RT: 13.986 min Scan# 2022
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:252 Resp: 353908

Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.0	76.6
126	9.8	8.4	12.6







#83
Chrysene
Concen: 39.214 ng
RT: 14.063 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 1052493

Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.1	36.1
229	20.0	15.7	23.5







#84
Bis(2-ethylhexyl)phthalate
Concen: 36.599 ng
RT: 14.010 min Scan# 2026
Delta R.T. -0.012 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:149 Resp: 770893

Ion	Ratio	Lower	Upper
149	100		
167	26.9	21.4	32.0
279	4.6	3.8	5.6







#85
Di-n-octyl phthalate
Concen: 37.251 ng
RT: 14.621 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 1091714

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.4	6.9	10.3







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.504 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:264 Resp: 394917

Ion	Ratio	Lower	Upper
264	100		
260	23.7	19.1	28.7
265	21.1	17.5	26.3







#87
Indeno(1,2,3-cd)pyrene
Concen: 40.580 ng
RT: 16.986 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 1036676

Ion	Ratio	Lower	Upper
276	100		
138	21.6	18.4	27.6
277	25.5	20.3	30.5







#88
Benzo(b)fluoranthene
Concen: 36.786 ng
RT: 15.074 min Scan# 2207
Delta R.T. -0.006 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:252 Resp: 995633

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.5	26.3
125	8.0	7.0	10.4







#89
Benzo(k)fluoranthene
Concen: 37.223 ng
RT: 15.098 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 862161

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	8.8	7.2	10.8







#90
Benzo(a)pyrene
Concen: 38.240 ng
RT: 15.439 min Scan# 2269
Delta R.T. -0.012 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:252 Resp: 809840

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.0
125	9.7	7.8	11.8







#91
Dibenzo(a,h)anthracene
Concen: 40.945 ng
RT: 17.004 min Scan# 2
Delta R.T. -0.018 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 863026

Ion	Ratio	Lower	Upper
278	100		
139	14.1	11.4	17.2
279	23.8	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 41.473 ng
RT: 17.439 min Scan# 2609
Delta R.T. -0.012 min
Lab File: BF141939.D
Acq: 13 Mar 2025 12:17

Tgt Ion:276 Resp: 863842

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.7	28.1
138	18.3	15.8	23.8



Sample Results: **BF141939.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2		1,4-Dioxane	0.502	0.494	1.6	87	0.00
3		Pyridine	1.232	1.192	3.2	87	0.00
4		n-Nitrosodimethylamine	0.587	0.573	2.4	87	0.00
5	S	2-Fluorophenol	1.198	1.154	3.7	89	0.00
6		Aniline	1.505	1.410	6.3	86	0.00
7	S	Phenol-d6	1.526	1.446	5.2	89	0.00
8		2-Chlorophenol	1.329	1.276	4.0	90	0.00
9		Benzaldehyde	0.853	0.751	12.0	86	0.00
10	C	Phenol	1.605	1.528	4.8	88	0.00
11		bis(2-Chloroethyl)ether	1.205	1.127	6.5	87	0.00
12		1,3-Dichlorobenzene	1.435	1.371	4.5	89	0.00
13	C	1,4-Dichlorobenzene	1.452	1.393	4.1	89	0.00
14		1,2-Dichlorobenzene	1.367	1.320	3.4	90	0.00
15		Benzyl Alcohol	1.233	1.170	5.1	88	0.00
16		2,2'-oxybis(1-Chloropropane	1.455	1.285	11.7	84	0.00
17		2-Methylphenol	1.057	0.994	6.0	88	0.00
18		Hexachloroethane	0.544	0.518	4.8	89	0.00
19	P	n-Nitroso-di-n-propylamine	0.980	0.878	10.4	85	0.00
20		3+4-Methylphenols	1.354	1.268	6.4	87	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
22		Acetophenone	0.479	0.459	4.2	88	0.00
23	S	Nitrobenzene-d5	0.355	0.344	3.1	87	0.00
24		Nitrobenzene	0.353	0.341	3.4	87	0.00
25		Isophorone	0.628	0.580	7.6	85	0.00
26	C	2-Nitrophenol	0.173	0.171	1.2	86	0.00
27		2,4-Dimethylphenol	0.239	0.230	3.8	87	0.00
28		bis(2-Chloroethoxy)methane	0.394	0.365	7.4	86	0.00
29	C	2,4-Dichlorophenol	0.296	0.284	4.1	87	0.00
30		1,2,4-Trichlorobenzene	0.327	0.315	3.7	89	0.00
31		Naphthalene	1.027	0.986	4.0	88	0.00
32		Benzoic acid	0.210	0.221	-5.2	93	0.00
33		4-Chloroaniline	0.363	0.341	6.1	85	0.00
34	C	Hexachlorobutadiene	0.213	0.208	2.3	90	0.00
35		Caprolactam	0.090	0.086	4.4	89	0.00
36	C	4-Chloro-3-methylphenol	0.331	0.312	5.7	86	0.00
37		2-Methylnaphthalene	0.687	0.653	4.9	88	0.00
38		1-Methylnaphthalene	0.663	0.630	5.0	89	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
40		1,2,4,5-Tetrachlorobenzene	0.609	0.593	2.6	87	0.00
41	P	Hexachlorocyclopentadiene	0.238	0.250	-5.0	87	0.00
42	S	2,4,6-Tribromophenol	0.254	0.253	0.4	88	0.00
43	C	2,4,6-Trichlorophenol	0.398	0.393	1.3	85	0.00
44		2,4,5-Trichlorophenol	0.403	0.393	2.5	87	0.00
45	S	2-Fluorobiphenyl	1.315	1.265	3.8	87	0.00
46		1,1'-Biphenyl	1.520	1.471	3.2	87	0.00
47		2-Chloronaphthalene	1.133	1.108	2.2	88	0.00

Sample Results: **BF141939.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.328	0.320	2.4	86	0.00
49		Acenaphthylene	1.681	1.639	2.5	87	0.00
50		Dimethylphthalate	1.401	1.334	4.8	87	0.00
51		2,6-Dinitrotoluene	0.292	0.285	2.4	86	0.00
52	C	Acenaphthene	1.181	1.161	1.7	88	0.00
53		3-Nitroaniline	0.296	0.285	3.7	84	0.00
54	P	2,4-Dinitrophenol	0.139	0.155	-11.5	95	0.00
55		Dibenzofuran	1.688	1.627	3.6	87	0.00
56	P	4-Nitrophenol	0.223	0.230	-3.1	87	0.00
57		2,4-Dinitrotoluene	0.381	0.373	2.1	86	0.00
58		Fluorene	1.302	1.243	4.5	87	0.00
59		2,3,4,6-Tetrachlorophenol	0.367	0.364	0.8	88	0.00
60		Diethylphthalate	1.397	1.328	4.9	86	0.00
61		4-Chlorophenyl-phenylether	0.679	0.656	3.4	88	0.00
62		4-Nitroaniline	0.288	0.285	1.0	85	0.00
63		Azobenzene	1.298	1.230	5.2	85	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
65		4,6-Dinitro-2-methylphenol	0.119	0.125	-5.0	89	0.00
66	c	n-Nitrosodiphenylamine	0.647	0.624	3.6	87	0.00
67		4-Bromophenyl-phenylether	0.251	0.245	2.4	87	0.00
68		Hexachlorobenzene	0.275	0.272	1.1	87	0.00
69		Atrazine	0.198	0.165	16.7	87	0.00
70	C	Pentachlorophenol	0.170	0.180	-5.9	88	0.00
71		Phenanthrene	1.080	1.034	4.3	86	0.00
72		Anthracene	1.084	1.037	4.3	86	0.00
73		Carbazole	0.935	0.896	4.2	85	0.00
74		Di-n-butylphthalate	1.240	1.178	5.0	87	0.00
75	C	Fluoranthene	1.146	1.074	6.3	84	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	79	0.00
77		Benzidine	0.279	0.440	-57.7#	97	0.00
78		Pyrene	1.730	1.703	1.6	84	0.00
79	S	Terphenyl-d14	1.353	1.361	-0.6	84	0.00
80		Butylbenzylphthalate	0.677	0.649	4.1	78	-0.01
81		Benzo(a)anthracene	1.312	1.251	4.6	78	0.00
82		3,3'-Dichlorobenzidine	0.379	0.392	-3.4	84	0.00
83		Chrysene	1.190	1.166	2.0	81	0.00
84		Bis(2-ethylhexyl)phthalate	0.934	0.854	8.6	75	-0.01
85	c	Di-n-octyl phthalate	1.299	1.210	6.9	76	-0.01
86	I	Perylene-d12	1.000	1.000	0.0	84	0.00
87		Indeno(1,2,3-cd)pyrene	1.294	1.313	-1.5	86	-0.01
88		Benzo(b)fluoranthene	1.371	1.261	8.0	88	0.00
89		Benzo(k)fluoranthene	1.173	1.092	6.9	76	-0.01
90	C	Benzo(a)pyrene	1.073	1.025	4.5	83	-0.01
91		Dibenzo(a,h)anthracene	1.067	1.093	-2.4	86	-0.02
92		Benzo(g,h,i)perylene	1.055	1.094	-3.7	86	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Sample Results: **BF141939.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2		1,4-Dioxane	40.000	39.334	1.7	87	0.00
3		Pyridine	40.000	38.717	3.2	87	0.00
4		n-Nitrosodimethylamine	40.000	39.034	2.4	87	0.00
5	S	2-Fluorophenol	80.000	77.057	3.7	89	0.00
6		Aniline	40.000	37.480	6.3	86	0.00
7	S	Phenol-d6	80.000	75.826	5.2	89	0.00
8		2-Chlorophenol	40.000	38.395	4.0	90	0.00
9		Benzaldehyde	40.000	35.193	12.0	86	0.00
10	C	Phenol	40.000	38.077	4.8	88	0.00
11		bis(2-Chloroethyl)ether	40.000	37.411	6.5	87	0.00
12		1,3-Dichlorobenzene	40.000	38.208	4.5	89	0.00
13	C	1,4-Dichlorobenzene	40.000	38.375	4.1	89	0.00
14		1,2-Dichlorobenzene	40.000	38.620	3.5	90	0.00
15		Benzyl Alcohol	40.000	37.946	5.1	88	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	35.329	11.7	84	0.00
17		2-Methylphenol	40.000	37.607	6.0	88	0.00
18		Hexachloroethane	40.000	38.069	4.8	89	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	35.812	10.5	85	0.00
20		3+4-Methylphenols	40.000	37.464	6.3	87	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	88	0.00
22		Acetophenone	40.000	38.332	4.2	88	0.00
23	S	Nitrobenzene-d5	80.000	77.391	3.3	87	0.00
24		Nitrobenzene	40.000	38.634	3.4	87	0.00
25		Isophorone	40.000	36.899	7.8	85	0.00
26	C	2-Nitrophenol	40.000	39.549	1.1	86	0.00
27		2,4-Dimethylphenol	40.000	38.499	3.8	87	0.00
28		bis(2-Chloroethoxy)methane	40.000	37.044	7.4	86	0.00
29	C	2,4-Dichlorophenol	40.000	38.322	4.2	87	0.00
30		1,2,4-Trichlorobenzene	40.000	38.527	3.7	89	0.00
31		Naphthalene	40.000	38.385	4.0	88	0.00
32		Benzoic acid	40.000	42.081	-5.2	93	0.00
33		4-Chloroaniline	40.000	37.629	5.9	85	0.00
34	C	Hexachlorobutadiene	40.000	39.060	2.3	90	0.00
35		Caprolactam	40.000	37.884	5.3	89	0.00
36	C	4-Chloro-3-methylphenol	40.000	37.769	5.6	86	0.00
37		2-Methylnaphthalene	40.000	38.045	4.9	88	0.00
38		1-Methylnaphthalene	40.000	38.016	5.0	89	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	86	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	38.972	2.6	87	0.00
41	P	Hexachlorocyclopentadiene	40.000	41.981	-5.0	87	0.00
42	S	2,4,6-Tribromophenol	80.000	79.856	0.2	88	0.00
43	C	2,4,6-Trichlorophenol	40.000	39.520	1.2	85	0.00
44		2,4,5-Trichlorophenol	40.000	39.009	2.5	87	0.00
45	S	2-Fluorobiphenyl	80.000	76.930	3.8	87	0.00
46		1,1'-Biphenyl	40.000	38.721	3.2	87	0.00
47		2-Chloronaphthalene	40.000	39.130	2.2	88	0.00

Sample Results: **BF141939.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	39.074	2.3	86	0.00
49		Acenaphthylene	40.000	39.005	2.5	87	0.00
50		Dimethylphthalate	40.000	38.106	4.7	87	0.00
51		2,6-Dinitrotoluene	40.000	39.096	2.3	86	0.00
52	C	Acenaphthene	40.000	39.326	1.7	88	0.00
53		3-Nitroaniline	40.000	38.530	3.7	84	0.00
54	P	2,4-Dinitrophenol	40.000	41.009	-2.5	95	0.00
55		Dibenzofuran	40.000	38.554	3.6	87	0.00
56	P	4-Nitrophenol	40.000	41.301	-3.3	87	0.00
57		2,4-Dinitrotoluene	40.000	39.193	2.0	86	0.00
58		Fluorene	40.000	38.195	4.5	87	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	39.708	0.7	88	0.00
60		Diethylphthalate	40.000	38.044	4.9	86	0.00
61		4-Chlorophenyl-phenylether	40.000	38.646	3.4	88	0.00
62		4-Nitroaniline	40.000	39.528	1.2	85	0.00
63		Azobenzene	40.000	37.888	5.3	85	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	87	0.00
65		4,6-Dinitro-2-methylphenol	40.000	41.862	-4.7	89	0.00
66	c	n-Nitrosodiphenylamine	40.000	38.576	3.6	87	0.00
67		4-Bromophenyl-phenylether	40.000	38.946	2.6	87	0.00
68		Hexachlorobenzene	40.000	39.530	1.2	87	0.00
69		Atrazine	40.000	33.164	17.1	87	0.00
70	C	Pentachlorophenol	40.000	42.480	-6.2	88	0.00
71		Phenanthrene	40.000	38.274	4.3	86	0.00
72		Anthracene	40.000	38.288	4.3	86	0.00
73		Carbazole	40.000	38.363	4.1	85	0.00
74		Di-n-butylphthalate	40.000	37.983	5.0	87	0.00
75	C	Fluoranthene	40.000	37.478	6.3	84	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	79	0.00
77		Benzidine	40.000	63.130	-57.8#	97	0.00
78		Pyrene	40.000	39.368	1.6	84	0.00
79	S	Terphenyl-d14	80.000	80.485	-0.6	84	0.00
80		Butylbenzylphthalate	40.000	38.339	4.2	78	-0.01
81		Benzo(a)anthracene	40.000	38.137	4.7	78	0.00
82		3,3'-Dichlorobenzidine	40.000	41.361	-3.4	84	0.00
83		Chrysene	40.000	39.214	2.0	81	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	36.599	8.5	75	-0.01
85	c	Di-n-octyl phthalate	40.000	37.251	6.9	76	-0.01
86	I	Perylene-d12	20.000	20.000	0.0	84	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	40.580	-1.4	86	-0.01
88		Benzo(b)fluoranthene	40.000	36.786	8.0	88	0.00
89		Benzo(k)fluoranthene	40.000	37.223	6.9	76	-0.01
90	C	Benzo(a)pyrene	40.000	38.240	4.4	83	-0.01
91		Dibenzo(a,h)anthracene	40.000	40.945	-2.4	86	-0.02
92		Benzo(g,h,i)perylene	40.000	41.473	-3.7	86	-0.01

Sample Results: **BF141939.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141939.D
Acq On : 13 Mar 2025 12:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 13:05:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG No.: Q1502

Instrument ID: BNA_P Calibration Date/Time: 03/13/2025 10:22

Lab File ID: BP024171.D Init. Calib. Date(s): 03/12/2025 03/12/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 16:17 21:02

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.253	1.221		-2.6	
Phenol-d6	1.675	1.634		-2.4	
Phenol	1.691	1.632		-3.5	20.0
2-Chlorophenol	1.353	1.312		-3.0	
2-Methylphenol	1.061	1.039		-2.1	
3+4-Methylphenols	1.457	1.437		-1.4	
Nitrobenzene-d5	0.354	0.352		-0.6	
2-Nitrophenol	0.167	0.169		1.2	20.0
2,4-Dimethylphenol	0.215	0.215		0.0	
2,4-Dichlorophenol	0.291	0.291		0.0	20.0
Benzoic acid	0.209	0.221		5.7	
4-Chloro-3-methylphenol	0.313	0.338		8.0	20.0
2,4,6-Trichlorophenol	0.367	0.367		0.0	20.0
2-Fluorobiphenyl	1.356	1.335		-1.5	
2,4,5-Trichlorophenol	0.403	0.406		0.7	
2,4-Dinitrophenol	0.156	0.164	0.050	5.1	
4-Nitrophenol	0.266	0.287	0.050	7.9	
4,6-Dinitro-2-methylphenol	0.119	0.117		-1.7	
2,4,6-Tribromophenol	0.252	0.256		1.6	
Pentachlorophenol	0.168	0.176		4.8	20.0
Terphenyl-d14	0.969	0.947		-2.3	
2,3,4,6-Tetrachlorophenol	0.356	0.358		0.6	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.764	152	330996	20.000	ng		0.00
21) Naphthalene-d8	10.546	136	1338415	20.000	ng		0.00
39) Acenaphthene-d10	14.410	164	811311	20.000	ng		0.00
64) Phenanthrene-d10	17.216	188	1560243	20.000	ng		0.01
76) Chrysene-d12	21.675	240	1695896	20.000	ng		0.02
86) Perylene-d12	25.063	264	1872320	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.375	112	1616314	77.949	ng		0.00
7) Phenol-d6	6.946	99	2163611	78.028	ng		0.00
23) Nitrobenzene-d5	8.911	82	1885293	79.475	ng		-0.01
42) 2,4,6-Tribromophenol	15.916	330	829790	81.197	ng		0.00
45) 2-Fluorobiphenyl	13.010	172	4332655	78.744	ng		0.00
79) Terphenyl-d14	19.939	244	6421493	78.162	ng		0.02
Target Compounds							Qvalue
2) 1,4-Dioxane	3.317	88	319012	38.812	ng		98
3) Pyridine	3.711	79	875478	38.962	ng		98
4) n-Nitrosodimethylamine	3.623	42	296615	39.518	ng		99
6) Aniline	7.099	93	996069	39.427	ng		99
8) 2-Chlorophenol	7.340	128	868343	38.770	ng		99
9) Benzaldehyde	6.917	77	487054	36.398	ng		99
10) Phenol	6.975	94	1080465	38.599	ng		99
11) bis(2-Chloroethyl)ether	7.193	93	844429	38.253	ng		98
12) 1,3-Dichlorobenzene	7.658	146	925362	38.313	ng		99
13) 1,4-Dichlorobenzene	7.799	146	938164	38.043	ng		99
14) 1,2-Dichlorobenzene	8.117	146	901143	38.079	ng		100
15) Benzyl Alcohol	8.005	79	692761	39.462	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.293	45	887389	39.387	ng		97
17) 2-Methylphenol	8.205	107	687510	39.167	ng		99
18) Hexachloroethane	8.834	117	334690	38.254	ng		98
19) n-Nitroso-di-n-propyla...	8.564	70	646163	40.168	ng		99
20) 3+4-Methylphenols	8.534	107	951529	39.457	ng		98
22) Acetophenone	8.581	105	1315436	38.875	ng	#	99
24) Nitrobenzene	8.958	77	926251	39.581	ng		99
25) Isophorone	9.475	82	1626564	40.000	ng		100
26) 2-Nitrophenol	9.663	139	451089	40.354	ng		99
27) 2,4-Dimethylphenol	9.722	122	574507	39.969	ng		100
28) bis(2-Chloroethoxy)met...	9.958	93	1123652	39.279	ng		99
29) 2,4-Dichlorophenol	10.199	162	779715	40.100	ng		99
30) 1,2,4-Trichlorobenzene	10.405	180	827250	38.508	ng		99
31) Naphthalene	10.599	128	2770811	39.054	ng		99
32) Benzoic acid	9.863	122	590853	42.149	ng		99
33) 4-Chloroaniline	10.705	127	1033809	40.717	ng		100
34) Hexachlorobutadiene	10.881	225	470663	38.068	ng		98
35) Caprolactam	11.493	113	264873	41.564	ng		98
36) 4-Chloro-3-methylphenol	11.834	107	905741	43.180	ng		97
37) 2-Methylnaphthalene	12.210	142	1866035	38.932	ng		99
38) 1-Methylnaphthalene	12.428	142	1840085	39.471	ng		99
40) 1,2,4,5-Tetrachloroben...	12.581	216	910390	38.818	ng		100
41) Hexachlorocyclopentadiene	12.563	237	332187	39.294	ng		99
43) 2,4,6-Trichlorophenol	12.822	196	595448	39.951	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.899	196	658754	40.302	ng	99
46) 1,1'-Biphenyl	13.228	154	2372057	38.908	ng	99
47) 2-Chloronaphthalene	13.269	162	1800527	39.264	ng	100
48) 2-Nitroaniline	13.475	65	494971	42.201	ng	97
49) Acenaphthylene	14.122	152	2785296	40.141	ng	100
50) Dimethylphthalate	13.857	163	2249113	39.619	ng	100
51) 2,6-Dinitrotoluene	13.975	165	486661	40.744	ng	98
52) Acenaphthene	14.475	154	1749122	39.454	ng	99
53) 3-Nitroaniline	14.304	138	539214	41.797	ng	97
54) 2,4-Dinitrophenol	14.522	184	266632	42.187	ng	98
55) Dibenzofuran	14.810	168	2848254	39.663	ng	100
56) 4-Nitrophenol	14.634	139	465657	43.198	ng	99
57) 2,4-Dinitrotoluene	14.781	165	658978	41.767	ng	99
58) Fluorene	15.475	166	2244338	40.112	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.040	232	580529	40.199	ng	99
60) Diethylphthalate	15.246	149	2293687	40.573	ng	100
61) 4-Chlorophenyl-phenyle...	15.469	204	1090107	39.723	ng	100
62) 4-Nitroaniline	15.493	138	582707	43.066	ng	98
63) Azobenzene	15.769	77	2182033	40.952	ng	98
65) 4,6-Dinitro-2-methylph...	15.551	198	364570	39.360	ng	98
66) n-Nitrosodiphenylamine	15.693	169	1855220	39.157	ng	100
67) 4-Bromophenyl-phenylether	16.387	248	665669	38.860	ng	98
68) Hexachlorobenzene	16.498	284	791231	39.025	ng	97
69) Atrazine	16.669	200	412759	36.093	ng	98
70) Pentachlorophenol	16.857	266	548868	41.861	ng	99
71) Phenanthrene	17.257	178	3325589	39.016	ng	100
72) Anthracene	17.357	178	3282068	39.535	ng	100
73) Carbazole	17.634	167	3241622	40.491	ng	100
74) Di-n-butylphthalate	18.222	149	3850848	40.519	ng	100
75) Fluoranthene	19.345	202	3926031	39.770	ng	100
77) Benzidine	19.545	184	1055531	56.927	ng	100
78) Pyrene	19.722	202	4107543	38.959	ng	100
80) Butylbenzylphthalate	20.675	149	1748418	40.213	ng	99
81) Benzo(a)anthracene	21.657	228	4146350	39.325	ng	100
82) 3,3'-Dichlorobenzidine	21.569	252	1449835	41.027	ng	100
83) Chrysene	21.722	228	3917287	38.647	ng	99
84) Bis(2-ethylhexyl)phtha...	21.592	149	2637923	40.209	ng	100
85) Di-n-octyl phthalate	22.898	149	4257281	40.502	ng	99
87) Indeno(1,2,3-cd)pyrene	28.939	276	5103421	39.055	ng	94
88) Benzo(b)fluoranthene	23.998	252	4409444	39.060	ng	98
89) Benzo(k)fluoranthene	24.080	252	4326170	39.177	ng	99
90) Benzo(a)pyrene	24.921	252	3836294	39.080	ng	99
91) Dibenzo(a,h)anthracene	29.015	278	4295311	39.510	ng	99
92) Benzo(g,h,i)perylene	30.145	276	4295501	38.869	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.764 min Scan# 7
Delta R.T. -0.005 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 330996

Ion	Ratio	Lower	Upper
152	100		
150	155.6	125.5	188.3
115	59.5	46.3	69.5







#2
1,4-Dioxane
Concen: 38.812 ng
RT: 3.317 min Scan# 39
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion: 88 Resp: 319012

Ion	Ratio	Lower	Upper
88	100		
58	63.7	49.8	74.6
43	22.2	17.3	25.9







#3
Pyridine
Concen: 38.962 ng
RT: 3.711 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 875478

Ion	Ratio	Lower	Upper
79	100		
52	56.6	43.8	65.8
51	27.8	21.6	32.4







#4
n-Nitrosodimethylamine
Concen: 39.518 ng
RT: 3.623 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion: 42 Resp: 296615

Ion	Ratio	Lower	Upper
42	100		
74	173.7	139.8	209.6
44	7.4	5.3	7.9







#5
2-Fluorophenol
Concen: 77.949 ng
RT: 5.375 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 1616314

Ion	Ratio	Lower	Upper
112	100		
64	52.8	41.3	61.9
63	26.7	20.6	31.0







#6
Aniline
Concen: 39.427 ng
RT: 7.099 min Scan# 682
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion: 93 Resp: 996069

Ion	Ratio	Lower	Upper
93	100		
66	35.8	28.2	42.2
65	18.0	14.0	21.0







#7
Phenol-d6
Concen: 78.028 ng
RT: 6.946 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 2163611

Ion	Ratio	Lower	Upper
99	100		
42	13.4	10.6	16.0
71	29.2	23.3	34.9







#8
2-Chlorophenol
Concen: 38.770 ng
RT: 7.340 min Scan# 723
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:128 Resp: 868343

Ion	Ratio	Lower	Upper
128	100		
130	32.4	12.0	52.0
64	42.0	21.6	61.6







#9
Benzaldehyde
Concen: 36.398 ng
RT: 6.917 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 487054

Ion	Ratio	Lower	Upper
77	100		
105	101.2	81.4	121.4
106	98.7	81.1	121.1







#10
Phenol
Concen: 38.599 ng
RT: 6.975 min Scan# 661
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion: 94 Resp: 1080465

Ion	Ratio	Lower	Upper
94	100		
65	24.8	4.9	44.9
66	31.4	11.9	51.9







#11
bis(2-Chloroethyl)ether
Concen: 38.253 ng
RT: 7.193 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 844429

Ion	Ratio	Lower	Upper
93	100		
63	69.1	47.4	87.4
95	32.5	12.1	52.1







#12
1,3-Dichlorobenzene
Concen: 38.313 ng
RT: 7.658 min Scan# 777
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:146 Resp: 925362

Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.6	77.4
75	28.3	22.8	34.2







#13
1,4-Dichlorobenzene
Concen: 38.043 ng
RT: 7.799 min Scan# 8
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 938164

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.9	77.9
111	40.2	31.4	47.2







#14
1,2-Dichlorobenzene
Concen: 38.079 ng
RT: 8.117 min Scan# 855
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:146 Resp: 901143

Ion	Ratio	Lower	Upper
146	100		
148	64.7	51.4	77.2
111	41.5	33.2	49.8







#15
Benzyl Alcohol
Concen: 39.462 ng
RT: 8.005 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 692761

Ion	Ratio	Lower	Upper
79	100		
108	82.3	66.5	99.7
77	63.9	50.9	76.3







#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.387 ng
RT: 8.293 min Scan# 885
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion: 45 Resp: 887389

Ion	Ratio	Lower	Upper
45	100		
77	17.9	0.0	39.0
79	14.4	0.0	35.6







#17
2-Methylphenol
Concen: 39.167 ng
RT: 8.205 min Scan# 8
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 687510

Ion	Ratio	Lower	Upper
107	100		
108	111.8	90.0	135.0
77	42.1	34.0	51.0
79	41.3	32.7	49.1







#18
Hexachloroethane
Concen: 38.254 ng
RT: 8.834 min Scan# 977
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:117 Resp: 334690

Ion	Ratio	Lower	Upper
117	100		
119	95.8	77.4	116.0
201	97.8	76.6	114.8







#19
n-Nitroso-di-n-propylamine
Concen: 40.168 ng
RT: 8.564 min Scan# 9
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 646163

Ion	Ratio	Lower	Upper
70	100		
42	45.5	36.0	54.0
101	10.2	8.3	12.5
130	23.4	18.6	28.0







#20
3+4-Methylphenols
Concen: 39.457 ng
RT: 8.534 min Scan# 926
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:107 Resp: 951529

Ion	Ratio	Lower	Upper
107	100		
108	92.0	69.7	109.7
77	31.2	10.3	50.3
79	27.8	6.1	46.1







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.546 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 1338415

Ion	Ratio	Lower	Upper
136	100		
137	10.7	9.0	13.4
54	6.6	5.3	7.9
68	6.4	5.0	7.4







#22
Acetophenone
Concen: 38.875 ng
RT: 8.581 min Scan# 934
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:105 Resp: 1315436

Ion	Ratio	Lower	Upper
105	100		
71	2.1	2.4	3.6#
51	20.8	16.2	24.2
120	24.0	19.0	28.6







#23
Nitrobenzene-d5
Concen: 79.475 ng
RT: 8.911 min Scan# 9
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 1885293

Ion	Ratio	Lower	Upper
82	100		
128	46.1	37.4	56.0
54	40.6	32.5	48.7







#24
Nitrobenzene
Concen: 39.581 ng
RT: 8.958 min Scan# 998
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion: 77 Resp: 926251

Ion	Ratio	Lower	Upper
77	100		
123	48.9	39.4	59.2
65	12.4	9.9	14.9







#25
Isophorone
Concen: 40.000 ng
RT: 9.475 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 1626564

Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.3	7.9
138	18.2	14.8	22.2







#26
2-Nitrophenol
Concen: 40.354 ng
RT: 9.663 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:139 Resp: 451089

Ion	Ratio	Lower	Upper
139	100		
109	26.9	21.7	32.5
65	45.6	35.8	53.8







#27
2,4-Dimethylphenol
Concen: 39.969 ng
RT: 9.722 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 574507

Ion	Ratio	Lower	Upper
122	100		
107	111.4	89.3	133.9
121	59.0	47.5	71.3







#28
bis(2-Chloroethoxy)methane
Concen: 39.279 ng
RT: 9.958 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion: 93 Resp: 1123652

Ion	Ratio	Lower	Upper
93	100		
95	33.0	25.9	38.9
123	11.8	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 40.100 ng
RT: 10.199 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 779715

Ion	Ratio	Lower	Upper
162	100		
164	64.8	44.2	84.2
98	35.3	15.1	55.1







#30
1,2,4-Trichlorobenzene
Concen: 38.508 ng
RT: 10.405 min Scan# 1244
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:180 Resp: 827250

Ion	Ratio	Lower	Upper
180	100		
182	96.5	76.1	114.1
145	30.6	24.2	36.2







#31
Naphthalene
Concen: 39.054 ng
RT: 10.599 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 2770811

Ion	Ratio	Lower	Upper
128	100		
129	10.6	8.7	13.1
127	13.2	10.4	15.6







#32
Benzoic acid
Concen: 42.149 ng
RT: 9.863 min Scan# 1152
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:122 Resp: 590853

Ion	Ratio	Lower	Upper
122	100		
105	121.4	100.3	140.3
77	87.1	69.1	109.1







#33
4-Chloroaniline
Concen: 40.717 ng
RT: 10.705 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 1033809

Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.8	38.6
65	27.0	21.2	31.8
92	18.0	14.4	21.6







#34
Hexachlorobutadiene
Concen: 38.068 ng
RT: 10.881 min Scan# 1325
Delta R.T. -0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:225 Resp: 470663

Ion	Ratio	Lower	Upper
225	100		
223	63.3	49.9	74.9
227	63.6	52.2	78.2







#35
Caprolactam
Concen: 41.564 ng
RT: 11.493 min Scan# 14
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 264873

Ion	Ratio	Lower	Upper
113	100		
55	131.1	113.6	153.6
56	98.5	80.0	120.0







#36
4-Chloro-3-methylphenol
Concen: 43.180 ng
RT: 11.834 min Scan# 1487
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:107 Resp: 905741

Ion	Ratio	Lower	Upper
107	100		
144	26.9	21.6	32.4
142	84.6	65.4	98.0







#37
2-Methylnaphthalene
Concen: 38.932 ng
RT: 12.210 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 1866035

Ion	Ratio	Lower	Upper
142	100		
141	89.4	70.3	105.5
115	32.9	26.0	39.0







#38
1-Methylnaphthalene
Concen: 39.471 ng
RT: 12.428 min Scan# 1588
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:142 Resp: 1840085

Ion	Ratio	Lower	Upper
142	100		
141	91.2	73.9	110.9
116	3.5	2.8	4.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 811311

Ion	Ratio	Lower	Upper
164	100		
162	100.9	80.6	121.0
160	45.0	35.4	53.2







#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.818 ng
RT: 12.581 min Scan# 1614
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:216 Resp: 910390

Ion	Ratio	Lower	Upper
216	100		
214	78.6	62.8	94.2
179	20.8	16.8	25.2
108	18.5	14.7	22.1







#41
Hexachlorocyclopentadiene
Concen: 39.294 ng
RT: 12.563 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 332187

Ion	Ratio	Lower	Upper
237	100		
235	62.6	41.8	81.8
272	13.9	0.0	33.5







#42
2,4,6-Tribromophenol
Concen: 81.197 ng
RT: 15.916 min Scan# 2181
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:330 Resp: 829790

Ion	Ratio	Lower	Upper
330	100		
332	97.4	77.6	116.4
141	31.7	25.0	37.6







#43
2,4,6-Trichlorophenol
Concen: 39.951 ng
RT: 12.822 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 595448

Ion	Ratio	Lower	Upper
196	100		
198	95.4	76.6	115.0
200	30.8	25.4	38.0







#44
2,4,5-Trichlorophenol
Concen: 40.302 ng
RT: 12.899 min Scan# 1668
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:196 Resp: 658754

Ion	Ratio	Lower	Upper
196	100		
198	95.9	76.5	114.7
97	49.1	38.7	58.1
132	28.1	22.2	33.2







#45
2-Fluorobiphenyl
Concen: 78.744 ng
RT: 13.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 4332655

Ion	Ratio	Lower	Upper
172	100		
171	35.5	28.3	42.5
170	23.6	19.0	28.4







#46
1,1'-Biphenyl
Concen: 38.908 ng
RT: 13.228 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:154 Resp: 2372057

Ion	Ratio	Lower	Upper
154	100		
153	39.8	20.3	60.3
76	13.8	0.0	32.9







#47
2-Chloronaphthalene
Concen: 39.264 ng
RT: 13.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 1800527

Ion	Ratio	Lower	Upper
162	100		
127	39.2	31.5	47.3
164	33.0	26.6	39.8







#48
2-Nitroaniline
Concen: 42.201 ng
RT: 13.475 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion: 65 Resp: 494971

Ion	Ratio	Lower	Upper
65	100		
92	76.6	62.1	93.1
138	119.8	99.5	149.3







#49
Acenaphthylene
Concen: 40.141 ng
RT: 14.122 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 2785296

Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.0	24.0
153	13.3	10.6	15.8







#50
Dimethylphthalate
Concen: 39.619 ng
RT: 13.857 min Scan# 1831
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:163 Resp: 2249113

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	10.4	8.2	12.4







#51
2,6-Dinitrotoluene
Concen: 40.744 ng
RT: 13.975 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 486661

Ion	Ratio	Lower	Upper
165	100		
63	46.2	37.7	56.5
89	51.0	41.9	62.9







#52
Acenaphthene
Concen: 39.454 ng
RT: 14.475 min Scan# 1936
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:154 Resp: 1749122

Ion	Ratio	Lower	Upper
154	100		
153	111.3	90.5	135.7
152	53.4	43.1	64.7







#53
3-Nitroaniline
Concen: 41.797 ng
RT: 14.304 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 539214

Ion	Ratio	Lower	Upper
138	100		
108	10.1	8.2	12.2
92	112.2	86.7	130.1







#54
2,4-Dinitrophenol
Concen: 42.187 ng
RT: 14.522 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:184 Resp: 266632

Ion	Ratio	Lower	Upper
184	100		
63	55.0	42.1	63.1
154	55.3	44.6	67.0







#55
Dibenzofuran
Concen: 39.663 ng
RT: 14.810 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 2848254

Ion	Ratio	Lower	Upper
168	100		
139	36.5	29.1	43.7
169	13.6	10.8	16.2







#56
4-Nitrophenol
Concen: 43.198 ng
RT: 14.634 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:139 Resp: 465657

Ion	Ratio	Lower	Upper
139	100		
109	63.6	43.3	83.3
65	83.1	61.2	101.2







#57
2,4-Dinitrotoluene
Concen: 41.767 ng
RT: 14.781 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 658978

Ion	Ratio	Lower	Upper
165	100		
63	37.4	29.0	43.6
89	67.9	54.2	81.4
182	2.7	2.3	3.5







#58
Fluorene
Concen: 40.112 ng
RT: 15.475 min Scan# 2106
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:166 Resp: 2244338

Ion	Ratio	Lower	Upper
166	100		
165	97.3	77.6	116.4
167	13.5	10.8	16.2







#59
2,3,4,6-Tetrachlorophenol
Concen: 40.199 ng
RT: 15.040 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 580529

Ion	Ratio	Lower	Upper
232	100		
131	45.7	35.9	53.9
130	2.3	1.8	2.6
166	30.7	24.2	36.2







#60
Diethylphthalate
Concen: 40.573 ng
RT: 15.246 min Scan# 2067
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:149 Resp: 2293687

Ion	Ratio	Lower	Upper
149	100		
177	22.1	17.7	26.5
150	12.6	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 39.723 ng
RT: 15.469 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 1090107

Ion	Ratio	Lower	Upper
204	100		
206	33.2	26.4	39.6
141	57.2	45.8	68.8







#62
4-Nitroaniline
Concen: 43.066 ng
RT: 15.493 min Scan# 2109
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:138 Resp: 582707

Ion	Ratio	Lower	Upper
138	100		
92	47.3	28.7	68.7
108	67.2	49.5	89.5







#63
Azobenzene
Concen: 40.952 ng
RT: 15.769 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 2182033

Ion	Ratio	Lower	Upper
77	100		
182	25.0	7.7	47.7
105	21.8	2.3	42.3
51	23.2	3.2	43.2







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.216 min Scan# 2402
Delta R.T. 0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:188 Resp: 1560243

Ion	Ratio	Lower	Upper
188	100		
94	9.9	8.1	12.1
80	11.0	8.9	13.3







#65
4,6-Dinitro-2-methylphenol
Concen: 39.360 ng
RT: 15.551 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 364570

Ion	Ratio	Lower	Upper
198	100		
51	33.8	14.7	54.7
105	42.5	23.8	63.8







#66
n-Nitrosodiphenylamine
Concen: 39.157 ng
RT: 15.693 min Scan# 2143
Delta R.T. 0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:169 Resp: 1855220

Ion	Ratio	Lower	Upper
169	100		
168	66.2	52.9	79.3
167	36.0	28.8	43.2







#67
4-Bromophenyl-phenylether
Concen: 38.860 ng
RT: 16.387 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 665669

Ion	Ratio	Lower	Upper
248	100		
250	97.4	78.4	117.6
141	70.6	59.4	89.0







#68
Hexachlorobenzene
Concen: 39.025 ng
RT: 16.498 min Scan# 2280
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:284 Resp: 791231

Ion	Ratio	Lower	Upper
284	100		
142	34.1	25.4	38.2
249	30.0	23.5	35.3







#69
Atrazine
Concen: 36.093 ng
RT: 16.669 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 412759

Ion	Ratio	Lower	Upper
200	100		
173	27.2	7.6	47.6
215	50.4	29.0	69.0







#70
Pentachlorophenol
Concen: 41.861 ng
RT: 16.857 min Scan# 2341
Delta R.T. 0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:266 Resp: 548868

Ion	Ratio	Lower	Upper
266	100		
268	64.3	51.3	76.9
264	62.7	50.6	76.0







#71
Phenanthrene
Concen: 39.016 ng
RT: 17.257 min Scan# 24
Delta R.T. 0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 3325589

Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.6	23.4
179	15.5	12.5	18.7







#72
Anthracene
Concen: 39.535 ng
RT: 17.357 min Scan# 2426
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:178 Resp: 3282068

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.2	22.8
179	15.5	12.6	18.8







#73
Carbazole
Concen: 40.491 ng
RT: 17.634 min Scan# 24
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 3241622

Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.1	25.7
139	12.9	10.2	15.4







#74
Di-n-butylphthalate
Concen: 40.519 ng
RT: 18.222 min Scan# 2573
Delta R.T. 0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:149 Resp: 3850848

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.6	11.4
104	4.9	3.9	5.9







#75
Fluoranthene
Concen: 39.770 ng
RT: 19.345 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 3926031

Ion	Ratio	Lower	Upper
202	100		
101	12.8	0.0	32.5
203	17.8	0.0	37.9







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.675 min Scan# 3160
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:240 Resp: 1695896

Ion	Ratio	Lower	Upper
240	100		
120	12.0	9.2	13.8
236	25.4	20.4	30.6







#77
Benzidine
Concen: 56.927 ng
RT: 19.545 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 1055531

Ion	Ratio	Lower	Upper
184	100		
185	14.6	11.8	17.6
183	11.6	9.1	13.7







#78
Pyrene
Concen: 38.959 ng
RT: 19.722 min Scan# 2828
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:202 Resp: 4107543

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.1	25.7
203	17.9	14.2	21.4







#79
Terphenyl-d14
Concen: 78.162 ng
RT: 19.939 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 6421493

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.7	8.5
122	12.2	9.6	14.4







#80
Butylbenzylphthalate
Concen: 40.213 ng
RT: 20.675 min Scan# 2990
Delta R.T. 0.024 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:149 Resp: 1748418

Ion	Ratio	Lower	Upper
149	100		
91	68.5	55.3	82.9
206	21.5	16.1	24.1







#81
Benzo(a)anthracene
Concen: 39.325 ng
RT: 21.657 min Scan# 3
Delta R.T. 0.024 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 4146350

Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.8	32.8
229	19.9	16.0	24.0







#82
3,3'-Dichlorobenzidine
Concen: 41.027 ng
RT: 21.569 min Scan# 3142
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:252 Resp: 1449835

Ion	Ratio	Lower	Upper
252	100		
254	64.3	51.4	77.2
126	12.1	9.3	13.9







#83
Chrysene
Concen: 38.647 ng
RT: 21.722 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 3917287

Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.5	36.7
229	20.0	16.2	24.2







#84
Bis(2-ethylhexyl)phthalate
Concen: 40.209 ng
RT: 21.592 min Scan# 3146
Delta R.T. 0.024 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:149 Resp: 2637923

Ion	Ratio	Lower	Upper
149	100		
167	28.1	22.5	33.7
279	5.0	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 40.502 ng
RT: 22.898 min Scan# 3
Delta R.T. 0.029 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 4257281

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	7.5	5.8	8.8







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.063 min Scan# 3736
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:264 Resp: 1872320

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.5	27.7
265	22.2	17.7	26.5







#87
Indeno(1,2,3-cd)pyrene
Concen: 39.055 ng
RT: 28.939 min Scan# 4
Delta R.T. 0.012 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 5103421

Ion	Ratio	Lower	Upper
276	100		
138	16.9	13.7	20.5
277	20.6	20.2	30.2







#88
Benzo(b)fluoranthene
Concen: 39.060 ng
RT: 23.998 min Scan# 3555
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:252 Resp: 4409444

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.2	25.8
125	11.1	8.2	12.4







#89
Benzo(k)fluoranthene
Concen: 39.177 ng
RT: 24.080 min Scan# 3
Delta R.T. 0.029 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 4326170

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	10.6	9.1	13.7







#90
Benzo(a)pyrene
Concen: 39.080 ng
RT: 24.921 min Scan# 3712
Delta R.T. 0.029 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:252 Resp: 3836294

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.0
125	11.6	9.6	14.4







#91
Dibenzo(a,h)anthracene
Concen: 39.510 ng
RT: 29.015 min Scan# 4
Delta R.T. 0.006 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 4295311

Ion	Ratio	Lower	Upper
278	100		
139	16.7	13.0	19.4
279	24.3	18.8	28.2






#92
Benzo(g,h,i)perylene
Concen: 38.869 ng
RT: 30.145 min Scan# 4600
Delta R.T. 0.018 min
Lab File: BP024171.D
Acq: 13 Mar 2025 10:22

Tgt Ion:276 Resp: 4295501

Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.2	28.8
138	22.7	17.4	26.2



Sample Results: **BP024171.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2		1,4-Dioxane	0.497	0.482	3.0	89	0.00
3		Pyridine	1.358	1.322	2.7	87	0.00
4		n-Nitrosodimethylamine	0.454	0.448	1.3	90	0.00
5	S	2-Fluorophenol	1.253	1.221	2.6	87	0.00
6		Aniline	1.527	1.505	1.4	88	-0.01
7	S	Phenol-d6	1.675	1.634	2.4	86	0.00
8		2-Chlorophenol	1.353	1.312	3.0	87	0.00
9		Benzaldehyde	0.809	0.736	9.0	81	0.00
10	C	Phenol	1.691	1.632	3.5	87	0.00
11		bis(2-Chloroethyl)ether	1.334	1.276	4.3	87	0.00
12		1,3-Dichlorobenzene	1.459	1.398	4.2	87	0.00
13	C	1,4-Dichlorobenzene	1.490	1.417	4.9	87	-0.01
14		1,2-Dichlorobenzene	1.430	1.361	4.8	87	0.00
15		Benzyl Alcohol	1.061	1.046	1.4	87	0.00
16		2,2'-oxybis(1-Chloropropane	1.361	1.340	1.5	90	-0.01
17		2-Methylphenol	1.061	1.039	2.1	86	-0.01
18		Hexachloroethane	0.529	0.506	4.3	86	-0.01
19	P	n-Nitroso-di-n-propylamine	0.972	0.976	-0.4	88	-0.01
20		3+4-Methylphenols	1.457	1.437	1.4	87	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
22		Acetophenone	0.506	0.491	3.0	86	0.00
23	S	Nitrobenzene-d5	0.354	0.352	0.6	87	-0.01
24		Nitrobenzene	0.350	0.346	1.1	88	0.00
25		Isophorone	0.608	0.608	0.0	88	-0.01
26	C	2-Nitrophenol	0.167	0.169	-1.2	86	0.00
27		2,4-Dimethylphenol	0.215	0.215	0.0	87	-0.01
28		bis(2-Chloroethoxy)methane	0.427	0.420	1.6	87	0.00
29	C	2,4-Dichlorophenol	0.291	0.291	0.0	85	0.00
30		1,2,4-Trichlorobenzene	0.321	0.309	3.7	86	-0.01
31		Naphthalene	1.060	1.035	2.4	87	0.00
32		Benzoic acid	0.209	0.221	-5.7	88	-0.01
33		4-Chloroaniline	0.379	0.386	-1.8	89	0.00
34	C	Hexachlorobutadiene	0.185	0.176	4.9	85	-0.01
35		Caprolactam	0.095	0.099	-4.2	89	0.00
36	C	4-Chloro-3-methylphenol	0.313	0.338	-8.0	94	0.00
37		2-Methylnaphthalene	0.716	0.697	2.7	86	0.00
38		1-Methylnaphthalene	0.697	0.687	1.4	89	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
40		1,2,4,5-Tetrachlorobenzene	0.578	0.561	2.9	88	0.00
41	P	Hexachlorocyclopentadiene	0.208	0.205	1.4	84	0.00
42	S	2,4,6-Tribromophenol	0.252	0.256	-1.6	88	0.00
43	C	2,4,6-Trichlorophenol	0.367	0.367	0.0	86	0.00
44		2,4,5-Trichlorophenol	0.403	0.406	-0.7	87	0.00
45	S	2-Fluorobiphenyl	1.356	1.335	1.5	88	0.00
46		1,1'-Biphenyl	1.503	1.462	2.7	87	0.00
47		2-Chloronaphthalene	1.130	1.110	1.8	88	0.00

Sample Results: **BP024171.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.289	0.305	-5.5	91	0.00
49		Acenaphthylene	1.710	1.717	-0.4	90	0.00
50		Dimethylphthalate	1.399	1.386	0.9	89	0.00
51		2,6-Dinitrotoluene	0.294	0.300	-2.0	89	0.00
52	C	Acenaphthene	1.093	1.078	1.4	89	0.00
53		3-Nitroaniline	0.318	0.332	-4.4	90	0.00
54	P	2,4-Dinitrophenol	0.156	0.164	-5.1	90	0.00
55		Dibenzofuran	1.770	1.755	0.8	89	0.00
56	P	4-Nitrophenol	0.266	0.287	-7.9	90	0.00
57		2,4-Dinitrotoluene	0.389	0.406	-4.4	90	0.00
58		Fluorene	1.379	1.383	-0.3	90	0.00
59		2,3,4,6-Tetrachlorophenol	0.356	0.358	-0.6	88	0.00
60		Diethylphthalate	1.394	1.414	-1.4	91	0.00
61		4-Chlorophenyl-phenylether	0.676	0.672	0.6	89	0.00
62		4-Nitroaniline	0.334	0.359	-7.5	91	0.00
63		Azobenzene	1.314	1.345	-2.4	91	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	91	0.01
65		4,6-Dinitro-2-methylphenol	0.119	0.117	1.7	88	0.00
66	c	n-Nitrosodiphenylamine	0.607	0.595	2.0	90	0.01
67		4-Bromophenyl-phenylether	0.220	0.213	3.2	89	0.01
68		Hexachlorobenzene	0.260	0.254	2.3	91	0.00
69		Atrazine	0.147	0.132	10.2	89	0.01
70	C	Pentachlorophenol	0.168	0.176	-4.8	90	0.01
71		Phenanthrene	1.093	1.066	2.5	89	0.01
72		Anthracene	1.064	1.052	1.1	90	0.02
73		Carbazole	1.026	1.039	-1.3	90	0.02
74		Di-n-butylphthalate	1.218	1.234	-1.3	89	0.01
75	C	Fluoranthene	1.265	1.258	0.6	90	0.02
76	I	Chrysene-d12	1.000	1.000	0.0	90	0.02
77		Benzidine	0.219	0.311	-42.0#	109	0.02
78		Pyrene	1.243	1.211	2.6	91	0.02
79	S	Terphenyl-d14	0.969	0.947	2.3	90	0.02
80		Butylbenzylphthalate	0.513	0.515	-0.4	89	0.02
81		Benzo(a)anthracene	1.243	1.222	1.7	91	0.02
82		3,3'-Dichlorobenzidine	0.417	0.427	-2.4	90	0.02
83		Chrysene	1.195	1.155	3.3	90	0.02
84		Bis(2-ethylhexyl)phthalate	0.774	0.778	-0.5	88	0.02
85	c	Di-n-octyl phthalate	1.240	1.255	-1.2	89	0.03
86	I	Perylene-d12	1.000	1.000	0.0	92	0.00
87		Indeno(1,2,3-cd)pyrene	1.396	1.363	2.4	91	0.01
88		Benzo(b)fluoranthene	1.206	1.178	2.3	91	0.02
89		Benzo(k)fluoranthene	1.180	1.155	2.1	94	0.03
90	C	Benzo(a)pyrene	1.049	1.024	2.4	91	0.03
91		Dibenzo(a,h)anthracene	1.161	1.147	1.2	92	0.00
92		Benzo(g,h,i)perylene	1.180	1.147	2.8	91	0.02

Sample Results: **BP024171.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Sample Results: **BP024171.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	86	0.00
2		1,4-Dioxane	40.000	38.812	3.0	89	0.00
3		Pyridine	40.000	38.962	2.6	87	0.00
4		n-Nitrosodimethylamine	40.000	39.518	1.2	90	0.00
5	S	2-Fluorophenol	80.000	77.949	2.6	87	0.00
6		Aniline	40.000	39.427	1.4	88	-0.01
7	S	Phenol-d6	80.000	78.028	2.5	86	0.00
8		2-Chlorophenol	40.000	38.770	3.1	87	0.00
9		Benzaldehyde	40.000	36.398	9.0	81	0.00
10	C	Phenol	40.000	38.599	3.5	87	0.00
11		bis(2-Chloroethyl)ether	40.000	38.253	4.4	87	0.00
12		1,3-Dichlorobenzene	40.000	38.313	4.2	87	0.00
13	C	1,4-Dichlorobenzene	40.000	38.043	4.9	87	-0.01
14		1,2-Dichlorobenzene	40.000	38.079	4.8	87	0.00
15		Benzyl Alcohol	40.000	39.462	1.3	87	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	39.387	1.5	90	-0.01
17		2-Methylphenol	40.000	39.167	2.1	86	-0.01
18		Hexachloroethane	40.000	38.254	4.4	86	-0.01
19	P	n-Nitroso-di-n-propylamine	40.000	40.168	-0.4	88	-0.01
20		3+4-Methylphenols	40.000	39.457	1.4	87	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22		Acetophenone	40.000	38.875	2.8	86	0.00
23	S	Nitrobenzene-d5	80.000	79.475	0.7	87	-0.01
24		Nitrobenzene	40.000	39.581	1.0	88	0.00
25		Isophorone	40.000	40.000	0.0	88	-0.01
26	C	2-Nitrophenol	40.000	40.354	-0.9	86	0.00
27		2,4-Dimethylphenol	40.000	39.969	0.1	87	-0.01
28		bis(2-Chloroethoxy)methane	40.000	39.279	1.8	87	0.00
29	C	2,4-Dichlorophenol	40.000	40.100	-0.3	85	0.00
30		1,2,4-Trichlorobenzene	40.000	38.508	3.7	86	-0.01
31		Naphthalene	40.000	39.054	2.4	87	0.00
32		Benzoic acid	40.000	42.149	-5.4	88	-0.01
33		4-Chloroaniline	40.000	40.717	-1.8	89	0.00
34	C	Hexachlorobutadiene	40.000	38.068	4.8	85	-0.01
35		Caprolactam	40.000	41.564	-3.9	89	0.00
36	C	4-Chloro-3-methylphenol	40.000	43.180	-7.9	94	0.00
37		2-Methylnaphthalene	40.000	38.932	2.7	86	0.00
38		1-Methylnaphthalene	40.000	39.471	1.3	89	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	87	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	38.818	3.0	88	0.00
41	P	Hexachlorocyclopentadiene	40.000	39.294	1.8	84	0.00
42	S	2,4,6-Tribromophenol	80.000	81.197	-1.5	88	0.00
43	C	2,4,6-Trichlorophenol	40.000	39.951	0.1	86	0.00
44		2,4,5-Trichlorophenol	40.000	40.302	-0.8	87	0.00
45	S	2-Fluorobiphenyl	80.000	78.744	1.6	88	0.00
46		1,1'-Biphenyl	40.000	38.908	2.7	87	0.00
47		2-Chloronaphthalene	40.000	39.264	1.8	88	0.00

Sample Results: **BP024171.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	42.201	-5.5	91	0.00
49		Acenaphthylene	40.000	40.141	-0.4	90	0.00
50		Dimethylphthalate	40.000	39.619	1.0	89	0.00
51		2,6-Dinitrotoluene	40.000	40.744	-1.9	89	0.00
52	C	Acenaphthene	40.000	39.454	1.4	89	0.00
53		3-Nitroaniline	40.000	41.797	-4.5	90	0.00
54	P	2,4-Dinitrophenol	40.000	42.187	-5.5	90	0.00
55		Dibenzofuran	40.000	39.663	0.8	89	0.00
56	P	4-Nitrophenol	40.000	43.198	-8.0	90	0.00
57		2,4-Dinitrotoluene	40.000	41.767	-4.4	90	0.00
58		Fluorene	40.000	40.112	-0.3	90	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	40.199	-0.5	88	0.00
60		Diethylphthalate	40.000	40.573	-1.4	91	0.00
61		4-Chlorophenyl-phenylether	40.000	39.723	0.7	89	0.00
62		4-Nitroaniline	40.000	43.066	-7.7	91	0.00
63		Azobenzene	40.000	40.952	-2.4	91	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	91	0.01
65		4,6-Dinitro-2-methylphenol	40.000	39.360	1.6	88	0.00
66	c	n-Nitrosodiphenylamine	40.000	39.157	2.1	90	0.01
67		4-Bromophenyl-phenylether	40.000	38.860	2.9	89	0.01
68		Hexachlorobenzene	40.000	39.025	2.4	91	0.00
69		Atrazine	40.000	36.093	9.8	89	0.01
70	C	Pentachlorophenol	40.000	41.861	-4.7	90	0.01
71		Phenanthrene	40.000	39.016	2.5	89	0.01
72		Anthracene	40.000	39.535	1.2	90	0.02
73		Carbazole	40.000	40.491	-1.2	90	0.02
74		Di-n-butylphthalate	40.000	40.519	-1.3	89	0.01
75	C	Fluoranthene	40.000	39.770	0.6	90	0.02
76	I	Chrysene-d12	20.000	20.000	0.0	90	0.02
77		Benzidine	40.000	56.927	-42.3#	109	0.02
78		Pyrene	40.000	38.959	2.6	91	0.02
79	S	Terphenyl-d14	80.000	78.162	2.3	90	0.02
80		Butylbenzylphthalate	40.000	40.213	-0.5	89	0.02
81		Benzo(a)anthracene	40.000	39.325	1.7	91	0.02
82		3,3'-Dichlorobenzidine	40.000	41.027	-2.6	90	0.02
83		Chrysene	40.000	38.647	3.4	90	0.02
84		Bis(2-ethylhexyl)phthalate	40.000	40.209	-0.5	88	0.02
85	c	Di-n-octyl phthalate	40.000	40.502	-1.3	89	0.03
86	I	Perylene-d12	20.000	20.000	0.0	92	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	39.055	2.4	91	0.01
88		Benzo(b)fluoranthene	40.000	39.060	2.3	91	0.02
89		Benzo(k)fluoranthene	40.000	39.177	2.1	94	0.03
90	C	Benzo(a)pyrene	40.000	39.080	2.3	91	0.03
91		Dibenzo(a,h)anthracene	40.000	39.510	1.2	92	0.00
92		Benzo(g,h,i)perylene	40.000	38.869	2.8	91	0.02

Sample Results: **BP024171.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024171.D
Acq On : 13 Mar 2025 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 11:49:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

QC SAMPLE DATA


Sample Results: **BF141896.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141896.D
Acq On : 10 Mar 2025 10:31
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Mar 10 15:46:22 2025




AutoFind: Scans 1625, 1626, 1627; Background Corrected with Scan 1617


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.1	121957	PASS
68	69	0.00	2	1.9	2599	PASS
69	198	0.00	100	32.0	134239	PASS
70	69	0.00	2	0.7	949	PASS
127	198	10	80	44.8	187627	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	419221	PASS
199	198	5	9	6.7	28280	PASS
275	198	10	60	31.2	130819	PASS
365	198	1	100	4.2	17803	PASS
441	198	0.01	100	19.9	83389	PASS
442	442	50	100	100.0	536640	PASS
443	442	15	24	19.2	102888	PASS


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141896.D
Acq On : 10 Mar 2025 10:31
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 10 16:16:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration







TIC: BF141896.D\data.ms

(70) Pentachlorophenol (C)

11.204min (-0.000) 118650.09 ng

response 544757


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	64.05
264.00	62.70	63.55
0.00	0.00	0.00


Sample Results: **BF141896.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
Data File : BF141896.D
Acq On : 10 Mar 2025 10:31
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 10 16:16:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration







TIC: BF141896.D\data.ms

(77) Benzidine

12.739min (+ 0.006) 203796.44 ng

response 2274676

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.38
183.00	12.10	12.32
0.00	0.00	0.00

Instrument :
BNA_F
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/10/2025	BNA_F	BF141896.D
Compound Name	**Response**	**Retention Time**
DDT	1358235	13.562
DDD	20094	13.292
DDE	930	12.921
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
21024	1379259	1.52


Sample Results: **BF141938.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141938.D
Acq On : 13 Mar 2025 11:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Mar 10 15:46:22 2025



AutoFind: Scans 1623, 1624, 1625; Background Corrected with Scan 1617


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.0	74819	PASS
68	69	0.00	2	1.8	1404	PASS
69	198	0.00	100	34.0	79429	PASS
70	69	0.00	2	0.6	451	PASS
127	198	10	80	47.3	110651	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	233877	PASS
199	198	5	9	6.6	15462	PASS
275	198	10	60	29.9	69864	PASS
365	198	1	100	4.1	9507	PASS
441	198	0.01	100	16.4	38323	PASS
442	442	50	100	100.0	251264	PASS
443	442	15	24	19.1	48053	PASS


Sample Results: **BF141938.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141938.D
Acq On : 13 Mar 2025 11:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 13 13:04:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration





TIC: BF141938.D\data.ms

(70) Pentachlorophenol (C)

11.192min (-0.012) 133427.78 ng

response 291016


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	62.65
264.00	62.70	61.35
0.00	0.00	0.00


Sample Results: **BF141938.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141938.D
Acq On : 13 Mar 2025 11:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 13 13:04:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration





TIC: BF141938.D\data.ms

(77) Benzidine

12.727min (-0.006) 527010.30 ng

response 1286738

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.29
183.00	12.10	12.13
0.00	0.00	0.00

Instrument :
BNA_F
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/13/2025	BNA_F	BF141938.D
Compound Name	**Response**	**Retention Time**
DDT	695260	13.551
DDD	12936	13.251
DDE	425	12.916
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
13361	708621	1.89


Sample Results: **BP024159.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024159.D
Acq On : 12 Mar 2025 15:36
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Mar 13 05:55:58 2025



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2472


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.5	117822	PASS
68	69	0.00	2	1.6	2280	PASS
69	198	0.00	100	33.9	139938	PASS
70	69	0.00	2	0.6	822	PASS
127	198	10	80	46.1	190399	PASS
197	198	0.00	2	0.4	1468	PASS
198	198	100	100	100.0	412716	PASS
199	198	5	9	6.8	28114	PASS
275	198	10	60	27.5	113624	PASS
365	198	1	100	3.8	15804	PASS
441	198	0.01	100	14.7	60679	PASS
442	442	100	100	100.0	391061	PASS
443	442	15	24	19.5	76081	PASS


Sample Results: **BP024159.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024159.D
Acq On : 12 Mar 2025 15:36
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Mar 13 05:56:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration







TIC: BP024159.D\data.ms

(70) Pentachlorophenol (C)

16.857min (0.000) 4607317.61 ng

response 542052


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.10	63.56
264.00	63.30	63.37
0.00	0.00	0.00


Sample Results: **BP024159.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031225\
Data File : BP024159.D
Acq On : 12 Mar 2025 15:36
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Mar 13 05:56:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration







TIC: BP024159.D\data.ms

(77) Benzidine

19.551min (0.000) 3117449.49 ng

response 3272103

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.81
183.00	11.40	11.59
0.00	0.00	0.00

Instrument :
BNA_P
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/12/2025	BNA_P	BP024159.D
Compound Name	**Response**	**Retention Time**
DDT	1535436	20.851
DDD	16722	20.439
DDE	443	19.827
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
17165	1552601	1.11


Sample Results: **BP024170.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024170.D
Acq On : 13 Mar 2025 09:41
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Mar 13 05:55:58 2025



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2472


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.0	118052	PASS
68	69	0.00	2	1.5	2218	PASS
69	198	0.00	100	35.2	143312	PASS
70	69	0.00	2	0.6	790	PASS
127	198	10	80	46.9	190703	PASS
197	198	0.00	2	0.1	548	PASS
198	198	100	100	100.0	406784	PASS
199	198	5	9	6.9	28034	PASS
275	198	10	60	27.6	112405	PASS
365	198	1	100	3.6	14767	PASS
441	198	0.01	100	14.2	57619	PASS
442	442	100	100	100.0	374803	PASS
443	442	15	24	19.2	71957	PASS


Sample Results: **BP024170.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024170.D
Acq On : 13 Mar 2025 09:41
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Mar 13 11:49:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration





TIC: BP024170.D\data.ms

(70) Pentachlorophenol (C)

16.863min (+ 0.018) 604215.30 ng

response 563611


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.10	63.70
264.00	63.30	62.79
0.00	0.00	0.00


Sample Results: **BP024170.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031325\
Data File : BP024170.D
Acq On : 13 Mar 2025 09:41
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Mar 13 11:49:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration





TIC: BP024170.D\data.ms

(77) Benzidine

19.551min (+ 0.029) 3326037.58 ng

response 3272849

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.62
183.00	11.40	11.51
0.00	0.00	0.00

Instrument :
BNA_P
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/13/2025	BNA_P	BP024170.D
Compound Name	**Response**	**Retention Time**
DDT	1542848	20.851
DDD	21899	20.375
DDE	1202	19.857
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
23101	1565949	1.48

Raw Data: **BF141940.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	
Project:	NJ Waste Water PT	Date Received:	
Client Sample ID:	PB167097BL	SDG No.:	Q1502
Lab Sample ID:	PB167097BL	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141940.D	1	03/12/25 08:45	03/13/25 12:47	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
108-95-2	Phenol	0.91	U	0.91	5.00	ug/L
95-57-8	2-Chlorophenol	0.58	U	0.58	5.00	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.10	U	1.10	10.0	ug/L
88-75-5	2-Nitrophenol	1.80	U	1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	1.90	U	1.90	5.00	ug/L
120-83-2	2,4-Dichlorophenol	0.52	U	0.52	5.00	ug/L
65-85-0	Benzoic acid	4.20	U	4.20	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	0.59	U	0.59	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.51	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	0.62	U	0.62	5.00	ug/L
51-28-5	2,4-Dinitrophenol	6.00	U	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	2.40	U	2.40	10.0	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	2.90	U	2.90	10.0	ug/L
87-86-5	Pentachlorophenol	1.60	U	1.60	10.0	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	0.72	U	0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	134		10 - 139	90%	SPK: 150
13127-88-3	Phenol-d6	130		10 - 134	86%	SPK: 150
118-79-6	2,4,6-Tribromophenol	154		44 - 137	102%	SPK: 150
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	177000	6.875			
1146-65-2	Naphthalene-d8	706000	8.157			
15067-26-2	Acenaphthene-d10	408000	9.91			
1517-22-2	Phenanthrene-d10	770000	11.392			
1719-03-5	Chrysene-d12	602000	14.027			
1520-96-3	Perylene-d12	427000	15.498			

Raw Data: **BF141940.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	
Project:	NJ Waste Water PT	Date Received:	
Client Sample ID:	PB167097BL	SDG No.:	Q1502
Lab Sample ID:	PB167097BL	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141940.D	1	03/12/25 08:45	03/13/25 12:47	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141940.D
Acq On : 13 Mar 2025 12:47
Operator : RC/JU
Sample : PB167097BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB167097BL

Quant Time: Mar 13 13:06:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	176773	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	705846	20.000	ng	0.00
39) Acenaphthene-d10	9.910	164	407739	20.000	ng	0.00
64) Phenanthrene-d10	11.392	188	769835	20.000	ng	-0.01
76) Chrysene-d12	14.027	240	601564	20.000	ng	-0.01
86) Perylene-d12	15.498	264	427331	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	1423729	134.418	ng	0.01
7) Phenol-d6	6.498	99	1748591	129.663	ng	0.00
23) Nitrobenzene-d5	7.434	82	1170774	93.344	ng	-0.01
42) 2,4,6-Tribromophenol	10.698	330	794330	153.562	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	2412754	89.992	ng	-0.01
79) Terphenyl-d14	12.986	244	3551821	87.293	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141940.D
Acq On : 13 Mar 2025 12:47
Operator : RC/JU
Sample : PB167097BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB167097BL

Quant Time: Mar 13 13:06:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Instrument :
BNA_F
ClientSampleId :
PB167097BL

Tgt Ion:152 Resp: 176773

Ion	Ratio	Lower	Upper
152	100		
150	156.9	127.4	191.2
115	65.2	51.9	77.9







#5
2-Fluorophenol
Concen: 134.418 ng
RT: 5.504 min Scan# 580
Delta R.T. 0.012 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Tgt Ion:112 Resp: 1423729

Ion	Ratio	Lower	Upper
112	100		
64	57.8	46.5	69.7
63	31.8	25.4	38.2







#7
Phenol-d6
Concen: 129.663 ng
RT: 6.498 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Instrument :
BNA_F
ClientSampleId :
PB167097BL

Tgt Ion: 99 Resp: 1748591

Ion	Ratio	Lower	Upper
99	100		
42	17.8	14.6	21.8
71	39.1	30.8	46.2







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Tgt Ion:136 Resp: 705846

Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.8	13.2
54	6.7	5.8	8.8
68	4.7	4.1	6.1







#23
Nitrobenzene-d5
Concen: 93.344 ng
RT: 7.434 min Scan# 9
Delta R.T. -0.012 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Instrument :
BNA_F
ClientSampleId :
PB167097BL

Tgt Ion: 82 Resp: 1170774

Ion	Ratio	Lower	Upper
82	100		
128	44.3	36.0	54.0
54	48.6	39.6	59.4







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.910 min Scan# 1329
Delta R.T. -0.006 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Tgt Ion:164 Resp: 407739

Ion	Ratio	Lower	Upper
164	100		
162	102.6	81.8	122.6
160	45.7	36.7	55.1







#42
2,4,6-Tribromophenol
Concen: 153.562 ng
RT: 10.698 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Instrument :
BNA_F
ClientSampleId :
PB167097BL

Tgt Ion:330 Resp: 794330

Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.6	116.4
141	28.1	24.7	37.1







#45
2-Fluorobiphenyl
Concen: 89.992 ng
RT: 9.228 min Scan# 1213
Delta R.T. -0.012 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Tgt Ion:172 Resp: 2412754

Ion	Ratio	Lower	Upper
172	100		
171	35.9	29.3	43.9
170	23.8	19.4	29.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.392 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Instrument :
BNA_F
ClientSampleId :
PB167097BL

Tgt Ion:188 Resp: 769835

Ion	Ratio	Lower	Upper
188	100		
94	8.7	6.8	10.2
80	9.8	7.6	11.4







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.027 min Scan# 2029
Delta R.T. -0.012 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Tgt Ion:240 Resp: 601564

Ion	Ratio	Lower	Upper
240	100		
120	10.2	8.4	12.6
236	25.3	20.5	30.7







#79
Terphenyl-d14
Concen: 87.293 ng
RT: 12.986 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Instrument :
BNA_F
ClientSampleId :
PB167097BL

Tgt Ion:244 Resp: 3551821

Ion	Ratio	Lower	Upper
244	100		
212	7.1	6.0	9.0
122	9.0	7.7	11.5






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.498 min Scan# 2279
Delta R.T. -0.012 min
Lab File: BF141940.D
Acq: 13 Mar 2025 12:47

Tgt Ion:264 Resp: 427331

Ion	Ratio	Lower	Upper
264	100		
260	22.9	19.1	28.7
265	21.7	17.5	26.3



Raw Data: **BF141941.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	
Project:	NJ Waste Water PT	Date Received:	
Client Sample ID:	PB167097BS	SDG No.:	Q1502
Lab Sample ID:	PB167097BS	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141941.D	1	03/12/25 08:45	03/13/25 13:17	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
108-95-2	Phenol	45.8		0.91	5.00	ug/L
95-57-8	2-Chlorophenol	46.9		0.58	5.00	ug/L
95-48-7	2-Methylphenol	46.7		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	46.3		1.10	10.0	ug/L
88-75-5	2-Nitrophenol	47.5		1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	57.6		1.90	5.00	ug/L
120-83-2	2,4-Dichlorophenol	46.1		0.52	5.00	ug/L
65-85-0	Benzoic acid	52.7		4.20	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	45.9		0.59	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	49.2		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	45.4		0.62	5.00	ug/L
51-28-5	2,4-Dinitrophenol	99.2	E	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	100	E	2.40	10.0	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	51.0		2.90	10.0	ug/L
87-86-5	Pentachlorophenol	99.9	E	1.60	10.0	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	46.3		0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	140		10 - 139	93%	SPK: 150
13127-88-3	Phenol-d6	137		10 - 134	91%	SPK: 150
118-79-6	2,4,6-Tribromophenol	160		44 - 137	107%	SPK: 150
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	168000	6.875			
1146-65-2	Naphthalene-d8	667000	8.157			
15067-26-2	Acenaphthene-d10	381000	9.916			
1517-22-2	Phenanthrene-d10	657000	11.398			
1719-03-5	Chrysene-d12	424000	14.039			
1520-96-3	Perylene-d12	371000	15.504			

Raw Data: **BF141941.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	
Project:	NJ Waste Water PT	Date Received:	
Client Sample ID:	PB167097BS	SDG No.:	Q1502
Lab Sample ID:	PB167097BS	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141941.D	1	03/12/25 08:45	03/13/25 13:17	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Sample Results: **BF141941.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141941.D
Acq On : 13 Mar 2025 13:17
Operator : RC/JU
Sample : PB167097BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 13 13:42:14 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.875	152	167779	20.000	ng		0.00
21) Naphthalene-d8	8.157	136	666660	20.000	ng		0.00
39) Acenaphthene-d10	9.916	164	381133	20.000	ng		0.00
64) Phenanthrene-d10	11.398	188	657343	20.000	ng		0.00
76) Chrysene-d12	14.039	240	424407	20.000	ng		0.00
86) Perylene-d12	15.504	264	370791	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.510	112	1406671	139.927	ng		0.02
7) Phenol-d6	6.504	99	1754393	137.067	ng		0.00
23) Nitrobenzene-d5	7.439	82	1168111	98.606	ng		0.00
42) 2,4,6-Tribromophenol	10.704	330	773860	160.048	ng		0.00
45) 2-Fluorobiphenyl	9.233	172	2451144	97.806	ng		0.00
79) Terphenyl-d14	12.986	244	3018229	105.142	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.787	88	183786	43.632	ng		99
3) Pyridine	3.534	79	435363	42.136	ng		100
4) n-Nitrosodimethylamine	3.481	42	219416	44.549	ng		97
6) Aniline	6.539	93	455935	36.109	ng		99
8) 2-Chlorophenol	6.663	128	522369	46.852	ng		99
9) Benzaldehyde	6.428	77	161518	22.563	ng		99
10) Phenol	6.516	94	616143	45.757	ng		96
11) bis(2-Chloroethyl)ether	6.616	93	460406	45.535	ng		98
12) 1,3-Dichlorobenzene	6.816	146	549049	45.613	ng		100
13) 1,4-Dichlorobenzene	6.892	146	558151	45.829	ng		99
14) 1,2-Dichlorobenzene	7.045	146	535333	46.667	ng		99
15) Benzyl Alcohol	7.016	79	464922	44.930	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.151	45	521817	42.748	ng		99
17) 2-Methylphenol	7.128	107	413870	46.686	ng		98
18) Hexachloroethane	7.392	117	206331	45.179	ng		94
19) n-Nitroso-di-n-propyla...	7.292	70	356374	43.331	ng		99
20) 3+4-Methylphenols	7.281	107	525851	46.311	ng	#	81
22) Acetophenone	7.286	105	794019	49.735	ng		98
24) Nitrobenzene	7.457	77	540933	45.933	ng		100
25) Isophorone	7.698	82	992286	47.387	ng		100
26) 2-Nitrophenol	7.775	139	274315	47.460	ng		100
27) 2,4-Dimethylphenol	7.804	122	458082	57.557	ng		99
28) bis(2-Chloroethoxy)met...	7.904	93	584252	44.484	ng		100
29) 2,4-Dichlorophenol	8.010	162	454991	46.058	ng		100
30) 1,2,4-Trichlorobenzene	8.098	180	489377	44.952	ng		100
31) Naphthalene	8.181	128	1521001	44.415	ng		100
32) Benzoic acid	7.928	122	368882	52.727	ng		99
33) 4-Chloroaniline	8.222	127	189925	15.710	ng		98
34) Hexachlorobutadiene	8.298	225	330110	46.496	ng		99
35) Caprolactam	8.598	113	149736m	49.717	ng		
36) 4-Chloro-3-methylphenol	8.704	107	506070	45.924	ng		100
37) 2-Methylnaphthalene	8.869	142	984016	42.989	ng		98
38) 1-Methylnaphthalene	8.969	142	975143	44.147	ng		100
40) 1,2,4,5-Tetrachloroben...	9.039	216	589567	50.807	ng		98
41) Hexachlorocyclopentadiene	9.022	237	787647	173.440	ng		100
43) 2,4,6-Trichlorophenol	9.145	196	373446	49.244	ng		100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141941.D
Acq On : 13 Mar 2025 13:17
Operator : RC/JU
Sample : PB167097BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 13 13:42:14 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025

	Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44)	2,4,5-Trichlorophenol	9.186	196	348841	45.418	ng		100
46)	1,1'-Biphenyl	9.333	154	1455375	50.256	ng		100
47)	2-Chloronaphthalene	9.363	162	993605	46.033	ng		99
48)	2-Nitroaniline	9.451	65	295534	47.288	ng		97
49)	Acenaphthylene	9.775	152	1553482	48.500	ng		100
50)	Dimethylphthalate	9.639	163	1224816	45.891	ng		100
51)	2,6-Dinitrotoluene	9.698	165	259612	46.717	ng		96
52)	Acenaphthene	9.951	154	1170255	51.989	ng		98
53)	3-Nitroaniline	9.863	138	143947	25.536	ng		100
54)	2,4-Dinitrophenol	9.975	184	318705	99.219	ng	#	46
55)	Dibenzofuran	10.122	168	1417170	44.061	ng		99
56)	4-Nitrophenol	10.022	139	439494	103.387	ng		98
57)	2,4-Dinitrotoluene	10.104	165	361052	49.745	ng		99
58)	Fluorene	10.463	166	1147740	46.273	ng		99
59)	2,3,4,6-Tetrachlorophenol	10.233	232	323625	46.303	ng		100
60)	Diethylphthalate	10.339	149	1194423	44.881	ng		99
61)	4-Chlorophenyl-phenyle...	10.457	204	597749	46.224	ng		96
62)	4-Nitroaniline	10.480	138	256469	46.734	ng		99
63)	Azobenzene	10.616	77	1083805	43.803	ng		98
65)	4,6-Dinitro-2-methylph...	10.510	198	199851	50.999	ng		97
66)	n-Nitrosodiphenylamine	10.574	169	997450	46.890	ng		100
67)	4-Bromophenyl-phenylether	10.945	248	382756	46.364	ng		97
68)	Hexachlorobenzene	11.010	284	437639	48.335	ng		96
69)	Atrazine	11.098	200	402029	61.650	ng		99
70)	Pentachlorophenol	11.204	266	557370	99.911	ng		99
71)	Phenanthrene	11.427	178	1669056	47.009	ng		100
72)	Anthracene	11.474	178	1712956	48.088	ng		100
73)	Carbazole	11.627	167	1439856	46.870	ng		100
74)	Di-n-butylphthalate	11.957	149	1842392	45.199	ng		100
75)	Fluoranthene	12.610	202	1728401	45.891	ng		100
77)	Benzidine	12.727	184	500180	84.472	ng		100
78)	Pyrene	12.839	202	1719348	46.834	ng		100
80)	Butylbenzylphthalate	13.457	149	660564	45.971	ng		99
81)	Benzo(a)anthracene	14.027	228	1314815	47.211	ng		100
82)	3,3'-Dichlorobenzidine	13.986	252	227079	28.215	ng		100
83)	Chrysene	14.062	228	1167619	46.252	ng		100
84)	Bis(2-ethylhexyl)phtha...	14.015	149	882521	44.545	ng		100
85)	Di-n-octyl phthalate	14.627	149	1244287	45.138	ng		99
87)	Indeno(1,2,3-cd)pyrene	16.992	276	1228415	51.214	ng		98
88)	Benzo(b)fluoranthene	15.074	252	1177030	46.317	ng		100
89)	Benzo(k)fluoranthene	15.104	252	942444	43.336	ng		100
90)	Benzo(a)pyrene	15.445	252	996055	50.093	ng		99
91)	Dibenzo(a,h)anthracene	17.015	278	1005280	50.797	ng		99
92)	Benzo(g,h,i)perylene	17.445	276	938047	47.965	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141941.D
Acq On : 13 Mar 2025 13:17
Operator : RC/JU
Sample : PB167097BS
Misc :
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Mar 13 13:42:14 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:152 Resp: 16777


Ion	Ratio	Lower	Upper
152	100		
150	157.7	127.4	191.2
115	65.4	51.9	77.9



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#2
1,4-Dioxane
Concen: 43.632 ng
RT: 2.787 min Scan# 118
Delta R.T. 0.112 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 88 Resp: 183786

Ion	Ratio	Lower	Upper
88	100		
58	72.1	57.4	86.2
43	26.3	21.5	32.3





#3
Pyridine
Concen: 42.136 ng
RT: 3.534 min Scan# 24
Delta R.T. 0.094 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Tgt Ion: 79 Resp: 43536


Ion	Ratio	Lower	Upper
79	100		
52	66.5	53.3	79.9
51	34.5	27.6	41.4

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#4
n-Nitrosodimethylamine
Concen: 44.549 ng
RT: 3.481 min Scan# 236
Delta R.T. 0.088 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 42 Resp: 219416

Ion	Ratio	Lower	Upper
42	100		
74	133.1	103.4	155.0
44	6.8	5.0	7.4






#5
2-Fluorophenol
Concen: 139.927 ng
RT: 5.510 min Scan# 5
Delta R.T. 0.018 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:112 Resp: 140667


Ion	Ratio	Lower	Upper
112	100		
64	56.9	46.5	69.7
63	31.0	25.4	38.2



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#6
Aniline
Concen: 36.109 ng
RT: 6.539 min Scan# 756
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 93 Resp: 455935

Ion	Ratio	Lower	Upper
93	100		
66	40.4	31.9	47.9
65	19.8	15.8	23.6






#7
Phenol-d6
Concen: 137.067 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 99 Resp: 175439


Ion	Ratio	Lower	Upper
99	100		
42	18.0	14.6	21.8
71	38.9	30.8	46.2



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#8
2-Chlorophenol
Concen: 46.852 ng
RT: 6.663 min Scan# 777
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:128 Resp: 522369

Ion	Ratio	Lower	Upper
128	100		
130	33.5	12.8	52.8
64	43.5	24.4	64.4






#9
Benzaldehyde
Concen: 22.563 ng
RT: 6.428 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 77 Resp: 16151


Ion	Ratio	Lower	Upper
77	100		
105	98.8	80.0	120.0
106	95.0	74.5	114.5



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#10
Phenol
Concen: 45.757 ng
RT: 6.516 min Scan# 752
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 94 Resp: 616143

Ion	Ratio	Lower	Upper
94	100		
65	31.2	10.5	50.5
66	50.0	25.8	65.8





#11
bis(2-Chloroethyl)ether
Concen: 45.535 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Tgt Ion: 93 Resp: 46040


Ion	Ratio	Lower	Upper
93	100		
63	70.2	53.0	93.0
95	32.9	12.9	52.9

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#12
1,3-Dichlorobenzene
Concen: 45.613 ng
RT: 6.816 min Scan# 803
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:146 Resp: 549049

Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.3	76.9
75	33.2	26.6	40.0






#13
1,4-Dichlorobenzene
Concen: 45.829 ng
RT: 6.892 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:146 Resp: 55815


Ion	Ratio	Lower	Upper
146	100		
148	63.1	51.0	76.6
111	44.7	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#14
1,2-Dichlorobenzene
Concen: 46.667 ng
RT: 7.045 min Scan# 842
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:146 Resp: 535333

Ion	Ratio	Lower	Upper
146	100		
148	63.8	50.8	76.2
111	46.7	38.0	57.0






#15
Benzyl Alcohol
Concen: 44.930 ng
RT: 7.016 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 79 Resp: 46492


Ion	Ratio	Lower	Upper
79	100		
108	74.0	59.7	89.5
77	63.0	51.0	76.6



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.748 ng
RT: 7.151 min Scan# 860
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 45 Resp: 521817

Ion	Ratio	Lower	Upper
45	100		
77	18.4	0.0	37.8
79	13.9	0.0	34.2






#17
2-Methylphenol
Concen: 46.686 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:107 Resp: 41387


Ion	Ratio	Lower	Upper
107	100		
108	114.4	90.2	135.2
77	50.5	39.6	59.4
79	50.2	39.1	58.7



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#18
Hexachloroethane
Concen: 45.179 ng
RT: 7.392 min Scan# 901
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:117 Resp: 206331

Ion	Ratio	Lower	Upper
117	100		
119	97.9	76.9	115.3
201	110.1	79.9	119.9






#19
n-Nitroso-di-n-propylamine
Concen: 43.331 ng
RT: 7.292 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 70 Resp: 35637


Ion	Ratio	Lower	Upper
70	100		
42	52.3	41.4	62.0
101	10.3	8.2	12.2
130	23.4	18.2	27.4



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#20
3+4-Methylphenols
Concen: 46.311 ng
RT: 7.281 min Scan# 882
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:107 Resp: 525851

Ion	Ratio	Lower	Upper
107	100		
108	90.1	68.4	108.4
77	129.0	69.7	109.7#
79	33.3	11.4	51.4






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1(
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:136 Resp: 66666(


Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	6.8	5.8	8.8
68	4.9	4.1	6.1



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#22
Acetophenone
Concen: 49.735 ng
RT: 7.286 min Scan# 883
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:105 Resp: 794019

Ion	Ratio	Lower	Upper
105	100		
71	6.9	5.8	8.8
51	30.4	23.4	35.0
120	22.0	17.3	25.9






#23
Nitrobenzene-d5
Concen: 98.606 ng
RT: 7.439 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 82 Resp: 116811


Ion	Ratio	Lower	Upper
82	100		
128	43.9	36.0	54.0
54	48.6	39.6	59.4



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#24
Nitrobenzene
Concen: 45.933 ng
RT: 7.457 min Scan# 912
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 77 Resp: 540933

Ion	Ratio	Lower	Upper
77	100		
123	45.8	37.0	55.4
65	14.1	11.3	16.9






#25
Isophorone
Concen: 47.387 ng
RT: 7.698 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 82 Resp: 99228


Ion	Ratio	Lower	Upper
82	100		
95	8.1	6.4	9.6
138	18.7	14.9	22.3



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#26
2-Nitrophenol
Concen: 47.460 ng
RT: 7.775 min Scan# 966
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:139 Resp: 274315

Ion	Ratio	Lower	Upper
139	100		
109	33.3	26.5	39.7
65	50.7	40.3	60.5






#27
2,4-Dimethylphenol
Concen: 57.557 ng
RT: 7.804 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:122 Resp: 45808


Ion	Ratio	Lower	Upper
122	100		
107	120.2	94.8	142.2
121	59.3	47.8	71.6



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#28
bis(2-Chloroethoxy)methane
Concen: 44.484 ng
RT: 7.904 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 93 Resp: 584252

Ion	Ratio	Lower	Upper
93	100		
95	33.0	26.6	40.0
123	12.7	10.2	15.4






#29
2,4-Dichlorophenol
Concen: 46.058 ng
RT: 8.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:162 Resp: 45499


Ion	Ratio	Lower	Upper
162	100		
164	63.9	43.8	83.8
98	38.6	18.1	58.1



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#30
1,2,4-Trichlorobenzene
Concen: 44.952 ng
RT: 8.098 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:180 Resp: 489377

Ion	Ratio	Lower	Upper
180	100		
182	96.5	77.5	116.3
145	31.9	25.8	38.6






#31
Naphthalene
Concen: 44.415 ng
RT: 8.181 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:128 Resp: 152100


Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.0	13.6
127	13.4	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#32
Benzoic acid
Concen: 52.727 ng
RT: 7.928 min Scan# 992
Delta R.T. -0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:122 Resp: 368882

Ion	Ratio	Lower	Upper
122	100		
105	127.5	107.9	147.9
77	96.3	79.0	119.0





#33
4-Chloroaniline
Concen: 15.710 ng
RT: 8.222 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Instrument :
BNA_F
ClientSampleId :
PB167097BS


Tgt Ion:127 Resp: 18992

Ion	Ratio	Lower	Upper
127	100		
129	32.0	26.2	39.4
65	30.4	25.4	38.2
92	20.6	16.2	24.4

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#34
Hexachlorobutadiene
Concen: 46.496 ng
RT: 8.298 min Scan# 1055
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:225 Resp: 330110

Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.6	75.8
227	64.2	52.2	78.2






#35
Caprolactam
Concen: 49.717 ng m
RT: 8.598 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:113 Resp: 14973


Ion	Ratio	Lower	Upper
113	100		
55	139.9	129.6	169.6
56	115.9	104.0	144.0



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#36
4-Chloro-3-methylphenol
Concen: 45.924 ng
RT: 8.704 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:107 Resp: 506070

Ion	Ratio	Lower	Upper
107	100		
144	24.0	19.0	28.6
142	74.3	59.2	88.8






#37
2-Methylnaphthalene
Concen: 42.989 ng
RT: 8.869 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:142 Resp: 98401


Ion	Ratio	Lower	Upper
142	100		
141	90.8	71.2	106.8



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#38
1-Methylnaphthalene
Concen: 44.147 ng
RT: 8.969 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:142 Resp: 975143

Ion	Ratio	Lower	Upper
142	100		
141	92.4	74.2	111.4





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Tgt Ion:164 Resp: 38113


Ion	Ratio	Lower	Upper
164	100		
162	103.3	81.8	122.6
160	45.6	36.7	55.1

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#40
1,2,4,5-Tetrachlorobenzene
Concen: 50.807 ng
RT: 9.039 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:216 Resp: 589567

Ion	Ratio	Lower	Upper
216	100		
214	78.4	61.9	92.9
179	21.8	18.1	27.1
108	20.4	14.7	22.1





#41
Hexachlorocyclopentadiene
Concen: 173.440 ng
RT: 9.022 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Instrument :
BNA_F
ClientSampleId :
PB167097BS


Tgt Ion:237 Resp: 78764

Ion	Ratio	Lower	Upper
237	100		
235	63.1	43.1	83.1
272	12.8	0.0	32.0

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#42
2,4,6-Tribromophenol
Concen: 160.048 ng
RT: 10.704 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:330 Resp: 773860

Ion	Ratio	Lower	Upper
330	100		
332	95.8	77.6	116.4
141	28.0	24.7	37.1






#43
2,4,6-Trichlorophenol
Concen: 49.244 ng
RT: 9.145 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:196 Resp: 37344


Ion	Ratio	Lower	Upper
196	100		
198	100.4	80.6	121.0
200	32.0	25.8	38.6



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#44
2,4,5-Trichlorophenol
Concen: 45.418 ng
RT: 9.186 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:196 Resp: 348841

Ion	Ratio	Lower	Upper
196	100		
198	101.4	80.7	121.1
97	53.0	42.1	63.1
132	29.2	23.4	35.2






#45
2-Fluorobiphenyl
Concen: 97.806 ng
RT: 9.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:172 Resp: 245114


Ion	Ratio	Lower	Upper
172	100		
171	36.3	29.3	43.9
170	24.0	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#46
1,1'-Biphenyl
Concen: 50.256 ng
RT: 9.333 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:154 Resp: 1455375

Ion	Ratio	Lower	Upper
154	100		
153	41.2	21.1	61.1
76	11.5	0.0	31.3






#47
2-Chloronaphthalene
Concen: 46.033 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:162 Resp: 99360


Ion	Ratio	Lower	Upper
162	100		
127	43.2	35.5	53.3
164	32.8	26.6	40.0



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#48
2-Nitroaniline
Concen: 47.288 ng
RT: 9.451 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion: 65 Resp: 295534

Ion	Ratio	Lower	Upper
65	100		
92	70.5	56.6	84.8
138	105.7	88.6	133.0






#49
Acenaphthylene
Concen: 48.500 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:152 Resp: 155348


Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.3	24.5
153	13.9	11.1	16.7



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#50
Dimethylphthalate
Concen: 45.891 ng
RT: 9.639 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:163 Resp: 1224816

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.1	4.7
164	10.1	8.2	12.4





#51
2,6-Dinitrotoluene
Concen: 46.717 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Tgt Ion:165 Resp: 25961


Ion	Ratio	Lower	Upper
165	100		
63	54.7	47.0	70.4
89	56.6	47.4	71.2

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#52
Acenaphthene
Concen: 51.989 ng
RT: 9.951 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:154 Resp: 1170255

Ion	Ratio	Lower	Upper
154	100		
153	111.4	90.7	136.1
152	50.9	41.1	61.7





#53
3-Nitroaniline
Concen: 25.536 ng
RT: 9.863 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Tgt Ion:138 Resp: 14394


Ion	Ratio	Lower	Upper
138	100		
108	11.4	9.3	13.9
92	117.2	93.5	140.3

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#54
2,4-Dinitrophenol
Concen: 99.219 ng
RT: 9.975 min Scan# 1340
Delta R.T. 0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:184 Resp: 318705

Ion	Ratio	Lower	Upper
184	100		
63	61.8	60.2	90.4
154	68.9	132.9	199.3#






#55
Dibenzofuran
Concen: 44.061 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:168 Resp: 141717


Ion	Ratio	Lower	Upper
168	100		
139	41.5	33.9	50.9
169	13.3	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#56
4-Nitrophenol
Concen: 103.387 ng
RT: 10.022 min Scan# 1348
Delta R.T. 0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:139 Resp: 439494

Ion	Ratio	Lower	Upper
139	100		
109	83.4	61.6	101.6
65	100.7	79.4	119.4






#57
2,4-Dinitrotoluene
Concen: 49.745 ng
RT: 10.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:165 Resp: 36105


Ion	Ratio	Lower	Upper
165	100		
63	45.3	35.4	53.2
89	70.4	56.6	84.8
182	2.8	2.2	3.4



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#58
Fluorene
Concen: 46.273 ng
RT: 10.463 min Scan# 1423
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:166 Resp: 1147740

Ion	Ratio	Lower	Upper
166	100		
165	97.3	79.0	118.4
167	13.5	11.0	16.4






#59
2,3,4,6-Tetrachlorophenol
Concen: 46.303 ng
RT: 10.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:232 Resp: 32362


Ion	Ratio	Lower	Upper
232	100		
131	46.9	37.3	55.9
130	2.4	1.9	2.9
166	31.0	24.8	37.2



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#60
Diethylphthalate
Concen: 44.881 ng
RT: 10.339 min Scan# 1402
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:149 Resp: 1194423

Ion	Ratio	Lower	Upper
149	100		
177	23.1	18.0	27.0
150	12.5	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 46.224 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:204 Resp: 59774


Ion	Ratio	Lower	Upper
204	100		
206	32.8	26.8	40.2
141	61.6	52.4	78.6



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#62
4-Nitroaniline
Concen: 46.734 ng
RT: 10.480 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:138 Resp: 256469

Ion	Ratio	Lower	Upper
138	100		
92	52.4	32.6	72.6
108	86.3	65.0	105.0





#63
Azobenzene
Concen: 43.803 ng
RT: 10.616 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Instrument :
BNA_F
ClientSampleId :
PB167097BS


Tgt Ion: 77 Resp: 108380

Ion	Ratio	Lower	Upper
77	100		
182	28.3	6.6	46.6
105	22.2	1.3	41.3
51	29.9	10.4	50.4

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.398 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:188 Resp: 657343

Ion	Ratio	Lower	Upper
188	100		
94	8.4	6.8	10.2
80	9.6	7.6	11.4





#65
4,6-Dinitro-2-methylphenol
Concen: 50.999 ng
RT: 10.510 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Instrument :
BNA_F
ClientSampleId :
PB167097BS


Tgt Ion:198 Resp: 19985

Ion	Ratio	Lower	Upper
198	100		
51	40.0	22.3	62.3
105	41.8	23.0	63.0

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#66
n-Nitrosodiphenylamine
Concen: 46.890 ng
RT: 10.574 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:169 Resp: 997450

Ion	Ratio	Lower	Upper
169	100		
168	66.7	53.7	80.5
167	36.5	29.2	43.8






#67
4-Bromophenyl-phenylether
Concen: 46.364 ng
RT: 10.945 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:248 Resp: 38275


Ion	Ratio	Lower	Upper
248	100		
250	98.2	77.8	116.8
141	73.6	62.6	93.8



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#68
Hexachlorobenzene
Concen: 48.335 ng
RT: 11.010 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:284 Resp: 437639

Ion	Ratio	Lower	Upper
284	100		
142	30.5	27.0	40.6
249	30.7	25.2	37.8





#69
Atrazine
Concen: 61.650 ng
RT: 11.098 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Instrument :
BNA_F
ClientSampleId :
PB167097BS


Tgt Ion:200 Resp: 40202

Ion	Ratio	Lower	Upper
200	100		
173	25.1	4.7	44.7
215	46.1	25.4	65.4

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#70
Pentachlorophenol
Concen: 99.911 ng
RT: 11.204 min Scan# 1549
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:266 Resp: 557370

Ion	Ratio	Lower	Upper
266	100		
268	64.3	51.0	76.6
264	63.4	50.2	75.2






#71
Phenanthrene
Concen: 47.009 ng
RT: 11.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:178 Resp: 166905


Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.8	23.8
179	15.8	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#72
Anthracene
Concen: 48.088 ng
RT: 11.474 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:178 Resp: 1712956

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.1	22.7
179	16.0	12.6	19.0






#73
Carbazole
Concen: 46.870 ng
RT: 11.627 min Scan# 1(
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:167 Resp: 143985(


Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.4	26.2
139	14.0	11.0	16.6



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#74
Di-n-butylphthalate
Concen: 45.199 ng
RT: 11.957 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:149 Resp: 1842392

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.6	11.4
104	5.4	4.4	6.6






#75
Fluoranthene
Concen: 45.891 ng
RT: 12.610 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:202 Resp: 172840


Ion	Ratio	Lower	Upper
202	100		
101	9.6	0.0	29.7
203	17.8	0.0	37.8



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.039 min Scan# 2031
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:240 Resp: 424407

Ion	Ratio	Lower	Upper
240	100		
120	8.9	8.4	12.6
236	25.7	20.5	30.7






#77
Benzidine
Concen: 84.472 ng
RT: 12.727 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:184 Resp: 50018


Ion	Ratio	Lower	Upper
184	100		
185	14.5	11.5	17.3
183	12.0	9.7	14.5



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#78
Pyrene
Concen: 46.834 ng
RT: 12.839 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:202 Resp: 1719348

Ion	Ratio	Lower	Upper
202	100		
200	21.3	17.3	25.9
203	18.1	14.4	21.6






#79
Terphenyl-d14
Concen: 105.142 ng
RT: 12.986 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:244 Resp: 301822


Ion	Ratio	Lower	Upper
244	100		
212	7.0	6.0	9.0
122	8.9	7.7	11.5



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#80
Butylbenzylphthalate
Concen: 45.971 ng
RT: 13.457 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:149 Resp: 660564

Ion	Ratio	Lower	Upper
149	100		
91	69.8	56.5	84.7
206	21.7	17.3	25.9






#81
Benzo(a)anthracene
Concen: 47.211 ng
RT: 14.027 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:228 Resp: 131481


Ion	Ratio	Lower	Upper
228	100		
226	27.5	21.8	32.8
229	20.0	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#82
3,3'-Dichlorobenzidine
Concen: 28.215 ng
RT: 13.986 min Scan# 2022
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:252 Resp: 227079

Ion	Ratio	Lower	Upper
252	100		
254	64.2	51.0	76.6
126	10.4	8.4	12.6






#83
Chrysene
Concen: 46.252 ng
RT: 14.062 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:228 Resp: 116761


Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.1	36.1
229	19.8	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 44.545 ng
RT: 14.015 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:149 Resp: 882521

Ion	Ratio	Lower	Upper
149	100		
167	26.8	21.4	32.0
279	4.7	3.8	5.6





#85
Di-n-octyl phthalate
Concen: 45.138 ng
RT: 14.627 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17


Tgt Ion:149 Resp: 124428


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.1	6.9	10.3

Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.504 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:264 Resp: 370791

Ion	Ratio	Lower	Upper
264	100		
260	23.2	19.1	28.7
265	21.5	17.5	26.3






#87
Indeno(1,2,3-cd)pyrene
Concen: 51.214 ng
RT: 16.992 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:276 Resp: 122841


Ion	Ratio	Lower	Upper
276	100		
138	21.2	18.4	27.6
277	25.5	20.3	30.5



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#88
Benzo(b)fluoranthene
Concen: 46.317 ng
RT: 15.074 min Scan# 2207
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:252 Resp: 1177030

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.5	26.3
125	8.8	7.0	10.4






#89
Benzo(k)fluoranthene
Concen: 43.336 ng
RT: 15.104 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:252 Resp: 94244


Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	8.8	7.2	10.8



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#90
Benzo(a)pyrene
Concen: 50.093 ng
RT: 15.445 min Scan# 2270
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:252 Resp: 996055

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.4	26.0
125	9.5	7.8	11.8






#91
Dibenzo(a,h)anthracene
Concen: 50.797 ng
RT: 17.015 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:278 Resp: 100528


Ion	Ratio	Lower	Upper
278	100		
139	13.3	11.4	17.2
279	23.6	18.8	28.2



Instrument :
BNA_F
ClientSampleId :
PB167097BS

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025



#92
Benzo(g,h,i)perylene
Concen: 47.965 ng
RT: 17.445 min Scan# 2610
Delta R.T. -0.006 min
Lab File: BF141941.D
Acq: 13 Mar 2025 13:17

Tgt Ion:276 Resp: 938047

Ion	Ratio	Lower	Upper
276	100		
277	23.9	18.7	28.1
138	18.4	15.8	23.8



Raw Data: **BF141942.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	
Project:	NJ Waste Water PT	Date Received:	
Client Sample ID:	PB167097BSD	SDG No.:	Q1502
Lab Sample ID:	PB167097BSD	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141942.D	1	03/12/25 08:45	03/13/25 13:47	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
108-95-2	Phenol	43.0		0.91	5.00	ug/L
95-57-8	2-Chlorophenol	44.2		0.58	5.00	ug/L
95-48-7	2-Methylphenol	44.5		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	44.0		1.10	10.0	ug/L
88-75-5	2-Nitrophenol	45.9		1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	54.8		1.90	5.00	ug/L
120-83-2	2,4-Dichlorophenol	43.8		0.52	5.00	ug/L
65-85-0	Benzoic acid	52.1		4.20	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	43.1		0.59	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	45.1		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	45.7		0.62	5.00	ug/L
51-28-5	2,4-Dinitrophenol	95.0	E	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	97.3	E	2.40	10.0	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	49.9		2.90	10.0	ug/L
87-86-5	Pentachlorophenol	95.6	E	1.60	10.0	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	43.9		0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	132		10 - 139	88%	SPK: 150
13127-88-3	Phenol-d6	129		10 - 134	86%	SPK: 150
118-79-6	2,4,6-Tribromophenol	151		44 - 137	101%	SPK: 150
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	178000	6.875			
1146-65-2	Naphthalene-d8	702000	8.157			
15067-26-2	Acenaphthene-d10	402000	9.916			
1517-22-2	Phenanthrene-d10	683000	11.398			
1719-03-5	Chrysene-d12	437000	14.033			
1520-96-3	Perylene-d12	393000	15.504			

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark			Date Collected:	
Project:	NJ Waste Water PT			Date Received:	
Client Sample ID:	PB167097BSD			SDG No.:	Q1502
Lab Sample ID:	PB167097BSD			Matrix:	Water
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	SVOCMS Group2
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141942.D	1	03/12/25 08:45	03/13/25 13:47	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Sample Results: **BF141942.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141942.D
Acq On : 13 Mar 2025 13:47
Operator : RC/JU
Sample : PB167097BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Mar 13 14:35:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	178366	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	702401	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	401749	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	683407	20.000	ng	0.00
76) Chrysene-d12	14.033	240	437316	20.000	ng	0.00
86) Perylene-d12	15.504	264	392710	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	1413436	132.254	ng	0.02
7) Phenol-d6	6.504	99	1756797	129.108	ng	0.00
23) Nitrobenzene-d5	7.439	82	1163264	93.200	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	771672	151.406	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	2478502	93.823	ng	0.00
79) Terphenyl-d14	12.980	244	2991728	101.143	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.781	88	187189	41.802	ng	99
3) Pyridine	3.534	79	435192	39.620	ng	99
4) n-Nitrosodimethylamine	3.481	42	222631	42.519	ng	100
6) Aniline	6.540	93	484108	36.064	ng	100
8) 2-Chlorophenol	6.663	128	523341	44.153	ng	99
9) Benzaldehyde	6.428	77	161299	21.195	ng	100
10) Phenol	6.516	94	614836	42.950	ng	96
11) bis(2-Chloroethyl)ether	6.610	93	460371	42.829	ng	99
12) 1,3-Dichlorobenzene	6.816	146	547988	42.823	ng	100
13) 1,4-Dichlorobenzene	6.892	146	556911	43.013	ng	100
14) 1,2-Dichlorobenzene	7.045	146	534852	43.858	ng	99
15) Benzyl Alcohol	7.016	79	471295	42.842	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.151	45	523711	40.357	ng	99
17) 2-Methylphenol	7.122	107	419477	44.510	ng	98
18) Hexachloroethane	7.387	117	207547	42.748	ng	98
19) n-Nitroso-di-n-propyla...	7.292	70	358439	40.995	ng	100
20) 3+4-Methylphenols	7.275	107	531297	44.013	ng	94
22) Acetophenone	7.287	105	793124	47.151	ng	99
24) Nitrobenzene	7.457	77	541451	43.637	ng	100
25) Isophorone	7.698	82	997958	45.232	ng	100
26) 2-Nitrophenol	7.775	139	279514	45.898	ng	99
27) 2,4-Dimethylphenol	7.804	122	459597	54.809	ng	97
28) bis(2-Chloroethoxy)met...	7.904	93	590404	42.665	ng	99
29) 2,4-Dichlorophenol	8.010	162	455855	43.797	ng	99
30) 1,2,4-Trichlorobenzene	8.098	180	495238	43.176	ng	98
31) Naphthalene	8.181	128	1520406	42.138	ng	100
32) Benzoic acid	7.934	122	383964	52.090	ng	99
33) 4-Chloroaniline	8.222	127	231327	18.162	ng	99
34) Hexachlorobutadiene	8.298	225	329952	44.109	ng	99
35) Caprolactam	8.598	113	155078m	48.870	ng	
36) 4-Chloro-3-methylphenol	8.704	107	500230	43.084	ng	98
37) 2-Methylnaphthalene	8.869	142	1001009	41.506	ng	100
38) 1-Methylnaphthalene	8.969	142	974037	41.853	ng	100
40) 1,2,4,5-Tetrachloroben...	9.039	216	583297	47.688	ng	97
41) Hexachlorocyclopentadiene	9.022	237	787555	164.521	ng	100
43) 2,4,6-Trichlorophenol	9.145	196	360707	45.123	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141942.D
Acq On : 13 Mar 2025 13:47
Operator : RC/JU
Sample : PB167097BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Mar 13 14:35:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.186	196	370018	45.703	ng		99
46) 1,1'-Biphenyl	9.334	154	1464225	47.967	ng		100
47) 2-Chloronaphthalene	9.363	162	994179	43.696	ng		99
48) 2-Nitroaniline	9.451	65	298045	45.243	ng		98
49) Acenaphthylene	9.775	152	1557169	46.120	ng		100
50) Dimethylphthalate	9.639	163	1194632	42.463	ng		100
51) 2,6-Dinitrotoluene	9.698	165	263529	44.989	ng		94
52) Acenaphthene	9.951	154	1177793	49.639	ng		98
53) 3-Nitroaniline	9.863	138	159486	26.841	ng		98
54) 2,4-Dinitrophenol	9.969	184	320573	94.980	ng	#	52
55) Dibenzofuran	10.122	168	1424322	42.011	ng		99
56) 4-Nitrophenol	10.022	139	435956	97.292	ng		99
57) 2,4-Dinitrotoluene	10.098	165	359870	47.038	ng		97
58) Fluorene	10.463	166	1136451	43.467	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.233	232	323646	43.930	ng		100
60) Diethylphthalate	10.339	149	1198598	42.727	ng		100
61) 4-Chlorophenyl-phenyle...	10.451	204	594729	43.631	ng		99
62) 4-Nitroaniline	10.480	138	255806	44.221	ng		99
63) Azobenzene	10.616	77	1090059	41.795	ng		98
65) 4,6-Dinitro-2-methylph...	10.510	198	203376	49.919	ng		95
66) n-Nitrosodiphenylamine	10.575	169	990749	44.799	ng		99
67) 4-Bromophenyl-phenylether	10.945	248	379973	44.271	ng		97
68) Hexachlorobenzene	11.010	284	442267	46.983	ng		95
69) Atrazine	11.098	200	400827	59.122	ng		99
70) Pentachlorophenol	11.204	266	554214	95.557	ng		99
71) Phenanthrene	11.427	178	1661922	45.023	ng		100
72) Anthracene	11.475	178	1697853	45.846	ng		100
73) Carbazole	11.627	167	1421538	44.509	ng		100
74) Di-n-butylphthalate	11.957	149	1824481	43.052	ng		100
75) Fluoranthene	12.610	202	1694947	43.287	ng		100
77) Benzidine	12.727	184	545926	89.476	ng		100
78) Pyrene	12.839	202	1692747	44.748	ng		99
80) Butylbenzylphthalate	13.457	149	646504	43.665	ng		98
81) Benzo(a)anthracene	14.021	228	1329501	46.329	ng		100
82) 3,3'-Dichlorobenzidine	13.986	252	247002	29.785	ng		99
83) Chrysene	14.063	228	1152190	44.293	ng		99
84) Bis(2-ethylhexyl)phtha...	14.016	149	866554	42.448	ng		99
85) Di-n-octyl phthalate	14.627	149	1235348	43.491	ng		99
87) Indeno(1,2,3-cd)pyrene	16.992	276	1234216	48.584	ng		98
88) Benzo(b)fluoranthene	15.074	252	1060224	39.392	ng		99
89) Benzo(k)fluoranthene	15.104	252	1055343	45.819	ng		99
90) Benzo(a)pyrene	15.445	252	995119	47.252	ng		99
91) Dibenzo(a,h)anthracene	17.009	278	1018402	48.588	ng		99
92) Benzo(g,h,i)perylene	17.445	276	942720	45.514	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141942.D
Acq On : 13 Mar 2025 13:47
Operator : RC/JU
Sample : PB167097BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Mar 13 14:35:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:152 Resp: 17836


Ion	Ratio	Lower	Upper
152	100		
150	157.0	127.4	191.2
115	64.9	51.9	77.9



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#2
1,4-Dioxane
Concen: 41.802 ng
RT: 2.781 min Scan# 117
Delta R.T. 0.106 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 88 Resp: 187189

Ion	Ratio	Lower	Upper
88	100		
58	70.1	57.4	86.2
43	26.8	21.5	32.3






#3
Pyridine
Concen: 39.620 ng
RT: 3.534 min Scan# 24
Delta R.T. 0.094 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 79 Resp: 43519


Ion	Ratio	Lower	Upper
79	100		
52	65.7	53.3	79.9
51	34.1	27.6	41.4



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#4
n-Nitrosodimethylamine
Concen: 42.519 ng
RT: 3.481 min Scan# 236
Delta R.T. 0.088 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 42 Resp: 222631

Ion	Ratio	Lower	Upper
42	100		
74	129.6	103.4	155.0
44	6.5	5.0	7.4






#5
2-Fluorophenol
Concen: 132.254 ng
RT: 5.510 min Scan# 5
Delta R.T. 0.018 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:112 Resp: 141343


Ion	Ratio	Lower	Upper
112	100		
64	56.3	46.5	69.7
63	30.9	25.4	38.2



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#6
Aniline
Concen: 36.064 ng
RT: 6.540 min Scan# 756
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 93 Resp: 484108

Ion	Ratio	Lower	Upper
93	100		
66	40.1	31.9	47.9
65	19.5	15.8	23.6






#7
Phenol-d6
Concen: 129.108 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 99 Resp: 175679


Ion	Ratio	Lower	Upper
99	100		
42	18.0	14.6	21.8
71	39.0	30.8	46.2



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#8
2-Chlorophenol
Concen: 44.153 ng
RT: 6.663 min Scan# 777
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:128 Resp: 523341

Ion	Ratio	Lower	Upper
128	100		
130	32.9	12.8	52.8
64	43.3	24.4	64.4






#9
Benzaldehyde
Concen: 21.195 ng
RT: 6.428 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 77 Resp: 16129


Ion	Ratio	Lower	Upper
77	100		
105	100.8	80.0	120.0
106	94.6	74.5	114.5



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#10
Phenol
Concen: 42.950 ng
RT: 6.516 min Scan# 752
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 94 Resp: 614836

Ion	Ratio	Lower	Upper
94	100		
65	31.7	10.5	50.5
66	49.7	25.8	65.8






#11
bis(2-Chloroethyl)ether
Concen: 42.829 ng
RT: 6.610 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 93 Resp: 46037


Ion	Ratio	Lower	Upper
93	100		
63	71.6	53.0	93.0
95	32.6	12.9	52.9



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#12
1,3-Dichlorobenzene
Concen: 42.823 ng
RT: 6.816 min Scan# 803
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:146 Resp: 547988

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.3	76.9
75	33.8	26.6	40.0






#13
1,4-Dichlorobenzene
Concen: 43.013 ng
RT: 6.892 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:146 Resp: 55691


Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.0	76.6
111	44.6	35.7	53.5



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#14
1,2-Dichlorobenzene
Concen: 43.858 ng
RT: 7.045 min Scan# 842
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:146 Resp: 534852

Ion	Ratio	Lower	Upper
146	100		
148	63.2	50.8	76.2
111	46.0	38.0	57.0





#15
Benzyl Alcohol
Concen: 42.842 ng
RT: 7.016 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Tgt Ion: 79 Resp: 47129


Ion	Ratio	Lower	Upper
79	100		
108	73.4	59.7	89.5
77	62.5	51.0	76.6

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.357 ng
RT: 7.151 min Scan# 860
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 45 Resp: 523711

Ion	Ratio	Lower	Upper
45	100		
77	18.3	0.0	37.8
79	14.0	0.0	34.2






#17
2-Methylphenol
Concen: 44.510 ng
RT: 7.122 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:107 Resp: 41947


Ion	Ratio	Lower	Upper
107	100		
108	110.2	90.2	135.2
77	49.1	39.6	59.4
79	49.5	39.1	58.7



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#18
Hexachloroethane
Concen: 42.748 ng
RT: 7.387 min Scan# 900
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:117 Resp: 207547

Ion	Ratio	Lower	Upper
117	100		
119	97.2	76.9	115.3
201	103.3	79.9	119.9






#19
n-Nitroso-di-n-propylamine
Concen: 40.995 ng
RT: 7.292 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 70 Resp: 35843


Ion	Ratio	Lower	Upper
70	100		
42	51.9	41.4	62.0
101	10.3	8.2	12.2
130	23.0	18.2	27.4



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#20
3+4-Methylphenols
Concen: 44.013 ng
RT: 7.275 min Scan# 881
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:107 Resp: 531297

Ion	Ratio	Lower	Upper
107	100		
108	87.6	68.4	108.4
77	101.5	69.7	109.7
79	31.1	11.4	51.4






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:136 Resp: 70240


Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	6.7	5.8	8.8
68	4.8	4.1	6.1



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#22
Acetophenone
Concen: 47.151 ng
RT: 7.287 min Scan# 883
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:105 Resp: 793124

Ion	Ratio	Lower	Upper
105	100		
71	6.9	5.8	8.8
51	30.3	23.4	35.0
120	22.0	17.3	25.9





#23
Nitrobenzene-d5
Concen: 93.200 ng
RT: 7.439 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Tgt Ion: 82 Resp: 116326


Ion	Ratio	Lower	Upper
82	100		
128	45.2	36.0	54.0
54	49.4	39.6	59.4

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#24
Nitrobenzene
Concen: 43.637 ng
RT: 7.457 min Scan# 912
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 77 Resp: 541451

Ion	Ratio	Lower	Upper
77	100		
123	46.4	37.0	55.4
65	14.0	11.3	16.9






#25
Isophorone
Concen: 45.232 ng
RT: 7.698 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 82 Resp: 99795


Ion	Ratio	Lower	Upper
82	100		
95	8.2	6.4	9.6
138	18.8	14.9	22.3



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#26
2-Nitrophenol
Concen: 45.898 ng
RT: 7.775 min Scan# 966
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:139 Resp: 279514

Ion	Ratio	Lower	Upper
139	100		
109	33.5	26.5	39.7
65	49.6	40.3	60.5






#27
2,4-Dimethylphenol
Concen: 54.809 ng
RT: 7.804 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:122 Resp: 45959


Ion	Ratio	Lower	Upper
122	100		
107	122.1	94.8	142.2
121	60.7	47.8	71.6



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#28
bis(2-Chloroethoxy)methane
Concen: 42.665 ng
RT: 7.904 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 93 Resp: 590404

Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.6	40.0
123	12.8	10.2	15.4






#29
2,4-Dichlorophenol
Concen: 43.797 ng
RT: 8.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:162 Resp: 45585


Ion	Ratio	Lower	Upper
162	100		
164	64.6	43.8	83.8
98	38.2	18.1	58.1



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#30
1,2,4-Trichlorobenzene
Concen: 43.176 ng
RT: 8.098 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:180 Resp: 495238

Ion	Ratio	Lower	Upper
180	100		
182	95.1	77.5	116.3
145	31.6	25.8	38.6





#31
Naphthalene
Concen: 42.138 ng
RT: 8.181 min Scan# 1(
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Tgt Ion:128 Resp: 152040(


Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.0	13.6
127	13.5	10.8	16.2

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#32
Benzoic acid
Concen: 52.090 ng
RT: 7.934 min Scan# 993
Delta R.T. 0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:122 Resp: 383964

Ion	Ratio	Lower	Upper
122	100		
105	128.0	107.9	147.9
77	97.8	79.0	119.0






#33
4-Chloroaniline
Concen: 18.162 ng
RT: 8.222 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:127 Resp: 23132


Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.2	39.4
65	30.5	25.4	38.2
92	20.0	16.2	24.4



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#34
Hexachlorobutadiene
Concen: 44.109 ng
RT: 8.298 min Scan# 1055
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:225 Resp: 329952

Ion	Ratio	Lower	Upper
225	100		
223	61.8	50.6	75.8
227	64.6	52.2	78.2






#35
Caprolactam
Concen: 48.870 ng m
RT: 8.598 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:113 Resp: 15507


Ion	Ratio	Lower	Upper
113	100		
55	141.1	129.6	169.6
56	113.8	104.0	144.0



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#36
4-Chloro-3-methylphenol
Concen: 43.084 ng
RT: 8.704 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:107 Resp: 500230

Ion	Ratio	Lower	Upper
107	100		
144	24.1	19.0	28.6
142	75.8	59.2	88.8






#37
2-Methylnaphthalene
Concen: 41.506 ng
RT: 8.869 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:142 Resp: 100100


Ion	Ratio	Lower	Upper
142	100		
141	88.7	71.2	106.8



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#38
1-Methylnaphthalene
Concen: 41.853 ng
RT: 8.969 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:142 Resp: 974037

Ion	Ratio	Lower	Upper
142	100		
141	92.5	74.2	111.4






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:164 Resp: 401749


Ion	Ratio	Lower	Upper
164	100		
162	101.8	81.8	122.6
160	45.4	36.7	55.1



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 47.688 ng
RT: 9.039 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:216 Resp: 583297

Ion	Ratio	Lower	Upper
216	100		
214	79.7	61.9	92.9
179	22.2	18.1	27.1
108	21.1	14.7	22.1





#41
Hexachlorocyclopentadiene
Concen: 164.521 ng
RT: 9.022 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Instrument :
BNA_F
ClientSampleId :
PB167097BSD


Tgt Ion:237 Resp: 78755

Ion	Ratio	Lower	Upper
237	100		
235	63.1	43.1	83.1
272	12.9	0.0	32.0

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#42
2,4,6-Tribromophenol
Concen: 151.406 ng
RT: 10.704 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:330 Resp: 771672

Ion	Ratio	Lower	Upper
330	100		
332	97.0	77.6	116.4
141	28.4	24.7	37.1






#43
2,4,6-Trichlorophenol
Concen: 45.123 ng
RT: 9.145 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:196 Resp: 36070


Ion	Ratio	Lower	Upper
196	100		
198	99.8	80.6	121.0
200	32.4	25.8	38.6



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#44
2,4,5-Trichlorophenol
Concen: 45.703 ng
RT: 9.186 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:196 Resp: 370018

Ion	Ratio	Lower	Upper
196	100		
198	100.3	80.7	121.1
97	52.7	42.1	63.1
132	29.8	23.4	35.2






#45
2-Fluorobiphenyl
Concen: 93.823 ng
RT: 9.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:172 Resp: 247850


Ion	Ratio	Lower	Upper
172	100		
171	36.5	29.3	43.9
170	24.2	19.4	29.0



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#46
1,1'-Biphenyl
Concen: 47.967 ng
RT: 9.334 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:154 Resp: 1464225

Ion	Ratio	Lower	Upper
154	100		
153	41.2	21.1	61.1
76	11.3	0.0	31.3






#47
2-Chloronaphthalene
Concen: 43.696 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:162 Resp: 99417


Ion	Ratio	Lower	Upper
162	100		
127	43.6	35.5	53.3
164	33.4	26.6	40.0



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#48
2-Nitroaniline
Concen: 45.243 ng
RT: 9.451 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 65 Resp: 298045

Ion	Ratio	Lower	Upper
65	100		
92	70.7	56.6	84.8
138	108.1	88.6	133.0






#49
Acenaphthylene
Concen: 46.120 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:152 Resp: 155716


Ion	Ratio	Lower	Upper
152	100		
151	20.5	16.3	24.5
153	14.0	11.1	16.7



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#50
Dimethylphthalate
Concen: 42.463 ng
RT: 9.639 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:163 Resp: 1194632

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.1	4.7
164	10.3	8.2	12.4






#51
2,6-Dinitrotoluene
Concen: 44.989 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:165 Resp: 26352


Ion	Ratio	Lower	Upper
165	100		
63	53.8	47.0	70.4
89	55.8	47.4	71.2



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#52
Acenaphthene
Concen: 49.639 ng
RT: 9.951 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:154 Resp: 1177793

Ion	Ratio	Lower	Upper
154	100		
153	110.4	90.7	136.1
152	50.6	41.1	61.7





#53
3-Nitroaniline
Concen: 26.841 ng
RT: 9.863 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Tgt Ion:138 Resp: 15948


Ion	Ratio	Lower	Upper
138	100		
108	11.2	9.3	13.9
92	114.6	93.5	140.3

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#54
2,4-Dinitrophenol
Concen: 94.980 ng
RT: 9.969 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:184 Resp: 320573

Ion	Ratio	Lower	Upper
184	100		
63	67.3	60.2	90.4
154	77.9	132.9	199.3#






#55
Dibenzofuran
Concen: 42.011 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:168 Resp: 142432


Ion	Ratio	Lower	Upper
168	100		
139	41.8	33.9	50.9
169	13.6	11.0	16.4



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#56
4-Nitrophenol
Concen: 97.292 ng
RT: 10.022 min Scan# 1348
Delta R.T. 0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:139 Resp: 435956

Ion	Ratio	Lower	Upper
139	100		
109	83.9	61.6	101.6
65	100.0	79.4	119.4






#57
2,4-Dinitrotoluene
Concen: 47.038 ng
RT: 10.098 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:165 Resp: 35987


Ion	Ratio	Lower	Upper
165	100		
63	45.6	35.4	53.2
89	73.6	56.6	84.8
182	2.7	2.2	3.4



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#58
Fluorene
Concen: 43.467 ng
RT: 10.463 min Scan# 1423
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:166 Resp: 1136451

Ion	Ratio	Lower	Upper
166	100		
165	99.2	79.0	118.4
167	13.4	11.0	16.4






#59
2,3,4,6-Tetrachlorophenol
Concen: 43.930 ng
RT: 10.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:232 Resp: 32364


Ion	Ratio	Lower	Upper
232	100		
131	46.4	37.3	55.9
130	2.4	1.9	2.9
166	31.2	24.8	37.2



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#60
Diethylphthalate
Concen: 42.727 ng
RT: 10.339 min Scan# 1402
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:149 Resp: 1198598

Ion	Ratio	Lower	Upper
149	100		
177	22.5	18.0	27.0
150	12.6	10.2	15.4





#61
4-Chlorophenyl-phenylether
Concen: 43.631 ng
RT: 10.451 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Instrument :
BNA_F
ClientSampleId :
PB167097BSD


Tgt Ion:204 Resp: 59472

Ion	Ratio	Lower	Upper
204	100		
206	32.8	26.8	40.2
141	64.4	52.4	78.6

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#62
4-Nitroaniline
Concen: 44.221 ng
RT: 10.480 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:138 Resp: 255806

Ion	Ratio	Lower	Upper
138	100		
92	52.1	32.6	72.6
108	86.8	65.0	105.0






#63
Azobenzene
Concen: 41.795 ng
RT: 10.616 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion: 77 Resp: 109005


Ion	Ratio	Lower	Upper
77	100		
182	28.0	6.6	46.6
105	21.6	1.3	41.3
51	29.4	10.4	50.4



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.398 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:188 Resp: 683407

Ion	Ratio	Lower	Upper
188	100		
94	8.8	6.8	10.2
80	9.2	7.6	11.4





#65
4,6-Dinitro-2-methylphenol
Concen: 49.919 ng
RT: 10.510 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Tgt Ion:198 Resp: 20337


Ion	Ratio	Lower	Upper
198	100		
51	39.1	22.3	62.3
105	40.4	23.0	63.0

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#66
n-Nitrosodiphenylamine
Concen: 44.799 ng
RT: 10.575 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:169 Resp: 990749

Ion	Ratio	Lower	Upper
169	100		
168	66.5	53.7	80.5
167	36.2	29.2	43.8






#67
4-Bromophenyl-phenylether
Concen: 44.271 ng
RT: 10.945 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:248 Resp: 37997


Ion	Ratio	Lower	Upper
248	100		
250	99.4	77.8	116.8
141	74.6	62.6	93.8



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#68
Hexachlorobenzene
Concen: 46.983 ng
RT: 11.010 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:284 Resp: 442267

Ion	Ratio	Lower	Upper
284	100		
142	29.5	27.0	40.6
249	30.7	25.2	37.8





#69
Atrazine
Concen: 59.122 ng
RT: 11.098 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Instrument :
BNA_F
ClientSampleId :
PB167097BSD


Tgt Ion:200 Resp: 40082

Ion	Ratio	Lower	Upper
200	100		
173	25.0	4.7	44.7
215	46.1	25.4	65.4

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#70
Pentachlorophenol
Concen: 95.557 ng
RT: 11.204 min Scan# 1549
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:266 Resp: 554214

Ion	Ratio	Lower	Upper
266	100		
268	64.0	51.0	76.6
264	63.6	50.2	75.2






#71
Phenanthrene
Concen: 45.023 ng
RT: 11.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:178 Resp: 166192


Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.8	23.8
179	15.7	12.6	19.0



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#72
Anthracene
Concen: 45.846 ng
RT: 11.475 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:178 Resp: 1697853

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.1	22.7
179	15.8	12.6	19.0






#73
Carbazole
Concen: 44.509 ng
RT: 11.627 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:167 Resp: 142153


Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.4	26.2
139	13.8	11.0	16.6



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#74
Di-n-butylphthalate
Concen: 43.052 ng
RT: 11.957 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:149 Resp: 1824481

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.6	11.4
104	5.3	4.4	6.6






#75
Fluoranthene
Concen: 43.287 ng
RT: 12.610 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:202 Resp: 169494


Ion	Ratio	Lower	Upper
202	100		
101	9.5	0.0	29.7
203	17.8	0.0	37.8



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.033 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:240 Resp: 437316

Ion	Ratio	Lower	Upper
240	100		
120	9.8	8.4	12.6
236	26.2	20.5	30.7






#77
Benzidine
Concen: 89.476 ng
RT: 12.727 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:184 Resp: 54592


Ion	Ratio	Lower	Upper
184	100		
185	14.4	11.5	17.3
183	11.8	9.7	14.5



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#78
Pyrene
Concen: 44.748 ng
RT: 12.839 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:202 Resp: 1692747

Ion	Ratio	Lower	Upper
202	100		
200	21.3	17.3	25.9
203	17.8	14.4	21.6






#79
Terphenyl-d14
Concen: 101.143 ng
RT: 12.980 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:244 Resp: 299172


Ion	Ratio	Lower	Upper
244	100		
212	7.2	6.0	9.0
122	9.7	7.7	11.5



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#80
Butylbenzylphthalate
Concen: 43.665 ng
RT: 13.457 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:149 Resp: 646504

Ion	Ratio	Lower	Upper
149	100		
91	68.9	56.5	84.7
206	21.8	17.3	25.9






#81
Benzo(a)anthracene
Concen: 46.329 ng
RT: 14.021 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:228 Resp: 132950


Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.8	32.8
229	20.1	16.0	24.0



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#82
3,3'-Dichlorobenzidine
Concen: 29.785 ng
RT: 13.986 min Scan# 2022
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:252 Resp: 247002

Ion	Ratio	Lower	Upper
252	100		
254	64.1	51.0	76.6
126	9.8	8.4	12.6






#83
Chrysene
Concen: 44.293 ng
RT: 14.063 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:228 Resp: 115219


Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.1	36.1
229	19.4	15.7	23.5



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 42.448 ng
RT: 14.016 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:149 Resp: 866554

Ion	Ratio	Lower	Upper
149	100		
167	27.0	21.4	32.0
279	5.0	3.8	5.6





#85
Di-n-octyl phthalate
Concen: 43.491 ng
RT: 14.627 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Tgt Ion:149 Resp: 123534


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	8.2	6.9	10.3

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.504 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:264 Resp: 392710

Ion	Ratio	Lower	Upper
264	100		
260	23.1	19.1	28.7
265	21.5	17.5	26.3






#87
Indeno(1,2,3-cd)pyrene
Concen: 48.584 ng
RT: 16.992 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:276 Resp: 123421


Ion	Ratio	Lower	Upper
276	100		
138	21.5	18.4	27.6
277	25.5	20.3	30.5



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#88
Benzo(b)fluoranthene
Concen: 39.392 ng
RT: 15.074 min Scan# 2207
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:252 Resp: 1060224

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.5	26.3
125	8.3	7.0	10.4






#89
Benzo(k)fluoranthene
Concen: 45.819 ng
RT: 15.104 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:252 Resp: 105534


Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	8.5	7.2	10.8



Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025




#90
Benzo(a)pyrene
Concen: 47.252 ng
RT: 15.445 min Scan# 2270
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:252 Resp: 995119

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.0
125	9.1	7.8	11.8





#91
Dibenzo(a,h)anthracene
Concen: 48.588 ng
RT: 17.009 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47


Tgt Ion:278 Resp: 101840


Ion	Ratio	Lower	Upper
278	100		
139	13.6	11.4	17.2
279	23.8	18.8	28.2

Instrument :
BNA_F
ClientSampleId :
PB167097BSD

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/14/2025
Supervised By :Jagrut Upadhyay 03/14/2025






#92
Benzo(g,h,i)perylene
Concen: 45.514 ng
RT: 17.445 min Scan# 2610
Delta R.T. -0.006 min
Lab File: BF141942.D
Acq: 13 Mar 2025 13:47

Tgt Ion:276 Resp: 942720

Ion	Ratio	Lower	Upper
276	100		
277	23.5	18.7	28.1
138	18.2	15.8	23.8



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF031025	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC080	BF141904.D	Caprolactam	anahy	3/11/2025 9:10:54 AM	Jagrut	3/11/2025 10:18:21 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF031325	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB167097BS	BF141941.D	Caprolactam	anahy	3/14/2025 11:28:10 AM	Jagrut	3/14/2025 4:35:23 PM	Peak Integrated by Software
PB167097BSD	BF141942.D	Caprolactam	anahy	3/14/2025 11:28:42 AM	Jagrut	3/14/2025 4:35:25 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BP031225	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BP024161.D	Benzaldehyde	anahy	3/13/2025 11:10:33 AM	Jagrut	3/13/2025 11:45:04 AM	Peak Integrated by Software
SSTDICC050	BP024165.D	Pyridine	anahy	3/13/2025 11:11:20 AM	Jagrut	3/13/2025 11:45:07 AM	Peak Integrated by Software
SSTDICC060	BP024166.D	Indeno(1,2,3-cd)pyrene	anahy	3/13/2025 11:38:58 AM	Jagrut	3/13/2025 11:45:09 AM	Peak Integrated by Software
SSTDICC060	BP024166.D	Pyridine	anahy	3/13/2025 11:38:58 AM	Jagrut	3/13/2025 11:45:09 AM	Peak Integrated by Software
SSTDICC080	BP024167.D	Pyridine	anahy	3/13/2025 11:39:57 AM	Jagrut	3/13/2025 11:45:11 AM	Peak Integrated by Software
SSTDICV040	BP024168.D	Benzo(g,h,i)perylene	anahy	3/13/2025 11:40:49 AM	Jagrut	3/13/2025 11:45:14 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BP031325	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF031025

Review By	anahy	Review On	3/11/2025 9:11:47 AM				
Supervise By	Jagrut	Supervise On	3/11/2025 10:18:40 AM				
SubDirectory	BF031025	HP Acquire Method	BNA_F		HP Processing Method	bf031025	

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF141896.D	10 Mar 2025 10:31	RC/JU	Ok
2	SSTDICC2.5	BF141897.D	10 Mar 2025 11:01	RC/JU	Ok
3	SSTDICC005	BF141898.D	10 Mar 2025 11:30	RC/JU	Ok
4	SSTDICC010	BF141899.D	10 Mar 2025 12:00	RC/JU	Ok
5	SSTDICC020	BF141900.D	10 Mar 2025 12:29	RC/JU	Ok
6	SSTDICCC040	BF141901.D	10 Mar 2025 12:58	RC/JU	Ok
7	SSTDICC050	BF141902.D	10 Mar 2025 13:28	RC/JU	Not Ok
8	SSTDICC060	BF141903.D	10 Mar 2025 13:57	RC/JU	Ok
9	SSTDICC080	BF141904.D	10 Mar 2025 14:27	RC/JU	Ok,M
10	SSTDICC050	BF141905.D	10 Mar 2025 15:20	RC/JU	Ok
11	SSTDICV040	BF141906.D	10 Mar 2025 15:53	RC/JU	Ok
12	PB167012BL	BF141907.D	10 Mar 2025 17:09	RC/JU	Ok
13	SP6752	BF141908.D	10 Mar 2025 17:39	RC/JU	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF031325

Review By	anahy	Review On	3/14/2025 11:36:22 AM				
Supervise By	Jagrut	Supervise On	3/14/2025 4:35:55 PM				
SubDirectory	BF031325	HP Acquire Method	BNA_F		HP Processing Method	bf031025	

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF141938.D	13 Mar 2025 11:48	RC/JU	Ok
2	SSTDCCC040	BF141939.D	13 Mar 2025 12:17	RC/JU	Ok
3	PB167097BL	BF141940.D	13 Mar 2025 12:47	RC/JU	Ok
4	PB167097BS	BF141941.D	13 Mar 2025 13:17	RC/JU	Ok,M
5	PB167097BSD	BF141942.D	13 Mar 2025 13:47	RC/JU	Ok,M
6	Q1539-01	BF141943.D	13 Mar 2025 16:14	RC/JU	Ok
7	Q1539-02	BF141944.D	13 Mar 2025 16:44	RC/JU	Ok
8	Q1547-01	BF141945.D	13 Mar 2025 17:14	RC/JU	Ok
9	Q1547-01MS	BF141946.D	13 Mar 2025 17:44	RC/JU	Ok,M
10	Q1547-01MSD	BF141947.D	13 Mar 2025 18:13	RC/JU	Ok,M
11	Q1547-05	BF141948.D	13 Mar 2025 18:42	RC/JU	Ok
12	Q1547-05MS	BF141949.D	13 Mar 2025 19:11	RC/JU	Ok,M
13	Q1547-05MSD	BF141950.D	13 Mar 2025 19:41	RC/JU	Ok,M
14	Q1547-10	BF141951.D	13 Mar 2025 20:10	RC/JU	Ok,M
15	Q1547-06	BF141952.D	13 Mar 2025 20:39	RC/JU	Ok
16	SSTDCCC040	BF141953.D	13 Mar 2025 21:09	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP031225

Review By	anahy	Review On	3/13/2025 11:41:15 AM				
Supervise By	Jagrut	Supervise On	3/17/2025 4:44:19 PM				
SubDirectory	BP031225	HP Acquire Method	BNA_P		HP Processing Method	bp031225	

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024159.D	12 Mar 2025 15:36	RC/JU	Ok
2	SSTDICC2.5	BP024160.D	12 Mar 2025 16:17	RC/JU	Ok
3	SSTDICC005	BP024161.D	12 Mar 2025 16:57	RC/JU	Ok,M
4	SSTDICC010	BP024162.D	12 Mar 2025 17:38	RC/JU	Ok
5	SSTDICC020	BP024163.D	12 Mar 2025 18:19	RC/JU	Ok
6	SSTDICCC040	BP024164.D	12 Mar 2025 19:00	RC/JU	Ok
7	SSTDICC050	BP024165.D	12 Mar 2025 19:41	RC/JU	Ok,M
8	SSTDICC060	BP024166.D	12 Mar 2025 20:21	RC/JU	Ok,M
9	SSTDICC080	BP024167.D	12 Mar 2025 21:02	RC/JU	Ok,M
10	SSTDICV040	BP024168.D	12 Mar 2025 21:43	RC/JU	Ok,M
11	PB167078BL	BP024169.D	12 Mar 2025 22:24	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP031325

Review By	anahy	Review On	3/17/2025 4:45:24 PM		
Supervise By	Jagrut	Supervise On	3/17/2025 4:49:31 PM		
SubDirectory	BP031325	HP Acquire Method	BNA_P	HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024170.D	13 Mar 2025 09:41	RC/JU	Ok
2	SSTDCCC040	BP024171.D	13 Mar 2025 10:22	RC/JU	Ok
3	PB167097BL	BP024172.D	13 Mar 2025 11:02	RC/JU	Not Ok
4	Q1502-03	BP024173.D	13 Mar 2025 11:43	RC/JU	Dilution
5	Q1502-03DL	BP024174.D	13 Mar 2025 12:37	RC/JU	Ok
6	Q1502-06	BP024175.D	13 Mar 2025 13:51	RC/JU	Dilution
7	Q1502-06DL	BP024176.D	13 Mar 2025 15:48	RC/JU	Ok
8	Q1502-22	BP024177.D	13 Mar 2025 16:40	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF031025

Review By	anahy	Review On	3/11/2025 9:11:47 AM		
Supervise By	Jagrut	Supervise On	3/11/2025 10:18:40 AM		
SubDirectory	BF031025	HP Acquire Method	BNA_F	HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141896.D	10 Mar 2025 10:31		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141897.D	10 Mar 2025 11:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141898.D	10 Mar 2025 11:30	Compound#9,32,54,65 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141899.D	10 Mar 2025 12:00		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF141900.D	10 Mar 2025 12:29	Compound #54 Kept on LR, Method is good for DOD . Method failed for compound #77.	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF141901.D	10 Mar 2025 12:58		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141902.D	10 Mar 2025 13:28	Not used	RC/JU	Not Ok
8	SSTDICC060	SSTDICC060	BF141903.D	10 Mar 2025 13:57	Compound#69 Removed from 60 ppm	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF141904.D	10 Mar 2025 14:27	Compound#9,69 Removed from 80 ppm	RC/JU	Ok,M
10	SSTDICC050	SSTDICC050	BF141905.D	10 Mar 2025 15:20		RC/JU	Ok
11	SSTDICV040	ICVBF031025	BF141906.D	10 Mar 2025 15:53		RC/JU	Ok
12	PB167012BL	PB167012BL	BF141907.D	10 Mar 2025 17:09		RC/JU	Ok
13	SP6752	SP6752	BF141908.D	10 Mar 2025 17:39	8270 SPIKE-SP6752	RC/JU	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF031325

Review By	anahy	Review On	3/14/2025 11:36:22 AM		
Supervise By	Jagrut	Supervise On	3/14/2025 4:35:55 PM		
SubDirectory	BF031325	HP Acquire Method	BNA_F	HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141938.D	13 Mar 2025 11:48		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF141939.D	13 Mar 2025 12:17		RC/JU	Ok
3	PB167097BL	PB167097BL	BF141940.D	13 Mar 2025 12:47		RC/JU	Ok
4	PB167097BS	PB167097BS	BF141941.D	13 Mar 2025 13:17		RC/JU	Ok,M
5	PB167097BSD	PB167097BSD	BF141942.D	13 Mar 2025 13:47		RC/JU	Ok,M
6	Q1539-01	TAPIAL3-MW03D-0310	BF141943.D	13 Mar 2025 16:14		RC/JU	Ok
7	Q1539-02	TAPFTA-MW01I-03102	BF141944.D	13 Mar 2025 16:44		RC/JU	Ok
8	Q1547-01	OR-620-JB-COMP-01	BF141945.D	13 Mar 2025 17:14		RC/JU	Ok
9	Q1547-01MS	OR-620-JB-COMP-01M	BF141946.D	13 Mar 2025 17:44		RC/JU	Ok,M
10	Q1547-01MSD	OR-620-JB-COMP-01M	BF141947.D	13 Mar 2025 18:13		RC/JU	Ok,M
11	Q1547-05	OR-620-JB-COMP-01	BF141948.D	13 Mar 2025 18:42		RC/JU	Ok
12	Q1547-05MS	OR-620-JB-COMP-01M	BF141949.D	13 Mar 2025 19:11		RC/JU	Ok,M
13	Q1547-05MSD	OR-620-JB-COMP-01M	BF141950.D	13 Mar 2025 19:41		RC/JU	Ok,M
14	Q1547-10	OR-620-JB-COMP-02	BF141951.D	13 Mar 2025 20:10		RC/JU	Ok,M
15	Q1547-06	OR-620-JB-COMP-02	BF141952.D	13 Mar 2025 20:39		RC/JU	Ok
16	SSTDCCC040	SSTDCCC040EC	BF141953.D	13 Mar 2025 21:09		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP031225

Review By	anahy	Review On	3/13/2025 11:41:15 AM		
Supervise By	Jagrut	Supervise On	3/17/2025 4:44:19 PM		
SubDirectory	BP031225	HP Acquire Method	BNA_P	HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024159.D	12 Mar 2025 15:36		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024160.D	12 Mar 2025 16:17		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024161.D	12 Mar 2025 16:57	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024162.D	12 Mar 2025 17:38		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BP024163.D	12 Mar 2025 18:19		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BP024164.D	12 Mar 2025 19:00	This calibration is good for 8270E and 8270E DOD methods	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024165.D	12 Mar 2025 19:41		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024166.D	12 Mar 2025 20:21	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024167.D	12 Mar 2025 21:02	Compound #32,54,69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP031225	BP024168.D	12 Mar 2025 21:43		RC/JU	Ok,M
11	PB167078BL	PB167078BL	BP024169.D	12 Mar 2025 22:24		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP031325

Review By	anahy	Review On	3/17/2025 4:45:24 PM		
Supervise By	Jagrut	Supervise On	3/17/2025 4:49:31 PM		
SubDirectory	BP031325	HP Acquire Method	BNA_P	HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024170.D	13 Mar 2025 09:41		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024171.D	13 Mar 2025 10:22		RC/JU	Ok
3	PB167097BL	PB167097BL	BP024172.D	13 Mar 2025 11:02	Not Used	RC/JU	Not Ok
4	Q1502-03	PT-BN-WP	BP024173.D	13 Mar 2025 11:43	PT Sample, Need 5X Dilution	RC/JU	Dilution
5	Q1502-03DL	PT-BN-WPDL	BP024174.D	13 Mar 2025 12:37		RC/JU	Ok
6	Q1502-06	PT-ACIDS-WP	BP024175.D	13 Mar 2025 13:51	PT Sample, Need 5X Dilution	RC/JU	Dilution
7	Q1502-06DL	PT-ACIDS-WPDL	BP024176.D	13 Mar 2025 15:48		RC/JU	Ok
8	Q1502-22	RR-TRIAZINE-WP	BP024177.D	13 Mar 2025 16:40	PT Sample	RC/JU	Ok

M : Manual Integration

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB167097

SOP ID: M3510C,3580A-Extraction SVOC-20

Clean Up SOP #: N/A

Matrix : Water

Weigh By: N/A

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: E3880

Extraction By: RS

Filter By: RS

pH Meter ID: N/A

Hood ID: 4,6,7

Extraction Start Date : 03/12/2025

Extraction Start Time : 08:45

Extraction End Date : 03/12/2025

Extraction End Time : 13:55

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: [√] Seperatory Funnel [ ] Continious Liquid/Liquid [ ] Sonication [ ] Waste Dilution [ ] Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6720
Surrogate	1.0ML	100/150 PPM	SP6638
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3878
Baked Na2SO4	N/A	EP2593
10N NaoH	N/A	EP2559
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5ML Vial lot# 2210673. pH Adjusted<2 with 1:1 H2SO4 & >11 with 10 N NaOH.

KD Bath ID: WATER BATH-1,2

KD Bath Temperature: 60 °C

Envap ID: NEVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
3/12/25	RS (Ext-Lab)	RclSvoc
14:00	**Preparation Group**	**Analysis Group**

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB167097

Analytical Method: M3510C,3580A-Extraction SVOC-20 **Concentration Date:** 03/12/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167097BL	SBLK097	SVOC-TCL BNA -20	1000	6	RUPESH	ritesh	1			SEP-1
PB167097BS	SLCS097	SVOC-TCL BNA -20	1000	6	RUPESH	ritesh	1			2
PB167097BSD	SLCSD097	SVOC-TCL BNA -20	1000	6	RUPESH	ritesh	1			3
Q1502-03	PT-BN-WP	SVOCMS Group1	1000	6	RUPESH	ritesh	1			4
Q1502-06	PT-ACIDS-WP	SVOCMS Group2	1000	6	RUPESH	ritesh	1			5
Q1502-22	RR-TRIAZINE-WP	SVOCMS Group6	1000	6	RUPESH	ritesh	1			6
Q1522-02	TW-WTS-04	SVOCMS Group4	990	6	RUPESH	ritesh	1	D		7
Q1539-01	TAPIAL3-MW03D-031025-00-T1	SVOC-TCL BNA -20	1000	6	RUPESH	ritesh	1	O		8
Q1539-02	TAPFTA-MW01I-031025-00-T2	SVOC-TCL BNA -20	930	6	RUPESH	ritesh	1	O		9

RS
3/12

* **Extracts relinguished on the same date as received.**

167097
8:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1522 **WorkList ID :** 188221 **Department :** Extraction **Date :** 03-12-2025 08:39:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1502-03	PT-BN-WP	Water	SVOCMS Group1	Cool 4 deg C	ALLI03	QA Of	03/03/2025	8270E
Q1502-06	PT-ACIDS-WP	Water	SVOCMS Group2	Cool 4 deg C	ALLI03	QA Of	03/03/2025	8270E
Q1502-22	RR-TRIAZINE-WP	Water	SVOCMS Group6	Cool 4 deg C	ALLI03	QA Of	03/07/2025	8270E
Q1522-02	TW-WTS-04	Water	SVOCMS Group4	Cool 4 deg C	ENTA05	H31	03/06/2025	8270E
Q1539-01	TAPIAL3-MW03D-031025-00-T	Water	SVOC-TCL BNA -20	Cool 4 deg C	WEST04	I31	03/10/2025	8270E
Q1539-02	TAPFTA-MW01I-031025-00-T2	Water	SVOC-TCL BNA -20	Cool 4 deg C	WEST04	I31	03/10/2025	8270E

Date/Time 3/12/25 8:40
Raw Sample Received by: RS (Ext Lab)
Raw Sample Relinquished by: CP SM

Date/Time 3/12/25 9:15
Raw Sample Received by: CP SM
Raw Sample Relinquished by: RS (Ext Lab)

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q1502

Test : SVOCMS Group2

Prepbatch ID : PB167097,

Sequence ID/Qc Batch ID: BF031325,BP031325,

Standard ID :
EP2559,EP2565,EP2593,SP6638,SP6685,SP6686,SP6717,SP6720,SP6721,SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729,

Chemical ID :
10ul/1000ul sample,E3551,E3657,E3815,E3828,E3846,E3871,E3874,E3878,M5173,S10104,S10246,S10397,S10584,S10978,S10979,S10980,S11004,S11005,S11006,S11007,S11008,S11009,S11010,S11074,S11087,S11143,S11161,S11487,S11495,S11650,S11781,S11782,S11783,S11784,S11785,S12114,S12142,S12143,S12144,S12145,S12146,S12187,S12188,S12189,S12207,S12208,S12270,S12276,S12327,S12469,S12470,S12471,S12472,S12473,S12474,S12475,S12476,S12477,S12478,S12517,S12518,S12519,S12520,S12521,S12522,S12523,S12524,S12525,S12649,S12655,S12791,S12963,S12964,S12965,S12966,W3112,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1874	10 N SODIUM HYDROXIDE SOLN	EP2559	11/14/2024	05/14/2025	Rajesh Parikh	Extraction_SCALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 11/14/2024

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
314	1.1 H2SO4 SOLN	EP2565	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 1000.00000ml of M5173 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2593	03/07/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 03/07/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
19	8270/CLP Surrogate Solution, 100 PPM BN/150 PPM ACID	SP6638	10/10/2024	04/04/2025	Jagrut Upadhyay	None	None	mohammad ahmed 10/18/2024

FROM 1930.00000ml of E3815 + 5.00000ml of S10978 + 5.00000ml of S10979 + 5.00000ml of S10980 + 5.00000ml of S11004 + 5.00000ml of S11005 + 5.00000ml of S11006 + 5.00000ml of S11007 + 5.00000ml of S11008 + 5.00000ml of S11009 + 5.00000ml of S11010 + 5.00000ml of S12187 + 5.00000ml of S12188 + 5.00000ml of S12189 + 5.00000ml of S12207 = Final Quantity: 2000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
18	Second Source Calibration Stock Standard, 100 PPM, (8270/625/CLP)	SP6685	11/15/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel 12/27/2024

FROM 0.04000ml of S12189 + 0.08000ml of S12208 + 0.10000ml of S11074 + 0.20000ml of S12142 + 0.20000ml of S12469 + 0.20000ml of S12517 + 1.18000ml of E3828 = Final Quantity: 2.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
416	40 ng BNA ICV, 40 PPM	SP6686	11/15/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel 12/27/2024

FROM 0.01000ml of S12327 + 0.60000ml of E3828 + 0.40000ml of SP6685 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3895	50 ug/ml DFTPP 8270E	SP6717	01/15/2025	03/31/2025	Rahul Chavli	None	None	Yogesh Patel 01/16/2025

FROM 1.00000ml of S10246 + 19.00000ml of E3871 = Final Quantity: 20.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
171	8270/625 Spike Solution, 50/100 PPM	SP6720	01/29/2025	04/30/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/06/2025

FROM 0.40000ml of S10397 + 0.40000ml of S10584 + 0.40000ml of S11143 + 0.40000ml of S11487 + 0.40000ml of S11650 + 0.40000ml of S12478 + 0.50000ml of S11781 + 0.50000ml of S12470 + 0.60000ml of S11785 + 0.90000ml of S12518 + 0.90000ml of S12966 + 1.30000ml of S11782 + 1.30000ml of S11783 + 1.30000ml of S11784 + 1.30000ml of S12143 + 1.30000ml of S12144 + 1.30000ml of S12145 + 1.30000ml of S12146 + 1.30000ml of S12471 + 1.30000ml of S12472 + 1.30000ml of S12473 + 1.30000ml of S12474 + 1.30000ml of S12475 + 1.30000ml of S12476 + 1.30000ml of S12477 + 1.30000ml of S12519 + 1.30000ml of S12520 + 1.30000ml of S12521 + 1.30000ml of S12522 + 1.30000ml of S12523 + 1.30000ml of S12524 + 1.30000ml of S12525 + 1.30000ml of S12963 + 1.30000ml of S12964 + 1.30000ml of S12965 + 163.00000ml of E3846 = Final Quantity: 200.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3764	8270/625 Stock solution 100 ng	SP6721	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.26700ml of S10104 + 0.40000ml of S11495 + 0.50000ml of S12114 + 1.00000ml of S11087 + 1.00000ml of S11161 + 1.00000ml of S12270 + 1.00000ml of S12276 + 1.00000ml of S12791 + 3.83300ml of E3874 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
413	80 ng BNA ICC, 80 PPM	SP6722	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.20000ml of E3874 + 0.80000ml of SP6721 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
412	60 ng BNA ICC, 60 PPM	SP6723	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.40000ml of E3874 + 0.60000ml of SP6721 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
411	50 ng BNA ICC, 50 PPM	SP6724	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.50000ml of E3874 + 0.50000ml of SP6721 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
410	40 ng BNA ICC, 40 PPM	SP6725	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.60000ml of E3874 + 0.40000ml of SP6721 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3678	20 ng BNA ICC, 20 PPM	SP6726	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.80000ml of E3874 + 0.20000ml of SP6721 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
408	10 ng BNA ICC, 10 PPM	SP6727	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.90000ml of E3874 + 0.10000ml of SP6721 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
407	5 ng BNA ICC, 5 PPM	SP6728	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.95000ml of E3874 + 0.05000ml of SP6721 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
175	2.5 ng BNA ICC, 2.5 PPM	SP6729	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.50000ml of E3874 + 0.50000ml of SP6728 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	04/04/2025	10/04/2024 / Rajesh	10/04/2024 / Rajesh	E3815

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24G0862003	05/09/2025	11/09/2024 / Rajesh	11/04/2024 / Rajesh	E3828

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	06/26/2025	12/26/2024 / Rajesh	12/13/2024 / Rajesh	E3846

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	07/14/2025	01/14/2025 / Rajesh	12/27/2024 / Rajesh	E3871

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	07/30/2025	01/30/2025 / Rajesh	01/20/2025 / Rajesh	E3874

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	08/14/2025	02/14/2025 / Rajesh	12/27/2024 / Rajesh	E3878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	07/30/2025	01/30/2025 / anahy	12/09/2021 / Christian	S10104

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL,	A0182667	03/31/2025	01/15/2025 / Rahul	03/18/2022 / Christian	S10246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4]	A0185300	05/31/2025	01/29/2025 / anahy	05/18/2022 / Christian	S10397

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555868 / Custom Standard, Benzidine Std [CS 5328-1]	A0186373	06/30/2025	01/29/2025 / anahy	07/05/2022 / Christian	S10584

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10978

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10979

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10980

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11004

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11005

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11006

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11007

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11008

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11009

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11010

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0187043	05/15/2025	11/15/2024 / Jagrut	02/06/2023 / Christian	S11074

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010074-07 / 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 ml, (Maximum Expiration: 180 days)	406703	07/30/2025	01/30/2025 / anahy	02/07/2023 / Christian	S11087

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555869 / Custom Standard, hexachlorocyclopentadiene Std [CS 5328-2]	A0194702	07/29/2025	01/29/2025 / anahy	02/20/2023 / Christian	S11143

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110817-01 / Custom 8270 Mix, 4-55, 1000 mg/L, 1 ml, (Maximum Expiration: 90 Days)	414125	06/21/2025	01/30/2025 / anahy	03/06/2023 / Christian	S11161

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555870 / Custom Standard, 2,4-dinitrophenol Std [CS 5328-3]	A0200549	08/31/2026	01/29/2025 / anahy	08/10/2023 / yogesh	S11487

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	05/12/2025	11/12/2024 / Jagrut	08/11/2023 / Yogesh	S11495

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	07/29/2025	01/29/2025 / anahy	11/09/2023 / Yogesh	S11650

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	06/26/2025	12/26/2024 / Jagrut	11/21/2023 / Rahul	S11781

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	07/29/2025	01/29/2025 / anahy	11/21/2023 / Rahul	S11782

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	07/29/2025	01/29/2025 / anahy	11/21/2023 / Rahul	S11783

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	07/29/2025	01/29/2025 / anahy	11/21/2023 / Rahul	S11784

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	07/29/2025	01/29/2025 / anahy	11/21/2023 / Rahul	S11785

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	05/12/2025	11/12/2024 / Jagrut	03/08/2024 / Rahul	S12114

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	04/30/2025	11/14/2024 / anahy	03/15/2024 / Rahul	S12142

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	04/30/2025	01/29/2025 / anahy	03/15/2024 / Rahul	S12143

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	04/30/2025	01/29/2025 / anahy	03/15/2024 / Rahul	S12144

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	04/30/2025	01/29/2025 / anahy	03/15/2024 / Rahul	S12145

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	04/30/2025	01/29/2025 / anahy	03/15/2024 / Rahul	S12146

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12187

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12188

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12189

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12207

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	05/15/2025	11/15/2024 / Jagrut	03/15/2024 / Rahul	S12208

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	520963	07/30/2025	01/30/2025 / anahy	05/24/2024 / Rahul	S12270

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days)	495833	05/12/2025	11/12/2024 / Jagrut	05/24/2024 / Rahul	S12276

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0206540	05/12/2025	11/12/2024 / anahy	05/30/2024 / Rahul	S12327

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	05/14/2025	11/14/2024 / anahy	07/23/2024 / RAHUL	S12469

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	05/26/2025	11/26/2024 / Jagrut	07/23/2024 / RAHUL	S12470

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12471

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12472

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12473

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12474

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12475

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12476

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12477

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12478

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	05/14/2025	11/14/2024 / anahy	07/23/2024 / RAHUL	S12517

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/03/2025	01/03/2025 / Jagrut	07/23/2024 / RAHUL	S12518

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12519

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12520

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12521

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12523

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12524

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12525

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0212266	07/21/2025	01/21/2025 / anahy	09/20/2024 / anahy	S12649

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0212266	09/13/2025	03/13/2025 / Jagrut	09/20/2024 / anahy	S12655

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	06/21/2025	01/30/2025 / anahy	05/24/2024 / Rahul	S12791

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	07/29/2025	01/29/2025 / anahy	12/11/2024 / anahy	S12963

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	07/29/2025	01/29/2025 / anahy	12/11/2024 / anahy	S12964

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	07/29/2025	01/29/2025 / anahy	12/11/2024 / anahy	S12965

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	07/29/2025	01/29/2025 / anahy	12/11/2024 / anahy	S12966

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010074-07	406703	≤ -10 °C	Methylene Chloride	3/30/2025	3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
3,3'-dichlorobenzidine	91-94-1	99.5	74.3.26P	989 ± 7.53

Received on
02/07/23
by
CG

S11084
to
S11088

*Not a certified value

Certified By: _____
Jacob Mulloy
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricllly.

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110817-01	414125	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-55, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acetophenone	98-86-2	99.2	85.8.1P	998 ± 11.5
benzoic acid	65-85-0	100	123.7.1P	1010 ± 5.88
biphenyl	92-52-4	99.9	366.29.1P	999 ± 5.82
1,2,4,5-tetrachlorobenzene	95-94-3	99.7	53.7.2P	993 ± 5.79

Received on
02/07/23
by
CG
S11089
to
S11093

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-112090-04	440246	≤ -10 °C	Methylene Chloride	2/16/2026	CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d_4	93951-73-6	99.3	248.12.7P	7487 ± 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 ± 17.26
phenol-d6	13127-88-3	99.9	949.120.8P	7481 ± 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 ± 17.17

Received on
02/25/21
by
CG

S9236
to
S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: Erica Castiglione

Erica Castiglione
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on 03/08/22 by CG S10242 to S10247

Catalog No. : 31615 **Lot No.:** A0182667

Description : GC/MS Tuning Mixture

GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2025 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Pentachlorophenol **CAS #** 87-86-5 (Lot 211229RSR) **Purity** 99%	1,003.6 µg/mL	+/- 5.8897 µg/mL +/- 45.7132 µg/mL +/- 66.0037 µg/mL	Gravimetric Unstressed Stressed
2	DFTPP (Decafluorotriphenylphosphine) **CAS #** 5074-71-5 (Lot Q117-147) **Purity** 95%	1,006.6 µg/mL	+/- 5.9074 µg/mL +/- 45.8508 µg/mL +/- 66.2023 µg/mL	Gravimetric Unstressed Stressed
3	Benzidine **CAS #** 92-87-5 (Lot 211228JLM) **Purity** 99%	1,008.4 µg/mL	+/- 5.9179 µg/mL +/- 45.9318 µg/mL +/- 66.3193 µg/mL	Gravimetric Unstressed Stressed
4	4,4'-DDT **CAS #** 50-29-3 (Lot 210916JLM) **Purity** 99%	1,007.6 µg/mL	+/- 5.9132 µg/mL +/- 45.8954 µg/mL +/- 66.2667 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022 **Balance:** B345965662

Merlina Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Gravimetric Certificate

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 05/19/22

S10393 to S10402

Catalog No. : 555871 **Lot No.:** A0185300

Description : Custom 4-Nitrophenol Standard

Custom 4-Nitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2025 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	4-Nitrophenol	25,060.0 µg/mL	+/- 231.9100 µg/mL	Gravimetric
	CAS # 100-02-7 (Lot MKCN1089)		+/- 753.2622 µg/mL	Unstressed
	Purity 99%		+/- 905.6020 µg/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Katelyn McGinnis - Operations Tech I

Date Mixed: 16-May-2022 **Balance:** 1128342314

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Gravimetric Certificate

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by
CG
on
07/05/22
S 10583
to
S 10592

Catalog No. : 555868 **Lot No.:** A0186373

Description : Custom Benzidine Standard

Custom Benzidine Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2025 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Benzidine	25,200.0 µg/mL	+/- 233.2055 µg/mL	Gravimetric
	CAS # 92-87-5 (Lot 220511RSR)		+/- 351.6606 µg/mL	Unstressed
	Purity 99%		+/- 512.6054 µg/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Tom Suckar - Mix Technician **Date Mixed:** 16-Jun-2022 **Balance:** 1122030677

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on 02/06/23 by CG

S11071 to S11075

Catalog No. : 31853 **Lot No.:** A0187043

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	1,4-Dioxane	2,019.0 µg/mL	+/- 11.8486	µg/mL	Gravimetric
	CAS # 123-91-1 (Lot SHBN5929)		+/- 43.2570	µg/mL	Unstressed
	Purity 99%		+/- 44.5129	µg/mL	Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)


Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Brittany Federinko - Operations Tech I

Date Mixed: 07-Jul-2022 **Balance:** 1128360905

Marlina Cowan – Operations Tech II ARM QC

Date Passed: 12-Jul-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 S10951 to S10980

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol **CAS #** 367-12-4 **Purity** 99%	(Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL +/- 294.4162 µg/mL +/- 357.2628 µg/mL		Gravimetric Unstressed Stressed
2	Phenol-d6 **CAS #** 13127-88-3 **Purity** 99%	(Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL +/- 293.0957 µg/mL +/- 355.6603 µg/mL		Gravimetric Unstressed Stressed
3	2,4,6-Tribromophenol **CAS #** 118-79-6 **Purity** 99%	(Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL +/- 292.1253 µg/mL +/- 354.4829 µg/mL		Gravimetric Unstressed Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 **Balance:** 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 S10951 to S10980

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol **CAS #** 367-12-4 **Purity** 99%	(Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL +/- 294.4162 µg/mL +/- 357.2628 µg/mL		Gravimetric Unstressed Stressed
2	Phenol-d6 **CAS #** 13127-88-3 **Purity** 99%	(Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL +/- 293.0957 µg/mL +/- 355.6603 µg/mL		Gravimetric Unstressed Stressed
3	2,4,6-Tribromophenol **CAS #** 118-79-6 **Purity** 99%	(Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL +/- 292.1253 µg/mL +/- 354.4829 µg/mL		Gravimetric Unstressed Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 **Balance:** 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 S10951 to S10980

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol **CAS #** 367-12-4 **Purity** 99%	(Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL +/- 294.4162 µg/mL +/- 357.2628 µg/mL		Gravimetric Unstressed Stressed
2	Phenol-d6 **CAS #** 13127-88-3 **Purity** 99%	(Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL +/- 293.0957 µg/mL +/- 355.6603 µg/mL		Gravimetric Unstressed Stressed
3	2,4,6-Tribromophenol **CAS #** 118-79-6 **Purity** 99%	(Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL +/- 292.1253 µg/mL +/- 354.4829 µg/mL		Gravimetric Unstressed Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 **Balance:** 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555869 **Lot No.:** A0194702

Description : Custom Hexachlorocyclopentadiene Standard

Custom Hexachlorocyclopentadiene Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)
1	Hexachlorocyclopentadiene	77-47-4	0012019	99%	25,008.0 µg/mL

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 15-Feb-2023 **Balance:** B442140311

Manufactured under Restek Registered Quality Certificate #FM

tified Reference Material Notes

:es:

n date valid for unopened ampul stored in compliance with the recommended conditions.

nty, concentration, and expiration of the CRM are based on the unopened product being stored according to the ended condition found in the storage field.

d/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, LC/MS, RI, and/or melting point.

nds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A n factor is used to calculate the amount of compound necessary to achieve the desired concentration of the mpound in solution.

isomeric compounds is reported as the sum of the isomers.

lues are rounded to the nearest whole number.

rtainty Value Notes:

rtainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded ty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability ty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

erage factor of 2, which gives a level of confidence of approximately 95%.

ged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure nimum packaged amount can be sufficiently transferred.

Notes:

tion is based upon gravimetric preparation using either a balance whose calibration has been verified daily traceable weights, and/or dilutions with Class A glassware.

:

the unopened product, when stored in compliance with the recommended conditions, is guaranteed through ion displayed on the product label and certificate. Contact Restek for additional opened product stability ı, with the knowledge/understanding that open product stability is subject to the specific handling and ntal conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with ards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom m. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, des complete instructions.

ssolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely

PRODUCTOS QUIMICOS MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na_2SO_4
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na_2SO_4)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO_4)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreing matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RG-02-01, Ed. 3

Recd. by RB on 7/24/23 E3551

VWR®

Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Pellets

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

E 3657	E 3659
E 3654	E 3660

PO: PO2-329 PRODUCT CODE: SHIP DATE: 9/30/2024

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3815

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™ | J.T.Baker

Material No.: 9266-A4
Batch No.: 24J0862003
Manufactured Date: 2024-09-12
Expiration Date:2025-12-12
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	2
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3828

JCroak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date:2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 12/13/24

E 3846

JCroak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

12/9-1/24

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date:2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.5 ppm
Titrable Acid (µeq/g)	<= 0.3	0.0
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3871

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

PO: PO2-1178.2 PRODUCT CODE: SHIP DATE: 1/20/2025

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date:2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	4
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	99.9 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.8 ppm
Titrable Acid (µeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3874

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date:2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.5 ppm
Titrable Acid (µeq/g)	<= 0.3	0.0
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3878

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

Hydrochloric Acid, 36.5-38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

Material No.: 9530-33
Batch No.: 0000281827
Manufactured Date: 2021/03/30
Retest Date: 2026/03/29
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS - Assay (as HCl) (by acid-base titrn)	36.5 - 38.0 %	37.6
ACS - Color (APHA)	<= 10	5
ACS - Residue after Ignition	<= 3 ppm	1
ACS - Specific Gravity at 60°/60°F	1.185 - 1.192	1.189
ACS - Bromide (Br)	<= 0.005 %	< 0.005
ACS - Extractable Organic Substances	<= 5 ppm	< 1
ACS - Free Chlorine (as Cl_2)	<= 0.5 ppm	< 0.5
Phosphate (PO_4)	<= 0.05 ppm	< 0.03
Sulfate (SO_4)	<= 0.5 ppm	< 0.3
Sulfite (SO_3)	<= 0.8 ppm	0.3
Ammonium (NH_4)	<= 3 ppm	< 1
Trace Impurities - Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities - Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities - Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities - Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities - Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities - Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities - Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities - Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities - Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities - Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities - Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities - Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Material No.: 9530-33
Batch No.: 0000281827

Test	Specification	Result
Trace Impurities - Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities - Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities - Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities - Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities - Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities - Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities - Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities - Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities - Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities - Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities - Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities - Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities - Selenium (Se), For Information Only	ppb	1.0
Trace Impurities - Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities - Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities - Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities - Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities - Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities - Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities - Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities - Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities - Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities - Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities - Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use
Product Information (not specifications):
Appearance (clear, fuming liquid)
Meets ACS Specifications

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC



For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

S11484 ↓ S11493 Y.P. 08/10/23

Catalog No. : 555870 **Lot No.:** A0200549

Description : Custom 2,4-Dinitrophenol Standard

Custom 2,4-Dinitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	25,008.0 µg/mL	+/- 777.3323

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Tom Suckar - Mix Technician

Date Mixed: 02-Aug-2023 **Balance:** 1128342314

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d_4	2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl	321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5	4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14	1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 to 511498 } Y.P. 08/11/2023

*Not a certified value

Certified By: _____

Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0201728

Description : Custom Pentachlorophenol Standard

Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

511649 ↓ 511658 Y.P. 11/13/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023 **Balance:** B251644995

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

S12112 ↓ S12116 } RC/ 03/08/24

*Not a certified value

Certified By: Melissa Workoff

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

S12187 ↓ S12206 } RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

S12187 ↓ S12206 } RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

S12187 ↓ S12206 } RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

S12207 ↓ S12221 RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

S12207 ↓ S12221 RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	520963	≤ -10 °C	Methylene Chloride	10/10/2028	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1010 ± 9.89
acenaphthylene	208-96-8	97.6	14.290.1P	1014 ± 9.93
aniline	62-53-3	99.97	64.1.4P	1001 ± 9.8
anthracene	120-12-7	99.5	15.7.1P	999.6 ± 9.79
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 9.8
benzo[a]anthracene	56-55-3	100	16.7.3P	1007 ± 9.86
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1011 ± 14.11
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 10.96
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 13.95
benzo[a]pyrene	50-32-8	97	20.286.2P	999.9 ± 22.24
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 9.82
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	1000 ± 14.69
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1003 ± 13.89
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.15P	999.4 ± 14.68
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 9.8
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 17.03
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1.1P	1000 ± 13.85
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 16.79
carbazole	86-74-8	99.4	239.7.2P	1000 ± 9.8

S12270 ↓ S12274 } RC 05/24/24

*Not a certified value

Kerry E Kane

Certified By: _____

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
4-chloroaniline	106-47-8	100	66.7.1P	1000 ± 9.79
4-chlorophenylphenyl ether	7005-72-3	98	37.158.2P	1001 ± 17.07
4-chloro-3-methylphenol	59-50-7	99	102.1.2P	1006 ± 17.16
2-chloronaphthalene	91-58-7	99.9	42.7.6P	1000 ± 9.79
2-chlorophenol	95-57-8	99.8	103.7.1P	1007 ± 13.96
chrysene	218-01-9	96	21.286.2P	998.4 ± 12.85
dibenz[a,h]anthracene	53-70-3	99.44	22.286.3P	1000 ± 9.74
dibenzofuran	132-64-9	100	67.7.2.1P	1002 ± 9.77
di-n-butyl phthalate	84-74-2	99.84	40.286.1P	1007 ± 24.48
1,2-dichlorobenzene	95-50-1	99.8	43.7.1P	1000 ± 9.79
1,3-dichlorobenzene	541-73-1	99.5	44.1.3P	999.4 ± 9.79
1,4-dichlorobenzene	106-46-7	99.9	45.29.2P	1000 ± 9.79
2,4-dichlorophenol	120-83-2	99.6	104.7.1.1P	1005 ± 13.93
diethyl phthalate	84-66-2	99.8	38.7.1P	1011 ± 14
2,4-dimethylphenol	105-67-9	99.6	105.7.1.1P	1009 ± 13.98
dimethyl phthalate	131-11-3	99.9	39.9.2P	996.5 ± 13.8
1,2-dinitrobenzene	528-29-0	99.86	86.7.3.1P	999.5 ± 9.75
1,3-dinitrobenzene	99-65-0	100	313.7.2P	998 ± 9.79
1,4-dinitrobenzene	100-25-4	100	907.7.1P	999.5 ± 9.8
2,4-dinitrophenol	51-28-5	99.9	106.1.6DP	1002 ± 13.89
2,4-dinitrotoluene	121-14-2	100	87.7.3P	999.8 ± 13.85
2,6-dinitrotoluene	606-20-2	99.4	88.7.2.1P	999.6 ± 13.85
di-n-octyl phthalate	117-84-0	99.1	41.7.5P	991.6 ± 13.74
diphenylamine	122-39-4	100	78.1.6P	998 ± 13.79
2,3,5,6-tetrachlorophenol	935-95-5	97	1112.286.1P	1004 ± 14.02
fluoranthene	206-44-0	98.6	23.7.4P	999.6 ± 9.79
fluorene	86-73-7	98.4	24.7.1P	999.7 ± 9.79

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
hexachlorobenzene	118-74-1	99	46.158.4P	999.9 ± 13.96
hexachlorobutadiene	87-68-3	97.4	47.1.4P	1000 ± 9.79
hexachlorocyclopentadiene	77-47-4	99.2	48.2.2P	1001 ± 9.8
hexachloroethane	67-72-1	99.9	49.1.4P	1003 ± 9.82
indeno[1,2,3-cd]pyrene	193-39-5	98	25.286.4P	999.4 ± 22.23
isophorone	78-59-1	98.9	90.1.4P	999.9 ± 13.85
2-methyl-4,6-dinitrophenol	534-52-1	99.6	107.421.2DP	991 ± 24.09
1-methylnaphthalene	90-12-0	97.1	249.7.5P	999.2 ± 13.95
2-methylnaphthalene	91-57-6	97.4	68.7.2P	1006 ± 22.38
2-methylphenol	95-48-7	99.6	114.7.3P	1001 ± 13.87
3-methylphenol	108-39-4	99.1	115.7.4P	499.7 ± 6.92
4-methylphenol	106-44-5	99.5	116.7.1P	501.2 ± 6.94
naphthalene	91-20-3	99.8	26.9.1P	1018 ± 9.97
2-nitroaniline	88-74-4	99.7	69.29.1P	999.6 ± 9.79
3-nitroaniline	99-09-2	100	70.7.3P	1000 ± 9.74
4-nitroaniline	100-01-6	99.7	71.29.1P	1001 ± 9.8
nitrobenzene	98-95-3	100	94.7.1P	1000 ± 13.85
2-nitrophenol	88-75-5	99.1	108.29.1P	996.5 ± 13.81
4-nitrophenol	100-02-7	100	109.7.1P	1000 ± 13.82
N-nitrosodimethylamine	62-75-9	99.5	57.3.19P	998.5 ± 14.67
N-nitrosodi-n-propylamine	621-64-7	99.8	59.286.1P	996.8 ± 17
pentachlorophenol	87-86-5	99	110.1.7P	1004 ± 13.92
phenanthrene	85-01-8	99.7	27.1.5P	999 ± 12.87
phenol	108-95-2	100	112.7.1P	998.5 ± 13.8
pyrene	129-00-0	99.2	28.9.2P	998.9 ± 9.78
pyridine	110-86-1	100	101.24.1P	999 ± 9.73
2,3,4,6-Tetrachlorophenol	58-90-2	91.8	120.421.1P	996.5 ± 13.92

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	999.6 ± 9.79
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	999.5 ± 13.85
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	996 ± 13.8

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010442-07	495833	≤ -10 °C	Methylene Chloride	1/16/2028	Benzaldehyde Solution, 1000 mg/L, 1.3 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
benzaldehyde	100-52-7	98.3	442.421.1P	996.8 ± 11.49

S12275 ↓ S12279 } RC/ 05/24/24

*Not a certified value

Certified By: S. Hunter

Scott Hunter
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

S12312 ↓ S12331 } RC/ 05/30/24

Catalog No. : 31206 **Lot No.:** A0206540

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan - Operations Technician I

Date Mixed: 12-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0212266

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,000.6 µg/mL	+/- 90.1075
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,000.3 µg/mL	+/- 90.0925
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,000.4 µg/mL	+/- 90.1000
4	Phenanthrene-d10	1517-22-2	PR-34099	99%	2,000.5 µg/mL	+/- 90.1037
5	Chrysene-d12	1719-03-5	PR-33506	99%	2,000.7 µg/mL	+/- 90.1112
6	Perylene-d12	1520-96-3	PR-33205	99%	2,000.6 µg/mL	+/- 90.1075

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12645 ↓ S12674 AC 10/1/24

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0212266

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,000.6 µg/mL	+/- 90.1075
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,000.3 µg/mL	+/- 90.0925
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,000.4 µg/mL	+/- 90.1000
4	Phenanthrene-d10	1517-22-2	PR-34099	99%	2,000.5 µg/mL	+/- 90.1037
5	Chrysene-d12	1719-03-5	PR-33506	99%	2,000.7 µg/mL	+/- 90.1112
6	Perylene-d12	1520-96-3	PR-33205	99%	2,000.6 µg/mL	+/- 90.1075

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12645 ↓ S12674 AC 10/1/24

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤-10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

~~S12280~~ ↓ ~~S12284~~ RCI 05/24/24

New numbers Generated.

S12790 ↓ S12794 RCI 11/12/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0219438

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,008.3 µg/mL	+/- 36.6849
2	N-Nitrosodimethylamine	62-75-9	S240313RSR	99%	1,008.6 µg/mL	+/- 36.6985
3	Phenol	108-95-2	MKCK1120	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,002.9 µg/mL	+/- 36.4893
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.0 µg/mL	+/- 36.4938
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.6 µg/mL	+/- 36.5894
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.1 µg/mL	+/- 36.5348
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,002.1 µg/mL	+/- 36.4620
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,003.5 µg/mL	+/- 36.5120
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,005.3 µg/mL	+/- 36.5757
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,008.4 µg/mL	+/- 36.6894
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,004.6 µg/mL	+/- 36.5530
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 18.2697
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.8 µg/mL	+/- 18.3288
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,006.5 µg/mL	+/- 36.6212
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.5 µg/mL	+/- 36.5484
17	Nitrobenzene	98-95-3	10224044	99%	1,002.5 µg/mL	+/- 36.4757

18	Isophorone	78-59-1	MKCR3249	99%	1,003.4	µg/mL	+/- 36.5075
19	2-Nitrophenol	88-75-5	RP230710	99%	1,002.5	µg/mL	+/- 36.4757
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,006.5	µg/mL	+/- 36.6212
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,006.6	µg/mL	+/- 36.6257
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,001.5	µg/mL	+/- 36.4393
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,006.4	µg/mL	+/- 36.6166
24	Naphthalene	91-20-3	STBL1057	99%	1,002.1	µg/mL	+/- 36.4620
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.4	µg/mL	+/- 36.5439
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.5	µg/mL	+/- 36.4771
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,004.5	µg/mL	+/- 36.5484
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.0	µg/mL	+/- 36.3847
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 36.4325
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,006.4	µg/mL	+/- 36.6166
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.6	µg/mL	+/- 36.5505
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,004.3	µg/mL	+/- 36.5393
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.4	µg/mL	+/- 36.5439
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,002.8	µg/mL	+/- 36.4847
36	Acenaphthylene	208-96-8	RP241029RSR	98%	1,000.0	µg/mL	+/- 36.3835
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,006.3	µg/mL	+/- 36.6121
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,008.9	µg/mL	+/- 36.7076
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,006.6	µg/mL	+/- 36.6257
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,002.5	µg/mL	+/- 36.4757
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5530
43	2,4-Dinitrophenol	51-28-5	D240927RSR	----%	1,005.6	µg/mL	+/- 36.5894
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,003.5	µg/mL	+/- 36.5120
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,008.3	µg/mL	+/- 36.6849
46	4-Nitrophenol	100-02-7	20241029-2-AN	99%	1,004.8	µg/mL	+/- 36.5575
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,005.8	µg/mL	+/- 36.5939
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP231219RSR	99%	1,006.4	µg/mL	+/- 36.6166
49	Fluorene	86-73-7	10246250	98%	1,000.7	µg/mL	+/- 36.4102
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,004.9	µg/mL	+/- 36.5621
51	Diethylphthalate	84-66-2	BCCJ6241	99%	1,003.9	µg/mL	+/- 36.5257
52	4-Nitroaniline	100-01-6	RP230111	99%	1,006.6	µg/mL	+/- 36.6257
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,001.3	µg/mL	+/- 36.4302



54	Diphenylamine	122-39-4	MKCT1512	99%	1,003.0	µg/mL	+/- 36.4938
55	Azobenzene	103-33-3	BCCK0887	99%	1,002.4	µg/mL	+/- 36.4711
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,008.8	µg/mL	+/- 36.7031
57	Hexachlorobenzene	118-74-1	15458400	99%	1,005.1	µg/mL	+/- 36.5712
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.9	µg/mL	+/- 36.5984
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.9	µg/mL	+/- 36.5621
60	Anthracene	120-12-7	101492T18R	99%	1,005.1	µg/mL	+/- 36.5712
61	Carbazole	86-74-8	15276700	99%	1,005.4	µg/mL	+/- 36.5803
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,006.3	µg/mL	+/- 36.6121
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,003.5	µg/mL	+/- 36.5120
64	Pyrene	129-00-0	BCCK2592	99%	1,002.0	µg/mL	+/- 36.4575
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,007.5	µg/mL	+/- 36.6576
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.9	µg/mL	+/- 36.5984
67	Benz(a)anthracene	56-55-3	I70012022BAA	99%	1,005.5	µg/mL	+/- 36.5848
68	Chrysene	218-01-9	RP241007RSR	99%	1,005.3	µg/mL	+/- 36.5757
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,007.5	µg/mL	+/- 36.6576
70	Di-n-octyl phthalate	117-84-0	15566400	99%	1,002.3	µg/mL	+/- 36.4666
71	Benzo(b)fluoranthene	205-99-2	052013B	99%	1,004.1	µg/mL	+/- 36.5348
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,002.8	µg/mL	+/- 36.4847
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,006.2	µg/mL	+/- 36.6108
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,001.8	µg/mL	+/- 36.4490
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,003.3	µg/mL	+/- 36.5029
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,003.8	µg/mL	+/- 36.5217

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0219438

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,008.3 µg/mL	+/- 36.6849
2	N-Nitrosodimethylamine	62-75-9	S240313RSR	99%	1,008.6 µg/mL	+/- 36.6985
3	Phenol	108-95-2	MKCK1120	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,002.9 µg/mL	+/- 36.4893
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.0 µg/mL	+/- 36.4938
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.6 µg/mL	+/- 36.5894
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.1 µg/mL	+/- 36.5348
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,002.1 µg/mL	+/- 36.4620
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,003.5 µg/mL	+/- 36.5120
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,005.3 µg/mL	+/- 36.5757
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,008.4 µg/mL	+/- 36.6894
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,004.6 µg/mL	+/- 36.5530
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 18.2697
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.8 µg/mL	+/- 18.3288
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,006.5 µg/mL	+/- 36.6212
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.5 µg/mL	+/- 36.5484
17	Nitrobenzene	98-95-3	10224044	99%	1,002.5 µg/mL	+/- 36.4757

18	Isophorone	78-59-1	MKCR3249	99%	1,003.4	µg/mL	+/- 36.5075
19	2-Nitrophenol	88-75-5	RP230710	99%	1,002.5	µg/mL	+/- 36.4757
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,006.5	µg/mL	+/- 36.6212
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,006.6	µg/mL	+/- 36.6257
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,001.5	µg/mL	+/- 36.4393
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,006.4	µg/mL	+/- 36.6166
24	Naphthalene	91-20-3	STBL1057	99%	1,002.1	µg/mL	+/- 36.4620
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.4	µg/mL	+/- 36.5439
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.5	µg/mL	+/- 36.4771
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,004.5	µg/mL	+/- 36.5484
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.0	µg/mL	+/- 36.3847
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 36.4325
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,006.4	µg/mL	+/- 36.6166
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.6	µg/mL	+/- 36.5505
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,004.3	µg/mL	+/- 36.5393
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.4	µg/mL	+/- 36.5439
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,002.8	µg/mL	+/- 36.4847
36	Acenaphthylene	208-96-8	RP241029RSR	98%	1,000.0	µg/mL	+/- 36.3835
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,006.3	µg/mL	+/- 36.6121
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,008.9	µg/mL	+/- 36.7076
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,006.6	µg/mL	+/- 36.6257
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,002.5	µg/mL	+/- 36.4757
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5530
43	2,4-Dinitrophenol	51-28-5	D240927RSR	----%	1,005.6	µg/mL	+/- 36.5894
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,003.5	µg/mL	+/- 36.5120
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,008.3	µg/mL	+/- 36.6849
46	4-Nitrophenol	100-02-7	20241029-2-AN	99%	1,004.8	µg/mL	+/- 36.5575
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,005.8	µg/mL	+/- 36.5939
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP231219RSR	99%	1,006.4	µg/mL	+/- 36.6166
49	Fluorene	86-73-7	10246250	98%	1,000.7	µg/mL	+/- 36.4102
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,004.9	µg/mL	+/- 36.5621
51	Diethylphthalate	84-66-2	BCCJ6241	99%	1,003.9	µg/mL	+/- 36.5257
52	4-Nitroaniline	100-01-6	RP230111	99%	1,006.6	µg/mL	+/- 36.6257
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,001.3	µg/mL	+/- 36.4302



54	Diphenylamine	122-39-4	MKCT1512	99%	1,003.0	µg/mL	+/- 36.4938
55	Azobenzene	103-33-3	BCCK0887	99%	1,002.4	µg/mL	+/- 36.4711
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,008.8	µg/mL	+/- 36.7031
57	Hexachlorobenzene	118-74-1	15458400	99%	1,005.1	µg/mL	+/- 36.5712
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.9	µg/mL	+/- 36.5984
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.9	µg/mL	+/- 36.5621
60	Anthracene	120-12-7	101492T18R	99%	1,005.1	µg/mL	+/- 36.5712
61	Carbazole	86-74-8	15276700	99%	1,005.4	µg/mL	+/- 36.5803
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,006.3	µg/mL	+/- 36.6121
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,003.5	µg/mL	+/- 36.5120
64	Pyrene	129-00-0	BCCK2592	99%	1,002.0	µg/mL	+/- 36.4575
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,007.5	µg/mL	+/- 36.6576
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.9	µg/mL	+/- 36.5984
67	Benz(a)anthracene	56-55-3	I70012022BAA	99%	1,005.5	µg/mL	+/- 36.5848
68	Chrysene	218-01-9	RP241007RSR	99%	1,005.3	µg/mL	+/- 36.5757
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,007.5	µg/mL	+/- 36.6576
70	Di-n-octyl phthalate	117-84-0	15566400	99%	1,002.3	µg/mL	+/- 36.4666
71	Benzo(b)fluoranthene	205-99-2	052013B	99%	1,004.1	µg/mL	+/- 36.5348
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,002.8	µg/mL	+/- 36.4847
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,006.2	µg/mL	+/- 36.6108
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,001.8	µg/mL	+/- 36.4490
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,003.3	µg/mL	+/- 36.5029
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,003.8	µg/mL	+/- 36.5217

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0219438

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,008.3 µg/mL	+/- 36.6849
2	N-Nitrosodimethylamine	62-75-9	S240313RSR	99%	1,008.6 µg/mL	+/- 36.6985
3	Phenol	108-95-2	MKCK1120	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,002.9 µg/mL	+/- 36.4893
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.0 µg/mL	+/- 36.4938
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.6 µg/mL	+/- 36.5894
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.1 µg/mL	+/- 36.5348
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,002.1 µg/mL	+/- 36.4620
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,003.5 µg/mL	+/- 36.5120
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,005.3 µg/mL	+/- 36.5757
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,008.4 µg/mL	+/- 36.6894
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,004.6 µg/mL	+/- 36.5530
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 18.2697
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.8 µg/mL	+/- 18.3288
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,006.5 µg/mL	+/- 36.6212
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.5 µg/mL	+/- 36.5484
17	Nitrobenzene	98-95-3	10224044	99%	1,002.5 µg/mL	+/- 36.4757

18	Isophorone	78-59-1	MKCR3249	99%	1,003.4	µg/mL	+/- 36.5075
19	2-Nitrophenol	88-75-5	RP230710	99%	1,002.5	µg/mL	+/- 36.4757
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,006.5	µg/mL	+/- 36.6212
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,006.6	µg/mL	+/- 36.6257
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,001.5	µg/mL	+/- 36.4393
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,006.4	µg/mL	+/- 36.6166
24	Naphthalene	91-20-3	STBL1057	99%	1,002.1	µg/mL	+/- 36.4620
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.4	µg/mL	+/- 36.5439
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.5	µg/mL	+/- 36.4771
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,004.5	µg/mL	+/- 36.5484
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.0	µg/mL	+/- 36.3847
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 36.4325
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,006.4	µg/mL	+/- 36.6166
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.6	µg/mL	+/- 36.5505
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,004.3	µg/mL	+/- 36.5393
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.4	µg/mL	+/- 36.5439
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,002.8	µg/mL	+/- 36.4847
36	Acenaphthylene	208-96-8	RP241029RSR	98%	1,000.0	µg/mL	+/- 36.3835
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,006.3	µg/mL	+/- 36.6121
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,008.9	µg/mL	+/- 36.7076
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,006.6	µg/mL	+/- 36.6257
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,002.5	µg/mL	+/- 36.4757
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5530
43	2,4-Dinitrophenol	51-28-5	D240927RSR	----%	1,005.6	µg/mL	+/- 36.5894
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,003.5	µg/mL	+/- 36.5120
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,008.3	µg/mL	+/- 36.6849
46	4-Nitrophenol	100-02-7	20241029-2-AN	99%	1,004.8	µg/mL	+/- 36.5575
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,005.8	µg/mL	+/- 36.5939
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP231219RSR	99%	1,006.4	µg/mL	+/- 36.6166
49	Fluorene	86-73-7	10246250	98%	1,000.7	µg/mL	+/- 36.4102
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,004.9	µg/mL	+/- 36.5621
51	Diethylphthalate	84-66-2	BCCJ6241	99%	1,003.9	µg/mL	+/- 36.5257
52	4-Nitroaniline	100-01-6	RP230111	99%	1,006.6	µg/mL	+/- 36.6257
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,001.3	µg/mL	+/- 36.4302



54	Diphenylamine	122-39-4	MKCT1512	99%	1,003.0	µg/mL	+/- 36.4938
55	Azobenzene	103-33-3	BCCK0887	99%	1,002.4	µg/mL	+/- 36.4711
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,008.8	µg/mL	+/- 36.7031
57	Hexachlorobenzene	118-74-1	15458400	99%	1,005.1	µg/mL	+/- 36.5712
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.9	µg/mL	+/- 36.5984
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.9	µg/mL	+/- 36.5621
60	Anthracene	120-12-7	101492T18R	99%	1,005.1	µg/mL	+/- 36.5712
61	Carbazole	86-74-8	15276700	99%	1,005.4	µg/mL	+/- 36.5803
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,006.3	µg/mL	+/- 36.6121
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,003.5	µg/mL	+/- 36.5120
64	Pyrene	129-00-0	BCCK2592	99%	1,002.0	µg/mL	+/- 36.4575
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,007.5	µg/mL	+/- 36.6576
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.9	µg/mL	+/- 36.5984
67	Benz(a)anthracene	56-55-3	I70012022BAA	99%	1,005.5	µg/mL	+/- 36.5848
68	Chrysene	218-01-9	RP241007RSR	99%	1,005.3	µg/mL	+/- 36.5757
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,007.5	µg/mL	+/- 36.6576
70	Di-n-octyl phthalate	117-84-0	15566400	99%	1,002.3	µg/mL	+/- 36.4666
71	Benzo(b)fluoranthene	205-99-2	052013B	99%	1,004.1	µg/mL	+/- 36.5348
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,002.8	µg/mL	+/- 36.4847
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,006.2	µg/mL	+/- 36.6108
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,001.8	µg/mL	+/- 36.4490
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,003.3	µg/mL	+/- 36.5029
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,003.8	µg/mL	+/- 36.5217

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0219438

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,008.3 µg/mL	+/- 36.6849
2	N-Nitrosodimethylamine	62-75-9	S240313RSR	99%	1,008.6 µg/mL	+/- 36.6985
3	Phenol	108-95-2	MKCK1120	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,002.9 µg/mL	+/- 36.4893
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.0 µg/mL	+/- 36.4938
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.6 µg/mL	+/- 36.5894
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.1 µg/mL	+/- 36.5348
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,002.1 µg/mL	+/- 36.4620
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,003.5 µg/mL	+/- 36.5120
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,005.3 µg/mL	+/- 36.5757
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,008.4 µg/mL	+/- 36.6894
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,004.6 µg/mL	+/- 36.5530
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 18.2697
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.8 µg/mL	+/- 18.3288
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,006.5 µg/mL	+/- 36.6212
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.5 µg/mL	+/- 36.5484
17	Nitrobenzene	98-95-3	10224044	99%	1,002.5 µg/mL	+/- 36.4757

18	Isophorone	78-59-1	MKCR3249	99%	1,003.4	µg/mL	+/- 36.5075
19	2-Nitrophenol	88-75-5	RP230710	99%	1,002.5	µg/mL	+/- 36.4757
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,006.5	µg/mL	+/- 36.6212
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,006.6	µg/mL	+/- 36.6257
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,001.5	µg/mL	+/- 36.4393
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,006.4	µg/mL	+/- 36.6166
24	Naphthalene	91-20-3	STBL1057	99%	1,002.1	µg/mL	+/- 36.4620
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.4	µg/mL	+/- 36.5439
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.5	µg/mL	+/- 36.4771
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,004.5	µg/mL	+/- 36.5484
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.0	µg/mL	+/- 36.3847
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 36.4325
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,006.4	µg/mL	+/- 36.6166
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.6	µg/mL	+/- 36.5505
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,004.3	µg/mL	+/- 36.5393
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.4	µg/mL	+/- 36.5439
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,002.8	µg/mL	+/- 36.4847
36	Acenaphthylene	208-96-8	RP241029RSR	98%	1,000.0	µg/mL	+/- 36.3835
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,006.3	µg/mL	+/- 36.6121
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,008.9	µg/mL	+/- 36.7076
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,006.6	µg/mL	+/- 36.6257
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,002.5	µg/mL	+/- 36.4757
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5530
43	2,4-Dinitrophenol	51-28-5	D240927RSR	----%	1,005.6	µg/mL	+/- 36.5894
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,003.5	µg/mL	+/- 36.5120
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,008.3	µg/mL	+/- 36.6849
46	4-Nitrophenol	100-02-7	20241029-2-AN	99%	1,004.8	µg/mL	+/- 36.5575
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,005.8	µg/mL	+/- 36.5939
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP231219RSR	99%	1,006.4	µg/mL	+/- 36.6166
49	Fluorene	86-73-7	10246250	98%	1,000.7	µg/mL	+/- 36.4102
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,004.9	µg/mL	+/- 36.5621
51	Diethylphthalate	84-66-2	BCCJ6241	99%	1,003.9	µg/mL	+/- 36.5257
52	4-Nitroaniline	100-01-6	RP230111	99%	1,006.6	µg/mL	+/- 36.6257
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,001.3	µg/mL	+/- 36.4302



54	Diphenylamine	122-39-4	MKCT1512	99%	1,003.0	µg/mL	+/- 36.4938
55	Azobenzene	103-33-3	BCCK0887	99%	1,002.4	µg/mL	+/- 36.4711
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,008.8	µg/mL	+/- 36.7031
57	Hexachlorobenzene	118-74-1	15458400	99%	1,005.1	µg/mL	+/- 36.5712
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.9	µg/mL	+/- 36.5984
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.9	µg/mL	+/- 36.5621
60	Anthracene	120-12-7	101492T18R	99%	1,005.1	µg/mL	+/- 36.5712
61	Carbazole	86-74-8	15276700	99%	1,005.4	µg/mL	+/- 36.5803
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,006.3	µg/mL	+/- 36.6121
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,003.5	µg/mL	+/- 36.5120
64	Pyrene	129-00-0	BCCK2592	99%	1,002.0	µg/mL	+/- 36.4575
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,007.5	µg/mL	+/- 36.6576
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.9	µg/mL	+/- 36.5984
67	Benz(a)anthracene	56-55-3	I70012022BAA	99%	1,005.5	µg/mL	+/- 36.5848
68	Chrysene	218-01-9	RP241007RSR	99%	1,005.3	µg/mL	+/- 36.5757
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,007.5	µg/mL	+/- 36.6576
70	Di-n-octyl phthalate	117-84-0	15566400	99%	1,002.3	µg/mL	+/- 36.4666
71	Benzo(b)fluoranthene	205-99-2	052013B	99%	1,004.1	µg/mL	+/- 36.5348
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,002.8	µg/mL	+/- 36.4847
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,006.2	µg/mL	+/- 36.6108
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,001.8	µg/mL	+/- 36.4490
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,003.3	µg/mL	+/- 36.5029
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,003.8	µg/mL	+/- 36.5217

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.



SHIPPING DOCUMENTS

phenova® Certified Reference Materials | A Phenomenex® Company

Packing List

Date	Order #
03/03/2025	333289

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsofsale

Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: SJ
3/5/2025 14:30

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
PO2-1517	Net 30	ZCM-100	1500470	FedEx 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
			PT-TMSET-WP	WP Trace Metals Set : (TM1, HG and SNTI)		
1	1	0	PT-TM1-WP	WP Trace Metals 1	WP0325	8264-04
1	1	0	PT-HG-WP	WP Mercury	WP0325	8264-05
1	1	0	PT-SNTI-WP	WP Tin & Titanium	WP0325	8264-38
1	1	0	PT-CR6-WP	WP Hexavalent Chromium	WP0325	8264-06
1	1	0	PT-DEM-WP	WP Demand	WP0325	8264-07
			PT-MINSET-WP	WP Minerals Set : (MIN1, MIN2 and COND)		
1	1	0	PT-MIN1-WP	WP Minerals 1 Only	WP0325	8264-08
1	1	0	PT-MIN2-WP	WP Minerals 2 Only	WP0325	8264-102
1	1	0	PT-COND-WP	WP Conductivity Only	WP0325	8264-72
1	1	0	PT-SOL-WP	WP Solids	WP0325	8264-09
			PT-NUTSET-WP	WP Nutrients Set : (NUT1, NUT2 and NUT3)		
1	1	0	PT-NUT1-WP	WP NUT1 Simple Nutrients Only	WP0325	8264-10
1	1	0	PT-NUT2-WP	WP NUT2 - Complex Nutrients	WP0325	8264-11
1	1	0	PT-NUT3-WP	WP NUT3 - Nitrite Only	WP0325	8264-69
1	1	0	PT-OGR1L-WP	WP Oil and Grease 1L	WP0325	8264-103
1	1	0	PT-CL-WP	WP Residual Chlorine	WP0325	8264-13
1	1	0	PT-PH-WP	WP pH	WP0325	8264-15
1	1	0	PT-CN-WP	WP Cyanide	WP0325	8264-14
1	1	0	PT-PHEN-WP	WP Phenolics	WP0325	8264-16

phenova® Certified Reference Materials | A Phenomenex® Company

Packing List

Date	Order #
03/03/2025	333289

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsofsale

Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: SJ
3/5/2025 14:30

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
PO2-1517	Net 30	ZCM-100	1500470	FedEx 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
1	1	0	PT-S2-WP	WP Sulfide	WP0325	8264-22
1	1	0	PT-SSOL-WP	WP Settleable Solids	WP0325	8264-17
1	1	0	PT-TURB-WP	WP Turbidity	WP0325	8264-20
1	1	0	PT-VOA-WP	WP Volatiles	WP0325	8264-26
1	1	0	PT-BN-WP	WP Base Neutrals	WP0325	8264-27
1	1	0	PT-ACIDS-WP	WP Acids	WP0325	8264-28
1	1	0	PT-PEST-WP	WP Pesticides	WP0325	8264-29
1	1	0	PT-CHLR-WP	WP Chlordane	WP0325	8264-30
1	1	0	PT-TXP-WP	WP Toxaphene	WP0325	8264-31
1	1	0	PT-PCBW-WP	WP PCBs in Water	WP0325	8264-32
1	1	0	PT-HERB-WP	WP Herbicides	WP0325	8264-36
1	1	0	RR-TPH1L-WP	WP TPH 1L	R40367	R40367-104
1	1	0	RR-VSOL-WP	WP Volatile Solids	R40367	R40367-18
1	1	0	RR-SIO2-WP	WP Silica	R40367	R40367-21
1	1	0	RR-COL-WP	WP Color	R40367	R40367-51
1	1	0	RR-GAS-WP	WP Gasoline Range Organics	R40367	R40367-62
1	1	0	RR-DIES-WP	WP Diesel Range Organics	R40367	R40367-63
1	1	0	RR-8011-WP	WP EDB/DBCP/TCP	R40367	R40367-98
1	1	0	RR-PAH-WP	WP PAH-Low Level	R40433	R40433-37

phenova® Certified Reference Materials | A Phenomenex® Company

Packing List

Date	Order #
03/07/2025	335989

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsofsale

Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: SJ
3/11/2025 9:55

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
Email: Sohil Jodhani	Net 30	ZCM-100	1500470	FedEx 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
1	1	0	RR-TRIAZINE-WP	WP Triazine Pesticides	R40480	R40480-108

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

QA Control Code: A2070148