

Surrogate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167097BL	PB167097BL	2-Fluorophenol	150	134	90		10	139
		Phenol-d6	150	130	86		10	134
		2,4,6-Tribromophenol	150	154	102		44	137
PB167097BS	PB167097BS	2-Fluorophenol	150	140	93		10	139
		Phenol-d6	150	137	91		10	134
		2,4,6-Tribromophenol	150	160	107		44	137
PB167097BSD	PB167097BSD	2-Fluorophenol	150	132	88		10	139
		Phenol-d6	150	129	86		10	134
		2,4,6-Tribromophenol	150	151	101		44	137
Q1502-06	PT-ACIDS-WP	2-Fluorophenol	150	145	97		10	139
		Phenol-d6	150	142	95		10	134
		2,4,6-Tribromophenol	150	151	101		44	137
Q1502-06DL	PT-ACIDS-WPDL	2-Fluorophenol	150	149	99		10	139
		Phenol-d6	150	148	99		10	134
		2,4,6-Tribromophenol	150	144	96		44	137