

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E DataFile: BF141941.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167097BS	Phenol	50	45.8	ug/L	92				66	118	
	2-Chlorophenol	50	46.9	ug/L	94				70	117	
	2-Methylphenol	50	46.7	ug/L	93				69	109	
	3+4-Methylphenols	50	46.3	ug/L	93				67	106	
	2-Nitrophenol	50	47.5	ug/L	95				70	115	
	2,4-Dimethylphenol	50	57.6	ug/L	115				42	142	
	2,4-Dichlorophenol	50	46.1	ug/L	92				66	115	
	Benzoic acid	50	52.7	ug/L	105				10	125	
	4-Chloro-3-methylphenol	50	45.9	ug/L	92				65	114	
	2,4,6-Trichlorophenol	50	49.2	ug/L	98				61	110	
	2,4,5-Trichlorophenol	50	45.4	ug/L	91				70	106	
	2,4-Dinitrophenol	100	99.2	ug/L	99				36	166	
	4-Nitrophenol	100	100	ug/L	100				45	147	
	4,6-Dinitro-2-methylphenol	50	51.0	ug/L	102				62	132	
	Pentachlorophenol	100	99.9	ug/L	100				47	114	
	2,3,4,6-Tetrachlorophenol	50	46.3	ug/L	93				63	116	