

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

DataFile: BF141942.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		RPD
							Qual	Qual	Low	High	
PB167097BSD	Phenol	50	43.0	ug/L	86	6			66	118	20
	2-Chlorophenol	50	44.2	ug/L	88	6			70	117	20
	2-Methylphenol	50	44.5	ug/L	89	5			69	109	20
	3+4-Methylphenols	50	44.0	ug/L	88	5			67	106	20
	2-Nitrophenol	50	45.9	ug/L	92	3			70	115	20
	2,4-Dimethylphenol	50	54.8	ug/L	110	5			42	142	20
	2,4-Dichlorophenol	50	43.8	ug/L	88	5			66	115	20
	Benzoic acid	50	52.1	ug/L	104	1			10	125	20
	4-Chloro-3-methylphenol	50	43.1	ug/L	86	6			65	114	20
	2,4,6-Trichlorophenol	50	45.1	ug/L	90	9			61	110	20
	2,4,5-Trichlorophenol	50	45.7	ug/L	91	1			70	106	20
	2,4-Dinitrophenol	100	95.0	ug/L	95	4			36	166	20
	4-Nitrophenol	100	97.3	ug/L	97	3			45	147	20
	4,6-Dinitro-2-methylphenol	50	49.9	ug/L	100	2			62	132	20
	Pentachlorophenol	100	95.6	ug/L	96	4			47	114	20
	2,3,4,6-Tetrachlorophenol	50	43.9	ug/L	88	5			63	116	20