

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/13/2025

Lab File ID: BP024171.D Time Analyzed: 10:22

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	330996	7.764	1338420	10.55	811311	14.41
UPPER LIMIT	661992	8.264	2676840	11.046	1622620	14.91
LOWER LIMIT	165498	7.264	669210	10.046	405656	13.91
EPA SAMPLE NO.						
01 PT-ACIDS-WP	332785	7.74	1357430	10.53	822876	14.42
02 PT-ACIDS-WPDL	333252	7.75	1381030	10.54	845510	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.