

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q1557-05MSD		Client Sample ID: BPOW6-9-20250312MSD		DataFile: BN036612.D							
n-Nitrosodimethylamine	0.4	0	0.20	ug/L	50	0			10	158	20
Hexachlorobutadiene	0.4	0	0.40	ug/L	100	5			14	140	20
Hexachlorobenzene	0.4	0	0.46	ug/L	115	4			31	160	20
Atrazine	0.4	0	0.56	ug/L	140	4			20	150	20
1,4-Dioxane	0.4	0	0.23	ug/L	58	5			10	175	20