

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270-Modified DataFile: BN036630.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	
								Qual		High	RPD
PB167128BS	n-Nitrosodimethylamine	0.4	0.39	ug/L	98				33	130	
	Hexachlorobutadiene	0.4	0.39	ug/L	98				33	114	
	Hexachlorobenzene	0.4	0.45	ug/L	113				62	127	
	Atrazine	0.4	0.41	ug/L	103				20	150	
	1,4-Dioxane	0.4	0.36	ug/L	90				42	127	