

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ALLI03

Lab Code: CHEM

SAS No.: Q1502

SDG NO.: Q1502

Lab File ID: BN036600.D

DFTPP Injection Date: 03/14/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:30

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	65.5
68	Less than 2.0% of mass 69	0.9 (1.6) 1
69	Mass 69 relative abundance	55.5
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	51.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	24.7
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	9.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036601.D	03/14/2025	10:09
PT-BN-WP	Q1502-05	BN036605.D	03/14/2025	13:05
PT-BN-WPDL	Q1502-05DL	BN036606.D	03/14/2025	13:57
BPOW6-9-20250312MS	Q1557-04MS	BN036611.D	03/14/2025	17:24
BPOW6-9-20250312MSD	Q1557-05MSD	BN036612.D	03/14/2025	18:00