

Surrogate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167128BL	PB167128BL	2-Fluorophenol	0.4	0.38	94		10	100
		Phenol-d6	0.4	0.30	74		10	100
		2,4,6-Tribromophenol	0.4	0.25	62		10	131
PB167128BS	PB167128BS	2-Fluorophenol	0.4	0.38	94		10	100
		Phenol-d6	0.4	0.35	88		10	100
		2,4,6-Tribromophenol	0.4	0.23	58		10	131
Q1502-08	PT-ACIDS-WP	2-Fluorophenol	150	182	121	*	10	100
		Phenol-d6	150	199	132	*	10	100
		2,4,6-Tribromophenol	150	235	156	*	10	131
Q1502-08DL	PT-ACIDS-WPDL	2-Fluorophenol	150	155	103	*	10	100
		Phenol-d6	150	164	109	*	10	100
		2,4,6-Tribromophenol	150	171	114		10	131
Q1557-04MS	BPOW6-9-20250312MS	2-Fluorophenol	0.4	0.17	43		10	100
		Phenol-d6	0.4	0.094	23		10	100
		2,4,6-Tribromophenol	0.4	0.38	95		10	131
Q1557-05MSD	BPOW6-9-20250312MSD	2-Fluorophenol	0.4	0.18	44		10	100
		Phenol-d6	0.4	0.11	26		10	100
		2,4,6-Tribromophenol	0.4	0.43	106		10	131