

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1557-04MS	Client Sample ID:	BPOW6-9-20250312MS					DataFile:	BN036611.D		
Naphthalene	0.4	0	0.41	ug/L	103				62	119	
Acenaphthylene	0.4	0	0.44	ug/L	110				53	123	
Acenaphthene	0.4	0	0.43	ug/L	108				57	122	
Fluorene	0.4	0	0.44	ug/L	110				58	123	
Phenanthrene	0.4	0	0.50	ug/L	125				64	129	
Anthracene	0.4	0	0.51	ug/L	128	*			67	127	
Fluoranthene	0.4	0	0.52	ug/L	130				55	140	
Pyrene	0.4	0	0.50	ug/L	125				59	137	
Benzo(a)anthracene	0.4	0	0.50	ug/L	125				51	146	
Chrysene	0.4	0	0.51	ug/L	128	*			61	126	
Benzo(b)fluoranthene	0.4	0	0.51	ug/L	128	*			60	123	
Benzo(k)fluoranthene	0.4	0	0.55	ug/L	138	*			61	122	
Benzo(a)pyrene	0.4	0	0.53	ug/L	133	*			54	129	
Indeno(1,2,3-cd)pyrene	0.4	0	0.52	ug/L	130				50	134	
Dibenz(a,h)anthracene	0.4	0	0.49	ug/L	123	*			53	121	
Benzo(g,h,i)perylene	0.4	0	0.48	ug/L	120				44	132	