

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1557-05MSD	Client Sample ID:	BPOW6-9-20250312MSD					DataFile:	BN036612.D		
Naphthalene	0.4	0	0.43	ug/L	108		5		62	119	20
Acenaphthylene	0.4	0	0.48	ug/L	120		9		53	123	20
Acenaphthene	0.4	0	0.47	ug/L	117		8		57	122	20
Fluorene	0.4	0	0.47	ug/L	117		6		58	123	20
Phenanthrene	0.4	0	0.53	ug/L	133	*	6		64	129	20
Anthracene	0.4	0	0.53	ug/L	133	*	4		67	127	20
Fluoranthene	0.4	0	0.55	ug/L	138		6		55	140	20
Pyrene	0.4	0	0.51	ug/L	128		2		59	137	20
Benzo(a)anthracene	0.4	0	0.54	ug/L	135		8		51	146	20
Chrysene	0.4	0	0.55	ug/L	138	*	8		61	126	20
Benzo(b)fluoranthene	0.4	0	0.54	ug/L	135	*	5		60	123	20
Benzo(k)fluoranthene	0.4	0	0.56	ug/L	140	*	1		61	122	20
Benzo(a)pyrene	0.4	0	0.58	ug/L	145	*	9		54	129	20
Indeno(1,2,3-cd)pyrene	0.4	0	0.55	ug/L	138	*	6		50	134	20
Dibenz(a,h)anthracene	0.4	0	0.53	ug/L	133	*	8		53	121	20
Benzo(g,h,i)perylene	0.4	0	0.50	ug/L	125		4		44	132	20