

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ALLI03
Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG No.: Q1502
Instrument ID: BNA_P Calibration Date(s): 03/12/2025 03/12/2025
Calibration Time(s): 16:17 21:02

LAB FILE ID:		RRF2.5 = BP024160.D		RRF005 = BP024161.D		RRF010 = BP024162.D		
		RRF020 = BP024163.D		RRF040 = BP024164.D		RRF050 = BP024165.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.158	1.194	1.307	1.211	1.291	1.253	5.1
Phenol-d6		1.518	1.581	1.747	1.631	1.749	1.675	5.9
Nitrobenzene-d5		0.330	0.338	0.366	0.351	0.362	0.354	4.3
2-Fluorobiphenyl		1.364	1.358	1.420	1.319	1.335	1.356	2.4
2,4,6-Tribromophenol		0.203	0.222	0.256	0.252	0.270	0.252	11.7
Atrazine		0.165	0.171	0.157	0.135	0.105	0.147	18.5
Terphenyl-d14		0.967	0.973	1.033	0.948	0.966	0.969	3.4