

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG No.: Q1502

Instrument ID: BNA_P Calibration Date/Time: 03/13/2025 10:22

Lab File ID: BP024171.D Init. Calib. Date(s): 03/12/2025 03/12/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 16:17 21:02

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.253	1.221		-2.6	
Phenol-d6	1.675	1.634		-2.4	
Nitrobenzene-d5	0.354	0.352		-0.6	
2-Fluorobiphenyl	1.356	1.335		-1.5	
2,4,6-Tribromophenol	0.252	0.256		1.6	
Atrazine	0.147	0.132		-10.2	
Terphenyl-d14	0.969	0.947		-2.3	

All other compounds must meet a minimum RRF of 0.010.