

Surrogate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167097BL	PB167097BL	Nitrobenzene-d5	100	93.3	93		49	133
		2-Fluorobiphenyl	100	90.0	90		52	132
		Terphenyl-d14	100	87.3	87		48	125
PB167097BS	PB167097BS	Nitrobenzene-d5	100	98.6	99		49	133
		2-Fluorobiphenyl	100	97.8	98		52	132
		Terphenyl-d14	100	105	105		48	125
PB167097BSD	PB167097BSD	Nitrobenzene-d5	100	93.2	93		49	133
		2-Fluorobiphenyl	100	93.8	94		52	132
		Terphenyl-d14	100	101	101		48	125
Q1502-22	RR-TRIAZINE-WP	Nitrobenzene-d5	100	109	109		49	133
		2-Fluorobiphenyl	100	110	110		52	132
		Terphenyl-d14	100	123	123		48	125