

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ALLI03
 Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502
 Lab File ID: VX045211.D Date Analyzed: 03/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	101357	5.54	176548	6.75	153652	10.05
UPPER LIMIT	202714	6.044	353096	7.251	307304	10.549
LOWER LIMIT	50678.5	5.044	88274	6.251	76826	9.549
EPA SAMPLE NO.						
PT-VOA-WP	98273	5.55	178256	6.76	158768	10.06
VX0311WBL01	69956	5.54	139206	6.76	126312	10.06
VX0311WBS01	94786	5.54	172566	6.76	151770	10.06
VX0311WBSD01	92082	5.54	166231	6.76	148942	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.