

Surrogate Summary

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1502-02	PT-VOA-WP	1,2-Dichloroethane-d4	30	32.7	109	91	110
		Toluene-d8	30	30.2	101	91	112
		4-Bromofluorobenzene	30	29.9	100	63	112
VX0319WBL01	VX0319WBL01	1,2-Dichloroethane-d4	30	31.2	104	91	110
		Toluene-d8	30	29.5	98	91	112
		4-Bromofluorobenzene	30	28.4	95	63	112
VX0319WBS01	VX0319WBS01	1,2-Dichloroethane-d4	30	32.0	107	91	110
		Toluene-d8	30	31.1	104	91	112
		4-Bromofluorobenzene	30	31.1	104	63	112
VX0319WBSD0	VX0319WBSD01	1,2-Dichloroethane-d4	30	31.9	106	91	110
		Toluene-d8	30	29.6	99	91	112
		4-Bromofluorobenzene	30	30.5	102	63	112