

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ALLI03
 Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502
 Lab File ID: VX045340.D Date Analyzed: 03/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:41
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	16449	4.90	86152	6.76	80895	10.05
UPPER LIMIT	32898	5.397	172304	7.257	161790	10.549
LOWER LIMIT	8224.5	4.397	43076	6.257	40447.5	9.549
EPA SAMPLE NO.						
PT-VOA-WP	15107	4.89	80457	6.75	73631	10.05
VX0319WBL01	17082	4.89	85703	6.76	77105	10.05
VX0319WBS01	18710	4.89	97348	6.76	87335	10.05
VX0319WBSD01	15405	4.89	77122	6.76	74535	10.05

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.