

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q1503

Contract: ALLI03

Lab Code: CHEM

Case No.: Q1503

SAS No.: Q1503

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167449BS Mercury	ug/L	4.0	4.08		102	85 - 115	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q1503

Contract: ALLI03

Lab Code: CHEM

Case No.: Q1503

SAS No.: Q1503

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167491BS							
Aluminum	ug/L	2000	1810		90	85 - 115	P
Antimony	ug/L	800	777		97	85 - 115	P
Arsenic	ug/L	800	749		94	85 - 115	P
Barium	ug/L	200	173		86	85 - 115	P
Beryllium	ug/L	200	187		94	85 - 115	P
Boron	ug/L	300	315		105	85 - 115	P
Cadmium	ug/L	200	182		91	85 - 115	P
Calcium	ug/L	1000	913	J	91	85 - 115	P
Chromium	ug/L	400	380		95	85 - 115	P
Cobalt	ug/L	200	186		93	85 - 115	P
Copper	ug/L	300	297		99	85 - 115	P
Iron	ug/L	3000	2750		92	85 - 115	P
Lead	ug/L	1000	916		92	85 - 115	P
Magnesium	ug/L	2000	1760	J	88	85 - 115	P
Manganese	ug/L	200	184		92	85 - 115	P
Molybdenum	ug/L	400	384		96	85 - 115	P
Nickel	ug/L	500	468		94	85 - 115	P
Potassium	ug/L	10000	8810		88	85 - 115	P
Selenium	ug/L	2000	1920		96	85 - 115	P
Silver	ug/L	75.0	71.6		96	85 - 115	P
Sodium	ug/L	3000	3000		100	85 - 115	P
Thallium	ug/L	2000	1870		94	85 - 115	P
Tin	ug/L	700	652		93	85 - 115	P
Titanium	ug/L	200	179		90	85 - 115	P
Vanadium	ug/L	300	270		90	85 - 115	P
Zinc	ug/L	200	191		96	85 - 115	P