

LB135287


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

29980



A= 0.0000e+000
 B= 3.3658e-004
 C= 3.5606e-002
 Rho= 0.9994423
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.027	0.027	-27	0.000	-27	0				7
0.2	0.200	0.214	0.014	531	0.0 %	531					7
2.5	2.500	2.633	0.133	7718	0.0 %	7718					5
5.0	5.000	4.775	-0.225	14080	0.0 %	14080					-4
7.5	7.500	7.425	-0.075	21953	0.0 %	21953					-1
10.0	10.000	10.126	0.126	29980	0.0 %	29980					1

LB135287 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Apr 2025 13:51:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SB	-27	-	0.0000	7470A	PPB	03 Apr 2025 13:57:54	S	Std
0.2 - 1	SB-2	531	-	0.2000	7470A	PPB	03 Apr 2025 14:00:10	S	Std
2.5 - 1	SB-5	7718	-	2.5000	7470A	PPB	03 Apr 2025 14:02:27	S	Std
5.0 - 1	SB	14080	-	5.0000	7470A	PPB	03 Apr 2025 14:04:44	S	Std
7.5 - 1	SB-5	21953	-	7.5000	7470A	PPB	03 Apr 2025 14:07:02	S	Std
10.0 - 1	SB	29980	-	10.0000	7470A	PPB	03 Apr 2025 14:12:03	S	Std
ICV75 - 1	ICV75	12618	4.2826	-	7470A	PPB	03 Apr 2025 14:14:53	U	SMPL
ICB75 - 1	ICB75	-239	-0.0448	-	7470A	PPB	03 Apr 2025 14:17:09	U	SMPL
CCV28 - 1	CCV28	13837	4.6929	-	7470A	PPB	03 Apr 2025 14:19:26	U	SMPL
CCB28 - 1	CCB28	-188	-0.0277	-	7470A	PPB	03 Apr 2025 14:21:41	U	SMPL
CRA - 1	CRA	407	0.1726	-	7470A	PPB	03 Apr 2025 14:24:00	U	SMPL
HighStd - 1	HighStd	30171	10.1907	-	7470A	PPB	03 Apr 2025 14:26:15	U	SMPL
ChkStd - 1	ChkStd	19812	6.7040	-	7470A	PPB	03 Apr 2025 14:28:31	U	SMPL
PB167452BL - 1	PBW	-407	-0.1014	-	7470A	PPB	03 Apr 2025 14:30:49	U	SMPL
PB167452BS - 1	LCSW	11965	4.0628	-	7470A	PPB	03 Apr 2025 14:33:08	U	SMPL
Q1504-07 - 1	PT-HG-WP	51229	17.2785	-	7470A	PPB	03 Apr 2025 14:42:32	U	SMPL
Q1504-07DLX5 - 1	PT-HG-WP	10199	3.4684	-	7470A	PPB	03 Apr 2025 14:52:02	U	SMPL
CCV29 - 1	CCV29	16174	5.4795	-	7470A	PPB	03 Apr 2025 15:12:15	U	SMPL
CCB29 - 1	CCB29	-151	-0.0152	-	7470A	PPB	03 Apr 2025 15:14:30	U	SMPL