



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1504
Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1504 **SAS No.:** Q1504
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Aluminum	22.9	20.0	114	70 - 130	P	04/01/2025	13:33	LB135261
	Antimony	2.12	2.0	106	70 - 130	P	04/01/2025	13:33	LB135261
	Arsenic	1.05	1.0	105	70 - 130	P	04/01/2025	13:33	LB135261
	Barium	10.5	10.0	105	70 - 130	P	04/01/2025	13:33	LB135261
	Beryllium	1.14	1.0	114	70 - 130	P	04/01/2025	13:33	LB135261
	Cadmium	1.06	1.0	106	70 - 130	P	04/01/2025	13:33	LB135261
	Calcium	545	500	109	70 - 130	P	04/01/2025	13:33	LB135261
	Chromium	2.08	2.0	104	70 - 130	P	04/01/2025	13:33	LB135261
	Cobalt	1.11	1.0	111	50 - 150	P	04/01/2025	13:33	LB135261
	Copper	1.99	2.0	100	70 - 130	P	04/01/2025	13:33	LB135261
	Iron	55.1	50.0	110	70 - 130	P	04/01/2025	13:33	LB135261
	Lead	1.02	1.0	102	70 - 130	P	04/01/2025	13:33	LB135261
	Magnesium	562	500	112	70 - 130	P	04/01/2025	13:33	LB135261
	Manganese	1.11	1.0	111	50 - 150	P	04/01/2025	13:33	LB135261
	Molybdenum	5.22	5.0	104	70 - 130	P	04/01/2025	13:33	LB135261
	Nickel	0.98	1.0	98	70 - 130	P	04/01/2025	13:33	LB135261
	Potassium	542	500	108	70 - 130	P	04/01/2025	13:33	LB135261
	Selenium	5.43	5.0	109	70 - 130	P	04/01/2025	13:33	LB135261
	Silver	1.06	1.0	106	70 - 130	P	04/01/2025	13:33	LB135261
	Sodium	592	500	118	70 - 130	P	04/01/2025	13:33	LB135261
	Strontium	1.08	1.0	108	70 - 130	P	04/01/2025	13:33	LB135261
	Thallium	1.01	1.0	101	70 - 130	P	04/01/2025	13:33	LB135261
	Tin	5.46	5.0	109	70 - 130	P	04/01/2025	13:33	LB135261
	Titanium	5.44	5.0	109	70 - 130	P	04/01/2025	13:33	LB135261
	Vanadium	5.31	5.0	106	70 - 130	P	04/01/2025	13:33	LB135261
	Zinc	4.77	5.0	95	50 - 150	P	04/01/2025	13:33	LB135261