

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q1504

Contract: ALLI03

Lab Code: CHEM

Case No.: Q1504

SAS No.: Q1504

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167399BS							
Aluminum	ug/L	10000	10300		103	80 - 120	P
Antimony	ug/L	500	505		101	80 - 120	P
Arsenic	ug/L	500	487		97	80 - 120	P
Barium	ug/L	2500	2540		102	80 - 120	P
Beryllium	ug/L	500	518		104	80 - 120	P
Cadmium	ug/L	500	518		104	80 - 120	P
Calcium	ug/L	50000	51000		102	80 - 120	P
Chromium	ug/L	500	510		102	80 - 120	P
Cobalt	ug/L	500	507		101	80 - 120	P
Copper	ug/L	5000	5060		101	80 - 120	P
Iron	ug/L	25000	26100		104	80 - 120	P
Lead	ug/L	2500	2500		100	80 - 120	P
Magnesium	ug/L	50000	52800		106	80 - 120	P
Manganese	ug/L	5000	5160		103	80 - 120	P
Nickel	ug/L	500	516		103	80 - 120	P
Potassium	ug/L	25000	25700		103	80 - 120	P
Tin	ug/L	500	506		101	80 - 120	P
Molybdenum	ug/L	5000	4960		99	80 - 120	P
Selenium	ug/L	500	491		98	80 - 120	P
Silver	ug/L	500	517		103	80 - 120	P
Sodium	ug/L	50000	53300		107	80 - 120	P
Thallium	ug/L	500	498		100	80 - 120	P
Vanadium	ug/L	500	512		102	80 - 120	P
Zinc	ug/L	5000	5090		102	80 - 120	P
Strontium	ug/L	500	503		101	80 - 120	P
Titanium	ug/L	5000	4990		100	80 - 120	P