


Instrument\		IC-1	Analyst :		NF	Method: 0.0 / 9056A		date time		alwt/Fina	Analyst
Ident	Con F-	Con CL-	Con NO2	Con BR-	Con NO3	Con HPO4	Con SO4	Method name			
STD1	0	0	0	0	0	0	0	IC1-022125	2/21/2025 11:05	10	NF/IZ
STD2	0.448	0.678	0.657	2.152	0.544	1.059	3.319	IC1-022125	2/21/2025 11:26	10	NF/IZ
STD3	0.821	1.223	1.216	4.052	1.014	2.009	6.104	IC1-022125	2/21/2025 11:48	10	NF/IZ
STD4	1.013	1.524	1.523	5.08	1.272	2.514	7.612	IC1-022125	2/21/2025 12:09	10	NF/IZ
STD5	1.919	2.879	2.894	9.683	2.411	4.896	14.423	IC1-022125	2/21/2025 12:31	10	NF/IZ
STD6	3.884	5.815	5.898	19.712	4.921	9.947	29.333	IC1-022125	2/21/2025 12:52	10	NF/IZ
STD7	5.115	7.682	7.613	25.321	6.339	12.575	37.71	IC1-022125	2/21/2025 13:13	10	NF/IZ
ICV	2.008	2.989	3.015	10.228	2.532	5.141	15.146	IC1-022125	2/21/2025 13:35	10	NF/IZ
ICB	0	0	0	0	0	0	0	IC1-022125	2/21/2025 13:56	10	NF/IZ
CCV	2.051	3.096	3.083	10.45	2.611	4.63	15.543	IC1-022125	3/11/2025 10:41	10	NF/IZ
CCB	0	0	0	0	0	0	0	IC1-022125	3/11/2025 11:02	10	NF/IZ
LB135004BLW	0	0	0	0	0	0	0	IC1-022125	3/11/2025 11:24	10	NF/IZ
LB135004BSW	2.055	3.102	3.11	10.478	2.617	4.736	15.49	IC1-022125	3/11/2025 11:45	10	NF/IZ
Q1505-05	2.953	279.3	0	9.129	0	0	97.641	IC1-022125	3/11/2025 12:10	10	NF/IZ
Q1505-11	0	8.158	0	0	21.249	4.821	0	IC1-022125	3/11/2025 12:32	10	NF/IZ
Q1505-15	0	0.186	2.238	0	0	0	0	IC1-022125	3/11/2025 12:53	10	NF/IZ
Q1505-05DLX10	0.355	20.499	0	1.13	0	0	9.034	IC1-022125	3/11/2025 13:15	10	NF/IZ
Q1505-05DL2X50	0.134	3.852	0	0.426	0	0	2.264	IC1-022125	3/11/2025 13:36	10	NF/IZ
Q1505-11DLX10	0	0.861	0	0	1.869	0.543	0	IC1-022125	3/11/2025 13:58	10	NF/IZ
CCV	2.077	3.108	3.1	10.505	2.622	5.312	15.508	IC1-022125	3/11/2025 14:19	10	NF/IZ
CCB	0	0	0	0	0	0	0	IC1-022125	3/11/2025 14:41	10	NF/IZ

Clear table

Instrument ID: IC-2											
		Analyst :		IZ		Method: 300.0 / 9056A					
ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time	Initial wt/ Final	Analyst
STD1	0	0	0	0	0	0	0	IC1-022125	2/21/2025 11:05	10	NF/IZ
STD2	0.448	0.678	0.657	2.152	0.544	1.059	3.319	IC1-022125	2/21/2025 11:26	10	NF/IZ
STD3	0.821	1.223	1.216	4.052	1.014	2.009	6.104	IC1-022125	2/21/2025 11:48	10	NF/IZ
STD4	1.013	1.524	1.523	5.08	1.272	2.514	7.612	IC1-022125	2/21/2025 12:09	10	NF/IZ
STD5	1.919	2.879	2.894	9.683	2.411	4.896	14.423	IC1-022125	2/21/2025 12:31	10	NF/IZ
STD6	3.884	5.815	5.898	19.712	4.921	9.947	29.333	IC1-022125	2/21/2025 12:52	10	NF/IZ
STD7	5.115	7.682	7.613	25.321	6.339	12.575	37.71	IC1-022125	2/21/2025 13:13	10	NF/IZ
ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4				
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000				
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000				
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000				
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000				
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000				
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000				
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000				
ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4				
STD1											
STD2	12.0000	13.0000	9.5000	7.6000	8.8000	5.9000	10.6333				
STD3	2.6250	1.9167	1.3333	1.3000	1.4000	0.4500	1.7333				
STD4	1.3000	1.6000	1.5333	1.6000	1.7600	0.5600	1.4933				
STD5	-4.0500	-4.0333	-3.5333	-3.1700	-3.5600	-2.0800	-3.8467				
STD6	-2.9000	-3.0833	-1.7000	-1.4400	-1.5800	-0.5300	-2.2233				
STD7	2.3000	2.4267	1.5067	1.2840	1.4240	0.6000	1.9189				


Anions



Peak number	Retention time min	Area ($\mu\text{S/cm}$) x min	Height $\mu\text{S/cm}$	Concentration ppm	Component name
1	3.400	0.7457	5.535	5.115	Fluoride
2	4.197	0.0046	0.012	invalid	
3	7.070	0.7592	4.602	7.682	Chloride
4	9.330	1.6810	8.120	7.613	Nitrite
5	11.483	1.0784	4.588	25.321	Bromide
6	13.222	1.5233	5.678	6.339	Nitrate
7	15.325	1.0441	2.611	12.575	Phosphate
8	16.760	2.7982	7.416	37.710	Sulfate

Fluoride (Anions)

($\mu\text{S}/\text{cm}$) x min



Function: $A = -0.0112463 + 0.0147980 \times Q$


Relative standard deviation 4.464159 %

Correlation coefficient 0.998989

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-02-21 11:05:25 UTC-5	used
Standard 2	1	0.400	10.0	1.0	1.0	0.055	STD2	2025-02-21 11:26:47 UTC-5	used
Standard 3	1	0.800	10.0	1.0	1.0	0.110	STD3	2025-02-21 11:48:11 UTC-5	used
Standard 4	1	1.000	10.0	1.0	1.0	0.139	STD4	2025-02-21 12:09:35 UTC-5	used
Standard 5	1	2.000	10.0	1.0	1.0	0.273	STD5	2025-02-21 12:31:01 UTC-5	used
Standard 6	1	4.000	10.0	1.0	1.0	0.564	STD6	2025-02-21 12:52:27 UTC-5	used
Standard 7	1	5.000	10.0	1.0	1.0	0.746	STD7	2025-02-21 13:13:53 UTC-5	used

Chloride (Anions)

($\mu\text{S}/\text{cm}$) x min




Function: $A = -0.0105167 + 0.0100207 \times Q$

Relative standard deviation 4.669067 %

Correlation coefficient 0.998888

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-02-21 11:05:25 UTC-5	used
Standard 2	1	0.600	10.0	1.0	1.0	0.057	STD2	2025-02-21 11:26:47 UTC-5	used
Standard 3	1	1.200	10.0	1.0	1.0	0.112	STD3	2025-02-21 11:48:11 UTC-5	used
Standard 4	1	1.500	10.0	1.0	1.0	0.142	STD4	2025-02-21 12:09:35 UTC-5	used
Standard 5	1	3.000	10.0	1.0	1.0	0.278	STD5	2025-02-21 12:31:01 UTC-5	used
Standard 6	1	6.000	10.0	1.0	1.0	0.572	STD6	2025-02-21 12:52:27 UTC-5	used
Standard 7	1	7.500	10.0	1.0	1.0	0.759	STD7	2025-02-21 13:13:53 UTC-5	used

Nitrite (Anions) $(\mu\text{S}/\text{cm}) \times \text{min}$ Function: $A = -0.0232250 + 0.0223870 \times Q$


Relative standard deviation 3.068725 %

Correlation coefficient 0.999519

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-02-21 11:05:25 UTC-5	used
Standard 2	1	0.600	10.0	1.0	1.0	0.124	STD2	2025-02-21 11:26:47 UTC-5	used
Standard 3	1	1.200	10.0	1.0	1.0	0.249	STD3	2025-02-21 11:48:11 UTC-5	used
Standard 4	1	1.500	10.0	1.0	1.0	0.318	STD4	2025-02-21 12:09:35 UTC-5	used
Standard 5	1	3.000	10.0	1.0	1.0	0.625	STD5	2025-02-21 12:31:01 UTC-5	used
Standard 6	1	6.000	10.0	1.0	1.0	1.297	STD6	2025-02-21 12:52:27 UTC-5	used
Standard 7	1	7.500	10.0	1.0	1.0	1.681	STD7	2025-02-21 13:13:53 UTC-5	used

Bromide (Anions)

($\mu\text{S}/\text{cm}$) $\times$ min



Function: $A = -0.0113202 + 4.30364E-3 \times Q$


Relative standard deviation: 2.628989 %

Correlation coefficient: 0.999641

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-02-21 11:05:25 UTC-5	used
Standard 2	1	2.000	10.0	1.0	1.0	0.081	STD2	2025-02-21 11:26:47 UTC-5	used
Standard 3	1	4.000	10.0	1.0	1.0	0.163	STD3	2025-02-21 11:48:11 UTC-5	used
Standard 4	1	5.000	10.0	1.0	1.0	0.207	STD4	2025-02-21 12:09:35 UTC-5	used
Standard 5	1	10.000	10.0	1.0	1.0	0.405	STD5	2025-02-21 12:31:01 UTC-5	used
Standard 6	1	20.000	10.0	1.0	1.0	0.837	STD6	2025-02-21 12:52:27 UTC-5	used
Standard 7	1	25.000	10.0	1.0	1.0	1.078	STD7	2025-02-21 13:13:53 UTC-5	used

Nitrate (Anions)

($\mu\text{S}/\text{cm}$) $\times$ min




Function: $A = -0.0190446 + 0.0243327 \times Q$

Relative standard deviation: 2.936394 %

Correlation coefficient 0.999557

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-02-21 11:05:25 UTC-5	used
Standard 2	1	0.500	10.0	1.0	1.0	0.113	STD2	2025-02-21 11:26:47 UTC-5	used
Standard 3	1	1.000	10.0	1.0	1.0	0.228	STD3	2025-02-21 11:48:11 UTC-5	used
Standard 4	1	1.250	10.0	1.0	1.0	0.290	STD4	2025-02-21 12:09:35 UTC-5	used
Standard 5	1	2.500	10.0	1.0	1.0	0.568	STD5	2025-02-21 12:31:01 UTC-5	used
Standard 6	1	5.000	10.0	1.0	1.0	1.178	STD6	2025-02-21 12:52:27 UTC-5	used
Standard 7	1	6.250	10.0	1.0	1.0	1.523	STD7	2025-02-21 13:13:53 UTC-5	used


Phosphate (Anions) $(\mu\text{S}/\text{cm}) \times \text{min}$ Function: $A = -6.44326E-3 + 8.35469E-3 \times Q$

Relative standard deviation 1.393773 %

Correlation coefficient 0.999897

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-02-21 11:05:25 UTC-5	used
Standard 2	1	1.000	10.0	1.0	1.0	0.082	STD2	2025-02-21 11:26:47 UTC-5	used
Standard 3	1	2.000	10.0	1.0	1.0	0.161	STD3	2025-02-21 11:48:11 UTC-5	used
Standard 4	1	2.500	10.0	1.0	1.0	0.204	STD4	2025-02-21 12:09:35 UTC-5	used
Standard 5	1	5.000	10.0	1.0	1.0	0.403	STD5	2025-02-21 12:31:01 UTC-5	used
Standard 6	1	10.000	10.0	1.0	1.0	0.825	STD6	2025-02-21 12:52:27 UTC-5	used
Standard 7	1	12.500	10.0	1.0	1.0	1.044	STD7	2025-02-21 13:13:53 UTC-5	used

Sulfate (Anions)

 $(\mu\text{S}/\text{cm}) \times \text{min}$ Function: $A = -0.0430107 + 7.53437E-3 \times Q$

Relative standard deviation 3.742757 %

Correlation coefficient 0.999282

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-02-21 11:05:25 UTC-5	used
Standard 2	1	3.000	10.0	1.0	1.0	0.207	STD2	2025-02-21 11:26:47 UTC-5	used
Standard 3	1	6.000	10.0	1.0	1.0	0.417	STD3	2025-02-21 11:48:11 UTC-5	used
Standard 4	1	7.500	10.0	1.0	1.0	0.531	STD4	2025-02-21 12:09:35 UTC-5	used
Standard 5	1	15.000	10.0	1.0	1.0	1.044	STD5	2025-02-21 12:31:01 UTC-5	used
Standard 6	1	30.000	10.0	1.0	1.0	2.167	STD6	2025-02-21 12:52:27 UTC-5	used
Standard 7	1	37.000	10.0	1.0	1.0	2.798	STD7	2025-02-21 13:13:53 UTC-5	used


Sample data

Ident STD1
Sample type Standard 1
Determination start 2025-02-21 11:05:25 UTC-5
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.05 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time	Area	Height	Concentration	Component name
	min	($\mu\text{S/cm}$) x min	$\mu\text{S/cm}$	ppm	
1	1.167	0.0607	0.309	invalid	


Sample data

Ident STD2
Sample type Standard 2
Determination start 2025-02-21 11:26:47 UTC-5
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.11 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.390	0.0550	0.405	0.448	Fluoride
2	6.723	0.0574	0.344	0.678	Chloride
3	8.940	0.1238	0.587	0.657	Nitrite
4	11.110	0.0813	0.336	2.152	Bromide
5	12.852	0.1132	0.410	0.544	Nitrate
6	14.923	0.0820	0.192	1.059	Phosphate
7	16.323	0.2071	0.531	3.319	Sulfate


Sample data

Ident STD3
Sample type Standard 3
Determination start 2025-02-21 11:48:11 UTC-5
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.94 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.390	0.1102	0.822	0.821	Fluoride
2	6.712	0.1120	0.683	1.223	Chloride
3	8.918	0.2491	1.193	1.216	Nitrite
4	11.075	0.1631	0.680	4.052	Bromide
5	12.813	0.2277	0.831	1.014	Nitrate
6	14.912	0.1614	0.387	2.009	Phosphate
7	16.345	0.4169	1.076	6.104	Sulfate


Sample data

Ident STD4
Sample type Standard 4
Determination start 2025-02-21 12:09:35 UTC-5
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.94 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.395	0.1386	1.042	1.013	Fluoride
2	6.707	0.1422	0.872	1.524	Chloride
3	8.907	0.3178	1.529	1.523	Nitrite
4	11.052	0.2073	0.870	5.080	Bromide
5	12.788	0.2904	1.064	1.272	Nitrate
6	14.908	0.2036	0.492	2.514	Phosphate
7	16.363	0.5305	1.376	7.612	Sulfate

Sample data

Ident STD5
 Sample type Standard 5
 Determination start 2025-02-21 12:31:01 UTC-5
 Method IC1-022125
 Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
 Channel Conductivity
 Recording time 19.0 min
 Integration Automatically
 Column type Metrosep A Supp 19 - 150/4.0
 Eluent composition not defined
 Flow 0.700 mL/min
 Maximum flow monitored yes
 Pressure 11.88 MPa
 Maximum pressure monitored yes
 Temperature ---- °C

Anions

Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.393	0.2728	2.069	1.919	Fluoride
2	6.708	0.2780	1.702	2.879	Chloride
3	8.917	0.6246	3.000	2.894	Nitrite
4	11.068	0.4054	1.702	9.683	Bromide
5	12.807	0.5675	2.085	2.411	Nitrate
6	14.907	0.4026	0.982	4.896	Phosphate
7	16.342	1.0436	2.711	14.423	Sulfate


Sample data

Ident STD6
Sample type Standard 6
Determination start 2025-02-21 12:52:27 UTC-5
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.94 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.392	0.5636	4.320	3.884	Fluoride
2	6.710	0.5722	3.536	5.815	Chloride
3	8.920	1.2971	6.268	5.898	Nitrite
4	11.065	0.8370	3.548	19.712	Bromide
5	12.797	1.1784	4.381	4.921	Nitrate
6	14.898	0.8246	2.050	9.947	Phosphate
7	16.333	2.1670	5.703	29.333	Sulfate


Sample data

Ident STD7
Sample type Standard 7
Determination start 2025-02-21 13:13:53 UTC-5
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.88 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area ($\mu\text{S/cm}$) x min	Height $\mu\text{S/cm}$	Concentration ppm	Component name
1	3.400	0.7457	5.535	5.115	Fluoride
2	4.197	0.0046	0.012	invalid	
3	7.070	0.7592	4.602	7.682	Chloride
4	9.330	1.6810	8.120	7.613	Nitrite
5	11.483	1.0784	4.588	25.321	Bromide
6	13.222	1.5233	5.678	6.339	Nitrate
7	15.325	1.0441	2.611	12.575	Phosphate
8	16.760	2.7982	7.416	37.710	Sulfate


Sample data

Ident ICV
Sample type Check standard 1
Determination start 2025-02-21 13:35:21 UTC-5
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.94 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.392	0.2859	2.174	2.008	Fluoride
2	6.703	0.2890	1.769	2.989	Chloride
3	8.907	0.6518	3.137	3.015	Nitrite
4	11.053	0.4289	1.799	10.228	Bromide
5	12.787	0.5971	2.200	2.532	Nitrate
6	14.897	0.4230	1.038	5.141	Phosphate
7	16.342	1.0982	2.855	15.146	Sulfate

Sample data

Ident ICB
Sample type Sample
Determination start 2025-02-21 13:56:49 UTC-5
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.88 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident CCV
Sample type Check standard 1
Determination start 2025-03-11 10:41:14 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.37 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.420	0.2923	2.246	2.051	Fluoride
2	6.673	0.2997	1.872	3.096	Chloride
3	8.822	0.6669	3.283	3.083	Nitrite
4	10.888	0.4384	1.890	10.450	Bromide
5	12.603	0.6162	2.328	2.611	Nitrate
6	14.738	0.3804	0.891	4.630	Phosphate
7	16.743	1.1281	2.937	15.543	Sulfate

Sample data

Ident CCB
Sample type Sample
Determination start 2025-03-11 11:02:44 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.32 MPa
Maximum pressure monitored yes
Temperature ---- °C


Anions

Sample data

Ident LB135004BLW
Sample type Sample
Determination start 2025-03-11 11:24:14 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.49 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident LB135004BSW
Sample type Check standard 1
Determination start 2025-03-11 11:45:46 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.32 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.417	0.2928	2.264	2.055	Fluoride
2	6.658	0.3003	1.883	3.102	Chloride
3	8.797	0.6730	3.325	3.110	Nitrite
4	10.850	0.4396	1.904	10.478	Bromide
5	12.560	0.6178	2.344	2.617	Nitrate
6	14.725	0.3893	0.952	4.736	Phosphate
7	16.755	1.1241	2.954	15.490	Sulfate


Sample data

Ident Q1505-05
Sample type Sample
Determination start 2025-03-11 12:10:32 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.15 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.408	0.4258	3.278	2.953	Fluoride
2	6.737	27.9770	158.699	279.297	Chloride
3	10.823	0.3815	1.659	9.129	Bromide
4	16.720	7.3136	19.897	97.641	Sulfate


Sample data

Ident Q1505-11
Sample type Sample
Determination start 2025-03-11 12:32:05 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.26 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	6.650	0.8069	5.132	8.158	Chloride
2	12.513	5.1514	20.463	21.249	Nitrate
3	14.733	0.3963	1.010	4.821	Phosphate


Sample data

Ident Q1505-15
Sample type Sample
Determination start 2025-03-11 12:53:39 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.15 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	4.058	0.0057	0.023	invalid	
2	4.770	0.0027	0.016	invalid	
3	6.648	0.0081	0.050	0.186	Chloride
4	8.777	0.4778	2.362	2.238	Nitrite


Sample data

Ident Q1505-05DLX10
Sample type Sample
Determination start 2025-03-11 13:15:13 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.54 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.407	0.0413	0.313	0.355	Fluoride
2	6.642	2.0437	13.309	20.499	Chloride
3	7.420	0.0040	0.011	invalid	
4	10.802	0.0373	0.161	1.130	Bromide
5	16.738	0.6376	1.661	9.034	Sulfate


Sample data

Ident Q1505-05DL2X50
Sample type Sample
Determination start 2025-03-11 13:36:48 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.32 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.412	0.0085	0.062	0.134	Fluoride
2	6.648	0.3755	2.364	3.852	Chloride
3	10.827	0.0070	0.031	0.426	Bromide
4	16.758	0.1276	0.330	2.264	Sulfate


Sample data

Ident Q1505-11DLX10
Sample type Sample
Determination start 2025-03-11 13:58:22 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.71 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area ($\mu\text{S/cm}$) x min	Height $\mu\text{S/cm}$	Concentration ppm	Component name
1	6.637	0.0757	0.474	0.861	Chloride
2	12.510	0.4357	1.657	1.869	Nitrate
3	14.705	0.0389	0.091	0.543	Phosphate


Sample data

Ident CCV
Sample type Check standard 1
Determination start 2025-03-11 14:19:55 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.37 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.415	0.2961	2.301	2.077	Fluoride
2	6.643	0.3009	1.893	3.108	Chloride
3	8.770	0.6708	3.331	3.100	Nitrite
4	10.810	0.4408	1.915	10.505	Bromide
5	12.517	0.6190	2.360	2.622	Nitrate
6	14.718	0.4374	1.118	5.312	Phosphate
7	16.775	1.1254	2.957	15.508	Sulfate

Sample data

Ident CCB
Sample type Sample
Determination start 2025-03-11 14:41:25 UTC-4
Method IC1-022125
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.43 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions

WORKLIST(Hardcopy Internal Chain)

LB135004

WorkList Name : ANIONS PT-03112025

WorkList ID : 188194

Department : Wet-Chemistry

Date : 03-11-2025 10:28:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-05	PT-MIN1-WP	Water	Anions Group1	Cool 4 deg C	ALLI03	QA Of	03/03/2025	300.0
Q1505-11	PT-NUT1-WP	Water	Anions Group2	Cool 4 deg C	ALLI03	QA Of	03/03/2025	300.0
Q1505-15	PT-NUT3-WP	Water	Anions Group4	Cool 4 deg C	ALLI03	QA Of	03/03/2025	300.0

Date/Time 03-11-2025, 10:55

Raw Sample Received by: NIF(wc)

Raw Sample Relinquished by: SJ(QAo)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: