

Analytical Summary Report

Analysis Method: 9050A
Parameter: Conductance
Run Number: LB135023

ANALYST: Niha
SUPERVISOR REVIEW BY: Iwona
Meter ID: WC CONDUCTANCE METER-1
Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP111962

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB135023BL		1	22.5	0.83	0.87	03/13/2025	13:00
2	LB135023BS	1413	1	22.7	1347.00	1408.89	03/13/2025	13:05
3	Q1505-09		1	22.8	322.00	336.12	03/13/2025	13:10
4	Q1505-09DUP		1	22.7	320.00	334.70	03/13/2025	13:15

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

WORKLIST(Hardcopy Internal Chain)

LB135023

WorkList Name : CONDUCTANCE PT - 03

WorkList ID : 188260

Department : Wet-Chemistry

Date : 03-13-2025 11:37:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-09	PT-COND-WP	Water	Conductance	Cool 4 deg C	ALLI03	QA 01	03/03/2025	9050A

Date/Time 03-13-2025 11:45
Raw Sample Received by: NF (WC)
Raw Sample Relinquished by: SJ(QA0)

Date/Time NA
Raw Sample Received by: W
Raw Sample Relinquished by: 1