

Analytical Summary Report

Analysis Method: SM2510 B
Parameter: Conductance
Run Number: LB135057

ANALYST: Niha
SUPERVISOR REVIEW BY: Iwona
Meter ID: WC CONDUCTANCE METER-1
Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP111962

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB135057BL		1	22.5	0.83	0.87	03/13/2025	13:00
2	LB135057BS	1413	1	22.7	1347.00	1408.89	03/13/2025	13:05
3	Q1505-07		1	22.8	322.00	336.12	03/13/2025	13:10
4	Q1505-07DUP		1	22.7	320.00	334.70	03/13/2025	13:15

Note : Conductivity result is corrected to 25°C

Result = (Reading)
1 + 0.0191 (Temperature - 25)

WORKLIST(Hardcopy Internal Chain)

43 13505 #

WorkList Name : CONDUCTANCE PT-031

WorkList ID : 188258

Department : Wet-Chemistry

Date : 03-13-2025 11:37:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-07	PT-COND-WP	Water	Conductance	Cool 4 deg C	ALLI03	QA Of	03/03/2025	SM2510 B

Date/Time 03-13-2025 11:45
Raw Sample Received by: NF(wc)
Raw Sample Relinquished by: SJ(QAo)

Date/Time NA
Raw Sample Received by: y
Raw Sample Relinquished by: y