

Analytical Summary Report

Analysis Method: 120.1
Parameter: Conductance
Run Number: LB135058

ANALYST: Niha
SUPERVISOR REVIEW BY: Iwona
Meter ID: WC CONDUCTANCE METER-1
Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP111962

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB135058BL		1	22.5	0.83	0.87	03/13/2025	13:00
2	LB135058BS	1413	1	22.7	1347.00	1408.89	03/13/2025	13:05
3	Q1505-08		1	22.8	322.00	336.12	03/13/2025	13:10
4	Q1505-08DUP		1	22.7	320.00	334.70	03/13/2025	13:15

Note : Conductivity result is corrected to 25°C

Result = (Reading)
1 + 0.0191 (Temperature - 25)

WORKLIST(Hardcopy Internal Chain)

LB135058

WorkList Name : CONDUCTANCE PT -03

WorkList ID : 188259

Department : Wet-Chemistry

Date : 03-13-2025 11:37:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-08	PT-COND-WP	Water	Conductance	Cool 4 deg C	ALLI03	QA Of	03/03/2025	120.1

Date/Time 03.13.2025 11:45
Raw Sample Received by: NF(WC)
Raw Sample Relinquished by: SJ(QAO)

Date/Time
Raw Sample Received by: NA
Raw Sample Relinquished by: V