

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB135207

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP112522
Turbidity Calibration std, 20NTU	WP112519
Turbidity Calibration std, 40NTU	WP112518
Turbidity Calibration std, 1NTU	WP112521
Turbidity Calibration std, 5NTU	WP112520
10 NTU Standard 500 ml	W3116

Intercept: -0.0584

Slope: 0.9274

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.142	0.22		03/27/2025	10:36
2	CAL2	1	1	1.010	1.15	15.2	03/27/2025	10:40
3	CAL3	5	1	4.320	4.72	-5.6	03/27/2025	10:44
4	CAL4	10	1	9.260	10.05	0.5	03/27/2025	10:48
5	CAL5	20	1	18.200	19.69	-1.6	03/27/2025	10:52
6	CAL6	40	1	37.200	40.18	0.4	03/27/2025	10:56

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB135207

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	8.510	9.239	03/27/2025	11:00
2	ICB		1	0.144	0.218	03/27/2025	11:04
3	CCV1	10	1	8.800	9.552	03/27/2025	11:08
4	CCB1		1	0.142	0.216	03/27/2025	11:12
5	LB135207BL		1	0.140	0.214	03/27/2025	11:16
6	Q1505-34		1	20.400	22.060	03/27/2025	11:20
7	CCV2	10	1	8.750	9.498	03/27/2025	11:24
8	CCB2		1	0.148	0.223	03/27/2025	11:28

WORKLIST(Hardcopy Internal Chain)

LB135207

WorkList Name : TURBIDITY PT -0327202

WorkList ID : 188595

Department : Wet-Chemistry

Date : 03-27-2025 10:08:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-34	PT-TURB-WP	Water	Turbidity	Cool 4 deg C	ALL103	QA Of	03/03/2025	180.1

Date/Time 03.27.2025, 10:10
Raw Sample Received by: NF (wc)
Raw Sample Relinquished by: i2 (wc)

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

NA

"

"