

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1506
Client: GFE LLC
Analytical Method: SWTO-15 **Datafile :** VL042117.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0310ABS01	Vinyl Chloride	10	10.4	ppbv	104			70	130	
	1,1-Dichloroethene	10	9.40	ppbv	94			70	130	
	cis-1,2-Dichloroethene	10	7.90	ppbv	79			70	130	
	1,1,1-Trichloroethane	10	7.90	ppbv	79			70	130	
	Trichloroethene	10	8.40	ppbv	84			70	130	
	Tetrachloroethene	10	8.40	ppbv	84			70	130	