

Preparation Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1513

Project: Linden Generating Station - LRSA

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB134926BL COD	mg/L	< 5.0000	5.0000	U	2.35	10.0	03/10/2025
Sample ID: LB134930BL BOD5	mg/L	< 0.2000	0.2000	U	0.17	2.0	03/07/2025
Sample ID: LB134975BL TSS	mg/L	1	2.0000	J	1	4	03/10/2025