

Laboratory Control Sample Summary

Client:	Portal Partners Tri-Venture	SDG No.:	Q1514
Project:	Amtrak Sawtooth Bridges 2025	Run No.:	LB134933

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134933BS							
Hexavalent Chromium	mg/L	0.5	0.51		103	1	90-111	03/06/2025

Laboratory Control Sample Summary

Client:	Portal Partners Tri-Venture	SDG No.:	Q1514
Project:	Amtrak Sawtooth Bridges 2025	Run No.:	LB134940

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167023BS							
Hexavalent Chromium	mg/Kg	20	20.1		101	1	84-110	03/07/2025