


LB134978

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.031	-0.031	46	0.000	46					
0.2	0.200	0.172	-0.028	1155	0.0 %	1155					-14
2.5	2.500	2.411	-0.089	13351	0.0 %	13351					-4
5.0	5.000	5.264	0.264	28894	0.0 %	28894					5
7.5	7.500	7.512	0.012	41139	0.0 %	41139					0
10.0	10.000	9.872	-0.128	53996	0.0 %	53996					-1

LB134978 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Mar 2025 08:11:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7471B	PPB	11 Mar 2025 08:19:59	S	Std
0.2 - 1	50.2	1155	-	0.2000	7471B	PPB	11 Mar 2025 08:22:16	S	Std
2.5 - 1	52.5	13351	-	2.5000	7471B	PPB	11 Mar 2025 08:24:33	S	Std
5.0 - 1	55	28894	-	5.0000	7471B	PPB	11 Mar 2025 08:26:50	S	Std
7.5 - 1	57.5	41139	-	7.5000	7471B	PPB	11 Mar 2025 08:29:09	S	Std
10.0 - 1	50	53996	-	10.0000	7471B	PPB	11 Mar 2025 08:31:29	S	Std
ICV36 - 1	ICV36	21190	3.8500	-	7471B	PPB	11 Mar 2025 08:34:33	U	SMPL
ICB36 - 1	ICB36	3	-0.0393	-	7471B	PPB	11 Mar 2025 08:39:28	U	SMPL
CCV80 - 1	CCV80	28164	5.1302	-	7471B	PPB	11 Mar 2025 08:41:42	U	SMPL
CCB80 - 1	CCB80	-171	-0.0713	-	7471B	PPB	11 Mar 2025 08:43:57	U	SMPL
CRA - 1	CRA	1326	0.2035	-	7471B	PPB	11 Mar 2025 08:51:40	U	SMPL
HighStd - 1	HighStd	54240	9.9169	-	7471B	PPB	11 Mar 2025 08:53:55	U	SMPL
ChkStd - 1	ChkStd	36792	6.7140	-	7471B	PPB	11 Mar 2025 08:56:10	U	SMPL
PB167064BL - 1	PBS	182	-0.0065	-	7471B	PPB	11 Mar 2025 09:01:21	U	SMPL
PB167064BS - 1	LCSS	21037	3.8219	-	7471B	PPB	11 Mar 2025 09:05:41	U	SMPL
Q1507-01 - 1	50-MIDDLESEX-AVE	81037	14.8360	-	7471B	PPB	11 Mar 2025 09:07:57	U	SMPL
Q1508-01 - 1	RBR251372	3328	0.5710	-	7471B	PPB	11 Mar 2025 09:10:16	U	SMPL
Q1514-01 - 1	ENV-105-SB01	130445	23.9058	-	7471B	PPB	11 Mar 2025 09:12:57	U	SMPL
Q1514-03 - 1	ENV-105-SB02	28506	5.1930	-	7471B	PPB	11 Mar 2025 09:15:16	U	SMPL
Q1514-05 - 1	ENV-103-SB01	2950	0.5017	-	7471B	PPB	11 Mar 2025 09:18:00	U	SMPL
CCV81 - 1	CCV81	26963	4.9097	-	7471B	PPB	11 Mar 2025 09:20:21	U	SMPL
CCB81 - 1	CCB81	25	-0.0353	-	7471B	PPB	11 Mar 2025 09:22:37	U	SMPL
Q1515-01 - 1	AU-06-030625	2443	0.4086	-	7471B	PPB	11 Mar 2025 09:24:58	U	SMPL
Q1524-01 - 1	72-11930-343-COMP	136292	24.9792	-	7471B	PPB	11 Mar 2025 09:27:13	U	SMPL
Q1524-02 - 1	VNJ-241	8967	1.6062	-	7471B	PPB	11 Mar 2025 09:29:29	U	SMPL
Q1524-02DUP - 1	VNJ-241DUP	10610	1.9078	-	7471B	PPB	11 Mar 2025 09:32:10	U	SMPL
Q1524-02MS - 1	VNJ-241MS	32001	5.8345	-	7471B	PPB	11 Mar 2025 09:37:09	U	SMPL
Q1524-02MSD - 1	VNJ-241MSD	31222	5.6915	-	7471B	PPB	11 Mar 2025 09:39:25	U	SMPL
Q1524-02LX5 - 1		1909	0.3106	-	7471B	PPB	11 Mar 2025 09:41:48	U	SMPL
Q1524-02A - 1		31659	5.7718	-	7471B	PPB	11 Mar 2025 09:44:11	U	SMPL
Q1507-01DLX5 - 1		16091	2.9139	-	7471B	PPB	11 Mar 2025 09:46:28	U	SMPL
Q1514-01DLX5 - 1		25844	4.7043	-	7471B	PPB	11 Mar 2025 09:48:51	U	SMPL
CCV82 - 1	CCV82	28169	5.1311	-	7471B	PPB	11 Mar 2025 09:51:19	U	SMPL
CCB82 - 1	CCB82	-61	-0.0511	-	7471B	PPB	11 Mar 2025 09:53:53	U	SMPL
Q1524-01DLX5 - 1		26837	4.8866	-	7471B	PPB	11 Mar 2025 09:56:13	U	SMPL
CCV83 - 1	CCV83	27480	5.0046	-	7471B	PPB	11 Mar 2025 09:58:29	U	SMPL
CCB83 - 1	CCB83	-56	-0.0501	-	7471B	PPB	11 Mar 2025 10:01:00	U	SMPL