



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	Q1514				
Contract:	PORT06	Lab Code:	CHEM		Case No.:	Q1514		SAS No.:	Q1514	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB35	Mercury	0.20	+/-0.20	U	0.20	CV	03/10/2025	10:26	LB134958	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1514

Contract: PORT06

Lab Code: CHEM

Case No.: Q1514

SAS No.: Q1514

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB75	Mercury	0.20	+/-0.20	U	0.20	CV	03/10/2025	10:39	LB134958
CCB76	Mercury	0.20	+/-0.20	U	0.20	CV	03/10/2025	11:11	LB134958
CCB77	Mercury	0.20	+/-0.20	U	0.20	CV	03/10/2025	11:53	LB134958
CCB78	Mercury	0.20	+/-0.20	U	0.20	CV	03/10/2025	12:26	LB134958
CCB79	Mercury	0.20	+/-0.20	U	0.20	CV	03/10/2025	12:47	LB134958
CCB80	Mercury	0.20	+/-0.20	U	0.20	CV	03/10/2025	13:12	LB134958

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	Q1514			
Contract:	PORT06	Lab Code:	CHEM		Case No.:	Q1514	SAS No.:	Q1514	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB36	Mercury	0.20	+/-0.20	U	0.20	CV	03/11/2025	08:39	LB134978

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1514

Contract: PORT06

Lab Code: CHEM

Case No.: Q1514

SAS No.: Q1514

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB80	Mercury	0.20	+/-0.20	U	0.20	CV	03/11/2025	08:43	LB134978
CCB81	Mercury	0.20	+/-0.20	U	0.20	CV	03/11/2025	09:22	LB134978
CCB82	Mercury	0.20	+/-0.20	U	0.20	CV	03/11/2025	09:53	LB134978
CCB83	Mercury	0.20	+/-0.20	U	0.20	CV	03/11/2025	10:01	LB134978

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1514							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1514	SAS No.: Q1514						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/12/2025	11:01	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	11:01	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	11:01	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	11:01	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	11:01	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	11:01	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	11:01	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	11:01	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	11:01	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	11:01	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	11:01	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	11:01	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	11:01	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	11:01	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	11:01	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	11:01	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	11:01	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	11:01	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	11:01	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	11:01	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	11:01	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	11:01	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** Q1514
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1514 **SAS No.:** Q1514

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/12/2025	11:34	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	11:34	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	11:34	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	11:34	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	11:34	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	11:34	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	11:34	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	11:34	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	11:34	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	11:34	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	11:34	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	11:34	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	11:34	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	11:34	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	11:34	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	11:34	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	11:34	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	11:34	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	11:34	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	11:34	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	11:34	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	11:34	LB135011
CCB02	Aluminum	100	+/-100	U	100	P	03/12/2025	12:24	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	12:24	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	12:24	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	12:24	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	12:24	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	12:24	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	12:24	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	12:24	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	12:24	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	12:24	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	12:24	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	12:24	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	12:24	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	12:24	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	12:24	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	12:24	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	12:24	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1514

Contract: PORT06

Lab Code: CHEM

Case No.: Q1514

SAS No.: Q1514

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	12:24	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	12:24	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	12:24	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	12:24	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	12:24	LB135011
CCB03	Aluminum	100	+/-100	U	100	P	03/12/2025	13:14	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	13:14	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	13:14	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	13:14	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	13:14	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	13:14	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	13:14	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	13:14	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	13:14	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	13:14	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	13:14	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	13:14	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	13:14	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	13:14	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	13:14	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	13:14	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	13:14	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	13:14	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	13:14	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	13:14	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	13:14	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	13:14	LB135011
CCB04	Aluminum	100	+/-100	U	100	P	03/12/2025	14:28	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	14:28	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	14:28	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	14:28	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	14:28	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	14:28	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	14:28	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	14:28	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	14:28	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	14:28	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	14:28	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	14:28	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1514							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1514	SAS No.: Q1514						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	14:28	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	14:28	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	14:28	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	14:28	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	14:28	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	14:28	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	14:28	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	14:28	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	14:28	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	14:28	LB135011
CCB05	Aluminum	100	+/-100	U	100	P	03/12/2025	15:29	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	15:29	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	15:29	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	15:29	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	15:29	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	15:29	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	15:29	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	15:29	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	15:29	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	15:29	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	15:29	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	15:29	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	15:29	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	15:29	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	15:29	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	15:29	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	15:29	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	15:29	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	15:29	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	15:29	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	15:29	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	15:29	LB135011
CCB06	Aluminum	100	+/-100	U	100	P	03/12/2025	16:30	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	16:30	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	16:30	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	16:30	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	16:30	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	16:30	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	16:30	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1514

Contract: PORT06

Lab Code: CHEM

Case No.: Q1514

SAS No.: Q1514

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	16:30	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	16:30	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	16:30	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	16:30	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	16:30	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	16:30	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	16:30	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	16:30	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	16:30	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	16:30	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	16:30	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	16:30	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	16:30	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	16:30	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	16:30	LB135011
CCB07	Aluminum	100	+/-100	U	100	P	03/12/2025	17:30	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	17:30	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	17:30	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	17:30	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	17:30	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	17:30	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	17:30	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	17:30	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	17:30	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	17:30	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	17:30	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	17:30	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	17:30	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	17:30	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	17:30	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	17:30	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	17:30	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	17:30	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	17:30	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	17:30	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	17:30	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	17:30	LB135011
CCB08	Aluminum	100	+/-100	U	100	P	03/12/2025	18:27	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	18:27	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1514							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1514	SAS No.: Q1514						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	18:27	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	18:27	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	18:27	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	18:27	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	18:27	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	18:27	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	18:27	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	18:27	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	18:27	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	18:27	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	18:27	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	18:27	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	18:27	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	18:27	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	18:27	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	18:27	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	18:27	LB135011
	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	18:27	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	18:27	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	18:27	LB135011
CCB09	Aluminum	100	+/-100	U	100	P	03/12/2025	18:53	LB135011
	Antimony	50.0	+/-50.0	U	50.0	P	03/12/2025	18:53	LB135011
	Arsenic	20.0	+/-20.0	U	20.0	P	03/12/2025	18:53	LB135011
	Barium	100	+/-100	U	100	P	03/12/2025	18:53	LB135011
	Beryllium	6.00	+/-6.00	U	6.00	P	03/12/2025	18:53	LB135011
	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	18:53	LB135011
	Calcium	2000	+/-2000	U	2000	P	03/12/2025	18:53	LB135011
	Chromium	10.0	+/-10.0	U	10.0	P	03/12/2025	18:53	LB135011
	Cobalt	30.0	+/-30.0	U	30.0	P	03/12/2025	18:53	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	18:53	LB135011
	Iron	100	+/-100	U	100	P	03/12/2025	18:53	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	18:53	LB135011
	Magnesium	2000	+/-2000	U	2000	P	03/12/2025	18:53	LB135011
	Manganese	20.0	+/-20.0	U	20.0	P	03/12/2025	18:53	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	18:53	LB135011
	Potassium	2000	+/-2000	U	2000	P	03/12/2025	18:53	LB135011
	Selenium	20.0	+/-20.0	U	20.0	P	03/12/2025	18:53	LB135011
	Silver	10.0	+/-10.0	U	10.0	P	03/12/2025	18:53	LB135011
	Sodium	2000	+/-2000	U	2000	P	03/12/2025	18:53	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1514

Contract: PORT06

Lab Code: CHEM

Case No.: Q1514

SAS No.: Q1514

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	03/12/2025	18:53	LB135011
	Vanadium	40.0	+/-40.0	U	40.0	P	03/12/2025	18:53	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	18:53	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1514

Contract: PORT06

Lab Code: CHEM

Case No.: Q1514

SAS No.: Q1514

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/13/2025	12:17	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	12:17	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	12:17	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	12:17	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	12:17	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	12:17	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	12:17	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	12:17	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	12:17	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	12:17	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	12:17	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	12:17	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	12:17	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	12:17	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	12:17	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	12:17	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	12:17	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	12:17	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	12:17	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	12:17	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	12:17	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	12:17	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1514

Contract: PORT06

Lab Code: CHEM

Case No.: Q1514

SAS No.: Q1514

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/13/2025	12:58	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	12:58	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	12:58	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	12:58	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	12:58	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	12:58	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	12:58	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	12:58	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	12:58	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	12:58	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	12:58	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	12:58	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	12:58	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	12:58	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	12:58	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	12:58	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	12:58	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	12:58	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	12:58	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	12:58	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	12:58	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	12:58	LB135035
CCB02	Aluminum	100	+/-100	U	100	P	03/13/2025	13:54	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	13:54	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	13:54	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	13:54	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	13:54	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	13:54	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	13:54	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	13:54	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	13:54	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	13:54	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	13:54	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	13:54	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	13:54	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	13:54	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	13:54	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	13:54	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	13:54	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1514							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1514	SAS No.: Q1514						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	13:54	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	13:54	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	13:54	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	13:54	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	13:54	LB135035
CCB03	Aluminum	100	+/-100	U	100	P	03/13/2025	14:48	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	14:48	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	14:48	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	14:48	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	14:48	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	14:48	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	14:48	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	14:48	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	14:48	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	14:48	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	14:48	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	14:48	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	14:48	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	14:48	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	14:48	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	14:48	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	14:48	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	14:48	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	14:48	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	14:48	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	14:48	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	14:48	LB135035
CCB04	Aluminum	100	+/-100	U	100	P	03/13/2025	15:42	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	15:42	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	15:42	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	15:42	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	15:42	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	15:42	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	15:42	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	15:42	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	15:42	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	15:42	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	15:42	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	15:42	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1514							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1514	SAS No.: Q1514						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	15:42	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	15:42	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	15:42	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	15:42	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	15:42	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	15:42	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	15:42	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	15:42	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	15:42	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	15:42	LB135035
CCB05	Aluminum	100	+/-100	U	100	P	03/13/2025	16:32	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	16:32	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	16:32	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	16:32	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	16:32	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	16:32	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	16:32	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	16:32	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	16:32	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	16:32	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	16:32	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	16:32	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	16:32	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	16:32	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	16:32	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	16:32	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	16:32	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	16:32	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	16:32	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	16:32	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	16:32	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	16:32	LB135035
CCB06	Aluminum	100	+/-100	U	100	P	03/13/2025	17:25	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	17:25	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	17:25	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	17:25	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	17:25	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	17:25	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	17:25	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1514							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1514	SAS No.: Q1514						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	17:25	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	17:25	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	17:25	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	17:25	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	17:25	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	17:25	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	17:25	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	17:25	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	17:25	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	17:25	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	17:25	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	17:25	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	17:25	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	17:25	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	17:25	LB135035
CCB07	Aluminum	100	+/-100	U	100	P	03/13/2025	18:35	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	18:35	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	18:35	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	18:35	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	18:35	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	18:35	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	18:35	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	18:35	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	18:35	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	18:35	LB135035
	Iron	45.6	+/-100	J	100	P	03/13/2025	18:35	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	18:35	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	18:35	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	18:35	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	18:35	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	18:35	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	18:35	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	18:35	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	18:35	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	18:35	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	18:35	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	18:35	LB135035
CCB08	Aluminum	100	+/-100	U	100	P	03/13/2025	19:03	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	19:03	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1514							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1514	SAS No.: Q1514						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	19:03	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	19:03	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	19:03	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	19:03	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	19:03	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	19:03	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	19:03	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	19:03	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	19:03	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	19:03	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	19:03	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	19:03	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	19:03	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	19:03	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	19:03	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	19:03	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	19:03	LB135035
	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	19:03	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	19:03	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	19:03	LB135035
CCB09	Aluminum	100	+/-100	U	100	P	03/13/2025	19:21	LB135035
	Antimony	50.0	+/-50.0	U	50.0	P	03/13/2025	19:21	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	19:21	LB135035
	Barium	100	+/-100	U	100	P	03/13/2025	19:21	LB135035
	Beryllium	6.00	+/-6.00	U	6.00	P	03/13/2025	19:21	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	19:21	LB135035
	Calcium	2000	+/-2000	U	2000	P	03/13/2025	19:21	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	19:21	LB135035
	Cobalt	30.0	+/-30.0	U	30.0	P	03/13/2025	19:21	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	19:21	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	19:21	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	19:21	LB135035
	Magnesium	2000	+/-2000	U	2000	P	03/13/2025	19:21	LB135035
	Manganese	20.0	+/-20.0	U	20.0	P	03/13/2025	19:21	LB135035
	Nickel	40.0	+/-40.0	U	40.0	P	03/13/2025	19:21	LB135035
	Potassium	2000	+/-2000	U	2000	P	03/13/2025	19:21	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	19:21	LB135035
	Silver	10.0	+/-10.0	U	10.0	P	03/13/2025	19:21	LB135035
	Sodium	2000	+/-2000	U	2000	P	03/13/2025	19:21	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	Q1514			
Contract:	PORT06	Lab Code:	CHEM		Case No.:	Q1514	SAS No.:	Q1514	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	03/13/2025	19:21	LB135035
	Vanadium	40.0	+/-40.0	U	40.0	P	03/13/2025	19:21	LB135035
	Zinc	40.0	+/-40.0	U	40.0	P	03/13/2025	19:21	LB135035