


LB134958

7470A

INSTRUMENT M: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.065	-0.065	-56	0.000	-56	0				6
0.2	0.200	0.213	0.013	1125	0.0 %	1125					-1
2.5	2.500	2.475	-0.025	10755	0.0 %	10755					4
5.0	5.000	5.190	0.190	22313	0.0 %	22313					-1
7.5	7.500	7.422	-0.078	31815	0.0 %	31815					6
10.0	10.000	9.964	-0.036	42635	0.0 %	42635					

LB134958 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Mar 2025 09:37:29

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	-56	-	0.0000	7470A	PPB	10 Mar 2025 09:59:19	S	Std
0.2 - 1	30-2	1125	-	0.2000	7470A	PPB	10 Mar 2025 10:01:36	S	Std
2.5 - 1	52.5	10755	-	2.5000	7470A	PPB	10 Mar 2025 10:03:54	S	Std
5.0 - 1	55	22313	-	5.0000	7470A	PPB	10 Mar 2025 10:06:11	S	Std
7.5 - 1	57.5	31815	-	7.5000	7470A	PPB	10 Mar 2025 10:13:57	S	Std
10.0 - 1	516	42635	-	10.0000	7470A	PPB	10 Mar 2025 10:19:05	S	Std
ICV35 - 1	ICV35	17294	4.0112	-	7470A	PPB	10 Mar 2025 10:22:03	U	SMPL
ICB35 - 1	ICB35	-25	-0.0574	-	7470A	PPB	10 Mar 2025 10:26:52	U	SMPL
CCV75 - 1	CCV75	19837	4.6086	-	7470A	PPB	10 Mar 2025 10:34:16	U	SMPL
CCB75 - 1	CCB75	-76	-0.0694	-	7470A	PPB	10 Mar 2025 10:39:29	U	SMPL
CRA - 1	CRA	1120	0.2116	-	7470A	PPB	10 Mar 2025 10:41:43	U	SMPL
HighStd - 1	HighStd	41894	9.7901	-	7470A	PPB	10 Mar 2025 10:43:58	U	SMPL
ChkStd - 1	ChkStd	28380	6.6155	-	7470A	PPB	10 Mar 2025 10:46:13	U	SMPL
PB167033BL - 1	PBW	-28	-0.0581	-	7470A	PPB	10 Mar 2025 10:53:44	U	SMPL
PB167033BS - 1	LCSW	17655	4.0960	-	7470A	PPB	10 Mar 2025 10:56:00	U	SMPL
Q1488-02 - 1	ENV-101-SB01	-200	-0.0985	-	7470A	PPB	10 Mar 2025 10:58:16	U	SMPL
Q1488-04 - 1	ENV-101-SB02	-245	-0.1091	-	7470A	PPB	10 Mar 2025 11:00:35	U	SMPL
Q1488-06 - 1	ENV-102-SB01	24	-0.0459	-	7470A	PPB	10 Mar 2025 11:02:52	U	SMPL
Q1488-08 - 1	ENV-102-SB02	11	-0.0489	-	7470A	PPB	10 Mar 2025 11:05:08	U	SMPL
Q1488-10 - 1	ENV-104-SB01	126	-0.0219	-	7470A	PPB	10 Mar 2025 11:07:24	U	SMPL
CCV76 - 1	CCV76	20945	4.8689	-	7470A	PPB	10 Mar 2025 11:09:39	U	SMPL
CCB76 - 1	CCB76	-189	-0.0959	-	7470A	PPB	10 Mar 2025 11:11:55	U	SMPL
Q1488-12 - 1	ENV-104-SB02	49	-0.0400	-	7470A	PPB	10 Mar 2025 11:14:12	U	SMPL
Q1490-02 - 1	CLAY-SLUDGE-DRUMS	-11	-0.0541	-	7470A	PPB	10 Mar 2025 11:16:27	U	SMPL
Q1494-04 - 1	ASPHALT-SOIL	-5	-0.0527	-	7470A	PPB	10 Mar 2025 11:18:42	U	SMPL
Q1494-06 - 1	SOIL	-28	-0.0581	-	7470A	PPB	10 Mar 2025 11:20:58	U	SMPL
Q1494-08 - 1	SOIL-COMP	1	-0.0513	-	7470A	PPB	10 Mar 2025 11:23:16	U	SMPL
Q1510-01 - 1	FMC-25-0001-0005	19	-0.0470	-	7470A	PPB	10 Mar 2025 11:25:34	U	SMPL
Q1514-02 - 1	ENV-105-SB01	15	-0.0480	-	7470A	PPB	10 Mar 2025 11:27:50	U	SMPL
Q1514-04 - 1	ENV-105-SB02	140	-0.0186	-	7470A	PPB	10 Mar 2025 11:30:07	U	SMPL
Q1514-06 - 1	ENV-103-SB01	1	-0.0513	-	7470A	PPB	10 Mar 2025 11:39:15	U	SMPL
Q1514-06DUP - 1	ENV-103-SB01DUP	-25	-0.0574	-	7470A	PPB	10 Mar 2025 11:41:30	U	SMPL
CCV77 - 1	CCV77	22194	5.1623	-	7470A	PPB	10 Mar 2025 11:48:13	U	SMPL
CCB77 - 1	CCB77	10	-0.0492	-	7470A	PPB	10 Mar 2025 11:53:10	U	SMPL
Q1514-06MS - 1	ENV-103-SB01MS	18531	4.3018	-	7470A	PPB	10 Mar 2025 11:55:25	U	SMPL
Q1514-06MSD - 1	ENV-103-SB01MSD	18123	4.2059	-	7470A	PPB	10 Mar 2025 11:57:40	U	SMPL
PB167034BL - 1	PBW	9	-0.0494	-	7470A	PPB	10 Mar 2025 11:59:57	U	SMPL
PB167034BS - 1	LCSW	17577	4.0776	-	7470A	PPB	10 Mar 2025 12:02:14	U	SMPL
Q1478-01 - 1	IDW-AQ-MW-19B-COMP-022825	-286	-0.1187	-	7470A	PPB	10 Mar 2025 12:04:30	U	SMPL
Q1478-01DUP - 1	IDW-AQ-MW-19B-COMP-022825DUP	-43	-0.0616	-	7470A	PPB	10 Mar 2025 12:06:49	U	SMPL
Q1478-01MS - 1	IDW-AQ-MW-19B-COMP-022825MS	15282	3.5385	-	7470A	PPB	10 Mar 2025 12:09:05	U	SMPL
Q1478-01MSD - 1	IDW-AQ-MW-19B-COMP-022825MSD	15003	3.4730	-	7470A	PPB	10 Mar 2025 12:11:22	U	SMPL
Q1478-03 - 1	IDW-AQ-IW-01-COMP-022825	-252	-0.1107	-	7470A	PPB	10 Mar 2025 12:13:41	U	SMPL
Q1478-05 - 1	IDW-AQ-IW-02-COMP-022825	-15	-0.0550	-	7470A	PPB	10 Mar 2025 12:15:58	U	SMPL
CCV78 - 1	CCV78	22362	5.2017	-	7470A	PPB	10 Mar 2025 12:18:14	U	SMPL
CCB78 - 1	CCB78	-48	-0.0628	-	7470A	PPB	10 Mar 2025 12:26:13	U	SMPL
Q1478-07 - 1	IDW-AQ-IW-03-COMP-022825	6	-0.0501	-	7470A	PPB	10 Mar 2025 12:28:28	U	SMPL
Q1494-01 - 1	PURGE-WATER	18	-0.0473	-	7470A	PPB	10 Mar 2025 12:30:44	U	SMPL
Q1514-07 - 1	ENV-105-GW01	259291	60.8606	-	7470A	PPB	10 Mar 2025 12:32:59	U	SMPL
Q1514-08 - 1	ENV-103-GW01	2433	0.5200	-	7470A	PPB	10 Mar 2025 12:35:15	U	SMPL
Q1514-09 - 1	FB03062025	-240	-0.1079	-	7470A	PPB	10 Mar 2025 12:38:16	U	SMPL
Q1516-03 - 1	GAS-AUD-25-0026	-19	-0.0560	-	7470A	PPB	10 Mar 2025 12:40:34	U	SMPL
Q1517-01 - 1	MOO-25-0062	188	-0.0073	-	7470A	PPB	10 Mar 2025 12:42:51	U	SMPL
CCV79 - 1	CCV79	21747	5.0573	-	7470A	PPB	10 Mar 2025 12:45:08	U	SMPL
CCB79 - 1	CCB79	141	-0.0184	-	7470A	PPB	10 Mar 2025 12:47:26	U	SMPL
PB67020TB - 1		-17	-0.0555	-	7470A	PPB	10 Mar 2025 12:49:43	U	SMPL
Q1514-06LX5 - 1		205	-0.0033	-	7470A	PPB	10 Mar 2025 12:55:15	U	SMPL
Q1514-06A - 1		17135	3.9738	-	7470A	PPB	10 Mar 2025 12:57:30	U	SMPL
Q1478-01LX5 - 1		-152	-0.0872	-	7470A	PPB	10 Mar 2025 12:59:46	U	SMPL
Q1478-01A - 1		16901	3.9188	-	7470A	PPB	10 Mar 2025 13:05:50	U	SMPL
Q1514-07DLX10 - 1		28760	6.7047	-	7470A	PPB	10 Mar 2025 13:08:06	U	SMPL
CCV80 - 1	CCV80	23339	5.4312	-	7470A	PPB	10 Mar 2025 13:10:23	U	SMPL
CCB80 - 1	CCB80	112	-0.0252	-	7470A	PPB	10 Mar 2025 13:12:41	U	SMPL