



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture SDG No.: Q1514
Contract: PORT06 Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	106	40 - 160	CV	03/10/2025	10:41	LB134958
CRI01	Arsenic	19.0	20.0	95	40 - 160	P	03/11/2025	11:02	LB134987
	Barium	95.1	100	95	40 - 160	P	03/11/2025	11:02	LB134987
	Cadmium	5.81	6.0	97	40 - 160	P	03/11/2025	11:02	LB134987
	Chromium	9.83	10.0	98	40 - 160	P	03/11/2025	11:02	LB134987
	Lead	11.9	12.0	99	40 - 160	P	03/11/2025	11:02	LB134987
	Selenium	21.9	20.0	110	40 - 160	P	03/11/2025	11:02	LB134987
	Silver	10.3	10.0	103	40 - 160	P	03/11/2025	11:02	LB134987
CRI01	Arsenic	20.1	20.0	100	40 - 160	P	03/12/2025	11:05	LB135011
	Barium	88.8	100	89	40 - 160	P	03/12/2025	11:05	LB135011
	Cadmium	5.85	6.0	98	40 - 160	P	03/12/2025	11:05	LB135011
	Chromium	10.3	10.0	103	40 - 160	P	03/12/2025	11:05	LB135011
	Lead	10.9	12.0	91	40 - 160	P	03/12/2025	11:05	LB135011
	Selenium	18.3	20.0	92	40 - 160	P	03/12/2025	11:05	LB135011
	Silver	11.1	10.0	111	40 - 160	P	03/12/2025	11:05	LB135011
CRI01	Arsenic	21.2	20.0	106	40 - 160	P	03/13/2025	12:21	LB135035
	Barium	94.0	100	94	40 - 160	P	03/13/2025	12:21	LB135035
	Cadmium	5.90	6.0	98	40 - 160	P	03/13/2025	12:21	LB135035
	Chromium	10.4	10.0	104	40 - 160	P	03/13/2025	12:21	LB135035
	Lead	11.7	12.0	98	40 - 160	P	03/13/2025	12:21	LB135035
	Selenium	21.3	20.0	107	40 - 160	P	03/13/2025	12:21	LB135035
	Silver	10.6	10.0	106	40 - 160	P	03/13/2025	12:21	LB135035