



- 7 -

Client:	Portal Partners Tri-Venture	SDG No.:	Q1514				
Contract:	PORT06	Lab Code:	CHEM	Case No.:	Q1514	SAS No.:	Q1514

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167033BS Mercury	ug/L	4.0	4.10		102	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** Q1514
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1514 **SAS No.:** Q1514

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167037BS							
Arsenic	ug/L	4000	4120		103	80 - 120	P
Barium	ug/L	1000	882		88	80 - 120	P
Cadmium	ug/L	1000	955		96	80 - 120	P
Chromium	ug/L	2000	2140		107	80 - 120	P
Lead	ug/L	5000	4890		98	80 - 120	P
Selenium	ug/L	10000	9950		100	80 - 120	P
Silver	ug/L	380	384		101	80 - 120	P