

Continuing Calibration Report for SequenceID : FC031125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068389.D

Aliphatic C9-C12	8625598.000	60.000	3.101	6.364	143759.967	146281.713	1.724
Aliphatic C12-C16	5422929.000	40.000	6.365	9.747	135573.225	141000.725	3.849
Aliphatic C16-C21	7841092.000	60.000	9.748	13.100	130684.867	136996.570	4.607
Aliphatic C21-C28	9997401.000	80.000	13.101	16.752	124967.513	131264.609	4.797
Aliphatic C28-C40	12578828.000	120.000	16.753	21.552	104823.567	115412.820	9.175
Aliphatic EPH	44465848.000	360.000	3.101	21.552	123516.244	130520.647	5.367