

Continuing Calibration Report for SequenceID : FD031125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049179.D

Aromatic C10-C12	6699385.000	40.000	4.074	5.792	167484.625	164744.953	-1.663
Aromatic C12-C16	10091225.000	60.000	5.793	8.394	168187.083	174935.818	3.858
Aromatic C16-C21	12523109.000	80.000	8.395	12.658	156538.863	158351.936	1.145
Aromatic C21-C36	24929730.000	180.000	12.659	18.072	138498.500	133632.291	-3.641
Aromatic EPH	54243449.000	360.000	4.074	18.072	150676.247	149466.429	-0.809