

Continuing Calibration Report for SequenceID : FD031125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049188.D

Aromatic C10-C12	6767053.000	40.000	4.074	5.792	169176.325	164744.953	-2.690
Aromatic C12-C16	10153428.000	60.000	5.793	8.394	169223.800	174935.818	3.265
Aromatic C16-C21	12677995.000	80.000	8.395	12.658	158474.938	158351.936	-0.078
Aromatic C21-C36	26006138.000	180.000	12.659	18.072	144478.544	133632.291	-8.116
Aromatic EPH	55604614.000	360.000	4.074	18.072	154457.261	149466.429	-3.339