

Continuing Calibration Report for SequenceID : FE031125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE052727.D

Aliphatic C9-C12	9718254.000	60.000	3.148	6.791	161970.900	155948.436	-3.862
Aliphatic C12-C16	6383599.000	40.000	6.792	10.243	159589.975	152098.489	-4.925
Aliphatic C16-C21	9144320.000	60.000	10.244	13.622	152405.333	146548.085	-3.997
Aliphatic C21-C28	12023860.000	80.000	13.623	17.293	150298.250	142451.851	-5.508
Aliphatic C28-C40	16778160.000	120.000	17.294	22.207	139818.000	127589.108	-9.585
Aliphatic EPH	54048193.000	360.000	3.148	22.207	150133.869	141501.588	-6.100