

Continuing Calibration Report for SequenceID : FF031125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015651.D

Aromatic C10-C12	5477031.000	40.000	4.503	6.375	136925.775	128361.211	-6.672
Aromatic C12-C16	8600013.000	60.000	6.376	9.068	143333.550	136800.309	-4.776
Aromatic C16-C21	10909525.000	80.000	9.069	13.389	136369.063	133038.999	-2.503
Aromatic C21-C36	22831546.000	180.000	13.390	18.846	126841.922	130055.838	2.471
Aromatic EPH	47818115.000	360.000	4.503	18.846	132828.097	131654.549	-0.891