

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1514 **SAS No :** Q1514 **SDG No:** Q1514
Sample No : PB167056BS **Datafile:** FC068391.D
Client ID : PB167056BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	158	53		(40-140)
Aliphatic C12-C16	200	147	74		(40-140)
Aliphatic C16-C21	300	259	86		(40-140)
Aliphatic C21-C28	400	348	87		(40-140)
Aliphatic C28-C40	600	556	93		(40-140)