

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1514 **SAS No :** Q1514 **SDG No:** Q1514
Sample No : PB167056BS **Datafile:** FD049181.D
Client ID : PB167056BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	145	72		(40-140)
Aromatic C12-C16	300	234	78		(40-140)
Aromatic C16-C21	500	499	100		(40-140)
Aromatic C21-C36	800	867	108		(40-140)