

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1514 **SAS No :** Q1514 **SDG No:** Q1514
Sample No : PB167058BSD **Datafile:** FE052710.D
Client ID : PB167058BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10	5.52	55		0.97 (40-140)	50
Aliphatic C12-C16	6.7	5.40	81		0.68 (40-140)	50
Aliphatic C16-C21	10	9.11	91		0.28 (40-140)	50
Aliphatic C21-C28	13.3	12.2	92		0.83 (40-140)	50
Aliphatic C28-C40	20.0	20.5	102		0.55 (40-140)	50