

**SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY**

**Lab Name:** Chemtech **Client:** Portal Partners Tri-Venture  
**Lab Code:** CHEM **Cas No:** Q1514 **SAS No :** Q1514 **SDG No:** Q1514  
**Sample No :** PB167058BSD **Datafile:** FF015631.D  
**Client ID :** PB167058BSD

| COMPOUND         | SPIKE<br>ADDED<br>mg/kg | LCS/LCSD<br>CONCENTRATION<br>mg/kg | % REC | Qual | RPD   QC LIMITS | QC Limit<br>Of RPD |
|------------------|-------------------------|------------------------------------|-------|------|-----------------|--------------------|
| Aromatic C10-C12 | 6.7                     | 5.04                               | 76    |      | 10.3   (40-140) | 50                 |
| Aromatic C12-C16 | 10                      | 8.92                               | 89    |      | 11.2   (40-140) | 50                 |
| Aromatic C16-C21 | 16.7                    | 17.2                               | 103   |      | 11.6   (40-140) | 50                 |
| Aromatic C21-C36 | 26.6                    | 26.7                               | 100   |      | 11.1   (40-140) | 50                 |