

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

Continuing Calib Date: 03/08/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 00:30 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.68       | 5.68      | 5.58      | 5.78  | 0.00       |
| Aroclor-1248-2 (2)   | 5.95       | 5.95      | 5.85      | 6.05  | 0.00       |
| Aroclor-1248-3 (3)   | 6.15       | 6.15      | 6.05      | 6.25  | 0.00       |
| Aroclor-1248-4 (4)   | 6.55       | 6.55      | 6.45      | 6.65  | 0.00       |
| Aroclor-1248-5 (5)   | 6.59       | 6.59      | 6.49      | 6.69  | 0.00       |
| Tetrachloro-m-xylene | 4.53       | 4.53      | 4.43      | 4.63  | 0.00       |
| Decachlorobiphenyl   | 10.25      | 10.26     | 10.16     | 10.36 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

Continuing Calib Date: 03/08/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 00:30 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.92       | 4.92      | 4.82      | 5.02 | 0.00       |
| Aroclor-1248-2       | (2) | 5.15       | 5.16      | 5.06      | 5.26 | 0.01       |
| Aroclor-1248-3       | (3) | 5.20       | 5.20      | 5.10      | 5.30 | 0.00       |
| Aroclor-1248-4       | (4) | 5.37       | 5.37      | 5.27      | 5.47 | 0.00       |
| Aroclor-1248-5       | (5) | 5.76       | 5.77      | 5.67      | 5.87 | 0.01       |
| Tetrachloro-m-xylene |     | 3.83       | 3.83      | 3.73      | 3.93 | 0.00       |
| Decachlorobiphenyl   |     | 8.88       | 8.89      | 8.79      | 8.99 | 0.01       |