

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109723.D Time Analyzed: 17:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.789	4.691	4.891	528.490	500.000	5.7
Aroclor-1016-2	4.809	4.711	4.911	534.270	500.000	6.9
Aroclor-1016-3	4.865	4.767	4.967	530.540	500.000	6.1
Aroclor-1016-4	4.986	4.888	5.088	528.650	500.000	5.7
Aroclor-1016-5	5.243	5.145	5.345	533.980	500.000	6.8
Aroclor-1260-1	6.285	6.188	6.388	462.800	500.000	-7.4
Aroclor-1260-2	6.475	6.377	6.577	464.780	500.000	-7.0
Aroclor-1260-3	6.843	6.746	6.946	480.600	500.000	-3.9
Aroclor-1260-4	7.103	7.005	7.205	474.640	500.000	-5.1
Aroclor-1260-5	7.345	7.247	7.447	455.720	500.000	-8.9
Decachlorobiphenyl	8.753	8.657	8.857	47.210	50.000	-5.6
Tetrachloro-m-xylene	3.696	3.598	3.798	52.860	50.000	5.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109723.D Time Analyzed: 17:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.777	4.678	4.878	559.280	500.000	11.9
Aroclor-1016-2	4.796	4.697	4.897	567.820	500.000	13.6
Aroclor-1016-3	4.973	4.872	5.072	564.830	500.000	13.0
Aroclor-1016-4	5.014	4.914	5.114	529.430	500.000	5.9
Aroclor-1016-5	5.227	5.127	5.327	564.140	500.000	12.8
Aroclor-1260-1	6.260	6.161	6.361	507.370	500.000	1.5
Aroclor-1260-2	6.447	6.348	6.548	501.990	500.000	0.4
Aroclor-1260-3	6.601	6.502	6.702	495.000	500.000	-1.0
Aroclor-1260-4	7.073	6.974	7.174	480.740	500.000	-3.9
Aroclor-1260-5	7.314	7.214	7.414	491.550	500.000	-1.7
Decachlorobiphenyl	8.706	8.607	8.807	48.200	50.000	-3.6
Tetrachloro-m-xylene	3.694	3.594	3.794	56.280	50.000	12.6