

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109737.D Time Analyzed: 22:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.789	4.691	4.891	540.020	500.000	8.0
Aroclor-1016-2	4.809	4.711	4.911	542.330	500.000	8.5
Aroclor-1016-3	4.865	4.767	4.967	543.280	500.000	8.7
Aroclor-1016-4	4.985	4.888	5.088	545.240	500.000	9.0
Aroclor-1016-5	5.244	5.145	5.345	568.490	500.000	13.7
Aroclor-1260-1	6.285	6.188	6.388	527.110	500.000	5.4
Aroclor-1260-2	6.474	6.377	6.577	516.030	500.000	3.2
Aroclor-1260-3	6.842	6.746	6.946	506.590	500.000	1.3
Aroclor-1260-4	7.102	7.005	7.205	494.530	500.000	-1.1
Aroclor-1260-5	7.344	7.247	7.447	488.740	500.000	-2.3
Decachlorobiphenyl	8.752	8.657	8.857	43.420	50.000	-13.2
Tetrachloro-m-xylene	3.696	3.598	3.798	53.250	50.000	6.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109737.D Time Analyzed: 22:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.777	4.678	4.878	561.250	500.000	12.3
Aroclor-1016-2	4.796	4.697	4.897	574.670	500.000	14.9
Aroclor-1016-3	4.971	4.872	5.072	572.490	500.000	14.5
Aroclor-1016-4	5.013	4.914	5.114	519.900	500.000	4.0
Aroclor-1016-5	5.226	5.127	5.327	592.260	500.000	18.5
Aroclor-1260-1	6.259	6.161	6.361	541.810	500.000	8.4
Aroclor-1260-2	6.447	6.348	6.548	542.930	500.000	8.6
Aroclor-1260-3	6.600	6.502	6.702	533.020	500.000	6.6
Aroclor-1260-4	7.072	6.974	7.174	520.050	500.000	4.0
Aroclor-1260-5	7.313	7.214	7.414	527.490	500.000	5.5
Decachlorobiphenyl	8.705	8.607	8.807	45.760	50.000	-8.5
Tetrachloro-m-xylene	3.694	3.594	3.794	56.600	50.000	13.2