

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL05 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP070374.D Time Analyzed: 20:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.678	5.581	5.781	483.600	500.000	-3.3
Aroclor-1016-2	5.700	5.602	5.802	489.820	500.000	-2.0
Aroclor-1016-3	5.762	5.665	5.865	482.920	500.000	-3.4
Aroclor-1016-4	5.860	5.762	5.962	488.620	500.000	-2.3
Aroclor-1016-5	6.153	6.056	6.256	484.370	500.000	-3.1
Aroclor-1260-1	7.272	7.175	7.375	502.220	500.000	0.4
Aroclor-1260-2	7.526	7.429	7.629	478.070	500.000	-4.4
Aroclor-1260-3	7.885	7.788	7.988	511.980	500.000	2.4
Aroclor-1260-4	8.109	8.012	8.212	496.830	500.000	-0.6
Aroclor-1260-5	8.430	8.333	8.533	499.330	500.000	-0.1
Decachlorobiphenyl	10.253	10.156	10.356	48.470	50.000	-3.1
Tetrachloro-m-xylene	4.526	4.427	4.627	53.510	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL05 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP070374.D Time Analyzed: 20:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.916	4.820	5.020	499.910	500.000	0.0
Aroclor-1016-2	4.935	4.839	5.039	503.950	500.000	0.8
Aroclor-1016-3	5.112	5.017	5.217	522.390	500.000	4.5
Aroclor-1016-4	5.155	5.058	5.258	500.390	500.000	0.1
Aroclor-1016-5	5.369	5.273	5.473	532.610	500.000	6.5
Aroclor-1260-1	6.407	6.312	6.512	483.720	500.000	-3.3
Aroclor-1260-2	6.595	6.500	6.700	494.560	500.000	-1.1
Aroclor-1260-3	6.749	6.654	6.854	460.460	500.000	-7.9
Aroclor-1260-4	7.221	7.126	7.326	508.970	500.000	1.8
Aroclor-1260-5	7.463	7.367	7.567	526.760	500.000	5.4
Decachlorobiphenyl	8.882	8.788	8.988	48.760	50.000	-2.5
Tetrachloro-m-xylene	3.828	3.730	3.930	50.130	50.000	0.3