

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022025

Review By	yogesh	Review On	2/21/2025 8:09:53 AM
Supervise By	Ankita	Supervise On	2/21/2025 9:30:37 AM
SubDirectory	PO022025	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO109424.D	20 Feb 2025 16:09	YP/AJ	Ok
2	I.BLK	PO109425.D	20 Feb 2025 16:28	YP/AJ	Ok
3	AR1660ICC1000	PO109426.D	20 Feb 2025 16:46	YP/AJ	Ok
4	AR1660ICC750	PO109427.D	20 Feb 2025 17:04	YP/AJ	Ok
5	AR1660ICC500	PO109428.D	20 Feb 2025 17:23	YP/AJ	Ok
6	AR1660ICC250	PO109429.D	20 Feb 2025 17:41	YP/AJ	Ok
7	AR1660ICC050	PO109430.D	20 Feb 2025 17:59	YP/AJ	Ok
8	AR1221ICC500	PO109431.D	20 Feb 2025 18:18	YP/AJ	Ok
9	AR1232ICC500	PO109432.D	20 Feb 2025 18:36	YP/AJ	Ok
10	AR1242ICC1000	PO109433.D	20 Feb 2025 18:55	YP/AJ	Ok
11	AR1242ICC750	PO109434.D	20 Feb 2025 19:13	YP/AJ	Ok
12	AR1242ICC500	PO109435.D	20 Feb 2025 19:31	YP/AJ	Ok
13	AR1242ICC250	PO109436.D	20 Feb 2025 19:50	YP/AJ	Ok
14	AR1242ICC050	PO109437.D	20 Feb 2025 20:08	YP/AJ	Ok,M
15	AR1248ICC1000	PO109438.D	20 Feb 2025 20:26	YP/AJ	Ok
16	AR1248ICC750	PO109439.D	20 Feb 2025 20:45	YP/AJ	Ok
17	AR1248ICC500	PO109440.D	20 Feb 2025 21:03	YP/AJ	Ok
18	AR1248ICC250	PO109441.D	20 Feb 2025 21:21	YP/AJ	Ok
19	AR1248ICC050	PO109442.D	20 Feb 2025 21:40	YP/AJ	Ok,M
20	AR1254ICC1000	PO109443.D	20 Feb 2025 21:58	YP/AJ	Ok
21	AR1254ICC750	PO109444.D	20 Feb 2025 22:17	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022025

Review By	yogesh	Review On	2/21/2025 8:09:53 AM
Supervise By	Ankita	Supervise On	2/21/2025 9:30:37 AM
SubDirectory	PO022025	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO109445.D	20 Feb 2025 22:35	YP/AJ	Ok
23	AR1254ICC250	PO109446.D	20 Feb 2025 22:53	YP/AJ	Ok
24	AR1254ICC050	PO109447.D	20 Feb 2025 23:12	YP/AJ	Ok,M
25	AR1262ICC500	PO109448.D	20 Feb 2025 23:30	YP/AJ	Ok
26	AR1268ICC1000	PO109449.D	20 Feb 2025 23:48	YP/AJ	Ok
27	AR1268ICC750	PO109450.D	21 Feb 2025 00:07	YP/AJ	Ok
28	AR1268ICC500	PO109451.D	21 Feb 2025 00:25	YP/AJ	Ok
29	AR1268ICC250	PO109452.D	21 Feb 2025 00:43	YP/AJ	Ok
30	AR1268ICC050	PO109453.D	21 Feb 2025 01:02	YP/AJ	Ok
31	PO022025ICV500	PO109454.D	21 Feb 2025 01:20	YP/AJ	Ok
32	AR1242ICV500	PO109455.D	21 Feb 2025 01:38	YP/AJ	Ok
33	AR1248ICV500	PO109456.D	21 Feb 2025 01:57	YP/AJ	Ok
34	AR1254ICV500	PO109457.D	21 Feb 2025 02:15	YP/AJ	Ok
35	AR1268ICV500	PO109458.D	21 Feb 2025 02:34	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030725

Review By	yogesh	Review On	3/10/2025 10:54:39 AM
Supervise By	Ankita	Supervise On	3/10/2025 10:55:05 AM
SubDirectory	PO030725	HP Acquire Method	HP Processing Method
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO109697.D	07 Mar 2025 08:40	YP/AJ	Ok
2	AR1660CCC500	PO109698.D	07 Mar 2025 08:58	YP/AJ	Ok
3	I.BLK	PO109699.D	07 Mar 2025 09:17	YP/AJ	Ok
4	Q1437-07DL	PO109700.D	07 Mar 2025 09:36	YP/AJ	Ok
5	Q1438-03DL	PO109701.D	07 Mar 2025 09:55	YP/AJ	Ok
6	Q1438-05DL	PO109702.D	07 Mar 2025 10:12	YP/AJ	Ok
7	Q1433-06DL	PO109703.D	07 Mar 2025 10:29	YP/AJ	Ok
8	Q1433-09DL	PO109704.D	07 Mar 2025 10:48	YP/AJ	Ok
9	Q1434-02DL	PO109705.D	07 Mar 2025 11:06	YP/AJ	Not Ok
10	Q1434-05DL	PO109706.D	07 Mar 2025 11:24	YP/AJ	Ok
11	Q1434-07DL	PO109707.D	07 Mar 2025 11:43	YP/AJ	Ok
12	AR1660CCC500	PO109708.D	07 Mar 2025 12:30	YP/AJ	Ok
13	AR1242CCC500	PO109709.D	07 Mar 2025 12:48	YP/AJ	Ok
14	AR1248CCC500	PO109710.D	07 Mar 2025 13:05	YP/AJ	Ok
15	AR1254CCC500	PO109711.D	07 Mar 2025 13:24	YP/AJ	Ok
16	I.BLK	PO109712.D	07 Mar 2025 13:42	YP/AJ	Ok
17	PB167025BL	PO109713.D	07 Mar 2025 14:00	YP/AJ	Ok
18	PB167025BS	PO109714.D	07 Mar 2025 14:19	YP/AJ	Ok
19	Q1509-01	PO109715.D	07 Mar 2025 14:37	YP/AJ	Ok,M
20	Q1509-02	PO109716.D	07 Mar 2025 14:56	YP/AJ	Ok,M
21	Q1520-01	PO109717.D	07 Mar 2025 15:14	YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030725

Review By	yogesh	Review On	3/10/2025 10:54:39 AM
Supervise By	Ankita	Supervise On	3/10/2025 10:55:05 AM
SubDirectory	PO030725	HP Acquire Method	HP Processing Method
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1520-02	PO109718.D	07 Mar 2025 15:32	YP/AJ	Ok,M
23	Q1520-03	PO109719.D	07 Mar 2025 15:51	YP/AJ	Ok,M
24	Q1520-04	PO109720.D	07 Mar 2025 16:09	YP/AJ	Ok,M
25	Q1520-05	PO109721.D	07 Mar 2025 16:28	YP/AJ	Ok,M
26	Q1520-06	PO109722.D	07 Mar 2025 16:46	YP/AJ	Ok,M
27	AR1660CCC500	PO109723.D	07 Mar 2025 17:33	YP/AJ	Ok
28	AR1242CCC500	PO109724.D	07 Mar 2025 17:50	YP/AJ	Ok
29	AR1248CCC500	PO109725.D	07 Mar 2025 18:08	YP/AJ	Ok
30	AR1254CCC500	PO109726.D	07 Mar 2025 18:27	YP/AJ	Ok
31	I.BLK	PO109727.D	07 Mar 2025 18:45	YP/AJ	Ok
32	PB167028BL	PO109728.D	07 Mar 2025 19:04	YP/AJ	Ok
33	PB167028BS	PO109729.D	07 Mar 2025 19:22	YP/AJ	Ok
34	PB167028BSD	PO109730.D	07 Mar 2025 19:40	YP/AJ	Ok
35	Q1514-07	PO109731.D	07 Mar 2025 19:59	YP/AJ	Ok,M
36	Q1514-08	PO109732.D	07 Mar 2025 20:17	YP/AJ	Ok
37	Q1514-09	PO109733.D	07 Mar 2025 20:35	YP/AJ	Ok
38	Q1516-03	PO109734.D	07 Mar 2025 20:54	YP/AJ	Ok
39	Q1517-01	PO109735.D	07 Mar 2025 21:12	YP/AJ	Ok,M
40	Q1434-02DL	PO109736.D	07 Mar 2025 21:31	YP/AJ	Ok,M
41	AR1660CCC500	PO109737.D	07 Mar 2025 22:17	YP/AJ	Ok
42	AR1242CCC500	PO109738.D	07 Mar 2025 22:36	YP/AJ	Ok
43	AR1248CCC500	PO109739.D	07 Mar 2025 22:54	YP/AJ	Ok
44	AR1254CCC500	PO109740.D	07 Mar 2025 23:11	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030725

Review By	yogesh	Review On	3/10/2025 10:54:39 AM
Supervise By	Ankita	Supervise On	3/10/2025 10:55:05 AM
SubDirectory	PO030725	HP Acquire Method	HP Processing Method
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	I.BLK	PO109741.D	07 Mar 2025 23:30	YP/AJ	Ok
----	-------	------------	-------------------	-------	----

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP069994.D	24 Feb 2025 14:27	YP\AJ	Ok
2	I.BLK	PP069995.D	24 Feb 2025 14:43	YP\AJ	Ok
3	AR1660ICC1000	PP069996.D	24 Feb 2025 14:59	YP\AJ	Ok
4	AR1660ICC750	PP069997.D	24 Feb 2025 15:15	YP\AJ	Ok
5	AR1660ICC500	PP069998.D	24 Feb 2025 15:32	YP\AJ	Ok
6	AR1660ICC250	PP069999.D	24 Feb 2025 15:48	YP\AJ	Ok
7	AR1660ICC050	PP070000.D	24 Feb 2025 16:04	YP\AJ	Ok,M
8	AR1221ICC500	PP070001.D	24 Feb 2025 16:20	YP\AJ	Ok,M
9	AR1232ICC500	PP070002.D	24 Feb 2025 16:37	YP\AJ	Ok
10	AR1242ICC1000	PP070003.D	24 Feb 2025 16:53	YP\AJ	Ok
11	AR1242ICC750	PP070004.D	24 Feb 2025 17:09	YP\AJ	Ok,M
12	AR1242ICC500	PP070005.D	24 Feb 2025 17:25	YP\AJ	Ok
13	AR1242ICC250	PP070006.D	24 Feb 2025 17:42	YP\AJ	Ok,M
14	AR1242ICC050	PP070007.D	24 Feb 2025 17:58	YP\AJ	Ok,M
15	AR1248ICC1000	PP070008.D	24 Feb 2025 18:14	YP\AJ	Ok
16	AR1248ICC750	PP070009.D	24 Feb 2025 18:30	YP\AJ	Ok
17	AR1248ICC500	PP070010.D	24 Feb 2025 18:46	YP\AJ	Ok
18	AR1248ICC250	PP070011.D	24 Feb 2025 19:03	YP\AJ	Ok
19	AR1248ICC050	PP070012.D	24 Feb 2025 19:19	YP\AJ	Ok,M
20	AR1254ICC1000	PP070013.D	24 Feb 2025 19:35	YP\AJ	Ok,M
21	AR1254ICC750	PP070014.D	24 Feb 2025 19:51	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP070015.D	24 Feb 2025 20:08	YP\AJ	Ok
23	AR1254ICC250	PP070016.D	24 Feb 2025 20:24	YP\AJ	Ok
24	AR1254ICC050	PP070017.D	24 Feb 2025 20:40	YP\AJ	Ok,M
25	AR1262ICC500	PP070018.D	24 Feb 2025 20:56	YP\AJ	Ok
26	AR1268ICC1000	PP070019.D	24 Feb 2025 21:12	YP\AJ	Ok
27	AR1268ICC750	PP070020.D	24 Feb 2025 21:29	YP\AJ	Ok,M
28	AR1268ICC500	PP070021.D	24 Feb 2025 21:45	YP\AJ	Ok
29	AR1268ICC250	PP070022.D	24 Feb 2025 22:01	YP\AJ	Ok,M
30	AR1268ICC050	PP070023.D	24 Feb 2025 22:17	YP\AJ	Ok,M
31	PP022425ICV500	PP070024.D	24 Feb 2025 22:34	YP\AJ	Ok
32	AR1242ICV500	PP070025.D	24 Feb 2025 22:50	YP\AJ	Ok
33	AR1248ICV500	PP070026.D	24 Feb 2025 23:06	YP\AJ	Ok
34	AR1254ICV500	PP070027.D	24 Feb 2025 23:22	YP\AJ	Ok
35	AR1268ICV500	PP070028.D	24 Feb 2025 23:38	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030725

Review By	yogesh	Review On	3/7/2025 1:29:40 PM
Supervise By	Ankita	Supervise On	3/10/2025 11:00:00 AM
SubDirectory	PP030725	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP070337.D	07 Mar 2025 08:41	YP\AJ	Ok
2	AR1660CCC500	PP070338.D	07 Mar 2025 08:57	YP\AJ	Ok
3	I.BLK	PP070339.D	07 Mar 2025 09:15	YP\AJ	Ok
4	Q1434-10DL	PP070340.D	07 Mar 2025 09:32	YP\AJ	Ok,M
5	Q1436-02DL	PP070341.D	07 Mar 2025 09:48	YP\AJ	Ok,M
6	Q1436-08DL	PP070342.D	07 Mar 2025 10:04	YP\AJ	Dilution
7	Q1436-08DL2	PP070343.D	07 Mar 2025 10:20	YP\AJ	Ok,M
8	AR1660CCC500	PP070344.D	07 Mar 2025 10:58	YP\AJ	Ok
9	AR1242CCC500	PP070345.D	07 Mar 2025 11:14	YP\AJ	Ok
10	AR1248CCC500	PP070346.D	07 Mar 2025 11:31	YP\AJ	Ok
11	AR1254CCC500	PP070347.D	07 Mar 2025 11:47	YP\AJ	Ok,M
12	I.BLK	PP070348.D	07 Mar 2025 12:03	YP\AJ	Ok
13	PB167005BL	PP070349.D	07 Mar 2025 12:19	YP\AJ	Ok
14	Q1472-01	PP070350.D	07 Mar 2025 12:36	YP\AJ	Ok,M
15	Q1437-01DL	PP070351.D	07 Mar 2025 12:52	YP\AJ	Ok,M
16	Q1437-04DL	PP070352.D	07 Mar 2025 13:08	YP\AJ	Dilution
17	Q1437-04DL2	PP070353.D	07 Mar 2025 13:24	YP\AJ	Ok,M
18	PB167029BL	PP070354.D	07 Mar 2025 13:41	YP\AJ	Ok
19	PB167029BS	PP070355.D	07 Mar 2025 13:57	YP\AJ	Ok
20	Q1437-02	PP070356.D	07 Mar 2025 14:13	YP\AJ	Ok,M
21	Q1437-05	PP070357.D	07 Mar 2025 14:29	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030725

Review By	yogesh	Review On	3/7/2025 1:29:40 PM
Supervise By	Ankita	Supervise On	3/10/2025 11:00:00 AM
SubDirectory	PP030725	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1437-08	PP070358.D	07 Mar 2025 14:46	YP\AJ	Ok,M
23	AR1660CCC500	PP070359.D	07 Mar 2025 15:45	YP\AJ	Ok
24	AR1242CCC500	PP070360.D	07 Mar 2025 16:01	YP\AJ	Ok,M
25	AR1248CCC500	PP070361.D	07 Mar 2025 16:18	YP\AJ	Ok,M
26	AR1254CCC500	PP070362.D	07 Mar 2025 16:34	YP\AJ	Ok,M
27	I.BLK	PP070363.D	07 Mar 2025 16:50	YP\AJ	Ok
28	Q1438-01	PP070364.D	07 Mar 2025 17:06	YP\AJ	Ok,M
29	Q1438-09	PP070365.D	07 Mar 2025 17:23	YP\AJ	Ok,M
30	Q1507-01	PP070366.D	07 Mar 2025 17:39	YP\AJ	Ok,M
31	Q1508-01	PP070367.D	07 Mar 2025 17:55	YP\AJ	Ok,M
32	Q1508-01MS	PP070368.D	07 Mar 2025 18:11	YP\AJ	Ok,M
33	Q1508-01MSD	PP070369.D	07 Mar 2025 18:27	YP\AJ	Ok,M
34	Q1514-01	PP070370.D	07 Mar 2025 18:44	YP\AJ	Ok
35	Q1514-03	PP070371.D	07 Mar 2025 19:00	YP\AJ	Ok,M
36	Q1514-05	PP070372.D	07 Mar 2025 19:16	YP\AJ	Ok,M
37	Q1515-01	PP070373.D	07 Mar 2025 19:32	YP\AJ	Ok,M
38	AR1660CCC500	PP070374.D	07 Mar 2025 20:32	YP\AJ	Ok,M
39	AR1242CCC500	PP070375.D	07 Mar 2025 20:48	YP\AJ	Ok,M
40	AR1248CCC500	PP070376.D	07 Mar 2025 21:04	YP\AJ	Ok,M
41	AR1254CCC500	PP070377.D	07 Mar 2025 21:21	YP\AJ	Ok,M
42	I.BLK	PP070378.D	07 Mar 2025 21:37	YP\AJ	Ok,M
43	Q1516-01	PP070379.D	07 Mar 2025 21:53	YP\AJ	Ok,M
44	Q1516-02	PP070380.D	07 Mar 2025 22:09	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030725

Review By	yogesh	Review On	3/7/2025 1:29:40 PM
Supervise By	Ankita	Supervise On	3/10/2025 11:00:00 AM
SubDirectory	PP030725	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	Q1518-01	PP070381.D	07 Mar 2025 22:26	YP\AJ	Ok
46	Q1510-01	PP070382.D	07 Mar 2025 22:42	YP\AJ	Ok,NR
47	Q1510-02	PP070383.D	07 Mar 2025 22:58	YP\AJ	Ok,M
48	AR1660CCC500	PP070384.D	07 Mar 2025 23:58	YP\AJ	Ok
49	AR1242CCC500	PP070385.D	08 Mar 2025 00:14	YP\AJ	Ok,M
50	AR1248CCC500	PP070386.D	08 Mar 2025 00:30	YP\AJ	Ok,M
51	AR1254CCC500	PP070387.D	08 Mar 2025 00:46	YP\AJ	Ok
52	I.BLK	PP070388.D	08 Mar 2025 01:03	YP\AJ	Ok

M : Manual Integration